

Medical infrared thermal imaging of syringomyelia in the Cavalier King Charles Spaniel

Michael Larkin

Long Island Veterinary Specialists <https://orcid.org/0000-0001-6632-2764>

Catherine Loughin

Long Island Veterinary Specialists

Dominic Marino (✉ dmarino@lvs.org)

Curtis Dewey

Cornell University

Scott Umbaugh

Southern Illinois University Edwardsville

Joseph Sackman

Long Island Veterinary Specialists

Research article

Keywords: medical infrared thermal imaging, thermography, Chiari-like malformation, syringomyelia, syrinx

Posted Date: January 8th, 2020

DOI: <https://doi.org/10.21203/rs.2.20382/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Abstract

Background Medical infrared thermal imaging (MITI) is a non-invasive imaging modality gaining popularity in the veterinary field. An infrared camera captures emission of heat and creates a color map in the form of a thermogram. Topical heat emission is influenced by localized disease processes as a result of autonomic nervous system imbalance. The purpose of this study was to determine the utility of using thermography to identify changes in thermographic patterns associated with syringomyelia (SM) presence or absence in Cavalier King Charles Spaniels (CKCS) with Chiari-like Malformation (CLM).

Results In CKCS with CLM, MITI was most accurate at a texture distance of 6. Optimizing imaging feature sets produced a highest accuracy of 69.9% (95% CI: 59.5-79.0%), with 81.3% sensitivity and 57.8% specificity for identifying the presence of syringomyelia.

Conclusion Thermographic image analysis is a successful non-invasive, diagnostic test that can be used to screen for syringomyelia presence in a CKCS with CLM.

Background

Chiari-like malformation (CLM) is a neurologic condition affecting the craniocervical junction in dogs. Increasing evidence suggests that CLM in dogs is a condition in which the entire skull is malformed, although the caudal aspect is the most obvious abnormality on MRI [1]. If the abnormality is severe enough, there can be cerebellar herniation through the foramen magnum[2]. Syringomyelia (SM) is often found in association with CLM. SM is characterized by one or more fluid-containing cavities (syrinx) within the parenchyma of the spinal cord [3, 4]. Syrinx formation secondary to CLM is not fully understood but is believed to be multifactorial including structural abnormalities coupled with changes in CSF flow, blood flow and pressures [5]. SM is often thought to be the most important factor in determining symptoms of patients with CLM [6]. Dogs affected with CLM/SM can experience debilitating neuropathic discomfort and neurologic abnormalities.

The CLM/SM malformation is found most commonly in the Cavalier King Charles Spaniel (CKCS). It has been determined to be heritable and as many as 85% of CKCS are affected [7–9]. The most common clinical sign noted is pain secondary to disruption of the fibers of the dorsal horn laminae [10–12]. Other clinical signs include; scratching (also termed “air-guitar”) [12], allodynia (pain from non-noxious stimuli), hyperesthesia sensitivity to touch, facial rubbing, seizures, ataxia, paresis, exercise intolerance and lethargy [2, 3, 6].

Diagnosis of CLM/SM is supported by history, signalment, clinical signs, and diagnostic imaging findings. Imaging modalities for dogs experiencing cervical pain include: contrast myelography, radiography, computed tomography (CT), and magnetic resonance imaging (MRI). MRI is considered the modality of choice in humans & veterinary medicine. MRI has significant limitations in veterinary medicine including: expense, requirement for general anesthesia, and limited availability.

Medical infrared thermal imaging (MITI), also known as medical infrared imaging (MII) or infrared thermography (IRT), is a noninvasive imaging modality used for screening purposes in human medicine [13] and gaining popularity in veterinary medicine [14]. MITI requires use of a specific camera that captures emission of heat from the surface of objects, such as skin. Local dermal microcirculation, directly controlled by the sympathetic portion of the autonomic nervous system, is responsible for the surface heat detected [15], not heat emanating from deeper tissues. Once the thermal radiation is captured, a thermal pattern is then generated in the form of a color map called a thermogram. Warm regions are typically associated with increased circulation or metabolic rates secondary to disease processes (clinically associated with inflammation). Cold regions are typically associated with reduced perfusion suggestive of infarction or autonomic system imbalance. Therefore, thermography can be used to quantitatively detect alterations in sympathetic tone (i.e., surface heat distribution) that typically occurs secondary to injury or other disease processes [16]. MITI has been used successfully in small animal veterinary medicine to evaluate numerous conditions such as intervertebral disc disease [17], canine elbow dysplasia [18], feline hyperthyroidism [19] and cranial cruciate disease [20]. MITI has also been shown to be a useful method for estimating or even detecting cases of syringomyelia in human medicine [21]. MITI has many advantages over MRI including the lack of requirement for general anesthesia or sedation, less expensive equipment, rapid image acquisition time and decreased client expense.

The purpose of this retrospective study was to determine the utility of using MITI to identify differences in thermographic patterns associated with presence or absence of syringomyelia in CKCS affected with CLM. We hypothesized that the presence of SM secondary to CLM will be associated with a recognizable change in thermographic pattern.

Results

Demographics

One hundred dogs were included in this study. Due to poor image quality (usually motion artifact) preventing proper interpretation on some of the MITI images, seven of the dogs were eliminated totaling 93 CKCS in this study. Of the 93 CKCS included, the average age was 33.7 months old (SD = 27.4 months) and the average weight of each patient was 8.1 kg (17.8 pounds). There were 35 female intact (37.6%), 24 female spayed (25.8%), 15 male intact (16.1%) and 19 male neutered (20.4%) CKCS. Of the 93 CKCS included in this study, all (100%) had CLM based on complete, full-body MR imaging; 48 (51.6%) were diagnosed with SM and 45 (48.4%) had no evidence of SM.

Image Pattern Analysis

There were 8,184 experiments of various parameters performed. Calculated statistics, average and standard deviation for all 8,184 permutations were performed. The top ten best results for each texture distance (6, 7, 8 and 9) were documented. During the experiment, different feature sets yielded different results as expected. The most effective features for syrinx identification were found to be histogram

mean, histogram energy, texture inertia, texture energy, texture correlation, and texture entropy. Similarly, the experiments show Euclidean distance to be the best comparison metric and K-nearest neighbor as the optimal classification method. Standard normal density and softmax scaling normalization methods were both shown to be useful. Figure 1 shows thermograms for two different CKCS without (left) and with (right) syringomyelia.

The best results were achieved using a texture distance of 6 (Table 1). The highest accuracy rate among feature sets at this texture distance was approximately 70% (65 out of the 93 images correctly identified as SM present or absent; 95% CI: 59.5–79.0%). Depending on features used, pattern recognition analysis achieved a sensitivity as high as 83.3% (correctly identifying 40 of the 48 images that had syrinx presence; 95% CI: 69.8–92.5%) in one trial, and also achieved a specificity as high as 73.3% (correctly identifying 33 of the 45 images with syrinx absence; 95% CI: 58.1–85.4%) in another trial.

Discussion

MITI using pattern recognition software was successfully used to differentiate between thermographic patterns of CKCS dogs with CLM and SM vs. those without SM. We found MITI using pattern recognition software was 70% accurate in correctly identifying CKCS dogs with CLM and SM with an 81% sensitivity and 58% specificity. Additional studies are being conducted with different mask sizes and histogram features to see if better improvements in sensitivity and specificity can be made. As such, these preliminary results reaffirm the view of MITI as a screening tool rather than a diagnostic tool given the high sensitivity and relatively low specificity. By reducing the mask size, the background data or “noise” can be reduced thus concentrating the more relevant data. Larger data sets including normal are being acquired as the information will enhance the accuracy of the normal “masks” making comparisons more distinguishable.

MITI may be helpful clinically for clients who decline advanced diagnostic imaging due to financial concerns, patient anesthetic risk concerns or other reasons. MITI might provide further insight for these non-surgical candidates and aid in appropriate, goal-directed therapy.

The diagnosis of CLM can only be made by MRI, which is also the preferred imaging modality for diagnosing SM. There has been some acceptance in the veterinary community of the use of “abbreviated” MRI studies utilizing views of the head and neck region with implementation of limited imaging sequences in an effort to reduce costs. SM has been reported to be most commonly located at the C1-4 and T2-L2 spinal cord segments in a prior study [22]. MITI can be used as a screening tool to further support the necessity of full spinal imaging.

Historically, a goal of MITI has been to identify changes associated with pathology prior to clinical signs [23–25]. MITI has this potential regarding CLM/SM as well. MITI could be offered as a low-cost screening tool for breeders in young patients. Young patients can be imaged easily for the development of syringomyelia in patients without clinical signs, with continued monitoring as growth occurs. MITI might

aid in determining CKCS breed reproduction standards and sales of healthy offspring in the future. Further investigation with regards to MITI in young CKCS dogs is needed.

Treatment options for CLM/SM involve surgical or medical management. The preferred method of treating humans with CLM/SM is surgical decompression via a procedure called foramen magnum decompression (FMD) [26–28]. The authors have reported surgical success in veterinary patients as well, with a FMD combined cranioplasty procedure that was able to improve or resolve clinical signs in approximately 80% of dogs [29, 30]. Further studies are on going to evaluate for syrinx resolution in these post-operative patients. Currently, the only method of reassessing syrinx status after surgery is to repeat a full spinal MRI. MITI could potentially be used as a screening for syringomyelia resolution in post-operative FMD patients. Further development of computer recognition pattern analysis to improve both sensitivity and specificity would be necessary.

This study revealed that MITI is a successful screening test for the presence of SM in CKCS with CLM. Compared to other imaging modalities, MITI is a quick, inexpensive modality that does not require sedation nor anesthesia, and eliminates radiation exposure (to patient & staff). MITI does not provide insight to syrinx location or severity nor should it be used as a sole diagnostic modality. While veterinary thermographic imaging continues to improve as advances in technology occur, MRI will remain necessary for definitive diagnosis of Chiari-like malformation and syringomyelia.

Limitations

Limitations of this study include the moderate sample/population size, and a limited number of thermographic views obtained.

In this study, dorsal craniocervical views were chosen based on a pilot study utilizing four views (front of head, top of head, left side of head, right side of head) that revealed that the top of the head view provided the best success rate (unpublished data). Secondly, the top of the head view includes the craniocervical junction most appropriately compared to other views. Imaging of this region was important, as it has been shown that CSF obstruction at the foramen magnum causes turbulent flow that is associated with SM presence and severity [4, 31]. It was, and still is, believed that this turbulent flow at the foramen magnum accounts for changes noted in our thermographic patterns seen with syringomyelia presence. Ongoing studies are currently investigating spinal thermographic imaging (dorsal full spinal views) in an attempt to improve overall success rate for SM identification in the CKCS.

Despite all dogs having an MRI confirmed diagnosis of CLM, these dogs were not further divided into different categories based on their CLM disease. Divided categories could include CLM compression severity (mild, moderate, severe), or craniocervical compression types (medullary kinking, dorsal compression, etc.) [32].

Future studies based on larger populations, data sets, and mask sizes may help further improve pattern recognition software success.

Conclusion

We concluded that MITI was successful in detecting thermographic pattern changes associated with syringomyelia in CKCS affected with CLM. MITI should serve as a screening test for the presence or absence of SM, and would work best in conjunction with MRI. Future studies should focus on improving overall success rate, identifying syrinx location and quantifying syrinx severity. Future studies using MITI to determine its viability for tracking SM disease progression and/or resolution could be beneficial.

Methods

Case Selection

The medical records of 216 dogs diagnosed with CLM from 2007–2010 at The Canine Chiari Institute at Long Island Veterinary Specialists (Plainview, NY) were reviewed. One hundred CKCS dogs that were diagnosed with CLM were included in this study.

All dogs had a complete physical & neurologic evaluation performed by a Diplomate of the American College of Veterinary Surgeons. All dogs had full brain/spinal MRI, a craniocervical CT scan, cranial cervical radiography, and craniocervical thermal imaging during that time. These dogs were also found to have no comorbidities (systemically or discovered on imaging) and also had no prior surgical interventions for any disease process. After imaging and recovery from anesthesia all dogs were discharged to the care of their respective owners.

Thermographic Imaging

All dogs had restricted exercise and were allowed to acclimate to the room temperature (21 °C) for 30 minutes prior to MITI. Images were obtained using an infrared camera (Med 2000 IRIS, Meditherm, Inc. Beaufort, NC) with a focal plane array amorphous silicone microbolometer. Images were obtained from the top of the head (dorsal view), approximately one meter away from the patient. All dogs were imaged unsedated in sternal recumbency. To minimize thermal artifact from handling, trained technicians wearing latex gloves positioned dogs without touching the head and cervical regions. To minimize background artifact, a uniform board was placed under each patient. All images were sent to the Department of Electrical and Computer Engineering, School of Engineering, Southern Illinois University, Edwardsville, IL for evaluation. After thermographic imaging, all dogs were then placed under general anesthesia for radiographic, MRI, and CT imaging.

MRI Imaging

Magnetic resonance imaging of the brain and spinal cord of each dog were acquired by use of an Achieva 3.0-T MR imaging unit (Philips, Andover, MA). Imaging was performed in dorsal recumbency with the neck in partial flexion (between 100° and 138°) to approximate the craniocervical angle in standing CKCSs as previously described [31]. Sagittal T2-weighted, axial T2-weighted, and axial T1-weighted fluid-attenuated inversion recovery images of the brain were acquired in each dog. In addition, axial T2-

weighted images of the cervical, thoracic and lumbar regions were acquired in each dog. Presence or absence of syringomyelia was confirmed with MRI.

CT Imaging

Computed tomographic images of the head and cranial vertebral column including C3 were acquired by use of a multidetector CT (Marconi Mx8000, Marconi Medical Systems Inc, Cleveland, OH) operating at 140 kV and 150 mA; a bone reconstruction filter was used during image acquisition. Dogs were placed in sternal recumbency with the neck in partial flexion (between 100° and 138°). Continuous 1 mm collimated images were obtained by use of helical acquisition.

Thermographic Software

Images were evaluated at the Computer Vision and Image Processing Laboratory at Southern Illinois University. A software program (WINTES2, Compix Inc., Lake Oswego, OR) was used to save, analyze and review image data. The different color maps were converted and remapped to 256-level gray scale images by linearly mapping the temperature range of 19⁰C to 40⁰C, as this temperature range covers the entire range of the body temperature of canines. Custom processing software (CVIPtools) was used for image pattern recognition and to analyze and evaluate thermographic patterns.

Imaging Pattern Analysis

Thermographic images were independently analyzed with image processing software and computer recognition pattern analysis (CRPA) to differentiate between the presence or absence of syringomyelia. The primary tools used include Analysis Features and Analysis Pattern Classification. Data normalization methods included soft-max with R = 1 and standard normal density normalization. Image features that were assessed included histogram features (mean, SD, skew, energy and entropy) and texture features (energy, inertia, correlation, inverse difference, entropy, and texture distances of 6, 7, 8 and 9). Two classification methods (K-Nearest Neighbor with K = 3 and Nearest Neighbor) were also assessed. Evaluation for optimal image, feature and classification method were based on accuracy, sensitivity, and specificity compared to MRI results, as calculated with the leave-one-out cross-validation method. Confidence intervals for proportions were calculated based on the binomial probability distribution.

Abbreviations

CKCS

Cavalier King Charles Spaniels

CLM

Chiari-like malformation

CSF

cerebral spinal fluid

CT

computed tomography

FMD
foramen magnum decompression
MITI
medical infrared thermal imaging
MRI
magnetic resonance imaging
SM
syringomyelia

Declarations

Ethics approval and consent to participate

Ethical approval was obtained from the Ethics Review Committee of Long Island Veterinary Specialists. Written consent to participate in this study was obtained from the owners of these privately owned dogs.

Consent for publication

Not applicable

Availability of data and materials

The datasets used and/or analyzed during the current study are available from the corresponding author upon reasonable request.

Competing interests

The authors declare that they have no competing interests

Funding

Not applicable

Authors contributions

ML: study design, data acquisition and analysis, drafted manuscript

CL, DM, CD: study design, data acquisition and analysis, reviewed and revised the manuscript

JS: data acquisition

SU: thermographic pattern analysis

All authors read and approved the final manuscript and have agreed to be accountable for the author's contribution.

Acknowledgements

Grateful acknowledgement of Dr. George Moore DVM, PhD for his assistance with reviewing the manuscript drafts.

References

1.
Knowler SP, Galea GL, Rusbridge C. Morphogenesis of Canine Chiari Malformation and Secondary Syringomyelia: Disorders of Cerebrospinal Fluid Circulation. *Front Vet Sci*. 2018;5:171.
2.
Loughin CA. Chiari-like Malformation. *Vet Clin North Am Small Anim Pract*. 2016;46:231–42.
3.
Rusbridge C, Greitz D, Iskandar BJ. Syringomyelia: current concepts in pathogenesis, diagnosis, and treatment. *J Vet Intern Med*. 2006;20:469–79.
4.
March PA, Abramson CJ, Smith M, et al. CSF flow abnormalities in caudal occipital malformation syndrome. *J Vet Intern Med*. 2005;19:418–9. (abstr).
5.
Milhorat TH, Capocelli AL Jr, Kotzen RM, et al. Intramedullary pressure in syringomyelia: clinical and pathophysiological correlates of syrinx distension. *Neurosurgery*. 1997;41:1102–10.
6.
Rusbridge C, MacSweeney JE, Davies JV, et al. Syringohydromyelia in Cavalier King Charles Spaniels. *J Am Anim Hosp Assoc*. 2000;36:34–41.
7.
Lewis T, Rusbridge C, Knowler P, et al. Heritability of syringomyelia in Cavalier King Charles spaniels. *Vet J*. 2010;183:345–7.
8.
Rusbridge C, Knowler P, Rouleau GA, et al. Inherited occipital hypoplasia/ syringomyelia in the Cavalier King Charles spaniel: experiences in setting up a worldwide DNA collection. *J Hered*. 2005;96:745–9.
- 9.

- Rusbridge C, Knowler SP. Hereditary aspect of occipital bone hypoplasia and syringomyelia (Chiari type 1 malformation) in cavalier King Charles spaniels. *Vet Rec.* 2003;153:107–12.
- 10.
- Churcher RK, Child G. Chiari 1/syringomyelia complex in a King Charles Spaniel. *Aust Vet J.* 2000;78:92–5.
- 11.
- Hu HZ, Rusbridge C, Constantino-Cases F, et al. Histopathological investigation of syringomyelia in the Cavalier King Charles spaniel. *J Comp Pathol.* 2012;146:192–201.
- 12.
- Rusbridge C, Carruthers H, Dube M, et al. Syringomyelia in cavalier King Charles spaniels: the relationship between syrinx dimensions and pain. *J Small Anim Pract.* 2007;48:432–6.
- 13.
- Feig SA, Shaber GS, Schwartz GF, et al. Thermography, mammography, and clinical examination in breast cancer screening. *Radiology.* 1977;122:123–7.
- 14.
- Rekant SI, Lyons MA, Pacheco JM, et al. Veterinary applications of infrared thermography. *Am J Vet Res.* 2016;77:98–107.
- 15.
- Leandro P. Meditherm manual of clinical thermology. Beaufort: Meditherm Inc.; 2004.
- 16.
- Fernández-Cuevas I, Marins JCB, Lastras JA, et al. Classification of factors influencing the use of infrared thermography in humans: A review. *Infrared Phys Technol.* 2015;71:28–55.
- 17.
- Grossbard BP, Loughin CA, Marino DJ, et al. Medical infrared imaging (thermography) of type I thoracolumbar disk disease in chondrodystrophic dogs. *Vet Surg.* 2014;43:869–76.
- 18.
- McGowan L, Loughin CA, Marino DJ, et al. Medical infrared imaging of normal and dysplastic elbows in dogs. *Vet Surg.* 2015;44:874–82.
- 19.
- Waddell RE, Marino DJ, Loughin CA, et al. Medical infrared thermal imaging of cats with hyperthyroidism. *Am J Vet Res.* 2015;76:53–9.
- 20.
- Infernuso T, Loughin CA, Marino DJ, et al. Thermal imaging of normal and cranial cruciate ligament-deficient stifles in dogs. *Vet Surg.* 2010;39:410–7.
- 21.
- Koyanagi I, Iwasaki Y, Isu T. et al. Thermographic findings of syringomyelia. *No Shinkei Geka.* 1988;16:1149–54.
- 22.
- Loderstedt S, Benigni L, Chandler K, et al. Distribution of syringomyelia along the entire spinal cord in clinically affected Cavalier King Charles Spaniels. *Vet J.* 2011;190:359–63.

23. Farrell C, Wallace JD, Edeiken J. Thermography and osteosarcoma. *Radiology*. 1968;90:792–3.
24. Clark JA, Cena K. The potential of infra-red thermography in veterinary diagnosis. *Vet Rec*. 1977;100:402–4.
25. Schaefer AL, Cook N, Tessaro SV, et al. Early detection and prediction of infection using infrared thermography. *Can J Anim Sci*. 2004;84:73–80.
26. Strayer A. Chiari I malformation: clinical presentation and management. *J Neurosci Nurs*. 2001;33:90–104.
27. Mariana C, Cislighi MG, Barbieri S, et al. The natural history and results of surgery in 50 cases of syringomyelia. *J Neurol*. 1991;238:433–8.
28. Tubbs RS, McGirt MJ, Oakes WJ. Surgical experience in 130 pediatric patients with Chiari I malformations. *J Neurosurg*. 2003;99:291–6.
29. Dewey CW, Berg JM, Barone G, et al. Foramen magnum decompression for treatment of caudal occipital malformation syndrome in dogs. *J Am Vet Med Assoc*. 2005;227:1270–5, 1250–1271.
30. Dewey CW, Marino DJ, Bailey KS, et al. Foramen magnum decompression with cranioplasty for treatment of caudal occipital malformation syndrome in dogs. *Vet Surg*. 2007;36:406–15.
31. Cerda-Gonzalez S, Olby NJ, Broadstone R, et al. Characteristics of cerebrospinal fluid flow in Cavalier King Charles Spaniels analyzed using phase velocity cine magnetic resonance imaging. *Vet Radiol Ultrasound*. 2009;50:467–76.
32. Marino DJ, Loughin CA, Dewey CW, et al. Morphometric features of the craniocervical junction region in dogs with suspected Chiari-like malformation determined by combined use of magnetic resonance imaging and computed tomography. *Am J Vet Res*. 2012;73:105–11.

Tables

Table 1

Thermographic image pattern recognition: the top ten best results from a texture distance of six.

Feature Set*	# correct (out of 93)	% correct (out of 93) - accuracy	Syrinx presence - # correct (out of 48)	Syrinx presence - % sensitivity	Syrinx absence - # correct (out of 45)	Syrinx absence - % specificity
A	65	69.89%	39	81.25%	26	57.78%
B	64	68.82%	40	83.33%	24	53.33%
C	64	68.82%	33	68.75%	31	68.89%
D	64	68.82%	32	66.67%	32	71.11%
E	64	68.82%	31	64.58%	33	73.33%
F	63	67.74%	39	81.25%	24	53.33%
G	63	67.74%	36	75.00%	27	60.00%
H	63	67.74%	35	72.92%	28	62.22%
I	63	67.74%	34	70.83%	29	64.44%
J	63	67.74%	33	68.75%	30	66.67%

- Feature sets included one or more of the following features: histogram mean, histogram energy, texture inertia, texture energy, texture correlation, and texture entropy.

Figures




Figure 1

Thermogram of a CKCS (dorsal head view). The image on the left represents a CKCS without syringomyelia. The image on the right represents a CKCS with syringomyelia.