

Untuned broadband spiral micro-coils achieve sensitive multi-nuclear NMR TX/RX from microfluidic samples

Hossein Davoodi

Karlsruhe Institute of Technology

Nurdiana Nordin

Karlsruhe Institute of Technology

Hirokazu Munakata

Tokyo Metropolitan University

Jan Korvink

Karlsruhe Institute of Technology

Neil MacKinnon

Karlsruhe Institute of Technology

Vlad Badilita (✉ vlad.badilita@kit.edu)

Karlsruhe Institute of Technology

Research Article

Keywords: Untuned broadband spiral micro-coils, sensitive multi-nuclear NMR TX/RX

Posted Date: November 24th, 2020

DOI: <https://doi.org/10.21203/rs.3.rs-112538/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published on April 8th, 2021. See the published version at <https://doi.org/10.1038/s41598-021-87247-2>.

Untuned broadband spiral micro-coils achieve sensitive multi-nuclear NMR TX/RX from microfluidic samples

Hossein Davoodi¹, Nurdiana Nordin^{1,2}, Hirokazu Munakata³, Jan G. Korvink¹, Neil MacKinnon^{1,*}, and Vlad Badilita^{1,*}

¹Institute of Microstructure Technology (IMT), Karlsruhe Institute of Technology (KIT), Eggenstein-Leopoldshafen 76344, Germany

²Department of Chemistry, Faculty of Science, University of Malaya, 50603 Kuala Lumpur, Malaysia

³Department of Applied Chemistry for Environment, Graduate School of Urban Environmental Sciences, Tokyo Metropolitan University, Tokyo 192-0397, Japan

*corresponding. neil.mackinnon@kit.edu, vlad.badilita@kit.edu

ABSTRACT

The low frequency plateau in the frequency response of an untuned micro-resonator permits broadband radio-frequency reception, albeit at the expense of optimal signal-to-noise ratio for a particular nucleus. In this contribution we determine a useful figure-of-merit for broadband microcoils, and thereby explore the parametric design space towards acceptable simultaneous excitation and reception of a microfluidic sample over a wide frequency band ranging from ¹³C to ¹H, i.e., from 125–500 MHz in an 11.74 T magnet. The detector achieves 37 % of the performance of a comparably sized, tuned & matched resonator, and a linewidth of 16.6 ppb using standard magnet shims. The use of broadband detectors circumvents numerous difficulties introduced by multi-tuned RF detector circuits, including sample loading effects in matching, channel isolation, and field distortion.

Introduction

Nuclear magnetic resonance (NMR) is a powerful and versatile analytical tool due to its capability to address and analyse with exquisite chemical specificity each isotope exhibiting a unique magnetic moment. Consequently, NMR methods can be applied to almost all elements in the periodic table. Building upon this feature, NMR is able to offer correlated information either within the chemical shift range of the same isotope (homonuclear) or from different species (heteronuclear), therefore being capable to deliver information related to the structure and dynamics of a molecular system^{1,2} with atomic resolution. This unique characteristic allows the simultaneous detection of multiple nuclei either by parallel^{3,4} or sequential⁵ acquisition, performing decoupling experiments^{6,7}, or employing indirect detection methods, such as polarization-transfer techniques^{8,9}, to study low gyromagnetic, i.e. less sensitive nuclei.

A multitude of NMR detection hardware variants have been developed in order to meet the needs for multi-nuclear NMR detection, either as multi-resonant or as broadband topologies¹⁰. Multi-resonant circuits provide tuning and matching at different frequencies utilizing additional capacitors and inductors^{11–20}. Despite their popularity, these circuits suffer from low power transfer efficiency compared to their single resonant counterparts. This is attributed to the additional current paths introduced by the extra elements^{11–15}, or to the infinite quality factor of those elements^{16–20}. In the receive mode, this issue translates to higher noise level and/or lower signal amplitude, therefore lower signal-to-noise ratio (SNR). Moreover, the resonant frequencies of these circuits are sensitive to the environment and sample loading, meaning that a small perturbation in the environment of the coil can *un*-match the coil, additional effort being required in order to *re*-match it. Modifying these circuits in order to capture a different resonant frequency is a cumbersome task, which ends up either in reconstructing the circuit or replacing the circuit elements.

Another approach for multi-nuclear NMR detection is a multi-coil topology, where each coil is tuned and matched at one or multiple frequencies^{21–27}. The additional hardware and the cross-talk between the coils are the main issues to be addressed when employing this topology. The challenge encountered when using a multi-coil structure for multi-nuclear detection originates from the overlap of different fields-of-view (FOV) of the participating coils. The efficiency of coherence transfer or decoupling methods is directly influenced by the FOV overlap. At the same time, FOV overlap among different coils is synonymous with a cross-talk among those coils and this causes power leakage. The leaked power reduces the efficiency and limits the effective dynamic range of the low noise amplifier (LNA). Although all coils in a multi-coil structure should share

the same FOV, their cross-talk should be minimised. In order to compensate for the isolation issue, extra trap elements are required to decouple the coils. Due to their non-ideal nature, trap circuits introduce noise and power loss, thus degrading the performance of the system as a whole. The necessity of their field orthogonality to the external magnetic field introduces further challenges to the experimental set-up. Besides, the outermost coil, being expanded as a result of additional coils, suffers from low filling factor.

NMR miniaturisation is a well-recognised technique to enhance sensitivity for mass-limited samples by employing miniaturised coils. However, this approach poses a new challenge in multi-resonant circuits and multi-coil configurations. The relatively low impedance of micro-coils degrades both the power delivered to the coil in excitation mode, and power collected from the coil in receive mode. One solution for power transfer efficiency constraint is to make the peripheral circuit more compact, i.e., shortening the tracks; however, with the price of introducing static magnetic field distortions. The other solution is to use inductive coupling to another coil in order to supply power to the detecting micro-coil. In this case, magnetic field distortions can be eliminated by keeping the supply coil and its circuit away from the FOV of the detecting micro-coil, nevertheless the power transfer efficiency is lower than in the case of wired connection. Several studies have tackled the issue of employing micro-coils for multi-nuclear NMR applications, maintaining both spectral resolution and SNR^{23,28–30}.

While the above-mentioned topologies rely on resonating detection schemes, several alternative approaches have considered the challenge of multi-nuclear NMR detection in a continuous, wide bandwidth^{31–36}. These coils offer a detection bandwidth comparable to the central frequency of the detection band (e.g. measuring ¹H and ¹³C at 11.7 T, central frequency is 315 MHz and bandwidth is 375 MHz). A widely studied subcategory of wideband detectors are delay and transmission lines with lumped and distributed elements, respectively^{31–34}. In this case, the detector is a two-port element, in a form of π - or T-network of inductors and capacitors, and is terminated to its characteristic impedance at both ports. The inductive elements in the network are exploited as Faraday induction NMR detection elements. A secondary subcategory is the one-port detector^{35,36}, i.e., coils being terminated to the ground at one end, as opposed to two-port element delay/transmission line coils. This coil does not require any external tuning matching elements.

In an important paper on the topic of broadband NMR detectors, Fratila *et al.*³⁶ have reported a planar spiral micro-coil that operates in a wide frequency range between 61 MHz and 400 MHz, being able to observe most NMR-active nuclei at 9.4 T with remarkable mass sensitivity. In spite of the exquisite performance of this detector, the literature is still lacking an in-depth analysis of the operation principles of broadband planar spiral coils. We conjecture that the planar spiral topology holds several intrinsic characteristics that makes it interesting for a wider use for broadband NMR detection.

As will be comprehensively discussed in the following sections, the distribution of the windings of the spiral coil with respect to the sample volume is a key feature enabling effective broadband performance. In a spiral coil, magnetic fields at different locations of the sample region are uniformly vulnerable to all turns of the coil, in comparison to a solenoidal coil where the effect of the nearby turns is dominant and different parts of the sample are subject to the influence of different segments of the coil.

In general, spiral coils offer a lower quality factor comparing to solenoids because of their contact with a lossy substrate^{37,38}. The substrate loss is a prominent issue, especially in standard NMR applications, as the SNR of the received signal can be drastically influenced by the quality factor of the coil. The literature contains several approaches to eliminate this issue, e.g., using substrates with higher dielectric constants³⁹ or using suspended coils⁴⁰. Nonetheless, in broadband applications, a low quality factor is a *sine qua non* condition, which means that a trade-off between quality factor and detection bandwidth must be considered for an optimum broadband coil. Moreover, the presence of distributed inter-turn capacitors and inductors whose impedance decreases as the radii of the turns decreases, is another feature of the spiral coil which influences the performance of the coil and makes it suitable for broadband applications.

Another advantage of spiral coils, which stems directly from their planar geometry, is the introduction of less B_0 distortions compared to the solenoidal coil topologies, since the static magnetic flux, which is parallel to the coil surface, experiences fewer material interfaces. B_0 inhomogeneity is especially important in micro-detectors where the sample volume is small^{41,42} and consequently the signal amplitude is low, and spectral broadening further lowers already small peak heights resulting into the loss of spectral information. Nevertheless, the penalty for the superior B_0 homogeneity offered by spiral coils is lower filling factor and poor spatial RF field homogeneity compared to solenoidal coils. Moreover, spiral coils are compatible with standard PCB and microfabrication techniques, as well as microfluidic integration.

This paper bridges the gap in understanding and adopting the broadband spiral coils in NMR applications by providing a thorough fundamental study of working principles accompanied by characterisation results. The discussion also covers the analysis of the parameters to be tuned by the NMR engineer in order to optimise the overall performance of a broadband spiral coil detector and presents the limitations related to the rest of the existing signal chain of the commercial NMR system influencing the broadband performance.

The broadband detector introduced and analysed in this work consists of a miniaturised spiral coil, an RF connection port, and a microfluidic channel. The study starts with introducing the figures of merit for an NMR detector and continues




Figure 1. A vertical-bore superconductive magnet together with the probe, LNA block, and electronic racks for signal processing. The detector, being mounted on the NMR probe is directly connected to the LNA block using a coaxial cable. The sample is loaded in the microfluidic channel of the detector.

with simulating and optimising the performance of the broadband detector considering these figures of merit at different frequencies. These are compared and discussed with reference to the figures of merit for tuned/matched detectors as an evaluation of the broadband detector. The theoretical discussion and optimisation in the two different regimes (broadband and tuned/matched) are concluded by presenting the optimum coil design in each regime and calculating the penalty to be paid for a broadband detector. We also report on the fabrication procedure for the broadband detector using standard micro-fabrication techniques. The experimental evaluation of the broadband detector is presented using: i) NMR experiments at seven different frequencies corresponding to ^1H (500 MHz), ^{19}F (471 MHz), ^{31}P (202 MHz), ^7Li (194 MHz), ^{11}B (160 MHz), ^{23}Na (132 MHz), and ^{13}C (126 MHz) at 11.74 T to verify the broadband performance; ii) ^1H MRI experiment on a water phantom to evaluate the RF field distribution; iii) impedance measurements (given in [Supplementary Information](#)). Normalised limit-of-detection (nLOD) is reported for the broadband detector at different frequencies and compared with the state of the art. In addition, the discussion ends with listing the challenges associated with the broadband coils and suggestions are given for further performance enhancement.

Results

Figure 1 shows the NMR setup consisting of a superconducting magnet, pre-amplifier block (also known as low noise amplifier, LNA), and further spectrometer electronics (excitation source, signal processing chain, etc). A probe places the detector and (in this case) the sample inside the magnet. A coaxial cable connects the coil to the LNA block. Although the coil is not matched to $50\ \Omega$, both the characteristic impedance of the coaxial cable and the input impedance of the LNA stage are $50\ \Omega$, similar to standard NMR systems.

Borrowing from the classical description of NMR, once the sample is placed inside the static magnetic field B_0 , a bulk magnetisation is created along the direction of the magnetic field. By applying a radio-frequency (RF) magnetic field (B_1) for time τ , the magnetisation is rotated to an angle of $\theta = \gamma \times B_1 \times \tau$, where γ is the gyromagnetic ratio of the isotope to be studied. Once the excitation B_1 is silenced, the magnetisation returns to equilibrium, precessing around the static field axis at the Larmor frequency, $\omega_L = \gamma \times B_0$. The precessing magnetic moments induce a voltage in the NMR coil. According to the principle of reciprocity⁴³, the voltage induced into the coil by an elementary magnetisation depends on the magnetisation itself and on the RF field generated by the coil at that specific magnetisation location for a unit current applied to the coil ($S(x, y, z) = \partial(\frac{B_1(x, y, z)}{i}) \cdot M_0(x, y, z) / (\partial t)$). The total voltage being induced in the coil is the integral of these elementary voltages over the sample volume ($v_{\text{FID}} = \int_{\text{sample volume}} S(x, y, z) dv$). The noise spectral density of the micro-coils is usually the thermal noise of the coil expressed by $N_{\text{coil}} = \sqrt{4 \times K_B \times R_{\text{coil}} \times T}$.

NMR figures of merit

For a given sample and B_0 , the excitation efficiency and SNR of the system are the figures of merit to be considered, as detailed in Table 1. The excitation efficiency determines the average B_1 field generated by the coil at the sample region per square root

Figures of merit	tuned-matched	broadband
excitation efficiency (ef)	$\gamma \times B_1 / i \times \frac{1}{\sqrt{2 \times R_{\text{coil}}}}$	$\gamma \times B_1 / i \times \frac{\sqrt{2 \times Z_0}}{ Z_{\text{coil}} + Z_0 }$
SNR	$\frac{V_{\text{FID}}}{\sqrt{F \times N_{\text{coil}}}}$	$\frac{V_{\text{FID}} \times \frac{Z_0}{ Z_0 + Z_{\text{coil}} }}{\sqrt{N_{\text{coil}}^2 \times \frac{Z_0^2}{ Z_0 + Z_{\text{coil}} ^2} + N_{Z_0}^2 \times \frac{F-1}{4}}}$

Table 1. NMR figures of merit for two different working regimes, applicable to a conventional tuned and matched coil, and a broadband coil. In all four cases, the figure of merit should be maximised.

of applied power to its terminal. SNR is a more complex parameter containing several terms, one part of which are predefined at the coil design level, i.e, the properties of the sample and nuclei being investigated ($\gamma^3 \times N \times I \times (I + 1) \times \hbar^2 / (6 \times k_B^{3/2})$) the other depending on the static magnetic field and the sample volume ($B_0^2 \times V_s / (T_{\text{sample}} \times T_{\text{coil}}^{1/2})$).

Excitation efficiency

The excitation efficiency is defined as the frequency of the nutation signal collected for 1 W excitation power. The excitation source has a predefined output impedance, usually 50 Ω . Therefore, optimum power transfer can be achieved by *matching* the coil impedance to the impedance of the excitation source. Deviations from this matching will decrease excitation efficiency.

In a conventional probe, the coil is connected to the coaxial cable through a passive tuning and matching network. In its simplest form, this circuit consists of two capacitors. At the resonant frequency, the impedance of the tuning/matching circuit together with the coil, seen from the coaxial cable port, equals the characteristic impedance of the cable (Z_0). Therefore, the maximum power transfer occurs at this specific frequency. Proper matching conditions being fulfilled and at the resonant frequency, half of the power is consumed in the output impedance of the excitation source and half is dissipated in the coil and the tuning/matching circuit, neglecting the dissipation in the cable. For the frequencies other than the resonant frequency, the power transfer efficiency drops because of impedance mismatch. As the quality factor of the capacitors is usually high (> 1000), almost all power dissipation in the circuit happens in the coil (typically with $Q < 100$). Therefore, the excitation efficiency at the resonant frequency can be calculated as shown in Table 1.

For a broadband coil, wideband performance is achieved by removing tuning/matching. As the result of mismatch, the amplitude of the current flowing in the coil for a certain applied power depends on the impedance of the coil (Z_0), and hence the excitation efficiency is different from the tuned/matched case as indicated in Table 1.

SNR

In the reception mode, an ideal tuning/matching network transfers both the voltage induced into the coil and the noise spectral density of the coil by an amplification factor of $\sqrt{Z_0 / R_{\text{coil}}}$ at the resonant frequency and maintains SNR constant. Another amplification step at the LNA brings the signal to a sufficient level for further processing. During the amplification of the input signal, extra noise is introduced by the LNA. This extra noise is usually defined by the noise factor (F) of the LNA. The SNR at the resonant frequency and at the output of the LNA is also given in Table 1. According to the Friis formula for cascaded noise factor⁴⁴, the noise introduced by the LNA is dominant and the additional noise generated by successive processing steps can be ignored.

As discussed by Massin *et al.*³⁷, at the coil level design and for a given sample and certain B_0 , the intrinsic SNR can be enhanced by increasing the RF magnetic field amplitude at the sample region per square root of the resistance of the coil through proper selection and design of the coil geometry based on the sample.

For a broadband coil in receive mode, the intrinsic SNR of the coil is the same as in the tuned/matched case. Considering the resonant frequencies of the transmission line far from the working frequencies and neglecting the power dissipation in the transmission lines, the SNR at the input of the LNA is also the same as the intrinsic SNR of the coil, as long as the cable is terminated to its characteristic impedance at the LNA side. Assuming the same LNA in both tuned/matched and broadband cases, the output SNR can be derived according to Table 1, where N_{Z_0} represents the noise corresponding to the characteristic impedance (Z_0).

An intrinsic characteristic of a coil operating in a broadband regime is the unavoidable degradation of the overall performance of the system compared to tuned/matched operation, due to the mismatch between the coil and the other excitation/detection circuits. In the excitation mode, lack of tuning/matching results in a longer excitation pulse for a given applied power or, equivalently, higher excitation power for a given excitation pulse length. On the other hand, since the N_{Z_0} is usually higher than the thermal noise of the coil, lack of the tuning/matching affects the overall detection performance. These drawbacks can be compensated by increased signal accumulation. Therefore, in a broadband coil, a compromise in sensitivity is exchanged for an increased flexibility to operate over multiple frequencies. The challenge of the coil designer is to optimise the performance as a function of application and hence the required bandwidth.

Optimisation of detector geometry

The spiral coil was defined on the backside of a 100 μm -thick glass substrate and the coil material was gold. The design parameters defining the coil are: inner radius (r_{in}), number of turns (n), the width of the tracks (w), the spacing between the tracks (s), and the thickness of the tracks (h). The inner turn of the coil was connected to the corresponding port using a wirebonded bridge. The wirebond requires a 100 μm pad at the centre of the coil, thus the inner radius of the coil was set to 125 μm as a good compromise between a small sample volume and high RF field. The values for s and w were optimised based on the analytical formula discussed in the [Supplementary Information](#). For this case, $w + s = 40 \mu\text{m}$ results in the highest B_1 field at the sample region. In order to use the cross-section of the tracks efficiently, the thickness and the width of the tracks should be more than twice the skin depth at the working frequencies. The skin depth of gold at 100 MHz is about 7.5 μm , thus w , h , and s were selected to be 20 μm .

Simulations and performance analysis were carried out at three different frequencies. These frequencies were selected based on the Larmor frequency values of the most popular isotopes, listed in a recent publication of our group¹⁰, i.e., 500 MHz, 200 MHz, and 150 MHz. These simulations were performed assuming the sample volume as a cylinder with 500 μm diameter and 500 μm height, positioned co-axially with respect to the coil.

Figure 2a depicts the profile of the RF magnetic field for a unit current (B_1/i) applied to the ports of the spiral coil. The results are plotted at 500 MHz; however, the relative values are very similar at the frequencies tested. By increasing the number of turns, B_1 field homogeneity within the sample volume (blue square) is enhanced, while B_1 field intensity is maximum for a coil with 25 turns. Simulated physical parameters of the coil (impedance, quality factor, and average RF magnetic field per unit current) as a function of the number of turns at different frequencies are plotted in Figure 2b. It is important to note that a coil with 29 turns has a self-resonance at about 500 MHz. For values of n smaller than 29, the inductance and the resistance of the coil increase with the number of turns. However, the inductance increases faster, which leads to an overall increase of the coil quality factor with the number of turns. The average magnetic field at the sample region also increases with the number of turns until n reaches the self-resonance value. Then, this value drops dramatically because of the inter-turn capacitive links bypassing the inductive links.

The simulation results were further processed in order to extract the nutation signals and to gain a better understanding of the performance of different designs at different frequencies. The nutation signals for different numbers of turns, plotted in Figure 2c, reveal an optimum number of turns in which the excitation efficiency is maximum. Unlike the nutation frequency, which depends on the excitation efficiency, the decay rate of the nutation signal corresponds to the RF field inhomogeneities and is independent of power transfer between the coil and electronics. These nutation signals also confirm RF field homogeneity enhancement with the number of turns.

The figures of merit were also calculated based on the simulation results and for two different scenarios, i.e., tuned/matched case and broadband performance, according to the formulae presented in Table 1. The results, presented in Figure 2d, are normalised with respect to the maximum value for each configuration for better comparison. The results clearly show the existence of an optimum number of turns for which excitation efficiency and SNR are maximised. Further quantitative comparisons of the optimum geometries are available in Table 2.

To calculate the current, the noise figures of the LNAs of ^1H channel and X-nuclei channel were assumed 1 dB and 2 dB, respectively, to be consistent with the commercially available LNAs introduced in the Method Section. The optimum number of turns are different for the two different operation modes (tuned/matched and broadband) and for each operating frequency. These optimum values decrease with increasing frequency. Considering these results, the NMR engineer has to decide on a certain penalty to be paid when defining one single coil for the entire frequency range. A compromise in the optimum number of windings was determined by averaging the calculated optimal number of winding values over both the excitation efficiency and SNR figures of merit, at the evaluated frequencies for each configuration (3 opt. n values for each figure of merit, Table 2). This resulted in 17 windings for the tuned/matched case and 8 for broadband case. A detailed evaluation of the optimal coil for each configuration is given in Table S1, and it is noted that the maximum cost to be paid for selecting a certain number of turns is 6 % in terms of excitation efficiency and 14 % in terms of SNR.

Table 2 summarises the cost for the broadband performance with respect to the tuned/matched case. It is important to note that, although the excitation efficiency suffers considerably from mismatch, it can be easily compensated by employing a longer pulse ($\sim 3 - 5\times$) or a higher applied power ($\sim 9 - 20\times$). Accordingly, the SNR of the broadband detector is 32–37 % of its tuned/matched counterpart. This implies that the total cost to be paid for a broadband detection is $\sim 10\times$ more averaging in comparison to an optimised single resonant detector. Nevertheless, for a multi-nuclear application, the SNR degradation, stemming from tank/trap circuits, should be included in order to have a fair comparison (not considered in this work). One solution to enhance the overall SNR is to use LNAs with better noise performance. For instance, a noise figure of 0.5 dB enhances the SNR up to $\sim 50\%$.

Another important feature of an NMR spectrum is its spectral resolution. Apart from the sample properties, i.e., transverse relaxation time T_2 , spectral resolution depends on the B_0 distribution at the detection zone and is independent of tuning




Figure 2. a) Simulation results using COMSOL. a) B_1/i at 500 MHz (corresponding to ^1H Larmor frequency at 11.74 T). The dashed rectangle represents the sample volume and the line plot, corresponding to the upper x-axis, shows the field at the axis of the coil; b) resistance, inductance, quality factor, and average B_1/i at three different frequencies and for different number of turns; c) Nutation curves for the predefined sample volume at three different frequencies and for different number of turns; d) Normalised figures of merit of the spiral coils for tuned/matched case and broadband performance; e) B_0 distribution of the coil for three different sample geometries and at three different distances from the coil together with their corresponding NMR spectra at 500 MHz using the optimised coil.

and/or matching. The NMR detector designed based on the optimised broadband coil geometry was further investigated using susceptibility simulations, aiming for an optimised microfluidic channel geometry to minimise B_0 perturbations at the coil field of view (FOV). NMR spectra were constructed summing all the signals collected from elementary samples (extracted from B_1 simulations) and considering their relative susceptibility (extracted from B_1 simulations). In order to minimise the interference of the connector and metallic structures with B_0 distribution in the sample region, both the coil and the channel have been shifted towards one edge of the chip.

According to B_0 field patterns, given in Figure 2e, a sample volume with a circular footprint (500 μm diameter) supplied with channels of 100 μm width structured in glass results in severe distortions. Replacing the glass with an arbitrary susceptibility-matched medium enhances the spectral resolution and, therefore, the quality of the spectrum. Residual imperfections are caused by the metallic structures, i.e., the coil and its corresponding tracks. To remove material boundaries near to the detection volume, the channel width was increased to 500 μm with the cost of increased dead sample volume. In this case, the simulation shows a spectrum which is similar to the ideal situation when the glass substrate has been replaced with a susceptibility matched material. The amplitude of the signal in the rectangular channel is slightly higher since a larger amount of sample is contributing to the signal.

Experimental analysis

NMR spectroscopy

The optimised detector, being fabricated according to the procedure given in Methods section, was tested to detect seven different nuclei. For this purpose, two different samples (aqueous and battery electrolyte material) have been investigated. The aqueous sample containing ^1H , ^{31}P , ^{11}B , ^{23}Na , and ^{13}C isotopes is a good representative for the biological samples, whereas the battery electrolyte sample demonstrates a different class of NMR applications. 1-ethyl-3-methylimidazolium bis(trifluoromethanesulfonyl)imide (EMI-TFSI) is one of the major ionic liquids having attracted attention as novel solvents due to their high thermal stability, non-volatility and extremely low vapor pressure. These characteristics offer advantages as

Frequency (MHz)	excitation efficiency			SNR		
	opt. n_{TM}	opt. n_{BB}	$\frac{e_{f_{BB}(n=8)}}{e_{f_{TM}(n=17)}} (\%)$	opt. n_{TM}	opt. n_{BB}	$\frac{SNR_{BB}(n=8)}{SNR_{TM}(n=17)} (\%)$
150	12	9	32	25	11	37
200	11	7	31	25	9	34
500	11	5	22	19	7	32

Table 2. The optimum number of turns in different tuning/matching situations and for both figures of merit. Comparing the relative figures of merit for optimised geometries highlights the total cost to be paid in order to achieve broadband performance. For example, the expected SNR of the eight turn broadband coil is 37 % of the SNR from the 17 turn coil tuned to 150 MHz and matched to 50 Ω . Comparisons amongst either tuned/matched or broadband configurations are provided in Table S1.

green solvents (recycling ability and easy product recovery) over traditional organic solvents. Furthermore, their viscosity, hydrophobicity, dielectric constant and other physicochemical properties are tunable by changing the combination of anionic and cationic constituents. Therefore, the use of ionic liquids is spreading into various fields of applications such as organic synthesis and extraction. In electrochemical energy conversion, ionic liquids are in the spotlight as non-flammable electrolytes to provide safer lithium-ion batteries, in which Li salt is added into the ionic liquids as Li^+ -ion conductor, and lithium bis(trifluoromethane-sulfonyl)imide (LiTFSI) is a typical example for the EMI-TFSI class.⁴⁵⁻⁴⁷ Experiments on the battery electrolyte sample were performed at frequencies corresponding to 1H , ^{19}F , and 7Li isotopes, the results being presented in Figure 3.

NMR results on the aqueous sample clearly demonstrate the advantage of employing a broadband coil, since no adjustment to the detector is made during the experiments. Considering the constraint of multi-resonant micro-coils, e.g., the efficiency of the trap/tank circuits and the geometrical restrictions imposed by the limited space inside the magnet bore, it is only the broadband coil approach that allows the detection of five or more nuclei, without interrupting the measurements for detector modification.

The 1H spectral linewidth, measured for the TSP signal (0 ppm), is 8.3 Hz. Although this value is generally accepted for the




Figure 3. NMR experiment using the optimised NMR detector to observe different nuclei: a) 1H NMR spectrum (at 500 MHz) of the aqueous sample. inset is the same spectrum with enlarged amplitude. b) 1H nutation signal (simulated nutation has also been added as the reference). c) ^{31}P NMR spectrum (at 202 MHz) of the aqueous sample. d) ^{11}B NMR spectrum (at 160 MHz) of the aqueous sample. e) ^{23}Na NMR spectrum (at 132 MHz) of the aqueous sample. f) ^{13}C NMR spectrum (at 126 MHz) of the aqueous sample. g) 1H NMR spectrum (at 500 MHz) of the electrolyte sample. h) ^{19}F NMR spectrum (at 471 MHz) of the electrolyte sample. i) 7Li NMR spectrum (at 194 MHz) of the electrolyte sample.

micro-samples, the main obstacle to further improve the spectral resolution is the sample chamber not being at the isocentre of the magnet and the limited current of the shim channels. The shims of the commercial spectrometer are constructed and calibrated for macroscopic samples and coils, therefore the range of shim currents can not compensate for very local field inhomogeneities. More than two times enhancement in spectral resolution is expected by employing higher shim current values. It is worth noting that the spectral resolution is influenced by the material interfaces around the region of interest. Indeed, broadband coils have a superior spectral resolution compared to the multi-resonant coils with the same geometry, as a result of eliminating trap and tank circuits.

The ^1H nutation signal is in good agreement with the simulated signal presented in Figure 2. The decay rate of the nutation signal also matches perfectly with the simulation results. The only difference between $\pi/2$ pulse lengths ($\tau_{\pi/2} = 17 \mu\text{s}$ in the experimental and $13 \mu\text{s}$ in the simulation) is attributed to the extra power dissipation in the connectors, the solder points, the coaxial cable, and the excitation source which have not been considered in the simulations.

^{31}P and ^{23}Na experiments each demonstrate a single peak corresponding to the phosphate buffer solution (PBS), prepared using the sodium phosphate salts. The ^{11}B experiment shows a wide signal which is attributed to the boric anhydride B_2O_3 content in the glass substrate (D 263® T eco). In addition, multiple signals were observed as a result of boric acid chemical reaction with glucose. The overlapping signal corresponds to the borate-glucose complex molecule^{48,49}. ^{13}C experiments could also successfully reveal the glucose content although with a relatively low SNR which is because of its low receptivity and is inevitable in a direct detection scheme.

In the battery electrolyte sample, ^1H signal originated from the EMI protons in the solvent. The triplet signal at $\sim 1.50 - 1.56$ ppm is assigned to the proton at position 1, followed by quartet signal at ~ 4.22 ppm to ~ 4.48 ppm for carbon 2. A singlet was observed at ~ 3.80 ppm is assigned to the methyl protons at carbon 5. Protons at carbon 3 and 4 are symmetric and have the same doublet signals at the chemical shift in the higher field from $\sim 7.40 - 7.50$ ppm. Finally, a singlet signal was observed at ~ 8.50 ppm from the proton at carbon 6. ^7Li spectrum shows the lithium in the LiTFSI and ^{19}F content in both LiTFSI and TFSI was detected as a single peak in the ^{19}F spectrum. The measurement results are summarised in Table 3.

isotopes	aqueous sample					electrolyte sample		
	^1H	^{31}P	^{11}B	^{23}Na	^{13}C	^1H	^{19}F	^7Li
Excitation power (W)	1	1	1	1	1	1	1	1
$\pi/2$ pulse width (μs)	17	17	17	15	17	17	17	15
number of averages	512	2048	2048	32768	32768	256	128	512
SNR	4295	112.5	-	54	7.7	830.5	597.9	82.5
nLOD ($\text{nmol s}^{1/2}$)	88	135	-	1025	-	-	36	66

Table 3. NMR experiment results at seven different frequencies prove the broadband performance of the detector. SNR values were scaled to their single-scan equivalent before calculating nLOD.

As a unified method for reporting NMR detector sensitivity, we report the normalised limit of detection (nLOD) at 14.1 T magnetic field^{50,51}, summarised in Table 3. Comparison for ^1H shows that nLOD of the proposed broadband detector is comparable to the other single resonant detectors reported elsewhere (around $50 \text{ nmol s}^{1/2}$ for 100 nl sample volume). It should be noted that by overcoming the limit of shim currents, the nLOD values would be remarkably enhanced. For this purpose, one can consider shim-on-chip structures⁵² which are perfectly compatible with the current detector.

These results are additionally limited by 50Ω input impedance of the LNA and characteristic impedance of the cable. By employing an LNA with higher input impedance and design the coil accordingly, the results would be further improved. As a trend in micro-NMR, several studies have reported transceiver circuits being integrated to the coil⁵³⁻⁵⁶. The synergy between broadband NMR detectors and ASICs (application specific integrated circuits) will play a big role in future micro-NMR.

MRI experiments

The RF field pattern, extracted from ^1H MRI experiments, is given in Figure 4. This profile is compared with the simulation results and analytical expression derived from Biot-Savart formula (discussed in [Supplementary Information](#)), the three results being in good agreement. These results on one hand confirm the accuracy of the simulations, and on the other hand validate the simplifying assumptions made for the Biot-Savart equation.

Discussion

In a standard NMR experiment with a coil tuned to a single frequency, removing the coil and sample and tuning the coil to a secondary frequency is time consuming, affects the static field homogeneity, and causes relative dislocation of the sample and the coil. Thus, for experiments requiring multi-nuclear detection, a multi-resonant coil would be a better option. However,




Figure 4. a) Map of the RF magnetic field generated by the coil at the sample region extracted from the MRI experiments; b) measured B_1 field, generated by the coil at its axis, is in good agreement with the simulation results and analytical expression derived from Biot-Savart formula (Equation S1). The MRI results are valid just within the sample volume (gray rectangle). The mismatch at $y = 0 \mu\text{m}$ is a measurement error because of partial sample at the boundary voxels.

the trap/tank circuits introduce power loss and degrade both excitation efficiency and SNR in multi-resonant coils. It is even more challenging when adapting such circuits to micro-coils, since the micro-coils exhibit relatively small impedances and therefore the power loss in the accompanying circuits is considerable. An alternative solution is based on the broadband concept. This report provides a new layer of understanding by an in-depth, systematic study of the broadband coils, considering their different aspects. This contribution introduces a toolbox to accurately analyse and predict the performance of the detectors, either in broadband or in tuned/matched mode, using a 'digital twin' of the NMR figures of merit. In addition to the theoretical discussions, we have optimised and exploited a broadband spiral coil. According to the simulation results, the optimum design is more than 32 % as efficient as a single resonant coil, in terms of SNR. It should be noted that the lower SNR values can be compensated by more averaging. In a fair comparison, the efficiency of the broadband micro-coil should be compared to a multi-resonant micro-coil (in this case seven resonances) considering all previously-mentioned sources of decreased efficiency.

The optimised geometry has been fabricated using standard microfabrication techniques. Microfabrication has also enabled us to tailor the sample chamber and the microfluidic channel for the coil, and to form a hybrid system as the NMR detector. This detector has been experimentally evaluated at seven different frequencies. Besides, ^1H MRI experiments have been performed to verify the simulations. The normalised limit of detection extracted from the measurement results proves that this detector, despite its being broadband, has a performance that is comparable to the micro-detectors reported in an earlier review⁵¹, bearing in mind that those detectors are single resonant and cannot work at frequencies other than ^1H . It is worth mentioning that, even for a multi-resonant macro-coil, detection of seven different frequencies without any interruption to modify the circuit seems to be a non-solved problem. One shall also notice that the performance of the broadband coil is not limited to seven nuclei presented here, but the very same coil can be exploited for nuclei within the frequency range tested. Moreover, the performance of this detector can be further enhanced by integrating other features such as shim-on-chip features or integrating ASICs to amplify the collected signal locally and immunise it to the signal loss by the impedance mismatch between the commercial LNAs and the coil.

Methods

Finite Element Method (FEM) simulations

In order to evaluate and optimise the coil geometry, the commercially available COMSOL MultiPhysics® 5.5 (COMSOL AB, Sweden) was used. The RF and susceptibility studies have been carried out using RF and AC/DC modules, respectively. The relative tolerances of the RF simulations were set to 1×10^{-4} and relative tolerances of the susceptibility were 1×10^{-10} . The simulation results were processed in MATLAB® R2018a (MathWorks Inc., USA) to generate the 'digital twin' of NMR and nutation spectra and calculate the NMR detector figures of merit. To calculate the nutation spectra figures of merit, B_0 was assumed to be perfectly homogeneous. The NMR spectra were generated at 90° flip-angle and with no additional static field correction. Permeability values were assigned to different materials based on the values discussed elsewhere^{57,58}.

Micro-fabrication process

As depicted in Figure 5, the broadband detector consists of an assembly of three D 263® T eco wafers bonded together. The middle wafer is $500 \mu\text{m}$ thick. The microfluidic channels were structured in this wafer with a nanosecond laser (PIRANHA®




Figure 5. Fabrication process started with structuring the channel in the middle wafer. The process was continued with bonding three wafers together and coating the top side of the assembly with the seed layer. Using UV-lithography, a mold was patterned on the top side and the metallic structures were deposited. After removing the mold and the seed layer, the inlet and outlets were drilled with laser and the assembly was diced into individual chips. The process was concluded by connecting the centre pad of the coil to its corresponding track.

ACSYS). The wafers were bonded employing 5 μm -thick not-crosslinked, spincoated, SU-8 photoresist layers as the glue and employing a compression-bonding machine (EVG@510 EV Group) with 5 kN force and at 55 $^{\circ}\text{C}$ for 4 h. The assembly was exposed with UV-light (300 mJ cm^{-2} of 365 nm) followed by post-exposure bake (95 $^{\circ}\text{C}$ for 1 h) to enhance the chemical and mechanical stability of the bond. The 100 μm thick uppermost wafer was coated with chromium/gold seed layers (20/60 nm). The coil and access pads and tracks were patterned using standard SU-8 lithography process to produce the mould and gold-electrodeposition (20 μm). The 500 μm thick lowermost layer seals the microfluidic channel and maintains the mechanical stability of the structure. The inlet and outlet were lasered through the top wafer and the assembly was diced into individual chips. The centre pad of the coil was wirebonded to its corresponding track using an automated ball-wedge wirebonder (3100@ ESEC, Switzerland).

Probe head manufacturing

An MMBX connector was soldered to the detector to connect the chip to the coaxial cable (HUBER + SUHNER, Switzerland) bypassing all the tuning/matching circuits embedded in the conventional probe. The cable was 2 m in length and had an MMBX plug at one end and an N-type connector at the other end. It had an attenuation of less than 0.605 dB per meter and could tolerate continuous power of more than 54 W. The detector was mounted on a PMMA adaptor which fixes the detector on the probe.

Sample selection and preparation

Aqueous sample

D-glucose ($\geq 99.5\%$), 13-C labelled D-glucose ($\geq 99.5\%$), 3-(trimethylsilyl)- propionic-2,2,3,3-d₄ acid sodium salt (99.9 atom %, TSP), deuterium oxide D₂O (99.9 atom % D), phosphate buffer saline (PBS) tablets, boric acid (ACS reagent, $\geq 99.5\%$), and sodium dihydrogen phosphate (BioReagent, for molecular biology, anhydrous, $\geq 98\%$) were purchased from Sigma-Aldrich without further purification.

The aqueous test sample was prepared as 500 mM 13-C labelled D-glucose, 500 mM boric acid, 500 mM NaH₂PO₄ and 250 mM TSP dissolved in 3 mL of a 500 mM PBS solution (pH 5.5, using PBS tablets) prepared using 50 % H₂O and 50 % D₂O.

Battery electrolyte sample

Lithium bis(trifluoromethane-sulfonyl)imide (LiTFSI) salt was dissolved in 1-ethyl-3-methylimidazolium bis(trifluoromethane-sulfonyl)imide (EMI-TFSI) to make 1 M solution. Both chemicals were purchased from Tokyo Kasei Co., Ltd (Tokyo, Japan)

and used without further purification.

NMR/MRI experiments

The NMR/MRI were carried out using a 11.74 T Avance III Bruker NMR system (Bruker BioSpin, Rheinstetten, Germany) equipped with a Micro5 micro-imaging NMR probe. The sample was loaded in the microfluidic channel and the gradient sleeve was inserted. The probe together with the detector and gradient were inserted inside the magnet so that the middle axis of the microfluidic channel was along the z -axis of the magnet, whereas the excitation RF field is along the nominal y -axis of the magnet frame of reference.

For the ^1H measurement, the signal was routed through a narrow-band pre-amplifier (1H LNA MODULE 500, Bruker BioSpin, Rheinstetten, Germany) with 1 dB noise figure. X-nuclei signals were directed through a broadband pre-amplifier (XBB19F 2HS MODULE 500, Bruker BioSpin, Rheinstetten, Germany) with 2 dB noise figure. NMR experiments were performed using TopSpin 3.5pl2, the operating and processing software for Bruker NMR spectrometers. The applied power was 1 W and shimming was performed manually on the ^1H spectrum and was kept the same for all the other experiments. The temperature was maintained at 30 °C using the water cooling system embedded in the gradient. All NMR spectra were measured using a $\pi/2$ pulse width reported in Table 3.

General procedure; In all cases, the resulting signals were cumulated and multiplied with an exponential function equivalent to a specified amount of line broadening, followed by Fourier transformation. Base line and phase adjustments were performed automatically.

^1H NMR: The spectra acquired from the aqueous sample and electrolyte sample consisted of 512 scans and 256 scans, respectively. 20480 data points were collected for each scan with 20 ppm spectral width for the aqueous sample and 15 ppm spectral width for the battery electrolyte sample. The relaxation delay was 10 s in both experiments. 0.3 Hz line broadening applied. B_1 field homogeneity was evaluated using the nutation experiment on the aqueous sample employing 76 single scans at 1 W applied power with an increment of 2 μs in the pulse length. The relaxation delay between two consecutive scans was set to 15 s. The data points were integrated and graphed as a bar plot to reduce the distortions introduced by relaxation damping or B_0 inhomogeneities.

^{19}F NMR: The spectrum consisted of 128 scans. 20480 data points were collected for each scan with 20 ppm spectral width, while the relaxation delay was 10 s. 10 Hz line broadening applied.

^{31}P NMR: The spectrum consisted of 2048 scans. 30720 data points were collected for each scan with 60 ppm spectral width while the relaxation delay was 4 s. 5 Hz line broadening applied.

^7Li NMR: The spectrum consisted of 512 scans. 5120 data points were collected for each scan with 20 ppm spectral width while the relaxation delay was 1 s. 2 Hz line broadening applied.

^{11}B NMR: The spectrum consisted of 32768 scans. 10240 data points were collected for each scan with 250 ppm spectral width while the relaxation delay was 0.1 s. 100 Hz applied. Phase adjustment was done manually.

^{23}Na NMR: The spectrum consisted of 2048 scans. 2048 data points were collected for each scan with 30 ppm spectral width while the relaxation delay was 5 s. 1 Hz line broadening applied.

^{13}C NMR: The spectrum consisted of 32768 scans. 20480 data points were collected for each scan with 200 ppm spectral width while the relaxation delay was 0.1 s. 10 Hz line broadening applied.

^1H MRI: MRI experiments were performed using ParaVision 6.0.1 and the aqueous sample. For this purpose, *FLASH* sequences with the following settings were performed: TR/TE - 1000 ms/4 ms, slice thickness - 100 μm along the nominal y -axis, in-plane resolution - 32 \times 32 μm , Tscan - 21 min, FOV 1.6 \times 1.6 mm^2 , bandwidth - 5.4 kHz, and 64 averages. The excitation pulse length was set to 1.1 ms and the power level was swept from 0.1 mW to 624.1 mW non-linearly to have a linear sweep over flip angle. In order to determine the exact B_1 field of each voxel, the collected data were post-processed in MATLAB® R2018a (MathWorks Inc., USA) to calculate the nutation spectrum of each voxel. B_1 map was reconstructed using those nutation spectra.

Data availability

The datasets generated during the current study are available from the corresponding authors on reasonable request.

References

1. Wüthrich, K. The way to NMR structures of proteins. *Nat. Struct. Biol.* **8**, 923–925, DOI: [10.1038/nsb1101-923](https://doi.org/10.1038/nsb1101-923) (2001).
2. Sugiki, T., Kobayashi, N. & Fujiwara, T. Modern Technologies of Solution Nuclear Magnetic Resonance Spectroscopy for Three-dimensional Structure Determination of Proteins Open Avenues for Life Scientists. *Comput. Struct. Biotechnol. J.* **15**, 328–339, DOI: [10.1016/j.csbj.2017.04.001](https://doi.org/10.1016/j.csbj.2017.04.001) (2017).

3. Kupče, E. NMR with Multiple Receivers. In *eMagRes*, 721–732, DOI: [10.1002/9780470034590.emrstm1404](https://doi.org/10.1002/9780470034590.emrstm1404) (John Wiley & Sons, Ltd, Chichester, UK, 2015).
4. Kupče, E., Freeman, R. & John, B. K. Parallel acquisition of two-dimensional NMR spectra of several nuclear species. *J. Am. Chem. Soc.* **128**, 9606–9607, DOI: [10.1021/ja0634876](https://doi.org/10.1021/ja0634876) (2006).
5. Salzmann, M., Pervushin, K., Wider, G., Senn, H. & Wüthrich, K. TROSY in triple-resonance experiments: New perspectives for sequential NMR assignment of large proteins. *Proc. Natl. Acad. Sci.* **95**, 13585–13590, DOI: [10.1073/pnas.95.23.13585](https://doi.org/10.1073/pnas.95.23.13585) (1998).
6. Zangger, K. Pure shift NMR. *Prog. Nucl. Magn. Reson. Spectrosc.* **86-87**, 1 – 20, DOI: [10.1016/j.pnmrs.2015.02.002](https://doi.org/10.1016/j.pnmrs.2015.02.002) (2015).
7. Helge Meyer, N. & Zangger, K. Enhancing the resolution of multi-dimensional heteronuclear NMR spectra of intrinsically disordered proteins by homonuclear broadband decoupling. *Chem. Commun.* **50**, 1488–1490, DOI: [10.1039/C3CC48135B](https://doi.org/10.1039/C3CC48135B) (2014).
8. Johnson, R. L. & Schmidt-Rohr, K. Quantitative solid-state ¹³C NMR with signal enhancement by multiple cross polarization. *J. Magn. Reson.* **239**, 44 – 49, DOI: [10.1016/j.jmr.2013.11.009](https://doi.org/10.1016/j.jmr.2013.11.009) (2014).
9. Yamamoto, K., Caporini, M. A., Im, S.-C., Waskell, L. & Ramamoorthy, A. Cellular solid-state NMR investigation of a membrane protein using dynamic nuclear polarization. *Biochimica et Biophys. Acta (BBA) - Biomembr.* **1848**, 342 – 349, DOI: [10.1016/j.bbamem.2014.07.008](https://doi.org/10.1016/j.bbamem.2014.07.008) (2015).
10. Davoodi, H., Jouda, M., Korvink, J. G., MacKinnon, N. & Badilita, V. Broadband and multi-resonant sensors for NMR. *Prog. Nucl. Magn. Reson. Spectrosc.* **112-113**, 34–54, DOI: [10.1016/j.pnmrs.2019.05.001](https://doi.org/10.1016/j.pnmrs.2019.05.001) (2019).
11. Tang, P., Chien, W. J. & Harbison, G. S. Double-resonance circuit for nuclear magnetic resonance spectroscopy. *Solid State Nucl Magn Reson.* **2**, 343–348, DOI: [10.1016/0926-2040\(83\)90006-1](https://doi.org/10.1016/0926-2040(83)90006-1) (1993).
12. Park, S. H., De Angelis, A. A., Nevzorov, A. A., Wu, C. H. & Opella, S. J. Three-Dimensional Structure of the Transmembrane Domain of Vpu from HIV-1 in Aligned Phospholipid Bicelles. *Biophys. J.* **91**, 3032–3042, DOI: [10.1529/biophysj.106.087106](https://doi.org/10.1529/biophysj.106.087106) (2006).
13. Kan, S., Fan, M. & Courtieu, J. A single-coil triple resonance probe for NMR experiments. *Rev. Sci. Instruments* **51**, 887–890, DOI: [10.1063/1.1136352](https://doi.org/10.1063/1.1136352) (1980).
14. Li, Y., Logan, T. M., Edison, A. S. & Webb, A. Design of small volume HX and triple-resonance probes for improved limits of detection in protein NMR experiments. *J. Magn. Reson.* **164**, 128–135, DOI: [10.1016/S1090-7807\(03\)00184-8](https://doi.org/10.1016/S1090-7807(03)00184-8) (2003).
15. Martin, R. W., Paulson, E. K. & Zilm, K. W. Design of a triple resonance magic angle sample spinning probe for high field solid state nuclear magnetic resonance. *Rev. Sci. Instruments* **74**, 3045–3061, DOI: [10.1063/1.1571951](https://doi.org/10.1063/1.1571951) (2003).
16. Spengler, N. *et al.* Heteronuclear micro-helmholtz coil facilitates μ m-range spatial and sub-Hz spectral resolution NMR of nL-volume samples on customisable microfluidic chips. *PLoS ONE* **11**, 1–16, DOI: [10.1371/journal.pone.0146384](https://doi.org/10.1371/journal.pone.0146384) (2016).
17. Kendrick, R. D. & Yannoni, C. S. High-power ¹H-¹⁹F excitation in a multiple-resonance single-coil circuit. *J. Magn. Reson. (1969)* **75**, 506–508, DOI: [10.1016/0022-2364\(87\)90105-3](https://doi.org/10.1016/0022-2364(87)90105-3) (1987).
18. Lin, E. C., Wu, C. H., Yang, Y., Grant, C. V. & Opella, S. J. ¹H-¹³C separated local field spectroscopy of uniformly ¹³C labeled peptides and proteins. *J. Magn. Reson.* **206**, 105–111, DOI: [10.1016/j.jmr.2010.06.011](https://doi.org/10.1016/j.jmr.2010.06.011) (2010).
19. Tadanki, S., Colon, R. D., Moore, J. & Waddell, K. W. Double tuning a single input probe for heteronuclear NMR spectroscopy at low field. *J. Magn. Reson.* **223**, 64–67, DOI: [10.1016/J.JMR.2012.06.008](https://doi.org/10.1016/J.JMR.2012.06.008) (2012).
20. Hu, L. *et al.* A generalized strategy for designing ¹⁹F/¹H dual-frequency MRI coil for small animal imaging at 4.7 Tesla. *J. Magn. Reson. Imaging* **34**, 245–252, DOI: [10.1002/jmri.22516](https://doi.org/10.1002/jmri.22516) (2011).
21. Subramanian, R. & Webb, A. G. Design of solenoidal microcoils for high-resolution ¹³C NMR spectroscopy. *Anal. chemistry* **70**, 2454–8, DOI: [10.1021/ac980299s](https://doi.org/10.1021/ac980299s) (1998).
22. Kc, R., Henry, I. D., Park, G. H. J. & Raftery, D. Design and construction of a versatile dual volume heteronuclear double resonance microcoil NMR probe. *J. Magn. Reson.* **197**, 186–192, DOI: [10.1016/j.jmr.2008.12.020](https://doi.org/10.1016/j.jmr.2008.12.020) (2009).
23. Collier, K. A. *et al.* Design and construction of a quadruple-resonance MAS NMR probe for investigation of extensively deuterated biomolecules. *J. Magn. Reson.* **285**, 8–17, DOI: [10.1016/j.jmr.2017.10.002](https://doi.org/10.1016/j.jmr.2017.10.002) (2017).
24. Goluch, S. *et al.* Proton-decoupled carbon magnetic resonance spectroscopy in human calf muscles at 7 T using a multi-channel radiofrequency coil. *Sci. Reports* **8**, 1–10, DOI: [10.1038/s41598-018-24423-x](https://doi.org/10.1038/s41598-018-24423-x) (2018).

25. Rispoli, J. V. *et al.* Trap design and construction for high-power multinuclear magnetic resonance experiments. *Concepts Magn. Reson. Part B: Magn. Reson. Eng.* **46B**, 162–168, DOI: [10.1002/cmr.b.21345](https://doi.org/10.1002/cmr.b.21345) (2016).
26. Serés Roig, E. *et al.* A double-quadrature radiofrequency coil design for proton-decoupled carbon-13 magnetic resonance spectroscopy in humans at 7T. *Magn. Reson. Medicine* **73**, 894–900, DOI: [10.1002/mrm.25171](https://doi.org/10.1002/mrm.25171) (2015).
27. Ramaswamy, V. *et al.* Development of a ¹³C-optimized 1.5-mm high temperature superconducting NMR probe. *J. Magn. Reson.* **235**, 58 – 65, DOI: [10.1016/j.jmr.2013.07.012](https://doi.org/10.1016/j.jmr.2013.07.012) (2013).
28. Oosthoek-De Vries, A. J. *et al.* Continuous Flow ¹H and ¹³C NMR Spectroscopy in Microfluidic Stripline NMR Chips. *Anal. Chem.* **89**, 2296–2303, DOI: [10.1021/acs.analchem.6b03784](https://doi.org/10.1021/acs.analchem.6b03784) (2017).
29. Brauckmann, J. O., Janssen, J. W. G. H. & Kentgens, A. P. M. High resolution triple resonance micro magic angle spinning NMR spectroscopy of nanoliter sample volumes. *Phys. Chem. Chem. Phys.* **18**, 4902–4910, DOI: [10.1039/C5CP07857A](https://doi.org/10.1039/C5CP07857A) (2016).
30. Ryan, H., Song, S.-H., Zaß, A., Korvink, J. & Utz, M. Contactless NMR Spectroscopy on a Chip. *Anal. Chem.* **84**, 3696–3702, DOI: [10.1021/ac300204z](https://doi.org/10.1021/ac300204z) (2012).
31. Murphree, D., Cahn, S. B., Rahmlow, D. & DeMille, D. An easily constructed, tuning free, ultra-broadband probe for NMR. *J. Magn. Reson.* **188**, 160–167, DOI: [10.1016/j.jmr.2007.05.025](https://doi.org/10.1016/j.jmr.2007.05.025) (2007).
32. Kubo, A. & Ichikawa, S. Ultra-broadband NMR probe: Numerical and experimental study of transmission line NMR probe. *J. Magn. Reson.* **162**, 284–299, DOI: [10.1016/S1090-7807\(03\)00014-4](https://doi.org/10.1016/S1090-7807(03)00014-4) (2003).
33. Tang, J. A., Wiggins, G. C., Sodickson, D. K. & Jerschow, A. Cutoff-free traveling wave NMR. *Concepts Magn. Reson. Part A* **38A**, 253–267, DOI: [10.1002/cmr.a.20226](https://doi.org/10.1002/cmr.a.20226) (2011).
34. Scharfetter, H., Petrovic, A., Eggenhofer, H. & Stollberger, R. A no-tune no-match wideband probe for nuclear quadrupole resonance spectroscopy in the VHF range. *Meas. Sci. Technol.* **25**, 125501, DOI: [10.1088/0957-0233/25/12/125501](https://doi.org/10.1088/0957-0233/25/12/125501) (2014).
35. Mandal, S., Utsuzawa, S. & Song, Y. Q. An extremely broadband low-frequency MR system. *Microporous Mesoporous Mater.* **178**, 53–55, DOI: [10.1016/j.micromeso.2013.03.040](https://doi.org/10.1016/j.micromeso.2013.03.040) (2013).
36. Fratila, R. M., Gomez, M. V., Sýkora, S. & Velders, A. H. Multinuclear nanoliter one-dimensional and two-dimensional NMR spectroscopy with a single non-resonant microcoil. *Nat. communications* **5**, 3025, DOI: [10.1038/ncomms4025](https://doi.org/10.1038/ncomms4025) (2014).
37. Massin, C. *et al.* High-Q factor RF planar microcoils for micro-scale NMR spectroscopy. *Sensors Actuators, A: Phys.* **97-98**, 280–288, DOI: [10.1016/S0924-4247\(01\)00847-0](https://doi.org/10.1016/S0924-4247(01)00847-0) (2002).
38. Massin, C. *et al.* Planar microcoil-based microfluidic NMR probes. *J. Magn. Reson.* **164**, 242–255, DOI: [10.1016/S1090-7807\(03\)00151-4](https://doi.org/10.1016/S1090-7807(03)00151-4) (2003).
39. Liu, S. *et al.* Physical Models of Planar Spiral Inductor Integrated on the High-Resistivity and Trap-Rich Silicon-on-Insulator Substrates. *IEEE Transactions on Electron Devices* **64**, 2775–2781, DOI: [10.1109/TED.2017.2700022](https://doi.org/10.1109/TED.2017.2700022) (2017).
40. Yang, M. Z., Dai, C. L. & Hong, J. Y. Manufacture and characterization of high Q-factor inductors based on CMOS-MEMS techniques. *Sensors* **11**, 9798–9806, DOI: [10.3390/s111009798](https://doi.org/10.3390/s111009798) (2011).
41. Kentgens, A. P. M. *et al.* High-resolution liquid- and solid-state nuclear magnetic resonance of nanoliter sample volumes using microcoil detectors. *The J. Chem. Phys.* **128**, 052202, DOI: [10.1063/1.2833560](https://doi.org/10.1063/1.2833560) (2008).
42. Bart, J., Janssen, J., van Bentum, P., Kentgens, A. & Gardeniers, J. Optimization of stripline-based microfluidic chips for high-resolution NMR. *J. Magn. Reson.* **201**, 175–185, DOI: [10.1016/j.jmr.2009.09.007](https://doi.org/10.1016/j.jmr.2009.09.007) (2009).
43. Hault, D. & Richards, R. The signal-to-noise ratio of the nuclear magnetic resonance experiment. *J. Magn. Reson. (1969)* **24**, 71–85, DOI: [10.1016/0022-2364\(76\)90233-X](https://doi.org/10.1016/0022-2364(76)90233-X) (1976).
44. Friis, H. T. Noise figures of radio receivers. *Proc. IRE* **32**, 419–422 (1944).
45. Fujii, K. *et al.* Liquid Structure of Room-Temperature Ionic Liquid, 1-Ethyl-3-methylimidazolium Bis-(trifluoromethanesulfonyl) Imide. *The J. Phys. Chem. B* **112**, 4329–4336, DOI: [10.1021/jp7105499](https://doi.org/10.1021/jp7105499) (2008).
46. Lassègues, J.-C., Grondin, J. & Talaga, D. Lithium solvation in bis(trifluoromethanesulfonyl)imide-based ionic liquids. *Phys. Chem. Chem. Phys.* **8**, 5629–5632, DOI: [10.1039/B615127B](https://doi.org/10.1039/B615127B) (2006).
47. Garcia, B., Lavallée, S., Perron, G., Michot, C. & Armand, M. Room temperature molten salts as lithium battery electrolyte. *Electrochimica Acta* **49**, 4583–4588, DOI: [10.1016/j.electacta.2004.04.041](https://doi.org/10.1016/j.electacta.2004.04.041) (2004).

48. Smith, P., Donovan, K., Szekely, O., Baias, M. & Frydman, L. Ultrafast nmr t1 relaxation measurements: Probing molecular properties in real time. *ChemPhysChem* **14**, 3138–3145, DOI: [10.1002/cphc.201300436](https://doi.org/10.1002/cphc.201300436) (2013).
49. Swyer, I. *et al.* Interfacing digital microfluidics with high-field nuclear magnetic resonance spectroscopy. *Lab on a Chip* **16**, 4424–4435 (2016).
50. Finch, G., Yilmaz, A. & Utz, M. An optimised detector for in-situ high-resolution nmr in microfluidic devices. *J. Magn. Reson.* **262**, 73 – 80, DOI: [10.1016/j.jmr.2015.11.011](https://doi.org/10.1016/j.jmr.2015.11.011) (2016).
51. Badilita, V. *et al.* Microscale nuclear magnetic resonance: a tool for soft matter research. *Soft Matter* **8**, 10583, DOI: [10.1039/c2sm26065d](https://doi.org/10.1039/c2sm26065d) (2012).
52. van Meerten, S. G. J., van Bentum, P. J. M. & Kentgens, A. P. M. Shim-on-Chip Design for Microfluidic NMR Detectors. *Anal. Chem.* **90**, 10134–10138, DOI: [10.1021/acs.analchem.8b02284](https://doi.org/10.1021/acs.analchem.8b02284) (2018).
53. Anders, J., Handwerker, J., Ortmanns, M. & Boero, G. A low-power high-sensitivity single-chip receiver for NMR microscopy. *J. Magn. Reson.* **266**, 41–50, DOI: [10.1016/j.jmr.2016.03.004](https://doi.org/10.1016/j.jmr.2016.03.004) (2016).
54. Grisi, M., Conley, G. M., Sommer, P., Tinembart, J. & Boero, G. A single-chip integrated transceiver for high field NMR magnetometry. *Rev. Sci. Instruments* **90**, 015001, DOI: [10.1063/1.5066436](https://doi.org/10.1063/1.5066436) (2019).
55. Grisi, M., Gualco, G. & Boero, G. A broadband single-chip transceiver for multi-nuclear NMR probes. *Rev. Sci. Instruments* **86**, 044703, DOI: [10.1063/1.4916206](https://doi.org/10.1063/1.4916206) (2015).
56. Jouda, M., Gruschke, O. G. & Korvink, J. G. A new fully integrated multichannel receiver design for magnetic resonance imaging. *Concepts Magn. Reson. Part B: Magn. Reson. Eng.* **46B**, 134–145, DOI: [10.1002/cmrb.21341](https://doi.org/10.1002/cmrb.21341) (2016).
57. Schenck, J. F. The role of magnetic susceptibility in magnetic resonance imaging: MRI magnetic compatibility of the first and second kinds. *Med. Phys.* **23**, 815–850, DOI: [10.1118/1.597854](https://doi.org/10.1118/1.597854) (1996).
58. Wapler, M. C. *et al.* Magnetic properties of materials for MR engineering, micro-MR and beyond. *J. Magn. Reson.* **242**, 233 – 242, DOI: [10.1016/j.jmr.2014.02.005](https://doi.org/10.1016/j.jmr.2014.02.005) (2014).
59. Eroglu, S., Gimi, B., Roman, B., Friedman, G. & Magin, R. L. NMR spiral surface microcoils: Design, fabrication, and imaging. *Concepts Magn. Reson.* **17B**, 1–10, DOI: [10.1002/cmrb.10068](https://doi.org/10.1002/cmrb.10068) (2003).

Acknowledgements

HD, VB, NM, and NN gratefully acknowledge the financial support from the DFG (Deutsche Forschungsgemeinschaft) for the project Bio-PRICE (DFG BA 4275/4-1, DFG MA 6653/1-1). JGK acknowledges support from the DFG, the EU, and the KIT VIRMAT initiative. HM acknowledges the financial support from Tokyo Metropolitan University for the Promotion of Joint International Research. HD acknowledges support by the Karlsruhe Nano Micro Facility (KNMF), a Helmholtz Research Infrastructure at Karlsruhe Institute of Technology. This work was partly carried out with the support of the Helmholtz Research Area Information, Materials Systems Engineering, Topic 5 Materials Information Discovery at the Karlsruhe Institute of Technology (KIT, www.kit.edu). Authors thank Dr. Claudia Muhle-Goll for ¹¹B benchmarking experiments.

Author contributions statement

HD, NN, VB, and NM conceived the experiments, HD performed the simulations, HD designed and fabricated the chips, HD and NN conducted the experiments, HM prepared the EMI-TFSI ionic liquid battery electrolyte, HD and NN processed the results, HD, NN, and VB analysed the results. VB, NM, and JGK acquired funding for this research. VB, NM and JGK provided supervision. All authors reviewed the manuscript.

Additional information

The authors declare no competing interests.

Supplementary Information

Coil geometry optimisation

Frequency (MHz)	tuned/matched				Broadband			
	excitation efficiency		SNR		excitation efficiency		SNR	
	opt. n_{TM}	$\frac{ef_{TM}(n=17)}{ef_{TM}(opt.n_{TM})}$	opt. n_{TM}	$\frac{SNR_{TM}(n=17)}{SNR_{TM}(opt.n_{TM})}$	opt. n_{BB}	$\frac{ef_{BB}(n=8)}{ef_{BB}(opt.n_{BB})}$	opt. n_{BB}	$\frac{SNR_{BB}(n=8)}{SNR_{BB}(opt.n_{BB})}$
150	12	97 %	25	94 %	9	98 %	11	89 %
200	11	96 %	25	94 %	7	99 %	9	94 %
500	11	94 %	19	98 %	5	86 %	7	99 %

Table S1. Evaluation of the optimum coil in each configuration. The optimum numbers of turns for each configuration (tuned/matched and broadband) and based on each figure of merit (excitation efficiency and SNR) and at different working frequency (150 MHz, 200 MHz, and 500 MHz) are calculated according to the simulation results. Selecting a single coil for each configuration corresponds to a performance compromise. For example, the broadband 8 turn coil is 89 % as efficient as the the 11 turn coil at 150 MHz in terms of SNR.

Coil characteristics

For a spiral coil working at frequencies far below its self-resonant frequency, the RF magnetic field at the axis of the coil generated for a unit current can be calculated⁵⁹ using Biot-Savart law as

$$B_{1u}(z_o) = \frac{\mu}{2(w+s)} \left[\frac{r}{\sqrt{z_o^2 + r^2}} - \frac{r + (w+s)n}{\sqrt{z_o^2 + [r + (w+s)n]^2}} + \ln \left(\frac{r + (w+s)n + \sqrt{z_o^2 + [r + (w+s)n]^2}}{r + \sqrt{z_o^2 + r^2}} \right) \right] \quad (S1)$$

where z_o is the distance between the observation point and the centre of the coil, and has been assumed much larger than the thickness of the tracks. μ is the magnetic permeability of the sample, r is the inner radius of the coil, and n is its number of the turns. w and s are the width of the tracks and the separation between them, respectively and are assumed to be constant.

Benchtop measurements

In order to verify the simulations and analyse the coil performance, the impedance of the optimised coil was measured and compared to the simulation results. For this purpose, a test structure containing a coil with the same number of turns as the optimised coil was designed and fabricated together with the detectors as shown in Figure S1. The impedance was measured using a probe station (MPS150, FormFactor® Thiendorf, Germany) connected to a Network Analyzer (E5071C, Keysight® Stuttgart, Germany) employing a Z-probe (Z0-20-K3N-GS-50, FormFactor® Thiendorf, Germany) and WinCal™ XE software (FormFactor® Thiendorf, Germany) as shown in Figure S1. In order to eliminate the parasitic effects of the metallic chunk on the field distribution of the coils and their impedance, it was replaced with a PMMA chunk after calibration. A similar geometry was simulated using COMSOL. The results are presented in Figure S1. These results confirm the accuracy of the simulations for sub-GHz frequencies. In higher frequencies, difference in the self resonance frequency of simulated structures with the measured self-resonance frequency is the main reason for inaccuracies.




Figure S1. a) test structure; b) impedance measurement setup; c) impedance of the test structure in comparison with the simulation results.

Figures




Figure 1

A vertical-bore superconductive magnet together with the probe, LNA block, and electronic racks for signal processing. The detector, being mounted on the NMR probe is directly connected to the LNA block using a coaxial cable. The sample is loaded in the microfluidic channel of the detector.




Figure 2

a) Simulation results using COMSOL. a) $B_1=i$ at 500MHz (corresponding to ^1H Larmor frequency at 11.74 T). The dashed rectangle represents the sample volume and the line plot, corresponding to the upper x-axis, shows the field at the axis of the coil; b) resistance, inductance, quality factor, and average $B_1=i$ at three different frequencies and for different number of turns; c) Nutation curves for the predefined sample volume at three different frequencies and for different number of turns ; d) Normalised figures of merit of the spiral coils for tuned/matched case and broadband performance; e) B_0 distribution of the coil for three different sample geometries and at three different distances from the coil together with their corresponding NMR spectra at 500MHz using the optimised coil.




Figure 3

NMR experiment using the optimised NMR detector to observe different nuclei: a) ^1H NMR spectrum (at 500 MHz) of the aqueous sample. inset is the same spectrum with enlarged amplitude. b) ^1H nutation signal (simulated nutation has also been added as the reference). c) ^{31}P NMR spectrum (at 202 MHz) of the aqueous sample. d) ^{11}B NMR spectrum (at 160 MHz) of the aqueous sample. e) ^{23}Na NMR spectrum (at 132 MHz) of the aqueous sample. f) ^{13}C NMR spectrum (at 126 MHz) of the aqueous sample. g) ^1H NMR spectrum (at 500 MHz) of the electrolyte sample. h) ^{19}F NMR spectrum (at 471 MHz) of the electrolyte sample . i) ^7Li NMR spectrum (at 194 MHz) of the electrolyte sample.




Figure 4

a) Map of the RF magnetic field generated by the coil at the sample region extracted from the MRI experiments; b) measured B_1 field, generated by the coil at its axis, is in good agreement with the simulation results and analytical expression derived from Biot-Savart formula (Equation S1). The MRI results are valid just within the sample volume (gray rectangle). The mismatch at $y = 0 \mu\text{m}$ is a measurement error because of partial sample at the boundary voxels.




Figure 5

Fabrication process started with structuring the channel in the middle wafer. The process was continued with bonding three wafers together and coating the top side of the assembly with the seed layer. Using UV-lithography, a mold was patterned on the top side and the metallic structures were deposited. After removing the mold and the seed layer, the inlet and outlets were drilled with laser and the assembly was

diced into individual chips. The process was concluded by connecting the centre pad of the coil to its corresponding track.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [FigureS1.jpg](#)