Restoration of Complex Movement in the Paralyzed Upper Limb

Brady Hasse

University of Arizona <https://orcid.org/0000-0002-5273-3092>

Drew Sheets

University of Arizona

Nicole Holly

University of Arizona

Katalin Gothard

University of Arizona

Andrew Fuglevand (✉ fuglevan@email.arizona.edu)

College of Medicine, University of Arizona <https://orcid.org/0000-0002-4349-6478>

Article

Keywords:

Posted Date: January 20th, 2022

DOI: <https://doi.org/10.21203/rs.3.rs-1255522/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Date: January 12, 2022

Title: **Restoration of Complex Movement in the Paralyzed Upper Limb**

Authors: Brady A. Hasse^{1,2}, Drew E.G. Sheets^{1,3}, Nicole L. Holly¹, Katalin M. Gothard^{1,4,5}, Andrew J. Fuglevand^{1,4},✉

Affiliations: ¹Department of Physiology, College of Medicine, University of Arizona, Tucson AZ, USA, ²Graduate Program in Neuroscience, University of Arizona, Tucson AZ, USA, ³Present Address: Program in Computational Neuroscience, University of Chicago, Chicago IL, USA, ⁴Department of Neuroscience, University of Arizona, Tucson AZ, USA, ⁵Department of Neurology, College of Medicine, University of Arizona, Tucson AZ, USA, ✉fuglevan@email.arizona.edu

ABSTRACT

Functional electrical stimulation (FES) involves artificial activation of skeletal muscles to reinstate motor function in paralyzed individuals. While FES applied to the upper limb has improved the ability of tetraplegics to perform activities of daily living, there are key shortcomings impeding its widespread use. One major limitation is that the range of motor behaviors that can be generated is restricted to a small set of simple, preprogrammed movements. This limitation stems from the substantial difficulty in determining the patterns of stimulation across many muscles required to produce more complex movements. Here we show that machine learning can be used to flexibly identify patterns of muscle stimulation needed to evoke a wide array of multi-joint arm movements. Those patterns were delivered concurrently to 29 upper limb muscles in temporarily paralyzed monkeys. In some cases, the evoked responses matched the desired movements with good fidelity. Because the range of movements that can be produced is virtually unlimited, this approach could greatly expand the repertoire of movements available to individuals with high level paralysis. [171/175 allowed words]

Current attempts to restore limb function following spinal cord injury (SCI) rests on four main strategies: surgical reconstruction - often involving tendon transfers from non-paralyzed to paralyzed muscle^[1], repair the injury using a variety of neuroprotective and neuroregenerative agents including cell transplantation^[2-6], artificially enable movement with powered exoskeletons^[7-9], or activate paralyzed muscles, either by exciting spinal circuits below the level of the lesion^[10-18] or by directly stimulating motor axons innervating muscle^[19-24]. Therapeutic approaches, such as movement training^[11, 25-28] and sustained electrical stimulation spanning (and upstream) of the injury site^[30-33] have also been used effectively to enhance recovery in SCI, presumably by promoting axonal outgrowth^[29] and synaptic plasticity^[33]. Most of these approaches have focused on reinstating patterned rhythmical movements of the legs (or hindlimbs) for locomotion. The challenges associated with attempts to restore upper limb function are more significant, given the inordinate mechanical and coordinative complexity, episodic nature, and abundant variety of arm movements. As such, functional electrical stimulation (FES), involving excitation of motor axons with implanted electrodes^[19-24] provides the only feasible means to selectively control individual muscles needed to enact a wide array of motor behaviours of the arm and hand.

Despite its promise, the total number of tetraplegics who have benefited from such FES is meager (< 300^[21]) compared to ~ 250,000 new cases of SCI-related tetraplegia each year worldwide^[34, 35]. Paradoxically, one reason underlying such limited use of upper limb FES is the restricted range of motor behaviors that can be elicited with these systems. Indeed, a key barrier to more versatile control of FES is identifying the complex spatio-temporal patterns of muscle stimulation needed to evoke a large repertoire of movements. To overcome this limitation, here we use machine learning to predict patterns of muscle stimulation needed to evoke a wide range of complex movements in paralyzed upper limbs.

RESULTS

Our approach combined four sequential stages as depicted in Fig. 1a. In stage 1, we recorded limb kinematics and the associated electromyographic (EMG) signals from 29 upper limb muscles (Suppl table 1) using chronically implanted intramuscular electrodes (Fig. 1b) while a

monkey made a variety of arm movements (Fig 1c). The recorded EMG signals (Fig. 1d) were

Figure 1. a) Steps involved in the approach. b) Radiograph showing titanium plate on skull to which electrode-connector encasement was mounted. Electrode bundle emerged posteriorly from encasement and was routed to a site between the scapulae. From there, electrodes (2 electrodes/muscle) were tunneled to target muscles where gold anchors were attached and inserted into target muscles. Disk over spinous processes is an implanted ground/return electrode. c) small segment (~ 30 s) of hand trajectory (measured relative to shoulder location, sampling rate 120 Hz) used for training artificial neural network (ANN). d) Raw EMG signals recorded during the movement shown in c from 29 muscles that act primarily on the digits of the hand, wrist, elbow, shoulder, and scapula. Acronyms of muscles defined in Supplement Table 1. e) Rectified, lowpass filtered, and normalized (to peak EMG recorded in a session) representations of the signals shown in d. These signals were used in the training of the ANN. All traces shown on the same normalized scale.

rectified, low pass filtered, and normalized to the peak value detected within a recording session (Fig 1e). Kinematics included 3-dimensional positions, velocities, and accelerations of hand and elbow relative to the shoulder, and pitch, roll, and yaw angular orientations, velocities, and accelerations of the hand. These signals were then used as inputs to train a machine-learning algorithm (an artificial neural network, ANN) that characterized the relationship between kinematics and muscle activity. In stage 2 (Fig. 1a), the trained ANN predicted muscle activity associated with a new set of desired movements not included in the training set. In Stage 3, predicted patterns of muscle activity were transformed into stimulus pulse trains. In Stage 4, the stimulus pulses were delivered to muscles through the implanted electrodes to evoke upper limb movements in two other implanted ‘test’ monkeys that were temporarily paralyzed (i.e., under general anesthesia). We used different ‘trainer’ and ‘test’ monkeys because paralyzed humans would not be able to provide the signals needed to train such an ANN.

The ability of the ANN to predict muscle activity was evaluated for a set of desired movements recorded from the two test monkeys. Nine movements were selected for each test monkey (Fig. 2a) as representative of the complexity, duration, and regions of the reach space explored from a large set of movements recorded in both monkeys. Each test movement started and ended with the hand resting on an instrumented start box. Fig. 2b shows examples traces of actual and predicted normalized EMG signals (‘activation’) for four muscles during the nine movements in one monkey. Activation signals associated with each movement (labeled 1 – 9 in

Fig. 2b) are concatenated for illustration purposes. Predictions based on ANNs trained on data

Figure 2. **a)** Nine test movements for one monkey used to evaluate capability of machine learning to predict patterns of muscle activity. Hand position relative to shoulder position (diamond) was sampled at 120 Hz and normalized to arm length (as shown for movement 5). Color indicates time of movement, with time progressing from cool to hot colors. All movements began and ended with hand on an instrumented start box. **b)** Actual (red) and predicted (green – within subject ANN, blue – across subject ANN) EMG (activation) signals in four proximal muscles for each of the 9 test movements. **c)** Mean (SD) R^2 values between actual and predicted activation signals. Two-way ANOVA significant for training data source ($p < 0.001$) and muscle group ($p < 0.001$) but no interaction ($p = 0.66$). Holm-Sidak post-hoc analysis indicated that scapular, shoulder, and elbow muscle groups had higher R^2 values ($p < 0.05$) than wrist or digit muscle groups. **d)** Mean (SD) root mean square error (RMSE) between actual and predicted activation signals. Two-way ANOVA significant for training data source ($p < 0.001$) and muscle group ($p < 0.001$) but no interaction ($p = 0.13$). Holm-Sidak post-hoc analysis indicated that RMSE for scapular muscle group was greater ($p < 0.05$) than other muscle groups.

collected from the same subject (*within* – green trace) and from the trainer monkey (*across* – blue trace) are compared to the actual activation signals (red trace).

For these proximal muscles (Fig. 2b), *within* predictions were excellent, having large coefficient of determination (R^2) values (range 0.78 – 0.92) and small root mean squared errors (RMSE, range 2.9 – 4.8% of peak activation). In itself, such accuracy is remarkable, given the complexity of the movements and muscle activity patterns. *Across* predictions (blue traces, Fig. 2b) for these muscles were not as accurate but were still quite good (R^2 range 0.29 – 0.85, RMSE range 5.6 – 9.2%). When averaged across both monkeys and all movements (Figs. 2c, d) – *within* predictions were significantly better than *across*, as we have reported previously for human subjects [36-39]. In addition, correlations between actual and predicted activations (Fig. 2c) were significantly greater for proximal muscle groups (those mainly acting on the scapula, shoulder, and elbow) than for distal muscles (those primarily operating at the wrist and digits). This likely reflects the more direct involvement of proximal muscles in producing the limb movements recorded in these experiments. Interestingly, the RMSE was greater for scapular than for other muscle groups (Fig. 2d). This seems related to the higher overall activity levels for scapular muscles than other muscles during the tested movements (Suppl. Fig. 1). In terms of kinematic parameters used to train the ANN, only a modest number (primarily those related to hand position and velocity) were needed to obtain the best predictions (Suppl. Figure 2).

We next converted predicted (and actual) muscle activations associated with the test movements into stimulus pulse trains to be delivered to the 29 implanted muscles in anesthetized test monkeys. For the present application, we modified our previous approach^[40] and used amplitude-modulated, biphasic current pulses running at 40 pulses/s. For each of the 58 implanted electrodes (2 in each muscle), we first needed to identify the range of stimulus intensities that activated the target muscle without exciting neighboring muscles. To do this, monkeys were sedated and secured in a modified infant car seat (Fig 3a). A transducer was fixed to the wrist/hand (Fig. 3b), forearm, or upper arm to record the 3-dimensional components of evoked isometric forces associated with muscles acting on each of those segments. Limb segments proximal to the tested segments were immobilized with Velcro straps (Fig. 3a).

Single pulses were then delivered (1 pulse/s) in an incrementing fashion through each electrode and the evoked twitches recorded. As examples, Figs. 3c and 3e show the time-courses of the resultant twitch forces at each stimulus intensity for teres major and flexor carpi radialis. Figs 3d and 3f show the force vectors associated with the peak resultant force at each intensity. For teres major (Fig. 3c), evoked responses showed two phases, that leading up to an initial plateau, and a subsequent rise to a second plateau. Such intermediate plateaus could be due to activation of distant nerve branches within the same muscle^[41] or to activation of neighboring muscles. To distinguish these possibilities, we examined the direction of the resultant force vectors. As shown in Fig. 3d, the direction of the resultant force began to change for stimulus intensities above 17 mA, which we attributed to excitation of neighboring muscles. We used the largest stimulus intensity prior to such changes in force direction as the upper limit of stimulus intensity for each muscle. Lower limits were discerned as the minimal current that just evoked a noticeable response.

The muscles in the forearm are relatively small and their stimulation can readily cause nearby muscles to be activated. Furthermore, detecting changes in force direction can be challenging, particularly for neighboring synergists. To address this, we used bipolar stimulation in forearm muscles (current passing between active and return electrodes both within the same muscle) rather than monopolar stimulation (current passing between active electrode in muscle and the large return electrode between the scapulae) used for larger, more proximal muscles. This

seemed to be an effective way to restrict current spread and minimize activation of neighboring muscles. Figs. 3d,e show an example of such bipolar stimulation of a forearm muscle wherein little change in force direction occurred up through the maximum current delivered. In these cases, we identified the lowest intensity needed to achieve $\sim 95\%$ of the maximum force as the upper limit for stimulation (bold line, Fig. 3d).

For each electrode, the peak resultant force was fit as a logistic function of stimulus intensity between the minimum and maximum current levels. Knowing that the relation between EMG activity and isometric force is roughly linear, we substituted normalized EMG (activation) for

Figure 3. a) Set up for identifying stimulus current range for each muscle electrode. Monkeys were sedated and secured upright in an infant car seat. An isometric transducer was fixed to the wrist/hand (b), forearm, or upper arm to record the 3-dimensional components of evoked isometric forces associated with muscles acting on each of those segments. Limb segments proximal to the tested segments were immobilized with Velcro straps. c) Example resultant twitch force responses in teres major and associated d) force vectors for stimulus pulses that incremented in amplitude in 0.2 mA steps from 0 to 2 mA and in 1.0 mA steps from 3 – 32 mA. The return electrode was a large subcutaneous disk between the scapulae (Fig. 1b). No responses were seen (blue traces, c) until the stimulus amplitude reached 1.4 mA (min). Evoked responses then increased and had a consistent direction up to a stimulus amplitude of 17 mA (max, bold line) whereupon the force vector began to change direction and showed a second rise in magnitude (red traces, c, d). Changes in force direction were attributed to the activation of nearby muscles. e, f) Same as for panels c and d except for a different muscle, flexor carpi radialis. In this forearm muscle, both active and return electrodes were situated within the muscle. In this case, little change in force direction was observed (f). The upper limit of stimulus intensity (max) was identified as the lowest current that evoked ~95% of the maximum force (bold trace).

force in the logistic stimulus intensity curves ^[40]. Then for each of the 9 test movements, actual and predicted (within and across) activation signals for all muscles were transformed offline into 40 Hz trains of stimulus pulses with current amplitude modulated as an inverse logistic function of activation.

Then to test the degree to which the 9 test movements (different movements for the two monkeys) could be elicited artificially, both test monkeys were sedated in five sessions each and secured to infant car seat. Their implanted arm was unrestrained, instrumented with movement sensors, and positioned similar to the start position during awake reaching (Fig. 4a). The intramuscular electrodes were connected to a bank of programmable stimulators holding the stimulus pulse patterns for each test movement associated with actual, within-predicted, and across-predicted activation signals. Then in each session, each stimulus pattern (9 movements x 3 sources of activation) was delivered to the muscles and the resulting movements recorded. The order of stimulus patterns was varied across sessions.

A video depicting a test movement and the associated evoked movement is provided in the Supplemental material. The evoked movement is smooth, complex (including a lateral reach, followed by transport of the hand near the mouth, and a return to start position), and generally captures the profile of the desired movement. Figs. 4b, c show example trajectories of a different desired and evoked movement. The time varying *x*, *y*, *z* positions of the hand (relative to the shoulder) for this example are depicted in Fig. 4d. For this case, while the vertical (*z*) dimension

was reasonably well re-produced, an initial lateral (positive y) displacement of the hand during electrical stimulation led to clear differences between desired and evoked trajectories in the first half of the movement. Likewise, a posteriorly directed (decreasing x direction) hand motion during stimulation in the second half of the trial also led to notable error. As this example highlights, it could be that even modest imbalances in muscle forces applied to the unloaded limb may lead to significant displacement errors.

When averaged across both test monkeys, all test movements, and x , y , z directions, the average R^2 values between evoked and desired trajectories were all > 0.3 (Fig. 4e). These were all significantly greater than that associated with movements evoked by random shuffling of the stimulus patterns such that each muscle received a stimulus pattern designated for a different muscle. Likewise, the RMSE was greater for the shuffled trials than for the other trials (reaching significance, however, only for the within and across conditions). Such shuffling provides a reasonably stringent comparison as the underlying temporal patterns were maintained and there were significant correlations in activity patterns across muscles. Surprisingly, there were no significant differences in the accuracy of evoked movements across the three main sources of activation (actual, within predicted, across predicted). Therefore, despite substantial differences

Figure 4. a) Set up for evoking movements in anesthetized monkeys. Monkeys were secured upright in infant car seat with implanted arm instrumented with movement sensors and free to move. Example set of stimulus pulses associated with actual activation patterns recorded for one of the test movements. Only above threshold, cathodic phases (shown positive for illustration purposes) of biphasic pulses are shown. All stimuli shown on the same scale. Example b) desired and c) evoked hand trajectories for stimulus pattern shown in a. Length of each scale bar represents 0.55 arm-length displacement. d) Time courses of desired and evoked hand displacement in x (anterior-posterior), y (medial-lateral), and z (vertical) directions for examples shown in b and c. e) Mean (SD) R^2 and RMSE values of evoked relative to desired movements, averaged across two test monkeys, nine test movements, and five test sessions using actual and predicted (within and across) activations as stimulus templates. Shuffled trials involved using actual activation signals but delivered to randomly selected muscles other than the target muscles. One-way ANOVA showed a significant effect ($p < 0.001$) of activation source on R^2 , with shuffled trials significantly less ($p < 0.05$ Dunn's post-hoc method) than the other three. There was a significant effect of activation source on RMSE ($p < 0.001$), with shuffled greater than within- and across-predicted ($p < 0.05$) but not actual.

in the quality of the activation predictions (Fig. 2), evoked movements based on these predictions were no worse than that using the actual activation signals. This implies that the dominant sources of error were those related to the fidelity with which delivered stimulus pulses replicate the targeted active states within muscles.

Some insight into the nature of those errors can be obtained by examining trajectories evoked by repeated delivery of the same stimulus pattern. Fig. 5a shows hand trajectories elicited upon 10 repeated trials in one session in one monkey. In this case, the evoked movements were consistent across trials. Despite this reliability, $< 50\%$ of the variability in the desired movement was accounted for in the evoked movement ($R^2 = 0.45 \pm 0.02$, mean $\pm$ SD). This implies that there exist systematic errors in the conversion of activation signals into desired muscle contraction with electrical stimulation. There can be several sources of such errors. For example, the relation between stimulus intensity and evoked force was obtained at one limb configuration only. However, as joint angles change during movement, the effectiveness of stimulation can vary due to changes in the distance between the electrode and motor nerve branches within a muscle ^[42]. Also, maximum stimulus intensity, identified as that which

Figure 5. Within and across session variability in evoked movements. **a)** evoked hand trajectories for 10 repeated trials of one test movement in monkey A and **b)** in monkey M. Trials were recorded within the same session with about 1 minute rest between trials. **c)** evoked hand trajectories for one test movement recorded in five separate sessions in monkey M. Cold colors indicate early trials, warm colors later trials. Diamond indicates shoulder location.

evoked maximal force prior to activation of other muscles (Fig. 3), was mapped to the peak EMG detected during a recording session. Such peak EMG values, however, may not actually reflect the maximal activation of the muscle, as implied in our mapping.

As shown in Fig. 5b, we also encountered less reliable responses to repeated stimulation within a session. In this case (obtained in the other monkey), there was a progressive reduction in the extent of movement with repeated trails, likely due to mounting muscle fatigue. A key muscle involved in these upper limb movements is the anterior deltoid, which supports the arm's weight against gravity. To enhance the magnitude of evoked forces and to help minimize fatigue, we used two electrodes in this muscle as independent stimulus sources, as we have done previously^[41,43]. Perhaps due to non-optimal electrode placements, stimulation may not have been effective in enlisting the full complement of muscle fibers in anterior deltoid (and other muscles). In such a case, muscle loads are carried by a smaller fraction of muscle fibers than would normally occur, increasing the susceptibility to fatigue.

Finally, variability in evoked responses across sessions to the same stimulation pattern could also be substantial (Fig. 5c). It might be that small differences in initial conditions (such

variations in body position or starting configuration of the limb) contributed to such variability. Indeed, remounting an animal in a test apparatus within a session has previously been shown to cause significant changes in isometric forces in response to nerve cuff stimulation ^[44]. It is also possible that changes over time in electrode position and degree of tissue encapsulation may have contributed to across session variability. However, our experiments involving evoked movements occurred at least 18 weeks following implant surgery, well past the ~8 weeks needed for stimulating electrodes to stabilize physically and electrically ^[44].

DISCUSSION

Here we have shown that machine learning can be used to predict EMG signals associated with complex upper limb movements with reasonable fidelity. When using such predicted (or actual) activation signals as stimulation templates, evoked movements matched desired trajectories well in some cases. Typically, however, evoked movements only moderately matched desired movements. Therefore, although promising, additional developments are needed to reduce errors before such an approach could be considered for use in paralyzed human patients.

Systematic errors and those leading to trial-by-trial and across-session errors could be addressed in different ways. For systematic errors, one promising approach might be to use the evoked movements (rather than those produced during voluntary movements) and the associated stimulation patterns (rather than EMG signals) to train machine learning algorithms ^[45]. In this way, the algorithms would directly learn the relation between stimulation and movement, effectively bypassing errors associated with stage 3 shown in Fig. 1a. Furthermore, this approach would be tailored to the idiosyncrasies of electrode placements and the particular deficits of each SCI individual (such as those associated with muscle atrophy, joint stiffness, etc.). Nevertheless, stimulation based on actual EMG patterns recorded in healthy subjects would likely be much more efficient than arbitrary stimulation for this approach. This is because it would use activity patterns already identified during the process of normal motor development

and stored in the CNS as natural solutions as to how multiple muscles are activated in the elaboration of complex movements.

For trial-by-trial and session-by-session errors, such as those due to fatigue, unexpected perturbations, and unaccounted body/limb position changes, it seems crucial to combine on-line feedback control [12,46,47] with the open loop control system used here. A significant challenge for feedback control in the present context, however, is enacting real-time adjustments of large numbers of muscles based on sensed discrepancies between actual and desired trajectories.

Finally, it is important to consider how a paralyzed individual would supply the desired movement trajectory as the input to a trained FES algorithm. The most notable approach would be to identify intended movements from neural recordings obtained through electrodes implanted in the brain [48-54]. Alternatively, various non-invasive approaches are being developed that enable individuals with high-level paralysis to convey movement intentions to robotic arms [e.g., 55-58]. Ultimately, such signals representing intended motion could be transformed into the appropriate patterns of muscle stimulation using the approach described here to produce diverse movements in a paralyzed limb rather than in a robotic device. Such an integrated system would reinstate a wide range of voluntary movements, would accrue health benefits associated with increased muscular activity, and would increase independence in paralyzed individuals.

METHODS

Surgical procedures

All procedures complied with NIH guidelines for the use of non-human primates in research and was approved by The University of Arizona's Institutional Animal Care and Use Committee. Under isoflurane anesthesia and sterile conditions, three adult male monkeys (*Macaca mulatta*, 10 – 13.5 kg, ages 7 – 11 yrs) had 29 muscles (see Suppl. Table 1) chronically implanted with intramuscular electrodes. These muscles included those that operate on the scapula, shoulder, elbow, wrist, and digits of the hand. We included three extrinsic digit muscles because they contribute to wrist flexion and extension. A total of 58 electrode leads (two for

each muscle to enable bipolar EMG recording), consisting of Teflon coated, multi-stranded stainless-steel wires (Cooner AS633, outside diameter 0.33 mm), were soldered to a 64-channel electrode interface board (Neuralynx EIB-36-16TT). The interface board was mounted within a protective cylindrical encasement and the entire implant was sterilized prior to surgery. In monkey E (trainer monkey), the encasement was attached to the exposed skull using dental acrylic anchored to bone screws. This encasement ultimately failed due to untreatable Methicillin-resistant *Staphylococcus aureus* infection possibly secondary to cytotoxic effects of the acrylic. In the other two monkeys (A and M, test monkeys), we used acrylic-less implants^[59], in which titanium baseplates with undersurfaces fabricated (Gray Matter, Inc.) to conform to the contours of the animals' crania were mounted to the skulls using titanium screws. The baseplates were then allowed 6 – 8 weeks to osseointegrate before electrode implantation surgery. The connector encasement was then secured with screws to a footprint in the baseplate during the implant surgery.

Leads originating from the encasement were routed posteriorly under the skin on the back of the skull to an incision between the scapulae. A ground/return electrode (1 cm diameter) was tucked into a subcutaneous pocket at this incision site (Fig. 1b). Electrode leads (identified by a four-band color code painted on the distal ends) were tunneled under the skin to incisions over target muscles. A low impedance, insulated, tungsten electrode was then used to deliver brief trains of 40-Hz intramuscular stimuli to evoke muscle contractions and identify optimal locations for implanting an electrode. The exteriorized lead was then cut to length, a small amount of insulation was removed from the end of the lead with a thermal wire stripper, and a gold anchor (~ 1 x 6 mm) was crimped to the lead. The anchor consisted of a crimp terminal pin (3922 Mill-Max) with the pin removed. Anchors were used to increase shear friction and surface area to help prevent electrode migration. The anchor was then fed into the opening of a custom-built insertion device that consisted of a 14-gauge needle with a narrow slot cut along on the length through which the lead passed. This device was then inserted alongside the tungsten electrode before the tungsten electrode was removed. Stimuli were then delivered through the insertion device (fully insulated except for the tip of the needle) to verify placement before deploying the anchor with a plunger pushed through the needle. The insertion device was removed, and stimuli

were then delivered through the lead to the anchor to verify robust contraction in the target muscle. This process was repeated for a second electrode inserted into the target muscle ~ 1 cm away from the other electrode. After closing incisions, and before recovering the animal from surgery, the implanted arm was immobilized. Immobilization was maintained for 7 – 14 days to allow fibrous tissue to encapsulate electrodes, helping to stabilize their placements ^[60].

Behavior

Prior to surgery, monkeys were trained to reach to food morsels with their right arm while seated in an open primate chair. The left arm was restrained with Velcro straps. Monkeys initiated each reach with their hand resting on a switch-instrumented start box. When the hand was on the box, a tone played, indicating to the monkey that a new trial could be initiated. After 1 – 2 s with the hand on the start box, the experimenter presented a small food morsel to the monkey. The monkey grasped the morsel, brought it to his mouth, and returned his hand to the start position. On each trial, food morsels were positioned at different locations within the reach space of the monkey. To increase the types of arm movements, on most trials the experimenter moved the food morsel through a complex trajectory. Monkeys invariably tracked the position of the experimenter's hand with their hand and grasped the morsel only when the motion of the experimenter's hand was halted. This tracking occurred without contact between the monkey's hand and the experimenter's hand. During the grasping phase, touch between the monkey and the experimenter's hand was avoided (as much as possible) to minimize unaccounted contact forces. Monkeys readily performed this task for 15 – 30 minutes until sated. These procedures were repeated over several sessions during which EMG and kinematic data were recorded. Data sampled in these sessions were used to train the ANN.

Kinematics and EMG

Electromagnetic sensors (Polhemus) were used to record (120 Hz/channel) six degrees-of-freedom (x, y, z positions, and roll, pitch, and yaw orientations) motion of the right hand and elbow. Small sensors (0.7 cm x 0.5 cm x 0.5 cm) were attached with elastic wrap to the back of the hand and lateral elbow, and with tape to the shoulder. The shoulder was used to represent the origin of a reference frame for measuring hand and elbow positions. An additional sensor was placed on the opposite shoulder to account for trunk rotations.

To record EMG activity, lightweight cables were attached to the connectors within the skull-mounted encasement. These cables were routed to a set of four eight-channel differential amplifiers (Neuralynx). EMG signals were amplified at a gain of 1000, band-pass filtered from 100 – 475 Hz, and digitally sampled at ~3000 Hz per channel using a computerized data acquisition system (Spike2). A TTL pulse generated by the Polhemus system was recorded by the data acquisition system to enable synchronization of kinematic with EMG data. In addition, the switch signal from the start box was recorded and used to indicate when the limb was in the start configuration.

Signal processing

In off-line processing (MATLAB, Mathworks), EMG signals were full-wave rectified, low-pass filtered (3 Hz), and down-sampled to 120 Hz/signal to match kinematic data. EMG amplitude was normalized to the maximum value recorded during a session. We refer such processed EMG signals as activation. Three-dimensional coordinates of the hand and elbow were lowpass filtered (6 Hz), calculated with respect to the right shoulder location, and normalized to arm length. Hand orientation data were converted from yaw-pitch-roll angular representations to quaternions to remove discontinuities. Velocities and accelerations were obtained through differentiation (and double differentiation) of position and orientation data using finite difference methods.

Artificial neural network (ANN)

Previously, we demonstrated the applicability of using ANNs to predict activation patterns from kinematics^[38, 39]. The structure of the ANN used here was a multilayer perceptron involving a feed-forward network created in the Neural Networks Toolbox of MATLAB. Inputs to the network were kinematic signals (3-dimensional positions, velocities, and accelerations of hand and elbow relative to the shoulder, and pitch, roll, and yaw angular orientations, velocities, and accelerations of the hand.) and outputs were the activation signals of the 29 muscles. The network was fully connected such that in every layer, all neural units received the outputs from units involved in the previous layer. The network possessed one hidden layer with 50 units. A tan-sigmoid was used as the activation function for each neural unit. In the output layer, all units were fully connected to all outputs using a linear activation function. Network initialization was

done with random weights and biases. Training used resilient backpropagation with gradient descent, momentum weight, and bias learning functions. Mean squared error was used as the performance function. Training data were obtained from one session for each monkey and included 8 minutes of concatenated complex reaches to food morsels (see Behavior above). Data obtained while monkeys were at rest in the start position were excluded. As shown in Suppl. Fig. 3, increasing the amount of training data beyond ~ 6 minutes yielded little improvement in predictions.

Data Analysis - Prediction of Activation from Kinematics

Nine trials from each of the two test monkeys (A and M) were selected to test the ability of the trained ANNs to predict activation patterns from kinematics. These test movements (see Fig. 2a) were selected from many trials as generally representative of the complexity, duration, and reach space explored across trials. These trials were not used in the training of the ANNs. The kinematics associated with these trials served as inputs to two ANN: that trained on data from the same monkey (within) and that trained on data obtained from a different monkey (monkey E) (across). Predictions of activation patterns for all 29 muscles were compared to the actual activation patterns recorded during the test movements using two measures: coefficient of variation (R^2) and the root mean squared error (RMSE –percentage of peak activation). Accurate predictions have R^2 value approaching one and RMSE values approaching zero.

Conversion of Activation to Stimulation

Predicted activation signals (based on training data obtained both from the same and different monkey) for the 9 test trials were transformed into amplitude modulated trains of 40 Hz stimulus pulses (biphasic, cathodic phase leading, 250 μ s/phase). Forty Hz stimuli were used because force-frequency curves obtained in the monkey indicated that this is the lowest frequency that produces near maximal force with minimal force fluctuations^[42]. For comparison, actual activation signals recorded during the test trials for both test monkeys were also transformed into trains of stimulus pulses.

To use activation as a template for FES, the relationship between stimulation intensity and evoked isometric force was first determined for each electrode using an approach we have used previously^[43]. Monkeys were sedated in their home cage with Ketamine HCL (10–15 mg

kg⁻¹ IM) and transported to the procedure room. Atropine (0.04 mg kg⁻¹ IM) was given to reduce hyper-salivation common with Ketamine sedation. Carprofen (2.2 mg kg⁻¹ SQ) was also given to reduce inflammation associated with endotracheal intubation. An intravenous catheter was placed in the saphenous or cephalic vein to deliver lactated Ringers (5–10 ml/kg/h) to maintain hydration. Anesthesia was induced with isoflurane via mask insufflation. Following induction, an endotracheal tube was inserted to maintain airway patency and deliver anesthesia (1–2% isoflurane in 100% oxygen, ~1 l min⁻¹). An esophageal thermometer measured core temperature and was maintained at ~36.7 °C via a forced warm air blanket, bubble wrap, and blankets placed over the torso. Heart rate, respiratory rate, electrocardiogram, SpO₂, end-tidal CO₂, and non-invasive blood pressure were monitored throughout the experiment.

Monkeys were placed into a modified infant car seat in a seated position (Fig. 3a). A neonatal cervical collar was used to maintain the head in an upright position. The cervical collar was secured to the car seat with cable ties. Straps situated midway between the neck and shoulder and across the torso secured the animal to the chair. An isometric transducer (JR3) was fixed to the wrist/hand (Fig. 3b), forearm, and upper arm to record the 3-dimensional components of evoked isometric forces associated with muscles acting on each of those segments. Limb segments proximal to the tested segments were immobilized with Velcro straps.

Stimuli were delivered through the skull-mounted connector to each electrode separately using a programmable multi-channel stimulator (STG4008, MultiChannel Systems). Stimuli were single 500 μs biphasic rectangular pulses delivered at 1 second intervals, from 0.2 mA to 2.0 mA in 0.2 mA steps, then from 3 mA to 32 mA in 1 mA steps. We used single stimuli and evoked twitches rather than responses to tetanic stimulation to reduce the time monkeys were under anesthesia and minimize the possibility of fatigue. Previous work has shown that the normalized relation between stimulus intensity and force is practically the same for twitches and tetanic responses^[61]. For scapular, shoulder, and elbow muscles, the return was the subcutaneous disk electrode situated between the scapulae, while for wrist and digit muscles, the other electrode situated in the same muscle acted as the return.

In postprocessing (using MATLAB), the force signals were bandpass filtered (0.05 – 20 Hz). The signal from a breathing sensor, worn during the experiments, was used to trigger and

average artifacts in the force signals recorded during periods without stimulation associated with respiration. This template was then subtracted from the force signals, aligned to each breath cycle during stimulation, to help remove this artifact. Resultant twitch forces (e.g., Figs. 3c,e) were visually examined and the minimum current that elicited a just noticeable response was identified. To identify the maximum current, the lowest stimulus intensity needed to achieve ~95% of the maximum resultant force and which did not cause a clear change in force direction (e.g., Fig. 3d) was used as the upper limit for stimulation. A logistic curve was fit to the peak resultant forces between the minimum and maximum currents. Because the relation between activation and isometric force is practically linear for most muscles, we substituted activation for force in these logistic curves. For muscles that used the subcutaneous disk electrode as the return electrode for stimulation (i.e., scapular, shoulder, elbow), we selected the one electrode of the two whose data best fit the logistic curve and that exhibited a reasonably wide range of stimulus intensities over which force was modulated. The inverse of the logistic curve identified for the selected electrode (or for the bipolar pair in forearm muscles) was then used to convert activation at each time sample into the associated amplitude of biphasic current pulses running at 40 Hz. We have previously shown that stimulation using this approach reproduces both isometric and free movement patterns with good accuracy in a simple joint system ^[40]. Stimulus pulses associated with actual, within-predicted, and across-predicted activation patterns for the 9 test movements were stored in computer memory for later payout.

Electrical stimulation to evoke movements

As described above (Conversion of Activation to Stimulation), test monkeys were sedated and secured in an infant car seat for experiments to evoke movements in the paralyzed upper limb. The unrestrained implanted arm was instrumented with Polhemus movement sensors, and the hand was placed in a position like the start position for voluntary reaching. A custom MATLAB script controlled the payout of stimulus patterns associated with predicted and actual activations of 29 muscles using four, 8-channel programmable stimulators (STG4008, MultiChannel Systems). As we have done previously ^[41,43], to help minimize fatigue and enhance the strength of three shoulder muscles (anterior, middle, posterior deltoids) crucial for the types of movements involved, we used two independent sources of stimulation, one arising

from each of the two electrodes implanted in those muscles. Prior to delivery of stimuli associated with the test movements, brief trains of stimuli were delivered separately to each muscle to verify evoked responses in the target muscles. In one monkey, one of the two implanted anterior deltoid electrodes produced only weak responses to strong stimulation. Therefore, in each stimulation session in this monkey, we inserted a percutaneous, temporary, hook-wire electrode into anterior deltoid to serve as the second source of stimulation.

In each of five sessions, 27 stimulus patterns (9 test movements x 3 sources of activation signals [actual, within-, across-predicted]) were delivered to the paralyzed upper limb and the evoked movements recorded. About one minute of rest was provided between each trial. The order of stimulus patterns was varied across sessions. At least one week separated consecutive sessions for each monkey and these experiments were initiated 18 – 22 weeks following the initial implant surgeries in the two monkeys. RMSE and R^2 values were calculated for the x, y, and z positions of the hand for each evoked movement relative to the desired trajectory. In one session for each of the two test monkeys, after completing the set of 27 stimulation trials, one pattern was replayed 10 times with ~ 1 minute between trials to evaluate the reliability of stimulation. In addition, in one other session for each of the test monkeys, stimulus patterns based on actual activation signals were shuffled such that stimuli were delivered to randomly selected muscles rather than to the target muscles. This was repeated nine times, once for each test movement.

ACKNOWLEDGEMENTS

This work was funded by NIH grant NS102259. We are grateful for the outstanding support provided by Lierin Cox, Chelsea Mahnk, Baldwin Goodell, and Drs. CJ Doane, Charles Gray, and David Margolis.

REFERENCES

1. Fridén, J. & Gohritz, A. Tetraplegia management update. *The Journal of Hand Surgery* **40**, 2489–2500 (2015).
2. Assinck, P., Duncan, G. J., Hilton, B. J., Plemel, J. R. & Tetzlaff, W. Cell transplantation therapy for spinal cord injury. *Nat Neurosci* **20**, 637–647 (2017).
3. Rao, J.-S. *et al.* NT3-chitosan enables de novo regeneration and functional recovery in monkeys after spinal cord injury. *Proc. Natl. Acad. Sci. U.S.A.* **115**, E5595–E5604 (2018).
4. Rosenzweig, E. S. *et al.* Restorative effects of human neural stem cell grafts on the primate spinal cord. *Nature Medicine* **24**, 484–490 (2018).
5. Tran, A. P., Warren, P. M. & Silver, J. The biology of regeneration failure and success after spinal cord injury. *Physiological Reviews* **98**, 881–917 (2018).
6. Álvarez, Z. *et al.* Bioactive scaffolds with enhanced supramolecular motion promote recovery from spinal cord injury. *Science* **374**, 848–856 (2021).
7. Awad, L. N., Esquenazi, A., Francisco, G. E., Nolan, K. J. & Jayaraman, A. The ReWalk ReStore™ soft robotic exosuit: a multi-site clinical trial of the safety, reliability, and feasibility of exosuit- augmented post-stroke gait rehabilitation. 1–11 (2020). doi:10.1186/s12984-020-00702-5.
8. Hong, E. *et al.* Mobility Skills With Exoskeletal-Assisted Walking in Persons With SCI: Results From a Three Center Randomized Clinical Trial. *Front. Robot. AI* **7**, 93 (2020).
9. Koljonen, P. A. *et al.* Outcomes of a multicenter safety and efficacy study of the SuitX Phoenix Powered Exoskeleton for ambulation by patients with spinal cord injury. *Front. Neurol.* **12**, 689751 (2021).
10. Chau, C., Barbeau, H. & Rossignol, S. Effects of intrathecal alpha1- and alpha2-noradrenergic agonists and norepinephrine on locomotion in chronic spinal cats. *Journal of Neurophysiology* **79**, 2941–2963 (1998).
11. Courtine, G. *et al.* Transformation of nonfunctional spinal circuits into functional states after the loss of brain input. *Nat Neurosci* 1–12 (2009). doi:10.1038/nn.2401
12. Wenger, N. *et al.* Closed-loop neuromodulation of spinal sensorimotor circuits controls refined locomotion after complete spinal cord injury. *Sci Transl Med* **6**, 255ra133 (2014).
13. Holinski, B. J. *et al.* Intraspinal microstimulation produces over-ground walking in anesthetized cats. 1–18 (2016). doi:10.1088/1741-2560/13/5/056016
14. Capogrosso, M. *et al.* A brain–spine interface alleviating gait deficits after spinal cord injury in primates. *Nature* **539**, 284–288 (2016).

15. Angeli, C. A. *et al.* Recovery of Over-Ground Walking after Chronic Motor Complete Spinal Cord Injury. *N Engl J Med* **379**, 1244–1250 (2018).
16. Wagner, F. B. *et al.* Targeted neurotechnology restores walking in humans with spinal cord injury. *Nature* 1–29 (2018). doi:10.1038/s41586-018-0649-2
17. Gill, M. L. *et al.* Neuromodulation of lumbosacral spinal networks enables independent stepping after complete paraplegia. *Nature Medicine* 1–11 (2018). doi:10.1038/s41591-018-0175-7
18. Greiner, N. *et al.* Recruitment of upper-limb motoneurons with epidural electrical stimulation of the cervical spinal cord. *Nature Communications* **12**, 435–19 (2021).
19. Kilgore, K. L., Peckham, P. H., Thrope, G. B., Keith, M. W. & Gallaher-Stone, K. A. Synthesis of hand grasp using functional neuromuscular stimulation. *IEEE Trans. Biomed. Eng.* **36**, 761–770 (1989).
20. Hoshimiya, N., Naito, A., Yajima, M. & Handa, Y. A multichannel FES system for the restoration of motor functions in high spinal cord injury patients: a respiration-controlled system for multijoint upper extremity. *IEEE Trans. Biomed. Eng.* **36**, 754–760 (1989).
21. Peckham, P. H. *et al.* Efficacy of an implanted neuroprosthesis for restoring hand grasp in tetraplegia: A multicenter study. *Archives of Physical Medicine and Rehabilitation* **82**, 1380–1388 (2001).
22. Kilgore, K. L. *et al.* An implanted upper-extremity neuroprosthesis using myoelectric control. *The Journal of Hand Surgery* **33**, 539–550 (2008).
23. Memberg, W. D. *et al.* Implanted neuroprosthesis for restoring arm and hand function in people with high level tetraplegia. *Archives of Physical Medicine and Rehabilitation* **95**, 1201–1211.e1 (2014).
24. Triolo, R. J. *et al.* Longitudinal Performance of a Surgically Implanted Neuroprosthesis for Lower-Extremity Exercise, Standing, and Transfers After Spinal Cord Injury. *YAPMR* **93**, 896–904 (2012).
25. Forssberg, H., Grillner, S. & Halbertsma, J. The locomotion of the low spinal cat. I. Coordination within a hindlimb. *Acta Physiologica* **108**, 269–281 (1980).
26. Barbeau, H. & Rossignol, S. Recovery of locomotion after chronic spinalization in the adult cat. *Brain Research* **412**, 84–95 (1987).
27. Wirz, M., Colombo, G. & Dietz, V. Long term effects of locomotor training in spinal humans. *J. Neurol. Neurosurg. Psychiatr.* **71**, 93–96 (2001).
28. Delivet-Mongrain, H., Dea, M., Gossard, J.-P. & Rossignol, S. Recovery of locomotion in cats after severe contusion of the low thoracic spinal cord. *Journal of Neurophysiology* **123**, 1504–1525 (2020).
29. Borgens, R. B., Blight, A. R. & McGinnis, M. E. Behavioral recovery induced by applied electric fields after spinal cord hemisection in guinea pig. *Science* **238**, 366–369 (1987).
30. Carmel, J. B., Berrol, L. J., Brus-Ramer, M. & Martin, J. H. Chronic Electrical Stimulation of the Intact Corticospinal System after Unilateral Injury Restores Skilled Locomotor Control and Promotes Spinal Axon Outgrowth. *Journal of Neuroscience* **30**, 10918–10926 (2010).
31. Jack, A. S., Hurd, C., Martin, J. & Fouad, K. Electrical Stimulation as a Tool to Promote Plasticity of the Injured Spinal Cord. *J Neurotrauma* **37**, 1933–1953 (2020).

32. Inanici, F., Brighton, L. N., Samejima, S., Hofstetter, C. P. & Moritz, C. T. Transcutaneous Spinal Cord Stimulation Restores Hand and Arm Function After Spinal Cord Injury. *IEEE Trans. Neural Syst. Rehabil. Eng.* **29**, 310–319 (2021).
33. Jo, H. J. & Perez, M. A. Corticospinal-motor neuronal plasticity promotes exercise-mediated recovery in humans with spinal cord injury. *Brain* **143**, 1368–1382 (2020).
34. Bichenbach J *et al.* *International perspectives on spinal cord injury*. World Health Organization (2013).
35. Anderson, K. D. Targeting recovery: priorities of the spinal cord-injured population. *J Neurotrauma* **21**, 1371–1383 (2004).
36. Seifert, H. M. & Fuglevand, A. J. Restoration of movement using functional electrical stimulation and Bayes' theorem. *J. Neurosci.* **22**, 9465–9474 (2002).
37. Anderson, C. V. & Fuglevand, A. J. Probability-Based Prediction of Activity in Multiple Arm Muscles: Implications for Functional Electrical Stimulation. *Journal of Neurophysiology* **100**, 482–494 (2008).
38. Johnson, L. A. & Fuglevand, A. J. Evaluation of probabilistic methods to predict muscle activity: implications for neuroprosthetics. *J. Neural Eng.* **6**, 055008 (2009).
39. Tibold, R. & Fuglevand, A. J. Prediction of muscle activity during loaded movements of the upper limb. *Journal of NeuroEngineering and Rehabilitation* **12**, 6–12 (2015).
40. Johnson, L. A. & Fuglevand, A. J. Mimicking muscle activity with electrical stimulation. *J. Neural Eng.* **8**, 016009 (2011).
41. Buckmire, A. J., Arakeri, T. J., Reinhard, J. P. & Fuglevand, A. J. Mitigation of excessive fatigue associated with functional electrical stimulation. *J. Neural Eng.* **15**, 066004–12 (2018).
42. Crago, P. E., Peckham, P. H. & Thrope, G. B. Modulation of muscle force by recruitment during intramuscular stimulation. *IEEE Trans. Biomed. Eng.* **27**, 679–684 (1980).
43. Buckmire, A. J., Lockwood, D. R., Doane, C. J. & Fuglevand, A. J. Distributed stimulation increases force elicited with functional electrical stimulation. *J. Neural Eng.* **15**, 026001–15 (2018).
44. Grill, W. M. & Mortimer, J. T. Stability of the input-output properties of chronically implanted multiple contact nerve cuff stimulating electrodes. *IEEE Trans Rehabil Eng* **6**, 364–373 (1998).
45. Lujan, J. L. & Crago, P. E. Automated Optimal Coordination of Multiple-DOF Neuromuscular Actions in Feedforward Neuroprostheses. *IEEE Trans. Biomed. Eng.* **56**, 179–187 (2009).
46. Crago, P. E., Nakai, R. J. & Chizeck, H. J. Feedback regulation of hand grasp opening and contact force during stimulation of paralyzed muscle. *IEEE Trans. Biomed. Eng.* **38**, 17–28 (1991).
47. Blana, D., Kirsch, R. F. & Chadwick, E. K. Combined feedforward and feedback control of a redundant, nonlinear, dynamic musculoskeletal system. *Med. Biol. Eng. Comput.* **47**, 533–542 (2009).
48. Wessberg, J. *et al.* Real-time prediction of hand trajectory by ensembles of cortical neurons in primates. *Nature* **408**, 361–365 (2000).
49. Taylor, D. M., Tillery, S. I. H. & Schwartz, A. B. Direct cortical control of 3D neuroprosthetic devices. *Science* **296**, 1829–1832 (2002).

50. Moritz, C. T., Perlmutter, S. I. & Fetz, E. E. Direct control of paralysed muscles by cortical neurons. *Nature* **456**, 639–642 (2008).
51. Hochberg, L. R. *et al.* Reach and grasp by people with tetraplegia using a neurally controlled robotic arm. *Nature* **485**, 372–375 (2012).
52. Collinger, J. L. *et al.* High-performance neuroprosthetic control by an individual with tetraplegia. *Lancet* **381**, 557–564 (2013).
53. Ethier, C., Oby, E. R., Bauman, M. J. & Miller, L. E. Restoration of grasp following paralysis through brain-controlled stimulation of muscles. *Nature* **485**, 368–371 (2013).
54. Ajiboye, A. B. *et al.* Restoration of reaching and grasping movements through brain-controlled muscle stimulation in a person with tetraplegia: a proof-of-concept demonstration. *Lancet* **389**, 1821–1830 (2017).
55. Corbett, E. A., Kording, K. P. & Perreault, E. J. Real-Time Evaluation of a Noninvasive Neuroprosthetic Interface for Control of Reach. *IEEE Trans. Neural Syst. Rehabil. Eng.* **21**, 674–683 (2013).
56. Williams, M. R. & Kirsch, R. F. Evaluation of head orientation and neck muscle EMG signals as three-dimensional command sources. *Journal of NeuroEngineering and Rehabilitation* **12**, 25–16 (2015).
57. Marković, M., Dosen, S., Popovic, D., Graimann, B. & Farina, D. Sensor fusion and computer vision for context-aware control of a multi degree-of-freedom prosthesis. 1–16 (2015). doi:10.1088/1741-2560/12/6/066022
58. Wöhle, L. & Gebhard, M. Towards Robust Robot Control in Cartesian Space Using an Infrastructureless Head- and Eye-Gaze Interface. *Sensors* **21**, 1798 (2021).
59. Blonde, J. D. *et al.* Customizable cap implants for neurophysiological experimentation. *Journal of Neuroscience Methods* **304**, 103–117 (2018).
60. Bhadra, N. & Mortimer, J. T. Extraction forces and tissue changes during explant of CWRU-type intramuscular electrodes from rat gastrocnemius. *Ann Biomed Eng* **25**, 1017–1025 (1997).
61. Fisher, L. E., Tyler, D. J., Anderson, J. S. & Triolo, R. J. Chronic stability and selectivity of four-contact spiral nerve-cuff electrodes in stimulating the human femoral nerve. *J. Neural Eng.* **6**, 046010–10 (2009).

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [SupplementalMaterials.pdf](#)
- [20201215RightTargetcaptioned.mp4](#)