On the Impact of Vessel Wall Stiffness on Quantitative Flow Dynamics in a Synthetic Model of the Thoracic Aorta

Judith Zimmermann (✉ juzim@stanford.edu)

Stanford University, Department of Radiology, Stanford, 94305, USA

Michael Loecher

Stanford University

Fikunwa Kolawole

Stanford University

Kathrin Bäumler

Stanford University

Kyle Gifford

Stanford University

Seraina Dual

Stanford University

Marc Levenston

Stanford University

Alison Marsden

Stanford University

Daniel Ennis

Stanford University

Research Article

Keywords: pulse wave velocity (PWV), rigid stiffnesses, 4D-flow MRI data, hemodynamics

Posted Date: January 4th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-134591/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published at Scientific Reports on March 23rd, 2021. See the published version at <https://doi.org/10.1038/s41598-021-86174-6>.

On the Impact of Vessel Wall Stiffness on Quantitative Flow Dynamics in a Synthetic Model of the Thoracic Aorta

Judith Zimmermann^{1,2,*}, Michael Loecher^{1,3}, Fikunwa O. Kolawole⁴, Kathrin Bäuml¹, Kyle Gifford¹, Seraina A. Dual¹, Marc Levenston⁴, Alison L. Marsden^{5,6,7}, and Daniel B. Ennis^{1,3,7}

¹Stanford University, Department of Radiology, Stanford, 94305, USA

²Technical University of Munich, Department of Computer Science, Garching, 85748, Germany

³Veterans Affairs Health Care System, Division of Radiology, Palo Alto, 94304, USA

⁴Stanford University, Department of Mechanical Engineering, Stanford, 94305, USA

⁵Stanford University, Department of Bioengineering, Stanford, 94305, USA

⁶Stanford University, Department of Pediatrics, Stanford, 94305, USA

⁷Stanford University, Cardiovascular Institute, Stanford, 94305, USA

*Corresponding author: Judith Zimmermann, Lucas MRI Center, 1201 Welch Road, Stanford, CA 94305, USA; E-mail: juzim@stanford.edu.

ABSTRACT

Aortic wall stiffening is a predictive marker for morbidity in hypertensive patients. Arterial pulse wave velocity (PWV) correlates with the level of stiffness and can be derived using non-invasive 4D-flow magnetic resonance imaging (MRI). The objectives of this study were twofold: to develop subject-specific thoracic aorta models embedded into an MRI-compatible flow circuit operating under controlled physiological conditions; and to evaluate how a range of aortic wall stiffness impacts 4D-flow-based quantification of hemodynamics, particularly PWV. Three aorta models were 3D-printed using a novel photopolymer material at two compliant and one nearly rigid stiffnesses and characterized via tensile testing. Luminal pressure and 4D-flow MRI data were acquired for each model and cross-sectional net flow, peak velocities, and PWV were measured. In addition, the confounding effect of temporal resolution on all metrics was evaluated. Stiffer models resulted in increased systolic pressures (112, 116, and 133 mmHg), variations in velocity patterns, and increased peak velocities, peak flow rate, and PWV (5.8 to 7.3 m/s). Lower temporal resolution (20 ms down to 62.5 ms per image frame) impacted estimates of peak velocity and PWV (7.31 down to 4.77 m/s). Using compliant aorta models is essential to produce realistic flow dynamics and conditions that recapitulated *in vivo* hemodynamics.

Introduction

Aortic wall stiffness is a strong predictor for all-cause and cardiovascular morbidity in patients with systemic arterial hypertension^{1–3}. Model-based studies estimate that a total of 1.56 billion people worldwide may be affected by systemic arterial hypertension by 2025^{4,5}. Consequently, monitoring aortic wall stiffness has become increasingly important and could guide treatment strategies and prevention of systemic arterial hypertension. Aortic wall stiffening is linked to an increase of pulse wave velocity (PWV) - the velocity at which the blood pressure pulse travels through the circulatory system. For a vessel of constant diameter, this relationship is modeled by the Moens-Korteweg equation:

$$PWV = \sqrt{\frac{Eh}{2\rho r}} \quad (1)$$

where E is the elasticity modulus, ρ blood density, h wall thickness, and r vessel radius. Several PWV measurement technologies exist, with the carotid-femoral PWV (cfPWV) approach considered the clinical gold-standard⁶. cfPWV is approximated using the foot-to-foot temporal shifts of two signal waveforms (e.g. Doppler) recorded transcutaneously at the common carotid and femoral artery. The technical challenges associated with cfPWV measurements, such as carotid-femoral path length measurement inaccuracies or difficulties in transcutaneous signal recording, have limited broader adoption.

Non-invasive 4D-flow magnetic resonance imaging (MRI) provides three-dimensional (3D) and time-resolved velocity vector maps that serve as basis for image-based quantitative flow characterization^{7,8}. In particular, 4D-flow PWV calculations

use a similar transit-time over fixed distance approach as in conventional cfPWV measurements, but in addition exploits the volumetric imaging data⁹ and analyzes temporal shifts in flow rate waveforms extracted at numerous cross-sectional image planes along the aorta. This approach enables both regionally specific and more robust PWV estimates compared to two-point methods. In addition, the 3D anatomical image information is used to precisely measure the path length. Previous studies suggest low intra- and inter-observer variability and moderate test-retest performance^{10,11}. The same studies affirm that PWV increases with age and in the presence of aortic atherosclerosis.

The otherwise lengthy clinical *in vivo* 4D-flow MRI scan must be accelerated using parallel imaging, compressed sensing, or fast readout techniques^{12–17}. Each of these techniques, however, trades-off spatio-temporal image resolution and signal-to-noise ratio (SNR), both of which impact flow quantification accuracy. 4D-flow MRI sampling requirements to report robust PWV values are missing.

In vitro 4D-flow MRI using subject-specific synthetic aorta models connected to a cardiovascular flow pump enables prolonged imaging, thereby allowing optimal image quality. Moreover, *in vitro* setups enable studying flow dynamics under controllable conditions. In particular, we can program physiological flow waveforms, tune flow volume splits via outlet resistance control, and tune systemic pulse pressure via integration of capacitor elements. The majority of previous studies simplify their setup using rigid wall materials, which neglects the compliant nature of the human vasculature^{18–20}. A limited number of studies embed compliant models, but do not report on how the compliance of the model compares to the human aorta^{21,22}.

Novel 3D-printing technology permits building models with realistic and varying compliance which we seek to leverage. Herein, this work exploits *in vitro* 4D-flow MRI with realistic and compliant models of the thoracic aorta to study quantitative flow dynamics. The two objectives of this study were: (1) to demonstrate feasibility of deploying compliant 3D-printed subject-specific aorta models in an MRI-compatible flow circuit setup that matches physiological flow and pressure conditions; and (2) to evaluate the impact of wall stiffness variations on cross-sectional flow metrics and PWV.

Methods

Compliant aorta models

An *in vivo* chest 4D-flow MRI dataset was acquired from a healthy subject (50 y/o, male) using a protocol that was in accordance with relevant guidelines and regulations, and approved by Stanford University Institutional Review Board. Informed consent was obtained from the subject prior to imaging. The 4D-flow MRI magnitude image was used to generate a subject-specific polygon mesh model of the thoracic aortic wall including the brachiocephalic trunk, left common carotid, and left subclavian artery. The wall domain mesh generation consisted of: (1) binary segmentation of the aortic lumen; (2) polygon surface meshing of lumen mask (edge size = 0.8 mm); (3) extrusion of the mesh nodes in the normal direction to define the outer wall surface mesh; (4) boolean differencing of the outer wall and lumen surface mesh. The resulting wall thickness ($h_{wall} = 2$ mm) was within the reported range of the average wall thickness of the human adult aorta²³. The mesh model was extended by cylindrical caps (length = 20 mm) at the ascending aorta inlet and at the four outlets to enable connection to customized barbed model-tubing transition elements. These steps were performed using SimVascular²⁴ and Meshmixer (Autodesk) open software tools.

A photopolymerization 3D printer (J735 PolyJet, Stratasys) with novel printing materials (Agilus30 and VeroClear, Stratasys) was used to manufacture two compliant and one nearly rigid model of the subject-specific aorta geometry (referred to as M_{c1} , M_{c2} , and M_r). Printed models were finished with a thin coating (Bectron, Elantas) to prevent fluid absorption (Fig. 1g). To characterize the material stiffness, standardized dumbbell shaped samples (ASTM D412 type A) were 3D-printed in the same batch and using the same material blend as the aorta models.

Uniaxial tensile testing (Instron 5848 Microtester, 10-KN load cell) was performed on three dumbbell samples per material elasticity. Samples were pre-conditioned with five loading and unloading cycles to 10 % peak strain followed by a sixth measurement cycle to 50 % peak strain. Testing was done at ambient conditions with a strain rate of 25 %/sec. This rate corresponds to the upper loading rate limit on the aorta models when embedded in the flow circuit, which was assessed by analyzing the dynamic wall circumference in the imaging data. 3D-printing direction anisotropy was evaluated by varying the sample's orientation on the print bed.

Flow circuit setup

An MRI-compatible flow circuit setup (Fig. 1a) was engineered to enable *in vitro* 4D-flow imaging of the aorta models under physiological and controllable flow and pressure conditions. The inlets and outlets of the models were sealed to tubing via custom-fit 3D-printed barbed connectors with tapered transitions and then embedded into a ballistics gel block (Fig. 1b, ClearBallistics). The gel block provided a fixed positioning reference and had a short T1 relaxation time, which facilitated using it as static "tissue" for eddy current induced phase offset correction. Aortic model and gel block were placed inside an

enclosed box that connects through five box-mounted flow ports to both the pump unit (CardioFlow 5000 MR, Shelley Medical Imaging Technologies) and to the fluid reservoir which supplies the pump unit via a return flow path.

The subject's previously performed MRI exam included aortic flow measurements from which the flow rate waveform was derived and programmed to the pump unit. To this end, the original waveform was spline-interpolated, down-scaled to meet the pump's peak flow rate limit (300 mL/s), and discretized ($\Delta t = 10$ ms) over a cardiac cycle length of $RR = 1000$ ms (heart rate 60 bpm). The resulting stroke volume was 71.2 mL and total flow was 4.3 L/min.

Flow volume splits between the DAo outlet and the merged arch branches were controlled via the ratio of clamping at $R1_{dist}$ and $R2_{dist}$. Flow split tuning was assessed using an ultrasonic flow probe (ME-PXL14, Transonic) which clamped-on at the DAo outlet to a data acquisition system (DAQ) via a flow module (TS410, Transonic). Based on preliminary bench top tests, we targeted a flow volume split of 70/30 for model M_{c1} .

Pulse pressure was controlled using two capacitance elements at the DAo outlet ($C1$, see Fig. 1a) and at the merged arch branches ($C2$). The capacitors were designed as cylindrical towers with sealed air compression chambers in which the enclosed air volume dictates the amount of downstream capacitance. Pressure transducers (Micro-Tip SPR-350S, Millar) were inserted at the model inlet and descending aorta outlet, and pressure signal was received at the DAQ through a bridge amplifier front-end (FE224 Quad Bridge, ADInstruments). The mean arterial pressure was elevated by increasing the system resistance through clamping distal to the capacitance elements ($R1_{dist}$, $R2_{dist}$). We defined 110-120 mmHg and 70-80 mmHg as target systolic (P_{sys}) and diastolic (P_{dias}) pressure ranges for model M_{c1} .

For the three models, identical inflow conditions were programmed, whereas pressure and flow split tuning was performed with model M_{c1} only. Subsequently, models M_{c2} and M_r were embedded under the identical periphery without re-tuning system capacitance or resistance. All DAQ data analysis was performed using dedicated DAQ software (LabChart 8, ADInstruments).

Imaging experiments

Imaging experiments were performed using a 3T MRI scanner (Skyra, Siemens Healthineers) with a 32-channel spine and a 18-channel chest coil. We used a total fluid volume of seven liters (glycerol-water mixture with ratio = 40/60) with T_1 -shortening contrast agent (ferumoxytol, concentration = 0.75 mL/Liter) for increased signal-to-noise ratio (SNR). Protocol steps were as follows: (1) set up "fluid-empty" circuit on the MRI scanner table; (2) flush and de-bubble all fluid lines (steady flow); (3) record and tune pressures and flow splits (pulsatile flow); (4) remove pressure and flow transducers before moving setup to MRI iso-center; (5) run high-resolution 3D spoiled gradient echo (SPGR) acquisition (steady flow, 71.2 ml/s, Fig. 1c, 1d, 1e); (5) run 4D-flow acquisitions (three series, all pulsatile flow); (6) run 2D-PC and 2D-cine-GRE at pre-defined landmark slices (14 series, all pulsatile flow). The pump trigger signal was used for retrospective cardiac gating (direct input to scanner) and to later synchronize signals from pressure and ultrasonic recordings (direct input to DAQ). Image acquisition time per model was 1 hour 45 minutes, and total end-to-end experiment time (including setting up and swapping models) was 10 hours.

2D-PC and 2D-cine-GRE MRI

Two-dimensional (2D) imaging through lumen cross-section was performed at the following landmarks (Fig. 1f): ascending aorta inlet ('inlet'), ascending aorta ('AAo'), arch just proximal to brachiocephalic trunk ('BCT'), and distal to left subclavian artery ('LSA'), mid-descending aorta ('DAo'), descending aorta outlet ('outlet'), and three arch branches ('b1', 'b2', 'b3'). 2D MRI sequences included: (1) 2D cine gradient echo (2D-cine-GRE) with in-plane resolution $0.9 \text{ mm} \times 0.9 \text{ mm}$, slice thickness = 6 mm, FoV = $240 \text{ mm} \times 150 \text{ mm}$, TE/TR = $3/4.75 \text{ ms}$, flip angle = 7 degree, averages = 2, retrospective gating, number of temporal frames = 50 (frame length = 20 ms), no parallel imaging acceleration; and (2) 2D phase-contrast (2D-PC) with $V_{enc} = 90\text{-}120 \text{ cm/s}$, in-plane resolution $1.1 \text{ mm} \times 1.1 \text{ mm}$, slice thickness = 6 mm, FoV = $220 \text{ mm} \times 123 \text{ mm}$, TE/TR = $3/21 \text{ ms}$, flip angle = 25 degree, averages = 2, retrospective gating, number of temporal frames = 50 (frame length = 20 ms), no parallel imaging acceleration.

4D-flow MRI

We used a conventional 4D-flow sequence with Cartesian k-space sampling, a velocity encoding range (V_{enc}) of 120 cm/s, and repeated scans at three temporal resolutions (20 ms, 40 ms, 62.5 ms), leading to a total of nine datasets (Tab. 1). To minimize phase offsets and to improve geometric fidelity, the image data were corrected for Maxwell terms (during reconstruction), gradient non-linearity, and eddy current (both post-reconstruction). Distortion correction due to gradient non-linearity was implemented as described by Markl et al.²⁵ 3D phase images were corrected for eddy current effects via linear fitting of 3D offset maps through the ballistics gel image region. No phase unwrapping was required. All subsequent image-based analyses was performed using software tools available as part of MEVISFlow package (Fraunhofer Institute for Digital Medicine).

Image analysis

2D-PC and 2D-cine-GRE analysis

For all cross-sectional landmarks (Fig. 1f) aorta lumen contours were manually drawn in the first cardiac frame ($t = 0$) of the 2D-cine-GRE slices and tracked through all subsequent frames ($t = 1 - 49$) using a phase-based motion tracking algorithm as

	high temp-res	baseline temp-res	low temp-res
FoV [mm ³]	360x260x100	360x260x100	360x260x100
Acquisition matrix	144x104x40	144x104x40	144x104x40
Spatial resolution [mm]	2.5 isotropic	2.5 isotropic	2.5 isotropic
Lines per segment	1	2	3
Reconstructed frames	50	25	16
Temporal resolution [ms]	20	40	62.5
TE/TR [ms]	2.8/5.2	2.8/5.2	2.8/5.2
V_{enc} [cm/s]	120	120	120
Flip angle [degree]	15	15	15
BW [Hz/px]	451	451	451
Scan time [mm:ss]	42:40	21:20	14:40

Table 1. 4D-flow MRI sequence parameters. Three data sets were acquired with each model, resulting in a total of nine datasets available for analysis. Variations in effective temporal resolution were controlled by the number of acquired k-space lines per segment, and the number of reconstructed cardiac frames was adapted accordingly. FOV, field of view; TE, echo time; TR, repetition time; V_{enc} , velocity encoding range; BW, bandwidth.

described by Tautz et al.²⁶. Lumen expansion was assessed by calculating relative contour area change over time. Identical contours were used to assess inflow conditions and net flow splits based on the acquired 2D-PC data.

4D-flow analysis

For each aorta model, we segmented a 3D aorta lumen mask $m[\mathbf{x}]$ in the 3D SPGR image data using an automated 3D region growing algorithm, and subsequently derived the lumen centerline (c_{full}) using a skeleton approach²⁷ on the 3D binary mask. The centerline was used to define cross-sectional planes for 4D-flow parameter quantification at landmarks identical to those defined as part of the 2D acquisitions (Fig. 1b) and to extract equidistant flow waveforms used for the PWV computation.

Based on the 4D-flow magnitude image output, time-resolved lumen contours were automatically tracked as described above. The following 4D-flow based metrics were computed: flow rate (mL/s), net flow per cycle (mL), and mean/max velocity (cm/s). All metrics were compared for all combinations of model wall stiffness and temporal resolution.

PWV calculation

PWV computation was focused on the descending aorta only (with centerline c_{DAo} , reaching from landmarks 'LSA' to 'outlet'), owing to flow effects at the arch branches that alter flow waveforms and compound the computation. Given the full velocity vector field $\mathbf{v}[\mathbf{x}, t]$, lumen mask $m[\mathbf{x}]$, and c_{DAo} , flow rate curves were computed as follows: (1) define N cross-sectional analysis planes with normal vector $\mathbf{n}_k$ with $k = [1, N]$ at equidistant points c_k along c_{DAo} (spacing = 5 mm); (2) retrieve oriented lumen cross-sections A_k at planes defined by $\mathbf{n}_k$, c_k , and lumen mask $m(\mathbf{x})$; (3) compute flow rate

$$Q_k[t] = \int \langle \mathbf{v}[\mathbf{x}, t], \mathbf{n}_k \rangle dA_k.$$

$Q_k[t]$ curves were interpolated using cubic-splines and the time-to-foot (TTF) approach was used to calculate PWV²⁸. Briefly, TTF_k for each $Q_k[t]$ curve was defined as the x-intercept of a line fitted through the waveform's upslope points at 20 % and 80 % of the peak flow rate. $TTF_k[c_k]$ was plotted as function of the centerline location and PWV_{TTF} was defined as the inverse slope of the linear regression line fitted to $TTF_k[c_k]$. Linear regression used a conventional least-square-error (LSE) approach as well as a random sampling consensus (RANSAC) algorithm to better handle outlier.

Results

Tensile testing

All material samples exhibited non-linear stress-strain behavior. The incremental Young's moduli were estimated by the tangent modulus (E_t) at nominal stress $\sigma = 0.053$ MPa, which was approximated by $\sigma = Pr/h_{wall}$ with P given by the recorded mean

pressure $P_{MAP} = 57$ mmHg (during pulsatile flow), wall thickness $h_{wall} = 0.002$ m, and average lumen radius $r = 0.014$ m. E_t for the compliant models M_{c1} and M_{c2} were 1.27 MPa (ranging 1.23 MPa to 1.31 MPa) and 4.3 MPa (ranging 3.7 MPa to 4.88 MPa), respectively. As for model M_r , no absolute elasticity estimates could be derived from the given stress-strain data, but differences in M_r elasticity were approximated to be at least 15-fold (>15 MPa) when compared to model M_{c1} . 3D-printing anisotropy was negligible with differences $<5\%$ within the relevant strain range for all three samples. Stress-strain plots are shown in Supplementary Fig. S1.

Pressure tuning

Pressure recordings were performed on the scanner bed prior to image acquisition. Pressure waveforms (Fig. 2, top row) recorded at the model inlet show increased peak pressures in models M_{c2} (116 mmHg) and M_r (133 mmHg) compared to the most compliant model M_{c1} (112 mmHg). At the model outlet (DAo branch) peak pressures dropped by 7 mmHg for both compliant models M_{c1} and M_{c2} , and by 11 mmHg for the nearly rigid model M_r . Diastolic pressure values were between 38 mmHg and 40 mmHg for all models at the inlet and dropped by 1 mmHg at the outlet. All pressure waveforms showed an oscillating behaviour both in systole and diastole, which was most dominant in model M_r .

Flow split tuning

Flow splits between all model outlets were consistent between the three models, with net flows of 48-49 ml measured with the ultrasonic probe at the descending aorta outlet prior to each acquisition (corresponding 68 % of the programmed inlet flow). Inlet net flow volumes calculated from 2D-PC flow rate waveforms (Fig. 2, row 2) ranged from 64.6 to 66.4 mL; 2D-PC DAo outlet net flow volumes ranged from 51.6 to 52.7 mL. Adding up 2D-PC measured net flow at all outlets (b1, b2, b3 and DAo) total outflow was 70.0, 70.8, 67.8 mL for M_{c1} , M_{c2} , and M_r , corresponding to relative differences of 1.7, 0.6, and 4.8 % from the programmed inflow (71.2 mL).

Aorta wall expansion

Aortic wall expansion was clearly visible in systole for models M_{c1} and M_{c2} and most pronounced in the ascending aorta. Supplementary Fig. S2 displays contour tracking results. Based on the tracked contours, the calculated cross-sectional area increased by $>5\%$ in models M_{c1} and M_{c2} , whereas no detectable area change ($<1\%$) was measured for model M_r (Fig. 2, row 3). For models M_{c1} and M_{c2} , relative area change over the cardiac cycle also depicted a small secondary lobe in early diastole which was in phase with the secondary lobes of the pressure and flow rate waveforms.

Velocities at cross-sections

Figure 4 shows a qualitative comparison between model M_{c1} and model M_r and their velocity vector profiles for cardiac frames at peak systole and end systole. While profiles in models M_{c1} and M_r were similar at peak systole, minor qualitative differences were observed at end systole, particularly at landmarks proximal to the arch branches. Here, the nearly rigid model M_r showed a more centered cross-sectional velocity profile with less backward flow components and less helical flow tendencies. Velocity profiles at landmarks distal to the arch branches mainly differed with regards to the velocity vector magnitude, with higher velocities in model M_r . Model dependent velocity differences in the descending aorta can also be observed in maps of traced particles, as visualized in Fig. 3.

Figure 5 shows mean and maximum velocity analysis results. Maximum velocities at peak systole were highest in model M_r (73.1 cm/s at AAo, 42.5 cm/s at BCT, 39.1 cm/s at LSA, and 43.2 cm/s at DAo) compared to both compliant models M_{c1} (73.0 cm/s at AAo, 38.5 cm/s at BCT, 35.4 cm/s at LSA, and 35.5 cm/s at DAo) and M_{c2} (70.5 cm/s at AAo, 36.6 cm/s at BCT, 35.5 cm/s at LSA, and 39.2 cm/s at DAo). Likewise, cross-sectional mean velocities in systole were higher in model M_r than in the compliant models. Decreasing temporal sampling from 50 frames down to 16 frames showed greatest effects at the AAo landmark for peak velocity which decreased by 16 %, 19 %, 14 % for M_{c1} , M_{c2} , and M_r .

Flow at cross-sections

Figure 6 shows flow rate and cumulative net flow results. Results were similar between different models or between 4D-flow data of different temporal sampling rates. Net flow values were 69.6 ± 1.9 ml for AAo, 63.4 ± 1.6 ml for BCT, 47.0 ± 0.7 ml for LSA, and 45.7 ± 1.0 ml for DAo (given as mean $\pm$ SD over three models and three temporal sampling rates). Net flow at the landmarks upstream and downstream of the arch branches were within 10 % of the programmed pump value (71.2 ml) and the measured ultrasonic value (48.5 ml), respectively.

Flow rate waveforms over the cardiac cycle (Fig. 6, solid lines) showed weaker peak flow rate dampening with increased model wall stiffness. This effect was most pronounced at landmarks further downstream (LSA, DAo). 4D-flow sampled at 50 frames per cycle revealed a double flow rate peak in systole at all cross-sections, which was much less apparent in the dataset sampled at 25 frames per cycle and not apparent in the dataset sampled at 16 frames per cycle. Moreover, a distinct second ($t = 0.66$ s) and third ($t = 0.85$ ms) flow rate peak were present in all derived waveforms, irrespective of model stiffness and

temporal sampling rate.

PWV

PWV values were estimated from time-to-foot (TTF) delays of flow rate waveforms at equidistantly spaced cross-sectional planes along the descending aorta (Fig. 7). Based on the datasets with highest temporal sampling (50 frames/cycle), PWV was 6.98 m/s (LSE) and 5.78 m/s (RANSAC) for model M_{c1} , and 7.31 m/s (LSE) and 7.31 m/s (RANSAC) for model M_{c2} . Both for model M_{c1} and M_{c2} , PWV values were lower with datasets sampled at 25 frames/cycle and further decreased for datasets sampled at 16 frames/cycle. The 16 frames/cycle dataset of model M_{c1} included four extreme outlier points posting large negative TTF values. These points were excluded prior to fitting the linear model. No linear relationship between TTF and centerline position was detectable for model M_r irrespective of temporal resolution. Consequently, while PWV was very high, no PWV values could be reported based on the data.

Discussion

This study demonstrated the feasibility of integrating a subject-specific aorta model with varying wall elasticity into an MRI-compatible flow circuit setup that operates under physiological flow and pressure conditions. Utilizing prolonged and highly-controlled *in vitro* 4D-flow imaging, we showed the influences of aortic wall compliance and temporal sampling rates on both cross-sectional flow parameters and 4D-flow derived PWV.

Stress-strain testing of the compliant 3D-printing material suggested that the derived tangent moduli E_t of models M_{c1} and M_{c2} are in the same range as the incremental Young's moduli that have been reported for a 'young' (more compliant) and 'old' (stiffer) human thoracic aorta, respectively²⁹. We did not attempt to report E_t for the non-realistic model M_r due to the material's substantially higher stiffness which challenged reliable tensile testing, but approximated the difference in elasticity to be at least 15-fold when compared to model M_{c1} .

Pressure tuning was performed for model M_{c1} only. Subsequently, model M_{c1} was interchanged with models M_{c2} and M_r , but resistance and downstream capacitance were kept constant. This approach allowed for isolated evaluation of the effect increased wall compliance on pressure and flow. While tuning model M_{c1} to physiological P_{sys} was successful, P_{dias} was below the target range of 70-80 mmHg. Previous work with advanced MRI-compatible flow circuit setups reported similar increased (i.e. >50 mmHg) pulse pressures^{21,30,31}. We note that two main factors determine successful pressure tuning: (1) P_{sys} or P_{MAP} can easily be elevated by increasing flow resistance distal to the capacitors for which the pulse pressure remains constant; (2) pulse pressure is governed by the available capacitance, i.e. compressible air volume ($C1$, $C2$) and by the ratio of distal to proximal resistance ($R1_{dist}/R1_{prox}$, $R2_{dist}/R2_{prox}$). A greater ratio provides a wider range for tuning pulse pressure. Despite P_{dias} being lower than physiological values, the achieved conditions were found to be sufficient to study the impact wall compliance on flow dynamics. Interchanging models under consistent resistance and capacitance settings led to two effects on pressure: (1) P_{sys} increased in model M_{c2} , and more so in model M_r , but no effects were seen on P_{dias} ; (2) inlet to outlet peak pressure differences increased with increasing wall stiffness. Both of these observations were as expected and they affirmed the validity of the setup.

Pressure and flow rate oscillations (Fig. 2) are expected to be caused by wave reflections at several branching points – natural arch branches, rigid flow connectors, flow valves, etc. – and under-damping in the system. Other studies with comparable flow circuit setups showed similar oscillating waveform shapes^{21,30,31}. Additional engineering efforts to mitigate this phenomenon may benefit analyses of pressure and flow waveform shapes in multiple vessel geometries and/or under varying boundary conditions.

4D-flow image-based visualizations of vector profiles and traced particles indicated that variations in wall compliance lead to variations in velocity amplitudes and profiles (Fig. 3, 4). The quantitative analyses showed that both mean and maximum velocities decreased for the compliant models when compared to the nearly rigid version. Likewise, flow rate waveform dampening was most pronounced in the most compliant model and the further downstream the centerline. Thus, in comparative *in vitro* to *in vivo* studies – regardless of efforts to match patient-specific inflow conditions – fully rigid aorta models are likely insufficient for direct comparison.

4D-flow based PWV calculations in compliant models M_{c1} and M_{c2} provided values within the range of PWV values that have been reported in *in vivo* 4D-flow studies^{11,32,33}. PWV in the model M_r was too high to reliably resolve TTF delays along the centerline. Thus, we did not attempt to report a 4D-flow derived PWV for model M_r . The referenced *in vivo* studies included healthy volunteers (young and old) and patients with atherosclerosis. Mean PWV among the respective cohorts ranged from 3.8 m/s to 6.4 m/s. However, they did not include PWV measurements based on multiple 4D-flow datasets with varying temporal resolutions, which ranged from 32-41 ms – a typical temporal resolution in *in vivo* 4D-flow acquisitions amid scan time limitations. Another previous study utilizing through-plane encoded 2D-PC MRI with higher temporal resolution (6-10 ms), reported PWV values ranging from 4.3 m/s (healthy and young controls) to 6.5 m/s (older patients).

In contrast to the reported *in vivo* 4D-flow based PWV values, theoretical PWV values based on Moens-Korteweg (Eq. 1) are 8-10.7 m/s for model M_{c1} , 14.9-20 m/s for model M_{c2} , and 31-41 m/s (assuming E_t of model M_r to be 15-fold over model M_{c1}). The given ranges correspond to the change of aortic diameter, which ranges from 36 mm in the ascending aorta to 20 mm in the distal descending aorta. These theoretical values may be debated, as the Moens-Korteweg equations assumes a constant vessel diameter, which is not true of the aorta.

Variations in 4D-flow temporal resolutions affected PWV considerably. Assuming that the presented 4D-flow data at highest temporal resolution ($\Delta t = 20$ ms) generates the most reliable PWV values, the present results suggest that lower temporal sampling rates underestimate absolute PWV (up to 35 %). Specifically, our data shows that the impact of temporal resolution on PWV calculation may be more dominant than the effect of varying wall compliance. One TTF plot (model M_{c1} , 16 frames) included distinct outlier points with negative TTF delay that were removed prior to linear model fitting. This emphasizes that reliable PWV calculations are highly dependent on accurate flow waveforms, particularly when derived from data with low temporal resolution. In that case, using the alternative iterative RANSAC approach for fitting a linear regression showed the effect on PWV while directly excluding these outlier points.

Three key limitations of this study were identified. First, only a single approach for PWV measurement (TTF) was used. In addition to the TTF, others derived PWV by time-to-peak (TTP), time-to-upstroke (TTU), and by correlation analysis of time-shifted flow waveforms (xCorr)^{28,33,34}. Wentland et al.³³ analyzed differences in PWV for these four approaches. PWV values were similar for TTF, TTU and xCorr, while TTP results deviated most due to challenges of detecting the true peak flow point in data with mediocre temporal resolution.

Second, synthetic aorta models were manufactured with uniform wall thickness and elasticity which simplifies the *in vivo* aorta. These local variations of the model may impact calculated PWV values. A subject-specific wall mesh directly built on a vessel wall segmentation – rather than segmenting the lumen and extruding the surface by a pre-defined and uniform wall thickness – may be an alternative approach. To this end, a 3D dark blood MRI protocol is able to provide the necessary image basis for building models with non-uniform wall thickness³⁵.

Third, the study design did not assess the effects of heart rate or pressure variations on PWV, which remains a controversial topic according to other previous studies. A pre-clinical study with rats reported a positive HR to PWV correlation, which was further pronounced at higher mean arterial pressures³⁶. Clinical studies that paced patients at different heart rates found either a positive HR to PWV correlation^{37–40} or no correlation^{41,42}. If heart rate to PWV dependencies were to be investigated with the presented *in vitro* setup, careful considerations need to be made on how to modify the inlet flow rate waveform and whether or not pressures should be regulated with programmed HR changes. Given this open research question, we consider the present flow circuit setup with compliant aorta models to be of high value to further investigate heart rate and pressure variations.

In conclusion, this work demonstrated 3D-printed subject-specific compliant models of the thoracic aorta integrated into a highly-controlled physiological flow circuit for assessment via *in vitro* MRI. Using compliant rather than rigid models of the aorta is essential to produce realistic flow dynamics and conditions that recapitulate *in vivo* hemodynamics.

Acknowledgements

We thank Nicole Schiavone, Tyler Cork, and Tabitha Bandy-Vizcaino for their technical advice and Anja Hennemuth for making available software tools. J.Z. receives personal funds through DAAD doctoral candidate scholarship grant. This project was funded, in part, by NIH R01 HL131823 to D.B.E.

Author contributions

J.Z. conceptualized the study, engineered the flow circuit setup, defined and led imaging experiments, performed image analysis and data interpretation, and wrote the manuscript. M.Lo. defined the imaging protocol and developed image pre-processing software. F.K. performed tensile testing and supported imaging experiments. K.B. and K.G. developed and manufactured the aorta model. S.A.D. supported imaging experiments. M.Le. and A.M. supervised and advised the research work. D.B.E. conceptualized the study and provided overall supervision and advice to the research. All authors reviewed and approved the manuscript.

Additional Information

Data availability: The subject-specific thoracic aorta model and custom-build model-specific connectors (.stl files), as well as all acquired MRI DICOM data will be made publicly available upon publication.

Competing interests: The authors declare no competing interests.

References

1. Blacher, J. *et al.* Impact of aortic stiffness on survival in end-stage renal disease. *Circulation* **99**, 2434–2439, DOI: [10.1161/01.CIR.99.18.2434](https://doi.org/10.1161/01.CIR.99.18.2434) (1999).
2. Laurent, S. *et al.* Aortic stiffness is an independent predictor of all-cause and cardiovascular mortality in hypertensive patients. *Hypertension* **37**, 1236–1241, DOI: [10.1161/01.HYP.37.5.1236](https://doi.org/10.1161/01.HYP.37.5.1236) (2001).
3. Asmar, R., Rudnichi, A., Blacher, J., London, G. M. & Safar, M. E. Pulse pressure and aortic pulse wave are markers of cardiovascular risk in hypertensive populations. *Am. J. Hypertens.* **14**, 91–97, DOI: [10.1016/S0895-7061\(00\)01232-2](https://doi.org/10.1016/S0895-7061(00)01232-2) (2001).
4. Kearney, P. M. *et al.* Global burden of hypertension: analysis of worldwide data. *The Lancet* **365**, 217–223, DOI: [10.1016/S0140-6736\(05\)17741-1](https://doi.org/10.1016/S0140-6736(05)17741-1) (2005).
5. Zhou, B. *et al.* Worldwide trends in blood pressure from 1975 to 2015: a pooled analysis of 1479 population-based measurement studies with 19·1 million participants. *The Lancet* **389**, 37–55, DOI: [10.1016/S0140-6736\(16\)31919-5](https://doi.org/10.1016/S0140-6736(16)31919-5) (2017).
6. Laurent, S. *et al.* Expert consensus document on arterial stiffness: Methodological issues and clinical applications, DOI: [10.1093/eurheartj/ehl254](https://doi.org/10.1093/eurheartj/ehl254) (2006).
7. Markl, M. *et al.* Time-resolved three-dimensional phase-contrast MRI. *J. Magn. Reson. Imaging* **17**, 499–506, DOI: [10.1002/jmri.10272](https://doi.org/10.1002/jmri.10272) (2003).
8. van der Geest, R. J. & Garg, P. Advanced Analysis Techniques for Intra-cardiac Flow Evaluation from 4D Flow MRI. *Curr. Radiol. Reports* **4**, DOI: [10.1007/s40134-016-0167-7](https://doi.org/10.1007/s40134-016-0167-7) (2016).
9. Markl, M. *et al.* In vivo wall shear stress distribution in the carotid artery: effect of bifurcation geometry, internal carotid artery stenosis, and recanalization therapy. *Circ. Cardiovasc. Imaging* **3**, 647–655, DOI: [10.1161/circimaging.110.958504](https://doi.org/10.1161/circimaging.110.958504) (2010).
10. Markl, M., Frydrychowicz, A., Kozerke, S., Hope, M. & Wieben, O. 4D flow MRI. *J. Magn. Reson. Imaging* **36**, 1015–36, DOI: [10.1002/jmri.23632](https://doi.org/10.1002/jmri.23632) (2012).
11. Dyverfeldt, P., Ebbers, T. & Länne, T. Pulse wave velocity with 4D flow MRI: Systematic differences and age-related regional vascular stiffness. *Magn. Reson. Imaging* **32**, 1266–1271, DOI: [10.1016/j.mri.2014.08.021](https://doi.org/10.1016/j.mri.2014.08.021) (2014).
12. Valvano, G. *et al.* Accelerating 4D flow MRI by exploiting low-rank matrix structure and hadamard sparsity. *Magn. Reson. Medicine* **78**, 1330–1341, DOI: [10.1002/mrm.26508](https://doi.org/10.1002/mrm.26508) (2017).
13. Ma, L. E. *et al.* Aortic 4D flow MRI in 2 minutes using compressed sensing, respiratory controlled adaptive k-space reordering, and inline reconstruction. *Magn. Reson. Medicine* **81**, 3675–3690, DOI: [10.1002/mrm.27684](https://doi.org/10.1002/mrm.27684) (2019).
14. Schnell, S. *et al.* K-t GRAPPA accelerated four-dimensional flow MRI in the aorta: Effect on scan time, image quality, and quantification of flow and wall shear stress. *Magn. Reson. Medicine* **72**, 522–533, DOI: [10.1002/mrm.24925](https://doi.org/10.1002/mrm.24925) (2014).
15. Garg, P. *et al.* Comparison of fast acquisition strategies in whole-heart four-dimensional flow cardiac MR: Two-center, 1.5 Tesla, phantom and in vivo validation study. *J. Magn. Reson. Imaging* **47**, 272–281, DOI: [10.1002/jmri.25746](https://doi.org/10.1002/jmri.25746) (2018).
16. Loecher, M., Magrath, P., Aliotta, E. & Ennis, D. B. Time-optimized 4D phase contrast MRI with real-time convex optimization of gradient waveforms and fast excitation methods. *Magn Reson. Med* **1**, 1–12, DOI: [10.1002/mrm.27716](https://doi.org/10.1002/mrm.27716) (2019).
17. Dillinger, H., Walheim, J. & Kozerke, S. On the limitations of echo planar 4D flow MRI. *Magn. Reson. Medicine* **84**, 1806–16, DOI: [10.1002/mrm.28236](https://doi.org/10.1002/mrm.28236) (2020).
18. Lorenz, R. *et al.* Closed circuit MR compatible pulsatile pump system using a ventricular assist device and pressure control unit. *Magn. Reson. Medicine* **67**, 258–268, DOI: [10.1002/mrm.22983](https://doi.org/10.1002/mrm.22983) (2012).
19. Canstein, C. *et al.* 3D MR flow analysis in realistic rapid-prototyping model systems of the thoracic aorta: Comparison with in vivo data and computational fluid dynamics in identical vessel geometries. *Magn. Reson. Medicine* **59**, 535–546, DOI: [10.1002/mrm.21331](https://doi.org/10.1002/mrm.21331) (2008).
20. Wen, C. Y., Yang, A. S., Tseng, L. Y. & Chai, J. W. Investigation of pulsatile flowfield in healthy thoracic aorta models. *Annals Biomed. Eng.* **38**, 391–402, DOI: [10.1007/s10439-009-9835-6](https://doi.org/10.1007/s10439-009-9835-6) (2010).
21. Urbina, J. *et al.* Realistic aortic phantom to study hemodynamics using MRI and cardiac catheterization in normal and aortic coarctation conditions. *J. Magn. Reson. Imaging* **44**, 683–697, DOI: [10.1002/jmri.25208](https://doi.org/10.1002/jmri.25208) (2016).

22. Jung, B. *et al.* Investigation of hemodynamics in an in vitro system simulating left ventricular support through the right subclavian artery using 4-dimensional flow magnetic resonance imaging. *J. Thorac. Cardiovasc. Surg.* **150**, 200–207, DOI: [10.1016/j.jtcvs.2015.02.048](https://doi.org/10.1016/j.jtcvs.2015.02.048) (2015).
23. Rosero, E. B. *et al.* Sex, race, and age distributions of mean aortic wall thickness in a multiethnic population-based sample. *J. Vasc. Surg.* **53**, 950–957, DOI: [10.1016/j.jvs.2010.10.073](https://doi.org/10.1016/j.jvs.2010.10.073) (2011).
24. Updegrove, A. *et al.* SimVascular: An Open Source Pipeline for Cardiovascular Simulation. *Annals Biomed. Eng.* **45**, 525–541, DOI: [10.1007/s10439-016-1762-8](https://doi.org/10.1007/s10439-016-1762-8) (2017).
25. Markl, M. *et al.* Generalized reconstruction of phase contrast MRI: Analysis and correction of the effect of gradient field distortions. *Magn. Reson. Medicine* **50**, 791–801, DOI: [10.1002/mrm.10582](https://doi.org/10.1002/mrm.10582) (2003).
26. Tautz, L., Hennemuth, A. & Peitgen, H. O. Motion analysis with quadrature filter based registration of tagged MRI sequences. *Stat. Atlases Comput. Model. Hear. Lect. Notes Comput. Sci.* **7085**, 78–87, DOI: [10.1007/978-3-642-28326-0_48](https://doi.org/10.1007/978-3-642-28326-0_48) (2012).
27. Selle, D., Preim, B., Schenk, A. & Peitgen, H.-O. Analysis of vasculature for liver surgical planning. *IEEE Transactions on Med. Imaging* **21**, 1344–57, DOI: [10.1109/TMI.2002.801166](https://doi.org/10.1109/TMI.2002.801166) (2002).
28. Markl, M. *et al.* Estimation of global aortic pulse wave velocity by flow-sensitive 4D MRI. *Magn. Reson. Medicine* **63**, 1575–1582, DOI: [10.1002/mrm.22353](https://doi.org/10.1002/mrm.22353) (2010).
29. Learoyd, B. M. & Taylor, M. G. Alterations with age in the viscoelastic properties of human arterial walls. *Circ. Res.* **18**, 278–292, DOI: [10.1161/01.RES.18.3.278](https://doi.org/10.1161/01.RES.18.3.278) (1966).
30. Gallarello, A. *et al.* Patient-Specific Aortic Phantom With Tunable Compliance. *J. Eng. Sci. Med. Diagn. Ther.* **2**, 041005, DOI: [10.1115/1.4044611](https://doi.org/10.1115/1.4044611) (2019).
31. Birjiniuk, J. *et al.* Pulsatile Flow Leads to Intimal Flap Motion and Flow Reversal in an In Vitro Model of Type B Aortic Dissection. *Cardiovasc. Eng. Technol.* **8**, 378–389, DOI: [10.1007/s13239-017-0312-3](https://doi.org/10.1007/s13239-017-0312-3) (2017).
32. Markl, M. *et al.* Analysis of pulse wave velocity in the thoracic aorta by flow-sensitive four-dimensional MRI: Reproducibility and correlation with characteristics in patients with aortic atherosclerosis. *J. Magn. Reson. Imaging* **35**, 1162–1168, DOI: [10.1002/jmri.22856](https://doi.org/10.1002/jmri.22856) (2012).
33. Wentland, A. L. *et al.* Aortic pulse wave velocity measurements with undersampled 4D flow-sensitive MRI: Comparison with 2D and algorithm determination. *J. Magn. Reson. Imaging* **37**, 853–859, DOI: [10.1002/jmri.23877](https://doi.org/10.1002/jmri.23877) (2013). [NIHMS150003](https://pubmed.ncbi.nlm.nih.gov/24150003/).
34. Fielden, S. W. *et al.* A new method for the determination of aortic pulse wave velocity using cross-correlation on 2D PCMR velocity data. *J. Magn. Reson. Imaging* **27**, 1382–1387, DOI: [10.1002/jmri.21387](https://doi.org/10.1002/jmri.21387) (2008).
35. Roes, S. D. *et al.* Aortic vessel wall magnetic resonance imaging at 3.0 tesla: A reproducibility study of respiratory navigator gated free-breathing 3D black blood magnetic resonance imaging. *Magn. Reson. Medicine* **61**, 35–44, DOI: [10.1002/mrm.21798](https://doi.org/10.1002/mrm.21798) (2009).
36. Tan, I., Butlin, M., Liu, Y. Y., Ng, K. & Avolio, A. P. Heart rate dependence of aortic pulse wave velocity at different arterial pressures in rats. *Hypertension* **60**, 528–533, DOI: [10.1161/HYPERTENSIONAHA.112.194225](https://doi.org/10.1161/HYPERTENSIONAHA.112.194225) (2012).
37. Haesler, E., Lyon, X., Pruvot, E., Kappenberger, L. & Hayoz, D. Confounding effects of heart rate on pulse wave velocity in paced patients with a low degree of atherosclerosis. *J. Hypertens.* **22**, 1317–1322, DOI: [10.1097/01.hjh.0000125447.28861.18](https://doi.org/10.1097/01.hjh.0000125447.28861.18) (2004).
38. Albaladejo, P. *et al.* Heart Rate, Arterial Stiffness, and Wave Reflections in Paced Patients. *Hypertension* **38**, 949–952, DOI: [10.1161/hy1001.096210](https://doi.org/10.1161/hy1001.096210) (2001).
39. Lantelme, P., Mestre, C., Lievre, M., Gressard, A. & Milon, H. Heart rate: An important confounder of pulse wave velocity assessment. *Hypertension* **39**, 1083–1087, DOI: [10.1161/01.HYP.0000019132.41066.95](https://doi.org/10.1161/01.HYP.0000019132.41066.95) (2002).
40. Liang, Y. L. *et al.* Effects of heart rate on arterial compliance in men. *Clin. Exp. Pharmacol. Physiol.* **26**, 342–346, DOI: [10.1046/j.1440-1681.1999.03039.x](https://doi.org/10.1046/j.1440-1681.1999.03039.x) (1999).
41. Wilkinson, I. B. *et al.* Heart rate dependency of pulse pressure amplification and arterial stiffness. *Am. J. Hypertens.* **15**, 24–30, DOI: [10.1016/S0895-7061\(01\)02252-X](https://doi.org/10.1016/S0895-7061(01)02252-X) (2002).
42. Albaladejo, P. *et al.* Influence of sex on the relation between heart rate and aortic stiffness. *J. Hypertens.* **21**, 555–562, DOI: [10.1097/00004872-200303000-00021](https://doi.org/10.1097/00004872-200303000-00021) (2003).

Figure 1. (a) Schematic of the MRI-compatible flow circuit setup. The pump unit was positioned at the end of the patient bed; the fluid reservoir was positioned on the patient bed and inside the MRI bore. The pump controller provided a pulse for triggering both image acquisition and DAQ signals. Ultrasonic flow transducers and pressure transducers (dotted lines) were disconnected after tuning and prior to moving the setup to the MRI iso-center. (b) Photograph of the model-specific gel block with embedded aorta model and ports (blue) at inlet and outlet to insert pressure transducers. (c, d, e) 3D spoiled-gradient echo MRI image data for three reformatted planes (XY, XZ, YZ) depicting the aorta model embedded into the gel. (f) The subject-specific model of the thoracic aorta with defined cross-sectional landmarks, full centerline (black), and descending aorta centerline (blue) that was used for PWV analysis. The original model was extended with cylindrical caps (length = 2 cm) at the inlet and all outlets to accommodate connection to customized barbed connectors that then connect to tubing. (g) Photographs of a finished 3D-printed model.

Figure 2. Experimental setup conditions for aorta models M_{c1} (red), M_{c2} (green), and M_r (blue), evaluated at selected landmarks (Fig. 1f). (Row 1) Pressure conditions recorded at inlet (solid) and outlet (dashed). (Row 2) Flow rate waveforms at inlet and all outlets with calculated net flow volumes, based on 2D PC-MRI data. (Row 3) Cross-sectional area change relative to area at cardiac cycle start, based on tracked lumen contours in 2D-cine-GRE data. (Row 4) Overlay of tracked lumen contour at cross-section AAO for all acquired time frames ($N = 50$). Animated contour tracking results are presented in Supplementary Fig. S2.

Figure 3. Particle tracing based on 4D-flow MRI data for the three aorta models of identical subject-specific geometry, but different wall stiffness. While particle traces matched among the three models, velocities along the descending aorta – as depicted by the color – were slightly higher in the nearly rigid model M_r . An animated version of traced particles is shown in Supplementary Fig. S3-5.

Figure 4. 4D-flow cross-sectional velocity profiles in models M_{c1} (top) and M_r (bottom) at peak-systolic ($t = 280$ ms) and end-systolic ($t = 400$ ms) frames. All profiles are 3D-rendered using the identical camera view and colored according to the 3D direction (red-green-blue arrow legend). Backward flow is visible at cross-sections prior to the arch branches (inlet, AAO, BCT), specifically at end-systole. One can appreciate the different vector profiles for the compliant model M_{c1} when compared to the nearly rigid model M_r .

Figure 5. 4D-flow cross-sectional mean (solid) and maximum (dashed) velocity at four landmarks (Fig. 1f) and three different temporal sampling rates. Highest peak-systolic velocities (see values in legends) were measured in model M_r (blue); and inter-model peak-systolic velocity differences were greater at landmarks further downstream. The temporal sampling rate impacted the measurement of peak velocities, which was most pronounced at AAo point. Spikes in diastole were attributed to noise in near-boundary pixels and inaccurate contouring.

Figure 6. 4D-flow derived flow rate waveforms (left ordinates) and cumulative net flow (right ordinates) at four landmarks (Fig. 1f) and three different temporal sampling rates. Grey horizontal lines show the expected net flow according to the programmed inflow (71.2 ml/cycle) for the AAO and BCT slices, and measured (via ultrasonic transducer during setup tuning) DAO branch outflow (48.5 ml/cycle) for the LSA and DAO slices.

Figure 7. PWV calculations for models M_{c1} (red), M_{c2} (green), and M_r (blue) for three temporal sampling rates (row 1 through 3). Scattered points depict TTF of flow rate waveforms extracted at equidistantly spaced cross sections along c_{DAo} (Fig. 1f). All TTF values are shown as TTF differences to TTF at centerline position 0. Conventional LSE linear regression (solid line, with R^2 given in legend) and RANSAC (dashed line with assigned outlier marked +) were used to derive PWV values, defined as the inverse slope of the respective line (given in legend). Differences in PWV were observed between models M_{c1} and M_{c2} . However, temporal sampling rates impacted these values, with decreasing PWV estimates for lower temporal resolution. It appears that PWV in model M_r is too fast such that plausible TTF along the centerline cannot be resolved with the 4D-flow based approach. LSE is very sensitive to outlier data which was apparent in data M_{c1} (50 frames) and M_{c1} (16 frames). For the latter one, four TTF points laid outside the displayed y-axis with linear regression deviations $>100 \times \text{RMSE}$ and were thus excluded prior to fitting the model. TTF, time-to-foot; LSE, least-squared-error; RANSAC, random sampling consensus; RMSE, root-mean-squared-error.

Figures

Figure 1

(a) Schematic of the MRI-compatible flow circuit setup. The pump unit was positioned at the end of the patient bed; the fluid reservoir was positioned on the patient bed and inside the MRI bore. The pump controller provided a pulse for triggering both image acquisition and DAQ signals. Ultrasonic flow transducers and pressure transducers (dotted lines) were disconnected after tuning and prior to moving the setup to the MRI iso-center. (b) Photograph of the model-specific gel block with embedded aorta model and ports (blue) at inlet and outlet to insert pressure transducers. (c, d, e) 3D spoiled-gradient echo MRI image data for three reformatted planes (XY, XZ, YZ) depicting the aorta model embedded into the

gel. (f) The subject-specific model of the thoracic aorta with defined cross-sectional landmarks, full centerline (black), and descending aorta centerline (blue) that was used for PWV analysis. The original model was extended with cylindrical caps (length = 2 cm) at the inlet and all outlets to accommodate connection to customized barbed connectors that then connect to tubing. (g) Photographs of a finished 3D-printed model.

Figure 2

Experimental setup conditions for aorta models Mc1 (red), Mc2 (green), and Mr (blue), evaluated at selected landmarks (Fig. 1f). (Row 1) Pressure conditions recorded at inlet (solid) and outlet (dashed). (Row 2) Flow rate waveforms at inlet and all outlets with calculated net flow volumes, based on 2D PC-MRI data. (Row 3) Cross-sectional area change relative to area at cardiac cycle start, based on tracked lumen contours in 2D-cine-GRE data. (Row 4) Overlay of tracked lumen contour at cross-section AAo for all acquired time frames (N = 50). Animated contour tracking results are presented in Supplementary Fig. S2.

Figure 3

Particle tracing based on 4D-flow MRI data for the three aorta models of identical subject-specific geometry, but different wall stiffness. While particle traces matched among the three models, velocities along the descending aorta – as depicted by the color – were slightly higher in the nearly rigid model Mr. An animated version of traced particles is shown in Supplementary Fig. S3-5.

Figure 4

4D-flow cross-sectional velocity profiles in models Mc1 (top) and Mr (bottom) at peak-systolic ($t = 280\text{ms}$) and end-systolic ($t = 400\text{ms}$) frames. All profiles are 3D-rendered using the identical camera view and colored according to the 3D direction (red-green-blue arrow legend). Backward flow is visible at cross-sections prior to the arch branches (inlet, AAo, BCT), specifically at end-systole. One can appreciate the different vector profiles for the compliant model Mc1 when compared to the nearly rigid model Mr.

Figure 5

4D-flow cross-sectional mean (solid) and maximum (dashed) velocity at four landmarks (Fig. 1f) and three different temporal sampling rates. Highest peak-systolic velocities (see values in legends) were measured in model M_r (blue); and inter-model peak-systolic velocity differences were greater at landmarks further downstream. The temporal sampling rate impacted the measurement of peak velocities, which was most pronounced at AAo point. Spikes in diastole were attributed to noise in near-boundary pixels and inaccurate contouring.

Figure 6

4D-flow derived flow rate waveforms (left ordinates) and cumulative net flow (right ordinates) at four landmarks (Fig. 1f) and three different temporal sampling rates. Grey horizontal lines show the expected net flow according to the programmed inflow (71.2 ml/cycle) for the AAO and BCT slices, and measured (via ultrasonic transducer during setup tuning) DAo branch outflow (48.5 ml/cycle) for the LSA and DAo slices.

Figure 7

PWV calculations for models M_{c1} (red), M_{c2} (green), and M_r (blue) for three temporal sampling rates (row 1 through 3). Scattered points depict TTF of flow rate waveforms extracted at equidistantly spaced cross sections along cDAo (Fig. 1f). All TTF values are shown as TTF differences to TTF at centerline position 0. Conventional LSE linear regression (solid line, with R^2 given in legend) and RANSAC (dashed line with assigned outlier marked +) were used to derive PWV values, defined as the inverse slope of the respective line (given in legend). Differences in PWV were observed between models M_{c1} and M_{c2} . However, temporal sampling rates impacted these values, with decreasing PWV estimates for lower temporal resolution. It appears that PWV in model M_r is too fast such that plausible TTF along the centerline cannot be resolved with the 4D-flow based approach. LSE is very sensitive to outlier data which was apparent in data M_{c1} (50 frames) and M_{c1} (16 frames). For the latter one, four TTF points laid

outside the displayed y-axis with linear regression deviations $>100 \times \text{RMSE}$ and were thus excluded prior to fitting the model. TTF, time-to-foot; LSE, least-squared-error; RANSAC, random sampling consensus; RMSE, root-mean-squared-error.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [SupplementaryInformation.zip](#)