Pharmacokinetics of Anti-Infective Agents During CytoSorb Hemoadsorption: An Experimental Study

Antoine Schneider (✉ antoine.schneider@chuv.ch)

Centre Hospitalier Universitaire Vaudois <https://orcid.org/0000-0003-0576-7094>

Pascal André

Centre Hospitalier Universitaire Vaudois

Joerg Scheier

Cytosorbents Inc

Monika Schmidt

HCx consulting GmbH

Heiko Ziervogel

CP Consulting

Thierry Buclin

Centre Hospitalier Universitaire Vaudois

Detlef Kindgen-Milles

University Hospital Dusseldorf: Universitätsklinikum Dusseldorf

Research

Keywords: Pharmacokinetics, anti-bacterial agents, antifungal agents, sepsis, drug monitoring, hemoadsorption, CytoSorb

Posted Date: December 30th, 2020

DOI: <https://doi.org/10.21203/rs.3.rs-136014/v1>

License: This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

PHARMACOKINETICS OF ANTI-INFECTIVE AGENTS DURING CYTOSORB HEMOADSORPTION: AN EXPERIMENTAL STUDY

Authors:

Antoine G. Schneider, MD, PhD ^{1-2*}	antoine.schneider@chuv.ch
Pascal André, PharmD, PhD ^{3*}	pascal.andre@chuv.ch
Joerg Scheier, MD ⁴	joerg.scheier@cytosorbents.com
Monika Schmidt ⁵	m.schmidt@medizin-im-gruenen.de
Heiko Ziervogel, PhD ⁵	h.ziervogel@medizin-im-gruenen.de
Thierry Buclin, MD ³	thierry.buclin@chuv.ch
Detlef Kindgen-Milles, MD ⁴	kindgen-milles@med.uni-duesseldorf.de

*: Both authors contributed equally to the work

1. Adult Intensive Care Unit, Centre Hospitalier Universitaire Vaudois (CHUV), Lausanne, Switzerland
2. Faculty of Biology and Medicine, University of Lausanne, Lausanne, Switzerland
3. Clinical Pharmacology, Centre Hospitalier Universitaire Vaudois (CHUV), Lausanne, Switzerland
4. CytoSorbents Europe GmbH, Berlin, Germany
5. Medical Competence Center Berlin/Brandenburg c/o HCx Consulting GmbH, Wendisch Rietz, Germany
6. Dept. of Anesthesiology, University Hospital Düsseldorf, Germany

Corresponding author:

Dr Antoine G. Schneider
BH08-610, Centre Hospitalier Universitaire Vaudois (CHUV)
46, Avenue du Bugnon 1011, Lausanne Switzerland
Tel: +41 21 314 16 32
Email: antoine.schneider@chuv.ch

Running Title: CytoSorb and pharmacokinetics of anti-infectives

Word Count: 2958

ABSTRACT

Background: Cytokine hemoadsorption might be effective in patients with sepsis. However, its effect on anti-infective agents' pharmacokinetics remains largely unknown. We sought to determine the influence of hemoadsorption on the pharmacokinetics of common anti-infective agents.

Methods: This is an interventional experimental study, conducted in 24 healthy pigs. Animals were randomly allocated to either hemoadsorption (cases) or sham procedure (controls) and to a drug combination (3 cases and 3 controls for each combination). Hemoadsorption was performed with CytoSorb® (CytoSorbents Corporation, USA). We evaluated 17 drugs (clindamycin, fluconazole, linezolid, meropenem, piperacillin, anidulafungin, ganciclovir, clarithromycin, posaconazole, teicoplanin, tobramycin, ceftriaxone, ciprofloxacin, metronidazole, liposomal amphotericin B, flucloxacillin and cefepime). Repeated blood sampling from the extracorporeal circulation (adsorber inlet/outlet, sham circulation) were performed within six hours of administration. Total clearance and adsorber-specific clearances were computed at each time point.

Results: Hemoadsorption was associated with increased clearance of all study drugs, except for ganciclovir. Its impact on total body clearance was major for fluconazole (+282%), linezolid (+115%) and amphotericin B (+75%). It was minor for posaconazole (+32%), teicoplanin (+31%), anidulafungin (+23%), piperacillin (+19%), flucloxacillin (+16%), metronidazole (+16%) and ciprofloxacin (+15%) and insignificant (<10%) for all other drugs. Hemoadsorber clearance declined over time with even delayed *desorption* for beta-lactams. It was moderately correlated with drug's lipophilicity ($p=0.01$; $r^2=0.43$).

Conclusions: Hemoadsorption with CytoSorb® has limited effect on the pharmacokinetics of most tested anti-infective drugs but appears to increase fluconazole, linezolid and liposomal amphotericin B clearance. Studies in humans with sepsis are required to confirm these findings.

Keywords: Pharmacokinetics; anti-bacterial agents; antifungal agents; sepsis; drug monitoring, hemoadsorption; CytoSorb

PHARMACOKINETICS OF ANTI-INFECTIVE AGENTS DURING CYTOSORB HEMOADSORPTION: AN EXPERIMENTAL STUDY

BACKGROUND

Sepsis is a major health problem worldwide, affecting close to 50 million individuals each year and leading to 11 million deaths [1]. Its recognition has been identified as a global health priority by the WHO in 2017. Sepsis results from a dysregulated host response to infection [2]. Activation of both complement and coagulation systems lead to the massive release of pro- and anti-inflammatory cytokines in the blood, a phenomenon sometimes referred to as "cytokine storm" [3]. This response, through systemic hypotension, microcirculation alterations, endothelial lesions, as well as metabolism modulation, ultimately leads to cellular apoptosis, organ failure and death.

CytoSorb[®] (CytoSorbents Corporation, NJ, USA), a recent hemoadsorption device, has been marketed and licensed for extracorporeal cytokine removal within the European Union since 2011[4]. CytoSorb[®] cartridges can easily be inserted in extra-corporeal circulation circuits. They contain biocompatible polystyrene divinylbenzene copolymer beads coated with polyvinylpyrrolidone capable of removing molecules of mid-molecular weight using a combination of hydrophobic or ionic interactions as well as hydrogen bonding [5, 6].

The level of evidence supporting the use of CytoSorb[®] in septic shock remains low and largely observational [7] [8-11]. In their latest statement, experts from the Surviving Sepsis Campaign prompted for further research and did not recommend for or against blood purification techniques[12]. In this context, safety parameters are of particular importance in the decision of initiating such therapy in a patient. So far, post marketing surveillance and data from published literature has not suggested major adverse events besides occasional thrombocytopenia. However, the potential removal of life-saving medications such as antibiotics in sepsis requires particular attention [4]. Indeed, little is known about the effect of CytoSorb[®] on antibiotics' pharmacokinetics. In vitro models have confirmed its ability to remove some antibiotics from the blood [13]-[14]. However, these one-compartment models have numerous limitations and their results might not be translatable to humans. Authors have reported cases suggesting a minor

influence on serum blood levels [15-17]. However, these reports lack consistency and reproducibility.

Accordingly, we have designed an experimental animal study to determine the influence of hemoadsorption with CytoSorb® on the pharmacokinetics of anti-infective agents commonly prescribed in sepsis.

MATERIALS AND METHODS

Ethics and legal aspects

The study was approved by the State Office for Occupational Safety, Consumer Protection and Health – Department of Consumer Protection (Brandenburg, Potsdam, Germany), approval number 2347-3-2018. According to the German animal protection law, no additional approval by an Ethics Committee was necessary. The present report was prepared following the ARRIVE guidelines for animal research[18]. The experimental study was conducted in a medical competence center located in Wendisch Rietz, (Germany), in compliance with German law for animal protection.

Animal preparation and monitoring

We included 24 healthy German landrace pigs (10 female, 14 male) with a body weight of 45–60 kg (mean 52.3 kg). Pigs were pre-medicated with an initial intramuscular injection of ketamine (15 mg/kg), midazolam (0.25 mg/kg) and azaperon (6 mg/kg). Sedation was maintained using a continuous intravenous infusion of ketamine (10 mg/kg/h) and midazolam (0.5 mg/kg/h). Pancuronium was administered for muscle relaxation as required. Anticoagulation with heparin was administered. Surgical tracheostomy was performed to allow conventional mechanical ventilation (Evita XL or Oxylog 3000; Draegerwerk AG & Co KG Luebeck, Germany). Surgical catheterization of the carotid artery and jugular vein was performed (respectively Avanti and Avanti plus sheath introducer; Cordis Miami, FL, USA) to respectively allow for blood pressure monitoring and enable fluid and medication administration. Vital signs were monitored throughout the experiment (Philips M3046A Philips IntelliVue MP5; Philips, Amsterdam, The Netherlands). Finally, a 12 Fr double lumen catheter (dualyse expert; Vygon GmbH & Co KG, Aachen, Germany) was inserted in either the femoral or jugular vein to enable extracorporeal

circulation. All animals were euthanized at the end of the experiment, while fully sedated, by simultaneous intravenous potassium chloride and pancuronium administration.

Experimental design

Four experiments were conducted corresponding to four antibiotics combinations (Table 1). During each experiment, six animals were prepared as described above. After the administration of the antibiotic combination, animals were randomly allocated to either hemoadsorption with CytoSorb® (cases) or sham hemoperfusion (control group) on a one to one ratio.

Extracorporeal circulation

A schematic representation of the extracorporeal circulation (ECC) is available in the supplementary material. It was established in all animals with a dedicated device [BM11a, Baxter, Deerfield, IL, USA) and corresponding circuit (BM11-Lines-BA-HP (tube system set BM11-hemoperfusion adult set; VE17; BLD-clamp/VE); Baxter, Deerfield, IL, USA]. For animals allocated to the intervention group (table 1), a CytoSorb® cartridge was inserted in the ECC. For those allocated to the control group no cartridge was integrated (sham hemoperfusion). ECC was started one hour after the start of study antibiotic administration. Blood flow rate was kept between 150 and 200 mL/min throughout the experiment.

Blood samples and laboratory analysis

Blood samples were collected before the initiation of the antibiotics infusion, then respectively 5, 30, 90, 250 and 330 minutes after adsorber / sham procedure initiation. For animals allocated to the intervention group, we collected two samples per time point: before (inlet) and after (outlet) the absorber. For those allocated in the control group, we collected a single systemic sample (Fig. 1). All samples (approx. 9 mL) were drawn into EDTA monovettes (Sarstedt AG & Co KG, Nuembrecht, Germany). They were then centrifuged at 30'000 rpm for 10 min, split into two 2 mL Eppendorf tubes (Eppendorf AG, Hamburg, Germany) and frozen at -20°C. Drugs plasma levels were determined quantitatively by liquid chromatography tandem mass spectrometry (LC-MS/MS) (MS/MS: API 2000 or API 4000; Sciex, Nieuwerkerk aan den IJssel, The Netherlands; HPLC pump: Agilent Technologies, Santa Clara, CA, USA; CTC autosampler: Sciex) in a human medical laboratory (Bioscientia GmbH, Ingelheim, Germany).

Pharmacokinetic parameters calculations

Pharmacokinetics parameters were computed individually for each study drug and each animal using standard non-compartmental calculations and considering first order elimination kinetics. All calculations were performed using Microsoft Excel 2016 (Microsoft Corp., Redmond, WA, USA).

Area under the plasma drug concentration curves were calculated using the log trapezoidal rule based on the last measurement (AUC_{Co-last}) and with extrapolation to infinity assuming a constant elimination rate (AUC_{Co-inf}). Total clearance (CL_{tot}) could be deduced from the dose (D):

$$CL_{tot} = D / AUC_{Co-inf}$$

Total amount eliminated during the study period (D_{TE}) was calculated using the following formula:

$$D_{TE} = D \times AUC_{Co-last} / AUC_{Co-inf}$$

In addition, for animals allocated to the adsorber arm, adsorber's instantaneous removal amount (D_{IR}) were evaluated based on the following formula:

$$D_{IR} = Q_p \times ([inlet] - [outlet])$$

where Q_p is the effective plasmatic flow (hematocrit considered to be 40%), [inlet] the drug concentration in the pre-adsorber sample and [outlet] the drug concentration in the post adsorber sample. Cumulative removal attributable to the adsorber during study period (D_{CC}) was calculated using the log trapezoidal rule on instantaneous absorption rates. Finally, overall adsorber clearance during study period (CL_C) could be estimated:

$$CL_C = CL_{tot} \times D_{CC} / D_{TE}$$

For each drug, we report mean clearance and standard deviation by study group. Finally, we calculated the relative part of the hemoadsorber on overall animal total clearance (percent):

$$CL_C / CL_{tot} \times 100$$

We considered the impact of the adsorber on overall clearance to be major if it was associated with a >50% increase in total clearance, minor if this increase was between 10 and 50% and insignificant if it was <10%.

Predictors of hemoadsorber clearance

To evaluate the impact of drugs' physico-chemical properties on adsorber removal, we assessed the correlation between such properties and calculated attributable clearances. For each study drug, we obtained octanol–water partition coefficient (logP), octanol–water distribution coefficient (logD), acid dissociation constant at logarithmic scale (pKas), molecular electric charge (positive, negative, neutral or zwitterion), protein binding or molecular weight. These data were obtained using Chemaxon® online molecular library.

Statistical analyses

Between-groups comparisons were conducted using bilateral student t-test. Correlation analyses were performed using Pearson test. The strength of the correlation was assessed according to the correlation coefficient (ρ) and considered to be strong ($\rho > 0.8$), fair ($0.6 < \rho \leq 0.8$), moderate ($0.4 < \rho \leq 0.6$) or weak ($\rho \leq 0.4$). For all analyses a p value < 0.05 was considered to be statistically significant. All analyses were performed using Graphpad Prism 8.3 (Graphpad Software Inc.).

RESULTS

Measured plasma concentrations and calculated clearances for all study drugs are reported in the supplemental material (Fig. S2-S5).

Overall hemoadsorber clearance

Figure 1 depicts mean overall hemoadsorber clearance measured during the study period for each study drug under our experimental conditions. Except for ganciclovir (-0.3L/h), all values were positive indicating removal of the drug by the procedure. Five drugs (linezolid (4.6, SD 0.4), posaconazole (4.2 SD 0.7), fluconazole (4.0 SD 0.4), clindamycin (3.9 SD 0.2) and clarithromycin (3.3 SD 0.8) had a clearance > 3 L/h. Clearance was less than 2 L/h in all other molecules.

Relative part of hemoadsorber over total body clearance

Figure 2 presents, for each study drug, the impact of the hemoadsorber's associated clearance on total clearance. The hemoadsorber was associated with a major ($> 50\%$) increase in total clearance for fluconazole (+282%), linezolid (+115%), liposomal amphotericin B (+75%). The observed increase was minor (10-50%) for posaconazole (+32%), teicoplanin (+31%), anidulafungin

(+23%), piperacillin (+19%), flucloxacillin (+16%), metronidazole (+16%) and ciprofloxacin (+15%). It was insignificant (<10%) for clindamycin (+7%), clarithromycin (+5%), meropenem (+7%), ceftriaxone (+5%), tobramycin (+6%) and cefepime (+1%). Total body clearance was decreased for ganciclovir (-3%).

Kinetics of adsorber associated clearance

As depicted in figure 3 and 4, antibiotics clearance by the adsorber was not constant throughout the study period. For beta-lactams (piperacillin, flucloxacillin, ceftriaxone, cefepime and meropenem), it decreased progressively, even reaching negative values at the end of the experiment. For most other drugs (fluconazole, posaconazole, liposomal amphotericine B, linezolid, clarithromycin and teicoplanin), adsorber clearance decreased slowly but remained positive throughout the experiment. The clearance of some molecules became almost null respectively after one (anidulafungin, tobramycin, ganciclovir), two (metronidazole) or six (ciprofloxacin) hours of therapy and remained so for the rest of the experiment.

Factors associated with overall hemoadsorber clearance

Overall hemoadsorber clearance appeared to increase proportionally with the octanol–water partition coefficient of drugs (logP) i.e. their lipophilicity ($p=0.01$; $r^2=0.43$) and to a lesser degree with the logD value ($p=0.01$; $r^2=0.36$). There was no significant association with pKa, molecular electric charge, protein binding or molecular weight.

DISCUSSION:

Key findings

We have conducted an experimental study on 24 pigs to assess the impact of hemoadsorption with CytoSorb® on the pharmacokinetics of a panel of anti-infective drugs commonly utilized in the management of patients with sepsis. We found that for most drugs the procedure was associated with a relatively high clearance. However, when endogenous clearance was considered, we found that the additional clearance provided was major for only a few drugs (fluconazole, linezolid and liposomal amphotericin) but minor (<50%) or insignificant (<10%) for most other tested drugs. Importantly, we observed that the kinetics of clearance were not stable over time, with highest clearance observed during the first hours of therapy, often followed by a rapid decline and even desorption for some drugs (in particular beta-lactams). Finally, we found

that lipophilicity was the only pharmacokinetic factor moderately associated with overall hemoadsorber clearance.

Comparison with previous studies

These first systematic data from in-vivo experiments have to be compared to in-vitro studies. Koenig et al. observed strong adsorption of anti-infectives to the CytoSorb[®] adsorber with normal saline or human albumin as perfusion fluid. Adsorption was markedly reduced for two anti-infectives, namely meropenem and ciprofloxacin, if reconstituted blood was used for perfusion. However, Koenig et al. integrated the adsorber into a continuous veno-venous hemodialysis, so the results do not resemble pure hemoadsorption alone [14]. In addition, such data does not consider endogenous clearance and therefore the real impact of CytoSorb[®] on in-vivo pharmacokinetics is difficult to infer.

Other available data consist of case reports. The low clindamycin clearance observed in a young patient with refractory septic shock caused by Panton Valentin leucocidin producing methicillin resistant *Staphylococcus aureus* is consistent with the minimal effect of CytoSorb[®] on clindamycin overall clearance (+4.7%) observed in our study[16]. Similarly, immediate removal of teicoplanin with a saturable process was described by Dimski et al[17]. This is in full agreement with our data which demonstrate a high initial removal followed by progressive decline. Other observations have reported substantially lower linezolid peak levels during CytoSorb[®] adsorption [15, 19]. This is consistent with our finding of a doubled clearance of this molecule during therapy.

A high affinity of CytoSorb[®] for lipophilic molecules particularly in the range up to approximately 55 kDa has also been suspected [14, 20]. We found a weak but significant correlation between the logP and hemoadsorber associated clearance. However, this property alone cannot be used to predict the pharmacological impact of the hemoadsorber on drugs' pharmacokinetics. Indeed, when total clearance is considered, CytoSorb[®]'s effect on total clearance was limited for some drugs with high logP values (i.e. posaconazole, clarithromycin, flucloxacillin and clindamycin) while it was relatively important for some drugs with low logP values (fluconazole, linezolid or liposomal amphotericine B). This further highlights the need to consider hemoadsorber clearance relative to the *overall* clearance.

Finally, in vitro studies have described that CytoSorb[®]'s adsorptive capacities were saturable [17] and even to be subject to desorption[21]. Our study confirms these findings for the first time *in*

vivo. The ability of polyvinylpyrrolidone, the substance covering CytoSorb®'s beads, to adsorb and desorb molecules has in principle been described in technical settings other than hemoadsorption, and the substance was even proposed to be used as a component of pharmaceutical drug delivery systems [22].

Strengths and limitations

To the best of our knowledge, this is the first *in vivo* experimental study conducted to evaluate the impact of CytoSorb on the pharmacokinetics of anti-infective drugs. We tested a wide panel of medications commonly used in clinical practice. We have conducted sound analyses with thorough pharmacokinetics models. However, it has several limitations worth discussing.

First, our experimental study was conducted in healthy animals and our results might not directly be translatable to humans with septic shock. Indeed, drugs pharmacokinetics might vary from species to species. For instance, the clearances of ceftriaxone and teicoplanin measured in our model were higher than values previously reported in humans (respectively 8.4 vs 1.0 L/h and 5.4 vs 0.7-1.0 L/h) while that of linezolid was lower (8.8 vs 3.7 L/h)[23, 24]. In addition, and more importantly, pharmacokinetic parameters are known to be massively altered in sepsis (increased dilution volume, decreased protein binding etc) particularly in case of associated acute kidney injury or liver failure. Besides, in patients with sepsis, CytoSorb®'s adsorptive capacities might be modified, typically by competitive adsorption of other molecules such as pro-inflammatory mediators. The net effect of such competition is unknown and might lead to desorption or decreased drug adsorption. Further studies in humans with sepsis are therefore required to confirm or refute our findings.

Second, our protocol, enabled us to compute clearances obtained during the first 6 hours of CytoSorb adsorption, a duration lower than the manufacturer's recommended therapy (24 hours). Hence, a delayed effect of CytoSorb therapy for instance total desorption (i.e. beta-lactams) or more significant binding (i.e. posaconazole) cannot be ruled out. However, our kinetics analyses strongly suggest that the majority of the adsorption process takes place in the first hours of therapy.

Third, medications were not administered separately but as a group, and their pharmacokinetics might have influenced each other through binding competition. However, our findings for drugs classes (beta-lactams and azoles) were robust and consistent even if these medications were administered in different group combinations.

Implications for clinicians and policy makers

Based on our data, some tentative recommendations can be formulated. Obviously, human studies should be performed before these recommendations are implemented in clinical practice.

The net influence of hemoadsorption with CytoSorb® on beta-lactams' pharmacokinetics appears to be minimal. In addition, the initial adsorption followed by desorption might even be beneficial in terms of pharmacodynamics. Indeed, the intervention might optimize antimicrobial exposure and the activity of such time dependents agents. Hence, no dose adaptation would theoretically be required for this class of medications.

On the other hand, the observed initial removal of tobramycin, a concentration dependent drug, might be associated with a decrease in its antibacterial clinical efficacy. Here, a dose increase, accounting for the additional clearance, would be recommended. A temporary interruption of the adsorption therapy prior to drug administration could represent an alternative. The latter would have the theoretical advantage of enabling a high peak level (efficacy) and rapid removal (decreased toxicity).

For other drugs included in this study, biologic activity is related to the area under curve (AUC) divided by the minimum inhibitory concentration. This parameter was significantly decreased by CytoSorb® hemoadsorption for three study drugs: fluconazole, linezolid and liposomal amphotericin B. Hence, decreased biological activity is likely and drug dosage adaptation appears advisable. However, the required dose modification is likely to be minor. Indeed, for a drug to be considered as an inductor, a >125% increase in clearance is required while a >200% increase is required for it to be considered as a *moderate* inductor [25]. For all other drugs, the impact of the therapy on their total clearance was small and no major impact on pharmacodynamics would be expected and no dose adaptation would be required at least in patients with normal renal and liver function.

In acute kidney injury, or liver failure, the impact of hemoadsorption on total clearance is likely to increase making drug adaptation necessary. In these situations, therapeutic drug monitoring is strongly advised.

CONCLUSIONS:

Hemoadsorption with CytoSorb[®] has limited effect on the pharmacokinetics of the majority of tested drugs. However, the clearance of fluconazole, linezolid and liposomal amphotericin B appears to be increased by the procedure. These data need to be confirmed in humans with septic shock. In the meantime, therapeutic drug monitoring remains advisable.

List of abbreviations

- HA Hemoadsorption
- ECC Extracorporeal circulation
- logP octanol–water partition coefficient
- LogD octanol–water distribution coefficient

DECLARATIONS

Ethics approval and consent to participate

The study was approved by the State Office for Occupational Safety, Consumer Protection and Health – Department of Consumer Protection (Brandenburg, Potsdam, Germany), approval number 2347-3-2018. According to the German animal protection law, no additional approval by an Ethics Committee was necessary.

Consent for publication

Not applicable

Availability of data and material

The datasets used and/or analyzed during the current study are available from the corresponding author on reasonable request.

Competing interests

AS has received speaking honorarium from CytoSorbents Europe
JS is Senior Director Medical & Clinical of CytoSorbents Europe
DKM has received speaking honorarium from CytoSorbents Europe
All other authors stated that they have no conflict of interest.

Funding

The study was funded by CytoSorbents Europe, Berlin, Germany. The company was involved with the study design. However, data interpretation was conducted independently by AS, PA, TB

and DKM. The company was allowed to read the draft of the manuscript before submission but had no influence on its content or decision for submission.

Authors' contributions

AGS: Participated in study design, data interpretation and analyses and drafted the manuscript

PA: Performed pharmacokinetic analyses, participated in data interpretation and critically reviewed the manuscript

JS: Participated in study design and critically reviewed the manuscript

MS: Performed the animal experiment and critically reviewed the manuscript

HZ: Planned and monitored the animal experiment and critically reviewed the manuscript

TB: Participated in pharmacokinetic analyses and critically reviewed the manuscript

DKM: Participated in study design, data interpretation and critically reviewed the manuscript

All authors read and approved the final manuscript and agree to be personally accountable for their contribution.

Acknowledgements

The authors thank Dr Michael Zoller (Munich, Germany) for advising on the interpretation of the results. They also want to thank Dr Sven Giersdorf from CytoSorbents Europe for the support.

FIGURES AND TABLES TITLES AND LEGENDS

Figure 1: Total Adsorber Clearance: Clearance attributable to CytoSorb (mean and SD). Values were calculated based on pre and post adsorber measurements at different timepoints. AmphoB lipo: liposomal amphotericin B.

Figure 2: Additional clearance provided by adsorber: The white areas of the bars represent the endogenous drug clearance (without adsorber) while the grey areas represent the additional clearance provided by the adsorber under the experimental conditions. Percentage refer to the relative increase in clearance associated with adsorber insertion. AmphoB lipo: liposomal amphotericin B.

Figure 3: Kinetics of adsorber clearance of beta-lactams (panel a) and antifungals (panel b): Bars represent instantaneous clearance at the different study time points: 30 min (0.5h), 1, 2 and 3 hours after therapy initiation as well as the last measure obtained (6 hours). Dotted lines represent total clearance (also represented in figure 2). Reported values are mean and standard deviation. AmphoB lipo: liposomal amphotericin B.

Figure 4: Kinetics of adsorber clearance of other anti-infective agents: Bars represent instantaneous clearance at the different study time points: 30 min (0.5h), 1, 2 and 3 hours after therapy initiation as well as the last measure obtained (6 hours). Dotted lines represent total clearance (also represented in figure 2). Reported values are mean and standard deviation.

Table 1: Experimental groups: pigs received different combinations of drugs, followed by initiation of an extra-corporeal circulation with (cases) or without (sham, control) a CytoSorb cartridge. Three animals were randomly allocated to each group. All drugs were administered intravenously one hour before extra-corporeal circulation initiation.

Supplemental Material

Figure S1: Extra-corporeal circuit: According to study protocol, 1hr following antibiotic administration, study animals were randomly allocated to either CytoSorb hemoadsorption or sham procedure on a 1:1 basis. For animals allocated to sham procedure, a similar circuit without hemoadsorber was used.

Figure S2: Plasma concentrations (panel A) and calculated clearances (panel B) for beta-lactams. Full symbols correspond to cases (CytoSorb Hemoadsorption) and open symbols to controls

Figure S3: Plasma concentrations (panel A) and calculated clearances (panel B) for antifungals. Full symbols correspond to cases (CytoSorb Hemoadsorption) and open symbols to controls

Figure S4: Plasma concentrations (panel A) and calculated clearances (panel B) for other drugs (part 1). Full symbols correspond to cases (CytoSorb Hemoadsorption) and open symbols to controls

Figure S5: Plasma concentrations (panel A) and calculated clearances (panel B) for other drugs (part 2). Full symbols correspond to cases (CytoSorb Hemoadsorption) and open symbols to controls

REFERENCES

1. Rudd KE, Johnson SC, Agesa KM, Shackelford KA, Tsoi D, Kievlan DR, Colombara DV, Ikuta KS, Kissoon N, Finfer S *et al*: **Global, regional, and national sepsis incidence and mortality, 1990-2017: analysis for the Global Burden of Disease Study**. *Lancet* 2020, **395**(10219):200-211.
2. Singer M, Deutschman CS, Seymour CW, Shankar-Hari M, Annane D, Bauer M, Bellomo R, Bernard GR, Chiche JD, Coopersmith CM *et al*: **The Third International Consensus Definitions for Sepsis and Septic Shock (Sepsis-3)**. *JAMA* 2016, **315**(8):801-810.
3. Tisoncik JR, Korth MJ, Simmons CP, Farrar J, Martin TR, Katze MG: **Into the eye of the cytokine storm**. *Microbiol Mol Biol Rev* 2012, **76**(1):16-32.
4. Poli EC, Rimmelé T, Schneider AG: **Hemoadsorption with CytoSorb((R))**. *Intensive Care Med* 2018.
5. Rimmelé T, Kellum JA: **Clinical review: Blood purification for sepsis**. *Critical Care* 2011, **15**(1):205-205.
6. Malard B, Lambert C, Kellum JA: **In vitro comparison of the adsorption of inflammatory mediators by blood purification devices**. *Intensive Care Medicine Experimental* 2018, **6**(1):12.
7. Schädler D, Pausch C, Heise D, Meier-Hellmann A, Brederlau J, Weiler N, Marx G, Putensen C, Spies C, Jörres A *et al*: **The effect of a novel extracorporeal cytokine hemoadsorption device on IL-6 elimination in septic patients: A randomized controlled trial**. *PLoS One* 2017, **12**(10):e0187015.
8. Friesecke S, Träger K, Schitteck GA, Molnar Z, Bach F, Kogelmann K, Bogdanski R, Weyland A, Nierhaus A, Nestler F *et al*: **International registry on the use of the CytoSorb® adsorber in ICU patients**. *Medizinische Klinik - Intensivmedizin und Notfallmedizin* 2017.
9. Friesecke S, Stecher S-S, Gross S, Felix SB, Nierhaus A: **Extracorporeal cytokine elimination as rescue therapy in refractory septic shock: a prospective single-center study**. *J Artificial Organs* 2017, **20**(3):252-259.
10. Kogelmann K, Jarczak D, Scheller M, Drüner M: **Hemoadsorption by CytoSorb in septic patients: a case series**. *Critical Care* 2017, **21**(1):74.
11. Brouwer WP, Duran S, Kuijper M, Ince C: **Hemoadsorption with CytoSorb shows a decreased observed versus expected 28-day all-cause mortality in ICU patients with septic shock: a propensity-score-weighted retrospective study**. *Crit Care* 2019, **23**(1):317.
12. Rhodes A, Evans LE, Alhazzani W, Levy MM, Antonelli M, Ferrer R, Kumar A, Sevransky JE, Sprung CL, Nunnally ME *et al*: **Surviving Sepsis Campaign: International Guidelines for Management of Sepsis and Septic Shock: 2016**. *Crit Care Med* 2017, **45**(3):486-552.
13. Reiter K, Bordoni V, Dall'Olio G, Ricatti MG, Soli M, Ruperti S, Soffiati G, Galloni E, D'Intini V, Bellomo R *et al*: **In vitro removal of therapeutic drugs with a novel adsorbent system**. *Blood Purif* 2002, **20**(4):380-388.
14. König C, Rohr AC, Frey OR, Brinkmann A, Roberts JA, Wichmann D, Braune S, Kluge S, Nierhaus A: **In vitro removal of anti-infective agents by a novel cytokine adsorbent system**. *Int J Artif Organs* 2019, **42**(2):57-64.

15. Zoller M, Dobbeler G, Maier B, Vogeser M, Frey L, Zander J: **Can cytokine adsorber treatment affect antibiotic concentrations? A case report.** *J Antimicrob Chemother* 2015, **70**(7):2169-2171.
16. Poli EC, Simoni C, Andre P, Buclin T, Longchamp D, Perez MH, Ferry T, Schneider AG: **Clindamycin clearance during Cytosorb((R)) hemoadsorption: A case report and pharmacokinetic study.** *Int J Artif Organs* 2019, **42**(5):258-262.
17. Dimski T, Brandenburger T, MacKenzie C, Kindgen-Milles D: **Elimination of glycopeptide antibiotics by cytokine hemoadsorption in patients with septic shock: A study of three cases.** *Int J Artif Organs* 2020, **43**(12):753-757.
18. Kilkenny C, Browne W, Cuthill IC, Emerson M, Altman DG, National Centre for the Replacement R, Reduction of Animals in R: **Animal research: reporting in vivo experiments--the ARRIVE guidelines.** *Journal of cerebral blood flow and metabolism : official journal of the International Society of Cerebral Blood Flow and Metabolism* 2011, **31**(4):991-993.
19. Morris C, Gray L, Giovannelli M: **Early report: The use of Cytosorb haemabsorption column as an adjunct in managing severe sepsis: initial experiences, review and recommendations.** *J Intensive Care Soc* 2015, **16**(3):257-264.
20. De Schryver N, Hantson P, Haufried V, Dechamps M: **Cardiogenic Shock in a Hemodialyzed Patient on Flecainide: Treatment with Intravenous Fat Emulsion, Extracorporeal Cardiac Life Support, and CytoSorb(R) Hemoadsorption.** *Case Rep Cardiol* 2019, **2019**:1905871.
21. Lang CN, Sommer MJ, Neukamm MA, Staudacher DL, Supady A, Bode C, Duerschmied D, Lothar A: **Use of the CytoSorb adsorption device in MDMA intoxication: a first-in-man application and in vitro study.** *Intensive Care Med Exp* 2020, **8**(1):21.
22. Franco P, De Marco I: **The Use of Poly(N-vinyl pyrrolidone) in the Delivery of Drugs: A Review.** *Polymers (Basel)* 2020, **12**(5).
23. Goodman L, Brunton L, Chabner B, SKnollmann C: **Goodman & Gilman's Pharmacological Basis of Therapeutics.** *New York: McGraw-Hill* 2011.
24. Targocid_information: <https://www.swissmedicinfo.ch/Accept.aspx?Lang=EN>. *Swissmedicinfo* 2017.
25. European_Medicine_Agency: https://www.ema.europa.eu/en/documents/scientific-guideline/guideline-investigation-drug-interactions-revision-1_en.pdf. 2012.

Figures

Figure 1

Total Adsorber Clearance: Clearance attributable to CytoSorb (mean and SD). Values were calculated based on pre and post adsorber measurements at different timepoints. AmphoB lipo: liposomal amphotericin B.

Figure 2

Additional clearance provided by adsorber: The white areas of the bars represent the endogenous drug clearance (without adsorber) while the grey areas represent the additional clearance provided by the adsorber under the experimental conditions. Percentage refer to the relative increase in clearance associated with adsorber insertion. AmphoB lipo: liposomal amphotericin B.

Figure 3

Kinetics of adsorber clearance of beta-lactams (panel a) and antifungals (panel b): Bars represent instantaneous clearance at the different study time points: 30 min (0.5h), 1, 2 and 3 hours after therapy initiation as well as the last measure obtained (6 hours). Dotted lines represent total clearance (also represented in figure 2). Reported values are mean and standard deviation. AmphoB lipo: liposomal amphotericin B.

Figure 4

Kinetics of adsorber clearance of other anti-infective agents: Bars represent instantaneous clearance at the different study time points: 30 min (0.5h), 1, 2 and 3 hours after therapy initiation as well as the last measure obtained (6 hours). Dotted lines represent total clearance (also represented in figure 2). Reported values are mean and standard deviation.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [SupplementalMaterial.docx](#)