

# A new 3D printing porous trabecular titanium metal acetabular cup for primary total hip arthroplasty: a minimum 2-year follow-up of 92 consecutive patients

**Xiao Geng**

Peking University Third Hospital

**Yang Li**

Peking University Third Hospital

**Feng Li**

Peking University Third Hospital

**Xinguang Wang**

Peking University Third Hospital

**Ke Zhang**

Peking University Third Hospital

**Zhongjun Liu**

Peking University Third Hospital

**Hua Tian** (✉ [tianhua@bjmu.edu.cn](mailto:tianhua@bjmu.edu.cn))

Peking University Third Hospital <https://orcid.org/0000-0001-7139-3372>

---

## Research article

**Keywords:** arthroplasty, hip, 3D printing, trabecular titanium acetabular cup, outcome

**Posted Date:** August 15th, 2020

**DOI:** <https://doi.org/10.21203/rs.3.rs-17827/v2>

**License:**   This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

---

**Version of Record:** A version of this preprint was published on September 4th, 2020. See the published version at <https://doi.org/10.1186/s13018-020-01913-1>.

# Abstract

**Background** Aseptic cup loosening is still one of the main reasons leading to acetabular cup failures. 3D printing porous trabecular titanium metal acetabular cup may provide good initial stability and secondary fixation because of its highly interconnected, porous structure. Few large sample studies have reported the clinical outcomes of Electron Beam Melting (EBM) porous titanium acetabular cup in Chinese population.

**Methods** We retrospectively collected and analyzed the clinical data of a total of 92 consecutive patients between January 2013 and November 2017, with an average follow-up of  $48.2 \pm 3.6$  months. Clinical outcomes included Harris Hip Score (HHS), the Western Ontario and McMaster Universities (WOMAC) Osteoarthritis Index, satisfaction rate and cup survival rate were evaluated. Radiographic assessments were conducted to evaluate osteointegration.

**Results** HHS scores improved significantly while the WOMAC score decreased significantly at the latest follow-up ( $p < 0.001$ ). The satisfaction rate (prevalence of satisfied or very satisfied) was 91.3%. No acetabular cup failures occurred. Radiolucent lines appeared in 15 cases (18 hips), and disappeared in six months. No cup loosening signs found until the last follow up. The overall survival rate of implantation is 99.1% (cup survival rate 100%).

**Conclusion** The new EBM-produced 3D ACT™ cup demonstrated us its favorable short to mid-term clinical outcomes in Chinese THA patients. It can provide high acetabular cup survival rate, great clinical improvements and excellent biological fixation. Further investigations are needed to confirm its long-term outcomes.

## Background

Cementless acetabular cups have been utilized more and more common in total hip arthroplasty (THA) in recent years<sup>1,2</sup>. Although it has provided remarkably successful clinical outcomes in primary THAs, every year many cup failures, especially aseptic cup loosening, occur and lead to intractable revisions and heavy medical burdens.<sup>3</sup> Thus, surgeons and engineers have been working on prosthetic designs and materials for many years to improve the acetabular cups' primary mechanical stability and secondary bone fixation<sup>4-6</sup>.

The mechanical fixation and bone integration between the surface of the cup and the host bone interface are important factors to maintain initial and long-term stability of the cups<sup>7,8</sup>. But it is technically limited because the traditional two-step porous coated cup surface cannot be designed integrated with the solid layer and optionally to achieve a better bone osseointegration with the host bone by ideal porous structure<sup>9</sup>.

As the 3D-printing technology widely applied to design various orthopedic implants, a few studies have showed excellent clinical and radiological outcomes and survivorship of the porous trabecular titanium

metal acetabular component via EBM, which is an important branch of 3D printing technology, especially when patients with poor bone quality in primary and revision THA surgeries<sup>10,11</sup>. With an average porosity of 65% and a mean pore diameter of 640  $\mu\text{m}$ , a kind of 3D printing porous titanium acetabular component, DELTA-TT™ cup, was reported to obtain satisfactory mid-term clinical and radiographic outcomes<sup>12</sup>. The continuity between the porous surface and solid parts broke through the limitations of the traditional porous coatings, which can provide a higher resistance to detachment and corrosion of the two parts. Also, the EBM trabecular titanium has showed good osteoinductive and osteoconductive performances by a series of vitro and animal experiments<sup>13-15</sup>.

The 3D ACT™ cups (Fig. 1) was started utilized in primary THAs and hip revisions since year 2013 and performed good clinical outcomes refer to our clinical practice. Few large sample studies have reported the clinical outcomes of this kind of EBM porous titanium acetabular cup. The aim of this study was to investigate the short to mid-term clinical and radiographic outcomes and satisfaction of 92 consecutive patients who underwent primary THA using a new 3D printing produced trabecular titanium cups, relatively for complex THA cases.

## Methods

### General Data

After the approval of the Institutional Review Board was obtained for this study, we retrospectively collected the data from a total of 103 patients (120 hips) who underwent primary total hip arthroplasty used the 3D ACT porous titanium (Ti6Al4V) trabecular acetabular cups between January 2013 and November 2017. Inclusion criteria: Males and females > 18 years old, with a body mass index(BMI)  $\leq$  40  $\text{kg}/\text{m}^2$ , diagnosed with primary hip osteoarthritis(OA), fracture of femoral neck(FFN) or secondary hip osteoarthritis due to developmental dysplasia of the hip(DDH), hip avascular necrosis (AVN), rheumatoid arthritis(RA), ankylosing spondylitis(AS) and coxa plana with a minimum 2-year clinical follow-up. Exclusion criteria: Severe vascular diseases of the lower extremity, neuromuscular diseases, infectious diseases, severe dysfunctions of major organs, severe neurosensory deficiencies caused by spinal diseases.

3 patients died of cancers and 8 patients were lost to follow-up before the 2-year follow-up. None of these 11 patients were deceased due to THA associated diseases or underwent revision until our last evaluation. Finally, 40 males (47 hips) and 52 females (61 hips) were enrolled in our study, with 2–6 year follow up. Patient demographics and preoperative clinical data were listed in Table 1. 76 unilateral procedures (40 left-sided and 36 right-sided) were performed and 14 patients underwent a bilateral two-stage procedure, while remain 2 patients underwent a simultaneous bilateral THA surgery. Original diagnosis included OA in 8 (7.4%) cases, DDH in 52 (48.1%) cases, RA in 3(2.7%) cases, AVN in 36 (33.3%) cases, FFN in 4 (3.7%) cases, AS in 4 (3.7%) cases and coxa plana in 1 (0.9%) cases.

Table 1  
Demographic and clinical data of the patients

		Mean	SD	p
Age(years)	Male (n = 40)	59.8	9.8	0.58
	Female (n = 52)	58.7	8.9	
BMI(Kg/m2)	Male (n = 40)	26.0	4.9	0.51
	Female (n = 52)	26.8	6.4	
Pre-Harris hip score		45.2	4.8	
Pre-WOMAC score		54.4	7.3	

## Components And Surgical Procedures

Cementless 3D ACT cups (Fig. 1), ML stems, BIOLOX® Delta ceramic (CeramTec GmbH, Plochingen, Germany) femoral heads and ceramic-on-polyethylene bearing surfaces were used in all cases. The structure of 3D ACT cup is characterized by a 1.5 mm thick porous layer, an average porosity of 80%, pore size of 600–800 µm, and lower modulus of elasticity which is equal to that of human cancellous bone. The rough surface, with coefficient of friction on cancellous bone of 1.08 and on cortical bone of 0.93, could theoretically obtain a better bone-implant initial fixation. There are 16 specifications of the 3D ACT cups, and the diameter of which is 40–70 mm, with 2 mm interval. Unlike the two-step coated surface of the traditional cup, the ACT cup was naturally integrated (Fig. 2), reducing the detachment risk of the metal solid layer and the surface.

56 posterolateral approaches and 52 modified Hardinge approaches were performed by 6 senior surgeons due to their preferences. All cups were implanted by 1 mm press-fit technology according to the manufacturers' recommendations with additional screw fixation if necessary. No intraoperative complications occurred, although there were several considerable tough cases, like Crowe type IV DDH and RA stiffness hip combined with severe osteoporosis. In one case, a patient with acetabular protrusion secondary to severe RA, two half-moon shaped 3D printing augments (50mm\*15 mm, Fig. 3), designed by the same EBM technology, were used to repair bone defect before the implantation of the cup. All cups obtained good primary stabilities after implantation.

## Clinical And Radiographic Assessments

Patients' clinical and radiographic data, as well as complications were collected preoperatively and 3, 6, 12 months, then annually after 1 year postoperatively. HHS<sup>16</sup>, WOMAC score<sup>17</sup> are evaluated as clinical outcomes. A self-assessment satisfaction questionnaire (sorted by very dissatisfied, dissatisfied, neutral,

satisfied or very satisfied) was used to assess patients' subjective satisfaction rate<sup>18</sup>. Postoperative complications were recorded as well.

An independent clinical staff who was blinded to each patient's clinical condition assessed the radiographic results, included radiolucent lines, osteolysis and sclerosis according to the DeLee and Charnley's<sup>19</sup> definition of three zones and the bone ingrowth criteria defined by the Anderson Orthopedic Research Institute<sup>20</sup> to evaluate bone osseointegration and acetabular component stability. Cup loosening criteria: a progressive radiolucent line wider than 2 mm or cup migration more than 5 mm or cup inclination change more than 5°. Cup osseointegration criteria: (1) the absence of radiolucent lines; (2) the presence of superolateral buttresses; (3) the presence of medial stress-shielding; (4) the presence of radial trabeculae; and (5) the presence of inferomedial buttresses. No less than 3 signs can be determined as radiographic osseointegration.

## Statistical Assessments

SPSS (Version 20, IBM Corp.) was used for the statistical analysis. All continuous variables were expressed as mean and standard deviation and compared by t-tests if they are normally distributed. An appropriate chi-square test or Fisher's exact test was carried out to compare categorical variables. Statistical significance level was defined at p value < 0.05.

## Results

A total of 92 consecutive patients (108 hips) were enrolled in statistical analysis. 75(81.5%) patients completed the clinical and radiographic evaluation by regular follow up and data of remain 17(18.5%) patients were collected via telephones or / and online communications.

The mean follow-up period was  $48.2 \pm 3.6$  months. 40(43.5%) were male and 52(56.5%) were female, with a mean age of  $59.8 \pm 9.8$  years old. Average Body Mass Index(BMI) was  $26.0 \pm 4.9$  kg/m<sup>2</sup>. HHS improved remarkably, from  $45.2 \pm 4.8$  preoperatively to  $95.8 \pm 6.0$  postoperatively, along with WOMAC score decreased significantly from  $54.4 \pm 7.3$  preoperatively to  $11.2 \pm 4.2$  at the last follow-up (Table 2). The satisfaction rate (prevalence of satisfied or very satisfied) was 91.3% (84 patients).

Table 2  
Follow-up of the clinical outcomes of patients using ACT cups(Mean ± SD)

	Pre-op*	6 months post-op**	1 year post-op	2 years post-op	last follow up
Harris Hip Score	54.4 ± 7.3	16.3 ± 6.2†	13.6 ± 5.8†	10.5 ± 3.7†	11.2 ± 4.2†
WOMAC Score	45.2 ± 4.8	90.5 ± 5.3†	95.1 ± 4.5†	96.7 ± 5.2†	95.8 ± 6.0†
†:p < 0.001 when compared with preoperative scores					
*Pre-op: preoperative					
**Post-op: postoperative					

Dislocation occurred in 1(0.9%) female patients with Crowe IV DDH immediately after surgery, and problem was solved after an open reduction on the same day. During the first 2 years follow up, none of the patients complained unbearable pain or dysfunction that stopped them to return to daily routine activities. 1(0.9%) female patient presented to us 30 months postoperatively because of the aseptic loosening of the femoral stem, then a revision of the femoral stem only was taken on her because we found the ACT cup fixed firmly during the surgical exploration. They were the only 2 patients who declared very dissatisfied and the other 106 (98.1%) of the patients were satisfied with the treatments. Besides, no cup revisions (cup survival rate was 100%) and other complications occurred during the follow up period, and the overall Kaplan-Meier cumulative survivorship was 99.1% at the latest follow-up (Fig. 4).

Radiographic assessments were carried out at the latest follow-up, all cups showed excellent osseointegration with at least 3 signs according to the Anderson criteria. Radiolucent lines appeared within 2 mm in 15 cases (18 hips), and disappeared in 6 months postoperatively (Fig. 5). No significant cup migration or inclination met the cup loosening criteria.

## Discussion

With the operation technology becoming more and more mature, surgeons and engineers are paying more and more attention to prosthesis designs and modifications with better biomechanics properties to improve THA clinical outcomes<sup>21,22</sup>. The absence of locking mechanism of bone cement may lead to high early aseptic loosening, so cementless acetabular and femoral components has become more and more popular in THA surgeries<sup>23</sup>. Multiples studies report short to mid-term clinical outcomes with a large variety of cementless acetabular cups<sup>24-26</sup>. As is known to all, both initial mechanical stability and long-term biological fixation are essential parts to achieve durable survivorship of acetabular cups. Thus, the biological coating technology becomes a highlight issue. Traditional coated surfaces of the acetabular

cup include cobalt chromium molybdenum sintered beads, diffusion bonded fibre metal mesh, cancellous structured titanium and titanium-sprayed plasma, demonstrated a certain osseointegration ability. However, the limitations of these techniques may occur in case of poor host bone qualities and conditions, such as osteonecrosis, osteoporosis, dysplasia and massive bone defect.

The manufacturing process of 3D printing acetabular cup is completely different from that of traditional cup. In traditional reduction casting process, the interface between the solid layer and the coated surface of the acetabular cups may cause detachment and corrosion, resulting to cup failures. But 3D printing, via additive manufacturing process, has made it easier to individualize product design and manufacturing<sup>27</sup>. Thus, the 3D ACT EBM cup, with more appropriate porosity, ideal porous diameter, lower modulus of elasticity and better coefficient of friction, may provide better solutions for various kinds of primary THAs.

It has been formerly confirmed that EBM manufactured implants, with interconnected porosity and rough surface, could provide a viable medium for bone ongrowth and ingrowth by creating a stable porous construct while also maintaining the mechanical strength in vitro and animal studies<sup>13-15</sup>. Perticarini<sup>12</sup> et al first reported the successful utilization of Trabecular Titanium acetabular cups in European patients, among which are mainly primary hip OA, as well as AVN and DDH patients, and significant functional recovery and pain relief were recorded in all 133 cases at minimum 5 years follow-up.

All cups obtained initial stabilities and long-term fixation in our study. No cup loosening determined according to the DeLee and Charnley's definition of three zones and excellent bone ingrowth defined by the Anderson Orthopedic Research Institute in our study. Early-stage radiolucent lines within 2 mm showed in 15 cases gradually disappeared at 3 to 6 months postoperatively, which represented efficient bone ingrowth to the porous metal structure. The interface micromotion can be reduced if the coefficient of friction get increased, which can enhance primary fixation and contribute to secondary fixation<sup>28,29</sup>. Great bone integration property is considered as the EBM-produced cups' peculiar advantage related to: its extremely more rough surface and higher coefficient of friction on cancellous bone of 1.08, compared with that in traditional sintered beads (0.5); and its porous, solid and special interconnected structure mimic the trabecular morphology of the natural bone. These properties do have the advantages of osteoinduction and osteoconduction by stimulate vascularisation, osteoblast proliferation and differentiation in previous study<sup>30</sup>.

It should be mentioned that in China, the biggest developing country in the world, we often encounter severe type of end-stage hip diseases, such as Crowe IV DDH and acetabular protrusion with great bone defect, due to China's national economical and healthcare conditions. In our study, we are glad to see encouraging great clinical and radiographic outcomes in difficult cases such as Crowe IV DDH, acetabular protrusion with large bone defect and RA patients with severe osteoporosis. Remarkable HHS and WOAMC score improvements and high satisfaction rate ensured the cup's efficiency among Chinese patients. No acetabular cup failures occurred in any cases, with 100% survival rate till the last follow up, together showed an excellent short to mid-term outcomes of the 3D ACT cups.

This study does have several limitations. First, in this retrospective study, 11/103 of the patients lost to follow up. Second, no controlled groups enrolled in this study, and we are about to carry out a prospective randomized controlled trial for higher-level evidence. Third, cases from 6 different surgeons enrolled which may confound the results. Fourth, computer tomography (CT) scans and bone densitometry evaluations as well as relevant laboratory examinations were not conducted.

## **Conclusions**

As far as be concerned, the application of EBM-produced 3D ACT cup demonstrated us its favorable short to mid-term clinical outcomes in Chinese THA patients. It can provide high acetabular cup survival rate, great clinical improvements and excellent biological fixation. More investigations of the outcomes of this EBM-produced porous trabecular titanium cup are needed in larger volume of patients and at longer term follow-up.

## **Abbreviations**

EBM: Electron Beam Melting; HHS: Harris Hip Score; WOMAC: the Western Ontario and McMaster Universities; BMI: body mass index; OA: osteoarthritis; FFN: fracture of femoral neck; DDH: developmental dysplasia of the hip; AVN: avascular necrosis; RA: rheumatoid arthritis; AS: ankylosing spondylitis

## **Declarations**

### **Ethics approval and consent to participate**

This study was approved by the hospital ethics committee, and assigned the protocol number 14-27-QX-GK. Each author certifies that all the investigations were conducted in conformity with ethical principles.

### **Consent for publication**

We obtained a written informed consent from the patients for publishing the de-identified images.

### **Availability of data and materials**

Upon request, raw data can be provided.

### **Competing interests**

The authors have no competing interests to declare.

### **Funding**

No funding was received for this work.

### **Authors' contributions**

Study conception and design: XG, YL, KZ, ZL, HT. Data analysis: XG, XW. Manuscript preparation: XG, YL, XW, HT.

## Acknowledgements

Not applicable.

## References

1. Illgen R, Rubash H E. The Optimal Fixation of the Cementless Acetabular Component in Primary Total Hip Arthroplasty. *The Journal of the American Academy of Orthopaedic Surgeons*. 2002;10(1):43-56.
2. Banerjee, S.; Issa, K.; Kapadia, B. H.; Pivec, R.; Khanuja, H. S.; and Mont, M. A.: Highly-porous metal option for primary cementless acetabular fixation. What is the evidence? *Hip international : the journal of clinical and experimental research on hip pathology and therapy*. 2013;23(6):509-21.
3. Howard JL, Kremers HM, Loechler YA, Schleck CD, Harmsen WS, Berry DJ, et al. Comparative survival of uncemented acetabular components following primary total hip arthroplasty. *J Bone Joint Surg Am*. 2011;93(17):1597–604.
4. Huo M H, Dumont G D, Knight J R, et al. What's New in Total Hip Arthroplasty. *J Bone Joint Surg Am*. 2012;94(18):1721-7.
5. Steno, B.Kokavec, M., and Necas, L.: Acetabular revision arthroplasty using trabecular titanium implants. *International Orthopaedics*. 2015;39(3): 389-95.
6. Wegrzyn J, Kaufman K R, Hanssen A D, et al. Performance of Porous Tantalum vs. Titanium Cup in Total Hip Arthroplasty: Randomized Trial with Minimum 10-Year Follow-Up. *The Journal of Arthroplasty*. 2015;30(6):1008-13.
7. Marin E, Fusi S, Pressacco M, et al. Characterization of cellular solids in Ti6Al4V for orthopaedic implant applications: Trabecular titanium. *J Mech Behav Biomed Mater*. 2010;3(5):373-81.
8. Macheras, G. A., Lepetsos, P., Leonidou, A. O., Anastasopoulos, P. P., Galanakos, S. P., & Poultsides, L. A. Survivorship of a Porous Tantalum Monoblock Acetabular Component in Primary Hip Arthroplasty With a Mean Follow-Up of 18 Years. *The Journal of Arthroplasty*. 2017;32(12):3680–4.
9. Li X, Wang C, Zhang W, et al. Fabrication and characterization of porous Ti6Al4V parts for biomedical applications using electron beam melting process. *Materials Letters*. 2009;63(3-4):403-5.
10. De Meo, F.; Cacciola, G.; Bellotti, V.; Bruschetta, A.; and Cavaliere, P.: Trabecular Titanium acetabular cups in hip revision surgery: mid-term clinical and radiological outcomes. *HIP International*. 2018;28(2): 61-5.
11. Massari L, Bistolfi A, Grillo P P, et al. Periacetabular Bone Densitometry after Total Hip Arthroplasty with Highly Porous Titanium Cups: A 2-Year Follow-Up Prospective Study. *HIP International*. 2017;27(6):551-7.
12. Perticarini L, Zanon G, Rossi S M P, et al. Clinical and radiographic outcomes of a trabecular titanium? acetabular component in hip arthroplasty: results at minimum 5 years follow-up. *BMC*


- Musculoskeletal Disorders. 2015;16(1):375-80.
13. Ponader S, Vairaktaris E, Heini P, et al. Effects of topographical surface modifications of electron beam melted Ti-6Al-4V titanium on human fetal osteoblasts. *Journal of Biomedical Materials Research Part A*. 2008;84(4):1111-9.
  14. Heini P, Lenka Müller, K?rner C , et al. Cellular Ti-6Al-4V structures with interconnected macro porosity for bone implants fabricated by selective electron beam melting. *Acta Biomaterialia*. 2008;4(5):1536-44.
  15. Taniguchi , Fujibayashi S , Takemoto M , et al. Effect of pore size on bone ingrowth into porous titanium implants fabricated by additive manufacturing: An in vivo experiment. *Materials Science & Engineering C*. 2015;59:690-701.
  16. Harris W H. Traumatic Arthritis of the Hip After Dislocation and Acetabular Fractures: Treatment by Mold Arthroplasty. An End-Result Study Using a New Method of Result Evaluation. *The Journal of Bone and Joint Surgery*. 1969;51(4):737-55.
  17. Bellamy N, Buchanan W W, Goldsmith C H, et al. Validation of WOMAC; a health status instrument for measuring clinically important patient Relevant outcomes to anti-rheumatic drug therapy in patients with osteoarthritis of the hip or knee. *Journal of Rheumatology*. 1988;15(12):1833-40.
  18. Tang H, Du H, Tang Q, et al. Chinese Patients Satisfaction With Total Hip Arthroplasty: What Is Important and Dissatisfactory? *The Journal of Arthroplasty*. 2014;29(12):2245-50.
  19. Delee J G, Charnley J. Radiological Demarcation of Cemented Sockets in Total Hip Replacement. *Clin Orthop Relat Res*. 1976;121(121):20-32.
  20. Moore MS, McAuley JP, Young AM, Engh CA, Sr. Radiographic signs of osseointegration in porous-coated acetabular components. *Clin Orthop Relat Res*. 2006;444:176-83.
  21. Engh C A, Claus A M, Hopper R H, et al. Long-Term Results Using the Anatomic Medullary Locking Hip Prosthesis. *Clin Orthop Relat Res*. 2001;393:137-46.
  22. Engh CA, Hopper RH Jr. The odyssey of porous-coated fixation. *The Journal of Arthroplasty*. 2002;17(suppl 1):102-7.
  23. Aigner C. 10 years results with the corund-blasted Zweymüller titanium alloy threaded acetabular cup. *Z Orthop Ihre Grenzgeb*. 1998;Mar-Apr (2): 110-4.
  24. Komarasamy B, Vadivelu R, Bruce A, et al. Clinical and radiological outcome following total hip arthroplasty with an uncemented trabecular metal monoblock acetabular cup. *Acta orthopaedica Belgica*. 2006;72(3):320-5.
  25. Mid-Term Clinical and Radiographic Outcomes of Porous Tantalum Modular Acetabular Components for Hip Dysplasia. *The Journal of Arthroplasty*. 2015;30(4):607-10.
  26. A Randomized Trial Comparing Acetabular Component Fixation of Two Porous Ingrowth Surfaces Using RSA. *The Journal of Arthroplasty*. 2013;28(8):48-52.
  27. Langridge B, Momin S, Coumbe B, et al. Systematic Review of the Use of 3-Dimensional Printing in Surgical Teaching and Assessment. *Journal of Surgical Education*. 2017;75(1):209-21.

28. Saleh K J, Bear B, Bostrom M, et al. Initial stability of press-fit acetabular components: an in vitro biomechanical study. *American journal of orthopedics* (Belle Mead, N.J.). 2008;37(10):519-22.
29. Pilliar R M, Lee J M, Maniopoulos C. Observations on the Effect of Movement on Bone Ingrowth into Porous-Surfaced Implants. *Clin Orthop Relat Res*. 1986; 208:108-13.
30. Devine D, Arens D, Burelli S, Bloch HR, Boure L. In vivo evaluation of the osteointegration of new highly-porous Trabecular Titanium™. *J Bone Joint Surg Br*. 2012;94-B(Suppl XXXVII):201.

## Figures




1a



1b

### Figure 1

The picture shows the 3D ACT EBM-produced trabecular titanium acetabular cup (a) and the SEM image of its cellular solid structure (b).




**Figure 2**

The picture shows the interface between the two layers of traditional cup (Left) and the integration EBM porous structure (Right).







**Figure 3**

The picture shows a 64-year-old male patient with acetabular protrusion caused by severe RA (a), underwent a 3D ACT primary THA with 3D printing augments for bone defect repair (b).



**Figure 4**

The overall Kaplan-Meier cumulative survivorship of implantation.



### Figure 5

The picture shows the radiographic assessments of a 61-year-old female patient with Crowe IV DDH, preoperatively(a) and 2 weeks (b), 3 months (c), 6 months (d) and 4 years(e) postoperatively. Radiolucent line (appeared on 4b and 4c, black arrow) vanished at about 6 months postoperatively, represents new bone ingrowth.