Multi-Volume Hemacytometer

Thunyaporn Ravangnam

Konyang University

Il Doh

Korea Research Institute of Standards and Science

Dong Woo Lee (✉ mems@konyang.ac.kr)

Konyang University

Research Article

Keywords: Hemacytometer, multi-volume hemacytometer, fixed volume, Fixed-Volume Cell Counting Model (FVCC model), Multi-Volume Cell Counting Model (MVCC model)

Posted Date: February 9th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-192770/v1>

License: This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published at Scientific Reports on July 8th, 2021. See the published version at <https://doi.org/10.1038/s41598-021-93477-1>.

Multi-Volume Hemacytometer

¹Ravangnam Thunyaporn, Il Doh², ¹Dong Woo Lee*

¹Department of Biomedical Engineering, Konyang University, Daejeon, 35365, Republic of Korea.

² Korea Research Institute of Standards and Science, Daejeon, Korea (KRISS)

* Corresponding Author

Abstract:

Cell counting has become an essential method for monitoring the viability and proliferation of cells. A hemacytometer is the standard device used to measure cell numbers in most laboratories which are typically automated to increase throughput. The principle of both manual and automated hemacytometers is to calculate cell numbers with a fixed volume within a set measurement range ($10^5 \sim 10^6$ cells/ml). If the cell concentration of the unknown sample is outside the range of the hemacytometer, the sample must be prepared again by increasing or decreasing the cell concentration. We have developed a new hemacytometer that has a multi-volume chamber with 4 different depths containing different volumes (0.1, 0.2, 0.4, 0.8 μ l respectively). A multi-volume hemacytometer can measure cell concentration with a maximum of 10^6 cells/ml to a minimum of 5×10^3 cells/ml. Compared to a typical hemacytometer with a fixed volume of 0.1 μ l, the minimum measurable cell concentration of 5×10^3 cells/ml on the multi-volume hemacytometer is twenty times lower. Additionally, the Multi-Volume Cell Counting model (cell concentration calculation with the slope value of cell number in multi-chambers) showed a wide measurement range ($5 \times 10^3 \sim 1 \times 10^6$ cells/ml) while reducing total cell counting numbers by 62.5% compared to a large volume (0.8 μ l-chamber) hemacytometer.

Keyword: Hemacytometer, multi-volume hemacytometer, fixed volume, Fixed-Volume Cell Counting Model (FVCC model), Multi-Volume Cell Counting Model (MVCC model)

1 Introduction

2 Cell counting is a method used to determine the concentration of cells in a sample volume. For
3 example, in cell culture experiments counting cells is vital to monitor cell viability, proliferation rate,
4 immortalization or transformation, seeding, or preparing cells in subsequent experiments, transfection,
5 or infection [1]. Diagnostically, the concentration of white blood cells and red blood cells can assist
6 those in the medical field to identify specific diseases and can predict severity of disease states such as
7 coronary atherosclerosis, etc. [2-4]

8 The hemacytometer was first invented over a century ago to count blood cells. Presently the
9 hemacytometer remains the gold standard for cell counting and is used by most laboratories to count
10 cells as it is both inexpensive and versatile [5]. A hemacytometer is made from optical glass for use
11 under a microscope and consists of 2 parts: a thick glass slide and cover glass with a small gap to contain
12 cell suspension in a 0.1 μ l small grid area size 1×1 mm [6].

13 Currently, automated hemacytometers have been developed to increase the throughput of
14 counting a large number of samples in a short period of time. Automation also reduces human error in
15 the cell counting process and automatic hemacytometers are commonly used in laboratories for these
16 reasons. In general, hemacytometers are composed of a digital camera and a slide or cartridge containing
17 the cell suspension to be counted. The digital camera captures cells on the slide in the fixed volume
18 and analyzes them using specialized software for cell counters to calculate cell concentration based on
19 the fixed volume [6]. Both manual and automated hemacytometers use a fixed volume to measure cell
20 concentration. Though manual and automated hemacytometers are standard in most laboratories, they
21 cannot measure low cell concentration samples and have a narrow measurement range (10^5 ~ 10^6
22 cells/ml) due to the fixed volume containing cell. Most hemacytometers count cells in 0.1 μ l fixed
23 volume where one cell exists at 10^4 cells/ml sample. The range of hemacytometer is approximately 10^5
24 cells/ml. For example, the Cedex HiRes Analyzer which is an automatic hemacytometer, can measure
25 cell density ranging from 3.13×10^5 to approximately 1.0×10^7 cells/ml [7]. The optimal range of
26 Cellometer cell counters is approximately 1×10^5 to 1×10^6 cells/ml [8]. The limit of the range of cell
27 concentration measurement is due to the fixed volume of the sample which causes the minimum and

maximum measurable cell concentration to decrease or increase. A typical 0.1 μl fixed volume (most used volume in hemacytometers) contains 10 ~ 500 cells at $10^5 \sim 5 \times 10^6$ cells/ml while a 0.2 μl fixed volume contains 10 ~ 500 cells at $5 \times 10^4 \sim 2.5 \times 10^6$ cells/ml. Therefore, if a sample of unknown concentration is out of range of the hemacytometer, the sample must be prepared again by increasing or decreasing the cell concentration.

To solve this problem, the current study proposed the use of a multi-volume hemacytometer. This device counts cells in four chambers at volumes of 0.1, 0.2, 0.4, 0.8 μl as shown in **Figure 1**. A multi-volume cell counter can measure cell concentrations ranging from 10^6 cells/ml to 5×10^3 cells/ml. Compared to a hemacytometer having a fixed volume of 0.1 μl , the minimum measurable cell concentration is 5×10^3 cells/ml which is twenty times lower. Additionally, the Multi-Volume Cell Counting model (cell concentration calculation with the slope value of the cell number in multi-chambers) was also validated to measured cell concentration and increases the measurement speed as a sample of only 1.5 μl can measure the wide range of cell concentrations.

Results

To compare the accuracy and range of cell concentration measurement between the conventional hemacytometer and the proposed multi-volume hemacytometer, we measured five samples at 10^3 , 5×10^3 , 10^4 , 10^5 , and 10^6 cells/ml. In the case of the multi-volume hemacytometer, there are two cell concentration calculation models necessary to determine the cell concentration. In the conventional Fixed-Volume Cell Counting (FVCC), the cell concentrations were calculated by counting cells in each chamber at different volumes. In the Multi-Volume Cell Counting (MVCC) model, the cell concentrations were calculated using the slope (cell number growth per volume) in the graph of cell number by volume as show in **Figure 1.(c)**.

Fixed-Volume Cell Counting (FVCC) model

A hemacytometer has a ruled surface below the cover glass of 0.1 mm which limits the volume of liquid suspension to 0.1 μl per one of nine squares and restricts it to accurately measure the concentration of cells in the range of $2 \times 10^5 \sim 2.5 \times 10^6$ cells/ml [9]. As shown in **Table 1**, in the current experiment the hemacytometer shows a large standard deviation (more than 15%) in the samples lower than 10^5 cells/ml. The proposed device has several chambers with different volumes such as 0.1, 0.2, 0.4, and 0.8 μl to contain the multi-volume of cells in suspension. **Figure 2** shows that the number of cells varied in conjunction with the volume. The number of cells increased with the volume of the chamber. As shown in **Table 1**, chamber 1, 2, 3, 4, and the MVCC model were used to measure cell concentration. In the current experiments, chamber 1 in area 1 mm^2 contained a volume of 0.1 μl equal to the volume of the standard hemacytometer. In the low volume chamber, the number of cells was not great enough to calculate the cell concentration. In contrast, as the volume of the chamber increased, the cells also increased enough to allow the cell concentration to be calculated. Therefore, chamber 4 has a wide measurement range and was able to be measured at the lowest concentration at 5×10^3 cells/ml. However, in the maximum cell concentration of 10^6 cells/ml, it was difficult to count a high number of cells in a large volume (0.8 μl) and repeat the measurement five times. Therefore, the user must select the appropriate chamber volume according to the sample.

Multi-Volume Cell Counting (MVCC) model

Figure 3 shows the cell number measurement by chamber volume and the slope in the Multi-Volume Cell Counting (MVCC) model. The slope indicates the cell number per volume, reflecting the cell concentration. The slope (cell numbers per volume) was calculated by the cell number in 0.1, 0.2, 0.4, and 0.8 μl chambers. Dotted lines indicate a fit for standard error. As the sample cell concentration decreased, the spacing between the dotted lines increased. In the repeated cell concentration measurement of 5×10^3 cells/ml sample, the MVCC model indicates the standard deviation of cell concentration is lower than 15%, while the FVCC model indicates a higher standard deviation. As shown in **Figure 4**, measurement using the conventional hemacytometer shows that when the sample

concentration was low, the relative standard deviation of measuring cell concentration increased significantly. The measurement range of the conventional hemacytometer is $10^6 \sim 10^5$ cells/ml, and using a $0.8 \mu\text{l}$ chamber showed a wide measurement range of $10^6 \sim 5 \times 10^3$ cells/ml. However, in 10^6 cells/ml, the hemacytometer using a $0.8 \mu\text{l}$ chamber had too many cells to count manually (approximately 4000). **Figure 4** shows how the proposed MVCC model reduced the number of cell counts needed to measure cell concentration without reducing the measurement range. The MVCC model measured $5 \times 10^3 \sim 10^6$ cells/ml within a 10% standard deviation and reduced the counting cell number by about 62.5%. The MVCC model reduced the counting burden in large volumes as a total volume of only $1.5 \mu\text{l}$ ($0.1+0.2+0.4+0.8 \mu\text{l}$) was needed to measure the cell concentration at a wide range of concentrations.

As well known, the hemacytometer is a basic device used to count cells in most laboratories. A lot of times the sample of cells must be prepared by increasing or decreasing the cell concentration because the unknown sample of cells is outside the range of the hemacytometer. Thus, our study proposed a novel device to measure cell concentration using a multi-volume hemacytometer. This device is designed to have a multi-chamber that can contain multi-volume such as 0.1, 0.2, 0.4, and $0.8 \mu\text{l}$, expanding the contained volume for a hemacytometer. Compared to a standard hemacytometer with a fixed volume of $0.1 \mu\text{l}$, the minimum measurable cell concentration for the multi-chamber hemacytometer is 5×10^3 cells/ml, which is twenty times lower.

Discussion

As well known, the hemacytometer is a basic device used to count cells in most laboratories. Many times a cell sample needs to be re-prepared by either increasing or decreasing the cell concentration because the unknown cell concentration is outside the range of the hemacytometer. Thus, our study proposed a novel device to measure cell concentration using a multi-volume hemacytometer. This device is designed to have a multi-chamber that can contain multi-volume such as 0.1, 0.2, 0.4, and $0.8 \mu\text{l}$, expanding the contained volume for a hemacytometer. Compared to a standard

hemacytometer with a fixed volume of 0.1 μl , the minimum measurable cell concentration for the multi-chamber hemacytometer is 5×10^3 cells/ml, which is twenty times lower.

This is because the multi-chamber hemacytometer has 4 chambers that can hold different amounts of cell samples from low to high volume (0.1, 0.2, 0.4, and 0.8 μl). Cell concentration in each chamber can be calculated by the average number of cells in the area divided by the fixed volume. The number of cells increases as the chamber volume increases, i.e. at fixed volume 0.1 μl , there are 10 ~ 500 cells at $10^5 \sim 5 \times 10^6$ cells/ml while at fixed volumes 0.2 μl , there are 10 ~ 500 cells at $5 \times 10^4 \sim 2.5 \times 10^6$ cells/ml. This multi-volume chamber expands the range of the cell concentration measurements compared to the standard hemacytometer. When unknown sample was loaded in multi-volume hemacytometer, the cell concentration can be measured by selecting an appropriate chamber with the number of cells (10 to 200 cells/chamber) out of four chambers.

In addition, the multi-volume hemacytometer proposed the Multi-Volume Cell Counting (MVCC) model for cell concentration measurement. Chamber 4 contains a maximum volume of 0.8 μl and has a wide measurement range that able to measure the lowest concentration at 5×10^3 cells/ml. Nonetheless, if cell concentration was high around 1×10^6 cells/ml, cell counting only in 0.8 μl -Chamber was difficult and used a lot of time to determine the accuracy of the cell concentration. So, we proposed and validated the Multi-Volume Cell Counting (MVCC) model that cell concentration calculation with the slope value of cell numbers in multi-chambers. The MVCC showed a wide measurement range of $5 \times 10^3 \sim 1 \times 10^6$ cells/ml while reducing total cell counting numbers by approximately 62.5% compared to a 0.8 μl -chamber hemacytometer.

Materials and Methods

Multi-volume hemacytometer design

The multi-volume hemacytometer is designed to address the problem of loading cell samples in unknown concentrations which can fall outside the range of a standard hemacytometer, requiring the sample to be re-prepared. The proposed design is able to define multiple volumes as it consists of 4 chambers with different depths as follows: chamber 1 depth 0.1 mm, chamber 2 depth 0.2 mm, chamber

3 depth 0.4 mm, and chamber 4 depth 0.8 mm which are connected within the same sheet. Each chamber has an area of $3 \times 3 \text{ mm}^2$, between the chambers there is a slope to avoid the cells sticking to the perimeter of the chamber which causes the cells to spread continuously. It also has an inlet and vent area. The counter chamber is made using the PDMS production process after which the PDMS sheets are bonded on a slide glass. (Figure 1)

Multi-volume hemacytometer fabrication

This cell counter sheet is made using polydimethylsiloxane (PDMS) with s SYLGARD 184 silicone elastomer base and SYLGARD 184 silicone elastomer curing agent. The mold was made of aluminium by computer numerical control (CNC) machining. In the first step silicone elastomer base and silicone curing agent were mixed at a 10:1 ratio and a desiccator was used to burst any bubbles for 15 minutes before the elastomer mix was poured into a cell counter aluminium mold and placed inside a desiccator for 15 minutes. The mold was then baked in the oven at 65°C for 4 hours. The PDMS cell counter sheet was then placed on a glass slide and sealed to the glass slide with an oxygen plasma bonder for 2 minutes under a pressure of 80 W.

Cell preparation

Cell line A-549 was counted in the current experiment. A-549 cells were cultured in Dulbecco's Modified Eagle's medium (DMEM) supplemented with 10% Fetal Bovine Serum (FBS) and 1% Anti-Anti (Antibiotic-Antimycotic (100x)) and incubated at 37°C in 5% CO_2 . For harvest, cells were rinsed with 0.25% Trypsin/0.53 mM EDTA to detach the cell layers. Trypsin (2 ml) was added, and the cells were placed in an incubator for 3 minutes. Cells were then washed with phosphate-buffered saline (PBS) and centrifuged. PBS was removed and 1 ml of DMEM was added. The number of cells in the suspension was determined by manual counting using a standard hemacytometer. Cell samples were diluted in a 9: 1 ratio (DMEM: cells suspension) to almost the target cell concentration range, such as 10^6 , 10^5 , 10^4 , 5×10^3 , 10^3 cells/ml.

Cell concentration measurement

For the multi-volume hemacytometer two cell concentration calculation models are needed. The Fixed-Volume Cell Counting (FVCC) model counted cells in a fixed volume of 0.1 μ l which was repeated five times. The FVCC model is the conventional method used in a hemacytometer. The Multi-Volume Cell Counting (MVCC) model counted cells in multiple different volumes at once by calculating cells over an increasing slope per volume to measuring cell concentration. A 10 μ l sample containing cells were loaded in the hemacytometer to standardize samples. After direct counting using a microscope with 5 squares (4 corners and the center squares), we calculated the mean and standard deviation of cell concentrations. The number of cells is divided by the chamber's volume of 0.1 μ l using the following formula:

$$\text{concentration (cells/ml)} = \frac{\text{Number of cells}}{0.1 \mu\text{l}} \quad (1)$$

In the proposed multi-volume hemacytometer, samples containing cells are loaded into four different volume chambers. The cell images were then collected using a microscope with a 40x objective to capture 5 images of different areas in each chamber. Each image has an area of 1 mm^2 . In the current experiment, we designed the four chambers to have heights of 0.1, 0.2, 0.4, and 0.8 mm. The equation used to calculate cell concentration is below.

$$\text{Chamber 1 (volume: 0.1 } \mu\text{l): } \text{concentration (cells/ml)} = \frac{\text{average of counted cells} \times 1000}{0.1 \mu\text{l}} \quad (2)$$

$$\text{Chamber 2 (volume: 0.2 } \mu\text{l): } \text{concentration (cells/ml)} = \frac{\text{average of counted cells} \times 1000}{0.2 \mu\text{l}} \quad (3)$$

$$\text{Chamber 3 (volume: 0.4 } \mu\text{l): } \text{concentration (cells/ml)} = \frac{\text{average of counted cells} \times 1000}{0.4 \mu\text{l}} \quad (4)$$

$$\text{Chamber 4 (volume: 0.8 } \mu\text{l): } \text{concentration (cells/ml)} = \frac{\text{average of counted cells} \times 1000}{0.8 \mu\text{l}} \quad (5)$$

Cell concentrations were calculated by taking five images in a chamber. According to the sample cell concentrations, chambers among the four were selected to measure the cell concentration. If the concentration was low, a larger volume chamber was needed. If the concentration was high, 0.1

1 μl would be enough to measure the cell concentration. However, there were too many cells to count in
2 a $0.8\ \mu\text{l}$ chamber. Therefore, the volume of the chamber was chosen according to the cell concentration.

3 In the proposed multi-chamber hemacytometer, the Multi-Volume Cell Counting model
4 (MVCC model) was applied to cell concentration. The cell numbers in each chamber from four images
5 were plotted according to chamber volume and the slope of a linear equation between counted cell
6 numbers and volume in multi-chambers was calculated to determine the concentration. The slope
7 indicates the number of cells per volume (ml) in the graph of cell number by volume. In the current
8 experiment, we counted four pictures in each chamber to find the cell concentration. The multi-volume
9 model is derived from calculating the slope of the linear equation $y = mx$; m = slope and setting the
10 intercept (x, y) at 0 as the counted cell number starts at 0 cell/ml. (**Figure 1**) In the FVCC model, the
11 large volume chamber has a wide cell concentration measurement range. (**Table 2**) However, there are
12 too many cells in the high concentration cell sample to be counted. In the proposed MVCC model, the
13 total cell count was being reduced by more than 50% while the cell concentration measurement range
14 remained the same.

References

1. Au - Ongena, K., et al., *Determining Cell Number During Cell Culture using the Scepter Cell Counter*. JoVE, 2010(45): p. e2204.
2. BRIGGS, C., *Quality counts: new parameters in blood cell counting*. International Journal of Laboratory Hematology, 2009. **31**(3): p. 277-297.
3. Ates, A.H., et al., *Total white blood cell count is associated with the presence, severity and extent of coronary atherosclerosis detected by dual-source multislice computed tomographic coronary angiography*. Cardiology journal, 2011. **18**(4): p. 371-377.
4. Lu, H., et al., *High throughput single cell counting in droplet-based microfluidics*. Scientific Reports, 2017. **7**(1): p. 1366.
5. Cadena-Herrera, D., et al., *Validation of three viable-cell counting methods: manual, semi-automated, and automated*. Biotechnology Reports, 2015. **7**: p. 9-16.
6. Absher, M., *Hemocytometer counting*, in *Tissue culture*. 1973, Elsevier. p. 395-397.
7. Huang, L.-C., et al., *Validation of cell density and viability assays using Cedex automated cell counter*. Biologicals, 2010. **38**(3): p. 393-400.
8. Chan, L.L.-Y., et al., *A novel image-based cytometry method for autophagy detection in living cells*. Autophagy, 2012. **8**(9): p. 1371-1382.
9. Vembadi, A., A. Menachery, and M.A. Qasaimeh, *Cell Cytometry: Review and Perspective on Biotechnological Advances*. Frontiers in bioengineering and biotechnology, 2019. **7**: p. 147-147.

Acknowledgment

This paper was supported by the Konyang University Research Fund in 2019.

Author contributions statement

Ravangnam Thunyaporn performed experiment and wrote paper. Il Doh advised experimental problems and performed data analysis. Dong Woo Lee supervised the experiments and wrote discussion.

Additional information

The authors declare no competing interests.

Figure 1 Schematic view of the multi-volume hemacytometer. (a) The counting chamber was attached to a glass slide that had an inlet for pipetting and an outlet for airflow. The cell counting chamber consisted of four chambers of varying depths including 0.1, 0.2, 0.4, 0.8 mm respectively. (b) The PDMS fabrication device. (c) The Multi-Volume Cell Counting model was used to determine cell concentration by calculating the slope of cell measurement in each chamber.

Figure 2 Microscopic cell images in multiple volumes (0.1, 0.2, 0.4, and 0.8 μ l) at 1.13×10^4 cells/ml liquid sample. (under $40\times$ magnification)

Figure 3 Graphs (a), (b), (c), (d), and (e) indicate finding using the Multi-Volume Cell Counting model (MVCC model) based on a linear equation for each concentration. The slopes from the linear equations were calculated as a measurement of cell concentration.

Figure 4 Comparison of the percent of relative standard deviation of cell number in each concentration between a standard hemacytometer and a multi-volume hemacytometer with four chambers and the MVCC model.

Table 1 Comparison of cell concentration measurements between a standard hemacytometer and a multi-volume hemacytometer.

Sample	Concentration Unit	Hemacytometer (volume: 0.5 µl)	Multi-volume hemacytometer				
			FVCC Model*				MVCC model** (volume: 1.5µl)
			Chamber 1 (volume: 0.5 µl)	Chamber 2 (volume: 1 µl)	Chamber 3 (volume: 2 µl)	Chamber 4 (volume: 4 µl)	
		Mean±SD(%RSD)	Mean±SD(%RSD)	Mean±SD(%RSD)	Mean±SD(%RSD)	Mean±SD(%RSD)	Fit values±SD (%RSD)
1	10 ⁶ cells/ml	1.08±0.11(9.7%)	1.14±0.02(1.6%)	0.99±0.07(7.0%)	1.14±0.07(5.7%)	1.02±0.03(3.3%)	1.04±0.03(2.5%)
2		1.18±0.13(10.8%)	1.14±0.11(9.3%)	0.94±0.06(6.7%)	1.08±0.10(9.6%)	1.14±0.05(4.5%)	1.12±0.04(3.6%)
3		1.05±0.09(8.5%)	1.07±0.16(14.9%)	1.09±0.15(17%)	1.08±0.16(14.4%)	1.07±0.01(1.3%)	1.07±0.03(3.1%)
4		1.16±0.12(10.5%)	0.88±0.06(6.5%)	0.94±0.04(4.6%)	1.08±0.15(14.0%)	1.12±0.07(6.5%)	1.10±0.03(3.0%)
5		1.19±0.14(11.3%)	1.12±0.17(15.2%)	0.92±0.06(6.8%)	0.94±0.07(7.2%)	1.05±0.13(13.3%)	1.03±0.11(10.4%)
	AVG	1.13±0.06	1.07±0.11	0.98±0.08	1.07±0.08	1.08±0.06	1.07±0.04
	%RSD	5.61	10.51	6.89	6.86	4.54	3.64
1	10 ⁵ cells/ml	1.24±0.42(34.0%)	1.12±0.34(30.5%)	1.07±0.15(14.3%)	1.18±0.30(25.7%)	1.26±0.08(6.4%)	1.23±0.09(7.7%)
2		1.38±0.35(25.3%)	1.26±0.27(21.4%)	1.14±0.20(17.4%)	0.94±0.12(11.5%)	1.22±0.10(8.3%)	1.17±0.02(1.5%)
3		1.20±0.43(35.8%)	1.16±0.49(42.1%)	1.28±0.47(36.9%)	1.20±0.27(22.3%)	1.21±0.11(9.0%)	1.21±0.11(9.4%)
4		1.22±0.35(28.6%)	1.16±0.43(37.4%)	1.21±0.19(15.3%)	1.28±0.29(22.5%)	1.26±0.12(9.6%)	1.26±0.13(10.2%)
5		1.20±0.53(44.1%)	1.00±0.16(15.8%)	1.28±0.27(21.1%)	1.13±0.23(19.9%)	1.30±0.10(8.0%)	1.26±0.10(8.1%)
	AVG	1.25±0.08	1.14±0.09	1.20±0.09	1.15±0.13	1.25±0.04	1.23±0.04
	%RSD	6.06	8.23	7.63	11.10	6.06	3.17
1	10 ⁴ cells/ml	1.00±0.71(70.7%)	1.00±1.00(100%)	1.10±0.65(59.3%)	1.15±0.42(36.4%)	1.20±0.19(15.8%)	1.22±0.12(9.5%)
2		1.40±1.14(81.4%)	1.40±1.14(81.4%)	1.20±0.57(47.5%)	1.15±0.14(11.9%)	1.18±0.14(12.1%)	1.17±0.10(8.7%)
3		0.60±0.55(91.3%)	1.40±0.89(63.9%)	1.20±0.57(47.5%)	1.15±0.38(33.0%)	1.15±0.27(23.6%)	1.16±0.22(19.0%)
4		1.00±0.71(70.7%)	1.20±0.83(69.7%)	1.20±0.57(47.5%)	1.10±0.29(25.9%)	1.28±0.29(22.4%)	1.24±0.18(14.3%)
5		1.00±0.71(70.7%)	0.80±0.45(55.9%)	1.00±0.50(50.0%)	1.15±0.52(45.1%)	1.18±0.59(22.1%)	1.16±0.23(19.9%)
	AVG	1.00±0.28	1.16±0.26	1.14±0.09	1.14±0.02	1.20±0.05	1.19±0.04
	%RSD	28.28	22.48	7.85	1.96	4.02	3.19
1	5×10 ³ cells/ml	2.00±4.47 (223.6%)	6.00±8.94 (149.1%)	7.00±8.37 (119.5%)	6.00±1.37 (22.8%)	6.25±1.77 (28.3%)	6.24±1.67 (28.8%)
2		6.00±5.48 (91.3%)	4.00±8.94 (223.6%)	9.00±9.62 (106.9%)	6.00±2.85 (47.5%)	5.25±2.85 (54.3%)	5.54±3.06 (55.3%)
3		2.00±4.47 (223.6%)	4.00±8.94 (223.6%)	4.00±4.18 (104.6%)	7.50±2.50 (33.3%)	6.00±1.63 (27.2%)	6.16±1.17 (19.0%)
4		4.00±5.58 (136.9%)	8.00±4.47 (55.9%)	2.00±2.74 (136.9%)	7.00±3.26 (46.6%)	5.50±1.43 (25.9%)	5.65±1.27 (22.4%)
5		8.00±8.37 (104.6%)	2.00±4.47 (223.6%)	6.00±2.24 (37.3%)	5.50±2.09 (38.0%)	5.25±2.05 (39.1%)	5.29±1.90 (36.0%)
	AVG	4.40±2.61	4.80±2.28	5.60±2.70	6.40±0.82	5.65±0.45	5.78±0.41
	%RSD	59.27	47.51	48.25	12.84	8.04	7.06
1	10 ³ cells/ml	2.00±4.47 (223.6%)	0	0	1.50±1.37 (91.3%)	1.25±0.88 (70.7%)	1.22±0.85 (69.9%)
2		0	0	0	1.50±1.37 (91.3%)	1.50±0.56 (37.3%)	1.41±0.33 (23.6%)
3		4.00±5.48 (136.9%)	2.00±4.47 (223.6%)	1.00±2.24 (223.6%)	2.00±1.77 (70.7%)	2.00±0.68 (34.2%)	2.05±0.83 (40.6%)
4		2.00±4.47 (223.6%)	2.00±4.47 (223.6%)	1.00±2.24 (223.6%)	2.50±3.26 (163%)	1.75±1.43 (81.4%)	1.76±0.70 (39.4%)
5		0	0	1.00±2.24 (223.6%)	2.50±3.06 (122%)	1.75±2.43 (81.4%)	1.84±0.89 (48.3%)
	AVG	1.60±1.67	0.80±1.10	0.60±0.55	2.00±0.50	1.65±0.29	1.66±0.33
	%RSD	104.58	136.93	91.28	25.00	17.28	20.10

*FVCC Model: Fixed-Volume Cell Counting Model

**MVCC Model: Multi-Volume Cell Counting Model

Legend: Cell concentration measurement quality control pass (%RSD <15%)

1 **Table 2** Total cell counts from cell concentration measurements.

Measure- ment method Concen- -tration (cells/ml)	Hema- cytometer (volume: 0.1×5μl)	Multi-volume Hemacytometer				
		FVCC Model *				MVCC Model ** (volume: 1.5 μl)
		Chamber 1	Chamber 2	Chamber 3	Chamber 4	
		(volume: 0.1×5μl)	(volume: 0.2×5μl)	(volume: 0.4×5μl)	(volume: 0.8×5μl)	
10 ⁶	500	500	1000	2000	4000	1500
10 ⁵	50	50	100	200	400	150
10 ⁴	5	5	10	20	40	15
5×10 ³	2.5	2.5	5	10	20	7.5
10 ³	0.5	0.5	1	2	4	1.5

2 *FVCC Model: Fixed-Volume Cell Counting Model

3 **MVCC Model: Multi-Volume Cell Counting Model

4 □□□: Cell concentration measurement range

5

Figures

Figure 1

Schematic view of the multi-volume hemacytometer. (a) The counting chamber was attached to a glass slide that had an inlet for pipetting and an outlet for airflow. The cell counting chamber consisted of four chambers of varying depths including 0.1, 0.2, 0.4, 0.8 mm respectively. (b) The PDMS fabrication device. (c) The Multi-Volume Cell Counting model was used to determine cell concentration by calculating the slope of cell measurement in each chamber.

Figure 2

Microscopic cell images in multiple volumes (0.1, 0.2, 0.4, and 0.8 μ l) at 1.13×10^4 cells/ml liquid sample. (under 40 $\times$ magnification)

Figure 3

Graphs (a), (b), (c), (d), and (e) indicate finding using the Multi-Volume Cell Counting model (MVCC model) based on a linear equation for each concentration. The slopes from the linear equations were calculated as a measurement of cell concentration.

Figure 4

Comparison of the percent of relative standard deviation of cell number in each concentration between a standard hemacytometer and a multi-volume hemacytometer with four chambers and the MVCC model.