

Effect of Initial Switch-On Within 24 Hours of Cochlear Implantation Using Slim Modiolar Electrodes: Potential Clinical Implications

Woongsang Sunwoo

Gachon University

Hyoung Won Jeon

Seoul National University Bundang Hospital

Byung Yoon Choi (✉ choiby@snuhb.org)

Seoul National University Bundang Hospital

Research Article

Keywords: modiolar electrodes, cochlear implants, surgery

Posted Date: March 10th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-257630/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License. [Read Full License](#)

Abstract

Reducing electrode impedance is an important factor in improving the functional benefits of cochlear implants (CIs). The immediate effect of early switch-on within 24 hours of surgery on impedance among CI recipients with various types of electrodes has been reported previously; however, the immediate change and the evolution of electrode impedances of slim modiolar electrodes after early switch-on within 24 hours of implantation has not. Therefore, the focus of this study was to compare the effect of early switch-on and conventional switch-on 3–4 weeks post-operation on impedance. Compared with impedance measured intraoperatively, our results demonstrate a significant drop in impedance for early switch-on, which sharply contrasted with elevated impedance values for conventional switch-on 3–4 weeks after implantation. Notably, a comparatively lower impedance than the conventional switch-on protocol was observed for up to 2 months post-operation. Most importantly, a much earlier stabilization of impedance can be achieved with the early switch-on protocol coupled with the slim modiolar electrode array compared to the conventional switch-on protocol, offering the advantage of reducing the number of required mapping sessions in the early stages of rehabilitation.

Introduction

In most cochlear implant (CI) centers currently, electrode impedance measurements are routinely conducted immediately after electrode insertion during CI and at the beginning of every programming session^{1,2}. Software-based impedance telemetry has evolved to assist in the objective assessment of internal implant device functionality. Although the internal CI device reliability is high, previous studies have determined that electrode failures can occur in up to 7–9% of recipients^{3–5}. In addition to detecting electrode abnormalities, measuring impedance allows for an enhanced understanding of the setting of the intracochlear electrodes. Electrode impedance is influenced by the physical properties of the electrode contacts as well as the medium surrounding the electrode, such as macrophages, proteins, fibrous tissues, and perilymph^{6,7}. Thus, intracochlear changes due to insertional trauma or an immunologic response to the electrode cause postoperative increases in impedance^{8,9}.

The use of an electrode with a high impedance value may compromise sound quality, speech perception, and the loudness perception¹⁰. Higher impedances require the generation of higher voltages across the electrode-tissue interface to achieve a satisfactory level of loudness. Owing to the compliance voltage limit of the CI device, the maximum current that can be delivered to the electrode contacts with high impedance values is necessarily reduced, which in turn decreases the dynamic range of the current stimulation. In addition, using the maximum current level may increase the likelihood of channel interaction and reduce battery durability. Therefore, reducing electrode impedance is an important aspect of CI surgery and subsequent mappings to improve functional benefits.

Atraumatic insertion is an important prerequisite for minimizing the increase in electrode impedance levels, since inflammatory responses after mechanical trauma have been considered the primary cause of cochlear fibrosis and ossification^{11,12}. Electrode design is also one of the major factors contributing to the damage to the cochlear structure during electrode insertion. Recently, a new type of slim modiolar electrode (CI532 and CI632) has been developed by Cochlear Ltd. (Sidney, Australia) to protect and preserve the delicate structures of the cochlea, while obtaining a tight modiolar proximity. Due to its pre-curved perimodiolar design without a stylet, injury to the basilar membrane and the lateral wall of the scala tympani can be avoided. Moreover, since its thin profile has 60% less volume than previous perimodiolar electrodes (CI512), a lower hydrodynamic load can be expected at the distal end of the cochlea¹³. Several studies have reported that insertion of the slim modiolar electrode can be less traumatic, as indicated by the higher rate of residual hearing preservation^{14,15}.

Electrical stimulation can also alter electrode impedance¹⁶. In general, the impedance of implanted electrodes decreases temporarily after stimulation and increases during periods of non-stimulation. Therefore, it can be inferred that the timing of device activation after implantation can also affect the change in impedance. Since the efficacy and safety of early switch-on within 24 h of surgery has previously been confirmed¹⁷, many CI centers, including ours, have changed the switch-on protocol over the past few years from the initial fitting at 3–6 weeks post-surgery to the day after surgery. Previous studies investigating the evolution of electrode impedance after early switch-on have shown inconsistent results according to the type of electrode array^{18–20}. To the best of our knowledge, this is the first study to describe changes in the impedance of slim modiolar electrodes over time in CI recipients who were rehabilitated with the conventional and early switch-on protocols. The aim of this study was to determine the changes in electrode impedance over time in a relatively large group of participants who were implanted with the slim modiolar electrode for up to 3–4 months after surgery. We also investigated whether the evolution of electrode impedances systematically varied between the two different switch-on protocols and assessed the effect of early initiation of stimulation on impedance changes.

Methods

Ethical statement

This study was approved by the institutional review board (IRB) of Seoul National University Bundang Hospital (No. B-2008/630-108). The IRB also approved a waiver of informed consent for this retrospective chart review, which involves no more than minimal risk. All methods carried out in this study were in accordance with the approved guidelines, regulations, and the Declaration of Helsinki.

Subjects

84 patients (108 ears) who underwent cochlear implantation by a single surgeon (BY Choi) at Seoul National University Bundang Hospital from July 2018 to July 2020 were recruited for this study. All patients underwent CI exclusively with a slim modiolar electrode (CI532 or CI632) (Cochlear Ltd., Sidney, Australia). The switch-on of the device and the first mapping were performed within 24 h of cochlear implantation for 36 ears (26 subjects), and this group was designated the early switch-on group. For the other 72 ears (58 subjects), device switch-on and the first mapping were performed two to three weeks after

surgery, and this group was classified as the conventional switch-on group. To be matched with the early switch on group (n = 36 ears), 72 ears, twice the number of ears of the early switch-on group, were randomly assigned to the conventional mapping group with age and gender matching between the two groups.

Impedance was measured once during surgery (intraoperative) and was performed at least four times thereafter, once for each of the four consecutive mappings. In both groups, the second mapping was performed 2–3 weeks after the first mapping, the third mapping was performed 3–4 weeks after the second mapping, and the fourth mapping was performed 4–8 weeks after the third mapping. We defined CH1-CH7 as basal electrodes, CH8-CH14 as mid-portion electrodes, and CH15-CH22 as apical electrodes.

Mode of impedance measurement

The impedance measurement modes include the monopolar, bipolar, and common ground (CG) modes. Among them, we measured the impedance level using both the monopolar 2 (MP2) mode, where the reference is the plate electrode, and the CG mode. Previously, impedance measurements using the CG mode provided an impedance level that was 13.3% lower than that using the MP2 mode¹⁸, and we intended to analyze the impedance change based on both modes to strengthen our argument.

Periods used for mapping

The early switch-on and conventional switch-on groups differed from each other according to the timing of the device switch-on. In order to effectively compare the impedance level between the groups at each mapping over the 4 months following implantation, the total follow-up period was divided into seven sections, and the impedance values in each of these sections were compared between the two groups. The seven sections were as follows: period 1, intraoperative period; period 2, 0–2 weeks after surgery; period 3, 3–4 weeks after surgery; period 4, 5–7 weeks after surgery; period 5, 8–10 weeks after surgery; period 6, 11–13 weeks after surgery; and period 7, more than 14 weeks after surgery.

Statistical analysis

For the statistical analysis, we compared the average impedance levels between the early switch-on group and the conventional switch-on group using a non-parametric test, the Mann Whitney U-test. Statistical significance was set at $p < 0.05$.

Results

Electrical impedance of slim modiolar electrodes at the initial switch-on within 24 hours of cochlear implantation

The intraoperative electrical impedance (MP2) and the switch-on impedance (MP2) measured within 24 h of CI were compared. The impedance was notably decreased for all 22 contact channels at switch-on ($p < 0.001$) (Fig. 1a & Table 1). This result was in sharp contrast with the statistically significant increase in impedance found when the initial switch-on was 2–3 weeks after CI, which is the conventional method (Fig. 1b & Table 1). The detailed impedance values under the CG mode are described in Supplementary Table S1.

Comparison of longitudinal changes in electrical impedance of slim modiolar electrodes depending on the timing of the initial switch-on

During the first 24 h, the impedance value decreased sharply; however, it gradually increased from the initial switch-on until the third to fourth week after surgery, when it reached a plateau (Fig. 2 & Table 1). This trend was similar between the two modes (MP2 vs. CG). The absolute value of impedance in the CG mode was lower than that in the MP2 mode.

This longitudinal change under the MP2 and CG modes was compared against that of the electrical impedance of the slim modiolar electrodes when the initial switch-on was made 2–3 weeks after the operation (Fig. 3a–b & Table 1). Statistically lower impedance was observed in the early switch-on group compared with the conventional switch-on group that persisted throughout the initial 7 weeks postoperatively. This tendency was identified even 8–10 weeks post-operation; however, the difference was not statistically significant. At 13 weeks or thereafter, there were no differences in the electrical impedance of the slim modiolar electrodes between the two groups.

Importantly, a large difference in the timing at which the impedance stabilized between the two groups was noted (Fig. 3). Specifically, the impedance values stabilized much earlier in the early switch-on group than in the conventional switch-on group. For the early switch-on group, the impedance value reached 8.91 ± 0.15 k Ω (MP2) at period 3 (3–4 weeks after surgery) and stabilized thereafter (Fig. 3). However, the impedance value from the conventional switch-on group did not stabilize until period 6 (13 weeks after surgery) (Fig. 3).

The pattern of longitudinal change in electrical impedance of slim modiolar electrodes in the case of early switch-on (< 24 h) seems to be slightly different among apical, middle, and basal electrodes, despite a lack of statistical difference (Fig. 4 & Table 1), which was also the case for the conventional switch-on group (Fig. 4). Specifically, electrode impedance values measured intraoperatively were highest for basal electrodes and lowest for apical electrodes. However, this trend was reversed in period 2 for both groups, in which the impedance of the apical electrodes was higher than that of the middle and basal electrodes (Fig. 4). The trend reversed again about two months after surgery, with the basal electrodes again showing the highest value (Fig. 4).

Comparison of changes in impedance over cochlear implant device usage time

We also statistically analyzed the changes in impedance according to the timing of CI device usage between the first and second mapping in the early switch-on group. Statistically, the analysis was conducted with simple correction analysis, and the change in impedance over the cochlear implant device usage time

did not produce statistically significant results for any channels.

Discussion

One of the main findings of this study was that impedance at the initial fitting session measured in the 26 subjects whose devices were activated on the day after surgery (early switch-on) were significantly lower than that in the subjects who underwent conventional switch-on between 2 and 3 weeks after implantation of slim modiolar electrodes. The impedance values obtained intraoperatively were similar between the groups. However, the two groups showed interesting differences in the evolution of impedance from surgery to the initial fitting session. The average impedance level (MP2) in the conventional group increased by 34.8% ($p = 0.001$) at the initial switch-on session compared to the intraoperative level, while that of the early switch-on group had decreased by 54.9% ($p < 0.001$). This increase in impedance level for the slim modiolar electrode array during the first 2–3 weeks before the initial switch-on in the conventional group was similar with that reported in previous studies with different types of electrode arrays, including the Nucleus 24 contour electrode array and the Nucleus standard straight array^{21–23}. As reported in a previous study using the Nucleus implant system, a significant reduction in impedance for all electrode contacts from slim modiolar electrodes was also observed from the early switch-on group in the present study^{18,24}. In contrast, other studies using the MED-EL or the Advanced Bionics implant system showed no significant decrease in impedance at early switch-on^{19,20,25}. Saunders et al. and Hu et al. have indicated that different electrode array designs show different trends in longitudinal impedance changes over time^{20,22}. To the best of our knowledge, this is the first study to report the time course of the postoperative impedance levels for a slim modiolar electrode array.

The second important finding of this study was that the significantly lower electrode contact impedance in the early switch-on group compared to the conventional switch-on group lasted nearly two months. Notably, in the early switch-on group, the impedance of all electrode contacts gradually increased during the first month after device activation. This gradual increase observed in our study was in sharp contrast with previous reports in which electrode impedance decreased gradually with the use of electrical stimulation^{7,21,22,26}. This discrepancy is presumed to stem from the degree of tissue that covered the electrode when electrical stimulation was initiated. After implantation, the adsorption of protein onto the electrode surface and cell adherence to a protein layer are elicited as a foreign body immune response²⁷. An *in vitro* study showed that the protein layer and cell growth increased electrode impedance²⁸ and also revealed no significant stimulation-induced impedance changes in clean electrodes without cell coverage²⁹. In addition, in both an *in vivo* study⁷ and a clinical study²⁶, the reduction in impedance caused by stimulation was found to be larger when the impedance levels at 2 weeks after surgery were high. The effects of an inflammatory response on stimulation-induced changes in impedance have also been clinically studied. Tykocinski et al. reported an increase in polarization impedance up to 2 weeks after implantation caused by changes in the electrode-electrode interface, and a reversal of those impedance changes once electrical stimulation was initiated⁶. Previous studies have suggested that electrical stimulation may cause protein desorption and/or cell loss via electroporation over the electrode^{7,28}, while it may have no effect on fibrous tissue growth around the electrode^{6,9}. Although impedance increased after early activation, the value measured at one month was significantly lower ($p < 0.001$) than the value measured during surgery. In addition, impedance at 1 month was significantly lower ($p < 0.001$) in the early switch-on group than in the conventional switch-on group. It is likely that the extent of the protein layer and cell growth in the early switch-on group continued to be lower than that in the conventional switch-on group for nearly 2 months. Thus, these findings indicate that early activation of the device partially prevented the increase in impedance associated with electrochemical reactions on the electrode surface. Empirically, providing low impedances in the early switch-on group at an early post-implant period was particularly beneficial for CI programming in a subgroup of patients who had abnormal findings on preoperative magnetic resonance imaging such as cochlear nerve deficiency (CND) or in elderly patients with a significantly smaller spiral ganglion neuron population. Indeed, in a bid to achieve auditory perception, CI recipients with CND required higher electrical charge per unit phase than those with normal anatomy³⁰. In many cases, lowering impedance with the early switch-on protocol in such patients made it possible to reduce the number of out-of-compliance electrodes and allow for increased electric dynamic ranges from the early programming sessions.

Interestingly, the current study showed that impedance values stabilized much earlier in the early switch-on group than in the conventional switch-on group. For the patient group that underwent early switch-on, the measured values remained as low as 8.91 ± 0.15 k Ω (MP2) in period 3 (3–4 weeks after surgery) and stabilized thereafter. Conversely, in the conventional switch-on group, impedances decreased significantly from 14.44 ± 0.2 k Ω (MP2) in period 2 to 8.3 ± 0.13 k Ω (MP2) in period 6, and then remained relatively stable. Although the differences in impedance between the two groups were not significant for period 5, impedance remained slightly higher in the conventional switch-on group than in the early switch-on group (Figure 2). Among the impedance components, the access resistance is generally determined by the electrolyte resistivity around the electrode and can be increased by the development of a fibrous tissue sheath. Since previous studies have indicated that the access resistance was not affected by electrical stimulation⁷, we speculated that changes in this impedance component over time may not be different between the two groups. Therefore, the lack of a significant difference in the impedance between the two groups in period 5 suggests that the reversible component, such as polarization impedance, in the conventional group was reduced after 2 months of electrical stimulation. As described above, assuming that early activation has the effect of preventing fluctuation of polarization impedance, when stabilization of the impedance had occurred as early as period 3, access resistance also did not increase 3–4 weeks after surgery. One possible factor that could lead to the early stabilization of access resistance is the electrode array type. The current study used a slim modiolar electrode array, which is designed to cause minimal intracochlear trauma during insertion and can be placed in a position close to the modiulus through a high degree of pre-curvedness. It is reasonable to presume that less traumatic insertion and closer perimodiolar proximity may lead to the development of less fibrous tissue around the electrode, especially at the medial aspect of the electrode array, where the electrode contacts are facing and electrical stimulation does occur. We further analyzed whether the amount of time spent wearing the CI device was associated with a difference in impedance evolution within the population in the early switch-on group. It was found that when the initial activation was performed early, the frequency of CI device usage did not significantly affect the impedance evolution. Based on this consistent result of early stabilization of impedances using the slim modiolar electrode with the early switch-on protocol, a stable map was usually achieved 3–4 weeks after activation of the device. As a result, early switch-on of the CI device not only provided a nonstop rehabilitation program, but also decreased the number of hospital visits for CI fitting.

Previous studies have identified that the time course of impedance values differ according to the type of electrode array, and even within the electrode array the trend of impedance changes varies across the electrode array^{18,20-22}. However, longitudinal changes in impedance over the long-term for slim modiolar electrode arrays have not been previously investigated. In the present study, electrode impedance values measured intraoperatively were highest for basal electrodes and lowest for apical electrodes. This could be due to the presence of air bubbles around the basal electrode contacts, which is related to the characteristic insertion technique of the slim modiolar electrode array³¹. Since this electrode array is loaded into the external polymer sheath before insertion, air trapped inside this sheath is commonly inserted into the cochlea with the electrode. Although impedance did not differ significantly across the array, this trend was reversed in both groups in period 2 (Fig. 4). On average, the impedance of the apical electrodes was higher than that of the middle and basal electrodes. This may be related to the differences in the geometric surface area of the electrode contacts across the array, which gradually decreases in an apical direction (CI532, 0.19 mm²–0.16 mm²; CI632, 0.16 mm²–0.15 mm²). It has been reported that impedance is inversely related to the geometric surface area of the electrodes⁶. Figure 4 shows that this trend is reversed again about two months after surgery, with the basal electrodes having the highest value. This may be due to the combination of the two factors. First, it has been reported that there was more extensive fibrous tissue growth in the basal turn of the cochlea compared to the apical turn after implantation^{9,26}. Second, the distance of the electrode in contact with the modiolar wall of the scala tympani was different. In this study, we used the round window insertion technique, in which the electrode was placed in a more medial position within the proximal basal turn by the crista fenestra at the inferior aspect of the round window compared to the electrode inserted via cochleostomy³². Although a favorable perimodiolar position of the electrode can be achieved by the slim modiolar electrode via the round window, a previous histological study has demonstrated that the distance of the electrode contact to the modiolar wall in the basal turn was more variable and farther apart than that in the apical and middle turns³³. Thus, the impedance of the basal electrodes may be more vulnerable to the development of fibrous tissue because of its lateral placement.

As the safety and feasibility of the early activation of the CI device have been previously demonstrated, many centers have been encouraged to change their strategy to early switch-on. However, other than the financial benefits of reducing the waiting period and the number of hospital visits³⁴, the functional benefits of early switch-on have not yet been clearly reported. In this study, we found that at the initial switch-on, significantly lower impedance was associated with early switch-on of the slim modiolar electrode than the conventional switch-on, and the effect lasted nearly 2 months. We also observed that impedance stabilization was achieved as early as one month after surgery with an early switch-on strategy, which was 2 months earlier than that with the conventional switch-on strategy. Since sound quality and power consumption can be affected by impedance, stable impedances are essential for efficient and safe electrical stimulation. However, we only investigated the impedance values in this study, which cannot lead to conclusions about the effects of the early switch-on strategy on clinical outcomes. Thus, further research with psychophysical data, including thresholds, comfort levels, and dynamic ranges, will be needed to validate the functional benefits of early impedance stabilization.

The present study demonstrates that when applying the early switch-on strategy as early as 1 d after surgery for the slim modiolar electrode array, a much earlier stabilization of impedance can be achieved compared to the conventional switch-on strategy performed 3–4 weeks postoperatively, potentially reducing the number of required mapping sessions in the early stages of rehabilitation. Additionally, the early switch-on strategy leads to comparatively lower impedance than the conventional switch-on strategy for at least 2 months postoperatively, offering a significant advantage in early mapping.

Declarations

Acknowledgements

We give special thanks to Amy Lee for her help in retrieval of impedance data of the patients and also to Hyeong Yun Choi from department of Information Science Major specializing in Data Science, University of Maryland, College Park, USA for his devotion to impedance data analysis.

This research was supported by the SNUBH Research Fund (Nos. 00-2020-13-2018-015, 00-2020-13-2016-014, 00-2020-13-2015-019, 02-2017-060 and, 18-2018-009 to B.Y.C).

Author contributions

The final manuscript has been seen and approved by all the authors, and they have given necessary attention to the manuscript to ensure the integrity of the work. H.W.J. and B.Y.C designed the study, W.S. and H.W.J. analyzed data and wrote the paper. H.W.J performed the statistical analysis; H.W.J and B.Y.C collected medical charts. W.S. and B.Y.C provided critical revision. All authors reviewed the manuscript.

Competing interests: The authors declare no competing interests.

References

1. Cosetti, M. K. *et al.* An evidence-based algorithm for intraoperative monitoring during cochlear implantation. *Otol Neurotol.* **33**, 169–176 <https://doi.org/10.1097/MAO.0b013e3182423175> (2012).
2. Wolfe, J. Cochlear implants: audiologic management and considerations for implantable hearing devices (Plural Publishing, 2018).
3. Carlson, M. L. *et al.* Prevalence and timing of individual cochlear implant electrode failures. *Otol Neurotol.* **31**, 893–898 <https://doi.org/10.1097/MAO.0b013e3181d2d697> (2010).
4. Schow, B., Friedland, D. R., Jensen, J., Burg, L. & Runge, C. L. Electrode failure and device failure in adult cochlear implantation. *Cochlear implants international.* **13**, 35–40 (2012).
5. Harris, J. M. *et al.* Incidence, Time Course, and Implications of Electrode Abnormalities in Pediatric Cochlear Implant Recipients. *Ear Hear.* <https://doi.org/10.1097/aud.0000000000000924> (2020).

6. Tykocinski, M., Cohen, L. T. & Cowan, R. S. Measurement and analysis of access resistance and polarization impedance in cochlear implant recipients. *Otol Neurotol.* **26**, 948–956 <https://doi.org/10.1097/01.mao.0000185056.99888.f3> (2005).
7. Newbold, C. *et al.* Impedance changes in chronically implanted and stimulated cochlear implant electrodes. *Cochlear Implants Int.* **15**, 191–199 <https://doi.org/10.1179/1754762813y.0000000050> (2014).
8. Kamakura, T. & Nadol, J. B. Jr. Correlation between word recognition score and intracochlear new bone and fibrous tissue after cochlear implantation in the human. *Hear Res.* **339**, 132–141 <https://doi.org/10.1016/j.heares.2016.06.015> (2016).
9. Wilk, M. *et al.* Impedance Changes and Fibrous Tissue Growth after Cochlear Implantation Are Correlated and Can Be Reduced Using a Dexamethasone Eluting Electrode. *PLoS One.* **11**, e0147552 <https://doi.org/10.1371/journal.pone.0147552> (2016).
10. Saoji, A. A., Adkins, W. J., Olund, A. P., Nelson-Bakkum, E. R. & Koka, K. Effect of exceeding compliance voltage on speech perception in cochlear implants. *Hear Res.* **400**, 108112 <https://doi.org/10.1016/j.heares.2020.108112> (2021).
11. Seyyedi, M. & Nadol, J. B. Jr. Intracochlear inflammatory response to cochlear implant electrodes in humans. *Otol Neurotol.* **35**, 1545–1551 <https://doi.org/10.1097/mao.0000000000000540> (2014).
12. Ishai, R., Herrmann, B. S., Nadol, J. B. Jr. & Quesnel, A. M. The pattern and degree of capsular fibrous sheaths surrounding cochlear electrode arrays. *Hear Res.* **348**, 44–53 <https://doi.org/10.1016/j.heares.2017.02.012> (2017).
13. Gomez Serrano, M., Patel, S., Harris, R. & Selvadurai, D. Initial surgical and clinical experience with the Nucleus CI532 slim modiolar electrode in the UK. *Cochlear Implants Int.* **20**, 207–216 <https://doi.org/10.1080/14670100.2019.1597461> (2019).
14. Nagy, R. *et al.* Long-term hearing preservation with slim perimodiolar CI532® cochlear implant array. *AMERICAN JOURNAL OF OTOLARYNGOLOGY* **1**, Azonosító: 1019-Terjedelem: 1016 p(2018).
15. Haber, K. *et al.* The influence of Slim Modiolar electrode on residual hearing in pediatric patients. *Eur Arch Otorhinolaryngol.* <https://doi.org/10.1007/s00405-020-06342-8> (2020).
16. Newbold, C. *et al.* An in vitro model for investigating impedance changes with cell growth and electrical stimulation: implications for cochlear implants. *J Neural Eng.* **1**, 218–227 <https://doi.org/10.1088/1741-2560/1/4/005> (2004).
17. Chen, J. K., Chuang, A. Y., Sprinzl, G. M., Tung, T. H. & Li, L. P. Safety and feasibility of initial frequency mapping within 24 hours after cochlear implantation. *Acta Otolaryngol.* **135**, 592–597 <https://doi.org/10.3109/00016489.2015.1006335> (2015).
18. Hu, H. C. *et al.* Evolution of impedance field telemetry after one day of activation in cochlear implant recipients. *PLoS One.* **12**, e0173367 <https://doi.org/10.1371/journal.pone.0173367> (2017).
19. Lin, D. P., Chen, J. K., Tung, T. H. & Li, L. P. Differences in the impedance of cochlear implant devices within 24 hours of their implantation. *PLoS One.* **14**, e0222711 <https://doi.org/10.1371/journal.pone.0222711> (2019).
20. Hu, H. C. *et al.* Evolution of impedance values in two different electrode array designs following activation of cochlear implants 1 day after surgery: A study of 58 patients. *Clin. Otolaryngol.* **45**, 584–590 <https://doi.org/10.1111/coa.13327> (2020).
21. Busby, P. A., Plant, K. L. & Whitford, L. A. Electrode impedance in adults and children using the Nucleus 24 cochlear implant system. *Cochlear Implants Int.* **3**, 87–103 <https://doi.org/10.1179/cim.2002.3.2.87> (2002).
22. Saunders, E. *et al.* Threshold, comfortable level and impedance changes as a function of electrode-modiolar distance. *Ear Hear.* **23**, 28–40 <https://doi.org/10.1097/00003446-200202001-00004> (2002).
23. Paasche, G., Bockel, F., Tasche, C., Lesinski-Schiedat, A. & Lenarz, T. Changes of postoperative impedances in cochlear implant patients: the short-term effects of modified electrode surfaces and intracochlear corticosteroids. *Otol Neurotol.* **27**, 639–647 <https://doi.org/10.1097/01.mao.0000227662.88840.61> (2006).
24. Chen, J. K., Chuang, A. Y., Sprinzl, G. M., Tung, T. H. & Li, L. P. Impedance and electrically evoked compound action potential (ECAP) drop within 24 hours after cochlear implantation. *PLoS One.* **8**, e71929 <https://doi.org/10.1371/journal.pone.0071929> (2013).
25. Alhabib, S. F., Abdelsamad, Y., Yousef, M., Alzhirani, F. & Hagr, A. Effect of early activation of cochlear implant on electrode impedance in pediatric population. *International Journal of Pediatric Otorhinolaryngology.* **140**, 110543 <https://doi.org/10.1016/j.ijporl.2020.110543> (2021).
26. Paasche, G., Tasche, C., Stöver, T., Lesinski-Schiedat, A. & Lenarz, T. The long-term effects of modified electrode surfaces and intracochlear corticosteroids on postoperative impedances in cochlear implant patients. *Otol Neurotol.* **30**, 592–598 <https://doi.org/10.1097/MAO.0b013e3181ab8fba> (2009).
27. Tang, L. & Eaton, J. W. Natural responses to unnatural materials: A molecular mechanism for foreign body reactions. *Mol Med.* **5**, 351–358 (1999).
28. Newbold, C. *et al.* Changes in biphasic electrode impedance with protein adsorption and cell growth. *J Neural Eng.* **7**, 056011 <https://doi.org/10.1088/1741-2560/7/5/056011> (2010).
29. Newbold, C. *et al.* Electrical stimulation causes rapid changes in electrode impedance of cell-covered electrodes. *J Neural Eng.* **8**, 036029 <https://doi.org/10.1088/1741-2560/8/3/036029> (2011).
30. Buchman, C. A. *et al.* Cochlear implantation in children with labyrinthine anomalies and cochlear nerve deficiency: implications for auditory brainstem implantation. *Laryngoscope.* **121**, 1979–1988 <https://doi.org/10.1002/lary.22032> (2011).
31. Hey, M. *et al.* Objective, audiological and quality of life measures with the CI532 slim modiolar electrode. *Cochlear Implants Int.* **20**, 80–90 <https://doi.org/10.1080/14670100.2018.1544684> (2019).
32. Briggs, R. J. *et al.* Comparison of round window and cochleostomy approaches with a prototype hearing preservation electrode. *Audiol Neurootol.* **11** Suppl (1), 42–48 <https://doi.org/10.1159/000095613> (2006).
33. Iso-Mustajärvi, M. *et al.* A New Slim Modiolar Electrode Array for Cochlear Implantation: A Radiological and Histological Study. *Otol Neurotol.* **38**, e327–e334 <https://doi.org/10.1097/mao.0000000000001542> (2017).

34. Hajr, E. A. & Almuhawes, F. Financial Benefits of the Early Fitting of a Cochlear Implant Speech Processor: Assessment of the Direct Cost. *Cureus*. **11**, e5684 <https://doi.org/10.7759/cureus.5684> (2019).

Tables

Table 1
Median impedance values of all channels of slim modiolar electrodes under MP2 mode.

| | Period 1 | | Period 2 | | Period 3 | | Period 4 | | Period 5 | | Period 6 |
|-------------------------------|---------------|---------------|---------------|-------------|---------------|-------------|--------------|-------------|--------------|-------------|--------------|
| | Conventional | Early | Conventional | Early | Conventional | Early | Conventional | Early | Conventional | Early | Conventional |
| Basal Electrodes | | | | | | | | | | | |
| CH 01 | 10.63 | 11.88 | 12.78 | 4.54 | 12.80 | 7.75 | 10.23 | 8.21 | 9.90 | 8.12 | 9.57 |
| CH 02 | 11.36 | 12.15 | 13.54 | 4.79 | 13.12 | 7.85 | 9.97 | 8.15 | 9.84 | 8.16 | 9.19 |
| CH 03 | 11.83 | 12.18 | 14.23 | 4.87 | 13.46 | 8.44 | 10.17 | 8.44 | 9.95 | 8.72 | 9.15 |
| CH 04 | 11.65 | 12.49 | 14.17 | 5.16 | 13.62 | 8.83 | 10.14 | 8.84 | 9.74 | 8.85 | 8.85 |
| CH 05 | 11.58 | 12.61 | 14.78 | 5.23 | 13.42 | 8.65 | 10.05 | 8.94 | 9.59 | 8.66 | 8.61 |
| CH 06 | 11.94 | 12.75 | 14.56 | 5.20 | 13.42 | 8.75 | 10.10 | 8.78 | 9.67 | 8.67 | 8.48 |
| CH 07 | 11.70 | 12.24 | 15.04 | 5.23 | 13.15 | 8.78 | 10.08 | 8.83 | 9.65 | 8.66 | 8.36 |
| Mean | 11.53 | 12.33 | 14.16 | 5.00 | 13.28 | 8.44 | 10.11 | 8.60 | 9.76 | 8.55 | 8.89 |
| (95% CI) | (11.31–11.74) | (12.02–12.63) | (13.75–14.56) | (4.89–5.12) | (13.01–13.56) | (8.14–8.73) | (9.86–10.35) | (8.32–8.88) | (9.53–9.99) | (8.19–8.91) | (8.66–9.12) |
| Mid-portion Electrodes | | | | | | | | | | | |
| CH 08 | 10.91 | 11.72 | 13.67 | 5.15 | 12.99 | 8.66 | 9.99 | 8.98 | 9.41 | 8.75 | 8.18 |
| CH 09 | 10.75 | 11.53 | 13.60 | 5.01 | 13.06 | 8.69 | 9.78 | 8.91 | 9.26 | 8.59 | 8.03 |
| CH 10 | 10.56 | 11.27 | 13.97 | 5.03 | 12.99 | 8.82 | 9.70 | 8.88 | 9.28 | 8.60 | 8.00 |
| CH 11 | 10.66 | 11.10 | 14.22 | 5.01 | 13.13 | 8.78 | 9.65 | 8.52 | 9.10 | 8.25 | 7.76 |
| CH 12 | 10.62 | 10.84 | 14.53 | 5.17 | 13.01 | 8.76 | 9.62 | 8.73 | 8.95 | 8.18 | 7.65 |
| CH 13 | 10.93 | 11.17 | 15.08 | 5.29 | 13.26 | 8.75 | 9.71 | 8.75 | 9.05 | 8.30 | 7.69 |
| CH 14 | 10.91 | 11.65 | 15.33 | 5.44 | 13.47 | 9.18 | 9.89 | 8.81 | 9.21 | 8.67 | 7.81 |
| Mean | 10.76 | 11.33 | 14.34 | 5.16 | 13.13 | 8.81 | 9.76 | 8.80 | 9.18 | 8.48 | 7.88 |
| (95% CI) | (10.58–10.95) | (11.05–11.6) | (14.08–14.6) | (5.05–5.26) | (12.88–13.38) | (8.57–9.04) | (9.53–9.99) | (8.56–9.03) | (8.95–9.41) | (8.16–8.79) | (7.67–8.08) |
| Apical Electrodes | | | | | | | | | | | |
| CH 15 | 10.84 | 11.66 | 15.24 | 5.45 | 13.60 | 9.39 | 10.12 | 9.07 | 9.40 | 8.79 | 8.04 |
| CH 16 | 10.39 | 11.34 | 14.25 | 5.36 | 13.62 | 9.48 | 9.99 | 9.24 | 9.31 | 8.69 | 8.10 |
| CH 17 | 10.38 | 11.16 | 14.66 | 5.23 | 13.61 | 9.38 | 10.07 | 9.11 | 9.26 | 8.57 | 8.08 |
| CH 18 | 10.46 | 11.48 | 14.70 | 5.40 | 13.67 | 9.57 | 10.26 | 9.30 | 9.35 | 8.65 | 8.21 |
| CH 19 | 9.82 | 10.79 | 14.67 | 5.32 | 13.49 | 9.38 | 9.84 | 9.48 | 9.21 | 8.68 | 8.12 |
| CH 20 | 9.52 | 10.47 | 14.84 | 5.32 | 13.36 | 9.31 | 9.96 | 9.29 | 9.24 | 8.27 | 8.00 |
| CH 21 | 8.57 | 9.92 | 14.74 | 5.40 | 13.58 | 9.46 | 10.06 | 9.48 | 9.45 | 8.24 | 8.47 |
| CH 22 | 9.65 | 10.19 | 15.01 | 5.25 | 13.75 | 9.42 | 10.21 | 8.70 | 9.34 | 8.08 | 8.34 |

| | Period 1 | | Period 2 | | Period 3 | | Period 4 | | Period 5 | | Period 6 |
|-------------------------|---------------|---------------|---------------|-------------|---------------|-------------|--------------|-------------|--------------|-------------|--------------|
| | Conventional | Early | Conventional | Early | Conventional | Early | Conventional | Early | Conventional | Early | Conventional |
| Mean | 9.95 | 10.87 | 14.76 | 5.34 | 13.59 | 9.42 | 10.06 | 9.21 | 9.32 | 8.50 | 8.17 |
| (95% CI) | (9.75–10.15) | (10.59–11.16) | (14.41–15.12) | (5.24–5.44) | (13.38–13.80) | (9.20–9.64) | (9.87–10.26) | (9.01–9.41) | (9.12–9.52) | (8.20–8.79) | (7.98–8.36) |
| Total Electrodes | | | | | | | | | | | |
| Mean | 10.71 | 11.48 | 14.44 | 5.18 | 13.35 | 8.91 | 9.98 | 8.88 | 9.42 | 8.51 | 8.3 |
| (95% CI) | (10.59–10.83) | (11.31–11.65) | (14.23–14.64) | (5.11–5.24) | (13.2–13.49) | (8.77–9.06) | (9.85–10.11) | (8.74–9.02) | (9.29–9.54) | (8.32–8.69) | (8.18–8.43) |

Figures




Figure 1
 Impedance values under monopolar 2 (MP2) modes measured at the initial switch-on in the early switch-on group (a, within 24 hours) and in the conventional switch-on group (b, at 2–3 weeks postoperatively). Error bars indicate ± 2 standard errors, which corresponds to a 95 % confidence interval.




Figure 2
 Longitudinal impedance changes of a slim modiolar electrode under common ground (CG) mode and monopolar 2 (MP2) mode in the early switch-on group. Error bars represent ± 2 standard errors, which corresponds to a 95 % confidence interval. IO: intraoperative period.




Figure 3
 Comparison of longitudinal impedance changes between early switch-on (< 24 hours) and conventional switch-on (2–3 weeks) under monopolar 2 (MP2) mode (a) and common ground (CG) mode (b). Error bars indicate ± 2 standard errors, which corresponds to a 95 % confidence interval. Asterisks indicate statistically significant difference between two groups ($p < 0.05$, Mann–Whitney U test). IO: intraoperative period.




Figure 4

Comparison of longitudinal impedance changes according to electrode segment between early switch-on and conventional switch-on under monopolar 2 (MP2) mode (a and b) and common ground (CG) mode (c and d). Different shapes of each data point indicate average impedances of different electrode segments. Δ : CH1-CH7 as basal electrodes, \square : CH8-CH14 as mid-portion electrodes, \bullet : CH15-CH22 as apical electrodes. IO: intraoperative period.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [SupplementaryTableS1.docx](#)