

The Development and Psychometric Validation of A Low-Cost Anthropomorphic 3D-Printout Simulator for Training Basic Skills Applicable to Office-Based Hysteroscopy

Ana Rita Peixoto Panazzolo (✉ anaritapanaz@hotmail.com)

Universidade Federal de Santa Catarina <https://orcid.org/0000-0002-9502-6240>

Leisa Beatriz Grando

Universidade Federal de Santa Catarina

Lia Karina Volpato

Universidade Federal de Santa Catarina

Getúlio R. de Oliveira Filho

Universidade Federal de Santa Catarina

Original Article

Keywords: hysteroscopy, office-based hysteroscopy, printing, 3-dimensional, simulation training, validation studies

Posted Date: May 14th, 2020

DOI: <https://doi.org/10.21203/rs.3.rs-26870/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Abstract

Background Hysteroscopy training requires the development of specific psychomotor skills. Few validated low-cost models exist in hysteroscopy. The main objective of this study is to determine face, content and construct validity of a simulator designed for training basic hysteroscopy skills applied to office-based hysteroscopy.

Methods Twenty-five hysteroscopy experts and 30 gynecology residents participated in this prospective observational study. The simulator consisted of three color-textured, silicone-coated anthropomorphic 3-dimensional (3-D) printout uterine models inside a box. Each uterine model in the simulator was designed to develop one of the following basic hysteroscopic skills: hysteroscopic navigation, direct biopsy, and foreign body removal. Participants performed five video-recorded simulation attempts on each model. Procedure-specific checklists were used to rate performance.

Results Median scores (25th – 75th percentiles; p-value) 4 (3–4; $p < 0.001$) and surgical experience associated to the simulated procedures 4 (3–4; $p < 0.001$) indicated positive perceptions as to the realism of the internal cavity of the simulator. Median scores of 4 (3–4; $p < 0.001$) were assigned to the realism and utility of the tasks performed in the simulator for enhancing novice training in hysteroscopy. Expert performance scores were significantly higher and task completion times were significantly lower than those of novices in the navigation exercise ($F_{(1,53)} = 56.66$; $p < 0.001$), the directed biopsy exercise ($F_{(1,53)} = 22.45$; $p < 0.001$), and the foreign body removal exercise ($F_{(1,53)} = 58.51$; $p < 0.001$). Novices' performance improved on all three exercises: navigation exercise ($F_{(1,53)} = 182.44$; $p < 0.001$); directed endometrial biopsy ($F_{(1,53)} = 110.53$; $p < 0.001$); and foreign body removal ($F_{(1,53)} = 58.62$; $p < 0.001$). Experts' task completion times were significantly lower when compared to that of novices in the five attempts ($p < 0.001$) of the exercises: navigation ($F_{(1,48)} = 25.46$; $p < 0.001$); directed biopsy ($F_{(1,46)} = 31.20$; $p < 0.001$); and foreign body removal ($F_{(1,50)} = 69.8$; $p < 0.001$). Novices' task completion times diminished significantly throughout the sequence of exercises.

Conclusions The low-cost simulator designed for the acquisition of basic skills in hysteroscopy demonstrated face, content and construct validity.

Background

Hysteroscopy is a minimally invasive procedure considered as the gold standard for the evaluation and treatment of intrauterine diseases [1]. Currently available, small-diameter scopes have permitted the performance of in-office hysteroscopic procedures, allowing non-traumatic insertion into the cervix, a vaginoscopy technique [2]. Therefore, hysteroscopic proficiency requires specific training for the development of new psychomotor skills. The importance of a simulator as a training tool has been extensively emphasized [3].

Simulators allow training in a safe environment, real-time expert feedback and repeated practice of each step of the procedure [4–6]. Recommendations have been made for institutions to provide dry labs for training endoscopic surgical skills [7–8].

The successful integration of a new simulation system into a training curriculum demands rigorous evaluation of its validity [9]. The first step is to demonstrate that new simulator resembles real-life situations, that is, face validity. The extent to which new simulator allows acquisition of abilities required for proficiently performing in real patients, content validity, must also be assured. Thereon, validity procedures aim to demonstrate that new simulator allows detection of distinct levels of operator proficiency, for example, performance of novices and experts [10]. It is also desirable to demonstrate that simulator allows detection of changes in performance, learning curve, during repeated attempts [11].

We hypothesized that a novel, inorganic, anthropomorphic simulator incorporates 3-D printing would demonstrate high levels of face, content and construct validity. The present study aimed to test these hypotheses by applying specific procedures for face, content, and construct validation.

Materials And Methods

Data were collected at two general hospitals. The IRB (Institutional review board) approved the study at the participant institutions. Written informed consent was obtained from participants before inclusion in the study. From October 2018 to April 2019, a convenience sample consisting of 30 gynecology residents with no previous experience in hysteroscopy and 25 experts participated in the study. Experts were gynecologic surgeons who had received formal training in hysteroscopic procedures and had been practicing diagnostic and operative hysteroscopy for more than three years at the time they were enrolled in the study.

The Simulator

The anthropomorphic anatomical models used in the simulator were manufactured in acrylonitrile butadiene styrene (ABS) on a fusion deposition modeling (FDM) 3-D printer (3-D Printer Prusa I3 Rework, ITCTERM, Brazil). The 3-D model was scanned from a 9 × 7 × 6 cm pyriform mold simulating a human uterus. The models contained two hemicavities internally lined with textured, rose-colored acetic silicone to mimic the appearance of the human endometrium. A 2-cm diameter opening was created to allow the atraumatic passage of the hysteroscope.

Three uterine models were embedded in latex foam inside an opaque 30 cm-wide, 28 cm-long, 18-cm high cardboard box. Three 3-cm circular apertures were made on the longest box wall. The three apertures were framed with ABS ring-shaped connectors to which red-colored male condoms were attached and connected to the opening of the uterine models inside the box simulating vaginas (Fig. 1). Each uterine model in the simulator was designed to develop one of the basic hysteroscopic skills.

The first model was designed to develop intra-uterine navigation skills. It was equipped with nine distinct adhesive images positioned on the anterior, posterior, left lateral, right lateral walls, on the uterine fundus, and on the anterior and posterior tubal ostia, represented by two red markings inside the model.

The second uterine model was designed to develop direct biopsy skills. Five electronic touch-sensors (Micro Switch, KW-1 3pins, Dongnan, China) covered by a non-slip adhesive tape were connected to a microcontroller board, Arduino Uno (Smarts Projects, Ivrea, Italy) which was connected to a microcomputer via a USB (Universal Serial Bus) port. The sensors were positioned through openings in the model's wall in the uterine fundus and on the anterior, posterior, left and right walls of the inner surface of the model. A visual sign on the computer screen indicated which sensor had been pressed at each attempt.

The third uterine model was designed for developing foreign body removal skills. The model was equipped with five 8-mm diameter apertures on the model wall through which 8-mm diameter, 4-cm long mini balloons were inserted five millimeters into the model cavity.

Face and content validity

Face and content validities were based on expert perceptions of the realism of tasks performed on the simulator compared to procedures performed on real patients. Face validity was evaluated by assessing the degree of expert concordance with the following statements: (1) the appearance of the simulator's internal cavity resembles that of the human uterine body cavity; and (2) the procedures performed on the simulator resemble the office-based hysteroscopy procedures performed on human patients. Responses were measured on four-point unidirectional, forced-response scales, where the extreme points 1 and 4 indicated full disagreement and full agreement, respectively. Also, the realism of the global experience in performing the tasks on the simulator was assessed on a 10-point rating scale (1 = absolutely non-realistic; 10 = absolutely realistic).

Content validity was based on experts' agreement with statements addressing the similarity between the tasks performed on the simulator, the hysteroscopic maneuvers performed in real patients, for example hand-eye coordination, accurate visualization through a 30° angle scope and the use of grasping forceps. Also, the utility of learning experiences in the simulator for training novices was used to assess the content validity of the simulator. Responses were measured on four-point unidirectional scales where the extreme points 1 and 4 indicated full disagreement and full agreement, respectively.

Construct validity

The ability to discriminate between expert and resident performance determined the construct validity of the simulator. Participants' performance was measured on procedure-specific checklists used for rating performance on the video-recorded simulation sessions.

For developing the procedure-specific checklists (Appendix), three authors (A.R.P., L.B.G. and L.K.V.) independently searched the literature to create technical performance checklists addressing the main

steps of the simulation tasks: hysteroscopic navigation, direct biopsy, and foreign body removal [12, 13]. Items in the checklists addressed ergonomics, image visualization, safe navigation and handling the grasping forceps [10]. Items of the checklists represented the elements or steps identified in the three tasks that could be measured and potentially differentiate amongst levels of technical competence. Three-point scales were added to the scoring rubrics, where zero indicated an unskilled performance, as expected from a novice; one indicated a somewhat skilled performance, as expected from an in-training novice; and two indicated a skilled performance, as expected from an experienced surgeon [14].

Simulation sessions

Simulation sessions took place in private rooms equipped with a video endoscopy equipment consisting of a high definition monitor (Karl Storz, Germany), a cold light source XENON 300 (Karl Storz, Germany) and a video-camera IMAGE 1 HUB™ (Karl Storz, Germany). Also available in the simulation room were HOPKINS® Forward-Oblique 30°, 2.9-mm Telescope (Karl Storz, Germany), BETTOCCHI® inner and outer sheaths (Karl Storz, Germany), and semirigid 5 French (Fr) grasping forceps (Karl Storz, Germany), which were used to complete the tasks in the simulator. The simulator box was positioned on a ninety-centimeter high table for the training sessions. Participants stood in front of the simulator during the experiment, having the video equipment positioned anteriorly and to their left, allowing a full view of the screen. The same investigator (A.R.P.) conducted all simulation sessions, which involved only one participant at a time. Formative feedback on participant performance was allowed [11].

Simulation sessions consisted of a preparatory and a hands-on phase. The preparatory phase aimed to provide the participant with relevant theoretical and practical guidance about the tasks to be simulated. The preparatory phase started soon after the arrival of the participant at the simulation room. The instructor started by providing a structured twenty-minute presentation addressing the definition, the indications, the equipment, the technique and the complications of office-based hysteroscopy. Next, participants were informed about the technical aspects of the tasks to be simulated, their metrics and goals. Participants were further instructed about how to handle the scope and the grasping forceps. The preparatory phase of the experiment finished after the instructor presented the simulator, demonstrated the tasks and allowed the participants to manipulate the simulator, the video and the hysteroscopy equipment for approximately five minutes before starting the hands-on phase of the simulation session.

Participants performed three exercises (tasks) during the hands-on phase of the simulation session. The navigation exercise (task 1) consisted of identifying, visualizing and centering the nine targets and simulated tubal ostia inside the first model followed by obtaining a panoramic view of the interior uterine model, before withdrawing the hysteroscope. The aim of the direct biopsy simulation exercise (task 2) was to press the sensors positioned inside model 2 with the tip of a grasping forceps. The foreign body removal simulation exercises (task 3) consisted of using grasping forceps to clamp and pull the mini balloons inserted on the wall of the third model at least 3 cm into the simulator cavity. Participants repeated each exercise five times. Task completion times were counted from the moment of insertion to the moment of withdrawal of the hysteroscope from the simulator. Participant performance was digitally

recorded (Canon EOS 60D 18 MP, Canon U.S.A., Inc., Huntington, NY, U.S.A) for further analyses. Three authors (A.R.P, L.B.G. and L.K.V.) rated the videos independently [15].

At the end of the simulation session, participants answered the demographic data questionnaire. Experts also completed the eight-item face and content validity questionnaire.

Statistical analysis

Statistical analysis was performed utilizing SPSS v 26.0 (SPSS Inc., Chicago, IL, U.S.A). Basic descriptive statistics were calculated for demographic data. All continuous variables were presented as mean (standard deviation) or median (range), and nominal variables were presented as frequency (percentage, %). Kolmogorov-Smirnov test was used to assess normality. Generalized linear mixed model (GLM mixed-model) ANOVA was used to compare scores and task completion times within and between groups. Median scores of the items used to assess the face and content validities were tested against the mathematical center of the scales by using one-sample Wilcoxon signed-rank tests.

Psychometric analyses of the technical performance checklists included exploratory factor analyses by the principal-component extraction method and varimax rotation with Kaiser Normalization to determine the factorial structure of the instruments and the estimation of Cronbach's alpha coefficients to assess the internal consistency of the technical performance checklists.

Inter-rater agreement was evaluated by estimating intra-class correlation coefficients and their 95% confidence intervals among the scores attributed by raters 1, 2, and 3 to the items in the checklists and among the average scores, under the following assumptions: the same raters assessed all videotaped procedures; the selected raters were representatives of a larger sample of raters (experts in ambulatorial hysteroscopy); the reliability of the mean value of multiple raters was the measure of interest, and we searched for consistency of ratings. Based on the assumptions mentioned above, two-way random-effects models were used [16].

The estimated sample size was 44 participants for five attempts with a power of 80% at an α level of 0.05. Calculations were based on the effect size (ϵ^2) equal to 0.057 observed in a previous study [10]. Fifty-five subjects were enrolled in the study to account for eventual losses.

Results

Twenty-five experts and thirty residents completed the study. Demographic data are shown in table 1. A total of 825 videos records of participants' performance were obtained and randomized (www.random.org).

	Experts	Residents
Number of participants ^a	25 (45.45%)	30 (54.54%)
Gender ^a		
Female	18 (72%)	24 (80%)
Male	7 (28%)	6 (20%)
Age (years) ^b		
Female	44 (38 – 52)	28 (27 – 30)
Male	46 (43.5 – 52.75)	29.5 (28 – 30)
Experience in hysteroscopy (years) ^b		
Female	16 (8 – 21)	N/A
Male	14 (9 – 20.5)	N/A
Post-graduate year ^a		
PGY-1	N/A	16 (53.3%)
PGY-2	N/A	7 (23.3%)
PGY-3	N/A	7 (23.3%)
Footnotes: a. Data expressed as frequency (percentage); b. Data expressed as median (25 th – 75 th percentiles); N/A – Not applicable.		

Table 1
Demographic data of the study subjects.

Face and content validity

Median scores (25th – 75th percentiles) assigned by experts to the statements that assessed face validity were significantly higher than the mathematical center of the scale ($p < .001$), indicating positive perceptions about the realism of the internal cavity of the simulator 4 (3 – 4; $p < .001$) and of the surgical experience associated the simulated procedures 4 (3 – 4; $p < .001$). Also, the median score of the global experience on the simulator was 9 (25th – 75th percentiles = 9 – 10), differing significantly from the mathematical center of the scale ($p < .001$). Median scores of 4 (3 – 4; $p < .001$) were assigned to the realism and utility of the tasks performed in the simulator as applicable to the training of novices in hysteroscopy.

Psychometric analyses of the technical performance procedure-specific checklists

A total of 825 video recordings of participants' performance were obtained, from which 27 (3.2%) were excluded due to low quality of images. Three authors rated 798 videos. Ratings were used for psychometric analyses of the technical performance checklists.

The factorial structure of the checklist used to assess participants' performance in task 1 comprised two factors, which explained 55.89% of score variance: factor 1 (items related to image visualization ability) (Eigenvalue = 7.42; explained variance = 49.48%), and factor 2 (ergonomics-related items) (Eigenvalue = 1.56; explained variance = 10.41%). The factorial structure of the scale used to assess participants' performance in task 2 comprised a single factor that explained 51.62% of score variance (Eigenvalue = 3.097). The factorial structure of the scale used to assess participants' performance in task 3 comprised a single factor that explained 52.92% of score variance (Eigenvalue = 3.176).

Cronbach's alpha coefficients of the checklist designed to assess task 1 (navigation) were 0.92 (95% Confidence interval, CI = 0.91 – 0.93), 0.77 (95% CI = 0.75 – 0.80), and 0.90 (95% CI = 0.89 – 0.91) for image visualization ability factors, ergonomics, and for the global scale, respectively. Based on these findings, the global score was chosen to measure participants' performance on the navigation tasks. The Cronbach's alpha coefficient estimated for the checklist designed to assess performance at tasks 2 and 3 were 0.80 (95% CI = 0.78 – 0.83) and 0.82 (95% CI = 0.80 – 0.84), respectively.

Intraclass correlation coefficients among raters' average scores on items addressing image visualization and ergonomics in task 1 checklist were 0.55 (95% CI = 0.45 – 0.64) and 0.76 (95% CI = 0.71 – 0.81). Intraclass correlation coefficients among raters' average scores were 0.75 (95% CI = 0.69 – 0.79) and 0.75 (95% CI = 0.69 – 0.80) for tasks 2 and 3 technical performance checklists, respectively.

Construct validity

Expert technical performance scores were constant during the five attempts and significantly higher than novices for navigation tasks ($F_{(1,53)} = 56.66$; $p < .001$; Figure 2-A); direct endometrial biopsy ($F_{(1,53)} = 22.45$; $p < .001$; figure 3-A); and foreign body removal ($F_{(1,53)} = 58.51$; $p < .001$; Figure 4-A). Residents' performance at the navigation tasks improved from the third to the fifth attempt ($F_{(1,53)} = 182.44$; $p < .001$; figure 2-A). Significant improvement was observed at the fourth attempt at the direct endometrial biopsy ($F_{(1,53)} = 110.53$; $p < .001$; figure 3-A). Residents' performance improved significantly from the second through the fifth attempts at the foreign body removal exercise ($F_{(1,53)} = 58.62$; $p < .001$; figure 4-A).

Experts' times to complete the three tasks were significantly lower than those of residents at all attempts at navigation ($F_{(1,48)} = 25.46$; $p < .001$); direct biopsy ($F_{(1,46)} = 31.20$; $p < .001$); and foreign body removal task ($F_{(1,50)} = 69.8$; $p < .001$) as shown in figures 4-B. Experts' and residents' times to perform the

navigation task decreased significantly throughout the five attempts ($F_{(1,48)} = 30.98$; $p < .001$; figure 2-B). Residents' times decreased significantly from the second through the fifth attempts during the sequence of simulated endometrial biopsy ($F_{(1,46)} = 4.35$; $p < .001$; figure 3-B). At the foreign body removal task, residents' times decreased significantly from the second through the fifth attempts ($F_{(1,50)} = 5.85$; $p < .019$; figure 4-B), while experts' times decreased significantly from the third to the fifth attempt ($F_{(1,46)} = 4.35$; $p < .001$) and ($F_{(1,50)} = 5.85$; $p < .019$; figure 4-B).

Discussion

The main results of this study were that the new simulator exhibited high face, content, and construct validity. Also, highly reliable instruments were used to produce the measures of technical performance.

Face validity of anthropomorphic simulators requires close resemblance between what the operator sees in the simulator and in real patients (realism). For this reason, the face validity of anthropomorphic simulators must rely on expert perception about the realism of the internal appearance of the simulator [17, 18]. To meet the requirement, the simulator developed for this present study was lined with colored texturized silicone to give it a realistic aspect, mimicking the endometrial cavity, and the tubal ostia were marked [19]. The silicone lining also aimed to protect the hysteroscope lens during the training sessions. Face validity includes the similarity between simulated and real-life maneuvers as perceived by expert users [20]. The present simulator shares high face validity with other recently developed hysteroscopy simulators [21].

High content validity indicates that the simulator allows the acquisition of the basic skills necessary to safely and proficiently complete the addressed procedures. Content validity grants the simulator the property of serving as a useful training tool. Office-based hysteroscopic experience is necessary to assess content validity, only experts' ratings were used in this study, as preconized elsewhere [18, 22]. Because the simulator includes three procedure-specific uterine models, users can acquire skills in the most frequent and challenging maneuvers performed during office-based hysteroscopy. Training navigation includes the manipulation of the hysteroscope within the uterine cavity to obtain a clear view through the 30°-angle optics; novices can also train the reaching of small targets inside the uterine cavity as required for the successful performance of hysteroscopic endometrial biopsy; the simulator also permits the acquisition of skills at manipulating grasping forceps, which are crucial for foreign body removal. Furthermore, the simulator provides haptic feedback that increases the realism of the task [23]. To the authors' knowledge, no other study addressed the content validity of simulators designed to the acquisition of skills for office-based hysteroscopy [3, 24].

High construct validity of a simulator indicates the ability to differentiate the performance of experts and novices practicing on it. Besides, the demonstration that training in the simulator improves skills over time is also an indicator of construct validity, which is a necessary feature of simulators designed to follow the learning curve of novices during their path to proficiency [25]. By strategically positioning targets on every uterine wall, the simulator posed distinct levels of difficulty to assess the targets, thus

allowing for differentiation between expert and novice performance. Furthermore, the success at progressively more difficult targets progressively increased the summative score, allowing for the detection of technical performance improvement among novices. Similarly, progressively decreased time to touch the targets might also be evidence of performance improvement, as demonstrated in a previous study [26]. Although significant improvement in technical ability and decreases in task completion time during five attempts, residents did not reach experts' performance level, indicating that learning curves of the simulated tasks are longer and deserve further studies. However, this finding corroborates the need for hysteroscopic simulation training before a performance in real patients [3, 27, 28].

In addition to its high face, content and construct validity, the current simulator is appealing for its affordable production cost (approximately US\$ 150.00), easy reproducibility and portability. These characteristics make the simulator an attractive tool for training and diffusing office-based hysteroscopy technique in developing countries where scarce resources challenge the acquisition of expensive simulators [29].

Low cost and easy reproducibility were the main concerns in the project of the simulator. Incorporating 3-D technology into the manufacturing of the uterine models fulfilled both requirements. However, equipping the model with resources for training vaginoscopy, the insertion of the hysteroscope through the uterine cervix and managing irrigation and complications would have substantially increased the cost of the simulator. These limitations have also been found in other studies [19, 21, 30].

Conclusion

Before its implementation as a training tool, a newly developed simulator requires independent validation of its face, content and construct validity. This study described a new simulator designed for the acquisition of skills relevant to the safe performance of office-based hysteroscopic procedures. Training on the simulator produced realistic experiences for the user, an indication of high face validity; experts also highly rated the usefulness of the exercises performed on the simulator for the training of novices, an indication of content validity. Performance on the simulator also differentiates levels of surgical experience, an indication of construct validity.

List Of Abbreviations

3-D – Three dimensional

p - p-value

F – Factor; Factor analysis

IRB – Institutional review board

ABS - Acrylonitrile butadiene styrene

FDM - Fusion deposition modeling

USB – Universal Serial Bus

Fr – French

GLM mixed-model - Generalized linear mixed model

e^2 - Effect size

CI – Confidence interval

Declarations

Ethical Approval and consent to participate

Approved by the IRB under CAAE - 88475118.7.0000.0121 on June 19, 2018, of the participating institutions. Written informed consent was obtained from participants before inclusion in the study.

Consent for publication

Not applicable.

Availability of data and material

Data will be made available on request.

Competing interests

The authors declare that they have no competing interests.

Source of funding

None

Authors' contributions

Ana Rita Peixoto Panazzolo: this author designed and managed the construction of the simulator, participated in the design of the study and in the construction of the measurement instruments, collected data, participated in statistical analyses, produced the manuscript and approved its final version.

Leisa Beatriz Grando, MD: this author participated in the design and construction of the measurement instruments, reviewed and rated the video recordings, read and approved the final version of the manuscript.

Lia Karina Volpato, MD: this author participated in the design and construction of the measurement instruments, reviewed and rated the video recordings, read and approved the final version of the manuscript.

Getúlio R. de Oliveira Filho: this author participated in the design the study, in statistical analyses, in the production of the manuscript, and approved its final version.

Statement of prior presentation or publications and/or abstract/poster presentation: The abstracts with partial results of this research were accepted for poster presentation at the 59 Brazilian Congress of Gynecology and Obstetrics in November 2019. These data have not been published elsewhere and the manuscript is not under review at any other journal.

DISCLOSURES

The authors declare that they have no conflicts of interest.

Statement of prior presentation or publications and/or abstract/poster presentation: The abstracts with partial results of this research were accepted for poster presentation at the 59 Brazilian Congress of Gynecology and Obstetrics in November 2019. These data have not been published elsewhere and the manuscript is not under review at any other journal.

Acknowledgements

The authors thank Roxana Knobel, MD, PHD for her support to build the simulator, and English teacher and translator Keila Campos de Oliveira for her linguistic support.

Appendix

MODULES	ITEMS	CLASSIFICATION		
		Unskilled 1	Somewhat skilled 2	Skilled 3
ERGONOMY	Posture	Poor posture, such as shoulder contraction, excessive body movement over 50% of the procedures	Alternate periods of good and bad posture during procedures	Good and relaxed posture during procedures
	Hand movement	Excessive and uncontrolled hand movements and no proper hysteroscope rotation on over 50% of movements	Controlled movements, but no proper hysteroscope rotation	Economic motion control and proper hysteroscope rotation
	Hand/eye coordination	Looked at the simulator or hands more than 50% of the time while performing the task	Looked at the simulator or hands less 50% the time	Did not look at the simulator or hands while performing the task
IMAGE QUALITY (only for the navigation exercise)	Camera alignment	Horizontal alignment loss on over 50% of movements	Horizontal alignment loss on less than 50% of movements	Good camera alignment
	Centering and good image view	Difficulty viewing and centering images	View images but had difficulty centering them	Easily viewed and centered images
SAFE NAVIGATION	Safe navigation	Touched the walls and/or did not have a logical sequence and/or difficulty in panoramic viewing	Did not have a logical sequence, but did not touch the walls, performed panoramic viewing	Executed a good sequence, did not touch the walls and performed panoramic viewing
HANDLING OF INSTRUMENTS	Handling of instruments	Difficulty to depress the sensors and pull the mini-balloons, big distance between the grasping forceps and hysteroscope	Depressed the sensors and pulled mini-balloons, but maintained big distance between the forceps and hysteroscope	Depressed the sensors, pulled mini-balloons and maintained good distance between forceps and hysteroscope

APPENDIX:

Technical performance checklist

References

1. Salazar CA, Isaacson KB (2018) Office operative hysteroscopy: an update. *J Minim Invasive Gynecol* 25(2):199–208. <https://doi.org/10.1016/j.jmig.2017.08.009>.
2. Di Spiezio SA, Bettocchi S, Spinelli M, et al (2010) Review of new office-based hysteroscopic procedures 2003-2009. *J Minim Invasive Gynecol* 17(4):436–48. <https://doi.org/10.1016/j.jmig.2010.03.014>.
3. Gambadauro P, Milenkovic M, Hadlaczky G (2018) Simulation for training and assessment in hysteroscopy: a systematic review. *J Minim Invasive Gynecol* 25(6):963–73. <https://doi.org/10.1016/j.jmig.2018.03.024>.
4. Ericsson KA, Krampe RT, Tesch-Römer C (1993) The role of deliberate practice in the acquisition of expert performance. *Psychol Rev* 100(3):363–406. <https://doi.org/10.1037/0033-295x.100.3.363>.
5. Rackow BW, Solnik MJ, Tu FF, Senapati S, Pozolo KE, Du H (2012) Deliberate practice improves obstetrics and gynecology residents' hysteroscopy skills. *J Grad Med Educ* 4(3):329–34. <https://doi.org/10.4300/JGME-D-11-00077.1>.
6. Munro MG (2012) Surgical Simulation: Where have we come from? Where are we now? Where are we going? *J Minim Invasive Gynecol* 19(3):272–83. <https://doi.org/10.1016/j.jmig.2012.01.012>.
7. Campo R, Puga M, Meier FR, Wattiez A, De Wilde RL (2014) Excellence needs training “Certified programme in endoscopic surgery”. *Facts, Views Vis ObGyn* 6(4):240–4.
8. Bharathan R, Setchell T, Miskry T, Darzi A, Aggarwal R (2014) Gynecologic endoscopy skills training and assessment: review. *J Minim Invasive Gynecol* 21(1):28–43. <https://doi.org/10.1016/j.jmig.2013.07.016>.
9. Gallagher AG, Ritter EM, Satava RM (2003) Fundamental principles of validation, and reliability: rigorous science for the assessment of surgical education and training. *Surg Endosc Other Interv Tech* 17(10):1525–9. <https://doi.org/10.1007/s00464-003-0035-4>.
10. Bajka MI, Tuchs Schmid S, Fink D, Székely G, Harders M (2010) Establishing construct validity of a virtual-reality training simulator for hysteroscopy via a multimetric scoring system. *Surg Endosc Other Interv Tech* 24(1):79–88. <https://doi.org/10.1007/s00464-009-0582-4>.
11. De Oliveira Filho GR, Mettrau FAC (2018) The effect of high-frequency, structured expert feedback on the learning curves of basic interventional ultrasound skills applied to regional anesthesia. *Anesth Analg* 126(3):1028–34. <https://doi.org/10.1213/ANE.0000000000002748>.
12. Tuchs Schmid S, Bajka M, Bachofen D, Székely G, Harders M (2007) Objective surgical performance assessment for virtual hysteroscopy. *Stud Health Technol Inform* 125:473–8.
13. Zevin B, Levy JS, Satava RM, Grantcharov TP (2012) A consensus-based framework for design, validation, and implementation of simulation-based training curricula in surgery. *J Am Coll Surg* 215(4):580-586.e3. <https://doi.org/10.1016/j.jamcollsurg.2012.05.035>.
14. Jacoby J, Michael SM (2014) Three-Point Likert Scales. *J Mark Res* 8(4):495–500.
15. Burchard ER, Lockrow EG, Zahn CM, Dunlow SG, Satin AJ (2007) Simulation training improves resident performance in operative hysteroscopic resection techniques. *Am J Obstet Gynecol* 197(5):542.e1-542.e4. <https://doi.org/10.1016/j.ajog.2007.07.023>.

16. Koo TK., Li MY (2016) A guideline of selecting and reporting intraclass correlation coefficients for reliability research. *J Chiropr Med* 15(2):155–63. <https://doi.org/10.1016/j.jcm.2016.02.012>.
17. Bright E, Vine S, Wilson MR, Masters RSW, McGrath JS (2012) Face validity, construct validity and training benefits of a virtual reality turp simulator. *Int J Surg* 10(3):163–6. <https://doi.org/10.1016/j.ijsu.2012.02.012>.
18. Dulan G, Rege RV, Hogg DC, Gilberg-Fisher KK, Tesfay ST, Scott DJ (2012) Content and face validity of a comprehensive robotic skills training program for general surgery, urology, and gynecology. *Am J Surg* 203(4):535–9. <https://doi.org/10.1016/j.amjsurg.2011.09.021>.
19. Janse JA, Tolman CJ, Veersema S, Broekmans FJM, Schreuder HWR (2014) Hysteroscopy training and learning curve of 30° camera navigation on a new box trainer: The HYSTT. *Gynecol Surg* 11(2):67–73. <https://doi.org/10.1007/s10397-014-0833-9>.
20. Verdaasdonk EGG, Stassen LPS, Monteny LJ, Dankelman J (2006) Validation of a new basic virtual reality simulator for training of basic endoscopic skills: The SIMENDO. *Surg Endosc Other Interv Tech* 20(3):511–8. <https://doi.org/10.1007/s00464-005-0230-6>.
21. Munro MG., Brown AN., Saadat S, et al (2019) Essentials in minimally invasive gynecology manual skills pilot validation trial. *J Minim Invasive Gynecol* <https://doi.org/10.1016/j.jmig.2019.04.018>.
22. De Oliveira Filho GR, Helayel PE, Conceição DB, Garzel IS, Pavei P, Ceccon MS (2008) Learning curves and mathematical models for interventional ultrasound basic skills. *Anesth Analg* 106(2):568–73. <https://doi.org/10.1213/ane.0b013e3181605412>.
23. Lin Y, Wang X, Wu F, Chen X, Wang C, Shen G (2014) Development and validation of a surgical training simulator with haptic feedback for learning bone-sawing skill. *J Biomed Inform* 48:122–9. <https://doi.org/10.1016/j.jbi.2013.12.010>.
24. Savran MM, Maya S, Sørensen D, et al (2016) Training and assessment of hysteroscopic skills: a systematic review. *J Surg Educ* 73(5):906–18. <https://doi.org/10.1016/j.jsurg.2016.04.006>.
25. Mokkink LB, Terwee CB, Gibbons E, et al (2010) Inter-rater agreement and reliability of the COSMIN (Consensus-based standards for the selection of health status measurement instruments) checklist. *BMC Med Res Methodol* 10(1):82. <https://doi.org/10.1186/1471-2288-10-82>.
26. Alici F, Buerkle B, Tempfer CB (2014) Objective Structured Assessment of Technical Skills (OSATS) evaluation of hysteroscopy training: a prospective study. *Eur J Obstet Gynecol Reprod Biol* 178:1–5. <https://doi.org/10.1016/j.ejogrb.2014.04.032>.
27. Janse JA., Driessen SRCC, Veersema S, Broekmans FJMM, Jansen FW, Schreuder Henk WRR (2014) Training of hysteroscopic skills in residency program: the Dutch experience. *J Surg Educ* 72(2):345–50. <https://doi.org/10.1016/j.jsurg.2014.09.003>.
28. Elessawy M, Skrzypczyk M, Eckmann-scholz C, et al (2016) Integration and Validation of Hysteroscopy Simulation in the Surgical Training Curriculum. *J Surg Educ* 74(1):84–90. <https://doi.org/10.1016/j.jsurg.2016.06.007>.
29. Sleiman Z, Rassi E, Tanos V, Khazzaka A (2018) From the dry lab to the operating room: the first transferability study of hysteroscopic skills. *Surg Endosc* 33: 1854-1857.

<https://doi.org/10.1007/s00464-018-6464-x>.

30. Kingston A, Abbott J, Lenart M, Vancaillie T (2004) Hysteroscopic training: The butternut pumpkin model. *J Am Assoc Gynecol Laparosc* 11(2):256–61. [https://doi.org/10.1016/S1074-3804\(05\)60210-1](https://doi.org/10.1016/S1074-3804(05)60210-1).

Figures



Figure 1

The simulator. An internal view of the simulator with the three uterine models embedded in a foam base and fixed with Velcro straps is shown on the left panel. Male condoms connect the aperture of the model to the wall of the simulator box. From left to right, the uterine models designed to the exercises of navigation, direct biopsy, and foreign body removal are shown. Sensor wires in the second model and balloons inserted through the wall of the third model can also be seen. Images of the internal surface of the models intended for training navigation, direct biopsy, and foreign body extraction are shown from top to bottom in the right panel of the figure.





Figure 2

Navigation task. Panel A - Global performance scores; Panel B – Task completion time. Brackets and the respective vertically oriented p-values refer to between-group comparisons. Horizontally oriented p-values refer to within-group comparisons to the first attempt.






Figure 3

Direct biopsy task - Panel A - Global performance scores; Panel B – Task completion time. Brackets and the respective vertically oriented p-values refer to between-group comparisons. Horizontally oriented p-values refer to within-group comparisons to the first attempt.

A - Task 3 - Global performance score at foreign body removal tasks



B - Task 3 - Time to complete foreign body removal tasks




Figure 4

Foreign body removal task - Panel A - Global performance scores; Panel B – Task completion time. Brackets and the respective vertically oriented p-values refer to between-group comparisons. Horizontally oriented p-values refer to within-group comparisons to the first attempt.