

Feasibility of Conebeam CT-based Online Adaptive Radiotherapy for Neoadjuvant Treatment of Rectal Cancer.

Rianne de Jong (✉ m.a.j.dejong@amsterdamumc.nl)

Amsterdam Universitair Medische Centra <https://orcid.org/0000-0003-3833-3479>

Jorrit Visser

Amsterdam UMC - Locatie AMC: Amsterdam UMC Locatie AMC

Niek van Wieringen

Amsterdam UMC - Locatie AMC: Amsterdam UMC Locatie AMC

Jan Wiersma

Amsterdam UMC - Locatie AMC: Amsterdam UMC Locatie AMC

Debby Geijsen

Amsterdam UMC - Locatie AMC: Amsterdam UMC Locatie AMC

Arjan Bel

Amsterdam UMC - Locatie AMC: Amsterdam UMC Locatie AMC

Research

Keywords: Adaptive radiotherapy, Adaptive treatment, Rectal cancer, Conebeam CT

Posted Date: March 13th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-292967/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published at Radiation Oncology on July 23rd, 2021. See the published version at <https://doi.org/10.1186/s13014-021-01866-7>.

Abstract

Background

Online adaptive radiotherapy has the potential to reduce toxicity for patients treated for rectal cancer because smaller planning target volumes (PTV) margins around the entire clinical target volume (CTV) are required. The aim of this study is to describe the first clinical experience of a Conebeam CT (CBCT)-based online adaptive workflow, evaluating timing of different steps in the workflow, plan quality, target coverage and patient compliance.

Methods

Twelve consecutive patients eligible for 5 x 5 Gy pre-operative radiotherapy were treated on a ring-based linear accelerator with a multidisciplinary team present at the treatment machine for each fraction. The accelerator is operated using an integrated software platform for both treatment planning and delivery. In all directions for all CTVs a PTV margin of 5 mm was used, except for the cranial/caudal borders of the total CTV where a margin of 8mm was applied. A reference plan was generated based on a single planning CT. After aligning the patient the online adaptive procedure started with acquisition of a CBCT. The planning CT scan was registered to the CBCT using deformable registration and a synthetic CT scan was generated. With the support of artificial intelligence, structure guided deformation and the synthetic CT scan contours were adapted by the system to match the anatomy on the CBCT. If necessary, these contours were adjusted before a new plan was generated. A second and third CBCT were acquired to validate the new plan with respect to CTV coverage just before and after treatment delivery, respectively. Treatment was delivered using volumetric modulated arc treatment (VMAT). All steps in this process defined and timed.

Results

On average the timeslot needed at the treatment machine was 34 minutes. The process of acquiring a CBCT, evaluating and adjusting the contours, creating the new plan and verifying the CTV on the CBCT scan took on average 20 minutes. Including delivery and post treatment verification this was 26 minutes. Manual adjustments of the target volumes were necessary in 50% of fractions. Plan quality, target coverage and patient compliance were excellent.

Conclusions

First clinical experience with CBCT-based online adaptive radiotherapy shows it is feasible for rectal cancer.

Trial registration

Medical Research Involving Human Subjects Act (WMO) does not apply to this study and was retrospectively approved by the Medical Ethics review Committee of the Academic Medical Center

(W21_087 # 21.097; Amsterdam University Medical Centers, Location Academic Medical Center, Amsterdam, The Netherlands).

Background

Radiotherapy is standard of care in neoadjuvant treatment of intermediate and locally advanced rectal cancer, generally followed by TME surgery¹. In intermediate risk tumors radiotherapy is applied with a prescribed dose of 5 x 5 Gy in order to sterilize surgery planes, while for locally advanced rectal cancer a dose of 25 x 2 Gy is given for downstaging to improve rates of complete resection. Radiotherapy comes inevitably at the cost of toxicity as a result of dose to healthy tissue, in case of rectal cancer mostly due to dose to bladder and small bowel^{2,3}.

To ensure target coverage during treatment the radiotherapy target volume as defined on a single reference CT is enlarged with a PTV margin. The size of this margin is based on all uncertainties in the treatment chain of radiotherapy⁴. For rectal cancer the daily change in volume and shape is the main contributor to the PTV margin and results in a required margin of up to 25 mm at the ventral side of the upper mesorectum⁵⁻⁷ in order for the CTV to receive at least 95% of the prescribed dose for 90% of the patients.

Multiple steps have been taken in the last two decades to minimize toxicity of radiotherapy treatment by reducing dose to the healthy tissue without compromising coverage of the target volume. First, treatment delivery and planning technique evolved to intensity modulated radiation therapy (IMRT) and volumetric modulated arc therapy (VMAT), which enabled tightly shaped dose distribution around the PTV with a steep dose gradient⁸⁻¹⁰. Secondly, integrated in-room 3D-kV imaging (Conebeam CT (CBCT)) made it possible to position the patient and assess target coverage on a daily basis¹¹. In case of insufficient target coverage due to systematic local anatomical changes or more global anatomical changes that would result in a suboptimal dose distribution, the plan could be adapted in an offline setting using repeat CT scan¹². Online CBCT also allows a plan selection strategy where a-priori generated plans are available and the plan that best fits the anatomy of the day is selected¹³⁻¹⁵. We previously demonstrated that this plan selection strategy can reduce dose to the organs at risk (OAR) significantly for individual patients, yet whereas the benefit is substantial for individual cases it is limited for the group of patients as a whole¹⁶.

Recently, technologies have become available that enable online adaptive radiotherapy (ART) such as in-room online MRI guidance. It combines improved image quality (with respect to CBCT) with software programs that allow automatic identification of the target volume and organs at risk and automatic re-planning¹⁷. In a previous study we showed that there is a large dosimetric advantage of online ART for rectal cancer¹⁸. Currently, also a more cost economic online adaptive system has become commercially available that uses CBCT for the online workflow. The goal of this study was to evaluate for the first time the feasibility of CBCT-based online ART procedure for rectal cancer in the neoadjuvant setting.

This study describes and evaluates a CBCT-based online adaptive workflow with respect to timing of different steps in the process, plan quality, target coverage and patient compliance.

Methods

Patients

Twelve consecutive rectal cancer patients eligible for a prescribed dose of 5 x 5 Gy according to Dutch guidelines¹⁹ were scheduled for external beam radiotherapy (Ethos, Varian Medical Systems, Palo Alto, USA) between June 2020 and February 2021. Exclusion criteria were hip prosthesis on both sides and an inability to lie still at the treatment table for a period of 30 minutes. Patient compliance with respect to bladder filling and total treatment time were monitored.

Reference CT and target volumes

A reference CT was acquired with a drinking instruction aiming at a comfortably filled bladder. To achieve this, patients were asked to void the bladder 1.5 hours before the scheduled CT appointment and subsequently drink 0.5 liter of fluid. All patients were positioned supine using a knee support and with arms raised over the head using a thorax support (Thorax support, MacroMedics). All patients received intra venous contrast and female patients with distal tumors vaginal contrast as well.

Clinical target volumes were delineated (Velocity 4.1, Varian Medical Systems) using the national delineation guidelines following Valentini et al. (Fig. 1)²⁰. Next to the GTV (tumor and positive lymph nodes) the radiation oncologist separately delineated the CTV upper and lower mesorectum (divided at the base of bladder), presacral space and left and right elective lymph node regions. Organs at risks (OAR) were delineated by Radiation Therapists (RTTs) according to RTOG guidelines as well as the rectum. Target volumes were reviewed by a second expert radiation oncologist.

Margins

In all directions for all CTVs a PTV margin of 5 mm was used, with the exception in the cranial direction of the presacral, mesorectum upper and lymph node CTV and for the caudal direction of the mesorectum lower CTV. Because it was expected to be difficult to discern the cranial and caudal borders of the target volumes on CBCT it was decided to use a PTV margin of 8 mm in these cases.

Treatment planning

The planning CT and delineations were exported from Velocity to Aria (Varian Medical System, version 16.00.00). In Ethos Treatment Management (Varian Medical Systems, version 02.00.10) a template was loaded with a departmental prioritized list of clinical goals. These clinical goals consisted of the evaluation objectives that were used for plan evaluation, as specified in the clinical protocol. Additionally, in order to achieve a more desirable dose distribution than achieved by only supplying the evaluation objectives, also optimization objectives were added as clinical goals in the template. The CT was

imported into Ethos from Aria, after which the body contour was automatically delineated as well as the bony structures. If present, gas pockets and regions with contrast fluid were delineated and a material assignment was applied, where water was used as material. Next a preview of the dose distribution was automatically generated by the system. This dose preview was generated using a fluence optimized nine beams IMRT plan (system assigned) with the provided clinical goals as input. Based on the dose preview the clinical goals that were used as optimization objectives were adjusted to further improve the plan as routine practice for all patients. Subsequently, the final clinical goals and material assignments were used by the system to generate multiple deliverable IMRT and VMAT plans with fixed beam setup. From these plans the most suitable plan was selected as reference plan and approved by a radiation oncologist. As part of the QA protocol, an independent dose calculation was performed (Mobius3D 3.1, Varian Medical Systems) on the reference plan. The pass rate was required to be larger than 90%, which indicated the percentage of voxels with $\gamma < 1$ (3%/3mm, where the percentage was relative to the maximum dose and only voxels with a dose of more than 10% of the maximum dose were taken into account). In addition, the reference plan was delivered to a phantom (Octavius 4D, PTW, Germany) and a pass rate of at least 90% was required (γ 3%/3 mm, where the percentage was relative to the local dose and only voxels with more than 10% of the maximum dose were taken into account).

Adaptive Workflow

For all patients and all fractions one physicist, one radiation oncologist and two RTTs were present at the treatment machine. RTTs were in charge of running the software and adjusted structures under supervision of the radiation oncologist. For every patient, prior to the first treatment session, a 30-minute timeslot was scheduled to discuss patient specific clinical target volumes and reference plan to avoid discussions during the online adaptive workflow. Because both radiation oncologist and two RTTs were evaluating contours on a single monitor a checklist was developed to streamline and order this process (Supplement 1). A flowchart of the adaptive workflow on the treatment machine is provided in Fig. 2.

After patient setup the first CBCT was acquired (scatter grid, 125 kV, 1080 mAs, iterative reconstruction, extended FOV if the CTV in CC direction exceeds 18 cm with a maximum of 25 cm, matrix 512x512).

The system generated a synthetic CT by deforming the pretreatment planning CT to the CBCT using mutual information. The resulting vector field was used by the system to propagate the body contour, bony structures, and material override structures from the pretreatment planning CT to the synthetic CT.

Subsequently, the system generated delineations using Artificial Intelligence of the following pelvic organs: rectum and bladder. In this system these structures are called 'influencer structures' as they influence the deformation of the target volumes using structure-guided deformable registration. If necessary, these delineations were adjusted by the RTT.

In the following step the system combined the deformation vector field and the influencer structures to automatically propagate the target volumes from the pretreatment planning CT to the CBCT. At the moment the system determined the target volumes of the patients current anatomy it presented these

target structures to the users and at the same time started 1) generating a newly optimized treatment plan using the beam setup and clinical goals of the reference plan and 2) calculating the dose distribution of the unaltered reference plan but using a isocenter translation aligning the target volumes on pretreatment planning CT and those propagated to the CBCT. In both calculations the system used the patients anatomy as represented by the synthetic CT including body contour and material assignments. The newly created plan was called the adapted plan, whereas and the reference plan recalculated on current anatomy was referred to as the scheduled plan.

If necessary, the RTT adjusted the propagated target volumes yielding a restart of 1) and 2) described above, at the moment the adjusted the target volumes were approved.

RTTs, together with the radiation oncologist and physicist, evaluated both plans by comparing the clinical goals and the dose distribution, after which the best plan was selected.

After approval of the selected treatment plan a second CBCT scan (same variables as first CBCT except for 540 mAs) was acquired to verify if the target volumes were still within the PTV. If the table displacement based on a bony anatomy registration to the first CBCT scan was more than 1 mm in any direction a table translation was applied.

Concurrently, an independent calculation of the dose distribution was done as part of our QA protocol (Mobius, Varian Medical Systems), where a pass rate of 90% was required (gamma 3%/3mm). Additionally, as a sanity check, the number of monitor units (MU) of the selected plan was compared to the number of MU of the reference plan.

After treatment delivery a third, post RT CBCT scan was acquired to again check if the target volume, as visible on the post treatment CBCT scan was inside the PTV of the plan used for treatment.

Timing of each step was captured to provide an overview as well as the number of times the target volumes needed to be adjusted by the users. Since this was a novel procedure for our department with no extensive clinical experience, unplanned events were recorded.

Results

An overview of the twelve patients included in this study can be found in Table 1, a summary of the timing of all steps of the online adaptive workflow can be found in Table 2.

In-room time, on-table time and patient compliance

The average total treatment time defined as the patient entering and exiting the treatment room was 34 minutes.

The complete online adaptive workflow including all CBCT imaging, adjustments of contours, plan calculation and treatment delivery (excluding patient alignment) took on average 26 minutes.

Patients all tolerated the time needed for the procedure.

Contour adjustments

The average time to spent to evaluate and adjust the system generated contours of bladder and rectum (influencers) needed for contour propagation of the target volumes was 4 minutes.

Subsequently, the time spent evaluating the target volumes when adjustments were not applied was on average 4 minutes. If the target volumes required adjustments the average time increased to 9 minutes. In 30 out of 60 fractions (50%) the target volumes were adjusted after visual inspection.

Plan calculation, plan quality and independent dose calculation

When selecting the most suitable reference plan a VMAT delivery technique was selected to limit delivery time to minimize the possibility of intra fraction motion. The average time needed to calculate scheduled and adapted plan was 8 minutes (synchronous calculation).

To assess the quality of the scheduled and the adapted plan the volume of the PTV receiving 95% of the prescribed dose or more (V95%) and the conformity of the 95% isodose to the PTV was checked and used for selecting the best plan. For all fractions the adapted plan was selected. For 55 out of 60 fractions V95% of PTV was less than the required 99% for the scheduled plan, while for the adapted plan this was the case for 3 fractions (Fig. 3).

The number of MU of the selected plan never deviated more than 15% from the number of MU of the reference plan (Fig. 4).

All patients passed the independent dose calculation.

Online procedure

Completing the online adaptive procedure, i.e. from the first CBCT followed by contour adjustments (influencers and target volumes), plan calculation up to and including the second CBCT for verification took on average 20 minutes.

Treatment delivery and intra fraction motion

Treatment delivery took on average 4 minutes.

All post RT CBCTs showed adequate coverage of the target volume.

Unplanned events

For a few fractions the adaptive procedure did not go according to plan as a result of the following:

- 1) the workflow was interrupted after first CBCT due to a very full bladder in one fraction, because it was expected that the patient required voiding the bladder before the end of the fraction. (Fig. 5a).
- 2) the workflow was completely restarted after the second CBCT because of insufficient target coverage due to intra fraction motion of a large gas pocket in one fraction.
- 3) after the second CBCT there was insufficient target coverage due to limited intra fraction motion of a gas pocket. Coverage was restored with a table shift in one fraction.
- 4) for one patient the synthetic CT had a small error in 3 out of 5 fractions with respect to bony anatomy registration that resulted in a need to extensively adapt the presacral delineation in that region (Fig. 5b).
- 5) for one patient the synthetic CT had an error in all fractions with respect to body contour definition (Fig. 5c). This error was ignored because limited impact on the dose distribution was expected.

Discussion

To our knowledge, this is the first study that describes an online CBCT-based adaptive radiotherapy workflow for rectal cancer in the neoadjuvant setting and reports the first clinical experience. All scheduled patients were treated as intended. The workflow worked well but was labor intensive as it required a multi-disciplinary team at the treatment machine and compared to our previously used plan selection protocol^{14,16} time slots at the treatment machine are prolonged (15 minutes versus 35 minutes). Patient compliance was not affected.

Treatment times

Intven et al.²¹ was the first to report daily adaptive radiotherapy for rectal cancer patients treated with 5 x 5 Gy and similar target volumes and showed it was a feasible strategy for MR guided radiotherapy.

Overall, the online adaptive workflow they described took longer in comparison to our workflow. Median of the total treatment time defined as the time between first MR scan and end of treatment delivery was longer with a median of 43 minutes (IQR 9 minutes), as compared to 26 minutes (range 16–46) in our study. They report contouring time with a median of 13 minutes (IQR 11). In our workflow contours can be evaluated and adapted in two steps: 1) the system generated structures of the pelvic organs (the influencers), i.e. rectum and bladder, and 2) the target volumes. This first step is not part of the MR guided workflow. When only the system generated delineations of bladder and rectum were evaluated and/or adjusted it took on average 4 minutes. If on top of that the target volumes needed to be adapted this increased to an average of 9 minutes.

When selecting a reference plan for our online workflow the VMAT technique was favored for its fast delivery over IMRT. The 5-field IMRT technique used for delivery by Intven et al.²¹ resulted in a median of 7 minutes (IQR 1) treatment delivery time compared to an average of 4 minutes (range 3–6) in our study.

Imaging

The system provided the possibility of iterative reconstructed and produced sufficient image quality for the evaluation and adjustment of the influencer structures and target volumes. Incidental an unfavorable patient anatomy, causing a lot of streak artefacts (moving gas in small bowel), increased the time needed to evaluate and adapt contours.

For now there is no comparing literature for using CBCT images in an online adaptive workflow, it is commonly performed with an MR-Linac. MR guided radiotherapy has the potential benefit of higher image quality compared to CBCT scans²¹. Possibly, the MRI image quality could result in more accurate re-delineation of contours which needs to be investigated. MRI guided imaging could also enable the visualization of the GTV and treatment response which is not possible using CBCT-based online adaptive radiotherapy.

Margins and intra fraction motion

We previously compared plan selection and online ART with respect to dose to the healthy tissue¹⁸. In that study a 3 mm PTV margin around the total target volume was used. As this was our first clinically implemented online adaptive workflow with the Ethos system we decided to start with a slightly larger margin: 5 mm in all directions with the exception of 8 mm in cranial direction for upper mesorectum, presacral space and elective lymph node regions and 8 mm in caudal direction for lower mesorectum because it was expected to be difficult to discern the cranial and caudal border of the target volume on CBCT. In this decision it was also taken into account that when using smaller margins possibly more adjustments of small deviations would be needed, consequently increasing the time needed for the online ART process. Conversely, larger margins would possibly mean less adjustments are needed of small deviations, but this would yield less advantage of the online ART process with respect to dose to the OARs.

To assess target coverage in the context of intra fraction motion, we also acquired a pretreatment and post treatment CBCT scan. As stated, the adapted plan needed to be shifted in one fraction and workflow was interrupted in one fraction because of insufficient target coverage. Post treatment CBCT scans showed good coverage for all performed adapted treatments. Whether margins can be reduced further needs to be investigated. Intven et al.²¹ started using 10 mm PTV margin around the mesorectum in all directions and 8 mm PTV margin around lymph node regions. After 25 patients they reduced margins to the mesorectum to 4 mm in all directions except for 6 mm in CC and ventral direction. For the lymph node regions the margins were reduced to 4 mm in all directions except for 6 mm in CC direction. These PTV margin reductions were based on an evaluation of adequate coverage to the target on the post treatment MRI scans.

Resources

When designing the adaptive protocol we aimed for an RTT led workflow from the start. Therefore, we protocolized that RTTs would drive the software under supervision of the radiation oncologist and the medical physicist for the first 12 patients. RTTs were already experienced in CBCT-based online IGRT and plan selection for rectal cancer. For the adaptive workflow, additional training for the RTTs was provided in the form of the ESTRO Falcon delineation course followed by a target definition workshop for departmental specific criteria. This was combined with 40 hours of individual training on a research version of the clinical software after individual instructions. A total of 5 RTTs were trained. To streamline and order the process of evaluation and adaptation of contours we developed a checklist, see Supplement 1. An offline QA protocol would help and is currently under investigation as well as the possibility to remotely view the screens to support a workflow without the physician and/or physicist physically present at the treatment machine.

Intven et al.²¹ report that their workflow started with recontouring performed by radiation oncologists for the first 12 patients. As of patient 13 RTTs were trained followed by a gradual shift in the responsibility of the recontouring to RTTs. This shift towards an RTT led online adaptive procedure is in line with the results of an international survey published by McNair et al.²².

In our workflow also the medical physicist was present at the treatment machine to approve the adapted plan and to evaluate the independent dose calculation. A traffic light protocol is would further enable an autonomous RTT led workflow in the future as is described by Betgen et al.²³.

The time slots at the treatment machine were 40 minutes. The difference between the on table time (26 minutes) and the total treatment time as defined by patient entering and leaving the treatment room (34 minutes) is large. This is most likely the result of a not fully booked patient schedule, leaving extra time for social interaction between RTTs and the patients.

Patient compliance

In general, rectal cancer patients receive a bladder filling instruction that aims at comfortable full bladders during treatment. We estimated the treatment time for online ART to be about 30 minutes and did not change drinking instructions as a (comfortable) full bladder is beneficial for dose to the bladder and it improves CBCT image quality. The average in-room time as defined by patient entering and leaving the treatment room was on average 34 minutes but with outliers up to 54 minutes. All patients were compliant with these treatment times, although for one fraction the workflow was interrupted based on an extremely full bladder on the first CBCT scan.

Prolonged treatment times (compared to previous plan selection time slots) can affect patient comfort as they are immobilized with the arms elevated. As a result we needed to alter arm position to arms crossed on the chest for one patient after the first treatment fraction.

Scheduled and adapted plan

The system automatically provides a scheduled plan next to the adapted plan. The user is able to compare both plans with respect to clinical goals and dose distribution. In this study the scheduled plan was never selected (Fig. 3).

Future directions

With a feasible CBCT-based online adaptive workflow a new resource has become available for radiotherapy departments to improve treatment. Whether the CBCT image quality suffices for other treatment sites needs to be investigated. To improve CBCT-based online ART effort has to be made to assess remaining and new uncertainties to calculate optimal margins. Also the supporting software tools in this process needs to be optimized to reduce time per fraction. In view of the considerable reduction of irradiated healthy tissue it is likely that ultimately it will be proven that this technique improves patients' quality of life.

Conclusion

CBCT-based online adaptive radiotherapy in the neo adjuvant treatment of intermediate and locally advanced rectal cancer is feasible and appears to be a promising strategy to reduce PTV margins and thus toxicity.

List Of Abbreviations

ART Adaptive Radiation Therapy

CBCT conebeam CT

CT computed tomography

CTV clinical target volume

GTV gross tumor volume

Gy gray

IMRT intensity modulated radiation therapy

MU Monitor unit

OAR Organs at Risk

PTV planning target volume

RTT radiation therapist

VMAT volumetric modulated arc therapy

Declarations

Ethics approval and consent to participate

The institutional ethics committee granted an exemption from requiring ethics approval in this study (PENDING; Amsterdam University Medical Centers, Location Academic Medical Center, Amsterdam, The Netherlands).

Consent for publication

Figures depicting individual person's data are unidentifiable and there are no details on individuals reported within the manuscript. Therefore, separate consent for publication is not deemed necessary for the figures shown this manuscript.

Availability of data and materials

The datasets generated and/or analyzed during the current study are not publicly available since the participants did not consent in sharing the data with third parties.

Competing interests

A.B. is involved in several industry-sponsored projects (Elekta and Varian) outside of this work. These companies had no involvement in study design, data collection and analysis, or writing of the manuscript. The other authors have no relevant conflicts of interest to disclose.

Funding

Not applicable.

Authors' contributions

R.d.J., J.V., N.v.W., J.W., D.G., A.B. made substantial contributions to the design of the clinical adaptive strategy used in this study. R.d.J. analyzed the data and drafted the manuscript. All authors have contributed significantly to the manuscript, have read the manuscript and have given final approval of the version to be published, and take public responsibility for appropriate portions of the content.

Acknowledgements

Not applicable.

References

1. van de Velde CJ, Boelens PG, Borrás JM, et al. EURECCA colorectal: multidisciplinary management: European consensus conference colon & rectum. *Eur J Cancer*. 2014;50(1):1 e1-1 e34.

2. Peeters KC, van de Velde CJ, Leer JW, et al. Late side effects of short-course preoperative radiotherapy combined with total mesorectal excision for rectal cancer: increased bowel dysfunction in irradiated patients—a Dutch colorectal cancer group study. *J Clin Oncol*. 2005;23(25):6199-6206.
3. Appelt AL, Bentzen SM, Jakobsen A, Vogelius IR. Dose-response of acute urinary toxicity of long-course preoperative chemoradiotherapy for rectal cancer. *Acta Oncol*. 2015;54(2):179-186.
4. van Herk M. Errors and margins in radiotherapy. *Semin Radiat Oncol*. 2004;14(1):52-64.
5. Nijkamp J, de Jong R, Sonke JJ, Remeijer P, van Vliet C, Marijnen C. Target volume shape variation during hypo-fractionated preoperative irradiation of rectal cancer patients. *Radiother Oncol*. 2009;92(2):202-209.
6. Nijkamp J, de Jong R, Sonke JJ, van Vliet C, Marijnen C. Target volume shape variation during irradiation of rectal cancer patients in supine position: comparison with prone position. *Radiother Oncol*. 2009;93(2):285-292.
7. Nijkamp J, Swellengrebel M, Hollmann B, et al. Repeat CT assessed CTV variation and PTV margins for short- and long-course pre-operative RT of rectal cancer. *Radiother Oncol*. 2012;102(3):399-405.
8. Duthoy W, De Gerssem W, Vergote K, et al. Clinical implementation of intensity-modulated arc therapy (IMAT) for rectal cancer. *Int J Radiat Oncol Biol Phys*. 2004;60(3):794-806.
9. Samuelian JM, Callister MD, Ashman JB, Young-Fadok TM, Borad MJ, Gunderson LL. Reduced acute bowel toxicity in patients treated with intensity-modulated radiotherapy for rectal cancer. *Int J Radiat Oncol Biol Phys*. 2012;82(5):1981-1987.
10. Guerrero Urbano MT, Henrys AJ, Adams EJ, et al. Intensity-modulated radiotherapy in patients with locally advanced rectal cancer reduces volume of bowel treated to high dose levels. *Int J Radiat Oncol Biol Phys*. 2006;65(3):907-916.
11. Smitsmans MH, Wolthaus JW, Artignan X, et al. Automatic localization of the prostate for on-line or off-line image-guided radiotherapy. *Int J Radiat Oncol Biol Phys*. 2004;60(2):623-635.
12. Sonke JJ, Aznar M, Rasch C. Adaptive Radiotherapy for Anatomical Changes. *Semin Radiat Oncol*. 2019;29(3):245-257.
13. Beekman C, van Triest B, van Beek S, Sonke JJ, Remeijer P. Margin and PTV volume reduction using a population based library of plans strategy for rectal cancer radiotherapy. *Med Phys*. 2018;45(10):4345-4354.
14. de Jong R, Lutkenhaus L, van Wieringen N, et al. Plan selection strategy for rectum cancer patients: An interobserver study to assess clinical feasibility. *Radiother Oncol*. 2016;120(2):207-211.
15. Lutkenhaus LJ, de Jong R, Geijssen ED, Visser J, van Wieringen N, Bel A. Potential dosimetric benefit of an adaptive plan selection strategy for short-course radiotherapy in rectal cancer patients. *Radiother Oncol*. 2016;119(3):525-530.
16. de Jong R, Visser J, Crama KF, et al. Dosimetric benefit of an adaptive treatment by means of plan selection for rectal cancer patients in both short and long course radiation therapy. *Radiat Oncol*. 2020;15(1):13.

17. Raaymakers BW, Lagendijk JJ, Overweg J, et al. Integrating a 1.5 T MRI scanner with a 6 MV accelerator: proof of concept. *Phys Med Biol*. 2009;54(12):N229-237.
18. de Jong R, Crama KF, Visser J, et al. Online adaptive radiotherapy compared to plan selection for rectal cancer: quantifying the benefit. *Radiat Oncol*. 2020;15(1):162.
19. www.richtlijndatabase.nl.
20. Valentini V, Gambacorta MA, Barbaro B, et al. International consensus guidelines on Clinical Target Volume delineation in rectal cancer. *Radiother Oncol*. 2016;120(2):195-201.
21. Intven MPW, de Mol van Otterloo SR, Mook S, et al. Online adaptive MR-guided radiotherapy for rectal cancer; feasibility of the workflow on a 1.5T MR-linac: clinical implementation and initial experience. *Radiother Oncol*. 2020;154:172-178.
22. McNair HA, Wiseman T, Joyce E, Peet B, Huddart RA. International survey; current practice in On-line adaptive radiotherapy (ART) delivered using Magnetic Resonance Image (MRI) guidance. *Tech Innov Patient Support Radiat Oncol*. 2020;16:1-9.
23. Betgen A, Bilderbeek J, Janssen T, et al. Real time IGART - Changing responsibilities for RTTs on the MR-Linac. *Radiotherapy and Oncology*. 2019;141:S5.

Tables

Table 1:

Patient characteristics

	Age	Sex	Tumor stage	GTV location	Chemo	Surgery	Remarks
1	66	F	cT3bN1M0 MRF-	mid	no	yes	
2	82	M	T4N0M0 MRF+ EMVI+	mid	no	yes	
3	39	M	T3N2M1	distal	yes	no	Reirradiation
4	62	M	cT3cN2M1 MRF+	distal	yes	yes	
5	49	M	cT4bN2M1 MRF+	distal	yes	yes	
6	46	F	cT4aN2bM1 MRF+	distal	yes	no	
7	50	M	cT3c-T4M1 MRF+ EMVI+	distal	yes		Surgery pending chemo
8	81	M	iT3aN0Mx MRF- EMVI-	distal	yes		+ Oesophagus tumor, Surgery pending chemo
9	75	F	cT4N1M0 MRF+	distal	yes	yes	+ Sigmoid tumor
10	47	M	cT3bN2M0 MRF-	distal	yes	yes	
11	69	M	cT3N1M0 MRF- EMVI+	proximal	no	yes	
12	62	M	cT3bN1M0 MRF- EMVI-	distal	no	yes	

Table 2:

Overview of steps and duration

Time (Minutes) needed for:	AVG	min	max
1 Evaluation and adjustment system generated contours (Influencers)	4	1	11
2 Evaluation target volume without adjustments	4	2	11
3 Evaluation and adjustment target volume	9	4	21
4 Adaptive procedure before delivery (CBCT2 - CBCT1)	20	11	40
5 Plan calculation	8	6	11
6 Treatment delivery	4	3	6
7 On table (CBCT3 - CBCT1)	26	16	46
8 Patient entering and leaving treatment room	34	20	54

Figures




Figure 1

Illustration of planning CT with target definition in blue. On the right, the axial slice shows the CTV of the upper mesorectum and lymph node region left and right while on left the sagittal slice shows the lower mesorectum, upper mesorectum and presacral space (bottom to top).




Figure 2

Flowchart of the online adaptive workflow.

PTV V95%




Figure 3

Boxplot showing PTV V95% of the scheduled and the adapted plan.

MU of adapted plan




Figure 4

Boxplot showing monitor units distribution of adapted plan relative to reference plan.




Figure 5

CBCT with a too full bladder at start of treatment (5a). A small error with respect to deformable bony anatomy registration of CT to CBCT. In pink the representation of the bony anatomy of the sCT with overlay on the CBCT(5b). A small error with respect to deformed body contour of CT to CBCT. In green the representation of the body contour of the sCT with overlay on the CBCT(5c).

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [Supplement1.docx](#)