

Cyclotron-based production of ^{68}Ga , $[^{68}\text{Ga}]\text{GaCl}_3$, and $[^{68}\text{Ga}]\text{Ga-PSMA-11}$ from a liquid target

Melissa E Rodnick

University of Michigan

Carina Sollert

GE Healthcare

Daniela Stark

Royal Prince Alfred Hospital

Mara Clark

University of Michigan

Andrew Katsifis

Royal Prince Alfred Hospital

Brian G Hockley

University of Michigan

D. Christian Parr

GE Healthcare

Jens Frigell

GE Healthcare

Bradford D Henderson

University of Michigan

Monica Abghari-Gerst

University of Michigan

Morand R Piert

University of Michigan

Michael J Fulham

Royal Prince Alfred Hospital

Stefan Eberl

Royal Prince Alfred Hospital

Katherine Gagnon

GE Healthcare

Peter J. H. Scott (✉ pjhscott@umich.edu)

University of Michigan <https://orcid.org/0000-0002-6505-0450>

Keywords: Gallium-68, cyclotron targetry, positron emission tomography, PSMA

Posted Date: August 18th, 2020

DOI: <https://doi.org/10.21203/rs.3.rs-38981/v2>

License:  This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published on November 12th, 2020. See the published version at <https://doi.org/10.1186/s41181-020-00106-9>.

Abstract

Purpose: To optimize the direct production of ^{68}Ga on a cyclotron, via the $^{68}\text{Zn}(p,n)^{68}\text{Ga}$ reaction using a liquid cyclotron target. We investigated the yield of cyclotron-produced ^{68}Ga , extraction of $[\text{}^{68}\text{Ga}]\text{GaCl}_3$ and subsequent $[\text{}^{68}\text{Ga}]\text{Ga-PSMA-11}$ labeling using an automated synthesis module.

Methods: Irradiations of a 1.0 M solution of $[\text{}^{68}\text{Zn}]\text{Zn}(\text{NO}_3)_2$ in dilute (0.2-0.3 M) HNO_3 were conducted using GE PETtrace cyclotrons and GE ^{68}Ga liquid targets. The proton beam energy was degraded to a nominal 14.3 MeV to minimize the co-production of ^{67}Ga through the $^{68}\text{Zn}(p,2n)^{67}\text{Ga}$ reaction without unduly compromising ^{68}Ga yields. We also evaluated the effects of varying beam times (50-75 min) and beam currents (27-40 μA). Crude ^{68}Ga production was measured. The extraction of $[\text{}^{68}\text{Ga}]\text{GaCl}_3$ was performed using a 2 column solid phase method on the GE FASTlab Developer platform. Extracted $[\text{}^{68}\text{Ga}]\text{GaCl}_3$ was used to label $[\text{}^{68}\text{Ga}]\text{Ga-PSMA-11}$ that was intended for clinical use.

Results: The decay corrected yield of ^{68}Ga at EOB was typically >3.7 GBq (100 mCi) for a 60 min beam, with irradiations of $[\text{}^{68}\text{Zn}]\text{Zn}(\text{NO}_3)_2$ at 0.3 M HNO_3 . Target/chemistry performance was more consistent when compared with 0.2 M HNO_3 . Radionuclidic purity of ^{68}Ga was typically $>99.8\%$ at EOB and met the requirements specified in the European Pharmacopoeia ($<2\%$ combined $^{66/67}\text{Ga}$) for a practical clinical product shelf-life. The activity yield of $[\text{}^{68}\text{Ga}]\text{GaCl}_3$ was typically $>50\%$ (~ 1.85 GBq, 50 mCi); yields improved as processes were optimized. Labeling yields for $[\text{}^{68}\text{Ga}]\text{Ga-PSMA-11}$ were near quantitative (~ 1.67 GBq, 45mCi) at EOS. Cyclotron produced $[\text{}^{68}\text{Ga}]\text{Ga-PSMA-11}$ underwent full quality control, stability and sterility testing, and was implemented for human use at the University of Michigan as an Investigational New Drug through the US FDA and also at the Royal Prince Alfred Hospital (RPA).

Conclusion: Direct cyclotron irradiation of a liquid target provides clinically relevant quantities of $[\text{}^{68}\text{Ga}]\text{Ga-PSMA-11}$ and is a viable alternative to traditional $^{68}\text{Ge}/^{68}\text{Ga}$ generators.

1 Introduction

The medicinal use of ^{68}Ga was first described over 4 decades ago albeit with a very small clinical footprint for much of that time (Eder et al. 2014; Graham et al. 2017; Lenzo et al. 2018; Velikyan 2018). Over the past 15 years, there has been a surge in ^{68}Ga radiopharmaceutical development, exceeding that of other radiotracers, with a 100-fold increase in the number of ^{68}Ga publications. Over the last decade, there has also been a marked increase in the clinical use of ^{68}Ga radiotracers that has been attributed to the ease of acquiring ^{68}Ga from $^{68}\text{Ge}/^{68}\text{Ga}$ generators and the development and approval of new theranostic tracers (see Figure 1 for some recent examples) (Baum and Kulkarni 2012). The diagnostic applications of ^{68}Ga vary across jurisdictions/countries, but initial development was driven by development of theranostic agents targeting somatostatin receptors (SSTRs) for the PET imaging (^{68}Ga , ^{64}Cu) and either alpha (^{213}Bi , ^{225}Ac) or beta (^{90}Y , ^{177}Lu) therapy of neuroendocrine tumors (NETs)

(Graham et al. 2017; Jackson et al. 2017). In the United States, DOTATATE labeled with ^{68}Ga (NETSPOT) and ^{177}Lu (Lutathera) by the U.S. Food and Drug Administration for NET diagnosis and therapy while, the European Union approved [^{68}Ga]DOTATOC (SomaKIT TOC) and Lutathera. Subsequent development of theranostic agents for infection/inflammation (Velikyan 2018), prostate cancer (Eder et al. 2014; Lenzo et al. 2018; Ruangma et al. 2018), C-X-C chemokine receptor type 4 (CXCR4) (Gourni et al. 2011; Herrmann et al. 2016) and, most recently, fibroblast activation protein inhibitors (FAPI) (Kratochwil et al. 2019) is further driving demand and highlights the need for access to a reliable (and economical) supply of gallium-68 that is the focus of this paper. Analogous development of a reliable pipeline of therapeutic radionuclides is also an urgent need for the nuclear medicine community (Herrmann et al. 2020), but beyond the scope of this article.

Gallium-68 is usually eluted from a $^{68}\text{Ge}/^{68}\text{Ga}$ generator, and thus can be readily implemented in PET facilities that do not own a cyclotron. There are also many additional attributes of ^{68}Ga that make it a desirable PET radionuclide and the first widely available PET radioactive metal ion (radiometal) for routine use globally. ^{68}Ga is a positron emitting (89% β^+) radionuclide with a 68 min half-life that, although relatively short, is compatible with distribution. The $^{68}\text{Ga}^{3+}$ cation is small with an ionic radius of 0.62 Å, which behaves as a relatively hard Lewis acid with an affinity for binding ligands containing oxygen and nitrogen donors, and is suitable for conjugation to various biomolecular vectors using bifunctional chelators and various macromolecules including small molecules with rapid pharmacokinetic profiles, such as peptides and peptidomimetics (Blower et al. 2019; Martiniova et al. 2016; Smith et al. 2013). This synthetic diversity provides the ability for ^{68}Ga kit development.

The most recent main contributor to the expansion of ^{68}Ga -based PET has been imaging of the prostate specific membrane antigen (PSMA) with a host of new radiopharmaceuticals (Ruangma et al. 2018). Of these new agents, [^{68}Ga]Ga-PSMA-11 has been most widely used to date (Hope et al. 2017). Prostate cancer is the second most common cancer found in men in the United States and the second most prevalent cause of cancer death in men (Prostate Cancer: Statistics, 2020). Survival rates depend on the type of prostate cancer and the stage at diagnosis. Men with localized disease have a 5-year survival rate of nearly 100%. However, 20-40% of these patients develop biochemical recurrence (BCR) and the recurrent disease can be loco-regional or more widespread. Patients with metastatic disease have a markedly decreased 5-year survival rate of 30% (Prostate Cancer: Statistics, 2020). The early and accurate identification of tumor recurrence and metastatic disease is essential for optimal patient management, but this remains a major challenge for traditional imaging methods with anatomical imaging and bone scintigraphy.

Figure 1. Theranostic Radiopharmaceuticals

The imaging of PSMA expression with [^{68}Ga]Ga-PSMA-11 and PET/CT has proven to be a highly effective and sensitive tool for patient management (Blower et al. 2019). While the primary use of [^{68}Ga]Ga-PSMA-11 has been for detecting recurrent disease, it has also been successful at staging

primary prostate cancer, and useful for guiding biopsies to improve sample accuracy, guiding surgery, and monitoring treatment response (Lenzo et al. 2018). There has been a positive clinical impact of [⁶⁸Ga]Ga-PSMA-11 at the University of Michigan with a change in patient care management in 72% of the scanned patient population. A similar high impact has been reported in a large Australian study that cited a 51% change in care management (62% for BCR patients and 21% for primary staging) (Roach et al. 2018). In 2016 a study from Belgium reported that in patients who underwent a [⁶⁸Ga]Ga-PSMA-11 scan there was a 76% impact in patient care management (Albisinni et al. 2017). A 2017 study from the University of California San Francisco reported a 53% change in patient management (Hope et al. 2017). Additionally, [⁶⁸Ga]Ga-PSMA-11 has been used theranostically in conjunction with complementary ¹⁷⁷Lu (i.e. β-) or ²²⁵Ac (i.e. α) therapeutic PSMA targeting agents. Such PSMA targeted therapies are currently undergoing evaluation in clinical trials in patients with castrate-resistant metastatic prostate cancer (Fendler et al. 2017; Kratochwil et al. 2018).

Since [⁶⁸Ga]Ga-PSMA-11 has higher accuracy and sensitivity in detecting metastatic disease than [¹⁸F]fluorocholine, [¹¹C]choline, and CT (Afshar-Oromieh et al. 2014; Blower et al. 2019; McCormick et al. 2019; Schwenck et al. 2017), a superior detection rate to [¹⁸F]fluciclovine (Calais et al. 2018), and overall superior clinical performance (Lenzo et al. 2018; McCormick et al. 2019; Schwenck et al. 2017), it is rapidly becoming the most commonly used radiotracer for prostate cancer management. This presents challenges in meeting the expected demand for the agent once regulatory approval is gained. To put this into context, [¹¹C]choline has been one of the most widely used prostate cancer radiotracers in the US since its FDA approval in 2012 (Evans et al. 2018), and the busiest cancer centers reportedly perform 10-15 [¹¹C]choline scans daily (Lowe and Kwon 2015). It has been possible to service this volume of patients given high yielding [¹¹C]choline syntheses (>7.4 GBq (200 mCi) /dose) (Shao et al. 2011; Shao et al. 2010), coupled with the ability to run production multiple times per day (limited only by cyclotron and synthesis module availability). Contrastingly, transferring such a patient population to exclusively [⁶⁸Ga]Ga-PSMA-11 PET is not feasible in a workflow relying entirely on ⁶⁸Ge/⁶⁸Ga generators. While ⁶⁸Ge/⁶⁸Ga generators offer workflow simplicity for tracer production but there are a number of limitations: a) current GMP generators have a maximum activity of 1.85 GBq (50 mCi) and are restricted to elutions every 3-4-hour increments, which in practice typically means 2 production runs per day with 2-4 doses per day; b) two or more generators increase the number of patients doses to 6 or more, but still less than the requirements of busy cancer centers; c) commercial supply has not kept pace with the clinical demand and lead times for generator delivery can be up to 18 months in some markets (Cutler and Minoshima, 2018); d) the eluted activity constantly declines over time and so to ensure a regular clinical supply of [⁶⁸Ga]Ga-PSMA-11, multiple sequential and overlapping generators must be purchased throughout the year and; e) there is the potential for long lived parent ⁶⁸Ge contamination and/or breakthrough. To this end, an additional source of ⁶⁸Ga needs to be explored and implemented into the clinical setting to meet the current and future patient demand (Cutler and Minoshima, 2018).

An attractive alternative to diversifying the supply of ^{68}Ga is the direct production of ^{68}Ga on a cyclotron, via the $^{68}\text{Zn}(p,n)^{68}\text{Ga}$ reaction. This alternative approach has garnered significant interest by the community, including publication of a European Pharmacopeia monograph for the direct accelerator-based production of [^{68}Ga]GaCl₃ which was published in draft form in 2018 and finalized in 2020 (Gallium (^{68}Ga) chloride (accelerator-produced) solution for radiolabeling, 2020) and a technical document published by the IAEA in support of direct production of ^{68}Ga via liquid and solid targets (International Atomic Energy Agency, 2019). There are two strategies for producing ^{68}Ga via the $^{68}\text{Zn}(p,n)^{68}\text{Ga}$ reaction on a cyclotron - namely, liquid (Alves et al. 2017; International Atomic Energy Agency, 2019; Jensen and Clark 2010; Gallium (^{68}Ga) chloride (accelerator-produced) solution for radiolabeling, 2020; Nair et al. 2017; Oehlke et al. 2015; Pandey et al. 2019; Pandey et al. 2014; Pandey and DeGrado 2019) and solid targets (Boschi et al. 2019; Engle et al. 2012; Lin et al. 2018; Sadeghi et al. 2009; Schweinsberg et al. 2019; Tolmachev and Lundqvist 1996; Zeisler et al. 2019). Liquid targets offer implementation simplicity for sites familiar with [^{18}F]FDG production as they present a similar workflow to production of [^{18}F]fluoride and are compatible with laboratory set-ups in existing PET radiopharmaceutical production centers. Solid targets, however, typically impose increased requirements on infrastructure and/or local site expertise but offer more than an order of magnitude higher ^{68}Ga yields (e.g. several GBq/Ci) (Lin et al. 2018; Schweinsberg et al. 2019). Regardless of opting for liquid or solid targets an efficient means for purifying the ^{68}Ga from the irradiated ^{68}Zn is required. The limitations of cyclotron produced ^{68}Ga are obviously: a) a cyclotron with suitable targets, b) the co-production of ^{67}Ga and ^{66}Ga and, c) the potential for residual levels of ^{68}Zn and other metal impurities affecting labeling efficiencies. These factors place stringent demands on the proton energy, the target material and reagent quality, and finally $^{68}\text{Zn}/^{68}\text{Ga}$ separation methods.

We present results of the liquid target-based production of ^{68}Ga on GE PETtrace cyclotrons, with focus on yield of ^{68}Ga and extraction of [^{68}Ga]GaCl₃ using the GE FASTlab Developer platform. Furthermore, to demonstrate the clinical relevance of this direct production method, a single FASTlab cassette was used to perform the $^{68}\text{Zn}/^{68}\text{Ga}$ purification and subsequent labeling of [^{68}Ga]Ga-PSMA-11. The cyclotron produced [^{68}Ga]Ga-PSMA-11 underwent full quality control, stability and sterility testing, and has been used in humans at the University of Michigan (UM, Ann Arbor Mi, USA) under the FDA's Investigational New Drug (IND) program, and at Royal Prince Alfred Hospital (RPA, Sydney, Australia) under exemption of the Therapeutic Goods Act (TGA) in a TGA GMP-licensed facility. The results from UM, GEMS (GE Healthcare Uppsala, Sweden) and RPA are presented.

2 Materials And Methods

2.1 Liquid target irradiations

The GE ^{68}Ga PETtrace Liquid Target (Figure 2) is a water-cooled, gridded target without requiring He cooling of foils, designed specifically for ^{68}Ga production. The target comprises a 200 μm thick aluminum energy degrader, a 25 μm Havar foil for support, and a 25 μm niobium foil for chemical inertness with the target media, thus rendering a nominal 14.3 MeV incident proton energy on the target media. Including the target lines/dead volume, the total target fill volume is approximately 2.2 mL.

Figure 2. GE Gallium-68 Liquid Target

The target media was prepared from isotopically enriched [^{68}Zn]ZnO (Isoflex, USA) with addition of water (Ultrapur or 18 M Ω -cm) and 70% nitric acid (>99.999% trace metal basis) to yield a 1.0 M solution of [^{68}Zn]Zn(NO $_3$) $_2$ with an excess 0.2 M or 0.3 M HNO $_3$ (both concentrations tested). All irradiations at UM and GEMS and the majority of irradiations at RPA employed the same lot of enriched ^{68}Zn – namely: ^{64}Zn (0.03%), ^{66}Zn (0.16%), ^{67}Zn (0.62%), ^{68}Zn (99.16%), and ^{70}Zn (0.03%), and from a chemical perspective, comprised 1 ppm iron. Recent irradiations at RPA used a different lot of enriched ^{68}Zn – namely: ^{64}Zn (0.1%), ^{66}Zn (0.18%), ^{67}Zn (0.96%), ^{68}Zn (98.20%), and ^{70}Zn (0.56%), and Fe 3.1 ppm.

Irradiations were performed on GE PETtrace cyclotrons using the ^{68}Ga Liquid Target and were typically 50-70 minutes in duration with beam currents of \sim 30 - 40 μA . Whenever possible, within the routine daily production schedule, a “cleaning” irradiation at 30-35 μA of typically 10 - 60 minutes was performed with dilute nitric acid (0.6 M) after irradiation of the [^{68}Zn]Zn(NO $_3$) $_2$ solution.

2.2 Chemical Isolation on the fastLAB

2.2.1 Delivery to the FASTLAB

To facilitate use of the same FASTlab for both ^{68}Ga and ^{18}F processing and the dilution of the delivered ^{68}Ga target solution, the irradiated target media was delivered from the cyclotron into an external 10 mL V-vial with connections to the FASTlab (Figure 3). Thus, delivery of ^{68}Ga target material completely bypasses the incoming activity plunger of the FASTlab module avoiding potential cross contamination between $^{18}\text{F}^-$ and ^{68}Ga target deliveries when the module is used for both types of targets. In this activity receiving vial, the ^{68}Ga target solution was automatically diluted with water from the synthesis unit to achieve a nitric acid concentration of <0.1 M required for subsequent processing. The diluted target solution is automatically loaded onto the cassette by nitrogen overpressure.

Figure 3. Activity receiving vial for collection of the irradiated ^{68}Zn solution

2.2.2 Chemical isolation of $[^{68}\text{Ga}]\text{GaCl}_3$

A primary goal of this effort was to develop a FASTlab cassette which allowed for $[^{68}\text{Ga}]\text{GaCl}_3$ extraction in a formulation comparable with existing generators. Additionally, on-line column conditioning, the use of minimum quantities of acid, and the exclusion of organic solvents or base-mediated pH adjustments were desired. Chemical isolation of $[^{68}\text{Ga}]\text{GaCl}_3$ was implemented on the GE FASTlab Developer platform. The process is based on the 2-column approach we have previously reported for liquid targets (Nair et al. 2017) and recently repeated by Riga and colleagues (Riga et al. 2018). The process is shown in Figure 4. Initial separation of ^{68}Ga from ^{68}Zn is performed by trapping the ^{68}Ga on a hydroxamate-based resin (ZR resin, Triskem) cartridge. Further purification, concentration and acid reduction is realized by using a TOPO-based resin (TK200 resin, Triskem) cartridge.

In our initial efforts, elution of the TK200 resin with water (Scheme A in Table 1) resulted in a $[^{68}\text{Ga}]\text{GaCl}_3$ solution containing approximately 0.6 M HCl due to residual HCl content in the cartridge. Implementation of a NaCl/HCl rinse (Mueller et al. 2012) for reduction of residual acid achieved a final $[^{68}\text{Ga}]\text{GaCl}_3$ formulation of 0.1 M HCl in 5 mL (Scheme B in Table 1). This formulation is directly comparable to commercially available $^{68}\text{Ge}/^{68}\text{Ga}$ generators and is compatible with formulations required for pharmaceutical cold kit labeling.

Figure 4. Two-column approach for ^{68}Ga chemical separation

Process steps:

1. Condition ZR cartridge with 0.1 M HNO_3 (7 mL) and TK200 resin with sterile water for injection (7 mL) followed by 1.75 M HCl (4 mL) prior to use.
2. Trapping of ^{68}Ga on a hydroxamate-based resin (2 mL [\sim 700 mg] ZR resin, Triskem)
3. Rinsing of the resin to remove residual zinc
4. Elution onto a TOPO-based resin (2 mL [\sim 700 mg] TK200 resin, Triskem)
5. Wash to decrease residual acid content, and
6. Final elution with water and dilute hydrochloric acid, volumes of which can be varied, to yield $[^{68}\text{Ga}]\text{GaCl}_3$ in the desired formulation (e.g. 5 mL of 0.1 M HCl)

The process was optimized over time (with regards to flow rates, volumes, cassette rinsing, etc), thus not all runs were identical with regards to time lists on the FASTlab. Nevertheless, the chemical process can be categorized into two primary schemes (as noted in Table 1). Building on Scheme A, Scheme B includes

a wash step of the TK200 resin in order to reduce the residual acid content in the $[^{68}\text{Ga}]\text{GaCl}_3$ eluate. The purification time is approximately 30 minutes.

Table 1. High level schemes of $[^{68}\text{Ga}]\text{GaCl}_3$ purifications.

	Scheme A*	Scheme B
□ ZR Load	<0.1 M HNO ₃	
▣ ZR Wash	15 mL 0.1M HNO ₃	
□ ZR Elution / Trapping on TK200	5-6 mL ~1.75 M HCl	
□ TK Wash	–	3.5 mL 2.0 M NaCl in 0.13 M HCl
□ TK Elution	H ₂ O	1-2 mL H ₂ O followed by dilute HCl to formulate

*Process as reported previously (Nair et al. 2017)

In comparison to recent work, this method requires less acid and does not involve organic solvents or base-mediated pH adjustments, which is highlighted in Table 2.

Table 2. Comparison of FASTlab $[^{68}\text{Ga}]\text{GaCl}_3$ purification vs. recent literature

Reference	HNO ₃ * [mmol]	HCl [mmol]	Organic solvents	Base-mediated pH adjustment?
This work	2.4	16	No	No
(Oehlke et al. 2015)	-	886	Yes (Methanol)	No
(Alves et al. 2018)	-	265	Yes (HBr/acetone)	No
(Pandey and DeGrado 2019)	0.25	38	Yes (Acetonitrile)	Yes

*Does not account for HNO₃ in the liquid target.

The cassette layout for the automated $[^{68}\text{Ga}]\text{GaCl}_3$ separation on the FASTlab is given in Figure 5, noting that the $[^{68}\text{Ga}]\text{GaCl}_3$ chemistry is reserved to the right-hand side of the cassette. The left-hand side was kept vacant to enable subsequent on-cassette labeling (e.g. PSMA, NET tracers, etc), including C18 cartridge purification (Figure 6). The line labeled “to activity source” is connected to the activity receiving vial (Figure 3). Where applicable during the process, the vials of 0.6 M HNO₃, 4 M HCl, and 3 M NaCl were automatically diluted and/or mixed to the desired concentrations by the FASTlab.

In advance of receipt of the activity the columns were automatically conditioned on the FASTlab, the ZR resin was conditioned with 0.1 M HNO₃ (7 mL) and the TK200 was conditioned with both water (7 mL) followed by 1.75 M HCl (4 mL). Recycling of ⁶⁸Zn is not presently being performed given the current availability and cost of ⁶⁸Zn (approximately US\$100 per target fill), however, the ⁶⁸Zn solution is collected separately to facilitate future recycling.

Figure 5. FASTlab cassette layout – applicable to Schemes “A&B” of Table 3.

2.2.3 Synthesis of [⁶⁸Ga]Ga-PSMA-11

The direct cyclotron-based production of [⁶⁸Ga]Ga-PSMA-11 was executed at the UM, GEMS and RPA using a single FASTlab cassette in a continuous process to perform both the [⁶⁸Ga]GaCl₃ isolation chemistry and subsequent PSMA-11 labeling, including C18 purification. Cassettes were prepared at each institution based on the FASTlab developer cassettes and accessories.

Initial labeling tests (UM) employed [⁶⁸Ga]GaCl₃ separation scheme “A”, with 10 µg PSMA-11 precursor in 1.5 M Hepes (1 mL) and 3M NaOAc (1.3 mL) buffer. Scheme “B” developed and implemented at GEMS and RPA used 10 µg PSMA-11 precursor in 1.5 mL/1.0 M NaOAc (GEMS) or 1.3 mL/1.5 M NaOAc buffer (RPA) adjusted to pH 4.5-4.8. Approximately 3-4 mg L-ascorbic acid was also added (GEMS/RPA) to the precursor vial to minimize radiolysis during synthesis. An additional 20-21 mg of L-ascorbic acid (0.44 mL; 0.25 M) is also added directly into the product line at RPA as stabilizer of the final product. Labeling occurred for 5 min at 50 °C. At UM and RPA, the final product was reformulated into Phosphate Buffered Saline (PBS) using a C18 cartridge (preconditioned with 1.2 mL EtOH followed by 4.5 mL sterile water for injection). Briefly, at UM the crude reaction mixture was diluted (5 mL EtOH/ 38 mL sterile water for injection) and passed through the C18 cartridge to trap [⁶⁸Ga]Ga-PSMA-11. The cartridge was rinsed with additional sterile water for injection (9 mL and 10 mL), and [⁶⁸Ga]Ga-PSMA-11 was eluted from the C18 cartridge with EtOH/H₂O (1:1 v:v, 2 mL), sterile filtered (Millipore Cathivex-GV, 0.22 µm) and diluted with PBS solution (14.5 mL) to give the final formulated product that was submitted for quality control testing.

Figure 6. Partitioning of FASTlab cassette: Right-hand side is reserved for [⁶⁸Ga]GaCl₃ purification, and the left-hand side accommodates the [⁶⁸Ga]Ga-PSMA-11 synthesis and C18 cartridge based purification.

2.2.4 Quality Control of [⁶⁸Ga]Ga-PSMA-11

Quality control testing of [⁶⁸Ga]Ga-PSMA-11 doses was conducted according to the guidelines outlined in the U.S. and European Pharmacopeias using standard methods (visual inspection, pH, radionuclidic identity, sterile filter integrity, bacterial endotoxin analysis, and sterility testing) previously described (Mossine et al. 2020), as well as procedures specific for [⁶⁸Ga]Ga-PSMA-11 and described below.

Representative TLC Analysis of [⁶⁸Ga]Ga-PSMA-11

At UM, radiochemical purity was determined by thin layer chromatography (TLC) using silica-backed 7.5cm glass or paper chromatography plates and an Eckert and Ziegler AR-2000 TLC scanner. The plate was spotted with a sample of [⁶⁸Ga]Ga-PSMA-11 dose. The developing solution (mobile phase), 77 g/L ammonium acetate in a 50/50 water/methanol. The plate was developed in this solvent system, dried using a warm laboratory hotplate, and placed on the TLC scanner to determine radiochemical purity, which must be $\geq 90\%$.

Representative Radio-HPLC Analysis of [⁶⁸Ga]Ga-PSMA-11

At UM, chemical and radiochemical purities/identities were analyzed using a Shimadzu LC2010 HPLC equipped with a radioactivity detector and an ultraviolet (UV) detector (column: Phenomenex C18(2) 250 x 4.6 mm column; mobile phase A: 0.1% trifluoroacetic acid in MeCN; mobile phase B: 0.1% trifluoroacetic acid in milli-Q water; Gradient: 0 min A:B = 5:95 – 10 min A:B = 30:70 – 11 min A:B = 30:70 – 11.1 min A:B = 5:95 – 15 min A:B = 5:95; flow rate: 1.0 mL/min; UV: 205 nm). On this QC HPLC system, [⁶⁸Ga]Ga-PSMA-11 has a retention time of ~11 min. Radiochemical purity for doses was confirmed to be >90%, and identity was confirmed by comparing the retention time of the radiolabelled product with that of the corresponding unlabelled reference standard (40 µg/mL).

Representative Gallium-66 And Gallium-67 Concentration (Radionuclidic Purity)

At UM, cyclotron-produced [⁶⁸Ga]Ga-PSMA-11 was tested for ⁶⁶Ga and ⁶⁷Ga contamination weekly to determine the fitness of the target produced ⁶⁸Ga for use in production of clinical doses of [⁶⁸Ga]Ga-PSMA-11 for the following week. A small sample of [⁶⁸Ga]Ga-PSMA-11 was tested using gamma spectroscopy to determine an initial radioactivity profile. The sample was then stored in a shielded environment for 24-72 hours to allow ⁶⁸Ga to decay to near-zero, and the sample analyzed by gamma

spectroscopy a second time. Decay-corrected values for the amount of ^{66}Ga and ^{67}Ga were determined and used in conjunction with the initial radioactivity profile value to determine the relative amount of both radionuclides in the final product. The recent EU Monograph on cyclotron-produced requires $\leq 2\%$ combined ^{66}Ga and ^{67}Ga in doses until the end of the shelf-life.

3 Chemical Isolation On The Fastlab

3.1 ^{68}Ga Yields

Total ^{68}Ga yields from the target were assessed by: (a) downloading the total irradiated target contents into a vial placed in a dose calibrator without chemical purification and ensuring suitable decay time (90 - 120 min) or curve fitting to avoid any ^{13}N contribution, or (b) measurement of residual activity of cassette components and product post $[^{68}\text{Ga}]\text{GaCl}_3$ isolation or post $[^{68}\text{Ga}]\text{Ga-PSMA-11}$ labeling chemistry. For the data presented at GEMS, this includes an early series of 9 consecutive 60-minute irradiations from 30 – 40 μA (entire target contents), and 20 consecutive irradiations (post-chemistry) following a target rebuild.

Radioactivity yields exceeding 100 mCi (3.7 GBq) at EOB are typical (see Table 3) with irradiation of $[^{68}\text{Zn}]\text{Zn}(\text{NO}_3)_2$ at 0.3 M HNO_3 yielding more consistent target/chemistry performance. Albeit higher acid concentrations have been reported in the literature (Pandey et al. 2019), we opted to maintain the excess nitric acid as low as possible to minimize corrosive wear on components and facilitate the subsequent chemistry (which requires <0.1 M HNO_3 for ZR resin loading).

While it is theoretically possible to increase the target yields by increasing the ^{68}Zn concentration, the 1.0 M solution used here facilitates transfer to the hot cell (i.e. the solution is not too viscous). Should multi-Ci yields of ^{68}Ga be desired, adoption of the proposed method to solid targets as has been reported previously by taking advantage of ^{68}Ga trapping on ZR resin in high HCl concentration loading conditions (Schweinsberg et al. 2019).

Table 3. Summary of ^{68}Ga productions and total ^{68}Ga radioactivity yield at EOB

Site	HNO ₃ [M]	N	I [μA]	Beam time [min]	EOB activity [GBq]	EOB activity [mCi]	Measurement
UM	0.2	13	30	60	4.1 ± 0.6	112 ± 16	Entire target contents
	0.2	6	35	60	3.9 ± 0.6	106 ± 17	Entire target contents
	0.2	6	40	60	3.8 ± 0.4	102 ± 11	Entire target contents
	0.3	12	34 ± 4	60	4.6 ± 0.4	126 ± 12	Entire target contents
GEMS	0.2	9	36 ± 5	60	4.5 ± 0.3	120 ± 9	Entire target contents
	0.2	14	30	69 ± 7	3.5 ± 0.9	94 ± 24	☒ of parts post chemistry
	0.3	6	29 ± 1	70 ± 13	4.3 ± 0.5	115 ± 14	☒ of parts post chemistry
RPA	0.3	25	36 ± 2.2	60	4.0 ± 0.6	107 ± 17	Entire target contents
	0.3	53	35	60	3.8 ± 0.5	104 ± 14	☒ of parts post chemistry

3.2 [68Ga]GaCl₃, [68Ga]Ga-PSMA-11 – Yields and Quality

Several hundred irradiations and purifications/labelings have been performed throughout the development efforts, however, for sake of brevity, we report herein on several representative subsets of experimental data. These data are summarized in Tables 4 to 6.

Table 4 High-level summary of ⁶⁸Ga runs reported herein for UM, GEMS and RPA.

	Site	N	comment
[68Ga]GaCl ₃	UM	27	60 min beam current
	GEMS	13	Consecutive productions, 0.2 or 0.3 M HNO ₃
	RPA	20	60 min 35 μA beam, 0.3 M HNO ₃
[68Ga]Ga-PSMA-11	UM	3 + 35	Validation + clinical
	GEMS	3	Consecutive productions
	RPA	8	Validation + clinical

Table 5. Overview of [⁶⁸Ga]GaCl₃ productions (EOS)

Site	Chemistry Scheme	HNO ₃	I	Beam time	N	Product activity	
		[mol/L]	[μA]	[min]		[GBq]	[mCi]
UM	A	0.2	30	60	15	2.0 ± 0.3	54 ± 8
			35		6	2.0 ± 0.3	55 ± 8
			40		6	1.9 ± 0.2	50 ± 5
GEMS	B	0.2	30	64 ± 6	10	1.7 ± 0.5	46 ± 13
		0.3	29 ± 1	73 ± 6	3	2.5 ± 0.1	67 ± 3
RPA	B	0.3	35	60	20	2.0 ± 0.2	55 ± 6

Table 6. Overview of [⁶⁸Ga]Ga-PSMA-11 productions (EOS)

Site	HNO ₃	I	Beam time	N	Product activity		Notes
	[mol/L]	[μA]	[min]		[GBq]	[mCi]	
UM	0.2	30-40	60	3	1.6 ± 0.3	43 ± 9	Validation runs
UM	0.2	30-40	60	35	1.7 ± 0.2	45 ± 6	Clinical
GEMS	0.3	30	64 ± 4	3	2.1 ± 0.4	57 ± 10	R&D efforts
RPA	0.3	35	60	14	1.6 ± 0.1	44 ± 3	Final validation and clinical (5) runs

Table 5 clearly demonstrates a robust routine production of ~1.85 GBq (~50 mCi) of [⁶⁸Ga]GaCl₃ via the liquid target cyclotron route. This compares favorably with approximately ~1.48 GBq (~40 mCi) of [⁶⁸Ga]GaCl₃ from a brand new, highest commercially available activity GMP generator with ~1.85 GBq (~50 mCi) of ⁶⁸Ge. Furthermore, the eluted ⁶⁸Ga activity steadily decreases over time due to the decay of the ⁶⁸Ge.

[⁶⁸Ga]Ga-PSMA-11 activity yields at EOS varied slightly across sites (Table 6) which may be at least partly attributed to beam parameters, state of target and slightly different labeling conditions used. At RPA, 3 patients can be readily scanned from a single batch of [⁶⁸Ga]Ga-PSMA-11 using 2 scanners, which is the same number of patients which can be scanned with [⁶⁸Ga]Ga-PSMA-11 produced from 2 staggered ⁶⁸Ge/⁶⁸Ga generators. As more than 2 productions runs can potentially be performed with the target, the number of patients able to be scanned per day is potentially increased.

Although activity yield is an important parameter to measure process performance, the quality of the cyclotron-produced [⁶⁸Ga]GaCl₃ is of even greater importance as high quality [⁶⁸Ga]GaCl₃ is critical to enable efficient labeling. Therefore, in addition to yield measurement and periodic quality control (QC) assessment, validation studies for [⁶⁸Ga]GaCl₃ and [⁶⁸Ga]Ga-PSMA-11 were carried out at UM and for [⁶⁸Ga]Ga-PSMA-11 at RPA. The test methods performed (half-life, radiochemical purity, pH, radionuclidic purity, metal analysis) were in accordance with the EUP monograph for [⁶⁸Ga]GaCl₃ (Gallium (⁶⁸Ga) chloride (accelerator-produced) solution for radiolabeling, 2020) and are described in Section 2.2.4, with the exception of testing the Fe and Zn content, for which semi-quantitative colorimetric test strips (e.g. EM-Quant, Merck) and/or ICP-MS were used.

Table 7. Quality Control Data for three [⁶⁸Ga]GaCl₃ validation runs (UM)^a

TEST	1	2	3	Avg & SD	Release Criteria (EUP)
Radiochemical Purity [⁶⁸ Ga]GaCl ₃ (iTLC-SG)	99	98	98	98.3 ± 0.3	≥ 95
Rf [⁶⁸ Ga]GaCl ₃ (TLC)	< 0.2	< 0.2	<0.2	< 0.2	≤ 0.2
Rf Ref B ^b (TLC)	> 0.7	> 0.7	> 0.7	> 0.7	≥ 0.7
pH	< 2	< 2	< 2	< 2	< 2
Visual Inspection	Passed	Passed	Passed	N/A	Clear, colorless, no visible particulate
Radionuclidic Identity (t _{1/2})	67.2	68.8	69.1	68.4 ± 0.8	64.6 - 71.4 min
Endotoxin Analysis	< 2	< 2	< 2	< 2	≤ 58.3 EU/mL
Fe µg/GBq	<5	<5	<5	<5	≤ 10 µg/GBq
Zn µg/GBq	<1.25	<1.25	<1.25	<1.25	≤ 10 µg/GBq
RNP at EOB (MCA)	99.8	99.8	99.8	99.8	≥ 98% (at time of use)

^a After FASTLab isolation; ^b Reference solution B (Pentetic acid solution) from the European Pharmacopoeia (Gallium (⁶⁸Ga) chloride (accelerator-produced) solution for radiolabeling, 2020)

Table 8. Quality Control Data for three [⁶⁸Ga]Ga-PSMA-11 validation runs (UM)

Tests	1	2	3	Avg & SD	Release Criteria (UM)
Radiochemical Purity (via TLC)	99.5	99.4	99.3	99.4 ± 0.1	≥ 90%
Relative Retention time (via HPLC)	1.004	1.005	1.005	1.0046 ± 0.0003	RRT: 0.9 - 1.1
pH	7.0	7.0	7.0	7.0	4.0 - 8.0
Visual Inspection	Passed	Passed	Passed	N/A	Clear, colorless, no visible particulate
Radionuclidic Identity (t _{1/2})	67.61	68.45	67.20	67.75 ± 0.53	64.6 - 71.4 min
Endotoxin Analysis	< 2	< 2	< 2	< 2	≤ 10.9 EU/mL
Bubble Point (PSI)	51	52	53	52 ± 1	≥ 50 PSI
Sterility	Passed	Passed	Passed	Passed	Complies with USP<71> ^a
RNP at EOB (MCA)	99.8	99.8	99.8	99.8	≥ 98% (at time of use)

^a See: USP 71 Microbiological Tests/Sterility Tests, 2012.

Table 9. Quality Control Data for three [⁶⁸Ga]Ga-PSMA-11 validation runs (RPA)

Tests	1	2	3	Avg & SD	Release Criteria (RPA)
Radiochemical Purity (via TLC)	99.94	99.99	99.94	99.96 ± 0.03	≥ 95%
Radiochemical Purity (via HPLC)	99.94	99.97	100.0	99.97 ± 0.03	≥ 95%
pH	5.0	5.5	5.5	5.3 ± 0.3	4.0 - 8.0
Visual Inspection	Passed	Passed	Passed	N/A	Clear, colorless, no visible particulate
Radionuclidic Identity (t _{1/2})	67.9	68.1	67.7	67.9 ± 0.20	62 - 74 min
Endotoxin Analysis	< 1	< 1	< 1	< 1	≤ 17.5 EU/mL
Bubble Point (bar)	4.1	4.2	4.1	4.1 ± 0.06	≥ 3.5 bar
Sterility	Passed	Passed	Passed	Passed	Sterile – no growth
RNP at EOS (Well Counter)	99.7	99.8	99.8	99.8 ± 0.06	≥ 98% (at time of use)

The validation data shown here demonstrates the high quality of cyclotron-produced [⁶⁸Ga]GaCl₃ and [⁶⁸Ga]Ga-PSMA-11, and highlights the reliability and reproducibility of both processes. Notably, the reported RNP satisfied the proposed EU Pharmacopoeia limits (≤2% of combined ⁶⁶Ga + ⁶⁷Ga for the shelf life of the product) (Gallium (⁶⁸Ga) chloride (accelerator-produced) solution for radiolabeling, 2020). Setting this criteria for duration of shelf life and the varying amounts of ⁶⁶Ga + ⁶⁷Ga in each batch mean that the shelf-life will vary batch to batch. In this work the RNP was typically 0.2% at EOS. This allows determination of the shelf life to be approximately 3.5 h, after which time the amount of ⁶⁶Ga + ⁶⁷Ga will exceed 2% due to radioactive decay of shorter-lived ⁶⁸Ga. The dosimetry of the 2% ⁶⁶Ga + ⁶⁷Ga limit has been previously reported using worst-case assumptions, such as no biological clearance and rapid organ uptake (Graves et al. 2018). For this scenario, a relative dose increase up to 20% is reported but is typically less than 10% when compared to “pure” ⁶⁸Ga (i.e. not comparing with generator ⁶⁸Ga which may contain ⁶⁸Ge). Overall, the obtained results provided a solid basis for the clinical evaluation of cyclotron-produced [⁶⁸Ga]Ga-PSMA-11.

3.3 Clinical production and use of [⁶⁸Ga]Ga-PSMA-11

The necessary FDA approval was obtained to translate cyclotron-based [^{68}Ga]Ga-PSMA-11 into clinical use at UM and evaluation began in 2019. At RPA, cyclotron-based ^{68}Ga used for clinical [^{68}Ga]Ga-PSMA-11 production began in 2020. The final manufacturing process takes ~2 h from start of irradiation to release of the dose to the clinic. For example, the process at UM proceeds as follows:

- Cyclotron irradiation (60 min)
- [^{68}Ga]Ga-PSMA-11 synthesis (35 min)
- Quality control testing (25 min)

To date, over 700 patients have been scanned with [^{68}Ga]Ga-PSMA-11 at UM under our IND approval. Initially, this was with generator-based [^{68}Ga]Ga-PSMA-11. We amended the IND to include cyclotron-based [^{68}Ga]Ga-PSMA-11, and the first clinical production of cyclotron-based [^{68}Ga]Ga-PSMA-11 from a single FASTlab cassette occurred in February 2019. As of March 2020, 50 clinical batches of cyclotron-produced [^{68}Ga]Ga-PSMA-11 have been manufactured and used to scan more than 90 patients (see Table 6). A representative image from the first patient scanned with cyclotron-based [^{68}Ga]Ga-PSMA-11 is shown in Figure 6. We have noted no differences in the quality of studies where ^{68}Ga was produced from a cyclotron when compared to studies using ^{68}Ga obtained from a generator, and no pharmacological or physiological changes have been observed after intravenous administration of either generator-based or cyclotron-based [^{68}Ga]Ga-PSMA-11.

Figure 7. Images from the first patient scanned with [^{68}Ga]Ga-PSMA-11 labeled with cyclotron produced ^{68}Ga at the University of Michigan

4 Results And Discussion

4.1 ^{68}Ga Yields

Total ^{68}Ga yields from the target were assessed by: (a) downloading the total irradiated target contents into a vial placed in a dose calibrator without chemical purification and ensuring suitable decay time (90–120 min) or curve fitting to avoid any ^{13}N contribution, or (b) measurement of residual activity of cassette components and product post [^{68}Ga]GaCl₃ isolation or post [^{68}Ga]Ga-PSMA-11 labeling chemistry. For the data presented at GEMS, this includes an early series of 9 consecutive 60-minute irradiations from 30–40 μA (entire target contents), and 20 consecutive irradiations (post-chemistry) following a target rebuild.

Radioactivity yields exceeding 100 mCi (3.7 GBq) at EOB are typical (see Table 3) with irradiation of [^{68}Zn]Zn(NO₃)₂ at 0.3 M HNO₃ yielding more consistent target/chemistry performance. Albeit higher acid concentrations have been reported in the literature (29), we opted to maintain the excess nitric acid as low

as possible to minimize corrosive wear on components and facilitate the subsequent chemistry (which requires < 0.1 M HNO₃ for ZR resin loading).

While it is theoretically possible to increase the target yields by increasing the ⁶⁸Zn concentration, the 1.0 M solution used here facilitates transfer to the hot cell (i.e. the solution is not too viscous). Should multi-Ci yields of ⁶⁸Ga be desired, adoption of the proposed method to solid targets as has been reported previously by taking advantage of ⁶⁸Ga trapping on ZR resin in high HCl concentration loading conditions (38).

Table 3
Summary of ⁶⁸Ga productions and total ⁶⁸Ga radioactivity yield at EOB

Site	HNO ₃ [M]	N	I [uA]	Beam time [min]	EOB activity [GBq]	EOB activity [mCi]	Measurement
UM	0.2	13	30	60	4.1 ± 0.6	112 ± 16	Entire target contents
	0.2	6	35	60	3.9 ± 0.6	106 ± 17	Entire target contents
	0.2	6	40	60	3.8 ± 0.4	102 ± 11	Entire target contents
	0.3	12	34 ± 4	60	4.6 ± 0.4	126 ± 12	Entire target contents
GEMS	0.2	9	36 ± 5	60	4.5 ± 0.3	120 ± 9	Entire target contents
	0.2	14	30	69 ± 7	3.5 ± 0.9	94 ± 24	☒ of parts post chemistry
	0.3	6	29 ± 1	70 ± 13	4.3 ± 0.5	115 ± 14	☒ of parts post chemistry
RPA	0.3	25	36 ± 2.2	60	4.0 ± 0.6	107 ± 17	Entire target contents
	0.3	53	35	60	3.8 ± 0.5	104 ± 14	☒ of parts post chemistry

4.2 [⁶⁸Ga]GaCl₃, [⁶⁸Ga]Ga-PSMA-11 – Yields and Quality

Several hundred irradiations and purifications/labelings have been performed throughout the development efforts, however, for sake of brevity, we report herein on several representative subsets of experimental data. These data are summarized in Tables 4 to 6.

Table 4

High-level summary of ^{68}Ga runs reported herein for UM, GEMS and RPA.

	Site	N	comment
$[^{68}\text{Ga}]\text{GaCl}_3$	UM	27	60 min beam current
	GEMS	13	Consecutive productions, 0.2 or 0.3 M HNO_3
	RPA	20	60 min 35 μA beam, 0.3 M HNO_3
$[^{68}\text{Ga}]\text{Ga-PSMA-11}$	UM	3 + 35	Validation + clinical
	GEMS	3	Consecutive productions
	RPA	8	Validation + clinical

Table 5

Overview of $[^{68}\text{Ga}]\text{GaCl}_3$ productions (EOS)

Site	Chemistry Scheme	HNO_3 [mol/L]	I [μA]	Beam time [min]	N	Product activity [GBq] [mCi]	
UM	A	0.2	30	60	15	2.0 ± 0.3	54 ± 8
			35		6	2.0 ± 0.3	55 ± 8
			40		6	1.9 ± 0.2	50 ± 5
GEMS	B	0.2	30	64 ± 6	10	1.7 ± 0.5	46 ± 13
			0.3	29 ± 1	73 ± 6	3	2.5 ± 0.1
RPA	B	0.3	35	60	20	2.0 ± 0.2	55 ± 6

Table 6
Overview of [⁶⁸Ga]Ga-PSMA-11 productions (EOS)

Site	HNO ₃	I	Beam time	N	Product activity	Notes
	[mol/L]	[μA]	[min]		[GBq] [mCi]	
UM	0.2	30– 40	60	3	1.6 ± 0.3	43 ± 9 Validation runs
UM	0.2	30– 40	60	35	1.7 ± 0.2	45 ± 6 Clinical
GEMS	0.3	30	64 ± 4	3	2.1 ± 0.4	57 ± 10 R&D efforts
RPA	0.3	35	60	14	1.6 ± 0.1	44 ± 3 Final validation and clinical (5) runs

Table 5 clearly demonstrate a robust routine production of ~ 50 mCi of [⁶⁸Ga]GaCl₃ via the liquid target cyclotron route. This compares favorably with approximately 40 mCi of [⁶⁸Ga]GaCl₃ from a brand new, highest commercially available activity GMP generator with 50 mCi of ⁶⁸Ge. Furthermore, the eluted ⁶⁸Ga activity steadily decreases over time due to the decay of the ⁶⁸Ge.

[⁶⁸Ga]Ga-PSMA-11 activity yields at EOS varied slightly across sites (Table 6) which may be at least partly attributed to beam parameters, state of target and slightly different labeling conditions used. At RPA, 3 patients can be readily scanned from a single batch of [⁶⁸Ga]Ga-PSMA-11 using 2 scanners, which is the same number of patients which can be scanned with [⁶⁸Ga]Ga-PSMA-11 produced from 2 staggered ⁶⁸Ge/⁶⁸Ga generators. As more than 2 productions runs can potentially be performed with the target, the number of patients able to be scanned per day is potentially is increased.

Although activity yield is an important parameter to measure process performance, the quality of the cyclotron-produced [⁶⁸Ga]GaCl₃ is of even greater importance as high quality [⁶⁸Ga]GaCl₃ is critical to enable efficient labeling. Therefore, in addition to yield measurement and periodic quality control (QC) assessment, validation studies for [⁶⁸Ga]GaCl₃ and [⁶⁸Ga]Ga-PSMA-11 were carried out at UM and for [⁶⁸Ga]Ga-PSMA-11 at RPA. The test methods performed (half-life, radiochemical purity, pH, radionuclidic purity, metal analysis) were in accordance with the EUP monograph for [⁶⁸Ga]GaCl₃ (25), with an exception for Fe and Zn, for which semi-quantitative colorimetric test strips (e.g. EM-Quant, Merck) and/or ICP-MS were used.

Table 7 reports on the QC results for the [⁶⁸Ga]GaCl₃ validation runs carried out at UM. QC testing of [⁶⁸Ga]Ga-PSMA-11 is shown in Table 8 for UM and Table 9 for RPA, with RCP assessment by radio-TLC

and HPLC. Endotoxin, 4-hour stability (data not shown), and sterility testing were also performed for the three validation runs. The 3 validation runs and all 4-hour stability time points passed all QC parameters.

Table 7
Quality Control Data for three [⁶⁸Ga]GaCl₃ validation runs (UM)

TEST	1	2	3	Avg & SD	Release Criteria (EUP)
Radiochemical Purity [⁶⁸ Ga]GaCl ₃ (iTLC-SG)	99	98	98	98.3 ± 0.3	≥ 95
Rf [⁶⁸ Ga]GaCl ₃ (TLC)	< 0.2	< 0.2	< 0.2	< 0.2	≤ 0.2
Rf Ref B* (TLC)	> 0.7	> 0.7	> 0.7	> 0.7	≥ 0.7
pH	< 2	< 2	< 2	< 2	< 2
Visual Inspection	Passed	Passed	Passed	N/A	Clear, colorless, no visible particulate
Radionuclidic Identity (t _{1/2})	67.2	68.8	69.1	68.4 ± 0.8	64.6–71.4 min
Endotoxin Analysis	< 2	< 2	< 2	< 2	≤ 58.3 EU/mL
Fe µg/GBq	< 5	< 5	< 5	< 5	≤ 10 µg/GBq
Zn µg/GBq	< 1.25	< 1.25	< 1.25	< 1.25	≤ 10 µg/GBq
RNP at EOB (MCA)	99.8	99.8	99.8	99.8	≥ 98% (at time of use)
*Ref B-Pentetic acid solution					

Table 8
Quality Control Data for three [⁶⁸Ga]Ga-PSMA-11 validation runs (UM)

Tests	1	2	3	Avg & SD	Release Criteria (UM)
Radiochemical Purity (via TLC)	99.5	99.4	99.3	99.4 ± 0.1	≥ 90%
Relative Retention time (via HPLC)	1.004	1.005	1.005	1.0046 ± 0.0003	RRT: 0.9–1.1
pH	7.0	7.0	7.0	7.0	4.0–8.0
Visual Inspection	Passed	Passed	Passed	N/A	Clear, colorless, no visible particulate
Radionuclidic Identity (t _{1/2})	67.61	68.45	67.20	67.75 ± 0.53	64.6–71.4 min
Endotoxin Analysis	< 2	< 2	< 2	< 2	≤ 10.9 EU/mL
Bubble Point (PSI)	51	52	53	52 ± 1	≥ 50 PSI
Sterility	Passed	Passed	Passed	Passed	Complies with USP < 71 > (45)
RNP at EOB (MCA)	99.8	99.8	99.8	99.8	≥ 98% (at time of use)

Table 9
Quality Control Data for three [⁶⁸Ga]Ga-PSMA-11 validation runs (RPA)

Tests	1	2	3	Avg & SD	Release Criteria (RPA)
Radiochemical Purity (via TLC)	99.94	99.99	99.94	99.96 ± 0.03	≥ 95%
Radiochemical Purity (via HPLC)	99.94	99.97	100.0	99.97 ± 0.03	≥ 95%
pH	5.0	5.5	5.5	5.3 ± 0.3	4.0–8.0
Visual Inspection	Passed	Passed	Passed	N/A	Clear, colorless, no visible particulate
Radionuclidic Identity (t _{1/2})	67.9	68.1	67.7	67.9 ± 0.20	62–74 min
Endotoxin Analysis	< 1	< 1	< 1	< 1	≤ 17.5 EU/mL
Bubble Point (bar)	4.1	4.2	4.1	4.1 ± 0.06	≥ 3.5 bar
Sterility	Passed	Passed	Passed	Passed	Sterile – no growth
RNP at EOS (Well Counter)	99.7	99.8	99.8	99.8 ± 0.06	≥ 98% (at time of use)

The validation data shown here demonstrates the high quality of cyclotron-produced [⁶⁸Ga]GaCl₃ and [⁶⁸Ga]Ga-PSMA-11 and highlights the reliability and reproducibility of both processes. Notably, the reported RNP satisfied the proposed EU Pharmacopoeia limits (radioactivity maximum 2% of combined ⁶⁶Ga + ⁶⁷Ga) (25). The dosimetry of such a limit has been previously reported using worst-case assumptions, such as no biological clearance and rapid organ uptake (46). For this scenario, a relative dose increase up to 20% is reported but is typically less than 10% when compared with “pure” ⁶⁸Ga (i.e. not comparing with generator ⁶⁸Ga which may contain ⁶⁸Ge). Overall, the obtained results provided a solid basis for the clinical evaluation of cyclotron-produced [⁶⁸Ga]Ga-PSMA-11.

4.3 Clinical production and use of [⁶⁸Ga]Ga-PSMA-11

To date, over 700 patients have been scanned with [⁶⁸Ga]Ga-PSMA-11 at UM under IND. Initially, this was with generator-based [⁶⁸Ga]Ga-PSMA-11, though amended to include cyclotron-based [⁶⁸Ga]Ga-PSMA-11, with the first clinical production of cyclotron-based [⁶⁸Ga]Ga-PSMA-11 from a single FASTlab cassette in February 2019. As of March 2020, 50 clinical batches of cyclotron-produced [⁶⁸Ga]Ga-PSMA-11 were used to scan more than 90 patients (see Table 6) and the image from the first patient scanned with cyclotron-produced [⁶⁸Ga]Ga-PSMA-11 in Fig. 6). There were no differences noted in the quality of studies where ⁶⁸Ga was produced from a cyclotron when compared to a generator. At RPA, cyclotron-based ⁶⁸Ga used for clinical [⁶⁸Ga]Ga-PSMA-11 began in 2020.

5 Conclusions And Outlook

A process for isolating high purity [^{68}Ga]GaCl₃ from cyclotron-produced ^{68}Ga and subsequent labeling of PSMA-11 on the GE FASTlab synthesizer with both steps being performed on a single cassette has been developed. The cyclotron-based method offers a reliable source of ^{68}Ga and delivers consistently higher yields than currently available commercial 1.85 GBq (50 mCi) $^{68}\text{Ge}/^{68}\text{Ga}$ generators. Furthermore, in contrast to generators, for which ^{68}Ga activity falls over time due to ^{68}Ge decay, cyclotron-based ^{68}Ga activity is consistent with time thereby simplifying patient scheduling. The FASTlab-derived [^{68}Ga]GaCl₃ solution for radiolabeling met the requirements in the European Pharmacopeia with the purity of reagents and ^{68}Zn enrichment and purity used at these sites, and validation of [^{68}Ga]GaCl₃ and [^{68}Ga]Ga-PSMA-11 for clinical application has been demonstrated by the UM and RPA. The total manufacturing time is approximately 2 h and over 90 patients have been scanned using cyclotron-based [^{68}Ga]Ga-PSMA-11 to date. The process is in routine use to meet the growing demands for PSMA-based PET imaging at UM, and the first clinical studies have also been conducted at RPA. Additional studies to broaden the applicability of the [^{68}Ga]GaCl₃ process for labeling with other commonly used chelators such as DOTA have been performed successfully at RPA, and clinical use of [^{68}Ga]Ga-DOTATATE labeled with cyclotron-produced ^{68}Ga is ongoing. Similar studies are also currently ongoing at other sites.

Declarations

Acknowledgements

The authors thank Dr. Ali Afshar-Oromieh MD PhD, Prof. Dr. Klaus Kopka PhD, and their colleagues at the University Hospital of Heidelberg and the German Cancer Research Center (DKFZ) Heidelberg for the valuable assistance in qualifying the University of Michigan for clinical production of [^{68}Ga]Ga-PSMA-11. In addition, the authors thank Steffen Happel at Triskem for valuable suggestions, feedback, and initial resin samples.

Compliance with ethical standards

Conflict of interest KG, JF, DCP, and CS are employees of GE Healthcare. SE, AK, DS, MJF, MER, MC, BGH, BDH, MA-G, MRP and PJHS declare no conflict of interest.

Ethical Approval This article does not contain any original studies with human or animal subjects performed by any of the authors.

Funding

PJHS acknowledges financial support from Michigan Medicine.

References

- Afshar-Oromieh A, Zechmann CM, Malcher A, Eder M, Eisenhut M, Linhart HG, et al. Comparison of PET imaging with a ^{68}Ga -labelled PSMA ligand and ^{18}F -choline-based PET/CT for the diagnosis of recurrent prostate cancer. *Eur. J. Nucl. Med. Mol. Imaging*. 2014;41:11–20.
- Albisinni S, Artigas C, Aoun F, Biaou I, Grosman J, Gil T, et al. Clinical impact of ^{68}Ga -prostate-specific membrane antigen (PSMA) positron emission tomography/computed tomography (PET/CT) in patients with prostate cancer with rising prostate-specific antigen after treatment with curative intent: preliminary analysis. *BJU Int*. 2017;120:197–203.
- Alves F, Alves VHP, Do Carmo SJC, Neves ACB, Silva M, Abrunhosa AJ. Production of copper-64 and gallium-68 with a medical cyclotron using liquid targets. *Mod. Phys. Lett. A*. 2017;32:1740013.
- Alves V, do Carmo S, Alves F, Abrunhosa A. Automated Purification of Radiometals Produced by Liquid Targets. *Instruments*. 2018;2:17.
- Baum RP, Kulkarni HR. THERANOSTICS: From Molecular Imaging Using Ga-68 Labeled Tracers and PET/CT to Personalized Radionuclide Therapy - The Bad Berka Experience. *Theranostics*. 2012;2:437–47.
- Blower JE, Cooper MS, Imberti C, Ma MT, Marshall C, Young JD, et al. The Radiopharmaceutical Chemistry of the Radionuclides of Gallium and Indium. In: *Radiopharmaceutical Chemistry* by Lewis J, Windhorst A, Zeglis B (Eds). Springer; 2019. p. 255–71.
- Boschi A, Martini P, Costa V, Pagnoni A, Uccelli L, Santini C, et al. Interdisciplinary Tasks in the Cyclotron Production of Radiometals for Medical Applications. The Case of ^{47}Sc as Example. *Molecules*. 2019;24:1–14.
- Calais J, Fendler WP, Herrmann K, Eiber M, Ceci F. Comparison of ^{68}Ga -PSMA-11 and ^{18}F -Fluciclovine PET/CT in a case series of 10 patients with prostate cancer recurrence. *J. Nucl. Med*. 2018;59:789–94.
- Cutler CS, Minoshima S. Shortage of Germanium68/gallium68 generators in the United States [Internet]. 2018; Available from: <https://s3.amazonaws.com/rdcms-snm/production/public/Ga68%20shortage%20letter.pdf>. Accessed 12-Aug-2020.
- Eder M, Neels O, Müller M, Bauder-Wüst U, Remde Y, Schäfer M, et al. Novel Preclinical and Radiopharmaceutical Aspects of [^{68}Ga]Ga-PSMA-HBED-CC: A New PET Tracer for Imaging of Prostate Cancer. *Pharmaceuticals*. 2014;7:779–96.

- Engle JW, Lopez-Rodriguez V, Gaspar-Carcamo RE, Valdovinos HF, Valle-Gonzalez M, Trejo-Ballado F, et al. Very high specific activity $^{66/68}\text{Ga}$ from zinc targets for PET. *Appl. Radiat. Isot.* 2012;70:1792–6.
- Evans JD, Jethwa KR, Ost P, Williams S, Kwon ED, Lowe VJ, et al. Basic Original Report Prostate cancer-specific PET radiotracers: A review on the clinical utility in recurrent disease. *Pract. Radiat. Oncol.* 2018;8:28–39.
- Fendler WP, Rahbar K, Herrmann K, Kratochwil C, Eiber M. ^{177}Lu -PSMA Radioligand Therapy for Prostate Cancer. *J. Nucl. Med.* 2017;58:1196–200.
- Gallium (^{68}Ga) Chloride (accelerator produced) solution for radiolabelling. *European Pharmacopoeia.* 2020;10(Suppl. 10.3):4864-4865.
- International Atomic Energy Agency. Gallium-68 Cyclotron Production. IAEA-TECDOC-1863. IAEA, Vienna; 2019.
- Gourni E, Demmer O, Schottelius M, D'Alessandria C, Schulz S, Dijkgraaf I, et al. PET of CXCR4 expression by a ^{68}Ga -labeled highly specific targeted contrast agent. *J. Nucl. Med.* 2011;52:1803–10.
- Graham MM, Gu X, Ginader T, Breheny P, Sunderland JJ. Ga-DOTATOC Imaging of Neuroendocrine Tumors: A Systematic Review and Metaanalysis. *J. Nucl. Med.* 2017;58:1452–8.
- Graves SA, Engle JW, Eriksson TE, Gagnon K. Dosimetry of cyclotron-produced [^{68}Ga]Ga-PSMA-11, [^{68}Ga]Ga-DOTA-TATE, and [^{68}Ga]Ga-DOTA-TOC. *J. Nucl. Med.* 2018;59(Suppl. 1):1003.
- Herrmann K, Schottelius M, Lapa C, Osl T, Poschenrieder A, Hänscheid H, et al. First-in-human experience of CXCR4-directed endoradiotherapy with ^{177}Lu - and ^{90}Y -labeled pentixather in advanced-stage multiple myeloma with extensive intra- and extramedullary disease. *J. Nucl. Med.* 2016;57:248–51.
- Herrmann K, Schwaiger M, Lewis JS, Solomon SB, McNeil BJ, Baumann M, et al. Radiotheranostics: a roadmap for future development. *Lancet Oncol.* 2020;21:e146–56.
- Hope TA, Aggarwal R, Chee B, Tao D, Greene KL, Cooperberg MR, et al. Impact of ^{68}Ga -PSMA-11 PET on management in patients with biochemically recurrent prostate cancer. *J. Nucl. Med.* 2017;58:1956–61.
- Jackson IM, Scott PJH, Thompson S. Clinical Applications of Radiolabeled Peptides for PET. *Semin. Nucl. Med.* 2017;47:493-523.
- Jensen M, Clark JC. Direct production of Ga-68 from proton bombardment of concentrated aqueous solutions of [^{68}Zn] Zinc Chloride. In *The 13th International Workshop on Targetry and Target Chemistry Proceedings* by Horoun S, Givskov and Jensen M (Eds). Risø DTU, National Laboratory for Sustainable Energy, Technical University of Denmark. Roskilde. 2011; 288-292. Available online:

https://inis.iaea.org/collection/NCLCollectionStore/_Public/43/006/43006750.pdf. Accessed 12-Aug-2020.

Kratochwil C, Bruchertseifer F, Rathke H, Hohenfellner M, Giesel FL, Haberkorn U, et al. Targeted α -therapy of metastatic castration-resistant prostate cancer with ^{225}Ac -PSMA-617: Swimmer-Plot Analysis Suggests efficacy regarding duration of tumor control. *J. Nucl. Med.* 2018;59:795–802.

Kratochwil C, Flechsig P, Lindner T, Abderrahim L, Altmann A, Mier W, et al. ^{68}Ga -FAPI PET/CT: Tracer Uptake in 28 Different Kinds of Cancer. *J. Nucl. Med.* 2019;60:801–5.

Lenzo N, Meyrick D, Turner J, Lenzo NP, Meyrick D, Turner JH. Review of gallium-68 PSMA PET/CT imaging in the management of prostate cancer. *Diagnostics.*; 2018;8:16.

Lin M, Waligorski GJ, Lepera CG. Production of curie quantities of ^{68}Ga with a medical cyclotron via the $^{68}\text{Zn}(p,n)^{68}\text{Ga}$ reaction. *Appl. Radiat. Isot.* 2018;133:1–3.

Lowe VJ, Kwon ED. PET in Prostate Cancer: A Focus on C-11 Choline. *Pathways Clin. Trials Netw. Newsl.* 2015;5:1-4. [Internet] Available from: http://s3.amazonaws.com/rdcms-snm/production/public/images/Publications/Jan%202015%20Pathways_web.pdf. Accessed 12-Aug-2020.

Martiniova L, De Palatis L, Etchebehere E, Ravizzini G. Gallium-68 in Medical Imaging. *Curr. Radiopharm.* 2016;9:187–207.

McCormick B, Mahmoud A, Williams S, Davis J. Biochemical recurrence after radical prostatectomy: Current status of its use as a treatment endpoint and early management strategies. *Indian J. Urol.* 2019;35:6–17.

Mossine A V, Tanzey SS, Brooks AF, Makaravage KJ, Ichiishi N, Miller JM, et al. Synthesis of high-molar-activity [^{18}F]6-fluoro-L-DOPA suitable for human use via Cu-mediated fluorination of a BPin precursor. *Nat. Protoc.* 2020;15:1742–59.

Mueller D, Klette I, Baum RP, Gottschaldt M, Schultz MK, Breeman WAP. Simplified NaCl based ^{68}Ga concentration and labeling procedure for rapid synthesis of ^{68}Ga radiopharmaceuticals in high radiochemical purity. *Bioconjug. Chem.* 2012;23:1712–7.

Nair M, Happel S, Eriksson T, Pandey M, DeGrado T, Gagnon K. Cyclotron production and automated new 2-column processing of [^{68}Ga]GaCl₃. *Eur J Nucl Med Mol Imaging* . 2017;44(Suppl 2):S119–956.

Oehlke E, Hoehr C, Hou X, Hanemaayer V, Zeisler S, Adam MJ, et al. Production of Y-86 and other radiometals for research purposes using a solution target system. *Nucl. Med. Biol.* 2015;42:842–9.

- Pandey MK, Byrne JF, Jiang H, Packard AB, DeGrado TR. Cyclotron production of ^{68}Ga via the $^{68}\text{Zn}(p,n)^{68}\text{Ga}$ reaction in aqueous solution. *Am. J. Nucl. Med. Mol. Imaging*. 2014;4:303–10.
- Pandey MK, Byrne JF, Schlasner KN, Schmit NR, DeGrado TR. Cyclotron production of ^{68}Ga in a liquid target: Effects of solution composition and irradiation parameters. *Nucl. Med. Biol.* 2019;74/75:49-55.
- Pandey MK, DeGrado TR. Rapid Isolation of Cyclotron-Produced Gallium-68. 2019; WO2018039662.
- Prostate Cancer: Statistics [Internet]. 2020; Available from: <https://www.cancer.net/cancer-types/prostate-cancer/statistics>. Accessed 12-August-2020.
- Riga S, Cicoria G, Pancaldi D, Zagni F, Vichi S, Dassenno M, et al. Production of Ga-68 with a General Electric PETtrace cyclotron by liquid target. *Phys. Medica*. 2018;55:116–26.
- Roach PJ, Francis R, Emmett L, Hsiao E, Kneebone A, Hraby G, et al. The impact of ^{68}Ga -PSMA PET/CT on management intent in prostate cancer: Results of an Australian prospective multicenter study. *J. Nucl. Med.* 2018;59:82–8.
- Ruangma A, Kijprayoon S, Ngokpol S. PSMA for PET imaging of prostate cancer. *Bangkok Med. J.* 2018;14:95–100.
- Sadeghi M, Kakavand T, Rajabifar S, Mokhtari L, Rahimi-Nezhad A. Cyclotron production of ^{68}Ga via proton-induced reaction on ^{68}Zn target. *Nukleonika*. 2009;54:25–8.
- Schweinsberg C, Johayem A, Llamazares A, Gagnon K. The first curie-quantity production of [^{68}Ga]Ga-PSMA-HBED-CC. *J. Label. Compd. Radiopharm.* 2019;62(S1):P121.
- Schwenck J, Rempp H, Reischl G, Kruck S, Stenzl A, Nikolaou K, et al. Comparison of ^{68}Ga -labelled PSMA-11 and ^{11}C -choline in the detection of prostate cancer metastases by PET/CT. *Eur. J. Nucl. Med. Mol. Imaging*. 2017;44:92–101.
- Shao X, Hoareau R, Runkle AC, Tluczek LJM, Hockley BG, Henderson BD, et al. Highlighting the versatility of the Tracerlab synthesis modules. Part 2: fully automated production of [^{11}C]-labeled radiopharmaceuticals using a Tracerlab FX_{C-PRO}. *J. Label. Compd. Radiopharm.* 2011;54:819–38.
- Shao X, Hockley BG, El Hoareau R, Schnau PL, Scott PJH. Fully automated preparation of [^{11}C]choline and [^{18}F]fluoromethylcholine using TracerLab synthesis modules and facilitated quality control using analytical HPLC. *Appl. Radiat. Isot.* 2010;69:403–9.
- Smith DL, Breeman WAP, Sims-Mourtada J. The untapped potential of Gallium 68-PET: The next wave of ^{68}Ga -agents. *Appl. Radiat. Isot.* Pergamon; 2013;76:14–23.

Tolmachev V, Lundqvist H. Rapid separation of gallium from zinc targets by thermal diffusion. Appl. Radiat. Isot. 1996;47:297–9.

Velikyan I. Prospective of ^{68}Ga Radionuclide Contribution to the Development of Imaging Agents for Infection and Inflammation. Contrast Media Mol. Imaging. Hindawi; 2018;2018:1–24.

Zeisler S, Limoges A, Kumlin J, Siikanen J, Hoehr C. Fused Zinc Target for the Production of Gallium Radioisotopes. Instruments. 2019;3:10.

USP 71 Microbiological Tests/Sterility Tests. United States Pharmacopeial Conv. 2012.

Figures



Figure 1

GE Gallium-68 Liquid Target




Figure 2

Activity receiving vial for collection of the irradiated ⁶⁸Zn solution




Figure 3

Two-column approach for ⁶⁸Ga chemical separation




Figure 4

FASTlab cassette layout – applicable to Schemes “A&B” of Table 3.




Figure 5

Partitioning of FASTlab cassette: Right-hand side is reserved for $[^{68}\text{Ga}]\text{GaCl}_3$ purification, and the left-hand side accommodates the $[^{68}\text{Ga}]\text{Ga-PSMA-11}$ synthesis and C18 cartridge based purification.




Figure 6

Images from the first patient scanned with $[^{68}\text{Ga}]\text{Ga-PSMA-11}$ labeled with cyclotron produced ^{68}Ga at the University of Michigan