Safety and Risk Factors of Tinavi Robot-assisted Percutaneous Pedicle Screw Placement in Spinal Surgery

Ren-Jie Zhang

First Affiliated Hospital of Anhui Medical University

Lu-Ping Zhou

First Affiliated Hospital of Anhui Medical University

Lai Zhang

First Affiliated Hospital of Anhui Medical University

Hua-Qing Zhang

First Affiliated Hospital of Anhui Medical University

Jian-Xiang Zhang

First Affiliated Hospital of Anhui Medical University

Cailiang Shen (✉ shencailiang@ahmu.edu.cn)

First Affiliated Hospital of Anhui Medical University <https://orcid.org/0000-0002-9835-6384>

Research article

Keywords: robot, pedicle screw placement, intra-pedicular accuracy, proximal facet joint, risk factor, spine surgery

Posted Date: April 28th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-451919/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Abstract

Objective To determine the rates and risk factors of pedicle screw placement accuracy and the proximal facet joint violation (FJV) using TINAVI robot-assisted technique.

Methods Patients with thoracolumbar fractures or degenerative diseases were retrospectively recruited from June 2018 and June 2020. The pedicle penetration and proximal FJV were compared in different instrumental levels to identify the safe and risk segments during insertion. Moreover, the factors were also assessed using univariate and multivariate analyses.

Results A total of 72 patients with 332 pedicle screws were included in the current study. The optimal and clinically acceptable screw positions were 85.8% and 93.4%. Of the 332 screws concerning the intra-pedicular accuracy, 285 screws (85.8%) were evaluated as grade A according to the Gertzbein and Robbins scale, with the remaining 25 (7.6%), 10 (3.0%), 6 (1.8%), and 6 screws (1.8%) as grade B, C, D, and E. Moreover, in terms of the proximal FJV, 255 screws (76.8%) screws were assessed as grade 0 according to the Babu scale, with the remaining 34 (10.3%), 22 (6.6%), and 21 screws (6.3%) as grade 1, 2, and 3. Furthermore, the univariate analysis showed significantly higher rate of penetration for patients with age ≥ 61 years old, sex of female, thoracolumbar insertion, shorter distance from skin to insertion point, and smaller facet angle. Meanwhile, the patients with the sex of female, BMI < 25.9 , grade I spondylolisthesis, lumbosacral insertion, longer distance from skin to insertion point, and larger facet angle had a significantly higher rate of proximal FJV. The outcomes of multivariate analyses showed that sex of male (adjusted OR 0.320, 95% CI 0.140–0.732; $p = 0.007$), facet angle $\geq 45^\circ$ (adjusted OR 0.266, 95% CI 0.090–0.786; $p = 0.017$), distance from skin to insertion point $\geq 4.5\text{cm}$ (adjusted OR 0.342, 95% CI 0.134–0.868; $p = 0.024$), and lumbosacral instrumentation (adjusted OR 0.227, 95% CI 0.091–0.566; $p = 0.001$) were independently associated with intra-pedicular accuracy; the L5 insertion (adjusted OR 2.020, 95% CI 1.084–3.766; $p = 0.027$) and facet angle $\geq 45^\circ$ (adjusted OR 1.839, 95% CI 1.026–3.298; $p = 0.041$) were independently associated with the proximal FJV.

Conclusion TINAVI robot-assisted technique was associated with a high rate of pedicle screw placement and a low rate of proximal FJV. This new technique showed a safe and precise performance for pedicle screw placement in spinal surgery. Facet angle $\geq 45^\circ$ is independently associated with both the intra-pedicular accuracy and proximal FJV.

Introduction

The pedicle screw placement has been widely applied in the treatment of spinal fractures, tumor, infections, deformities, and degenerative diseases for the improvement of stability after the spinal reconstruction. However, due to the spinal cord, nerve root and blood vessels adjacent to the pedicle, the malposition of screws might contribute to severe complications, the rate of which was reported as 4.2% of patients.[1] Besides, the proximal facet joint violation (FJV), one of the most common complications of malposition, has been regarded as a crucial risk factor for adjacent segment degeneration (ASD) after

surgery.[2, 3] Moreover, Levin et al.[4] reported that proximal FJV was independently associated with a higher reoperation rate and diminished improvement in quality of life.

Initially, the pedicle screws were placed in open methods with inevitable need for paravertebral muscle dissection resulting in excessive bleeding and trauma of soft tissues. With the wide acceptance of minimally invasive concepts, various insertion techniques, such as the percutaneous fluoroscopy-guided, patient-specific template-guided, the computer-assisted, and robot-assisted (RA) navigation techniques, have been designed to guide the pedicle screw placement for the enhancement of insertion accuracy, protection of proximal facet joint, and reduction of severe complication including neurological, vascular, and muscular injuries.[5, 6] Among them, the widely used robots of SpineAssist, Renaissance, and Mazor X (Mazor Robotics Ltd., Caesarea, Israel) were reported to enable less damage to soft tissues, higher intra-pedicular accuracy of 85.0-98.2%[7-13], and lower rates of proximal FJV of 0-5%[9, 13, 14] with intraoperative navigation, automatic trajectory determination, and reduced manual errors[15]. With the clinical application of orthopedic robots, spine surgery has entered the era of intelligence.

Recent years, a new robot system named TiRobot (TINAVI Medical Technologies Co. Ltd.) has been introduced in orthopedic surgery, which received China Food and Drug Administration approval in 2016. [16] A series of studies conducted by Tian et al. [16-19] reported that the TiRobot system led to higher intra-pedicular accuracy rates of 93.2-98.2%[16-18] and lower proximal FJV rates of 0-13.0%[16, 18, 19] in a single trial center. However, the accuracy of RA pedicle screw placement was unstable and controversial. Meanwhile, the studies regarding the intra-pedicular and proximal facet joint accuracy of TiRobot system are still rare, and the risk factors of the pedicle screw placement accuracy and proximal FJV with TiRobot system need further identification. Thus, this study was performed to evaluate the radiographic outcomes of intra-pedicular accuracy and proximal FJV in TINAVI RA technique, and analyze the risk factors affecting the inaccuracy of pedicle screw placement and proximal FJV to provide references for surgeons.

Method

Study design and patient selection

A consecutive series of patients who underwent TINAVI RA percutaneous pedicle screw placement for the treatment of thoracolumbar fractures or the degenerative diseases between June 2018 and June 2020, were respectively recruited, and all the operations were conducted by a surgical team. The study was approved by hospital's institutional review board with all the included patients signing informed consent forms.

The inclusion criteria were as follows: (1) patients older than 18 years old; (2) patients diagnosed of thoracolumbar fractures or degenerative diseases including lumbar disc herniation, spinal stenosis, grade I spondylolisthesis, and mild degenerative scoliosis; (3) patients undergoing TINAVI RA percutaneous insertion, regardless of purely instrumental fixation or decompression along with intervertebral fusion; (4) Insertion segments ranging from T11 to S1. (5) Patient receiving postoperative three-dimensional (3D) CT

examination. The exclusion criteria consisted of the following: (1) patients with a history of previous surgery on the thoracolumbar segments, and revision surgery in the same segments; (2) patients with destruction of the vertebral body caused by tuberculosis infection or tumor, and unclear anatomy of the surgical segments; (3) patients with active local or systematic infection, spinal tumor, or the tuberculosis.

Patients information on age, gender, body mass index (BMI), and diagnosis before surgery were obtained from medical records with the consent from patients.

Robot-assisted percutaneous insertion technique

The percutaneous pedicle screw placements were performed with the TiRobot system composed of three basic elements: an optical tracking device, a workstation for preoperative planning and intraoperative controlling, and a robotic arm.[16]

After general anesthesia with endotracheal intubation, the patient took the prone position. Then the position plate was located and the projection of the planned vertical body pedicle on skin was marked. The pedicle screws were inserted in according with the original method under the guidance of TiRobot system.[16]

As for the patients suffering from fractures, the pedicle screw was inserted directly along the guide wire, and the incision was closed after the fracture rehabilitation performed with the screw rod. In contrast, regarding the patients suffering from the degenerative diseases, the screw is placed on the non-decompression side after the insertion of guide wire under the robotic guidance. Whereas the trajectory of screw is prepared along the guide wire on the decompression side, and then the screw is placed after transforaminal lumbar interbody fusion (TLIF).[20]

Outcome measures

The accuracy of pedicle screw placement was assessed in accordance with the Gertzbein and Robbins scale[21] with the following grades: A, the screw completely contained in the pedicle; B, the cortical penetration of less than 2 mm; C, the cortical penetration of 2 mm or more but less than 4 mm; D, the cortical penetration of 4 mm or more but less than 6 mm; E, the cortical penetration of 6 mm or more. **(Figure 1)** Grades A was regarded as optimal screw position, Grades A and B indicated clinically acceptable screw position, whereas Grades C, D, and E were considered as misplacement. Besides, the evaluation of proximal FJV was according to the Babu scale[22] with the following grades: 0, screw not in facet; 1, screw in lateral facet but not in facet articulation; 2, penetration of facet articulation less than 1 mm; 3, penetration of facet articulation more than 1 mm or traveling within facet articulation. **(Figure 2)** Moreover, the pedicle penetration and proximal FJV were also compared in different instrumental levels to identify the safe and risk segments during insertion.

The following factors that possibly affected the accuracy of the pedicle screw placement and proximal FJV were also collected during the assessment of the postoperative CT scan: proximal facet joint degeneration, facet angle, axial pedicle angle, sagittal pedicle angle, distance from skin to insertion point.

The proximal facet joint degeneration was measured according to the Pathria classification[23]. The radiographic data were independently evaluated by a radiologist and a surgeon on the postoperative CT scans with a 1-mm thickness slice. If the two evaluators were inconsistent on the results, the final decision was made by the third surgeon.

Statistical analysis

The statistical analyses were performed with SPSS version 23.0. The continuous variables were described as the mean and standard deviation and the categorical variables as absolute (no.) and relative (%) frequencies. For statistical purpose, the rates and different grades of the accuracy of pedicle insertion and violation of proximal facet joint were analyzed in each instrumented level. Moreover, possible factors associated with intra-pedicular and proximal facet joint accuracy were gender, age, BMI, side of screw, diagnosis of grade I spondylolisthesis and mild degenerative scoliosis, facet degeneration, instrumented levels (thoracolumbar or lumbosacral instruments), depth of surgical field (distance of entry point from skin), facet angle, axial pedicle angle, and sagittal angle. After the univariate analyses of the above factors, the remaining issues with a significant level < 0.05 in univariate analyses were summarized to the final logistics regression modal of multivariate analyses. As for the intra-pedicular accuracy, the variables included in the multivariate analyses were gender, age, instrumented levels, depth of surgical field, and facet angle. While for the proximal FJV, the variables were gender, BMI, diagnosis of grade I spondylolisthesis, instrumented levels, depth of surgical field, and facet angle. The level of significance was defined as $\alpha = 0.05$.

Result

A total of 72 patients with 332 pedicle screws were included in the study. The demographic characteristics and screws-related data were summarized in **Table 1** and **2**, respectively. The overall patient mean age of the 32 females and 40 males was 50.17 ± 10.93 years old (range 28-72) and the mean value of the BMI was 23.54 ± 3.98 (range 18.3-36.7). The diagnoses were thoracolumbar fracture in 34 cases, lumbar disc hernia in 24 cases, lumbar canal stenosis in 2 cases, grade I spondylolisthesis in 7 cases, and mild degenerative scoliosis in 5 cases. Moreover, of the 322 screws insertions, the most frequently inserted level was L5 (70 screws) followed by L4 (60 screws), L1 (54 screws), and T12 (44 screws). All the screws were inserted in bilateral sides. Meanwhile, 188 proximal facet joints (56.6%) were regarded as mild osteoarthritis. The mean values of axial and sagittal pedicle angles were $14.02 \pm 4.92^\circ$ and $2.98 \pm 6.63^\circ$, respectively. Furthermore, the mean value distance from skin to insertion point was 5.35 ± 0.90 cm.

Of the 332 screws concerning the pedicle screw accuracy, 285 screws (85.8%) were evaluated as grade A according to the Gertzbein and Robbins scale, with the remaining 25 (7.6%), 10 (3.0%), 6 (1.8%), and 6 screws (1.8%) as grade B, C, D, and E, respectively. The optimal and clinically acceptable screw positions were 85.8% and 93.4%. The cortical encroachments in pedicles were found for 47 screws. Among the penetrations, the lateral cortex was the most frequently breached (74.5%), followed by the medial (21.2%)

and inferior (4.3%) cortexes. Meanwhile, the risk of pedicle penetration in L5 level was significantly lower to 0.222 times than other segments ($p = 0.013$).

In terms of the proximal FJV, 255 screws (76.8%) screws were assessed as grade 0 according to the Babu scale, with the remaining 34 (10.3%), 22 (6.6%), and 21 screws (6.3%) as grade 1, 2, and 3. The risk of proximal FJV in T12 insertion was 0.295 times than other levels ($p = 0.029$), whereas L5 insertion was associated with significantly higher risk of 3.153 times than others ($p \leq 0.001$). Besides, the screw head was the most frequent region responsible for proximal FJV (80.5%), followed by screw shaft (19.5%). There is no violation from the rod. The rates of intra-pedicular and proximal facet joint accuracy, and the related raw ORs in different instrumented levels were shown in **Table 3** and **4**.

The summary and characteristics of adverse events were displayed in **Supplementary file 1**. Regarding the complications, no patients underwent intraoperative revision due to screw malposition or postoperative revision due to screw malposition, and no neurological symptoms resulted from screw insertion. Furthermore, one patient suffered from wound infection after surgery with no need for another debridement surgery.

Factors that affected the intra-pedicular and proximal facet joint accuracy of robotic guidance percutaneous approach were explored with univariate analyses (**Supplementary file 2**). Regarding the intra-pedicular accuracy, the patients with the age ≥ 61 years old, sex of female, thoracolumbar insertion, shorter distance from skin to insertion point, and smaller facet angle had a significantly higher rate of penetration. As for the proximal FJV, the patients with the sex of female, BMI < 25.9 , grade I spondylolisthesis, lumbosacral insertion, longer distance from skin to insertion point, and larger facet angle had a significantly higher rate.

After the control of confounders, the multivariate analyses of associated with the intra-pedicular and proximal facet joint accuracy were performed concerning the issues in univariate analyses with a significant level < 0.05 (**Table 5**). The result showed that the sex of male (adjusted OR 0.320, 95% CI 0.140–0.732; $p = 0.007$), facet angle $\geq 45^\circ$ (adjusted OR 0.266, 95% CI 0.090–0.786; $p = 0.017$), distance from skin to insertion point $\geq 4.5\text{cm}$ (adjusted OR 0.342, 95% CI 0.134–0.868; $p = 0.024$), and lumbosacral instrumentation (adjusted OR 0.227, 95% CI 0.091–0.566; $p = 0.001$) were independently associated with intra-pedicular accuracy. Meanwhile, the L5 insertion (adjusted OR 2.020, 95% CI 1.084–3.766; $p = 0.027$) and facet angle $\geq 45^\circ$ (adjusted OR 1.839, 95% CI 1.026–3.298; $p = 0.041$) were independently associated with proximal FJV.

Discussion

The accuracy of pedicle screw insertion is especially crucial because of the severe consequences followed by malposition including neurological and vascular injuries, revision surgery, and spinal instability.[13, 16] Meanwhile, the proximal FJV has been regarded as an independent risk factor for ASD, finally resulting in higher reoperation rate and diminished improvement in quality of life.[3, 4] Thus, the safety of pedicle screw placement has increasingly drawn attention from spinal surgeons. With the

application of robotic guidance systems, higher rate of intra-pedicular accuracy and lower rate of proximal FJV in RA technique than those in percutaneous fluoroscopy-guided and freehand techniques have been reported in recent years.[5, 7-9, 13, 14, 16-19] However, the rates of intra-pedicular accuracy and proximal FJV were quite different in previously published studies, and compared with the Mazor RA techniques, the studies on TINAVI systems are rare. Furthermore, the risk factors that affected the outcomes of pedicle screw placement remain controversial and no consensus has been reached in clinical research. Here, we evaluated the safety of pedicle screw placement with TINAVI RA techniques in terms of the intra-pedicular accuracy and proximal FJV, and analyzed the risk factors associated with the two issues in screw implantation.

Many studies have concluded that RA technique showed improvement of intra-pedicular accuracy at the rates of 86.16-98.2% than conventional techniques. [8, 16-18, 24, 25] However, the previously published studies also showed that the RA technique performed no clear advantage in terms of intra-pedicular accuracy rates of 78.8-93.7% over the conventional techniques. [5, 9, 13, 26] Hyun et al.[9] conducted a RCT of Mazor robotic-guided versus freehand surgery, and found that all screws were optimal positions in RA group and 98.6% screws were graded as optimal positions, showing no remarkable difference. Kim et al.[13] also concluded that the intra-pedicular rate of 93.7% in the Mazor robotic-guided group was significantly similar to that of 91.9% in the freehand group. In the current study, we found that the optimal (grade A) and clinically acceptable (grade A+B) screw position were 85.8% and 93.4%, respectively, indicating promising results. Moreover, the main direction of screw deviations in pedicle was lateral, followed by medial and inferior breaches. The results might be explained as follows. First, based on the intraoperative 3D images where the anatomical structures, especially the abnormal and degenerative parts were clearly displayed, the RA system could automatically identify the optimal entry point and trajectory direction for the safety of the pedicle cortex. Then the implantation procedure was automatically performed according to the preplanned satisfactory trajectory in order to ensure the intra-pedicular accuracy. Second, the robot can repeat complex tasks precisely, thereby avoiding the limitation of human manual errors. Third, RA technique can minimize the resistance from the soft tissues around the spine compared to freehand method. After the guidewire was inserted under the guidance of a robotic arm, the screws were placed through the guiding tube into the pedicle. In contrast, the selection of entry point and the adjustment of trajectory would be biased due to the pulling and blocking of the paravertebral muscles in freehand surgery, leading to higher risk of screw malposition. Fourth, the RA insertion has a short learning curve because of the automatic procedures and the precision of implantation. Previous studies reported that the surgical outcomes from the initial experience to practiced experience showed no remarkable difference.[27, 28] Thus, surgical skills are relatively less demanding, thereby indicating that RA technique may prevent malposition by reducing unnecessary manual mistakes. Fifth, due to the precise insertion of RA technique, the rate of medial penetrations was greatly reduced, avoiding the relevant severe complications, such as spinal cord and nerve root injuries.

The importance of proximal facet joint protection has been increasingly realized in recent years and the violation might lead to the spinal instability in adjacent segments and reduce the load-bearing ability, eventually resulting in the radiographic ASD.[3, 29] The studies regarding the superiority of RA versus

conventional freehand techniques remained controversial. On the one hand, Some studies revealed that the RA pedicle screw placement was associated with significantly fewer FJV (0.00%-2.84%) than the conventional trajectory freehand implantation (15.85-100.00%).[2, 13, 14, 16, 18, 30-33] On the other hand, in a prospective RCT conducted by Hyun et al.[9], no remarkably statistical difference was noted regarding the rates of proximal FJV between the RA and freehand techniques (0.00% vs 0.71%). In a similar study, Archavlis et al.[14] also reported that the rate of proximal FJV in the RA group was similar to that in the freehand group (5% vs 6%). In the present study, we found that the proximal FJV was 23.2%, which was lower than that in freehand method but relatively higher than that in RA method from most of the previous published studies. Moreover, majority of the proximal FJV resulted from the screw head, followed by the screw shaft. The results might be attributed to the fact that the reduction of proximal FJV might result from the precision of pedicle screw placement. With the 3D images, the anatomical landmarks on the vertebra with hypoplastic or degenerative diseases can be displayed clearly with RA method. The RA system can identify the optimal entry point and trajectory and insert screws more accurately for the safety of facet joint by enlarging the distance between the facet and screws. Moreover, the advantages of muscle resistance and the short learning curve via RA method are also beneficial to facet protection. Furthermore, we found that the proximal FJV in the current study was higher than other trials. The results might be interpreted as that majority of pedicle screws were inserted in thoracolumbar levels with the patients diagnosed of vertebral fractures. However, we concluded in the current study that the FJVs in the thoracolumbar levels were remarkably higher than those in the lumbosacral levels due to the thinner diameter of pedicle in the thoracolumbar segments. Thus, the overall rate of FJV may be affected by the higher rate of thoracolumbar FJV.

The factors that affected the intra-pedicular accuracy and proximal FJV were also explored with univariate and multivariate analyses. As for the pedicle screw placement accuracy, our multivariate regression analysis demonstrated that lumbosacral instrumentation was 0.227 times likely to breach the cortex because of a larger diameter of pedicle in the lower lumbar segments. Thus, the occurrence of screw deviation causing severe penetration in lumbosacral segments might decrease than that in thoracolumbar segments. Moreover, we also found that the facet angle $\geq 45^\circ$ and the depth of surgical field $\geq 4.5\text{cm}$ are the protective factors in pedicle screw placement, which might result from the fact that in the lumbosacral levels, the facet angle was larger and the mean distance from the entry point to the skin was further than those in the thoracolumbar levels. Because of the higher rate of accuracy in lumbosacral instrumentation,

The distance from entry point to screw $\geq 4.5\text{cm}$ was less likely to breach the cortex. Moreover, we also found that the male patients suffered from less screw penetration.

In terms of the independent factors that affected the cranial FJV, our multivariate analysis showed that the L5 insertion was 2.02 times more likely to breach the facet. Tian et al.[34] identified an increasing trend of proximal FJV rates ranging from 1.5%, 7.4%, 17.6%, 30.9% to 42.6% with the levels of vertebrae dropping from L1 to L5. Besides, Moshirfar et al.[35] inserted the screws with freehand method, and it was found that the proximal FJV in the L5 level (48%) was significantly higher than that of other levels. In

L5 segment, the ideal entry point was deviated to the outside with the increasement of valgus angle. Meanwhile, the articular surface of the facet joint also shifted to the lateral and coronal planes, blocking the insertion trajectory and ultimately increasing proximal FJV.[22, 31] Moreover, patients with facet angle $\geq 45^\circ$ were 1.89 times more likely have a proximal FJV than patients with facet angle $< 45^\circ$. When the facet angle increases, the surface of facet joint gradually deviates to the coronal position, thereby blocking the trajectory of screws.[29] Another explanation might be that the facet angle in the lower lumbar is larger than that in the upper lumbar, and the lumbar lordosis is mainly manifested in the lower lumbar. Thus, there is more overlap between the facet joints and the pedicle direction, which also blocks the pathway of screws.[19]

There are some potential limitations in the current study. First, no control group was included in the study and only the data of RA technique was reported without comparative results. Although we compared the outcomes with those from previously published literatures, yet the higher biases appeared. Second, the overall intra-pedicular accuracy and proximal FJV consisted of thoracic and lumbar segments. The results may be affected due to the different anatomical characteristics. Third, we only reported the radiographic results in a short-term follow-up period. Nevertheless, the postoperative short-term outcomes are crucial for a novel technique regarding the accuracy of pedicle screw placement. Continued assessment of safety and accuracy of screw placement in long-term follow-up using TINAVI RA technique is needed in the future work.

Conclusion

TINAVI robot-assisted technique was associated with a high rate of pedicle screw placement and a low rate of proximal FJV. This technique showed a safe and precise performance for pedicle screw placement in spinal surgery.

Abbreviations

ASD adjacent segment degeneration; FJV proximal facet joint violation; RA robot-assisted; BMI body mass index; TLIF transforaminal lumbar interbody fusion.

Declarations

Disclosures

Conflict of interest The authors declare that they have no conflict of interest concerning the materials or methods used in this study or the findings specified in this paper.

Conflicts of Interest and Source of Funding

None of the authors have any conflict of interest.

Availability of data and materials

All data generated or analyzed during this study are included in this published article and its supplementary information files.

Ethics approval and consent to participate

Ethical approval was provided by the ethics committees of the First Affiliated Hospital of Anhui Medical University Ethics Committee.

Consent for publication

Not applicable.

Authors' contributions

Ren-Jie Zhang: Conceptualization, Data curation, Formal analysis, Investigation, Methodology, Software, Validation, Visualization, Writing - original draft, Writing - editing. **Lu-Ping Zhou:** Conceptualization, Data curation, Formal analysis, Investigation, Methodology, Software, Validation, Visualization, Writing - review. **Lai Zhang:** Data curation, Formal analysis, Investigation, Methodology, Software, Writing – editing. **Hua-Qing Zhang:** Data curation, Formal analysis, Investigation, Methodology, Software, Writing – editing. **Jian-Xiang Zhang:** Formal analysis, Software. **Cai-Liang Shen:** Conceptualization, Project administration, Investigation, Methodology, Supervision, Writing - review.

References

1. Hicks JM, Singla A, Shen FH, Arlet V: **Complications of pedicle screw fixation in scoliosis surgery: a systematic review.** *Spine (Phila Pa 1976)* 2010, **35**(11):E465-470.
2. Patel JY, Kundnani VG, Merchant ZI, Jain S, Kire N: **Superior facet joint violations in single level minimally invasive and open transforaminal lumbar interbody fusion: A comparative study.** *Asian Spine J* 2019, **2019 Oct 4**[Online ahead of print].
3. Bagheri SR, Alimohammadi E, Froushani AZ, Abdi A: **Adjacent segment disease after posterior lumbar instrumentation surgery for degenerative disease: Incidence and risk factors.** *Journal of Orthopaedic Surgery* 2019, **27**(2):1-6.
4. Levin JM, Alentado VJ, Healy AT, Steinmetz MP, Benzel EC, Mroz TE: **Superior segment facet joint violation during instrumented lumbar fusion is associated with higher reoperation rates and diminished improvement in quality of life.** *Clinical spine surgery* 2018, **31**(1):E36-E41.
5. Fan Y, Du J, Zhang J, Liu S, Xue X, Huang Y, Zhang J, Hao D: **Comparison of accuracy of pedicle screw insertion among 4 guided technologies in spine surgery.** *Med Sci Monit* 2017, **23**:5960-5968.
6. Cawley D, Dhokia R, Sales J, Darwish N, Molloy S: **Ten techniques for improving navigated spinal surgery.** *Bone Joint J* 2020, **102-B**(3):371-375.

7. Yang JS, He B, Tian F, Liu TJ, Liu P, Zhang JN, Liu SC, Tuo Y, Chu L, Hao DJ: **Accuracy of Robot-Assisted Percutaneous Pedicle Screw Placement for Treatment of Lumbar Spondylolisthesis: A Comparative Cohort Study.** *Med Sci Monit* 2019, **25**:2479-2487.
8. Ringel F, Stüer C, Reinke A, Preuss A, Behr M, Auer F, Stoffel M, Meyer B: **Accuracy of Robot-Assisted Placement of Lumbar and Sacral Pedicle Screws.** *Spine (Phila Pa 1976)* 2012, **37**(8):E496-501.
9. Hyun S-J, Kim K-J, Jahng T-A, Kim H-J: **Minimally invasive robotic versus open fluoroscopic-guided spinal instrumented fusions: a randomized controlled trial.** *Spine (Phila Pa 1976)* 2017, **42**(6):353-358.
10. Wallace DJ, Vardiman AB, Booher GA, Crawford NR, Rigglesman JR, Greeley SL, Ledonio CG: **Navigated robotic assistance improves pedicle screw accuracy in minimally invasive surgery of the lumbosacral spine: 600 pedicle screws in a single institution.** *The international journal of medical robotics + computer assisted surgery* 2020, **16**(1):e2054.
11. Vardiman AB, Wallace DJ, Crawford NR, Rigglesman JR, Ahrendtsen LA, Ledonio CG: **Pedicle screw accuracy in clinical utilization of minimally invasive navigated robot-assisted spine surgery.** *J Robot Surg* 2019.
12. Zhang JN, Fan Y, Hao DJ: **Risk factors for robot-assisted spinal pedicle screw malposition.** *Sci Rep* 2019, **9**(1):3025.
13. Kim HJ, Jung WI, Chang BS, Lee CK, Kang KT, Yeom JS: **A prospective, randomized, controlled trial of robot-assisted vs freehand pedicle screw fixation in spine surgery.** *The international journal of medical robotics + computer assisted surgery* 2017, **13**(3).
14. Archavlis E, Amr N, Kantelhardt SR, Giese A: **Rates of upper facet joint violation in minimally invasive percutaneous and open instrumentation: A comparative cohort study of different insertion techniques.** *Journal of Neurological Surgery, Part A: Central European Neurosurgery* 2018, **79**(1):1-8.
15. D'Souza M, Gendreau J, Feng A, Kim LH, Ho AL, Veeravagu A: **Robotic-assisted spine surgery: History, efficacy, cost, and future trends.** *Robot Surg* 2019, **6**:9-23.
16. Han X, Tian W, Liu Y, Liu B, He D, Sun Y, Han X, Fan M, Zhao J, Xu Y *et al*: **Safety and accuracy of robot-assisted versus fluoroscopy-assisted pedicle screw insertion in thoracolumbar spinal surgery: a prospective randomized controlled trial.** *Journal of neurosurgery Spine* 2019, **8**:1-8.
17. Feng S, Tian W, Wei Y: **Clinical Effects of Oblique Lateral Interbody Fusion by Conventional Open versus Percutaneous Robot-Assisted Minimally Invasive Pedicle Screw Placement in Elderly Patients.** *Orthopaedic Surgery* 2020, **12**(1):86-93.
18. Zhang Q, Han X-G, Xu Y-F, Liu Y-J, Liu B, He D, Sun Y-Q, Tian W: **Robot-assisted versus fluoroscopy-guided pedicle screw placement in transforaminal lumbar interbody fusion for lumbar degenerative disease.** *World Neurosurgery* 2019, **125**:E429-E434.
19. Le X-F, Shi Z, Wang Q-L, Xu Y-F, Zhao J-W, Tian W: **Rate and risk factors of superior facet joint violation during cortical bone trajectory screw placement: A comparison of robot-assisted approach with a conventional technique.** *Orthop Surg* 2020, **12**(1):133-140.

20. Harms J, Rolinger H: **A one-stager procedure in operative treatment of spondylolistheses: dorsal traction-reposition and anterior fusion (author's transl).** *Z Orthop Ihre Grenzgeb* 1982, **120**(3):343-347.
21. Gertzbein SD, Robbins SE: **Accuracy of pedicular screw placement in vivo.** *Spine (Phila Pa 1976)* 1990, **15**(1):11-14.
22. Babu R, Park JG, Mehta AI, Shan T, Grossi PM, Brown CR, Richardson WJ, Isaacs RE, Bagley CA, Kuchibhatla M *et al*: **Comparison of superior-level facet joint violations during open and percutaneous pedicle screw placement.** *Neurosurgery* 2012, **71**(5):962-970.
23. Pathria M, Sartoris DJ, Resnick D: **Osteoarthritis of the facet joints: accuracy of oblique radiographic assessment.** 1987, **164**:227-230.
24. Yang J-S, He B, Tian F, Liu T-J, Liu P, Zhang J-N, Liu S-C, Tuo Y, Chu L, Hao D-J: **Accuracy of robot-assisted percutaneous pedicle screw placement for treatment of lumbar spondylolisthesis: A comparative cohort study.** *Med Sci Monit* 2019, **25**:2479-2487.
25. Mao G, Gigliotti MJ, Myers D, Yu A, Whiting D: **Single-surgeon direct comparison of O-arm neuronavigation versus mazor x robotic-guided posterior spinal instrumentation.** *World Neurosurg* 2020, **S1878-8750**(20):30193-30195.
26. Laudato PA, Pierzchala K, Schizas C: **Pedicle screw insertion accuracy using O-arm, robotic guidance, or freehand technique: A comparative study.** *Spine (Phila Pa 1976)* 2018, **43**(6):E373-E378.
27. Kam JKT, Gan C, Dimou S, Awad M, Kavar B, Nair G, Morokoff A: **Learning curve for robot-assisted percutaneous pedicle screw placement in thoracolumbar surgery.** *Asian Spine J* 2019, **13**(6):920-927.
28. Urakov TM, Chang KH-K, Burks SS, Wang MY: **Initial academic experience and learning curve with robotic spine instrumentation.** *Neurosurg Focus* 2017, **42**(5):E4.
29. Teles AR, Paci M, Abduljabbar FH, Ouellet JA, Golan JD: **Anatomical and technical factors associated with superior facet joint violation in lumbar fusion** *Journal of neurosurgery Spine* 2017, **28**(2):173-180.
30. Patel RD, Graziano GP, Vanderhave KL, Patel AA, Gerling MC: **Facet violation with the placement of percutaneous pedicle screws.** *Spine (Phila Pa 1976)* 2011, **36**(26):E1749-E1752.
31. Herren C, Reijnen M, Pishnamaz M, Lichte P, Andruszkow H, Nebelung S, Siewe J, Hildebrand F, Kobbe P: **Incidence and risk factors for facet joint violation in open versus minimally invasive procedures during pedicle screw placement in patients with trauma.** *World Neurosurgery* 2018, **112**:E711-E718.
32. Tannous O, Jazini E, Weir TB, Banagan KE, Koh EY, Anderson G, Gelb DE, Ludwig SC: **Facet joint violation during percutaneous pedicle screw placement: A comparison of two techniques.** *Spine (Phila Pa 1976)* 2017, **42**(15):1189-1194.
33. Chen Z, Zhao J, Xu H, Liu A, Yuan J, Wang C: **Technical factors related to the incidence of adjacent superior segment facet joint violation after transpedicular instrumentation in the lumbar spine.** *Eur Spine J* 2008, **17**(11):1476-1480.
34. Tian W, Xu Y, Liu B, Liu Y, He D, Yuan Q, Lang Z, Lyu Y, Han X, Jin P: **Lumbar spine superior-level facet joint violations: percutaneous versus open pedicle screw insertion using intraoperative 3-dimensional**

computer-assisted navigation. *Chinese medical journal* 2014, **127**(22):3852-3856.

35. Moshirfar A, Jenis LG, Spector LR, Burke PJ, Losina E, Katz JN, Rand FF, Tromanhauser SG, Banco RJ: **Computed tomography evaluation of superior-segment facet-joint violation after pedicle instrumentation of the lumbar spine with a midline surgical approach.** *Spine (Phila Pa 1976)* 2006, **31**(22):2624-2629.

Tables

Table 1 Demographics and anatomical characteristics

Variables	N = 72
Mean age (year)	50.17 ± 10.93
Gender	
Male	32 (44.4%)
Female	40 (55.6%)
Body mass index (m/kg ²)	23.54 ± 3.98
Diagnosis	
Fracture	34 (47.2%)
Lumbar disc hernia	24 (33.3%)
Lumbar canal stenosis	2 (2.8%)
Spondylolisthesis (grade I)	7 (9.7%)
Mild degenerative scoliosis	5 (6.9%)

Table 2 Screw-related variables

Variables	N = 332
Instructed level	
T11	16 (4.8%)
T12	44 (13.3%)
L1	54 (16.3%)
L2	44 (13.3%)
L3	20 (6.0%)
L4	60 (18.1%)
L5	70 (21.1%)
S1	24 (7.2%)
The side of screw	
Right	166 (50%)
Left	166 (50%)
Facet joint osteoathritis	
Normal	45 (13.6%)
Mild	188 (56.6%)
Moderate	79 (23.8%)
Severe	20 (6.0%)
Axial pedicle angle (°)	14.02 ± 4.92
Sagittal pedicle angle	2.98 ± 6.63
Distance from skin to insertion point (cm)	5.35 ± 0.90

Table 3 Instrumented levels and risks for intra-pedicular accuracy

Level	A	B	A+B	C	D	E	Intact (%)	Breached (%)	Raw OR (95%CI)	p
T11	12	1	13	1	0	2	12 (3.6)	4 (1.2)	2.166 (0.653-6.862)	0.364
T12	33	5	38	2	2	2	33 (9.9)	11 (3.3)	2.333 (1.084-5.022)	0.270
L1	42	3	45	6	2	1	42 (12.7)	12 (3.6)	1.984 (0.953-4.128)	0.063
L2	34	7	41	0	2	1	34 (10.2)	10 (3.0)	1.995 (0.910-4.374)	0.080
L3	17	3	20	0	0	0	17 (5.1)	3 (0.9)	1.075 (0.302-3.820)	0.911
L4	56	4	60	0	0	0	56 (16.9)	4 (1.2)	0.380 (0.131-1.104)	0.066
L5	67	2	69	1	0	0	67 (20.2)	3 (0.9)	0.222 (0.067-0.737)	0.013
S1	24	0	24	0	0	0	24 (7.2)	0 (0.0)	0	0.998
Total	285	25	310	10	6	6	285 (85.8)	47 (14.2)	-	-

A: optimal position; A+B: clinically acceptable position

Table 4 Instrumented levels and risks for proximal facet joint violation

Level	0	1	2	3	Intact (%)	Violated (%)	Raw OR (95%CI)	p
T11	14	1	0	1	14 (4.2)	2 (0.6)	0.459 (0.102-2.066)	0.462
T12	40	3	0	1	40 (12.0)	4 (1.2)	0.295 (0.102-0.851)	0.029
L1	45	5	3	1	45 (13.6)	9 (2.7)	0.618 (0.287-1.329)	0.214
L2	36	4	2	2	36 (10.8)	8 (2.4)	0.705 (0.313-1.589)	0.398
L3	15	3	0	2	15 (4.5)	5 (1.5)	1.111 (0.390-3.162)	0.843
L4	42	8	4	6	42 (12.7)	18 (5.4)	1.547 (0.830-2.885)	0.168
L5	41	10	11	8	41 (12.3)	29 (8.7)	3.153 (1.785-5.572)	0.001
S1	22	0	2	0	22 (6.6)	2 (0.6)	0.282 (0.065-1.229)	0.124
Total	255	34	22	21	255 (76.8)	77 (23.2)	-	-

Table 5 Logistic regression model of variables associated with intra-pedicular and proximal facet joint accuracy, respectively

Variables group	Raw OR (95% CI)	Adjusted OR (95% CI)	p
Pedicle cortex penetration			
Sex of male	0.494 (0.263-0.925)	0.320 (0.140-0.732)	0.007
Facet angle $\geq 45^\circ$	0.241 (0.091-0.642)	0.266 (0.090-0.786)	0.017
Depth of surgical field $\geq 4.5\text{cm}$	0.173 (0.088-0.340)	0.342 (0.134-0.868)	0.024
Lumbosacral instrumentation	0.199 (0.095-0.417)	0.227 (0.091-0.566)	0.001
Facet joint violation			
L5 insertion	3.153 (1.785-5.572)	2.020 (1.084-3.766)	0.027
Facet angle $\geq 45^\circ$	2.197 (1.268-3.807)	1.839 (1.026-3.298)	0.041

Figures

Figure 1

Intra-pedicular accuracy according to the Gertzbin–Robbins scale in robot-assisted technique. A. Grade A, screw position completely within the pedicle (a); Grade B, cortical breach of 0.94 mm (b); B. Grade B, cortical breach of 1.78 mm (c); Grade C, cortical breach of 2.63 mm (d); C. grade D, cortical breach of 5.29 mm (e); D. Grade E, cortical breach of 7.53 mm (f).

Figure 2

Proximal facet joint violation in accordance with the Babu scale in robot-assisted technique. A. Grade 0, screw not in facet, transverse position (a); B. Grade 1, screw in lateral facet but not in facet articulation, transverse position (b); C. Grade 2, penetrating facet articulation less than 1 mm, transverse position (c); D. Grade 3, traveling within facet articulation, transverse position (d); E. Grade 0, screw not in facet, sagittal position (e); F. Grade 1, screw in lateral facet but not in facet articulation, sagittal position (f). G. Grade 2, penetrating facet articulation less than 1 mm, sagittal position (g); H. Grade 3, traveling within facet articulation, sagittal position (h).

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [SupplementaryMaterial1.docx](#)
- [SupplementaryMaterial2.docx](#)