Performance Characterization of the SAFIR Prototype PET Insert

Parisa Khateri (✉ pkhateri@phys.ethz.ch)

ETH Zurich <https://orcid.org/0000-0002-8563-3450>

Werner Lustermann

ETH Zurich

Christian Ritzer

ETH Zurich

Charalampos Tsoumpas

University of Leeds

Guenther Dissertori

ETH Zürich

Original research

Keywords: PET, PET/MRI, dynamic imaging, preclinical imaging, NEMA performance

Posted Date: May 6th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-458644/v1>

License: This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

RESEARCH

Performance Characterization of the SAFIR Prototype PET Insert

Parisa Khateri^{1*}, Werner Lustermann¹, Christian Ritzer¹, Charalampos Tsoumpas² and Günther Dissertori¹

*Correspondence:

pkhateri@phys.ethz.ch

¹Institute for Particle Physics and Astrophysics, ETH Zürich, Switzerland

Full list of author information is available at the end of the article

Abstract

Background: The SAFIR prototype insert is a preclinical Positron Emission Tomography (PET) scanner built to acquire dynamic images simultaneously with a 7 T Bruker Magnetic Resonance Imaging (MRI) scanner. The insert is designed to perform with an excellent coincidence resolving time of 194 ps Full Width Half Maximum (FWHM) and an energy resolution of 13.8% FWHM. These properties enable it to acquire precise quantitative images at activities as high as 500 MBq suitable for studying fast biological processes within short time frames (< 5 s). In this study, the performance of the SAFIR prototype insert is evaluated according to the NEMA NU 4-2008 standard while the insert is inside the MRI without acquiring MRI data.

Results: Applying an energy window of 391 – 601 keV and a coincidence time window of 500 ps the following results are achieved. The average spatial resolution at 5 mm radial offset is 2.6 mm FWHM when using the Filtered Backprojection 3D Reprojection (FBP3DRP) reconstruction method, improving to 2.3 mm when using the Maximum Likelihood Expectation Maximization (MLEM) method. The peak sensitivity at the center of the scanner is 1.06%. The Noise Equivalent Count Rate (NECR) is 799 kcps at the highest measured activity of 537 MBq for the mouse phantom and 121 kcps at the highest measured activity of 624 MBq for the rat phantom. The NECR peak is not yet reached for any of the measurements. The scatter fractions are 10.9% and 17.8% for the mouse and rat phantoms, respectively. The uniform region of the image quality phantom has a 3.0% STD, with a 4.6% deviation from the expected number of counts per voxel. The spill-over ratios for the water and air chambers are 0.18 and 0.17, respectively.

Conclusions: The results satisfy all the requirements initially considered for the insert, proving that the SAFIR prototype insert can obtain dynamic images of small rodents at high activities (~ 500 MBq) with a high sensitivity and an excellent count-rate performance.

Keywords: PET, PET/MRI, dynamic imaging, preclinical imaging, NEMA performance.

1 Background

SAFIR (Small Animal Fast Insert for mRi) is a preclinical Positron Emission Tomography (PET) insert designed for a 7 T Bruker BioSpin 70/30 Magnetic Resonance Imaging (MRI) scanner [1, 2]. It has been designed following the need to measure fast biological processes such as cerebral blood flow of small rodents [3]. The main requirements are precise quantitative PET images of $[^{15}\text{O}]\text{H}_2\text{O}$ with a spatial reso-

lution of ~ 2 mm, a temporal resolution of < 5 s and a quantitative voxel accuracy better than 10%. This implies measurements with activities as high as ~ 500 MBq for sufficient counting statistics. High rate capability in turn requires short coincidence resolving time (< 500 ps) and small coincidence time window (~ 500 ps) in order to reduce the number of random coincidences [2, 4].

Since these requirements cannot be fulfilled by any other preclinical PET scanners [5–16], we have designed the Small Animal Fast Insert for mRi (SAFIR) system. The prototype version of SAFIR has been built and initially characterized, showing an excellent time resolution of 194 ps and an energy resolution of 13.8% [17–19]. In this study, we evaluate the SAFIR prototype performance according to the National Electrical Manufacturers Association NU 4-2008 standard [20] which we refer to as NEMA in the rest of this paper. In addition, the energy resolution and the Coincidence Resolving Time (CRT) of the scanner are studied as a function of activity.

2 Materials and Methods

2.1 The SAFIR Prototype PET Insert

The SAFIR prototype PET insert comprises 2880 lutetium-yttrium oxyorthosilicate (LYSO) crystals arranged into 16 rings with a crystal-to-crystal distance of 128.1 mm. It covers an axial field of view of 35.6 mm. The insert is designed as a dodekagon with 12 identical sectors. Each sector hosts two detector modules. Each detector module comprises two LYSO crystal matrices, one made of 8×7 and one made of 8×8 crystals with $2.12 \times 2.12 \times 13.0$ mm³ size arranged on a grid with 2.2 mm pitch, using enhanced specular reflector foil (3M Vikuiti Enhanced Reflector Films) as spacer. Hence, the detector head in each sector is made of 2×2 parallel crystal matrices with 0.6 mm gap, aligning 16 crystals in axial direction times 15 crystals in transaxial direction. The crystal matrices are one-to-one coupled to Silicon Photo-Multiplier (SiPM) arrays (Hamamatsu S13361-2050 AE-08 SPL MPPC) with 2.2 mm pitch. Two crystal matrix and SiPM array assemblies are in turn mounted onto one detector module board hosting at the same time four PETA6SE Application-Specific Integrated Circuits (ASICs) [21].

2.2 Data Acquisition and Data Processing

The PETA6SE ASICs provide digitised energy and timing information of hits in the crystals, which are continuously read out by means of Field Programmable Gate Arrays (FPGAs) and transferred to the Data Acquisition (DAQ) computer via 12 optical Ethernet links (1 Gbit/s each).

The data are acquired with identical settings for the overvoltage of the SiPM arrays (6 V), the readout frequency (280 MHz) and the energy threshold (30 LSB, corresponding to 100 keV) of the PETA6SE ASIC. For all measurements, we applied a relative timing threshold of 75 LSB (corresponding to 45 mV), except for the scatter fraction and count rate measurements, where timing thresholds of 150 LSB (corresponding to 90 mV) and 250 LSB (corresponding to 150 mV) were used. This is required for the proper functioning of the SAFIR prototype at high activities up to 500 MBq.

The acquired raw data are processed off-line. We apply energy and timing calibrations converting time counter information into time stamps in picosecond and

Charge to Digital Converter (QDC) values into energies in electron volt [19]. These calibrated hit data are filtered by an energy window of 391 – 601 keV and sorted into coincidence events using the single window method and a coincidence time window of 500 ps. Coincidence events with more than two singles as well as those with a tangential angle between the singles smaller than 90° are eliminated. The resulting coincidence data set is stored in list mode. The same processing parameters are applied to all data reported below. For the peak sensitivity measurement, we create a second coincidence data set, by filtering and sorting the data using a larger energy window of 250 – 650 keV in addition to the one mentioned above.

We use [Software for Tomographic Image Reconstruction \(STIR\)](#) for the image reconstruction [22]. The coincidence data sets, stored in list mode, are sorted into three-dimensional (3D) projection data, which are then reconstructed into images. We employ two reconstruction methods: 1) [Filtered Backprojection 3D Reprojection \(FBP3DRP\)](#) [23] and 2) [Maximum Likelihood Expectation Maximization \(MLEM\)](#) [24]. The voxel size is $0.55 \times 0.55 \times 1.1 \text{ mm}^3$. When we perform MLEM reconstructions, we use 30 iterations and apply a Gaussian filter with a [Full Width Half Maximum \(FWHM\)](#) of $1.1 \times 1.1 \times 2.2 \text{ mm}^3$ after each iteration, except for the reconstruction of the spatial resolution data, where NEMA requires a reconstruction without any filtering.

In the [FBP3DRP](#) method, a cylindrical scanner model with an equidistant spacing of crystals in axial and transaxial direction is used while in the [MLEM](#) method, the scanner is modeled with the exact generic geometry [25]. It has been shown that using a more accurate model of the scanner improves the image quality [26].

2.3 Characterization of the SAFIR prototype insert

The performance of the [SAFIR](#) prototype insert is characterized according to the [NEMA](#) protocol while it is inside the [MRI](#) scanner without acquiring [MRI](#) data. In addition, the energy resolution and the [CRT](#) of the insert are evaluated at different activities up to 537 MBq using the [NEMA](#) mouse scatter phantom.

2.3.1 Spatial resolution

We measure the spatial resolution using a ^{22}Na point source ([Eckert & Ziegler Isotope Products](#), MMS09-022), with an activity of 0.487 MBq and a source diameter of 0.25 mm centered in an acrylic cube of 10 mm edge length. We acquire data from the point source for two axial positions of 0.0 mm and 8.9 mm (equivalent to a quarter of the axial field of view (FOV)) and for 10 radial positions from 0.0 mm to 45.0 mm in steps of 5.0 mm. At least 10^5 coincidence events are collected per source position.

We reconstruct the images of the point source using a) [FBP3DRP](#) and b) [MLEM](#) without any smoothing. The [FWHM](#) and [Full Width Tenth Maximum \(FWTM\)](#) of the images are obtained according to [NEMA](#).

2.3.2 Sensitivity

We measure the sensitivity using the the same point source as for the spatial resolution measurements (section 2.3.1). The source is located on the central axis of the scanner and is axially moved in steps of 1 mm starting from the axial offset of

–15 mm and ending with the axial offset of 15 mm. We collect 8×10^4 coincidence events per source position. The sensitivity at each axial position is calculated according to NEMA. The system sensitivities for mouse and rat are not calculated as SAFIR has a shorter axial FOV than required by NEMA.

2.3.3 Count Rate Performance and Scatter Fraction

We use NEMA mouse and rat scatter phantoms to measure count rate performance and scatter fraction. We use an ^{18}F labeled radiotracer as the radioactive source. The start and end activities are 537 MBq and 0.22 MBq for the mouse phantom, and 624 MBq and 1 MBq for the rat phantom, respectively. At least 5×10^5 coincidence events are collected per acquisition. Data analysis is performed according to NEMA to obtain the total, true, scattered and random count rates as well as the Noise Equivalent Count Rate (NECR) and the system Scatter Fraction (SF).

2.3.4 Energy Resolution and Coincidence Resolving Time

Using the data set collected with the mouse scatter phantom for the count rate measurement (section 2.3.3), we evaluate as well energy resolution and CRT as a function of activity. We report FWHM of a Gaussian fit (480 – 580 keV) to the coincidence energy spectrum as the energy resolution. The maximum of the coincidence timing spectrum is obtained by a parabolic fit through the highest bin and its two neighbours. We measure the FWHM of the spectrum by linearly interpolating between the bins at half the maximum and report the width as coincidence timing resolution [19].

2.3.5 Image Quality Study

We use the NEMA image quality phantom comprising: 1) two cold chambers, one filled with water and one filled with air, 2) five hot rods of (1, 2, 3, 4 and 5) mm diameters and 3) a uniform region. We fill the phantom with an ^{18}F labeled radiotracer. According to NEMA, the measurement should be done with an initial activity of 3.7 MBq and an acquisition time of 20 min. Since the SAFIR prototype insert does not cover the whole length of the phantom, we run the measurement in two bed positions and modify activity and acquisition time such that the same number of annihilations are produced in both bed positions, taking into account the decay of the activity. In the first bed position, the cold rods together with the uniform region are measured with an initial activity of 4.2 MBq for 25 min and in the second bed position, the hot rods are measured with an initial activity of 3.3 MBq for 32 min.

We reconstruct the data using MLEM with random, attenuation, scatter and normalization corrections embedded into the reconstruction algorithm. In addition, we calibrate the reconstructed image providing absolute voxel values and thus quantitative PET data. Both, data corrections and the quantitative calibration are described below.

The reconstructed images are analyzed according to NEMA to obtain the following measures:

- *Uniformity*: The mean, maximum, minimum and Standard Deviation (STD) of the counts in a cylindrical Region of Interest (ROI) of 22.5 mm diameter and 10 mm length are measured in the center of the uniform region.

- *Recovery coefficient*: The recovery coefficient and its STD are calculated for each hot rod.
- *Spill-over ratio*: The spill-over ratio and its STD are evaluated for each cold chamber.

In addition to the above mentioned values and to evaluate the accuracy of the image, we calculate the deviation of the absolute voxel value as follows:

$$Deviation(\%) = \frac{E - m}{E} \times 100\% \quad (1)$$

where m is the measured mean value of the uniform region and E is the expected voxel value.

We apply the following corrections to obtain quantitative PET data:

- *Random correction*: We estimate the number of random coincidences per Line of Response (LOR) using the singles-prompt method introduced by J.F. Oliver et al. [27]. This method is an extension of the singles-rate method [28], outperforming the singles-rate method at high activities.
- *Attenuation correction*: The attenuation maps of the phantoms and the bed are generated based on their known geometries, the material compositions and their corresponding attenuation coefficients taken from the National Institute of Standards and Technology (NIST) reference database [29]. The attenuation correction factors are calculated using these maps.
- *Scatter correction*: We use the Single Scatter Simulation (SSS) method implemented in STIR to estimate the number of scattered events that occurred in the phantom [30,31]. In this method, the attenuation map and the scanner geometry are down-sampled in order to accelerate the computation. We use a down-sampling factor of two.
- *Detector normalization*: We employ the direct normalization method to obtain the normalization factors [32,33]. A cylinder phantom of 74 mm diameter and 50 mm length uniformly filled with an ^{18}F labeled tracer is located in the center of the scanner. The data are acquired with a starting activity of 141 MBq for ~ 14 h, resulting in a total number of 2.65×10^9 coincidences and an average number of 1768 counts per LOR. The normalization factors are obtained by correcting for different lengths of intersection between the LOR and the cylinder volume and the effects of attenuation, scatter and randoms in the collected data.
- *Quantitative calibration of voxel values*: A cylinder of 40 mm diameter and 20 mm length uniformly filled with an ^{18}F labeled tracer located in the center of the scanner is measured for 30 min with a starting activity of 4.4 MBq. We reconstruct the image including normalization, attenuation, random, and scatter corrections. The average voxel value is computed inside a cylindrical ROI of 35 mm diameter and 15 mm length in the center of the image. The absolute calibration factor is calculated as the ratio of the expected number of counts per voxel to the average voxel value in the ROI.

3 Results and Discussion

3.1 Spatial Resolution

Table 1 and figure 1 show the spatial resolution results for the axial positions of 0.0 mm and 8.9 mm using FBP3DRP and MLEM.

Table 1 Spatial resolution results of the SAFIR prototype PET insert. All values are in mm.

At the center of axial FOV												
RO*	FBP3DRP						MLEM					
	Radial		Tangential		Axial		Radial		Tangential		Axial	
	HM*	TM*	HM	TM	HM	TM	HM	TM	HM	TM	HM	TM
0	1.28	2.69	2.38	4.12	2.63	4.59	2.18	3.90	2.57	3.86	2.18	3.97
5	2.82	5.61	2.04	3.82	2.83	5.15	3.50	5.72	2.18	3.80	2.00	3.91
10	3.25	6.99	2.26	4.42	2.85	5.21	3.44	5.58	2.68	4.93	1.82	3.83
15	3.37	11.45	2.91	6.13	2.87	5.28	3.06	5.23	3.23	5.21	1.89	3.89
20	2.91	7.07	3.07	5.52	2.89	5.32	3.51	6.24	3.17	5.24	1.84	3.87
25	3.42	12.75	3.11	5.90	2.89	5.33	3.49	7.02	2.87	5.33	1.85	3.88
30	3.42	11.46	2.82	6.36	2.91	5.37	4.08	7.46	3.01	5.59	1.80	3.85
35	3.75	11.68	3.02	6.98	2.91	5.38	4.85	7.72	2.98	6.36	1.74	3.82
40	4.33	11.55	3.15	8.99	2.96	5.49	5.04	8.34	3.04	6.44	1.93	3.67
45	4.68	12.06	3.22	8.46	2.91	5.29	2.61	4.76	2.22	3.57	2.13	3.96

At 1/4 of the axial FOV												
RO	FBP3DRP						MLEM					
	Radial		Tangential		Axial		Radial		Tangential		Axial	
	HM	TM	HM	TM	HM	TM	HM	TM	HM	TM	HM	TM
0	1.88	3.54	2.01	3.62	2.86	5.19	2.52	4.28	2.52	4.28	2.34	4.10
5	2.88	5.78	2.21	4.17	2.93	5.27	3.87	6.23	2.57	4.59	2.26	4.03
10	3.63	7.28	2.37	4.64	2.94	5.28	3.55	6.14	3.00	5.55	2.24	4.00
15	3.30	7.62	2.97	6.16	2.96	5.31	3.34	6.01	3.22	5.59	2.27	4.01
20	3.16	13.60	3.06	5.51	2.96	5.33	3.71	6.97	3.24	5.48	2.24	3.99
25	3.65	13.81	2.95	5.86	2.98	5.35	3.85	7.68	2.95	5.53	2.24	3.99
30	3.94	14.14	2.71	6.05	2.98	5.35	4.29	7.90	2.91	5.57	2.16	3.92
35	4.05	13.74	2.69	6.30	2.99	5.35	4.81	8.34	2.92	5.89	2.09	3.82
40	4.64	12.47	2.99	7.36	2.99	5.36	5.04	9.29	2.88	5.99	2.06	3.77
45	4.78	13.02	2.96	7.48	3.02	5.42	2.92	5.35	2.50	4.23	2.41	4.13

* RO = Radial Offset; HM = FWHM; TM = FWTM.

The MLEM algorithm yields better spatial resolution results than the FBP3DRP method. For instance, the average FWHM over radial, tangential and axial directions at the 5 mm radial offset and 0.0 mm axial offset is 2.6 mm for the FBP3DRP method and 2.3 mm for the MLEM method. Using FBP3DRP, the spatial resolution especially degrades with respect to FWTM due to the streak artifact that is present in the reconstructed image. The FWHM in radial direction degrades toward the edge of the scanner due to the parallax effect.

We compare the SAFIR prototype results with a group of preclinical PET scanners, referred to as *reference scanners* in this paper [5–15]. Among these scanners, Hyperion II^D and nanoScan are for combined PET-MRI systems and Bruker is designed for the same MRI system for which SAFIR is designed. Table 2 presents the spatial resolution of SAFIR and the *reference scanners*. PET scanners of similar crystal size yield a slightly better spatial resolution than the SAFIR prototype when using FBP3DRP reconstruction. This is related to the reconstruction algorithm in combination with the geometry of the scanner. For scanners of relatively large detector heads such as SAFIR with $(8 + 7) \times 8$ crystals per module, the FBP3DRP method introduces error, as this method requires regularly spaced projection data which in turn requires regularly spaced detector elements. Interpolating the crystal positions into a regular space causes artifacts in the image and degrades the spatial resolution. The iterative algorithm does not require interpolation. It uses the exact *generic* geometry and thus yields better results. In addition, all scanners in table 2 with a similar crystal size as SAFIR have a shorter crystal length, except

Figure 1 Spatial resolution in radial, tangential and axial directions for the point source located in the center of the axial FOV (a, b) and at one forth of the axial FOV (c, d). The image reconstruction methods are FBP3DRP (a, c) and MLEM (b, d).

for LabPET 8 which uses depth of interaction information and has a continuous arrangement of the crystals on a ring.

3.2 Sensitivity

Figure 2 shows the total sensitivity calculated for the point source data measured at different axial positions. The maximum sensitivity at the center is 1.06 % for the energy window of 391 – 601 keV. It decreases to 0.2 % at 15 mm axial offset. The expected triangular profile is clearly visible.

Table 3 compares the peak sensitivity of SAFIR and the *reference scanners*. The scanners have different axial FOVs and use different energy windows. Given its short axial FOV, the SAFIR prototype yields a high sensitivity which is in line with our goal for SAFIR.

3.3 Count Rate Performance and Scatter Fraction

Count rate results as a function of the activity in the phantom are plotted in figure 3, for the mouse and rat phantoms and for two relative timing thresholds of 90 mV and 150 mV. The NECR peak is not reached for any of the measurements. This proves that the SAFIR prototype insert is capable of handling activities higher than 500 MBq.

The higher relative timing threshold of 150 mV leads to higher count rates and NECR. The highest NECR is 799 kcps at 537 MBq (corresponding to an activity

Table 2 Comparison of the spatial resolution of different preclinical PET scanners. Data are taken from [5–15].

System	Recon. method	crystal size (mm ³)	FWHM (mm) ^a		
			R	T	A
SAFIR	FBP3DRP	2.1 × 2.1 × 13.0	2.77	1.89	2.83
SAFIR	MLEM	2.1 × 2.1 × 13.0	2.45	2.18	2.07
Hyperion II ^D	FBP	0.93 × 0.93 × 12	1.7	1.8	1.4
NanoScan	SSRB FBP	1.1 × 1.1 × 13.0	1.50	1.32	0.91
MuPET	SSRB FBP	1.24 × 1.4 × 9.5	1.48	1.34	0.99
Inveon	FORE ^b + 2D FBP	1.5 × 1.5 × 10.0	1.6	1.6	2.3
IRIS	MLEM	1.6 × 1.6 × 12.0	1.05	1.05	1.25
ClearPET	3D FBP	2.0 × 2.0 × 10.0	1.94	2.00	3.24
Mosaic HP	FBP3DRP	2.0 × 2.0 × 10.0	2.32	2.32	2.64
LabPET 8 TM	SSRB FBP	2.0 × 2.0 × 14.0	1.65	1.70	1.40
microPET R4	FORE ^b FBP	2.1 × 2.1 × 10.0	2.13	2.21	2.72
Xtrim-PET	SSRB FBP	2.1 × 2.1 × 10.0	2.01	1.95	1.74
Bruker	MLEM	50.0 × 50.0 × 10.0 ^c	0.87	0.78	0.77

^a FWHM at 5 mm radial offset in radial (R), tangential (T) and axial (A) directions.^b Fourier rebinning algorithm [34].^c Monolithic crystals. **Figure 2** Sensitivity at different axial positions.**Table 3** Comparison of the peak sensitivity of different preclinical PET scanners. Data are taken from [5–15].

System	Energy win. (keV)	Time win. (ns)	Axial FOV (mm)	Sensitivity (%)
SAFIR	391 – 601	0.5	36	1.06
SAFIR	250 – 650	0.5	36	2.57
Hyperion II ^D	250 – 625	200	96.7	4.0
NanoScan	250 – 750	5	94	8.4
MuPET	350 – 650	3.4	116	6.35
Inveon	350 – 625	3.4	127	6.72
IRIS	250 – 750	5.2	95	8
ClearPET	250 – 650	12	110	1.87
Mosaic HP	385 – 665	7	119	1.77
LabPET 8 TM	250 – 650	20	75	1.33
microPET R4	350 – 650	6	78	2.4
Xtrim-PET	250 – 650	10	50.3	2.99
Bruker	—	—	150	11.0

concentration of 2690 MBq/ml) for the mouse scatter phantom and 121 keps at 624 MBq (corresponding to an activity concentration of 1390 MBq/ml) for the rat scatter phantom. Due to the relatively short FOV of the SAFIR prototype, the detector receives many single γ -rays for the rat scatter phantom from outside the FOV, resulting in many randoms.

The scatter fractions are given in table 4. They are in line with results obtained by others (table 5).

Table 4 Scatter Fractions (SFs) for the measurements of the mouse and rat scatter phantoms at two different relative timing threshold of 90 mV and 150 mV.

Phantom	90 mV		150 mV	
	Mouse	Rat	Mouse	Rat
SF (%)	8.9	17.9	10.9	17.8

Table 5 presents the NECR peak and the scatter fraction of SAFIR and the *reference scanners*. Compared to other preclinical PET scanners, the SAFIR prototype shows an excellent count rate performance. All scanners reach their NECR peak at activities much less than 500 MBq. The SAFIR prototype has the highest measured NECR in comparison with other scanners with the exception of the Inveon [8] and MuPET [7] systems which have longer axial FOV and use larger energy windows.

3.4 Energy Resolution and Coincidence Resolving Time

The energy resolution and CRT versus activity are plotted in figure 4 for the mouse scatter phantom measurements. The pile-up increases with activity, thus energy resolution and CRT degrade by increasing activity. However, both values remain in a range fully sufficient for our application. The 90 mV relative timing threshold yields smaller CRTs than the 150 mV threshold, especially at lower activities.

Table 5 Comparison of NECR and Scatter Fraction (SF) of different preclinical PET scanners. Data are taken from [5–15].

System	Mouse phantom			Rat phantom		
	NECR peak (kcps)	Activity (MBq) ^b	SF (%)	NECR peak (kcps)	Activity (MBq)	SF (%)
SAFIR ^a	799	537	10.9	121	624	17.8
Hyperion II ^D	407	46	13	—	—	—
NanoScan	406	30	17.3	119	25	34
MuPET	1100	57	11.9	352	65	28.0
Inveon	1670	131	7.8	592	110	17.2
IRIS	185	14	15.6	40	10	22.4
ClearPET	73	18	31	—	—	—
Mosaic HP	555	92	5.4	244	87	12.7
LabPET 8 TM	279	82	15.6	94	91	29.5
microPET R4	168	91	18	89	81	28
Xtrim-PET	113.2	17	12.5	82.8	15	25.8
Bruker	486	23	—	240	23	—

^a For SAFIR, the highest NECR values at the highest measured activities are reported as the NECR peaks are not reached.

^b Activity at which the NECR peak is measured.

Figure 4 (a) Coincidence Resolving Time (CRT) and (b) energy resolution versus activity in the mouse scatter phantom measurement.

3.5 Image Quality Analysis

Figure 5 shows transverse, coronal and sagittal cross-sections of different regions in the NEMA image quality phantom. In short:

- All images are artifact-free.
- The uniform region has a 3.0% STD (table 6).
- The Spill-Over Ratio for the water/air chamber is 0.18/0.17 (table 7).
- The recovery coefficients for the smallest and largest rods are 0.13 and 0.88, respectively (table 8).
- The smallest hot rod (1 mm diameter) is not visible in the image.

The deviation of the absolute voxel value in the uniform region is $4.6\% \pm 6.5\%$. For the uncertainty, we only propagated the uncertainties of the measurement of the activity and the volume of the image quality phantom.

The Spill-Over Ratios of the air and water chambers are almost identical (0.01 difference), which is a direct result of including the data corrections into the reconstruction. Especially, the attenuation and scatter corrections influence the amount of background noise in the cold chambers with different attenuation properties.

Figure 5 Different cross sections of the NEMA image quality phantom reconstructed using MLEM with 30 iterations, including random, scatter, attenuation and normalization corrections into the reconstruction. (a) Transverse view of cold rods, (b) transverse view of the uniform region, (c) transverse view of hot rods, (d) coronal view of cold rods and uniform region and (e) sagittal view of hot rods.

Table 6 Uniformity measurement for the uniform region of the NEMA image quality phantom.

Mean (count)	Minimum (count)	Maximum (count)	STD (%)	Deviation (%)
93796	86060	102456	3.0	4.6

Table 7 Spill-Over Ratios (SORs) for the cold chambers in the NEMA image quality phantom.

	SOR (%)	STD (%)
Air chamber	17.3	19.5
Water chamber	18.5	20.1

Table 8 Recovery coefficients for the hot rods in the NEMA image quality phantom.

Rod diameter (mm)	Recovery coefficient	%STD
1	0.13	148.4
2	0.29	55.3
3	0.49	22.8
4	0.65	14.3
5	0.88	8.2

There is a strong correlation between the crystal size and the spatial resolution and thus with the recovery coefficient. SAFIR’s performance in terms of recovery coefficient is comparable with the scanners of similar crystal size (table 9). However, scanners of smaller crystals such as [5,8] and the Bruker scanner [15] with monolithic LYSO crystals outperform SAFIR.

Comparing Spill-Over Ratios is more difficult, as they depend on the reconstruction algorithm and whether or not attenuation and scatter corrections have been included into the reconstruction. Table 9 presents the Spill-Over Ratios and uniformities of SAFIR and the reference scanners. The uniformity achieved with the SAFIR prototype insert is the best among these reported in the table.

Table 9 Comparison of the recovery coefficient (RC) for the 3 mm diameter rod, uniformity and Spill-Over Ratio (SOR) of different preclinical PET scanners. Data are taken from [5–15, 35].

System	Recon. method	Corrected for AC/SC ^b	RC	STD (%)	SOR (%)	
				Uniform region	Water chamber	Air chamber
SAFIR	3D MLEM	yes/yes	0.49	3.0	17.3	18.5
Hyperion II ^D	3D MLEM	yes/yes	0.91	3.7	5.4	6.3
NanoScan	penalized MLEM	yes/yes	0.9	3.5	6.2	5.8
MuPET	FBP3DRP	yes/no	0.75	6.5	9	5
Inveon	FORE + 2D FBP	yes/yes	0.72	5.3	1.7	-0.6
IRIS	OSEM	yes/no	0.73	7	11	11
ClearPET	3D OSEM ^a	no/no	0.42	10.9	36.9	26.7
Mosaic HP	3D RAMLA ^a	yes/yes	0.56	5.1	6.3	2.7
LabPET 8 TM	2D MLEM	no/no	0.58	7.0	20	11
microPET R4	FORE ^a + 2D FBP	yes/no	0.60	4.5	6.2	4.6
Xtrim-PET	FORE + 2D OSEM	no/no	0.68	3.8	25	35
Bruker	3D MLEM	yes/no	0.91	4.5	6.2	4.6

^a FORE = Fourier Rebinning; RAMLA = Row-Action Maximum-Likelihood Algorithm; OSEM = Ordered Subset Expectation Maximization

^b AC = Attenuation Correction; SC = Scatter correction

4 Conclusion

The performance of the SAFIR prototype insert has been evaluated according to the NEMA NU 4-2008 standard while the insert was inside the 7 T MRI scanner without acquiring MRI data. The MRI-compatibility of the insert has been tested in previous studies [19, 36].

The results satisfy all requirements initially considered for the insert. The SAFIR prototype yields a high sensitivity for its short axial coverage. The count rate measurement results in an excellent NECR value of 799kcps at the highest measured activity of 537 MBq using the mouse phantom, while not yet reaching the NECR peak. This demonstrates the prototype capability to handle high rate measurements, appropriate for dynamic imaging of fast biological processes. The spatial resolution has been shown to be as good as for other preclinical scanners with similar crystal size. The tests performed using the image quality phantom present a high uniformity and accuracy for the reconstructed images, suitable for quantitative PET imaging.

The final version of the SAFIR insert is being developed with the same components as of the prototype but a quadrupled axial coverage, allowing dynamic whole body imaging of a mouse in a single bed position. This will also results in a higher sensitivity and NECR, especially for the rat scatter phantom as the contribution of the out-of-FOV γ -rays to the random count rate decreases.

Authors contributions

PK performed all experiments and data analysis and drafted the manuscript. WL contributed to the methodology of the investigation, participated in the analysis and discussion of the results and proofread the manuscript. CR wrote parts of the data analysis software, assisted in the NECR measurements and proofread the manuscript. CT was involved in various methodological approaches of the investigation and provided advice on adaptations of the STIR source code. GD carried the overall responsibility for the SAFIR project, supervised the PhD work of PK and proofread the manuscript.

Acknowledgements

We would like to thank Pascal Bebié, Jan Debus, Afroditi Eleftheriou and Geoffrey Warnock for their assistance in the measurements.

Funding

This work was supported by the ETH Zürich Foundation through ETH Research Grant ETH-30 14-2. Charalampos Tsoumpas is supported by a Royal Society Industry Fellowship (IF170011).

Abbreviations

Availability of data and materials

List mode data generated in this study can be provided upon request.

Ethics approval and consent to participate

Not applicable.

Competing interests

The authors declare that they have no competing interests.

Consent for publication

Not applicable.

Author details

¹Institute for Particle Physics and Astrophysics, ETH Zürich, Switzerland. ²Leeds Institute of Cardiovascular and Metabolic Medicine, University of Leeds, United Kingdom.

References

1. Becker, R., Casella, C., Dissertori, G., Fischer, J., Howard, A., Jeitler, A., Lustermann, W., Roeser, U., Wang, Q., Weber, B.: The SAFIR project: an innovative high rate preclinical PET/MR detector towards dynamic multimodal imaging. In: EJMNM Physics, vol. 2, pp. 1–1 (2015). SpringerOpen
2. Becker, R., Buck, A., Casella, C., Dissertori, G., Fischer, J., Howard, A., Ito, M., Khateri, P., Lustermann, W., Oliver, J.F., et al.: The SAFIR experiment: Concept, status and perspectives. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment **845**, 648–651 (2017)
3. Weber, B., Späth, N., Wyss, M., Wild, D., Burger, C., Stanley, R., Buck, A.: Quantitative cerebral blood flow measurements in the rat using a beta-probe and H215O. Journal of cerebral blood flow & metabolism **23**(12), 1455–1460 (2003)
4. Becker, R., Cachemiche, J.-P., Casella, C., Dissertori, G., Fischer, J., Howard, A.S., Kramer, K., Lustermann, W., Morel, C., Oliver, J.F., et al.: Monte-Carlo simulation based estimation of NECR, sensitivity, and spatial resolution of a novel preclinical PET insert for MR. In: 2015 IEEE Nuclear Science Symposium and Medical Imaging Conference (NSS/MIC), pp. 1–3 (2015). IEEE
5. Hallen, P., Schug, D., Weissler, B., Gebhardt, P., Salomon, A., Kiessling, F., Schulz, V.: Pet performance evaluation of the small-animal hyperion IID PET/MRI insert based on the NEMA NU-4 standard. Biomedical Physics & Engineering Express **4**(6), 065027 (2018)
6. Nagy, K., Tóth, M., Major, P., Patay, G., Egri, G., Häggkvist, J., Varrone, A., Farde, L., Halldin, C., Gulyás, B.: Performance evaluation of the small-animal nanoscan PET/MRI system. Journal of Nuclear Medicine **54**(10), 1825–1832 (2013)
7. Wong, W.-H., Li, H., Baghaei, H., Zhang, Y., Ramirez, R.A., Liu, S., Wang, C., An, S.: Engineering and performance (NEMA and animal) of a lower-cost higher-resolution animal PET/CT scanner using photomultiplier-quadrant-sharing detectors. Journal of Nuclear Medicine **53**(11), 1786–1793 (2012)
8. Bao, Q., Newport, D., Chen, M., Stout, D.B., Chatziioannou, A.F.: Performance evaluation of the inveon dedicated PET preclinical tomograph based on the NEMA NU-4 standards. Journal of Nuclear Medicine **50**(3), 401–408 (2009)
9. Belcari, N., Camarlinghi, N., Ferretti, S., Iozzo, P., Panetta, D., Salvadori, P.A., Sportelli, G., Del Guerra, A.: NEMA NU-4 performance evaluation of the IRIS PET/CT preclinical scanner. IEEE Transactions on Radiation and Plasma Medical Sciences **1**(4), 301–309 (2017)
10. Ziemons, K., Auffray, E., Barbier, R., Brandenburg, G., Bruyndonckx, P., Choi, Y., Christ, D., Costes, N., Declais, Y., Devroede, O., et al.: The ClearPET™ project: development of a 2nd generation high-performance small animal PET scanner. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment **537**(1–2), 307–311 (2005)
11. Surti, S., Karp, J.S., Perkins, A.E., Cardi, C.A., Daube-Witherspoon, M.E., Kuhn, A., Muehllehner, G.: Imaging performance of A-PET: a small animal PET camera. IEEE transactions on medical imaging **24**(7), 844–852 (2005)
12. Prasad, R., Ratib, O., Zaidi, H.: NEMA NU-04-based performance characteristics of the LabPET-8™ small animal PET scanner. Physics in Medicine & Biology **56**(20), 6649 (2011)
13. Knoess, C., Siegel, S., Smith, A., Newport, D., Richerzhagen, N., Winkeler, A., Jacobs, A., Goble, R.N., Graf, R., Wienhard, K., et al.: Performance evaluation of the microPET R4 PET scanner for rodents. European journal of nuclear medicine and molecular imaging **30**(5), 737–747 (2003)
14. Amirshadi, M., Sarkar, S., Ghafarian, P., Hashemi Shahraki, R., Geramifar, P., Zaidi, H., Ay, M.R.: NEMA NU-4 2008 performance evaluation of Xtrim-PET: A prototype SiPM-based preclinical scanner. Medical Physics **46**(11), 4816–4825 (2019)
15. Gsell, W., Molinos, C., Correcher, C., Belderbos, S., Wouters, J., Junge, S., Heidenreich, M., Velde, G.V., Rezaei, A., Nuyts, J., et al.: Characterization of a preclinical PET insert in a 7 Tesla MRI scanner: beyond NEMA testing. Physics in Medicine & Biology **65**(24), 245016 (2020)
16. Vigneaude, J.-M., Mcgrath, J., Courteau, A., Pegg, R., Gomis, A.S.-P., Camacho, A., Martin, G., Schramm, N., Brunotte, F.: Initial performance evaluation of a preclinical PET scanner available as a clip-on assembly in a sequential PET/MRI system. Physics in Medicine & Biology **63**(12), 125007 (2018)
17. Becker, R., Casella, C., Corrodi, S., Dissertori, G., Fischer, J., Howard, A., Ito, M., Lustermann, W.: Studies of the high rate coincidence timing response of the STiC and TOFPET ASICs for the SAFIR PET scanner. Journal of Instrumentation **11**(12), 12001 (2016)

18. Ahnen, M., Becker, R., Buck, A., Casella, C., Commichau, V., Di Calafiori, D., Dissertori, G., Eleftheriou, A., Fischer, J., Howard, A.S., *et al.*: Performance measurements of the SAFIR prototype detector with the STiC ASIC readout. *IEEE Transactions on Radiation and Plasma Medical Sciences* **2**(3), 250–258 (2018)
19. Ritzer, C., Becker, R., Buck, A., Commichau, V., Debus, J., Djambazov, L., Eleftheriou, A., Fischer, J., Fischer, P., Ito, M., *et al.*: Initial characterisation of the SAFIR prototype PET-MR scanner. *IEEE Transactions on Radiation and Plasma Medical Sciences* (2020)
20. Association, N.E.M., *et al.*: Performance measurements of small animal positron emission tomographs. NEMA Standards Publication, NU4-2008, 1–23 (2008)
21. Dohle, R., Sacco, I., Rittweg, T., Friedrich, T., Henning, G., Goßler, J., Fischer, P.: LTCC-based highly integrated SiPM module with integrated liquid cooling channels for high resolution molecular imaging. *Journal of Microelectronics and Electronic Packaging* **15**(2), 86–94 (2018)
22. Thielemans, K., Tsoumpas, C., Mustafovic, S., Beisel, T., Aguiar, P., Dikaio, N., Jacobson, M.W.: STIR: software for tomographic image reconstruction release 2. *Physics in Medicine & Biology* **57**(4), 867 (2012)
23. Kinahan, P.E., Rogers, J.: Analytic 3D image reconstruction using all detected events. Technical report, Triumf (1988)
24. Shepp, L.A., Vardi, Y.: Maximum likelihood reconstruction for emission tomography. *IEEE transactions on medical imaging* **1**(2), 113–122 (1982)
25. Röthlisberger, M.: Implementation and Evaluation of SAFIR Dual Ring Prototype Geometry Into STIR
26. Khateri, P., Fischer, J., Lustermann, W., Tsoumpas, C., Dissertori, G.: Implementation of cylindrical PET scanners with block detector geometry in STIR. *EJNMMI physics* **6**(1), 15 (2019)
27. Oliver, J.F., Rafecas, M.: Modelling random coincidences in positron emission tomography by using singles and prompts: a comparison study. *PLoS one* **11**(9) (2016)
28. Evans, R.D., Evans, R.: The atomic nucleus (1955)
29. Hubbell, J.H., Seltzer, S.M.: Tables of X-ray mass attenuation coefficients and mass energy-absorption coefficients 1 keV to 20 MeV for elements Z= 1 to 92 and 48 additional substances of dosimetric interest. Technical report, National Inst. of Standards and Technology-PL, Gaithersburg, MD (United States) (1995). <https://www.nist.gov/pml/x-ray-mass-attenuation-coefficients>
30. Watson, C.C.: New, faster, image-based scatter correction for 3D PET. *IEEE Transactions on Nuclear Science* **47**(4), 1587–1594 (2000)
31. Tsoumpas, C., Aguiar, P., Nikita, K., Ros, D., Thielemans, K.: Evaluation of the single scatter simulation algorithm implemented in the STIR library. In: *IEEE Symposium Conference Record Nuclear Science 2004.*, vol. 6, pp. 3361–3365 (2004). IEEE
32. Bailey, D.L., Townsend, D.W., Kinahan, P.E., Grootenboer, S., Jones, T.: An investigation of factors affecting detector and geometric correction in normalization of 3-D pet data. *IEEE Transactions on Nuclear Science* **43**(6), 3300–3307 (1996)
33. Hoffman, E.J., Guerrero, T.M., Germano, G., Digby, W., Dahlbom, M.: PET system calibrations and corrections for quantitative and spatially accurate images. *IEEE Transactions on Nuclear Science* **36**(1), 1108–1112 (1989)
34. Defrise, M., Kinahan, P.E., Townsend, D.W., Michel, C., Sibomana, M., Newport, D.F.: Exact and approximate rebinning algorithms for 3D PET data. *IEEE transactions on medical imaging* **16**(2), 145–158 (1997)
35. Goertzen, A.L., Bao, Q., Bergeron, M., Blankemeyer, E., Blinder, S., Cañadas, M., Chatziioannou, A.F., Dinelle, K., Elhami, E., Jans, H.-S., *et al.*: NEMA NU 4-2008 comparison of preclinical PET imaging systems. *Journal of Nuclear Medicine* **53**(8), 1300–1309 (2012)
36. Ritzer, C., Commichau, V., Dhawan, S.K., Fischer, J., Lustermann, W., Sumner, R., Warnock, G., Dissertori, G.: Compact MR-compatible DC-DC converter module. *Journal of Instrumentation* **14**(09), 09016 (2019)

Figures

(a) FBP3DRP, $z=0.0$ mm

(b) MLEM, $z=0.0$ mm

(c) FBP3DRP, $z=8.9$ mm

(d) MLEM, $z=8.9$ mm

Figure 1

Spatial resolution in radial, tangential and axial directions for the point source located in the center of the axial FOV (a, b) and at one forth of the axial FOV (c, d). The image reconstruction methods are FBP3DRP (a, c) and MLEM (b, d).

Figure 2

Sensitivity at different axial positions.

(a) mouse - 90 mV

(b) mouse - 150 mV

(c) rat - 90 mV

(d) rat - 150 mV

Figure 3

Count rate results for (a) the mouse scatter phantom, with relative timing threshold of 90mV, (b) the mouse scatter phantom, with relative timing threshold of 150mV, (c) the rat scatter phantom, with relative timing threshold of 90mV and (d) the rat scatter phantom, with relative timing threshold of 150mV. The highest activity for the mouse scatter phantom, 537 MBq, corresponds to an activity concentration of 2690MBq=ml. The highest activity for the rat scatter phantom, 624MBq, corresponds to an activity concentration of 1390MBq=ml.

(a)

(b)

Figure 4

(a) Coincidence Resolving Time (CRT) and (b) energy resolution versus activity in the mouse scatter phantom measurement.

Figure 5

Different cross sections of the NEMA image quality phantom reconstructed using MLEM with 30 iterations, including random, scatter, attenuation and normalization corrections into the reconstruction. (a) Transverse view of cold rods, (b) transverse view of the uniform region, (c) transverse view of hot rods, (d) coronal view of cold rods and uniform region and (e) sagittal view of hot rods.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [nemaperbmc.pdf](#)