Advanced Trap Lateral Flow Immunoassay Sensor for the Detection of Cortisol in Human Bodily Fluids

Hyun-Kyung Oh

Gwangju Institute of Science and Technology (GIST)

Kihyeun Kim

Gwangju Institute of Science and Technology (GIST)

Hyungjun Jang

Gwangju Institute of Science and Technology (GIST)

Jinhee Park

GMD Biotech Co. Ltd

Min-Gon Kim (✉ mkim@gist.ac.kr)

Gwangju Institute of Science and Technology (GIST)

Research Article

Keywords: biosensors , lateral flow immunoassay (LFI), POC diagnosis, signal amplification

Posted Date: May 7th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-478568/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published at Scientific Reports on November 19th, 2021.
See the published version at <https://doi.org/10.1038/s41598-021-02084-7>.

Abstract

Paper-based biosensors based on lateral flow immunoassay (LFI) are promising candidates for POC diagnosis because of their ease of use and rapid target detection. However, the low sensitivity of LFI limits its application, and signal amplification has been used in numerous studies to increase its sensitivity. We developed an advanced trap LFI (α -trapLFI), a simple-to-use sensor, with an additional step for signal amplification. Here, signal amplification is automatically implemented following delayed release of enhancement solution induced by water-soluble polyvinyl alcohol tape. As the polyvinyl alcohol tape is exposed to water, its polymer structure is perturbed (within 5 min), allowing ions to pass through. This new sensor was designed to have a short time delay between the flow of solutions used for the immunoassay and signal amplification. The α -trapLFI was subsequently used to detect cortisol with high sensitivity ($9.1 \text{ pg} \cdot \text{mL}^{-1}$) over a broad detection range ($0.01 - 1000 \text{ ng} \cdot \text{mL}^{-1}$) in bodily fluids. Furthermore, an excellent correlation was obtained by analyzing 20 human saliva samples using this sensor and a conventional ELISA ($R^2 = 0.90$). The new sensor will be helpful in detecting various small molecules for simple, rapid, and portable POC diagnosis of stress disorders.

Introduction

The stress hormone cortisol has emerged as a biomarker for the rapid diagnosis and treatment of post-traumatic stress disorders [1, 2]. Excessive and persistent cortisol secretion is known to cause serious side effects, such as increased blood pressure, decreased immune response, chronic fatigue, and insomnia [3]. A recent study showed that high cortisol concentrations in patients with coronavirus disease-2019 are associated with increased mortality and a decreased average survival rate [4]. Thus, quantification of cortisol levels is essential for evaluating and managing mental, physical, and physiological stresses. Among the various methods of monitoring cortisol secretion, point-of-care (POC) testing is desirable because cortisol levels change depending on the surrounding environment and behavior of individuals during the day.

Biosensors based on a lateral flow immunoassay (LFI) are promising candidates for POC diagnosis, as they are easy to use by non-specialized individuals and enable rapid target detection without specialized equipment [5-7]. However, most LFI-based biosensors show low sensitivity and cannot provide quantitative results [8]. To overcome these limitations, several signal amplification methods have been evaluated for use with LFI, including the application of enzymes for additional color-developing reactions [9-12] and application of nanoparticles for fluorescent signal enhancement [13-16]. Although these methods can improve the detection limit of the assay, several limitations prevent operational POC diagnosis, including stability issues and the requirement for extra equipment to detect fluorescent signals.

In 2011, Wei *et al.* suggested using metal ions to enhance the sensitivity of LFI [17]. After performing a general immunoassay using metal nanoparticles, Wei *et al.* amplified the signal by inducing metal nanoparticle growth through addition of a metal ion and reducing agent. Specifically, detection

antibody/gold nanoparticle (AuNP) conjugates bound to the analyte were immobilized on a detection antibody-loaded membrane and then reacted with gold ions and a reducing agent in the reaction buffer to enlarge the metal nanoparticle, leading to a superior signal amplification effect. Since then, several LFIs with gold ion signal amplification have been developed [18-23].

However, these systems exhibit several disadvantages. For example, the reagents used for signal amplification must be added manually, and there is a long lag-time (≥ 10 min) after the immunoassay, rendering the system unsuitable for POC diagnosis (time lag-related complexity is not easily managed by nonexperts while performing an immunoassay). To make the LFI more user-friendly, various fluid flow control methods (also known as delayed-release of reagent-contained solution for enhancement methods) have been employed, including the use of wax filler [24], dissolvable bridge [25], sugar barrier [26], delayed-release membrane [27], valving system [28], pressed-physical barrier [29, 30], pullulan film [31], water-soluble nanofibers [32], wax barrier [33], and wax-printed channels [34]. Nevertheless, the complexity of these methods limits their application, reproducibility, and large-scale production. Thus, a highly sensitive LFI method based on a simple structure to delay the release of the signal-amplification solution, without long lag time procedures, is warranted.

We previously reported the development of a trap LFI sensor ("trapLFI") for sensitively and quantitatively detecting cortisol in saliva [35]. The trapLFI sensor has deletion and detection zones loaded with target-protein conjugate and an anti-mouse IgG antibody, respectively. By lowering the amount of AuNP-antibody conjugate, conjugates that do not react with the free analytes are completely "trapped" in the deletion zone, whereas only the AuNP-antibody conjugate associated with the free analyte can escape from the deletion zone and react with immobilized secondary antibody on the detection zone, leading to a signal display. Although the platform showed high sensitivity and selectivity for detecting cortisol via ratio calculation of the quantified deletion and detection zones, an enhancement step with a long lag-time procedure was employed to amplify the signal. Thus, a trapLFI sensor integrating delayed release in the signal amplification step could be highly advantageous for POC cortisol detection.

Herein, we report the development of an advanced trapLFI sensor, the " α -trapLFI" sensor. In this sensor, automatic signal amplification is achieved using water-soluble tape, and long lag-time procedures are not required for signal amplification. Polyvinyl alcohol (PVA) tape, which is commercially available, has excellent adhesive properties and is water-soluble. This tape is placed between the signal amplification pad and membrane where the immune reaction occurs. When the PVA tape is exposed to water, it dissolves and the structure collapses, allowing ions to pass through. This process takes approximately 5 min, facilitating the time delay between the immune reaction and signal amplification reaction. These properties enabled signal amplification in our sensor system. Although the sample solution and signal amplification solution were applied consecutively, the PVA tape between the conjugate pad and membrane for the immune reaction delayed the flow of signal-amplification solution through the membrane of the LFI. As a consequence, signal amplification only occurred after the immunoassay had proceeded for a set amount of time. This method can "simultaneously and automatically" induce signal amplification, which allows for target detection with high sensitivity in a user-friendly manner. The PVA

tape-insulated LFI was successfully applied to detect cortisol in bodily fluids, such as saliva, serum, and urine, with high sensitivities. Our novel α -trapLFI provides a new benchmark for simple, rapid, and portable POC diagnosis of stress disorders.

Results And Discussion

A. Principle of α -trapLFI sensor

The proposed α -trapLFI sensor for cortisol analysis is shown schematically in Scheme 1. The sensor consists of a sample pad, conjugate pad, signal amplification pad, PVA tape, absorbent pad, and detection/deletion zones on the membrane. As the devised method involves the enlargement of AuNPs for signal amplification [36] (Scheme S1), a signal amplification pad loaded with gold ions required for signal amplification is placed on the PVA tape, creating a delayed-release effect.

When the sample and signal amplification solutions were consecutively injected into the strip (without a long time lag between injections), the PVA tape delayed the release of the gold ions loaded onto the signal amplification pad, whereas the target molecule in the sample solution and AuNP–antibody conjugate moved towards and reached the detection zone. After approximately 5 min of exposure to water, subsequent dissolution of the PVA tape allowed the gold ions to flow from the signal amplification pad to the membrane, resulting in automatic sequential signal amplification (Movie S1). As a consequence, although conventional LFI is detectable at a high concentration of target only, α -trapLFI is detectable at both low and high concentrations of target by automatic signal amplification (Scheme 1b).

B. Characteristics of α -trapLFI sensor

1. *Structure of PVA tape*

To investigate the structures of dry and wetted PVA tape, we performed field emission scanning electron microscopy and light microscopy. The surface of the PVA tape before exposure to water was polished (Fig. 1a) and rigid (Fig. 1c). In contrast, the surface of the PVA tape was pitted after exposure to water (Fig. 1b), and sizable holes were observed (Fig. 1d). These results confirm that the PVA tape dissolves upon exposure to water, and that its rigid structure is perturbed to produce a porous structure.

2. *Delayed-release effect of PVA tape*

To confirm the delayed-release effect, the effect of PVA tape was compared to that of conventional membranes, including NC membranes, plasma separation (Vivid) membranes, and asymmetric (MMM) membranes. Thus, the various types of membranes were applied to the LFI sensor for comparative testing. Bare AuNP solution was immobilized on each membrane ($1\ \mu\text{L}\cdot\text{cm}^{-1}$), and then a signal amplification pad ($8.2\ \mu\text{L}\cdot 3.8\ \text{mm}^{-2}$) loaded with gold(III) chloride trihydrate solution, 0.1 M in 10 mM

phosphate buffer (pH 7.4), was applied onto each membrane. To initiate LFI, a sample solution, 10% DMSO/PBS (v/v), and signal amplification solution were consecutively loaded. Using a bare membrane, NC membrane, Vivid membrane, and MMM membrane, the signal was rapidly amplified within 1 min. In contrast, a delayed-release effect was observed when using PVA tape, and the signal was only amplified after approximately 5 min (Fig. 2). These results confirm that the PVA tape induced a superior delayed-release effect compared to the other membranes, and thus could be appropriately used as a signal amplification platform based on its delayed-release property.

The PVA tape was also compared to a general double-sided tape to further investigate the observed ion diffusion effect. Using the prepared strips described in Section 2.D the double-sided and PVA tapes were placed between the signal amplification pad and membrane. The sample and amplification solutions were applied to the sample and signal amplification pads, respectively. No signal appeared over time on the strip with the general double-sided tape, whereas the signal intensity increased on the strip with the PVA tape at approximately 5–7 min after loading the sample and amplification solutions (Fig. S1). A comparison of the resulting signal intensities (Fig. S1) confirmed that only the PVA tape permitted ion diffusion.

The ion diffusion permissible property of PVA tape may be attributed to its water-solubility. To confirm this, the general double-sided and PVA tapes (size $3.8 \times 8.0 \text{ mm}^2$) were exposed to water, and the weight was measured at different time points. Each tape was completely dried in an oven to reduce deviation, and the weight was measured using an ultra-microbalance. The general double-sided tape showed no change in weight after exposure to water. However, the weight of the PVA tape decreased with the time of exposure to water, decreasing by 19.5% after 5 min (Fig. S1). Thus, the ion diffusion permissible property of PVA tape was attributed to water dissolution.

3. *Signal amplification through PVA tape*

To verify that the signal could be amplified via the PVA tape, the signal intensities of a conventional LFI and PVA tape-applied LFI were compared. Anti-mouse IgG antibody was immobilized on a membrane ($1 \mu\text{L} \cdot \text{cm}^{-1}$), and AuNP–antibody conjugate (1-fold) was applied on a conjugate pad. For the PVA tape-applied LFI, a PVA tape ($3.8 \times 8.0 \text{ mm}^2$) and gold(III) chloride trihydrate solution-loaded signal amplification pad ($8.2 \mu\text{L} \cdot 3.8 \text{ mm}^2$) were applied to the membrane. Although only the sample solution was loaded for conventional LFI, the sample solution and signal amplification solution were consecutively loaded for PVA tape-applied LFI.

Comparison of the signal intensities obtained using conventional LFI and PVA tape-applied LFI is shown in Figure 3. In PVA tape-applied LFI, the signal was amplified approximately 5–6 min after sample and amplification solution injection, and a higher signal (compared with conventional LFI) was achieved (Fig. 3a and 3b). As shown in Figure 3b, the signal achieved using PVA tape-applied LFI was approximately 6.9-fold higher than that achieved via conventional LFI.

4. *α -trapLFI sensor for detecting small molecules*

In a previous study of the trapLFI sensor developed by our group ³⁵, we found that detection of cortisol with high sensitivity could be achieved with deletion and detection zones, rather than the test and control zones used in typical LFI. Moreover, because we used a small amount of AuNP–antibody conjugate, the conjugate could not react with antigen "trapped" in the deletion zone, minimizing the false-positive signal in the detection zone. However, the amount of conjugate could not be reduced further for the trapLFI sensor, as the signal needed to be detectable to measure its intensity. Here, our previous trapLFI sensor was adopted as a platform to which the PVA tape-mediated delayed release method could be appended. In our modified sensor, the advanced trap-LFI (α -trapLFI), the advantages of the trapLFI sensor are combined with signal amplification by delayed release of gold ions to further increase sensitivity. As shown in Figure 3, although no signal was detected in a conventional LFI using the lowest amount of AuNP–antibody conjugate (Fig. 4 left), a signal was detected using the α -trapLFI sensor because of signal amplification by delayed-release gold ions and a reducing agent (Fig. 4 middle). In the α -trapLFI platform, the delayed-release effect induced by PVA tape was essential. Thus, in the absence of PVA tape, the gold ions were released in advance, and the signal was not detected because the ions reacted earlier with AuNP–antibody conjugate that had not undergone an immunoreaction (Fig. 4 right). Notably, the PVA tape was confirmed to have a negligible effect on the immunoreaction (Fig. S2).

Comparison of the conventional LFI sensor and α -trapLFI sensor for cortisol analysis was also conducted to evaluate sensitivity (Fig. 5). The target-BSA conjugate and anti-mouse IgG antibody was immobilized on a membrane ($1\ \mu\text{L}\cdot\text{cm}^{-1}$), and then the AuNP-antibody conjugate (1-fold for conventional LFI; 0.001-fold for α -trapLFI) was applied to a conjugate pad. For the α -trapLFI sensor, a signal amplification pad, PVA tape, and signal amplification solution were applied, and the assay was carried out as described above. The results demonstrate that the conventional LFI shows the signal changes from about 1 ng/mL, whereas the α -trapLFI shows a different signal even at a lower concentration of cortisol. Thus, cortisol was detected with high sensitivity over a broad-detection range with α -trapLFI sensor.

C. Sensitivity and selectivity of the α -trapLFI sensor

Our α -trapLFI sensor with simultaneous signal amplification developed using the delayed release effect of PVA tape was successfully optimized. Next, cortisol was detected in buffer, human saliva, human serum, and human urine, to evaluate the performance of the new sensor. The sensitivity of the α -trapLFI sensor was tested using cortisol standard solutions at various concentrations in DMSO, saliva, serum, and urine/buffer (10:90, v/v). The colorimetric signal intensities were determined at 15 min after consecutively adding the loading sample and signal amplification solutions, and the results were obtained by calculating the ratios of signals observed in the deletion and detection zones. As shown in Figure 6 and Figure S3, the α -trapLFI sensor was highly sensitive to cortisol in various human bodily fluids. The limit of detection (LOD) for the α -trapLFI was calculated as the sum of the blank signal plus 3 $\times$ the standard deviation of a control cortisol sample. LOD values for cortisol in DMSO, saliva, serum, and

urine were 6.9, 9.1, 14.9, and 15 pg•mL⁻¹, respectively, which were better than those previously reported for our trapLFI sensor (LOD 9.9 pg•mL⁻¹)³⁵.

Table 1 shows a comparison of the characteristics of the α -trapLFI sensor with previously reported LFI. The α -trapLFI sensor demonstrated a better LOD with a broader detection range compared to the other sensors, and it could successfully detect cortisol in various human bodily fluids.

TABLE 1

Comparison of the Specificities of *alpha*-trapLFI with Previously Developed Lateral Flow Immunoassay (LFI)

Ref.	Limit of detection (LOD)	Assay time	Detection range	Enhance-ment method	Additional steps for signal amplification	Applied human fluids
[37]	0.3 ng•mL ⁻¹	25 min	~60 ng•mL ⁻¹	enzyme enhancement	addition of washing and enhancing solutions 10–15 min after sample solution injection	Saliva
[38]	0.1 ng•mL ⁻¹	15 min	~100 ng•mL ⁻¹	gold ion enhancement	addition of enhancing solutions 15 min after sample solution injection	Saliva
[39]	2.5 ng•mL ⁻¹	15 min	~50 ng•mL ⁻¹	-	-	Saliva
[35]	9.9 pg•mL ⁻¹	15 min	~100 ng•mL ⁻¹	enzyme enhancement	addition of enhancing solution 10 min after sample solution injection	Saliva
α -trapLFI	6.9 pg•mL ⁻¹	10–15 min	~1000 ng•mL ⁻¹	gold ion enhancement	addition of enhancing solution and sample solution consecutively	saliva, serum, urine

Cross-reactivity of the α -trapLFI sensor was evaluated by exposure to other steroid hormones (Fig. 7a), including cortisone, CORT, P4, and 17-OHP4, at 10 ng•mL⁻¹. The results of these assays are shown in Figure 7b and Figure S4. The cortisol specificity of the α -trapLFI sensor was high, and no cross-reactivity with the other steroid hormones was observed. Thus, the α -trapLFI sensor can be used as a rapid and simple detection method for cortisol diagnosis.

D. Analysis of cortisol in human saliva

Twenty human saliva samples were tested using the α -trapLFI sensor. The values obtained from the saliva test samples analyzed on the α -trapLFI sensor were compared with those analyzed by ELISA (Fig. 8), revealing the following correlation: $y = -0.0661x + 0.5114$ ($R^2 = 0.90$). The above correlation indicated

reliable agreement with the ELISA results. Hence, the α -trapLFI sensor can be used as a highly reliable sensor platform for detecting cortisol in saliva samples. Together, our results support that our α -trapLFI sensor is a simple and practical platform for cortisol detection in human bodily fluids with high sensitivity.

Conclusion

In this study, we developed an α -trapLFI sensor for simultaneous signal amplification without long time lag procedures, as a highly sensitive LFI platform for cortisol detection based on a trapLFI sensor previously reported by our group. The application of PVA tape, which exhibits material delayed release effect, makes it easy to amplify the signal without additional steps. In addition, the performance of the α -trapLFI sensor was superior because the conjugates that did not react with the target were 'trapped', resulting in no signal in the detection zones when there was no target.; this is an inherent advantage of a trapLFI sensor. The performance of proposed sensor was verified using human bodily fluids including saliva, serum, and urine, and the correlation of the assay with the ELISA results was $R^2 = 0.90$ in the saliva samples. Particularly, the PVA tape used for simultaneous signal amplification is inexpensive and easy to procure and apply to signal amplification sensors; it also has a simple structure. Therefore, it can be easily applied to all LFI platforms to detect not only small molecules, but also various other targets.

Methods

A. Chemicals and reagents

Free cortisol, bovine serum albumin (BSA), cortisol-BSA, and surfactant 10G (S10G) were purchased from Fitzgerald Industries International (Acton, MA, USA). The anti-cortisol antibody was purchased from Abcam (Cambridge, UK), and the nitrocellulose (NC) membrane (FF80HP) was purchased from GE Healthcare (Chicago, IL, USA). Salivette[®] Cortisol, which was used to collect saliva samples, was purchased from Sarstedt (Numbrecht, Germany). The absorbent pad (Grade222), conjugate pad (6613), signal amplification pad (8964), and sample pads (Grade222) were purchased from Ahlstrom-Munksjö (Helsinki, Finland). The NC membrane without backing (HF180UBXSS) was purchased from EMD Millipore (Burlington, MA, USA), and the Vivid plasma separation-GX membrane (T9EXPPA0200S00X) and asymmetric membrane - MMM0.45 (T9PA045W000M) were purchased from Pall Corporation (Port Washington, NY, USA). PVA tape (126057) and a general double-sided tape (DTS-310) were purchased from Buho, Inc. (Daegu, Korea) and DUCKSUNG HITECH, Inc. (Gwangju, Korea), respectively. Gold nanoparticles (EM.GC15) and the TMB substrate reagent set (555214) were purchased from BB International (Cardiff, UK) and BD Biosciences (San Jose, CA, USA), respectively. The ELISA microplate (655061) was purchased from Greiner Bio-One International GmbH (Mühlkreis, Austria). Phosphate-buffered saline (PBS), sodium phosphate (monobasic), and sodium phosphate (dibasic) were purchased from Biosesang Co. (Sunghnam, Korea). Protein saver was purchased from TOYOBIO Co., Ltd. (Osaka, Japan). An anti-mouse IgG antibody produced in goat, boric acid, sodium tetraborate decahydrate,

trisodium citrate dihydrate, citric acid, polyvinylpyrrolidone (10K) (PVP 10K), D-(+)-trehalose dehydrate, Triton™ X-100, Tween 20, dimethyl sulfoxide (DMSO), gold(III) chloride trihydrate, hydroxylamine hydrochloride, sulfuric acid, cortisone, corticosterone (CORT), progesterone (P4), 17 α -hydroxyprogesterone (17-OHP4), human serum, and all other chemicals were purchased from Sigma-Aldrich (St. Louis, MO, USA). Normal human saliva fluid and normal human urine fluid were purchased from MyBioSource (San Diego, CA, USA). Water was purified using a lab water purification system (PURELAB® Option-Q, ELGA LabWater Ltd., Lane End, UK).

B. Characterization of PVA tape

The structure of PVA tape was characterized by field emission scanning electron microscopy and light microscopy. The PVA tape characteristics were also compared to a general double-sided tape to determine its delayed-release effect. AuNP colloid (15 nm in diameter, 1 $\times$ AuNP, $\lambda_{\max}$ O.D. = 1.0) was immobilized on an NC membrane (1 $\mu\text{L}\cdot\text{cm}^{-1}$), and then gold(III) chloride trihydrate (0.1 M) was applied on the signal amplification pad. The general double-sided and PVA tapes were attached beneath the signal amplification pad. The sample solution, 10% (v/v) DMSO/PBS, and signal amplification solution, 10 mM hydroxylamine in 10 mM citrate (pH 3.8), were applied onto the sample and signal amplification pads, respectively. The weight of the PVA tape was measured at five time points after loading the buffer. The PVA and general double-sided tapes (size, 3.8 $\times$ 8.0 mm²) were placed in an e-tube containing citrate buffer (10 mM, pH 3.8) and vigorously vortexed. After 15 min, the tapes were dried completely, and their weights were measured using an ultra-microbalance.

Effective delayed release from the PVA tape was confirmed as described above with different membranes. To determine whether the PVA tape impacts the immunoassay, anti-mouse IgG antibodies (1 mg $\cdot\text{mL}^{-1}$) were immobilized on an NC membrane (1 $\mu\text{L}\cdot\text{cm}^{-1}$), and then AuNP–antibody conjugate (three-fold) was applied on a conjugate pad. The PVA tape was attached beneath the signal amplification pad, and the sample and signal amplification solutions were applied to the sample and signal amplification pads, respectively.

C. Preparation of AuNP conjugate

The anti-cortisol antibody (5 μL , 2 mg $\cdot\text{mL}^{-1}$) was added to a mixture of 1 mL of AuNP colloid (15 nm in diameter, 1 $\times$ AuNP, $\lambda_{\max}$ O.D. = 1.0) and 100 μL of 0.1 M borate buffer (pH 8.5). After incubation for 30 min at 21°C, 10 μL of protein saver (100 mg $\cdot\text{mL}^{-1}$ in PBS) was added to the mixture to block non-specific binding sites on the AuNP surface. After incubation for 1 h at 21°C, the mixture was centrifuged at 15 602 $\times g$ for 15 min at 10°C. The supernatant was discarded and the AuNP conjugate was resuspended in 1 mL of 10 mM borate buffer (pH 8.5), and the AuNP conjugate was concentrated by 10-fold and stored in storage buffer composed of 5% (w/v) D-(+)-trehalose, 0.5% (w/v) Protein saver, 0.2% (v/v) Tween 20, and 1% (v/v) Triton™ X-100 in PBS, at 4°C until use.

D. Preparation of α -trapLFI sensor using PVA tape

To prepare an α -trapLFI sensor, 250 $\mu\text{g}\cdot\text{mL}^{-1}$ cortisol–BSA and 1 $\text{mg}\cdot\text{mL}^{-1}$ anti-mouse IgG antibody were immobilized on an NC membrane (1 $\mu\text{L}\cdot\text{cm}^{-1}$) to create two deletion zones and one detection zone, respectively; the distance between each zone was approximately 2.5 mm. An absorbent pad was attached on top of the membrane with a 2-mm overlap. The AuNP conjugate was diluted by 1000-fold with a solution containing 1% (w/v) PVP 10K and 0.5% (w/v) S10G in PBS, and 3.61 μL of diluted AuNP conjugate was applied to the conjugate pad. Next, 8.2 μL of solution containing 0.1 M gold(III) chloride trihydrate in 10 mM phosphate buffer (pH 7.4) was applied to the signal amplification pad

The AuNP conjugate-loaded conjugate pad was attached to the bottom of the prepared strip, and a sample pad was placed underneath it to load the sample. The PVA tape was attached to the membrane at 2-mm above the conjugate pad, and a gold(III) chloride trihydrate-loaded signal amplification pad was placed in the middle of the PVA tape. The assembled membrane was cut into 3.8-mm wide strips and stored in a humidity-controlled chamber (21°C; 23% relative humidity) until use. Before use, the strip was placed in a plastic case provided by Infopia Co., Ltd. (Anyang, Gyeonggi, Korea).

E. Cortisol analysis in buffer, saliva, serum, and urine samples

Cortisol solutions at various concentrations in DMSO, human saliva, serum, and urine were diluted 10-fold in 1× PBS buffer (10% concentration). The signal amplification solutions were prepared at 10 mM in citrate buffer (pH 3.8). Subsequently, 100 μL of the prepared sample solution and 20 μL of the signal amplification solution were applied to the sample and signal amplification pads of the strip, respectively. The strip was then incubated at 21°C for 20 min. In addition, several other steroid hormones were also analyzed with same methods described above. All images were obtained using a ChemiDoc XPS+ imaging system (Bio-Rad, Hercules, CA, USA).

F. Collection of saliva samples

Human saliva samples were obtained from 20 volunteers from the Gwangju Institute of Science and Technology (GIST) in Gwangju, Korea. We collected samples of saliva from healthy individuals between 08:00 and 09:00 using a saliva collecting kit. The participants were asked to refrain from eating, drinking, and performing oral hygiene procedures until the morning of saliva collection. All participants provided informed consent to participate in the study. The study was approved by GIST (Approval number: 20190510-BR-45-08-02). The collected saliva samples were centrifuged at $1274 \times g$ for 5 min to remove cell debris and other proteins, and the collected saliva samples were immediately used for testing. All experiments were performed in accordance with relevant guidelines and regulations.

G. Analysis of cortisol in human saliva samples

The results obtained using the sensor platform were compared to those obtained with an enzyme-linked immunosorbent assay (ELISA). Briefly, ELISA was performed in 96-well microplates. The anti-cortisol antibody (100 μL , 4 $\mu\text{g}\cdot\text{mL}^{-1}$) was added to each well, and the plates were incubated at 4°C overnight. After removing unbound antibodies by washing the wells with PBS-Tween (137 mM NaCl, 0.05% (v/v) Tween 20, pH 7.4), the wells were blocked with 200 μL of BSA solution (0.5 $\text{mg}\cdot\text{mL}^{-1}$ in PBS) at 21°C for 2 h. Next, individual wells were incubated with 100 μL of the diluted saliva samples and cortisol-HRP conjugate (1 $\mu\text{g}\cdot\text{mL}^{-1}$) mixture at 21°C for 5 h. After removing all unbound proteins, the samples were incubated with TMB substrate solution (50 μL) in the dark at 21°C for 10 min. Finally, sulfuric acid (50 μL , 2 N) was added to immediately stop the enzyme reaction. The absorbance of the solution was determined at 450 nm using a spectrometry system (Cytation 5; BioTek Instruments, Winooski, VT, USA).

H. Instruments

A microcentrifuge (smart R17 plus; Hanil Scientific Inc., Gimpo, Korea) was used to separate the AuNP conjugate, and a multi-purpose centrifuge (1580R; LABOGENE Co., Ltd., Lillerød, Denmark) was used to collect saliva samples. A drying oven (KO-100; LK Lab Co., Namyangju, Korea) was used to dry the membranes and all pads following antibody and sample loading. A dispenser (DCI100; Zeta Co., Gunpo, Korea) was used to immobilize samples on a membrane, and a cutting device (TBC-50Ts; Taewoo Co., Namyangju, Korea) was used to cut the membranes. A field emission scanning electron microscope (Hitachi S-4700, Tokyo, Japan) and a light microscope (Olympus BX43, Tokyo, Japan) were used to characterize the structure of PVA tape. An XPR ultra-microbalance (XPR6UD5[®]; Mettler Toledo, Masstron, OH, USA) was used to weigh tapes. All signals from the strips were measured and analyzed with a ChemiDoc XPS+ imaging system and Image lab software (6.0.1; Bio-Rad Laboratories), respectively. The ELISA results were evaluated using a cell imaging multi-mode reader (Cytation 5).

Declarations

DATA AVAILABILITY

All data generated or analyzed during this study are included in this published article (and its Supplementary Information files).

Acknowledgements

This work was supported by BioNano Health-Guard Research Center funded by the Ministry of Science, ICT & Future Planning (MSIP) of Korea as Global Frontier Project (Grant Number H-GUARD_NRF-2019M3A6B2060097); and GIST Young Scientist Research Project grant funded by the GIST in 2021.

Author contributions

M.K. and H.O. planned and designed the experiments, and H.O. performed and optimized the experiments included screening membranes and immunoassay. K.K. supported taking scanning electron microscope images, J.P. and H.J. gave scientific advice, and H.O. wrote the manuscript. All authors reviewed the manuscript.

Competing Interests Statement

The author(s) declare no competing interests.

References

1. Yehuda, R., Teicher, M. H., Trestman, R. L., Levengood, R. A. & Siever, L. J. Cortisol regulation in posttraumatic stress disorder and major depression: a chronobiological analysis. *Biol. Psychiatry***40**, 79-88 (1996).
2. Meewisse, M.-L., Reitsma, J. B., De Vries, G.-J., Gersons, B. P. & Olff, M. Cortisol and post-traumatic stress disorder in adults: systematic review and meta-analysis. *Br. J. Psychiatry* **191**, 387–392 (2007).
3. Karlamangla, A. S., Singer, B. H., McEwen, B. S., Rowe, J. W. & Seeman, T. E. Allostatic load as a predictor of functional decline: MacArthur studies of successful aging. *J. Clin. Epidemiol.***55**, 696–710 (2002).
4. Tan, T. et al. Association between high serum total cortisol concentrations and mortality from COVID-19. *Lancet DiabetesEndocrinol.* **8**, 659–660 (2020).
5. Posthuma-Trumpie, G. A., Korf, J. & van Amerongen, A. Lateral flow (immuno) assay: Its strengths, weaknesses, opportunities and threats. A literature survey. *Anal. Bioanal. Chem.***393**, 569–582 (2008).
6. The Immunoassay Handbook: Theory and Applications of Ligand Binding, ELISA, and Related Techniques (ed. Wild, D.) 89-95 (Elsevier, 2013).
7. Sajid, M., Kawde, A.-N. & Daud, M. Designs, formats and applications of lateral flow assay: A literature review. *J. Saudi Chem. Soc.***19**, 689–705 (2015).
8. Bahadır, E. B. & Sezgintürk, M. K. Lateral flow assays: Principles, designs and labels. *TrendsAnalyt. Chem.***82**, 286–306 (2016).
9. Busa, L. S. A., Maeki, M., Ishida, A., Tani, H. & Tokeshi, M. Simple and sensitive colorimetric assay system for horseradish peroxidase using microfluidic paper-based devices. *Sens. Actuators B Chem.* **236**, 433–441 (2016).
10. Chen, Y et al. A dual-readout chemiluminescent-gold lateral flow test for multiplex and ultrasensitive detection of disease biomarkers in real samples. *Nanoscale***8**, 15205–15212 (2016).

11. Gao, X. et al. An enzyme-amplified lateral flow strip biosensor for visual detection of microRNA-224. *Talanta***146**, 648–654 (2016).
12. Park, S., Kim, H., Paek, S. H., Hong, J. W. & Kim, Y. K. Enzyme-linked immuno-strip biosensor to detect Escherichia coli O157: H7. *Ultramicroscopy***108**, 1348–1351 (2008).
13. Liu, Y. et al. A highly sensitive and flexible magnetic nanoprobe labeled immunochromatographic assay platform for pathogen Vibrio parahaemolyticus. *Int. J. Food Microbiol.***211**, 109–116 (2015).
14. Loynachan, C. N. et al. Platinum nanocatalyst amplification: Redefining the gold standard for lateral flow immunoassays with ultrabroad dynamic range. *ACS Nano***12**, 279–288 (2018).
15. Park, J. M. et al. Chemiluminescence lateral flow immunoassay based on Pt nanoparticle with peroxidase activity. *Anal. Chim. Acta.***853**, 360–367 (2015).
16. Taranova, N. A., Berlina, A. N., Zherdev, A. V. & Dzantiev, B. B. 'Traffic light' immunochromatographic test based on multicolor quantum dots for the simultaneous detection of several antibiotics in milk. *Biosens. Bioelectron.***63**, 255–261 (2015).
17. Yang, W. et al. A colloidal gold probe-based silver enhancement immunochromatographic assay for the rapid detection of abrin-a. *Biosens. Bioelectron.***26**, 3710–3713 (2011).
18. Fridley, G. E., Le, H. & Yager, P. Highly sensitive immunoassay based on controlled rehydration of patterned reagents in a 2-dimensional paper network. *Anal. Chem.* **86**, 6447–6453 (2014).
19. Fu, E. et al. Enhanced sensitivity of lateral flow tests using a two-dimensional paper network format. *Anal. Chem.* **83**, 7941–7946 (2011).
20. Fu, E. et al. Two-dimensional paper network format that enables simple multistep assays for use in low-resource settings in the context of malaria antigen detection. *Anal. Chem.* **84**, 4574–4579 (2012).
21. Razo, S. C. et al. Enlargement of gold nanoparticles for sensitive immunochromatographic diagnostics of potato brown rot. *Sensors (Basel)***19**, 153 (2019).
22. Rodriguez, M. O., Covian, L. B., Garcia, A. C. & Blanco-Lopez, M. C. Silver and gold enhancement methods for lateral flow immunoassays. *Talanta***148**, 272–278 (2016).
23. Rohrman, B. A., Leautaud, V., Molyneux, E. & Richards-Kortum, R. R. A lateral flow assay for quantitative detection of amplified HIV-1 RNA. *PLOSOne***7**, e45611 (2012).
24. Fridley, G. E., Le, H. Q., Fu, E. & Yager, P. Controlled release of dry reagents in porous media for tunable temporal and spatial distribution upon rehydration. *Lab. Chip.***12**, 4321–4327 (2012).
25. Houghtaling, J., Liang, T., Thiessen, G. & Fu, E. Dissolvable bridges for manipulating fluid volumes in paper networks. *Anal. Chem.***85**, 11201–11204 (2012).
26. Lutz, B. et al. Dissolvable fluidic time delays for programming multi-step assays in instrument-free paper diagnostics. *Lab. Chip.***13**, 2840–2847 (2013).
27. Joung, H. A., Oh, Y. K. & Kim, M. G. An automatic enzyme immunoassay based on a chemiluminescent lateral flow immunosensor. *Biosens. Bioelectron.***53**, 330–335 (2014).

28. Toley, B. J. et al. A versatile valving toolkit for automating fluidic operations in paper microfluidic devices. *Lab. Chip.***15**, 1432–1444 (2015).
29. Park, J. & Park, J.-K. Pressed region integrated 3D paper-based microfluidic device that enables vertical flow multistep assays for the detection of C-reactive protein based on programmed reagent loading. *Sens. Actuators B Chem.***246**, 1049–1055 (2017).
30. Park, J., Shin, J. H. & Park, J. K. Pressed paper-based dipstick for detection of foodborne pathogens with multistep reactions. *Anal. Chem.***88**, 3781–3788 (2016).
31. Jahanshahi-Anbuhi, S. et al. Automating multi-step paper-based assays using integrated layering of reagents. *Lab. Chip.***17**, 943–950 (2017).
32. Kim, W., Lee, S. & Jeon, S. Enhanced sensitivity of lateral flow immunoassays by using water-soluble nanofibers and silver-enhancement reactions. *Sens. Actuators B Chem.***273**, 1323–1327 (2018).
33. Ishii, M. et al. Wax-assisted one-step enzyme-linked immunosorbent assay on lateral flow test devices. *Anal. Sci.***34**, 51–56 (2018).
34. Preechakasedkit, P., Siangproh, W., Khongchareonporn, N., Ngamrojanavanich, N. & Chailapakul, O. Development of an automated wax-printed paper-based lateral flow device for alpha-fetoprotein enzyme-linked immunosorbent assay. *Biosens. Bioelectron.***102**, 27–32 (2018).
35. Oh, H. K., Kim, J. W., Kim, J. M. & Kim, M. G. High sensitive and broad-range detection of cortisol in human saliva using a trap lateral flow immunoassay (trapLFI) sensor. *Analyst***143**, 3883–3889 (2018).
36. Ma, Z. & Sui, S. F. Naked-eye sensitive detection of immunoglobulin G by enlargement of Au nanoparticles in vitro. *Angew. Chem. Int. Ed. Engl.* **41**, 2176–2179 (2002).
37. Zangheri, M. et al. A simple and compact smartphone accessory for quantitative chemiluminescence-based lateral flow immunoassay for salivary cortisol detection. *Biosens. Bioelectron.***64**, 63–68 (2015).
38. Yang, J.-S., Shin, J., Choi, S. & Jung, H.-I. Smartphone Diagnostics Unit (SDU) for the assessment of human stress and inflammation level assisted by biomarker ink, fountain pen, and origami holder for strip biosensor. *Sens. Actuators B Chem.* **241**, 80–84 (2017).
39. Rey, E., Jain, A., Abdullah, S., Choudhury, T. & Erickson, D. Personalized stress monitoring: a smartphone-enabled system for quantification of salivary cortisol. *Pers. Ubiquitous Comput.* **22**, 867–877 (2018).

Figures

Figure 1.

Figure 1

Images of polyvinyl alcohol (PVA) tape with (a, b) Scanning electron microscope (SEM) and (c, d) optical microscopy. (a, c) Prior to loading solution; and (b, d) after loading solution. The SEM images are 60° tilted views.

Figure 2.

Figure 2

Delayed-release effect of the polyvinyl alcohol (PVA) tape compared with release using other membranes. (a) Time-dependent images obtained from assays. (b) Intensity plot. The error bars indicate the standard deviation from three independent experiments (p value < 0.05).

Figure 3.

Figure 3

Comparison of signal intensities from a conventional lateral flow immunoassay (LFI) and a polyvinyl alcohol (PVA) tape-applied LFI. (a) Time-dependent images obtained from assays. (b) Intensity plot. Error bars indicate the standard deviation from three independent experiments (p value < 0.05).

Figure 4.

Figure 4

Comparison of conventional lateral flow immunoassay (LFI) and α -trapLFI sensor. (a) Images of conventional LFI (left); α -trapLFI without a polyvinyl alcohol (PVA) tape (middle); and α -trapLFI with PVA tape (right). (*) AuNP-antibody conjugate 0.001-fold; (**) AuNP-antibody conjugate 1-fold.

Figure 5.

Figure 5

Sensitivity comparisons of the conventional lateral flow immunoassay (LFI) sensor and α -trapLFI sensor.

Figure 6.

Figure 6

Sensitivity of α -trapLFI sensor for cortisol determination in human bodily fluids. Cortisol standard in (a) buffer, (b) saliva, (c) serum, and (d) urine. Error bars indicate the standard deviation from three independent experiments (p value < 0.05).

Figure 7.

Figure 7

Cross-reactivity of α -trapLFI sensor in human bodily fluids. (a) Structure of cortisol and other steroid hormones. (b) Selectivity results comparing cortisol and other steroid hormones. Error bars indicate the standard deviation from three independent experiments.

Figure 8.

Figure 8

Human salivary cortisol level measured using conventional ELISA and α -trapLFI sensor.

Figure 9

Schematic illustration of the new α -trapLFI sensor. (a) Signal amplification pad is placed on water-soluble polyvinyl alcohol (PVA) tape. After injecting sample and signal amplification solutions consecutively into the sensor, the detection antibody/gold nanoparticles conjugate passes first (the release of the gold ions is delayed by the PVA tape). After the conjugate passes, the PVA tape dissolves allowing the diffusion of ions, thus increasing the signal of the conjugate. (b) Although conventional LFI sensors can detect the target at high concentrations, the α -trapLFI sensor can detect the target at both low and high concentrations; because lower amounts of AuNP-antibody conjugate are used together with signal amplification.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [SupplementarymovieMinGonKimGIST.mp4](#)
- [supplementarymaterialMinGonKimGIST.docx](#)