Nanosecond-Resolution Photothermal Dynamic Imaging via MHz Digitization and Match Filtering

Jiaze Yin

Boston University <https://orcid.org/0000-0001-6080-3073>

Lu Lan

Boston University

Yi Zhang

Boston University

Hongli Ni

Boston University

Yuying Tan

Department of Biomedical Engineering, Boston University

Meng Zhang

Boston University

Yeran Bai

Boston University <https://orcid.org/0000-0001-7550-6821>

Ji-Xin Cheng (✉ jxcheng@bu.edu)

Boston University <https://orcid.org/0000-0002-5607-6683>

Article

Keywords: photothermal microscopy, photothermal dynamic imaging (PDI), MHz digitization

Posted Date: June 8th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-496390/v1>

License: This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published at Nature Communications on December 1st, 2021. See the published version at <https://doi.org/10.1038/s41467-021-27362-w>.

Nanosecond-Resolution Photothermal Dynamic Imaging via MHz Digitization and Match Filtering

Jiaze Yin^{1,2,+}, Lu Lan^{1,2,+}, Yi Zhang³, Hongli Ni^{1,2}, Yuying Tan⁴, Meng Zhang^{1,2}, Yeran Bai^{1,2}, Ji-Xin Cheng^{1,2,3,4*}

¹Department of Electrical and Computer Engineering, Boston University, Boston, MA 02215, USA.

²Photonics Center, Boston University, Boston, MA 02215, USA.

³Department of Physics, Boston University, Boston, MA 02215, USA.

⁴Department of Biomedical Engineering, Boston University, Boston, MA 02215, USA.

*Authors contributed equally to this work.

*Corresponding author. E-mail: jxcheng@bu.edu

Abstract

Photothermal microscopy has enabled highly sensitive label-free imaging of absorbers, from metallic nanoparticles to chemical bonds. Photothermal signals are conventionally detected via modulation of excitation beam and demodulation of probe beam using lock-in amplifier. While convenient, the wealth of thermal dynamic is not revealed. Here, we present a lock-in free, mid-infrared photothermal dynamic imaging (PDI) system by MHz digitization and match filtering at harmonics of modulation frequency. Thermal-dynamic information is acquired at nanosecond resolution within single pulse excitation. Our method not only increases the imaging speed by two orders of magnitude, but also obtains four-fold enhancement of signal-to-noise ratio over lock-in counterpart, enabling high-throughput metabolism analysis at single-cell level. Moreover, by harnessing the thermal decay difference between water and biomolecules, water background is effectively separated in mid-infrared PDI of living cells. This ability to nondestructively probe chemically specific photothermal dynamics offers a valuable tool to characterize biological and material specimens.

Introduction

Photothermal microscopy is a versatile analytical tool to gauge the optical absorption with extremely high sensitivity¹. Unlike conventional spectroscopic methods that measure the light attenuation, photothermal detection acquires the absorption information via probing the thermal effect by using a second light beam outside the absorption band. Its high sensitivity majorly benefits from the much-reduced background by employing a modulated heating beam and heterodyne detection of a frequency-shifted probe beam with a lock-in amplifier^{2,3}. Using such a detection scheme, shot-noise limited imaging of single gold nanoparticles of 1.4 nm diameter has been demonstrated², and the single-molecule detection limit has been reported⁴. Recently, by employing a mid-infrared (mid-IR) laser as the pump source and visible light as the probe, a label-free vibrational spectroscopic imaging modality termed mid-infrared photothermal (MIP) microscopy has been demonstrated⁵. It obtains the mid-IR absorption contrast from a transient thermal field that is tightly confined in the vicinity of the absorber. By probing such a field with focused visible light, a submicron spatial resolution as good as 300 nm is achieved⁶. This new imaging modality not only enriches the photothermal techniques with enormous molecular fingerprint information, but also overcomes the limitations in conventional mid-IR absorption microscopy and near-field IR approaches. With the capability of submicron-resolution mapping of chemical bonds in an aqueous environment, the mid-infrared photothermal imaging field quickly expands with innovations, including wide-field detection⁷, optical phase detection⁸⁻¹⁰, photoacoustic detection^{11,12}, synergistic integration with Raman spectroscopy¹³. Mid-infrared photothermal microscopy has found broad applications, including non-contact material characterization¹⁴⁻¹⁶, bio-molecular mapping^{13,17}, protein dynamics and aggregation in neurons^{18,19}, bacterial response to antibiotics²⁰, single virus detection²¹ and imaging of metabolism in living cells and other organisms^{5,12,22}.

Yet, despite the success in the development and applications of photothermal microscopy¹, the wealth and valuable information about an object's thermal dynamics along with the transient photothermal process is rarely explored. In principle, lock-in based photothermal heterodyne imaging (PHI) can reveal thermal diffusivity of the medium through phase detection in the lock-in amplifier. This method^{23,24} has enabled various applications, including observing superconducting transition²⁵, tissue differentiation^{24,26}, and revealing membrane interfaces²⁷. However, lock-in demodulation at the fundamental modulation frequency loses all the higher-order harmonics of the photothermal signal and consequently offers a poor temporal resolution. Thus, it is hard to use PHI to interpret a mid-infrared photothermal signal that originates not only from the absorbers but also from the embedding medium. In the temporal domain, a time-gated approach employing a short pulse probe can resolve the dynamics by tuning the delay between the probe and pump pulses^{7,8,28}. Nevertheless, to acquire a complete thermal dynamic spectrum depicting temperature rise and decay at nanosecond resolution would require thousands of repetitive measurements, making it unsuitable for non-repetitive transient processes.

In this work, we demonstrate a mid-IR photothermal dynamic imaging (PDI) system with nanosecond-scale temporal resolution and covering a bandwidth larger than 25 MHz. By using a wideband voltage amplifier and a megahertz digitizer, the photothermal dynamic response to a single IR pulse excitation is recorded. Combined with digital signal processing to filter out the noise outside the fundamental IR modulation frequency and its harmonics, PDI achieves more than a four-fold improvement in signal-to-noise ratio (SNR) than lock-in based PHI. With these improvements, mid-IR PDI allowed high throughput metabolism analysis at the single bacterium

level with an imaging speed nearly two orders of magnitude faster than the lock-in counterpart. Moreover, mid-IR PDI retrieves the transient thermal field evolution and gives out information on the target's physical properties and microenvironment. Using PDI, we mapped the photothermal dynamics of various organelles inside a cancer cell. Distinct from the macroscopic observation on the homogeneous thermal response of tissue or cell as a whole^{24,28}, we show a highly heterogeneous, chemical-dependent thermal environment inside a cell. Furthermore, by harnessing the thermal decay difference between water and biomolecules, cellular components that are buried by the water background in conventional PHI can be differentiated in PDI based on their time-resolved signatures. Collectively, PDI enables direct detection of the transient photothermal process with nanosecond-level temporal resolution. Together with mid-IR excitation, PDI as a new approach allows for nondestructive investigation of a sample's intrinsic chemical and physical properties and is broadly applicable to biology and material science.

Results

Modeling the transient mid-infrared photothermal effect

The photothermal phenomenon originates from the transformation of absorbed photon energy into heat through a nonradiative relaxation²⁸. Under a pulsed laser excitation with a duration shorter than the thermal relaxation time, the absorbed energy forms a localized thermal field around the absorber. This thermal field induces concurrent thermoelastic deformation that modifies the local optical refractive index through local density change, which can be detected as a time-dependent photothermal signal through optical scattering. Compared with PHI detection of nanoparticles²⁹, the MIP thermal dynamics are distinct in two aspects. First, the MIP absorbers, specifically in the living system, cannot be simply modeled as point heat sources in a medium. For instance, targets like lipid droplets, protein aggregation, cytoplasm are bulky. The thermal dynamics are affected by both the absorbers and local medium collectively. Second, given the water absorption in the mid-IR range, both the absorbers and the aqueous medium will experience temperature elevation, which affects the contrast.

In a thermo-conductive environment, the evolution of an absorber's local temperature (T) under a mid-IR pump is governed by the heat transfer equation³⁰:

$$mC_s \frac{dT}{dt} = Q_{\text{abs}} - Q_{\text{diss}} \quad (1)$$

where m and C_s represent the mass and specific heat capacity of the absorber; dT/dt is the temperature change over time; $(Q_{\text{abs}} - Q_{\text{diss}})$ denotes the energy flux, representing the rate difference between the absorbed and dissipated energies. Q_{abs} can be approximated by $I_{\text{IR}}(t)\sigma_{\text{abs}}$, where $I_{\text{IR}}(t)$ represents the incident IR intensity over the IR pulse; σ_{abs} represents the IR absorption cross-section. The heat dissipation follows Newton's law, i.e., Q_{diss} is driven by the temperature gradient and given by $(hS[T(t) - T_0])$, where h and S represent the heat transfer coefficient and effective transfer surface area from specimen to environment, respectively. $(T(t) - T_0)$ is the time-dependent temperature difference between the absorber and ambient environment T_0 .

During the heating process, $T(t)$ can be derived by solving equation (1) with the initial condition $T(0) = T_0$ and assuming constant IR irradiation:

$$T(t) = T_0 + \frac{I_{IR}\sigma_{abs}}{hS} (1 - e^{-\frac{hS}{mC_s}t}) \quad (2)$$

When the IR pulse is finished, Q_{abs} becomes zero. The temperature change is only driven by Q_{diss} . $T(t)$ is then solved as:

$$T(t) = T_0 + (T_{max} - T_0) e^{-\frac{hS}{mC_s}t} \quad (3)$$

Here, T_{max} is the maximal temperature of the absorber after heating ends. From this model, both the heating and cooling process can be described as exponential processes. As an analogy to a resistor-capacitor circuit, mC_s can be considered as the thermal capacitor and $1/hS$ as the thermal resistor³⁰. The interplay of these two items results in a time constant $\tau = mC_s/hS$ and -3dB cutoff frequency at $1/(2\pi\tau)$. A simulation of this transient process is illustrated in **Fig. 1a** middle row, assuming $\tau=400$ ns.

For absorbers with large heat capacity, such as bulky water and large particles, their time constants are expected to be large. Meanwhile, hS is mostly related to the heat transfer capability of the embedding medium and the geometry of absorbers. Therefore, the thermal response is tightly connected to the physical properties of both the absorber and the environment, which can vary greatly in a heterogeneous system like a biological cell.

In the frequency domain, the photothermal signal produced by IR pulses at a repetition rate of f_{IR} can be treated as a Fourier synthesis of equations (2) and (3) and contains harmonics of f_{IR} , as shown in the simulation of absorbers with τ of 400 ns (**Fig. 1a**, bottom row). Harmonics are spread out widely in the frequency domain, with a bandwidth determined by the cutoff frequency. Large τ results in low cutoff frequency and thus weak harmonics, whereas fast decay signals exhibit strong harmonic components. Nevertheless, the lock-in only recovers the amplitude of the first harmonic and misses all higher harmonics, which not only sacrifices the sensitivity³¹, but also causes contrast distortion in a heterogeneous sample with different thermal responses. In the PHI of living cells using lock-in detection, the water background maximizes at the first harmonic, and it overwhelms the weak signal from the small organelles. To address these difficulties, we have developed a PDI system that is able to detect both fundamental and harmonic components in the frequency domain.

PDI instrumentation and data processing

A schematic of our PDI system is shown in **Fig. 1b**. A pulsed mid-IR pump beam is provided by a quantum cascade laser (QCL) and focused on the sample with a reflective objective lens (RL). Counter-propagated probe beam from a continuous-wave 532nm laser is focused by a water immersion objective lens (OL). Backscattered probe photons are collected with the same OL and separated with a 50/50 beam splitter; forward scattered probe photons are collected by the reflective objective and separated by a dichroic mirror. Both forward and backward probe photons are collected and sent to silicon photodiodes (PD) connected with a wideband voltage amplifier. In the following experiments, the photothermal signal was detected on backward scattering for nanoparticle and bacteria samples; on forward scattering for cell samples. The voltage signal is filtered with a low pass filter with a cut-off frequency of 25 MHz and sent to a high-speed digitizer (DAQ) with a sampling rate of 50 million samples per second. Meanwhile, a mercury cadmium

telluride (MCT) detector is placed to monitor the IR pulse, and the signal is digitized by the same DAQ synchronously. The electronics diagram is shown in **Fig. 1c**.

In PDI, we process the acquired data following the procedure shown in **Fig. 1d**. The PDI raw signal (left panel) was acquired from the center of the polymethyl methacrylate (PMMA) particle with a diameter of 500 nm under the IR pump at its absorption peak of 1729 cm^{-1} . Benefited from the broadband detection scheme, a single pulse photothermal signal can be clearly resolved with an SNR over 43 without averaging. During the imaging process, signal acquired per pixel is a segment with multiple pulses. We further performed a match filtering on each segment with a comb-like passband in the frequency domain. In the comb passband, each small window has its center position colocalized at the harmonic frequencies to reject most of the non-modulated noise (middle panel). After filtering and inverse Fourier transform, the photothermal dynamic signal was obtained in the temporal domain with a much-improved SNR of 570 (right panel). By scanning the sample, an X-Y-t stack with each spatial pixel extended in the temporal domain is obtained. The single pulse resolved capability in PDI enables an unprecedented imaging speed with pixel dwell time as short as 50 microseconds, practically limited by the stage scanning speed.

Figure 1. Principle and schematic of photothermal dynamic imaging. (a) The simulated photothermal dynamics of a particle absorber under pulsed IR pump. The absorber has a decay constant of 400 ns; the IR source has a repetition rate of 100 kHz and a pulse duration of 900 ns. (b) PDI setup schematics. A pulsed mid-IR beam is provided by a QCL and focused on the sample with a reflective objective lens (RL). Counter propagated probe beam provided by a continuous-wave visible laser is focused by a water immersion objective lens (OL). Backscattered probe photons are collected with a 50/50 beam splitter (BS), forward scattered probe photons are collected by the reflective objective and separated by a dichroic mirror (DM). Both forward and backward probes are collected and sent to silicon photodiodes (PD) connected with a wideband voltage amplifier. (c) Electronics diagram. A high-speed data acquisition (DAQ) card is used to collect photodiode signal after voltage amplifier (AMP), MCT signal and sample position synchronously. A computer (PC) is used to control and synchronize different instruments. (d) Digital signal processing procedure. The sample was a PMMA particle with 500-nm diameter under 1729 cm^{-1} excitation.

Left panel, raw signal trace; middle panel, frequency domain distribution and pass band windows; right panel, filtered signal. Match filtering is performed on the segment with 200 μ s. Probe power on sample: 25 mW; IR power on sample: 5 mW.

Mid-infrared photothermal dynamic imaging of nanoparticles

To characterize PDI's performance and evaluate its capability of scrutinizing transient thermal dynamics with chemical specificity, we performed PDI on PMMA particles with a nominal diameter of 300 nm. The QCL is set with a repetition rate of 100 kHz and a pulse duration of 600 ns. By tuning the QCL to 1729 cm^{-1} , corresponding to the absorption peak of the C=O bond in PMMA, we acquired the photothermal dynamics signal (**Supplementary Movie 1**). The IR pump starts at $t_1=0$ ns. As in **Fig. 2a**, most of the single particles with absorption reached their highest temperature at $t_2=440$ ns. The photothermal contrast disappears under an off-resonance IR pump at 1600 cm^{-1} (**Fig. 2b**). To compare PDI with PHI, we acquired the intensity image under 1729 cm^{-1} at the same field of view by demodulating at 100kHz using the lock-in (**Fig. 2c**). The line profiles across the indicated particle in **Fig. 2a** and **Fig. 2c** are shown in **Fig. 2d**. For PDI, an SNR of 230 was achieved with a pixel dwell time of 200 μ s, while it was 52 using PHI, corresponding to more than four-fold improvement in detection sensitivity.

Apart from the sensitivity improvement, PDI concurrently revealed dynamics that enable accurate thermal property characterization. The photothermal dynamics of the nanoparticle indicated in **Fig. 2a** under 1729 cm^{-1} excitation is presented as the red curve in **Fig. 2e**. The photothermal induced scattering modulation is known to linear dependent on temperature^{6,32}. The raw signal has a negative sign, indicating diminished backscattering as temperature rising. Here, an inversed waveform is used for better illustration. The temporal resolution in our system is ultimately limited by the photodiode response time, which is a few nanoseconds. With that high temporal resolution, the time-resolved energy flux function $[Q_{\text{abs}}(t) - Q_{\text{diss}}(t)]$ in equation (1) could be directly evaluated by taking the derivative of PDI signal to time, shown as the blue curve in **Fig. 2e**. The thermal dynamic process is composed of three stages. At the beginning of heating (from t_1 to t_2), the heat dissipation was negligible. $Q_{\text{abs}}(t)$ related to energy absorption was dominant, and it resulted in a waveform similar to the IR pulse shape $I_{\text{IR}}(t)$ shown as the black curve in **Fig. 2e**. The temperature kept rising until the heat dissipation term equaled to heat influx; at this point, the net energy flux became zero and the absorber entered a thermal equilibrium state. From t_2 to t_3 , with the IR intensity reduced gradually, the dissipated energy became dominant, and the energy flux function became negative, showing that the absorber entered the cooling stage. After the IR pulse ($>t_3$), the heat flux function only had the heat dissipation term $Q_{\text{diss}}(t)$, shown as an exponential decay.

Subsequently, we measured the thermal decay constant of the pure dissipation process after t_3 . The fitted exponential decay function is shown as the green dashed line in **Fig. 2e**, which has a decay constant of 300 ns. From equation (3), this time constant is given by mC_s/hS . With the material's density ρ and C_s from reference³³, we can determine the heat transfer parameter hS . For this 300 nm PMMA particle on a CaF_2 substrate in air, the heat transfer parameter is derived to be 7.78E-8 W/K. To verify the experimental result, this parameter is independently calculated to be 7.65E-8 W/K using the finite-element method (FEM) (**Supplementary Fig. S1**), which closely matches the experimental measurement. The thermal decay constants of PMMA particles with various sizes were statistically studied by PDI, as shown in **Supplementary Fig. S2**. The particles with diameters of 300 nm and 500 nm showed distinct thermal decay constants of 280 ns and 495 ns, respectively. To verify the spectral fidelity, we tuned the IR wavelength and

recorded the photothermal spectrum of the indicated particle as shown in **Fig. 2f**. When comparing it with the spectrum of PMMA film acquired with FTIR, good consistency, such as the peak ratio, was observed in the entire fingerprint region. Taken together, with the capability of characterizing thermal decay properties with sub-micron spatial resolution and spectral fidelity, PDI offering a new dimension for molecular analysis.

Importantly, the retrieved decay constant of absorbers can be used for optimizing the IR modulation frequency and increasing the imaging speed accordingly. For 500 nm PMMA particles with a decay constant of 495 ns, more than 95% heat drains out at the initial 1500 ns of the cooling process. Thus, modulation frequency higher than 100 kHz used in the previous work⁸ can be employed without influencing the photothermal signal amplitude. **Fig. 2g** depicts PDI on PMMA particles with a diameter of 500 nm at an IR repetition rate of 500 kHz and pulse duration of 400 ns. The pixel dwell time of PDI was reduced to 100 μ s, and single-particle SNR over 261 was achieved. This imaging speed is 30 times faster than the previous report work¹³ while achieving similar SNR. With such improvement, we further explored high-speed PDI of biological specimens, as shown below.

Figure 2. Mid-infrared photothermal dynamic imaging and spectroscopy of PMMA nanoparticles. (a,b) PDI-acquired photothermal intensity image of 300-nm diameter PMMA particles at the absorption peak 1729 cm^{-1} and off-resonant wavelength at 1600 cm^{-1} , respectively. (c) MIP intensity image at 1729 cm^{-1} acquired with the lock-in amplifier in the same field of view. (d) Line profiles across the particle

indicated in (a) and (c). Pixel dwell time: 200 μs ; Step size: 100 nm; Probe power on sample: 30 mW; IR power on sample: 5 mW at 1729 cm^{-1} , 16 mW at 1600 cm^{-1} ; IR repetition rate: 100 kHz; IR pulse duration: 900 ns. Scale bars: 5 μm . **(e)** Photothermal dynamics of the particle indicated in (a). The red curve is the transient PDI signal. The green dash line is the fitted exponential decay function, with a decay constant of 300 ns. The blue curve is the derivative of photothermal intensity over time. The dash-dot line indicates a thermal equilibrium state. The black curve is the synchronously acquired IR pulse. **(f)** Spectrum fidelity. FTIR spectrum of a standard PMMA film (top) and MIP spectrum profile measured at the indicated 300 nm PMMA particle (bottom). Acquisition speed: 50 cm^{-1} per second. **(g)** Large area PDI of 500-nm diameter PMMA particles at 1729 cm^{-1} . Pixel dwell time: 100 μs ; Step size: 150 nm; Probe power on sample: 30 mW; IR power on sample: 11 mW; IR repetition rate: 500 kHz; IR pulse duration: 400 ns. Scale bar: 20 μm .

Figure 3. PDI enables high-throughput metabolic imaging at a single bacterium level. **(a)** Photothermal dynamics of single *E. coli* under 1658 cm^{-1} excitation. **(b)** High-speed PDI of *E. coli* at the protein Amide I band. Pixel dwell time: 50 μs ; Scale bars: 20 μm . Probe power on sample: 10 mW; IR power on sample: 35 mW at 1658 cm^{-1} ; IR repetition rate: 1.0 MHz; IR pulse duration: 300 ns. **(c)** Line profile of indicated line in (b). **(d)** PDI of ^{12}C glucose-treated *E. coli* at Amide I band. **(e)** PDI of ^{13}C glucose-treated *E. coli* at Amide I band. Amide I peak is shifted from 1658 cm^{-1} to 1615 cm^{-1} due to the isotope substitution. Scale bars: 5 μm . **(f)** MIP spectra acquired from ^{12}C and ^{13}C glucose-treated *E. coli*.

PDI enables high-speed metabolic analysis at a single bacterium level

Given the cell-to-cell heterogeneity, high-speed imaging is essential to understand cell functions and pathways³⁴. For example, sensitive single bacterium metabolism analysis plays an important role in synthetic biology where rare species of interest need to be differentiated from a large population³⁵, or determination of antimicrobial susceptibility from few cells³⁶, such as cells from a small volume of cerebrospinal fluid. Here, we explored PDI imaging of carbohydrate conversion into biomass at a single bacterium level using *E. coli* as a testbed.

We first implemented PDI of *E. coli* at 100 kHz IR modulation frequency, and 200 μ s pixel dwell time. The thermal response of a single *E. coli* is shown in **Fig. 3a**. The decay constant is found to be 280 ns, where 95% heat drains out within 800 ns. Such thermal property allows IR modulation at up to 1 MHz without overheating. At 1 MHz frequency and with a pixel dwell time of 50 μ s, we performed PDI of *E. coli* under 1658 cm^{-1} excitation, corresponding to the Amide I band, as shown in **Fig. 3b**. An SNR over 140 is achieved for a single cell (**Fig. 3c**). This speed is 600 times faster than the previous lock-in based MIP imaging of single *E. coli*⁶. The PDI image covering more than 350 individual cells was acquired within 70 s, showing its high throughput.

Next, we implemented PDI of *E. coli* treated with normal and isotope-labeled glucose. Under the ^{12}C -glucose carbon source, the Amide I absorption peak is localized at 1658 cm^{-1} , as shown in **Fig. 3d**. By substituting the carbon source with ^{13}C -glucose in the growth medium, the Amide I band is shifted to 1615 cm^{-1} , as shown in **Fig. 3e**. The PDI spectral profiles of these two different groups of *E. coli* are shown in **Fig. 3f**. It is clear that the Amide I peak is shifted from 1658 cm^{-1} to 1615 cm^{-1} , and the Amide II peak is shifted from 1546 cm^{-1} to 1533 cm^{-1} . Collectively, these data demonstrate that PDI has high throughput and high sensitivity for single-cell metabolism imaging.

PDI separates biomolecular signal from water background in infrared photothermal imaging of living cells

Due to water absorption in the entire fingerprint region, the culture medium experiences temperature modulation when performing MIP imaging of living cells. Bulky water is known to have a large heat capacity, which helps increase the MIP contrast between bio-samples and background^{5,8} to some extent. However, in the conventional lock-in based MIP system, using signal amplitude alone is not sufficient for separating weak biomolecular signals from water background. By capturing the photothermal dynamics, PDI can be used to separate the water contribution from the organelles' signal based on their distinct thermal responses.

To investigate the transient thermal responses of various organelles inside cells, we performed PDI of U87 brain cancer cells in D_2O PBS. The D_2O PBS was used to maintain cell morphology and reduce the water absorption of mid-IR around 1650 cm^{-1} . By tuning the IR to 1655 cm^{-1} corresponding to the Amide I band, protein-rich contents inside the cell give strong contrast in the photothermal intensity map shown as **Fig. 4a**. Due to the residual water absorption at this wavenumber, the background was observed, but is relatively weak compared to the cells' signal^{5,7}. In this protein map, uniformly distributed protein contents in the cytoplasm and a strong signal from the nucleolus are observed. By tuning the IR to 1750 cm^{-1} corresponding to the C=O band from lipids, individual lipid droplets show a strong signal in **Fig. 4b**. To verify the chemical contents,

we performed photothermal spectroscopy in the fingerprint region at lipid droplet, nucleus, cytoplasm, and medium, respectively, shown in **Fig. 4c**. The spectra at the nucleus and cytoplasm show a strong peak in the Amide I band at 1655 cm^{-1} and a shifted Amide II band at 1450 cm^{-1} due to the deuterium substitution of N-H bond³⁷. The spectrum for lipid droplets shows a strong peak at 1750 cm^{-1} contributed by the C=O bond in esterified lipids. Interestingly, a broad peak centered at 1650 cm^{-1} showed up in the spectra of lipid as well. It matched the result of the intensity map at 1655 cm^{-1} (**Fig. 4a**), where lipid droplets are visible as well. This abnormal contrast at the protein band was found in previously reported scattering-based photothermal imaging but without in-depth explanation^{9,38}.

Here, enabled by PDI, we investigated the origin of this abnormal signal by studying the transient photothermal dynamics of various subcellular components indicated in **Fig. 4a, b**. The results in **Fig. 4d** show distinct thermal responses among organelles. The lipid droplets inside cells show a relatively fast decayed signal with a time constant of 300 ns. Nucleolus and cytoplasm with rich protein contents show slower thermal dynamics on the level of $2.5\text{ }\mu\text{s}$. The water background at 1655 cm^{-1} has the longest decay with a decay constant larger than $5\text{ }\mu\text{s}$, largely because of the large water heat capacity. For a more intuitive illustration, we generated the decay constant map of 1655 cm^{-1} (**Fig. 4e**) and 1750 cm^{-1} (**Fig. 4f**). From the decay map, the background and cellular structures are clearly differentiated for their distinct thermal response. The cytoplasm and nucleus have decay constant on the level of $2.5\text{ }\mu\text{s}$ with 1655 cm^{-1} excitation (**Fig. 4e**). Lipid droplets have decay constants ranging from 150 ns to 500 ns under 1750 cm^{-1} excitation (**Fig. 4f**).

Importantly, those “lipids” seen in the intensity image at 1655 cm^{-1} (**Fig. 4a**) did not show fast dynamics (**Fig. 4e**). Instead, they exhibit a decay constant similar to that of the background. Since the detected signal originates from the modulation of the scattering field, the signal intensity is proportional to $(n_s - n_m)$, where n_s and n_m are the refractive index of organelle and medium, respectively. In the presence of water absorption, $n_m(t)$ becomes time-dependent, and such change results in a pseudo-MIP signal without organelle’s absorption. To better understand the origin of “lipids” seen under 1655 cm^{-1} excitation, underlying thermal dynamics at indicated positions in **Fig. 4e** and **Fig. 4f** are plotted in **Fig. 4g** and **Fig. 4h**, respectively. At 1750 cm^{-1} , lipids’ signal has a fast response and decays in a few hundred nanoseconds. However, the signal at 1655 cm^{-1} from the same regions of interest shows a relatively slow process with a decay constant higher than $5\text{ }\mu\text{s}$, similar to that of the water background, as shown in the dash lines in **Fig. 4g**. In summary, by analysis of photothermal dynamics, we conclude that the 1655 cm^{-1} peak in lipid (**Fig. 4c**) comes from the water background rather than the organelles.

Next, by utilizing the distinct thermal properties between water medium and lipid droplets, we successfully extracted those water-induced signals at lipid droplet positions with the acquired thermal dynamics via a simple algorithm that evaluates their photothermal intensity and decay constants, as shown in **Fig. 4i**. Those strong signals with background type thermal dynamics at 1655 cm^{-1} are displayed in green color, and lipid droplets image acquired under 1750 cm^{-1} excitation is shown in red color. The yellow contrast indicates highly matched colocalization in the merged image. After removing those water-induced pseudo signals at 1655 cm^{-1} from the intensity image, a corrected content map between lipids and proteins is produced, as shown in **Fig. 4j**.

Figure 4. Mid-infrared photothermal dynamic imaging of U87 brain cancer cells. (a) PDI of U87 brain cancer cells at 1655 cm^{-1} Amide I band. (b) PDI of U87 brain cancer cell at 1750 cm^{-1} lipid C=O band. (c) MIP spectra of locations indicated in (a,b). All spectrum intensities are normalized according to the peak of cytoplasm at 1655 cm^{-1} . Acquisition speed: 50 cm^{-1} per second. (d) Thermal dynamics at the positions indicated in (a,b). Black curves are synchronously acquired IR pulse. (e) Decay constant map at 1655 cm^{-1} . (f) Decay constant map at 1750 cm^{-1} . (g) Thermal dynamics of positions indicated in (e) under 1655 cm^{-1} excitation. (h) Thermal dynamics of lipid droplets (LD) indicated in (f) under 1750 cm^{-1} excitation. (i) Merged intensity image at 1750 cm^{-1} (red) with background type signal at 1655 cm^{-1} (green). The yellow color shows a highly overlapped spatial distribution. (j) Merged lipid contents at 1750 cm^{-1} with protein contents at 1655 cm^{-1} after removing the background type signal. Pixel dwell time: 200 μs ; probe power on sample: 20 mW; IR power on sample: 22 mW at 1655 cm^{-1} , 2 mW at 1750 cm^{-1} ; Scale bars: 10 μm .

PDI allows extraction of extremely weak signal from water background

The above photothermal dynamic results show that lipid droplets exhibit a fast decay on the order of a few hundreds of nanoseconds, while the water background is much slower, on the order of a few microseconds. By capturing the high-order harmonic signals, PDI further enabled us to visualize the small lipids (**Fig. 5a**). These small lipids were completely buried in the water background when lock-in detection was used (**Fig. 5b**). To better understand this capability, we performed a Fourier analysis of the photothermal dynamic signals.

The transient photothermal signals of background and lipid droplets at indicated positions in **Fig. 5a** are shown in **Fig. 5c**. Their fundamental components (100 kHz) are acquired by Fourier transform and plotted as orange curves. Those signals of small particles have fast responses and load high-order harmonic components. On the contrary, the water background is majorly localized at the fundamental modulation frequency. Thus, lock-in demodulation at the fundamental frequency minimizes the contrast between lipid droplets and background.

The frequency responses of signals in **Fig. 5c** are shown in **Fig. 5d**. The slow background (blue square) majorly has components in the first and second harmonics. As a comparison, the fast lipid signal (red square) is widely spread out in the frequency domain, with the first harmonic only containing less than one-fifth of the total energy. Consequently, the lipid to background ratio (purple square) increases till the 21st harmonic. At the first harmonic that lock-in demodulates, the contrast is the lowest (**Fig. 5d**). Instead, the photothermal image (**Fig. 5e**) with the 21st harmonic (2.1 MHz) shows clear contrast for small lipids with minimal background. The intensity profiles of the line indicated in **Fig. 5e** at different frequencies are plotted in **Fig. 5f**. At the first harmonic (100 kHz), which is equivalent to lock-in extraction, the lipid signal is hardly resolved from the background signal, while a good signal to background ratio is shown in higher-order harmonics (0.7 MHz to 2.1 MHz).

Figure 5. PDI detection of weak lipid signal from water background. (a) PDI of U87 cancer cell at 1750 cm⁻¹, displayed with log scale color bar. (b) MIP image of the same field of view at 1750 cm⁻¹ acquired by the lock-in method, displayed with log scale color bar. Indicated lipid droplets (LD) in (a) are hardly seen. (c) Normalized thermal dynamics of background (top) and lipid droplet (bottom) at the position indicated in (a). Black curves are synchronously acquired IR pulse. (d) Frequency domain representation of signals in (c). (e) The 21st harmonic amplitude image was acquired by taking the Fourier transform of the PDI stack. (f) Intensity profiles of the line indicated in (e) at different frequencies. PDI pixel dwell time, 200 μs; PHI pixel dwell time 500 μs; Probe power on sample: 20 mW; IR power on sample: 2 mW. Scale bars: 10 μm.

Discussion

We have demonstrated a photothermal dynamic imaging microscope that can sense the transient photothermal modulation with nanosecond temporal resolution. This advanced technology enables concurrent detection of chemically specific IR absorption and physically specific thermal

dynamics at submicron spatial resolution. We retrieved the thermal response of various organelles inside a cell for the first time. Our data shows that cytoplasm, nucleus, and lipid droplets exhibit distinct time-resolved signatures. Based on the time-resolved signatures, PDI enabled the differentiation of small signals from water medium contribution.

Compared with conventional PHI via lock-in amplifier, PDI increases the sensitivity by more than four-fold for low duty cycle photothermal signals. This improvement leverages the broad detection bandwidth for capturing all the harmonics components induced by the short pulse IR pump. Notably, this scheme largely benefits the mid-IR photothermal microscope with a powerful optical parametric oscillator (OPO) source^{7,39}, which has a pulse duration of few nanoseconds and a fixed repetition rate of tens kilohertz. Such a short-pulsed and high peak power excitation source is highly preferred for generating large modulation depth of small objects on a thermo-conductive substrate or in an aqueous environment, where heat dissipation is relatively rapid. In such a case, the photothermal signal has a duty cycle of less than 1% and the lock-in amplifier can only capture a tiny portion of modulation. By capturing all the harmonics, PDI can improve the SNR over one order of magnitude as illustrated in **Supplementary S2**.

When imaging liquid specimens, lock-in detection at the fundamental modulation frequency diminishes the contrast of fast decay signals. This is the case for detecting the small lipid droplets inside cells. These small organelles' signals are buried in the water background in the lock-in method, but are detectable with PDI and give higher contrast in higher-order harmonics. We note that this issue can be mitigated in a lock-in based MIP microscope with tunable modulation frequency IR sources. If knowing the thermal dynamics of a particular type of absorber of interest, one can change the modulation frequency according to their transfer function for optimal detection. For those lipid droplets in water, which quickly decay in few hundreds of nanoseconds, modulation frequency at 500 kHz or higher is more suitable. Yet, the strategy of tuning modulation frequency is no longer valid if absorbers of different decay constants need to be imaged simultaneously.

We note that by retrieving the thermal-physical properties, such as thermal diffusion in a specimen, IR camera-based non-destructive thermal imaging has enabled advanced applications for characterizing materials⁴⁰, including detection of hidden defects like cavities^{41,42}, points of failure in composites, solar cells, semiconductors, and electronics⁴³. Also, with a near-field scanning approach, a time-resolved thermal microscope building on photothermal induced resonance (PTIR) system with 1 MHz bandwidth was reported⁴⁴. By measuring the thermal relaxation time, the thermal conductivity of PMMA is precisely retrieved. The bandwidth of the PTIR system is improved to 25 MHz subsequently by substituting the cantilever with cavity resonator⁴⁵. Yet, restricted by the inherent drawback of tip-based imaging modality of low acquisition speed and physical contact, it is limited to surface characterization.

Beyond the mid-IR photothermal process, PDI can fertilize research related to imaging of short lifetime events that includes but not limits to visible, near-infrared, and other transient physical perturbations^{46,47}. Lock-in free PDI offers nanoseconds temporal resolution that is ultimately limited by the photon detector response. For example, it can be used to study the nonlinear photothermal phenomenon, including the nanobubble generation within a life span of hundreds of nanoseconds^{48,49}, higher-order thermal wave modulation on heat capacity and conductivity^{50,51}; revealing resistive heating profile of nanostructure¹⁴ and detecting photothermal induced pressure wave. Those photothermal processes have a relatively short lifetime due to the rapidly dissipated heat. In biology, this new imaging scheme can be applied to scrutinize intracellular organelles' thermal response together with chemical composition, or monitor transient cell response to fast

perturbation⁴⁶, providing an entirely new perspective to understand the cell machinery. Collectively, the reported PDI system provides a novel far-field label-free imaging tool to scrutinize a sample's chemical composition and physical dynamics simultaneously.

Methods and Materials

Experimental setup.

The mid-IR PDI system (**Fig. 1b**) was built on an inverted microscope frame (IX71, Olympus). A pulsed mid-IR beam, generated by a tunable (from 900 to 1790 cm^{-1}) quantum cascade laser (MIRcat-2400, Daylight Solutions) is used as the pump source. The mid-IR beam passes through a Germanium dichroic mirror and then is focused onto the sample through a gold-coating reflective objective lens (52 $\times$; NA, 0.65; #66589, Edmund Optics). The residual mid-IR reflected by dichroic mirror is monitored with an MCT detector. The pulse repetition rate is tuned from 100 kHz to 1 MHz depending on the decay property of sample; the output pulse duration ranges from 300 to 1000 ns. Counter propagated probe beam provided by a continuous-wave 532 nm laser (Samba 532 nm, Cobolt) is focused by a water immersion objective lens (60 $\times$; NA, 1.2; water immersion; UPlanSApo, Olympus). The probe beam is aligned to be collinear to the mid-IR pump beam to ensure the overlap of the two foci to maintain a good signal level. A scanning piezo stage (Nano-Bio 2200, Mad City Labs) with a minimum pixel dwell time of 50 μs and moving range of 200 μm is used to scan the sample. Backscattered probe photons are collected with a 50/50 beam splitter (BS013, Thorlabs), forward scattered probe photons are collected by the reflective objective and separated by the dichroic mirror. Both forward and backward probes are collected and sent to silicon photodiodes (S3994-01, Hamamatsu) connected with a wideband voltage amplifier with 40 dB gain (DHPVA-101, FEMTO). The voltage signal is AC coupled and filtered with a low pass filter with a cut-off frequency of 25 MHz (BLP-25+, Minicircuit). Amplified signal is sent to a four-channel digitizer (Oscar 14, Gage Applied) installed on a computer, which has a sampling rate of 50 million samples per second and 14-bit A/D resolution. Acquired data is directly streamed to the computer memory for signal processing. During the imaging, the position feedback from scanning stage, mid-IR pulse shape from MCT are acquired synchronously by the same digitizer for image reconstruction and photothermal dynamics analysis. Another set of photodiode (S3994-01, Hamamatsu) connected with laboratory-built resonant circuit is sent to a lock-in amplifier (HF2LI, Zurich Instruments) for conventional MIP detection as comparison experiment and system calibration.

Data acquisition and processing

Acquired raw PDI data per frame is transferred from digitizer to computer memory after sample scanning finished and processed with custom-coded Matlab software. The whole temporal trace is segmented according to the assigned pixel dwell time. Then each segment is filtered in the frequency domain with a comb-like passband with each passing window at the harmonic of pump laser repetition rate (For laser running at 100 kHz, the pass windows are choosing at 100 kHz, 200 kHz, ..., 2 MHz, 2.1 MHz). The window size is defined by the spectrum resolution according to pixel dwell time. The number of passing harmonics decides the thermal dynamic bandwidth and influences the SNR. In this work, we used 16 order harmonics (1.6 MHz) for depicting absorption contrast which give the highest image SNR. On the other hand, for depicting complete photothermal dynamic profile, we used the bandwidth of 25 MHz.

With acquired temporal dynamic behind each pixel, the X-Y-t hyper image is then reconstructed according to the pixel position acquired from the scanning stage. For preserving positioning precision, we perform the reconstruction in an enlarged mesh with four times of acquired pixels by linear interpolation.

Spectra acquisition

The MIP spectra are acquired from PDI signal at different wavelengths and normalized according to the effective pulse energy. To minimize the heat dissipation influence and IR pulse duration variance, the absorption cross-section is evaluated at the initial 300 ns of the photothermal dynamic. The photothermal intensities at $t = 300$ ns at each wavelength are firstly evaluated noted as $S_{raw}(\lambda)$. Subsequently, the effective heating energy at each wavelength is evaluated by integrating the IR pulse shape from 0 ns to 300 ns, which is noted as $Q_{IR}(\lambda)$. The final spectrum is then derived from normalizing $S_{raw}(\lambda)/Q_{IR}(\lambda)$.

Thermal decay fitting

The thermal decay constant is acquired by fitting exponential function on the free decay process (IR pulse is finished) with a custom-coded Matlab program. The nonlinear regression model is used to retrieve the function with the formula: $y = y_0 + A_1 e^{-x/t}$.

The display decay constant map is acquired firstly by fitting the decay constant at every pixel in the PDI stack and then smoothed with a median filter.

PMMA particles

The PMMA particles with a nominate diameter of 300 nm (MMA300, Phosphorex) or with diameter of 500 nm (MMA500, Phosphorex) in solution form were first diluted with deionized water. Around 2 μ L of the solution was then dropped on the surface of a calcium fluoride (CaF_2) substrate with 0.2 mm thickness for imaging. The photothermal signal was detected from backward scattering.

The 500-nm diameter PMMA particles (MMA500, Phosphorex) and 300-nm diameter PMMA particles (MMA300, Phosphorex) mixture sample was prepared by firstly mixing their solution together with 1:1 ratio, and then spreading one droplet of the solution on the surface of a CaF_2 substrate with 0.2 mm thickness for imaging. The photothermal signal is detected from backward scattering.

E. coli

Bacteria were harvested during the log phase. ^{12}C or ^{13}C -glucose was used as the only carbon source in the growth media. The growth media was removed from the bacterial samples by centrifugation and washing four times with 2 mL of deionized water. The bacterial pellet was then suspended in 0.25 mL of water, and 1 μ L of the resulting bacterial suspension was dropped on CaF_2 with 0.2 mm thickness for imaging. The photothermal signal was detected from backward scattering.

U87 cancer cells

The U87 brain cancer cells were cultured on CaF_2 substrate for 24 hours and fixed with formalin. The medium was washed three times by phosphate-buffered saline and replaced by 0.9% NaCl/D₂O solution. It was later sandwiched with another thin calcium fluoride window with 0.2 mm thickness for imaging. The photothermal signal was detected from forward scattering.

Competing interests: JXC declares a financial interest with Photothermal Spectroscopy Corp and Vibronix Inc. The rest of the authors declare no competing interest.

Acknowledgements: The work is supported by R35GM136223 to JXC. L. L., J.Y. and JXC conceived the idea. J.Y. and L.L. carried out modeling, PDI experiments, and analyzed the data. H.N. contributed to theoretical analysis. Y.T. prepared the cancer cells, M.Z. prepared the E. coli samples, Y.B. provided an initial demonstration of metabolism imaging in E. coli. J.Y. drafted the manuscript and all authors contributed to the manuscript writing. The authors thank Celalettin Yurdakul for the helpful discussion.

References

- 1 Adhikari, S. *et al.* Photothermal microscopy: imaging the optical absorption of single nanoparticles and single molecules. *ACS nano* **14**, 16414-16445 (2020).
- 2 Berciaud, S., Cognet, L., Blab, G. A. & Lounis, B. Photothermal heterodyne imaging of individual nonfluorescent nanoclusters and nanocrystals. *Physical review letters* **93**, 257402 (2004).
- 3 Berciaud, S., Cognet, L., Tamarat, P. & Lounis, B. Observation of intrinsic size effects in the optical response of individual gold nanoparticles. *Nano letters* **5**, 515-518 (2005).
- 4 Gaiduk, A., Yorulmaz, M., Ruijgrok, P. & Orrit, M. Room-temperature detection of a single molecule's absorption by photothermal contrast. *Science* **330**, 353-356 (2010).
- 5 Zhang, D. *et al.* Depth-resolved mid-infrared photothermal imaging of living cells and organisms with submicrometer spatial resolution. *Science advances* **2**, e1600521 (2016).
- 6 Li, Z., Aleshire, K., Kuno, M. & Hartland, G. V. Super-resolution far-field infrared imaging by photothermal heterodyne imaging. *The Journal of Physical Chemistry B* **121**, 8838-8846 (2017).
- 7 Bai, Y. *et al.* Ultrafast chemical imaging by widefield photothermal sensing of infrared absorption. *Science advances* **5**, eaav7127 (2019).
- 8 Zhang, D. *et al.* Bond-selective transient phase imaging via sensing of the infrared photothermal effect. *Light: Science & Applications* **8**, 1-12 (2019).
- 9 Tamamitsu, M. *et al.* Label-free biochemical quantitative phase imaging with mid-infrared photothermal effect. *Optica* **7**, 359-366 (2020).
- 10 Schnell, M. *et al.* All-digital histopathology by infrared-optical hybrid microscopy. *Proceedings of the National Academy of Sciences* **117**, 3388-3396 (2020).
- 11 Shi, J. *et al.* High-resolution, high-contrast mid-infrared imaging of fresh biological samples with ultraviolet-localized photoacoustic microscopy. *Nature photonics* **13**, 609-615 (2019).
- 12 Pleitez, M. A. *et al.* Label-free metabolic imaging by mid-infrared optoacoustic microscopy in living cells. *Nature biotechnology* **38**, 293-296 (2020).
- 13 Li, X. *et al.* Fingerprinting a living cell by Raman integrated mid-infrared photothermal microscopy. *Analytical chemistry* **91**, 10750-10756 (2019).
- 14 Aleshire, K. *et al.* Far-field midinfrared superresolution imaging and spectroscopy of single high aspect ratio gold nanowires. *Proceedings of the National Academy of Sciences* **117**, 2288-2293 (2020).
- 15 Chatterjee, R., Pavlovetc, I. M., Aleshire, K., Hartland, G. V. & Kuno, M. Subdiffraction infrared imaging of mixed cation perovskites: Probing local cation heterogeneities. *ACS Energy Letters* **3**, 469-475 (2018).
- 16 Pavlovetc, I. M. *et al.* Suppressing Cation Migration in Triple-Cation Lead Halide Perovskites. *ACS Energy Letters* **5**, 2802-2810 (2020).

- 17 Lima, C., Muhamadali, H., Xu, Y., Kansiz, M. & Goodacre, R. Imaging Isotopically Labeled Bacteria at the Single-Cell Level Using High-Resolution Optical Infrared Photothermal Spectroscopy. *Analytical chemistry* (2021).
- 18 Lim, J. M. *et al.* Cytoplasmic protein imaging with mid-infrared photothermal microscopy: cellular dynamics of live neurons and oligodendrocytes. *The journal of physical chemistry letters* **10**, 2857-2861 (2019).
- 19 Klementieva, O. *et al.* Super-Resolution Infrared Imaging of Polymorphic Amyloid Aggregates Directly in Neurons. *Advanced Science* **7**, 1903004 (2020).
- 20 Xu, J., Li, X., Guo, Z., Huang, W. E. & Cheng, J.-X. Fingerprinting Bacterial Metabolic Response to Erythromycin by Raman-Integrated Mid-Infrared Photothermal Microscopy. *Analytical Chemistry* **92**, 14459-14465 (2020).
- 21 Zhang, Y. *et al.* Vibrational Spectroscopic Detection of a Single Virus by Mid-Infrared Photothermal Microscopy. *Analytical Chemistry* **93**, 4100-4107 (2021).
- 22 Bai, Y., Zhang, D., Li, C., Liu, C. & Cheng, J.-X. Bond-selective imaging of cells by mid-infrared photothermal microscopy in high wavenumber region. *The Journal of Physical Chemistry B* **121**, 10249-10255 (2017).
- 23 Selmkke, M., Heber, A., Braun, M. & Cichos, F. Photothermal single particle microscopy using a single laser beam. *Applied Physics Letters* **105**, 013511 (2014).
- 24 Dada, O. O., Feist, P. E. & Dovichi, N. J. Thermal diffusivity imaging with the thermal lens microscope. *Applied optics* **50**, 6336-6342 (2011).
- 25 McLaren, R. & Dovichi, N. J. Spatially resolved differential resistance of bulk superconductors by laser-induced heating. *Journal of applied physics* **68**, 4882-4884 (1990).
- 26 Burgi, D. S. & Dovichi, N. J. Submicrometer resolution images of absorbance and thermal diffusivity with the photothermal microscope. *Applied optics* **26**, 4665-4669 (1987).
- 27 Samolis, P. D. *et al.* Label-free imaging of fibroblast membrane interfaces and protein signatures with vibrational infrared photothermal and phase signals. *Biomedical optics express* **12**, 303-319 (2021).
- 28 Zharov, V. P. & Lapotko, D. O. Photothermal imaging of nanoparticles and cells. *IEEE Journal of selected topics in Quantum Electronics* **11**, 733-751 (2005).
- 29 Selmkke, M., Braun, M. & Cichos, F. Photothermal single-particle microscopy: detection of a nanolens. *Acs Nano* **6**, 2741-2749 (2012).
- 30 Bergman, T. L., Incropera, F. P., DeWitt, D. P. & Lavine, A. S. *Fundamentals of heat and mass transfer*. (John Wiley & Sons, 2011).
- 31 Muir, R. D., Sullivan, S. Z., Oglesbee, R. A. & Simpson, G. J. Synchronous digitization for high dynamic range lock-in amplification in beam-scanning microscopy. *Review of Scientific Instruments* **85**, 033703 (2014).
- 32 Sullenberger, R., Redmond, S., Crompton, D., Stolyarov, A. & Herzog, W. Spatially-resolved individual particle spectroscopy using photothermal modulation of Mie scattering. *Optics letters* **42**, 203-206 (2017).
- 33 Ali, U., Karim, K. J. B. A. & Buang, N. A. A review of the properties and applications of poly (methyl methacrylate)(PMMA). *Polymer Reviews* **55**, 678-705 (2015).
- 34 Wang, Y. *et al.* Raman–deuterium isotope probing to study metabolic activities of single bacterial cells in human intestinal microbiota. *Microbial biotechnology* **13**, 572-583 (2020).
- 35 Rodionova, M. V. *et al.* Biofuel production: challenges and opportunities. *International Journal of Hydrogen Energy* **42**, 8450-8461 (2017).
- 36 Wiegand, I., Hilpert, K. & Hancock, R. E. Agar and broth dilution methods to determine the minimal inhibitory concentration (MIC) of antimicrobial substances. *Nature protocols* **3**, 163 (2008).

- 37 Lad, M. D., Birembaut, F., Matthew, J. M., Frazier, R. A. & Green, R. J. The adsorbed conformation of globular proteins at the air/water interface. *Physical Chemistry Chemical Physics* **8**, 2179-2186 (2006).
- 38 Spadea, A., Denbigh, J., Lawrence, M. J., Kansiz, M. & Gardner, P. Analysis of Fixed and Live Single Cells Using Optical Photothermal Infrared with Concomitant Raman Spectroscopy. *Analytical Chemistry* (2021).
- 39 Pavlovetc, I. M. *et al.* Infrared photothermal heterodyne imaging: Contrast mechanism and detection limits. *Journal of Applied Physics* **127**, 165101 (2020).
- 40 Vavilov, V. P. & Burleigh, D. D. Review of pulsed thermal NDT: Physical principles, theory and data processing. *Ndt & E International* **73**, 28-52 (2015).
- 41 Poudel, A., Shrestha, S. S., Sandhu, J. S., Chu, T. P. & Pergantis, C. G. Comparison and analysis of Acoustography with other NDE techniques for foreign object inclusion detection in graphite epoxy composites. *Composites Part B: Engineering* **78**, 86-94 (2015).
- 42 Toivanen, J. *et al.* 3D thermal tomography with experimental measurement data. *International Journal of Heat and Mass Transfer* **78**, 1126-1134 (2014).
- 43 Christofferson, J. *et al.* Microscale and nanoscale thermal characterization techniques. *Journal of Electronic Packaging* **130** (2008).
- 44 Katzenmeyer, A. M. *et al.* Mid-infrared spectroscopy beyond the diffraction limit via direct measurement of the photothermal effect. *Nanoscale* **7**, 17637-17641 (2015).
- 45 Chae, J. *et al.* Nanophotonic atomic force microscope transducers enable chemical composition and thermal conductivity measurements at the nanoscale. *Nano letters* **17**, 5587-5594 (2017).
- 46 Jiang, Y. *et al.* Optoacoustic brain stimulation at submillimeter spatial precision. *Nature communications* **11**, 1-9 (2020).
- 47 Ling, T. *et al.* High-speed interferometric imaging reveals dynamics of neuronal deformation during the action potential. *Proceedings of the National Academy of Sciences* **117**, 10278-10285 (2020).
- 48 Nedosekin, D. A., Galanzha, E. I., Dervishi, E., Biris, A. S. & Zharov, V. P. Super-resolution nonlinear photothermal microscopy. *Small* **10**, 135-142 (2014).
- 49 Zharov, V. P. Ultrasharp nonlinear photothermal and photoacoustic resonances and holes beyond the spectral limit. *Nature photonics* **5**, 110-116 (2011).
- 50 Rajakarunanayake, Y. & Wickramasinghe, H. Nonlinear photothermal imaging. *Applied physics letters* **48**, 218-220 (1986).
- 51 He, J., Miyazaki, J., Wang, N., Tsurui, H. & Kobayashi, T. Label-free imaging of melanoma with nonlinear photothermal microscopy. *Optics letters* **40**, 1141-1144 (2015).

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [Supplementaryinformation.docx](#)
- [SupplementaryMovie1.avi](#)
- [SupplementaryMovie2.avi](#)