

Scar quality examination comparing titanium coated suture material and non-coated suture material on flap donor sites in reconstructive surgery

Laura Katharina Berninghausen

Universitätsklinikum Leipzig

Georg Osterhoff

Universitätsklinikum Leipzig

Stefan Langer

Universitätsklinikum Leipzig

Lukas Herbert Kohler (✉ l.h.kohler@gmail.com)

Universitätsklinikum Leipzig <https://orcid.org/0000-0002-7429-3630>

Research article

Keywords: suture, titanium coated, wound healing, scar quality, POSAS score

Posted Date: October 30th, 2020

DOI: <https://doi.org/10.21203/rs.3.rs-52294/v2>

License:  This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published on November 3rd, 2020. See the published version at <https://doi.org/10.1186/s12893-020-00932-3>.

Abstract

Background

Wound healing and scar quality after trauma are subject to impairment through excessive wound healing, chronic wound or even surgical site infections. Optimizing the process of scar formation and skin healing is crucial in virtually all fields of medicine. In this regard, we tested the possible usage and advantages of titanium coated suture material.

METHODS

We performed a prospective observational cohort study including 30 patients who underwent soft tissue reconstruction. One half of the donor flap site was sutured with titanium coated suture material, while the other half was closed with non-coated sutures. Scar quality of the donor flap site was assessed by photographs and POSAS scores on days 2-5, 14, 42, 72 and 180 postoperatively.

RESULTS

No difference between the titanium coated sutures and non-coated sutures was seen in the POSAS assessment, neither for the patient scale at 14, 42, 72 and 180 days, nor for the observer scale on the same dates. Comorbidities like diabetes, chronic renal failure and smoking as well as the BMI of each patient affected the wound healing process to an equal degree on both sides of the suture.

CONCLUSIONS

No difference between the titanium coated and non-titanium-coated suture material was seen in the POSAS assessment in regard to scar quality and wound healing. The titanium-coated suture material can be considered to be equally as effective and safe in all qualities as the non-titanium-coated suture material, even in patients with comorbidities.

Clinical Trial Register

This study is registered at the German Clinical Trials Register (DRKS) under the registration number DRKS00021767. (https://www.drks.de/drks_web/navigate.do?navigationId=trial.HTML&TRIAL_ID=DRKS00021767)

Background

As a result of trauma, either by accident or intent procedure, the skin is subject to a wound healing process resulting in the formation of a mature scar and therefore maintaining the integrity of the skin [1, 2].

Deviation from the physiological wound healing process, such as excessive wound healing, chronic wound or even surgical site infections, can impair the scar quality and the adequate physical function of

the skin [2-4]. A surgical suture's tissue integration and biocompatibility are decisive factors for ideal wound healing, therefore the optimal choice of suture material is still up for discourse [5, 6].

Surgical site infection has been assessed to arise in 5% of all medical procedures in hospitals [7], resulting in pain and discomfort for the patients and prolonged hospitalization and additional costs to the health system [8]. Not only implanted foreign bodies, yet also suture material have to be taken into consideration as a host for biofilm and therefore a potential source of infection [7, 9, 10]. In order to surmount this issue, the coating of suture material not only with chemicals such as triclosan [11-16], yet also with metal legions is not uncommon in medicine. Silver nanoparticle-coatings on silk suture for example have proven to be successful in the prevention of surgical site infections [17-20].

Titanium as a metal legion is renowned for excellent bone to implant bonding [21, 22], high biocompatibility due to low allergic potential [23], resistance to corrosion [24] and limited complications like wound dehiscence, infection and pain [25-27]. Therefore, titanium is commonly used for medical products, whether it be as orthopedic implants [28], titanium clips in cardiology and neurology [29-31], auditory ossicle replacements [32], endoprosthesis

surgery and osteosynthesis in dentistry [24, 33], nickel-titanium wire for closure in cleft lip procedures [25], titanium surgical tacks in gynecology [34] or as titanium coated meshes in abdominoplasty [35-38]. In the latter for example, it has shown to provoke less severe late inflammatory processes, greater tissue maturation and collagen disposition in comparison to a non-titanium-coated polypropylene mesh [37].

These qualities of titanium in medical products raise the question whether titanium, if used as a coating for suture material, could also be used to improve wound healing and scar quality.

Wound healing is an immensely difficult and interference-prone process which needs to be assisted at its best and optimized constantly, especially in patients with comorbidities that could compromise ideal healing and scar formation [1-3].

In this study, we therefore aimed to further evaluate the possible usage and advantages of titanium coated suture material with regard to wound healing, surgical site infections and scar quality in reconstructive and plastic surgery. This was exemplified on flap surgery donor sites.

Methods

Patient collective

We performed a prospective observational cohort study. The materials of this study have been drafted from patients aged 18 and older who had given their informed consent. All patients received soft tissue

reconstruction via free flap surgery at the University Hospital Leipzig from August 2018 to October 2019. Flaps included ALT-, latissimus dorsi-, DIEAP-, parascapular- and gastrocnemius flaps. In total, 30 patients (7 females, 23 males) with a median age of 60 (26 to 92) were included. Baseline data comprised of gender, BMI, comorbidities, duration of hospitalization after surgical intervention and wound healing with Patient and Observer Scar Assessment Score/POSAS (Suppl. 1, 2).

After raising the flap, the donor site wound was sutured continuously and intracutaneously with titanium coated suture material (Seratan[®] 2-0, titanium coated [Serag Wiessner GmbH & Co. KG., Naila, Germany]) on one half and non-titanium-coated suture material (Seralon[®] 2-0, non-titanium-coated [Serag Wiessner GmbH & Co. KG., Naila, Germany]) on the other (Supp. 3). The study was performed in a single blinded design with patients not knowing which half was sutured with titanium coated or non-titanium-coated material. Seratan[®] is priced at 264.94€ per unit (24 pieces) including tax and shipping, Seralon[®] at 168.20€ per unit (24 pieces).

Scar assessment via POSAS scores and clinical photography

On days 2-5 post-surgery, wound visits were performed. On day 10, the stitches were removed and on day 14, POSAS (Patient and Observer Scar Assessment Scale) scores for both sides of the suture were executed by the patient and an observer. Wound visits and POSAS scores were again carried out on day 42, 72 and 180. Photographical documentation was performed continuously during inpatient and outpatient follow up dates (Suppl. 3, 4). The POSAS consists of both a Patient Scale and an Observer Scale. Both scales contain six items that are scored numerically on a ten-step scale with 10 indicating the worst imaginable scar or sensation and 1 corresponding to the situation of normal skin. Together they make up the total score of the scale.

Statistical analysis

Post-test analysis was done using SPSS for Windows V24.0 (IBM, Chicago, IL, USA). All data is reported as frequencies with percentages (%) or mean and standard deviation (SD) with ranges.

A paired t-test was used to detect differences in means between Seratan[®] and Seralon[®] for continuous data. The level of significance was defined as $p < .05$.

Results

In total, 30 patients with a mean age of 60 years (SD 16, range, 26 to 92; 7 females, 23 males) were included into the final analysis. The most frequent flap entity was the anterior lateral thigh flap (n = 23),

followed by latissimus dorsi flaps (n = 3), deep inferior epigastric artery perforator flaps (n = 2), gastrocnemius flaps (n = 1), and parascapular flaps (n = 1). The mean hospital stay after surgery was 13 days (SD 7, range 5 to 37).

The patients' mean Body Mass Index was 27.0 kg/m² (SD 5.9, range 18 to 44). Preexisting comorbidities that are known to affect wound healing were diabetes in 15 patients (50 %), smoking in 13 (43 %) and chronic renal failure in 10 patients (33 %). No patient received immunosuppressive medication.

All patients' donor site wounds were closed with above named suture material in a 50:50 fashion. The mean wound length per patient sutured with titanium-coated material was 11.2 cm (SD 2.6) and 10.9 cm (SD 2.1) for non-coated sutures (p = 0.293). Complications occurred in four patients. This included one suture fistula (Seratan[®]), one superinfected hematoma (Seralon[®]), one superficial wound necrosis (Seralon[®] + Seratan[®]), and one hematoma which had to be revised (Seratan[®]). All other complications were managed in a conservative fashion.

No significant difference between the titanium coated and non-titanium coated sutures was seen in the POSAS assessment, neither for the patient scale at 14 days (p = 0.161), at 42 days (p = 0.787), at 72 days (p = 0.433) or at 180 days (p = 0.293), nor for the observer scale at 14 days (p = 0.787), at 42 days (p = 0.522), at 72 days (p = 0.184) or at 180 days (p = 0.375). However, there is a tendency for a slightly better overall opinion in all follow-ups for the titanium coated material in both groups, patients and observers (Figure 1). When taking into consideration comorbidities like diabetes (Seratan[®]: p = 0.808, Seralon[®]: p = 0.484), chronic renal failure (Seratan[®]: p = 0.297, Seralon[®]: p = 0.244) or smoking

(Seratan[®]: p = 0.459, Seralon[®]: p = 0.562) while comparing the POSAS scores on day 180, no significant difference between the two suture materials could be detected (Figure 2-4). Furthermore, comparing overall patient satisfaction on day 180 on coated and non-coated wound closure sites in correlation to the BMI of each patient did not show any significant difference in scar quality (Seratan[®]: p = 0.541, Seralon[®]: p = 0.647, Figure 5).

Discussion

Wound healing and scar quality are an important part of virtually every field of medicine. The choice of suture material, as an immense factor of impact for the scar formation and healing abilities, is of great importance when it comes to optimizing these processes. This is valid especially in patients with comorbidities which compromise ideal healing and scar formation, like diabetes mellitus, chronic renal failure or the indulgence in nicotine [1-3]. Our findings indicate that titanium coated suture material can be seen as equivalently effective and safe as non-titanium-coated suture material.

In this pilot study with 30 patients, scar quality was assessed on flap surgery donor sites with the help of POSAS score evaluation and photography, while comparing titanium coated sutures with non-titanium-

coated sutures. The coating of medical material with titanium material can be seen as leadoff technique in medicine, therefore scientific literature on this topic is rare. Notwithstanding, our findings can be seen as consistent with the advantages of coating suture material with chemicals [11-16] or metal legions, like the prevention of surgical site infections [17-20]. They were also in accordance with beneficial qualities of titanium in medical products, like excellent bone to implant bonding, high biocompatibility due to low allergic potential, resistance to corrosion and limited complications like wound dehiscence, infection and pain [21-27].

The titanium coated suture material was equivalent to the non-coated-suture material in regard to scar quality and wound healing. Even the presence of comorbidities did not have any impact on these qualities. However, it should be emphasized that high satisfaction in regard to scar quality was achieved in both groups which also increased over the course of time. Additionally, titanium coated sutures presented slightly better results in regard to wound healing than non-titanium-coated sutures. This data suggests the possibility of using titanium coated material in surgical procedures for wound closure in the future. Our findings support the results of Saalabian et al. in which titanium coated suture material showed significantly lower signs of inflammation in small wounds of the hand and forearm[39].

The limitations of this study include its small sample size and the limited variability in surgical sites. A more extensive clinical study and a cost efficiency analysis would have to be planned to not only show clinical evidence, but also practicability in the long term. Nevertheless, titanium coated suture material can be considered as equally effective and safe as non-titanium-coated suture material in regard to wound healing, scar quality and surgical site infection. Combining the advantages of titanium legions in medical products with benefits of coating suture material in order to minimize surgical site infections and therefore receiving optimum wound healing and scar formation was the intention behind the development of the suture material and this study. However, our research suggests that the production of the titanium coated suture material will not be able to deliver superior results to the non-titanium-coated suture material. Considering the higher price as mentioned above, we currently see no reason for a standardized wound closure with titanium coated sutures in our patients.

Conclusion

Titanium coated suture material brings forward as equally adequate results in scar quality and wound healing in flap surgery donor sites as non-titanium-coated suture material. The coating of medical products in general with titanium seems to have a positive impact on wound healing and provides

decreased complications like wound dehiscence and surgical site infection [25-27]. Nevertheless, the utilization of titanium coated suture material on flap surgery donor sites, even in patients with comorbidities or immunocompromization, will not be superior to non-titanium-coated suture material.

Abbreviations

POSAS = Patient and Observer Scar Assessment Scale

et al. = et alii. And others.

GmbH and Co. Kg. = Gesellschaft mit beschränkter Haftung & Compagnie KG (GmbH & Co. KG) is a limited partnership with, typically, the sole general partner being a limited liability company.

Declarations

Ethics approval and consent to participate

The ethics committee of affiliated Leipzig University Hospital of Leipzig University has approved the feasibility of the study prior to study initiation. In this study all patients signed a preoperative informed consent form and agreed to participate in the accompanying scientific research.

Consent for publication

All patients have signed a written informed consent form at least one day before surgery. In this study, there is no any identifying personal or clinical details along with any identifying images to be published.

Availability of data and materials

All data is contained within the manuscript. The datasets used and analyzed during the current study available from the corresponding author on reasonable request.

Competing interests

The authors declare that they have no competing interests.

Funding

Funding for 15,581.00€ was received by Serag Wiessner GmbH & Co. KG in order to conduct this clinical study. The funding company had no influence on the design of the study and collection, analysis, and interpretation of data and in writing the manuscript.

Treatments, material and medical decisions have not and will not be influenced by the results of this study in the future.

Authors' contributions

All authors have approved the final version of the manuscript. LHK and SL designed the research; LB and LHK obtained the clinical data; LHK and SL were responsible for surgical sutures; GO prepared the tables and figures; LB prepared the manuscript; and LHK, GO and SL revised the manuscript.

Acknowledgements

Not applicable.

References

- [1] Potter DA, Veitch D, Johnston GA. Scarring and wound healing. *Br J Hosp Med (Lond)*. 2019;80:C166-c71.
- [2] Wang PH, Huang BS, Horng HC, Yeh CC, Chen YJ. Wound healing. *J Chin Med Assoc*. 2018;81:94-101.
- [3] Gantwerker EA, Hom DB. Skin: histology and physiology of wound healing. *Facial Plast Surg Clin North Am*. 2011;19:441-53.
- [4] Ferguson MW, O'Kane S. Scar-free healing: from embryonic mechanisms to adult therapeutic intervention. *Philos Trans R Soc Lond B Biol Sci*. 2004;359:839-50.
- [5] Lambertz A, Schroder KM, Schob DS, Binnebosel M, Anurov M, Klinge U, Neumann UP, Klink CD. Polyvinylidene Fluoride as a Suture Material: Evaluation of Comet Tail-Like Infiltrate and Foreign Body Granuloma. *Eur Surg Res*. 2015;55:1-11.

- [6] Masini BD, Stinner DJ, Waterman SM, Wenke JC. Bacterial adherence to suture materials. *J Surg Educ.* 2011;68:101-4.
- [7] Kathju S, Nistico L, Tower I, Lasko LA, Stoodley P. Bacterial biofilms on implanted suture material are a cause of surgical site infection. *Surg Infect (Larchmt).* 2014;15:592-600.
- [8] Harrington P. Prevention of surgical site infection. *Nurs Stand.* 2014;28:50-8.
- [9] Juhl AA, Koudahl V, Damsgaard TE. Deep sternal wound infection after open heart surgery—reconstructive options. *Scand Cardiovasc J.* 2012;46:254-61.
- [10] Franco AR, Fernandes EM, Rodrigues MT, Rodrigues FJ, Gomes ME, Leonor IB, Kaplan DL, Reis RL. Antimicrobial coating of spider silk to prevent bacterial attachment on silk surgical sutures. *Acta Biomater.* 2019;99:236-46.
- [11] Henriksen NA, Deerenberg EB, Venclauskas L, Fortelny RH, Garcia-Alamino JM, Miserez M, Muysoms FE. Triclosan-coated sutures and surgical site infection in abdominal surgery: the TRISTAN review, meta-analysis and trial sequential analysis. *Hernia.* 2017;21:833-41.
- [12] Darvin VV, Lobanov DS, Krasnov EA, Gvozdetsky AN. [Assessment of the effectiveness of the suture with triclosan coated in emergency surgery]. *Khirurgiia (Mosk).* 2017;70-75.
- [13] Masood R, Hussain T, Umar M, Azeemullah, Areeb T, Riaz S. In situ development and application of natural coatings on non-absorbable sutures to reduce incision site infections. *J Wound Care.* 2017;26:115-20.
- [14] Guo J, Pan LH, Li YX, Yang XD, Li LQ, Zhang CY, Zhong JH. Efficacy of triclosan-coated sutures for reducing risk of surgical site infection in adults: a meta-analysis of randomized clinical trials. *J Surg Res.* 2016;201:105-17.
- [15] Meghil MM, Rueggeberg F, El-Awady A, Miles B, Tay F, Pashley D, Cutler CW. Novel Coating of Surgical Suture Confers Antimicrobial Activity Against *Porphyromonas gingivalis* and *Enterococcus faecalis*. *J Periodontol.* 2015;86:788-94.
- [16] Chen SY, Chen TM, Dai NT, Fu JP, Chang SC, Deng SC, Chen SG. Do antibacterial-coated sutures reduce wound infection in head and neck cancer reconstruction? *Eur J Surg Oncol.* 2011;37:300-4.
- [17] De Simone S, Gallo AL, Paladini F, Sannino A, Pollini M. Development of silver nano-coatings on silk sutures as a novel approach against surgical infections. *J Mater Sci Mater Med.* 2014;25:2205-14.
- [18] Liu X, Gao P, Du J, Zhao X, Wong KKY. Long-term anti-inflammatory efficacy in intestinal anastomosis in mice using silver nanoparticle-coated suture. *J Pediatr Surg.* 2017;52:2083-87.

- [19] Zhang S, Liu X, Wang H, Peng J, Wong KK. Silver nanoparticle-coated suture effectively reduces inflammation and improves mechanical strength at intestinal anastomosis in mice. *J Pediatr Surg*. 2014;49:606-13.
- [20] Ho CH, Odermatt EK, Berndt I, Tiller JC. Long-term active antimicrobial coatings for surgical sutures based on silver nanoparticles and hyperbranched polylysine. *J Biomater Sci Polym Ed*. 2013;24:1589-600.
- [21] Walsh WR, Bertollo N, Christou C, Schaffner D, Mobbs RJ. Plasma-sprayed titanium coating to polyetheretherketone improves the bone-implant interface. *Spine J*. 2015;15:1041-9.
- [22] Rieger E, Dupret-Bories A, Salou L, Metz-Boutigue MH, Layrolle P, Debry C, Lavallo P, Vrana NE. Controlled implant/soft tissue interaction by nanoscale surface modifications of 3D porous titanium implants. *Nanoscale*. 2015;7:9908-18.
- [23] Wood MM, Warshaw EM. Hypersensitivity reactions to titanium: diagnosis and management. *Dermatitis*. 2015;26:7-25.
- [24] Fukushima A, Mayanagi G, Nakajo K, Sasaki K, Takahashi N. Microbiologically induced corrosive properties of the titanium surface. *J Dent Res*. 2014;93:525-9.
- [25] Li H, Song T. Nickel-Titanium Wire as Suture Material: A New Technique for the Fixation of Skin. *J Craniofac Surg*. 2018;29:e343-e44.
- [26] Otto A, Mehl J, Obopilwe E, Cote M, Lacheta L, Scheiderer B, Imhoff AB, Mazzocca AD, Siebenlist S. Biomechanical Comparison of Onlay Distal Biceps Tendon Repair: All-Suture Anchors Versus Titanium Suture Anchors. *Am J Sports Med*. 2019;47:2478-83.
- [27] Patel SS, Aruni W, Inceoglu S, Akpolat YT, Botimer GD, Cheng WK, Danisa OA. A comparison of *Staphylococcus aureus* biofilm formation on cobalt-chrome and titanium-alloy spinal implants. *J Clin Neurosci*. 2016;31:219-23.
- [28] Zhao L, Hu Y, Xu D, Cai K. Surface functionalization of titanium substrates with chitosan-lauric acid conjugate to enhance osteoblasts functions and inhibit bacteria adhesion. *Colloids Surf B Biointerfaces*. 2014;119:115-25.
- [29] Lee CY, Johnson CA, Jr., Siordia JA, Lehoux JM, Knight PA. Comparison of Automated Titanium Fasteners to Hand-Tied Knots in Open Aortic Valve Replacement. *Innovations (Phila)*. 2018;13:29-34.
- [30] Marin Laut FM, Gomez Cardenas EA, Dormido JR, Molina NM, Lopez Lopez JA. Spinal dural closure without suture: Minimizing the risk of CSF leakage with a flat non-penetrating titanium U-clip. *Neurocirugia (Astur)*. 2019;30:173-78.

- [31] Nguyen KP, Teruya T, Alabi O, Sheng N, Bianchi C, Chiriano J, Dehom S, Abou-Zamzam A. Comparison of Nonpenetrating Titanium Clips versus Continuous Polypropylene Suture in Dialysis Access Creation. *Ann Vasc Surg.* 2016;32:15-9.
- [32] Gostian AO, Pazen D, Luers JC, Huttenbrink KB, Beutner D. Titanium ball joint total ossicular replacement prosthesis—experimental evaluation and midterm clinical results. *Hear Res.* 2013;301:100-4.
- [33] Chanchareonsook N, Tideman H, Lee S, Hollister SJ, Flanagan C, Jansen JA. Mandibular reconstruction with a bioactive-coated cementless Ti6Al4V modular endoprosthesis in *Macaca fascicularis*. *Int J Oral Maxillofac Surg.* 2014;43:758-68.
- [34] Shatkin-Margolis A, Merchant M, Margulies RU, Ramm O. Titanium Surgical Tacks: Are They Safe? Do They Work? *Female Pelvic Med Reconstr Surg.* 2017;23:36-38.
- [35] Khaled I, Priego P, Faisal M, Cuadrado M, Garcia-Moreno F, Ballestero A, Galindo J, Lobo E. Assessment of short-term outcome with TiO2 mesh in laparoscopic repair of large paraesophageal hiatal hernias. *BMC Surg.* 2019;19:156.
- [36] Prassas D, Rolfs TM, Sirothia N, Schumacher FJ. Lightweight Titanium-coated Mesh Versus Standard-Weight Polypropylene Mesh in Totally Extraperitoneal Inguinal Hernia Repair (TEP): A Cohort Analysis. *Surg Laparosc Endosc Percutan Tech.* 2016;26:e113-e16.
- [37] Pereira-Lucena CG, Artigiani-Neto R, Lopes-Filho GJ, Frazao CV, Goldenberg A, Matos D, Linhares MM. Experimental study comparing meshes made of polypropylene, polypropylene + polyglactin and polypropylene + titanium: inflammatory cytokines, histological changes and morphometric analysis of collagen. *Hernia.* 2010;14:299-304.
- [38] Tamme C, Garde N, Klingler A, Hampe C, Wunder R, Kockerling F. Totally extraperitoneal inguinal hernioplasty with titanium-coated lightweight polypropylene mesh: early results. *Surg Endosc.* 2005;19:1125-9.
- [39] Saalabian A, Steinhäuser K, Kneser U, Horch R. Verbesserung der Narbenästhetik durch titanbeschichtetes Nahtmaterial (Seratan) durch optimierte Wundheilung. *Thieme-German: Handchirurgie, Mikrochirurgie, Plastische Chirurgie.* 2013;45 - V0031.

Figures




Figure 1

POSAS Overall Opinion of Observers (a) and Patients (b) on follow-up days 14, 42, 72 and 180 for titanium-coated suture material (blue bar) and non-coated suture material (red bar).




Figure 2

Correlation analysis between POSAS Overall Opinion for observers (a) and patients (b) and diabetes mellitus type II on follow-up days 14, 42, 72 and 180 for titanium-coated suture material and non-coated suture material.




Figure 3

Correlation analysis between POSAS Overall Opinion for observers (a) and patients (b) and chronic renal failure on follow-up days 14, 42, 72 and 180 for titanium-coated suture material and non-coated suture material.




Figure 4

Correlation analysis between POSAS Overall Opinion for observers (a) and patients (b) and smoking on follow-up days 14, 42, 72 and 180 for titanium-coated suture material and non-coated suture material.




Figure 5

Correlation analysis between POSAS Overall Opinion and Body Mass Index (BMI) on follow up day 180 for titanium-coated suture material and non-coated suture material.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [SupplementalMaterial.docx](#)