De-noising of EEG Signals With Shift Based Cycle Spinning on Wave Atoms

M. Purnachandra Rao (✉ purnamailroom@gmail.com)

KKR and KSR Institute of Technology and Sciences

E. Srinivasa Reddy

Nagarjuna University: Acharya Nagarjuna University

Research Article

Keywords: EEG, DWT, Wave Atom (WA), cycle spinning, SNR, MSE

Posted Date: July 30th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-726997/v1>

License: This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published at Soft Computing on March 8th, 2022. See the published version at <https://doi.org/10.1007/s00500-021-06708-y>.

De-noising of EEG Signals with shift based Cycle Spinning on Wave Atoms

M. Purnachandra Rao^{1*}, Dr. E. Srinivasa Reddy²

¹Department of CSE, KKR & KSR Institute of Technology & Sciences, Guntur, AP, India.

²Department of CSE, University College of Engineering & Technology, Acharya Nagarjuna University, Guntur, AP, India.

*Corresponding Author: purnamailroom@gmail.com

Abstract: To predict our brain disorders, EEG signals need to be analyzed. However, most EEG signals were affected by different kinds of noise during their acquisition. So, signal analysis become most difficult due to the contamination of various noise. An appropriate technique is necessary to remove the noise from the signal. Wavelet Transform is one the most widely used technique for removing the noise from EEG signals. Wave atom is one of the new multiscale-multidirectional transforms, which is better than both wavelet as well as curvelet transforms. This wave atom transform has good orientation characteristic by which it preserves the edges in an efficient manner. This paper introduced a new method for denoising of EEG signal by shift-based cycle spinning on wave atom transform. Cycle spinning is a technique can be used to enhance the capability of wave atoms. An original EEG signals from public EEG database were used for this experiment. The results are analysed based on the performance measurements like SNR and MSE. The experimental results show that cycle spinning technique with appropriate shifts could be the better choice to denoise an EEG signals.

Keywords: EEG, DWT, Wave Atom (WA), cycle spinning, SNR, MSE

I. INTRODUCTION

EEG signal is an important brain signal from clinical point of view. The diagnosis of EEG is required to detect the disorders of brain functionality. But various kinds of noises can corrupt the signal. So, the signal diagnosis becomes more difficult because of noise. Therefore, these noises need to be removed from the signal for better diagnosis. Wavelet transform is widely used in the field of signal processing but has some limitations in the processing of signals [1][2]. DWT is one of the widely used technique for denoising EEG signals. Several works propose the use of different wavelets and thresholding techniques of DWT [3]. Thresholding technique that can be applied are soft thresholding, hard thresholding, sigmoid thresholding, etc. [4].

Coifman and Donoho suggested an improvement to basic wavelet thresholding known as cycle spinning [5]. It is well known that wavelet transform is periodically time-variant. By using different shifts of noise signal, one can compute different estimates of unknown signal. The idea behind is that estimates at different shifts have different errors, which will be minimized by averaging them. In wave atoms, oscillation function is sparser than that wavelet, Gabor atoms or curve wave [6]. Wave atoms applies to any local direction of mode and can sparsely spread in the anisotropy mode in the axis direction. By comparing with curve wave, wave atoms can not only capture the vibration mode, but can characterise the pattern through the oscillation. Due to lack of translation invariance, the artificial visual distortion will be introduced at the time of being applied to denoising, especially at the edge part, referred as Pseudo-Gibbs phenomenon. The cycle spinning technique can be used to

avoid this visual distortion well [7][8]. Hence wave atom can produce better results using cycle spinning.

II. LITERATURE SURVEY:

Form this literature survey, cycle spinning with appropriate number of shifts could give better results for removal of noise from EEG signals. Miguel A. Rodriguez-Hernandez [9], proposed an algorithm to implement one of undecimated wavelet transform (UWT) which is cycle spinning (CS) to analyse the signal. He applied a variant of cycle spinning to ultrasonic trace denoising. He assured that his technique produced better results in denoising the ultrasonic signals. Zhang et al. [10], found that denoising of image by thresholding the coefficients may lead to Pseudo-Gibbs phenomena. This is because of lack of translation invariance in wave atom transformation. Therefore, they use cycle spinning method to remove noise and remain edges by controlling Pseudo-Gibbs phenomena efficiently. Alyson et al. [11] introduced a new algorithm recursive cycle spinning. This algorithm repeatedly translates and denoising the signal with wavelet denoising technique translating back recursively and applied convergence properties of projections. Finally, the denoising projections converge to the original signal.

Jiao [12] suggested a method that use wave atoms with cycle spinning for removal of heavy background noise from an image. He concluded that wave atoms are an emerging new multiscale transformation used for removal of noise from images. Eddine and Hassene [13] decided that the wave atom transformation use more parameters like dimension, location, orientation, etc. Because of good orientation characteristic, the wave atom transform is better than wavelet transform and even curvelet transform. Their proposed algorithm is based on hard thresholding that optimizes the performance of wave atom transform. Shruthi and Kumar [14], proposed an algorithm used for denoising the images by decomposing them using wave atom transform and wavelet transforms. The algorithm used constrained least square filtering which is followed by denoising process. In denoising, the filtered image is decomposed into texture and cartoon parts using curvelet transform. Then wave atom transform is used to denoise texture part and wavelet transform is used to denoise cartoon part. They used hard thresholding to prevent edge details. John et al. [15], proposed an algorithm which uses cuckoo search algorithm to optimize coefficients of wave atom transform. The optimization is performed for minimization of error values in the signal. Then they used adaptive thresholding method to denoise the signal.

III. WAVE ATOM TRANSFORM

Wave atom transform is an effective denoising technique for signals containing oscillatory patterns and textures. It follows parabolic scaling relation i.e.,

$$\text{Wavelength} \sim (\text{diameter})^2$$

So, it has better performance in denoising the signals than wavelets and curvelets [16][17].

Different transforms based on wavelet packet like Gabor, Ridgelets, Curvelets and wave atoms are created using two parameters α and β . The parameter α determines whether the decomposition is multistate($\alpha=1$) or not ($\alpha=0$) and the parameter β determines whether the basic elements are localized. If $\beta=1$ then we have poorly directional and if $\beta=0$ then we have fully directional. We can classify the transform based on these parameters α and β , for

wavelets $\alpha=\beta=1$, for ridgelets $\alpha=1$ and $\beta=0$, for Gabor $\alpha=\beta=0$, for curvelets $\alpha=1$ and $\beta=1/2$ and wave atom corresponds to $\alpha=\beta=1/2$. The following figure summarise this classification.

Fig1: Various transforms of wave packet families

Wave atoms are constructed from 1D wave packets. Dimensional wave packets can be represented as [18]

- $\psi_u^j(X)$, $j, m \geq 0$ and $n \in \mathbb{Z}$.
- u represents scale parameter, rotation and translation. For 1D systems this u will be $u=(d, m, n)$ and for 2D systems this u will be $(j, m1, m2, n1, n2)$ where $j, m1, m2, n1, n2$ are integers and index of point $(X_{j,n}, \pm \Omega_{j,m})$ in the phase space.
- $\pm \Omega_{j,m} = \pm \pi 2^j m$: Frequency Restrictions is defined as the centre of the frequency domain with $C_1 2^j \leq m \leq C_2 2^j$.
- The centre of space is defined as $X_{j,n} = 2^j n$.

The framework elements of wave packets are called wave atoms when:

- $|\varphi_u(x)| \leq C_M 2^j (1 + 2^j |x - x_u|)^{-M}$
- $|\varphi_u(x)| \leq C_M 2^j (1 + 2^j |\omega - \omega_u|)^{-M} + C_M 2^{-j} (1 + 2^{-j} |\omega + \omega_u|)^{-M}$ where $M > 0$.

The following figures represents wave atom tiling of both spatial and frequency planes.

Fig 2: one dimensional wave atom in spatial domain for $j=3, m=3$

Fig 3: one dimensional wave atom in spatial domain for $j=4$, $m=5$

Fig 4: one dimensional wave atom in spatial domain for $j=5$, $m=8$

Fig 5: one dimensional wave atom in frequency domain for $j=3$, $m=3$

Fig 6: one dimensional wave atom in frequency domain for $j=4$, $m=5$

Fig 7: one dimensional wave atom in frequency domain for $j=5$, $m=8$

IV. PROPOSED METHOD

a) Shift based Cycle Spinning

Because of lack of translation invariance of wave atoms Pseudo-Gibbs phenomenon will be produced at edges and textures. This phenomenon is controlled by the concept of cycle spinning when it is used on wave atoms. Hence cycle spinning is a better technique to avoid distortions that happen during denoising process [18][19]. For each shift, the wave atom coefficients obtained by performing cycle spinning technique. Denoising is applied over wave atom coefficients at this analysis stage, over coefficients $d_{s,j}$, and $a_{s,j}$ as shown in the following figure.

Fig 8: Cycle spinning analysis stage

A thresholding is applied to these coefficients which produces new denoised coefficients $d'_{s,j}$ and $a'_{s,j}$ as shown in the following figure, which act as input for synthesis stage. An inverse wave atoms transformation and reverse cycle spinning is applied over these coefficients. Finally, denoised signal can be obtained by averaging all these results.

Fig 9: Cycle spinning synthesis stage

b) Threshold Selection

For thresholding of wave atom coefficients various kinds of threshold values can be used. The most used threshold values are Universal threshold, SURE threshold, and Minimax threshold values.

The following formula is used to calculate universal threshold value

$$T_{s,j}^U = \hat{\sigma}_{s,j} \sqrt{2 \ln N_j}$$

Where N_j is number of wave atom coefficients $d_{s,j}$ at decomposition level j and $\sigma_{s,j}$ is given as

$$\hat{\sigma}_{s,j} = \frac{\text{median}(|d_{s,j}|)}{0.6745}$$

The Minimax threshold can be calculated using the following formula,

$$T_{s,j}^{Mm} = \hat{\sigma}_{s,j} \lambda^*(N_j)$$

The Stein's unbiased risk estimation (SURE) threshold value can be calculated as follows,

$$\text{SURE}(T; X) = N_j - 2 \cdot \# \{i : |X_i| \leq T\} + \sum_{i=1}^{N_j} [\min(|X_i|, T)]^2$$

Where $\#\{i: |X_i| \leq T\}$ is the cardinality of the set brackets. The value that minimizes the risk function of wave atom coefficients is given as

$$T_{s,j}^S = \underset{T \geq 0}{\text{argmin}} \left\{ \text{SURE} \left(T; \frac{d_{s,j}}{\hat{\sigma}_{s,j}} \right) \right\}$$

c) Proposed Algorithm

The proposed algorithm is given as follows,

- Step 1:** Perform cycle spinning on a noisy signal u and obtained $S(u)$.
- Step 2:** Wave atom transformation is applied on $S(u)$ and get the transformation coefficients $TS(u)$.
- Step 3:** Applying threshold value Λ_h and obtain transformation coefficients $\Lambda_h(TS(u))$ after denoising.
- Step 4:** Carry out inverse wave atom transformation on the coefficients $\Lambda_h(TS(u))$ and get denoised signal $T^{-1}(\Lambda_h(TS(u)))$.
- Step 5:** The restored signal is $u' = S^{-1}T^{-1}(\Lambda_h(TS(u)))$ can be obtained by conducting reverse cycle spinning on signal obtained in step 4.
- Step 6:** Averaging of all signals obtained in step 5 resulting final denoised signal.

V. RESULTS ANALYSIS

For this work, a pure EEG signal is taken from public EEG database and a random noise is added to it. The above algorithm is implemented over it. Different thresholding values have been used to denoise the wave atom coefficients and results are compared to study the performance of cycle spinning with various shifts. A comparative study is made among wavelet transform, curvelet transform and wave atom transform. We determined that wave atom transform is optimal to remove the noise from the signals. These results are analysed by taking SNR and MSE values as quality parameters of denoised signal.

Fig 10: Original EEG signal taken from public EEG database

Fig 11: Denoised signals with a) Universal threshold b) Sure threshold and c) Minimax threshold values for even shifts

Fig 12: Denoised signals with a) Universal threshold b) Sure threshold and c) Minimax threshold values for odd shifts

Table 1: comparison of MSE and SNR values for various threshold values

S. No.	Threshold Value	Even Shifts		Odd Shifts	
		MSE	SNR	MSE	SNR
1	Universal Threshold value	1.2323	73.1364	1.2308	73.1369
2	SURE	0.0033	75.7079	0.0033	75.7125
3	MiniMaxi	0.3465	73.6874	0.3520	73.6806

Fig 13: Performance of various Threshold values

Table 2: Previous performance of wave atoms for denoising the EEG signals

Sno	Author	Method	Parameter	Value
1	S Shruthi and KG Satheesh Kumar[14]	Wave atom and Wavelet transforms	PSNR	31.59
2	Khelil Seif Eddine and Seddik Hassene[13]	Wave atom	PSNR	27.47
			SSIM	0.999
		Wavelet transform	PSNR	25.04
			SSIM	0.989
3	Pauline John, Therese Yamuna Mahesh and Bindu Sebastian[15]	Wavelet transform	SNR	10.7701
			MSE	0.0251
		Wave atom transform	SNR	20.2054
			MSE	0.006853
4	Wei-qiang Zhang, Yi-mei Song and Ji-qiang Feng[10]	Wave atom transform	PSNR	28.41
		Wave atom with cycle spinning	PSNR	29.53

Table 3: Results of Adaptive filter with wavelet decomposition

S No	Step value	Haar		Db1		Sym2		Coif1	
		MSE	SNR	MSE	SNR	MSE	SNR	MSE	SNR
1	0.1	0.2491	73.8306	0.2496	73.8298	0.2496	73.8299	0.249	73.8309
2	0.2	0.0061	75.4447	0.0060	75.4467	0.0061	75.4443	0.0061	75.4444
3	0.3	7.3135e-4	76.3630	7.3381e-4	76.3615	7.3306e-4	76.3620	7.3087e-4	76.3632
4	0.4	2.0228e-4	76.9211	2.0272e-4	76.9202	2.0225e-4	76.9212	2.0329e-4	76.9190
5	0.5	8.8695e-5	77.2792	8.8917e-5	77.2781	8.9048e-5	77.2775	8.8897e-5	77.2782
6	0.6	5.01780e-5	77.5129	5.2174e-5	77.5096	5.1797e-5	77.5128	5.2002e-5	77.5111
7	0.7	4.07954e-5	77.5463	4.764e-5	77.5491	4.8121e-5	77.5448	4.7559e-5	77.5499
8	0.8	1.1255e-4	77.1757	1.1239e-4	77.1764	1.1252e-4	77.1759	1.1153e-4	77.1797
9	0.9	0.001	76.2219	0.001	76.2225	0.001	76.2219	0.001	76.2221

10	1.0	743	70.3561	754	70.3492	5.3931e4	68.4953	3.5943e5	67.6715
----	-----	-----	---------	-----	---------	----------	---------	----------	---------

Table 4: Results of wavelet and coverlet thresholding for various threshold values

S. No.	Signal	Wavelet Thresholding		Curvelet Thresholding	
		MSE	SNR	MSE	SNR
1	EEG Signal 1	0.8440	68.8911	0.0405	70.2097
2	EEG Signal 2	0.6783	70.3956	0.0401	71.6242
3	EEG Signal 3	0.7563	70.0161	0.0401	71.2919
4	EEG Signal 4	0.5254	69.8901	0.0396	71.0134
5	EEG Signal 5	0.7205	69.9608	0.0396	71.2099

VI. CONCLUSION

In this paper, we propose an algorithm for denoising of EEG signal based on wave atom transform in combination with cycle spinning using different number of shifts. Different thresholding values were used for each type of shift and results were analysed in terms of SNR and MSE. We observed that SURE threshold value can produce better results in denoising of signal with wave atom transformation using odd shift-based cycle spinning technique.

Funding: No Funding

Conflict of interest: The authors declare that they have no conflict of interest.

Ethical approval: This article does not contain any studies with human participants or animals performed by any of the authors.

Author Contribution: All Authors are equally contributed and have read and agreed to the published version of the manuscript.

REFERENCES

- [1] E.J. Candes, L. Demanet, The curvelet representation of wave propagators is optimally sparse[J], Comm. Pure Appl. Math. 2005, 58(11): 1472 – 1528.
- [2] E.J. Candes, D.L. Donoho, New tight frames of curvelets and optimal representation of objects with piecewise – C2 singularities[J], Comm. Pure Appl. Math. 2004, 57(2): 219 – 266.
- [3] Uzzal Biswas, Biswajit Sana and Md. Maniruzzaman, “Denoising ECG Signal Using Different Wavelet Families and Comparison with Other Techniques”, IEEE 2015.
- [4] Juhi Sharma. Prof. Neha Verma, “ECG Signal De-Noising Techniques: A Review”, International Journal of Engineering Research and General Science Volume 4, Issue 1, January-February 2016.
- [5] R.R Coifman and D.L. Donoho. Translation-invariant de-noising. In A. Antoniadis and G. Oppenheim, editors, Wavelets and Statistics, volume 103 of Springer Lecture Notes in Statistics, pages 125-150, New York, 1995, Springer-Verlag.
- [6] Demanet L, Ying L X, “Wave Atoms and Sparsity of Oscillatory Patterns”, Appl. Comput. Harmon. Anal., Vol. 23, No.3, pp. 368-387, 2007.

- [7] Wei-qiang Zhang, Yi-mei Song, Ji-qiang Feng, "A New Image Denoising Method Based on Wave Atoms and Cycle Spinning", Journal of Software, Vol. 9, No. 1, January 2014.
- [8] G. Plonka, J. Ma. "Nonlinear regularized reaction-diffusion filters for denoising of images with textures", IEEE Trans. Image Process, 17, pp. 1283-1294, 2008.
- [9] Miguel A Rodriguez-Hernandez, "Shift selection influence in Partial Cycle Spinning denoising of biomedical signals", Biomedical Signal Processing and Control 26(2016): 64-68
- [10] Wei-qiang Zhang, Yi-mei Song and Ji-qiang Feng, "A new Image Denoising Method Based on Wave Atoms and Cycle Spinning", Journal of Software, vol. 9, No. 1, January 2014
- [11] Alyson K Fletcher and Kannan Ramchandran, " Wavelet Denoising by Recursive Cycle Spinning", IEEE International Conference on Image Processing (ICIP) 2002
- [12] JIAO Shu-qing, "Image Denoising Based on Wave Atoms", IEEE International Conference on Computational Intelligence and Security, 2019
- [13] Khelil Seif Eddine and Seddik Hassene, "Image Denoising using Wave Atom Transform", IEEE International Conference on Systems, Signals & Devices(SSD),2017
- [14] S Shruthi and KG Sathesh Kumar, "Constrained least squares filtering followed by denoising of decomposed images using wave atom and wavelet transform", Science Direct, 2017, p 526 - 532
- [15] Pauline John, Therese Yamuna Mahesh and Bindu Sebastian, "ECG Signal De-noising, Optimization and classification by Wave Atom Transform", IEEE International Conference on Intelligent Computing, Instrumentation and Control Technologies (ICICICT), 2017
- [16] L. Demanet, L. Ying. Wave atoms and sparsity of oscillatory patterns, Applied and Computational Harmonic Analysis 2007; 23(3): 368-387.
- [17] E. Candes, L. Demanet, D. Donoho and L. Ying, Fast discrete curvelet transforms, SIAM Multiscale Model. Simul. 2006; 5(3):861-899.
- [18] Leila. Boutella and Amina. Serir, "Fingerprint Orientation Map Based on Wave Atoms Transform", Journal of Image and Graphics Vol. 1, No.3 September 2013.
- [19] M. Misiti, Y. Misiti, G. Oppenheim, J. M. Poggi. Wavelet toolbox user's guide, September 2015