Novel Method for Evaluating Spatial Resolution of Magnetic Resonance Images

tomokazu takeuchi (✉ tm.takeuchi67@gmail.com)

Gunma Prefectural College of Health Sciences <https://orcid.org/0000-0002-0623-9974>

Norio Hayashi

Gunma Prefectural College of Health Sciences

Yuta Asai

Nippon Medical School Hospital

Yuka Kayaoka

Osaka City University Hospital

Kiichi Yoshida

Nippon Medical School Hospital

Research Article

Keywords: MRI, Image quality assessment, spatial resolution, MTF, ladder method

Posted Date: August 23rd, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-823953/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published at Physical and Engineering Sciences in Medicine on March 1st, 2022. See the published version at <https://doi.org/10.1007/s13246-022-01114-6>.

Novel method for evaluating spatial resolution of magnetic resonance images

Tomokazu Takeuchi 1)

Norio Hayashi 1)

Yuta Asai 2)

Yuka Kayaoka 3)

Kiichi Yoshida 2)

1) Graduate School, Gunma Prefectural College of Health Sciences

2) Department of radiology, Nippon Medical School Hospital

3) MedCity21, Division of Premier Preventive Medicine, Osaka City University
Hospital

Corresponding Author: Tomokazu Takeuchi

Graduate School, Gunma Prefectural College of Health Sciences

1-22-1, Mitumata-machi, Maebashi, Gunma, JAPAN

TEL: +81 90 8187-9988

E-mail address: tm.takeuchi67@gmail.com

Acknowledgements

I am grateful to Dr. A. Ogura for giving me constructive comments and warm encouragement. I also thank Dr. T. Miyati for clear advice and comments.

30 **Abstract**

31 **Purpose**

32 Recently, several methods for evaluating the spatial resolution of magnetic
33 resonance imaging (MRI) have been reported. However, these methods are not simple
34 and can only be used for specific devices. The International Electrotechnical
35 Commission (IEC) 62464-1 recommends a method that uses a periodic array pattern to
36 evaluate the spatial resolution of an MRI device. In this study, we develop a new
37 method (the ladder method) and evaluate its measurement accuracy by adapting the IEC
38 method to evaluate the spatial resolution.

39 **Methods**

40 First, the adaptation of the IEC method is analyzed by simulating the ladder
41 method using a phantom with a periodic pattern, which is constructed using acrylic
42 plates and a nickel sulfate aqueous solution. Subsequently, the ladder method is
43 evaluated in terms of spatial resolution by dividing the standard deviation (SD) by the
44 average signal in the region of interest (ROI) on the ladder phantom image. To evaluate
45 the precision of the ladder method, it is compared with the modulation transfer function
46 (MTF) using a magnitude image with the partial volume effect of the edge image.

47 **Results**

The simulation result shows that the evaluation of the spatial resolution using the ladder method is viable, in which a coefficient of correlation of 0.90 or higher is obtained for all evaluations using the ladder and MTF methods.

Conclusions

The ladder method can be assessed using only the signal mean value and SD in the ROI on the target image. Therefore, the ladder method is a promising method as a substitute for the MTF.

Keywords: MRI, Image quality assessment, spatial resolution, MTF, ladder method

Declarations

Funding: No finding was received.

Conflicts of interest/Competing interests: The author declared that they have no conflict of interest.

Availability of data and material: Not applicable.

Code availability: Not applicable.

Author's contributions: NH made the substantial contributions to the conception. YA, YK and KY interpreted the phantom data.

66 Ethics approval: Not applicable.

67 Consent to participate: Not applicable.

68 Consent for publication: Not applicable.

69

70

71

72

73

74 **1. Introduction**

75 Recently, imaging methods for magnetic resonance imaging (MRI) that are

76 optimized for clinical use have been developed. In these processes, image quality

77 assessment is often performed, where the signal-to-noise ratio, contrast properties, and

78 spatial resolution are considered. The modulation transfer function (MTF) is used the

79 most owing to its definition properties, which are quantifiable and highly precise [1-6].

80 However, the magnitude signal obtained from MRI using the MTF cannot satisfy

81 systems whose output scales linearly with input (linearity). This problem renders it

82 difficult to determine the MTF of MRI [7-10]. In some studies, signals obtained prior to

83 the absolute value arithmetic operation were used to calculate the correct MTF; however,

84 these methods are not simple [10].

85 The International Electrotechnical Commission (IEC) 62464-1 reported a method
 86 that uses the magnitude image to assess the spatial resolution of MRI [11]. This method
 87 uses periodic pattern images to set the region of interest (ROI) on the image.
 88 Subsequently, $m(v)$ at the spatial resolution v [cycles/mm] is calculated using the pixel
 89 signal mean (S) and SD inside the ROI, as follows:

$$90 \qquad \qquad \qquad m(v) = SD / S \qquad \qquad \qquad (1)$$

91 IEC 62464-1 stipulates that when the calculated $m(v)$ exceeds 0.56, the required
 92 spatial resolution level is achieved. Although this method is easy, it only assesses a
 93 single spatial frequency because it is used to assess the periodic resolution of a specific
 94 MRI device. Hence, it is unsuitable for assessing the resolution of different imaging
 95 sequences.

96 In this study, we developed a method to evaluate the spatial resolution of various
 97 MRI images using a new application of IEC 62464-1. We named this method the
 98 “ladder method.” By comparing the ladder method with the MTF, we aim to confirm the
 99 validity of the ladder method and establish it as a viable method for evaluating the
 100 spatial resolution of MRI.

101

2. Material and methods

2.1. Analysis of simulation using ladder method

Our proposed method for evaluating spatial resolution is a modification of the IEC 62464-1 method. The method recommended by the IEC includes determining the image quality to ascertain and maintain the performance evaluation of medical devices. The phantom used for this method contains both high and low magnetic resonance (MR) signal layers to obtain a periodic pattern image. An elliptical ROI is set on the periodic pattern inside the most lateral plate. The S and SD of the ROI are calculated. The $m(v)$ at spatial frequency v [cycles/mm] is defined as shown in Equation (1).

The spatial frequency v is calculated using the thickness (L) via Equation (2).

$$v = 1 / L \quad (2)$$

It is noteworthy that IEC 62464-1 restricted the phantom plate thickness and phantom periodicity. In the IEC method, the plate thickness is indicated by d , the distance between adjacent plates is d_a , and the sum of d and d_a is L , as illustrated in Figure 1. The IEC recommends a d_a/L ratio of between 0.61 and 0.70.

To study the effects of these values on the results, we performed a simulation and varied each of the terms. The simulation was performed using $d_a/L = 0.50, 0.67$, and 0.80 and plate thicknesses d of 1, 2 and 3 mm for each d_a/L case. By performing this

simulation, the suitable da/L will be obtained. In the simulation, the signal was assumed to be that of the pixel arrayed in one dimension, and we accounted for the ideal and blurry images. Simulations based on different da/L values and plate thicknesses were performed with a high signal of 100 and a low signal of 0. Figure 2 shows the signal array for a plate thickness of 2 mm and different da/L values used in the simulation. When the one-pixel signal was set as x_i , the adjoining pixel signals was represented as $x_{i+1}, x_{i+2}, \dots, x_{i+(n-1)}$. Using these pixel signals, the signal mean value $\bar{x}$ was calculated using Equation (3).

$$\bar{x} = \frac{1}{n} \sum_{i=1}^n x_i \quad (3)$$

The SD, σ_x , was calculated using Equation (4), as follows:

$$\sigma_x = \sqrt{\frac{\sum_{i=1}^n (x_i - \bar{x})^2}{n}} \quad (4)$$

An blurry image was obtained by convoluting the convolution filter to the ideal signal. The convolution filter was a three-term filter whose weight ratio was 1:2:1.

The results of these simulations showed that by changing the spatial frequency (v), $m(v)$ indicated the most significant change at $da/L = 0.50$. This implies that $da/L = 0.50$ can present a detailed resolution difference. Subsequently, another simulation was performed in which the phantom $da/L = 0.50$, and the plate thickness was set to 1.0, 1.5, 2.0, 2.5, 3.0, and 4.0 mm (0.50, 0.33, 0.25, 0.20, 0.16, and 0.13 cycles/mm). In this

arrangement, sharp and blurry images were assumed. A blurry image was assumed by convoluting the convolution filter. The weight ratios of the convolution filter were changed to four stages, i.e., 1:1.5:1, 1:2:1, 1:3:1, and 1:4:1, where the 1:4:1 ratio presented a sharp image, and the 1:1.5:1 presented the most blurred image (Figure 3).

2.2. Ladder method

The phantoms containing periodic patterns, as shown in Figure 1, were used to verify the ladder method. The phantoms comprised 10 parallel acrylic plates. The plate thickness d and the distance between plates d_a were set to the same value. To achieve parallel acrylic plates and to maintain a certain distance between the plates, a 15 mm wide acrylic plate was sandwiched on both sides of the plates and then adhered using an acrylic resin (Figure 1). Similar to the IEC method, the plate width (a) was set as twice the slice thickness, and the plate length (b) was set ten times the length L , as shown in Figure 1. Six different phantoms with plate thicknesses of 1, 1.5, 2, 2.5, 3, and 4 mm were fixed in a nickel sulfate aqueous solution ($1.25 \text{ g NiSO}_4 \cdot 6\text{H}_2\text{O} + 1000 \text{ g H}_2\text{O}$). The details of the custom-developed phantom are shown in Table 1, and a custom-developed phantom with a plate thickness of 2 mm is shown in Figure 4a. Each phantom was constructed using sections that were cut from the same acrylic plates.

Hence, the accuracy of the phantom construction was high, as the different phantoms did not yield significantly different results. The angle between the phantom's major axis (x-axis) and the magnetic bore was 13° in the initial setup. The frequency-encoding direction was matched to the x-direction of the phantom, as shown in Figure 4b. For the imaging, the slice was set on the x-z plane, as shown in Figure 1, to obtain the image shown in Figure 4c. The scanning parameters are shown in Section 2.4 below.

For the analysis, the ROI was set on the periodic images. The size of the ROI was set to the maximum; however, it did not exceed the circumference of the plates of the phantom and encompassed less than 90% of the phantom length. The S and SD of all pixels in the ROI were determined. The ladder method evaluates the spatial resolution by dividing the SD by the S in the ROI on the ladder phantom image using Equation (1). The S and SD in the ROI were measured using ImageJ (National Institutes of Health, Bethesda, MD), whereas Excel (Microsoft Corporation, Washington, USA) was used to calculate $m(v)$.

The ladder method uses the same periodic pattern phantom and involves an analysis based on IEC 62464-1. However, the da/L ratio of the phantom was 0.50, and by changing the plate thickness, this method reflects the spatial frequency characteristics.

2.3. Calculation of MTF

The MTF was calculated for a comparison with the ladder method. To obtain the correct MTF, we used the partial volume effect [10].

An acrylic plate with a thickness of 20 mm was fixed in a nickel sulfate aqueous solution to evaluate the MTF (Figure 5a). Figure 5b shows the details of the edge phantom and the location of the slice used for imaging, whereas Figure 5c shows a photograph of the edge phantom. The slice was set in the middle of the acrylic plate and nickel sulfate aqueous solution on the y-axis to offset the edge spread function (ESF) baseline and eliminate all negative lobes. This setting based on the partial volume effect yielded MTF linearity [10]. For imaging, the phantom edge, which was located in the middle of the z-axis, was set parallel to the magnetic bore (Figure 5b). The x- and frequency-encoding directions of the images were matched (Figure (5b)). Images of the edges were obtained using the scanning parameters, as shown in Section 2.4, and Figure 5d shows the obtained MR image. For the analysis, the square of the ROI was set on the edge of the image, and the ESF was obtained from the profile inside the ROI. The ROI measured 40 mm along the x-axis, and its length was 150 mm (Figure 5e). To obtain the MTF, we first differentiated the ESF using Equation (5), i.e., we performed

finite-element differentiation (FED), and obtained a line spread function (LSF) [10].

$$LSF(X_j) = ESF(X_j) - ESF(X_{j-1}), \quad (5)$$

where X is the spatial location, and j is the number of samples. Finally, a Fourier transform was performed in the LSF operation, and the MTF was calculated. The adjustment was performed for the FED using Equations (6) and (7), and the true MTF was calculated [12].

$$MTF(F) = MTF_e(f) \times \alpha(f) \quad (6)$$

$$\alpha(f) = \frac{1}{\text{sinc}\left(\frac{\pi f}{2f_c}\right)}, f_c = \frac{f_s}{2}, \quad (7)$$

where $MTF(f)$ = true MTF, $MTF_e(f)$ = experimental MTF, f_c is the Nyquist frequency, and f_s is the sampling frequency. ImageJ was used to set the ROI on the image and calculate the edge profile. The MTF was calculated using Excel.

2.4. Imaging parameter

All scans were performed using the abovementioned phantom with a loading device in a 15-channel head array coil on a 1.5 T MR system (Ingenia Philips Medical Systems, Best, The Netherlands). Data were collected using single slices via a spin echo sequence. Ten images of different resolutions were obtained for each of the six phantoms used in this study. The scanning parameters for the ladder method were based

on conventional magnitude imaging, with Repetition Time (TR) = 1000 ms, Echo Time (TE) = 50 ms, slice thickness = 5 mm, and field of view = 256 mm × 256 mm. IEC 62464-1 recommends an Signal to Noise Ratio (SNR) greater than 50. All images used for the ladder method had SNRs of 50 or higher when the number of excitations (NEX) was 1. Hence, the NEX was set to 1. The matrix size was varied to obtain images with different resolutions. The 10 different matrix sizes used were 112, 128, 160, 192, 256, 288, 352, 384, 448 and 512 (Table 2). Using Equation (2), the spatial resolution was evaluated at spatial frequencies of 0.125, 0.167, 0.200, 0.250, 0.333, and 0.500 cycles/mm using the six different phantoms. The results are presented in Table 3.

The imaging parameters for the edge phantom were similar to those used in the ladder method. However, the NEX was set to 15 to reduce noise.

2.5. Correlation between ladder method and MTF

The $m(v)$ and MTF were calculated for each of the 10 different resolutions. To compare between the $m(v)$ and MTF, a cutoff frequency, which corresponds to a specific value, was used [13].

Using the calculated $m(v)$ and MTF, the cutoff frequencies at $m(v)$ and MTF values of 0.3, 0.4, 0.5, 0.6, 0.7, and 0.8 were calculated using linear interpolation. The

correlation coefficient between $m(v)$ and MTF cutoff frequency was compared for each value. The correlation coefficient was calculated using Excel, and then the correlation between $m(v)$ and MTF cutoff frequency was evaluated.

3. Results

3.1. Simulation using ladder phantom

The $m(v)$ values measured at three different ratios of da/L (0.50, 0.67, and 0.80) and three plate thicknesses d (1.0, 2.0, and 3.0 mm) are shown in Figure 6a. The results indicate that the $m(v)$ of the phantom with $da/L = 0.50$ varied the most with different plate thicknesses. Therefore, the phantom with $da/L = 0.50$ was used in the ladder method.

Furthermore, the simulation results obtained by assuming a completely sharp image based on $m(v)$ at different spatial resolutions when $da/L = 0.5$ are shown in Figure 6b. The evaluated results of $m(v)$ did not vary from the maximum value of 1 at different spatial frequencies owing to the changes in the plate thickness. Figure 6c shows the results of the blurry pattern using four different convolution filters. These results show that $m(v)$ decreased at high spatial frequencies and increased at high resolutions.

3.2. Ladder method

The $m(v)$ values obtained using the ladder method for each plate thickness in the images of the 10 different matrix sizes are shown in Figure 7. The $m(v)$ value decreased as the spatial frequency increased. The $m(v)$ was low in the low-resolution images, similarly for the small matrix sizes.

3.3. MTF

Figure 8 shows the MTF of the 10 different resolution images obtained by changing the matrix size. For the low-resolution image, the MTF indicated a low value. Each MTF was calculated until the Nyquist frequency was reached; however, to match the evaluated $m(v)$, the MTF curve is shown until a spatial frequency of 0.6 cycles/mm.

3.4. Correlation between ladder method and MTF

Figures 9a–9c show the relationship between $m(v)$ and the MTF cutoff frequency at $m(v)$ and MTF values of 0.3, 0.5, and 0.8, respectively, in the 10 different resolution images. The regression lines and correlation coefficients (r) were calculated and are shown. The correlation coefficient was 0.974 for the $m(v)$ and MTF value of 0.3, 0.989

for 0.5, and 0.797 for 0.8. These values indicate a strong correlation between $m(v)$ and the MTF.

Table 4 shows the correlation coefficient between $m(v)$ and the MTF cutoff frequency at 0.3, 0.4, 0.5, 0.6, 0.7, and 0.8. At all spatial frequencies, the correlation coefficient exceeded 0.90, indicating strong correlation.

4. Discussion

Spatial resolution has been evaluated extensively using various methods [1,2]. However, these methods are complex. Therefore, we propose a simpler “ladder method” to evaluate the spatial resolution of MRI. The ladder method evaluates the spatial resolution by dividing the SD by the average signal in the ROI on the phantom image of the ladder. We compared the ladder method and MTF to validate the ladder method.

The simulation results show that the $m(v)$ of the phantom with $da/L = 0.50$ varied the most with different plate thicknesses. The phantom with $da/L = 0.50$ can be assumed to be rectangular waves with a duty ratio of 1:1. This rectangular wave contained the repetition frequency and the frequency of odd harmonics. Meanwhile, at $da/L = 0.67$ and 0.80, the duty ratio was not 1:1. At this time, the rectangular wave contained the

repetition frequency and the frequency of odd and even harmonics. The ladder method evaluates the spatial resolution at a single spatial frequency based on a single phantom; hence, using a phantom that contains many frequencies is not suitable for evaluating the spatial frequency of a pin point. As such, the $m(v)$ value of the phantom with $da/L = 0.50$ varied the most with different plate thicknesses. Therefore, it was clear that the phantom with $da/L = 0.50$ was the most suitable for evaluating the spatial resolution in different imaging sequences. With regard to the ladder phantom characteristics, the IEC suggests a da/L of 0.61–0.70, based on the relationship with the MTF. However, the simulation results indicate that $da/L = 0.50$ is applicable and renders it easy to understand the phantom characteristics, i.e., the phantom periodicity is 1:1.

The $m(v)$ value was obtained by dividing the SD by S in the ROI. The edge profile of the phantom plate became blurred, and gradation occurred in the low-resolution images. This gradation decreased the SD and hence $m(v)$. For images of the same resolution with thin plate thicknesses, the phantom cycles become finer and increased the gradation. Hence, $m(v)$ decreased. For phantoms with thin plate thicknesses, the ladder method can evaluate high-frequency spatial resolutions based on Equation (2).

Theoretically, the maximum value of $m(v) = 1$ can be obtained using Equation (8) below and the phantom used in this study. This result is exhibited in the simulation

results for the ideal case.

$$m(v) = \sqrt{\frac{(L-d)}{d}} \quad (8)$$

Therefore, $m(v)$ might be correlated to the MTF at low spatial frequencies.

Based on the comparison, the correlation coefficient between $m(v)$ and the MTF cutoff frequency exceeded 0.90 for all values of $m(v)$ and MTF, which indicated a strong correlation. Therefore, the ladder method is a useful yet easy method for evaluating spatial resolution.

In this study, we used plate thicknesses of 1, 1.5, 2, 2.5, 3, and 4 mm. However, it is unclear whether these configurations are sufficient for the spacing requirements. Hence, different plate thicknesses should be used to evaluate the spatial resolution in future studies.

Whereas several different resolution images were evaluated in this study, images with different contrasts or distortions were not analyzed. Some issues may emerge when using different imaging sequences; therefore, more investigations are required to expand the applicability of this method.

5. Conclusion

In this study, we demonstrated that the ladder method exhibited a strong correlation with the MTF, and that it can be used to evaluate the spatial resolution of MRI. An additional benefit of the ladder method is that it is measured only using the S and SD in the ROI on the target image. Therefore, the ladder method is a useful method that can be a substitute for the MTF.

Compliance with Ethical Standards

Conflict of Interest: The authors declare that they have no conflict of interest.

Statement of Human Rights: This article does not contain any studies with human participants performed.

Statements of Animal Rights: This article does not contain any studies with animals performed.

References

1. Delakis L, Xanthis C, Kitney RI. Assessment of the limiting spatial resolution of an MRI scanner by direct analysis of the edge spread function. Med Phys.

- 337 2009;36:1637-1642.
- 338 2. Borri M, Scurr ED, Richardson C, Usher M, Leach MO, Schmidt MA. A novel
339 approach to evaluate spatial resolution of MRI clinical images for optimization and
340 standardization of breast screening protocols. *Med Phys*. 2016;43:6354-6363.
- 341 3. Lim WT, Jung HR. Evaluation on availability of MTF measurement using the ACR
342 phantom in MRI. *Indian Journal of Science and Technology*. 2016;9:1-5.
- 343 4. Mohapatra SM, Turley JD, Prince JR, Blechinger JC, Wilson DA. Transfer function
344 measurement and analysis for a magnetic resonance imager. *Med Phys*.
345 1991;18:1141–1144.
- 346 5. Steckner MC, Drost DJ, Plato FS. A cosine modulation artifact in modulation
347 transfer function computations caused by the misregistration of line spread profiles.
348 *Med Phys*. 1993;20:469-473.
- 349 6. Cunningham IA, Reid BK. Signal and noise in modulation transfer function
350 determinations using the slit, wire, and edge techniques. *Med Phys*.
351 1992;19:1037-1044.
- 352 7. Miyati T, Fujita H, Kasuga T, et al. Measurements of MTF and SNR(f) using a
353 subtraction method in MRI. *Phys Med Biol*. 2002;47:2961-2972.
- 354 8. Steckner MC. Magnitude reconstruction distorts Gibb's phenomenon in magnetic

resonance imaging. Proceedings of the Society of Magnetic Resonance Imaging (SMRM), 9th Annual Meeting (SNRM, New York, 1990), p. 565.

9. Steckner MC, Drost DJ, Prato FS. Comments and reply: ‘ “Transfer function measurements and analysis for a magnetic resonance imager” ’ (Mohapatra et al. 1991). Med. Phys. 1992;19;511-512.

10. Steckner MC, Drost DJ, Prato FS. Computing the modulation transfer function of a magnetic resonance imager. Med Phys. 1994;21:483-489.

11. IEC 62464-1, Magnetic resonance equipment for medical imaging – Part 1. Determination of essential image quality parameters. 2018

12. Cunningham IA and Fenster A. A method for modulation transfer function determination from edge profile with correction for finite-element differentiation. Med Phys. 1987;14;533-537.

13. Miyoung Kwon, Gordon Legge. Spatial-frequency cutoff requirements for pattern recognition in central and peripheral vision. Vision Research. 2011;51;1995-2007.

Figures

Figure 1

Diagram of phantom used for ladder method

$da/L=0.50$	$da/L=0.67$
$da/L=0.80$	

Figure 2

Signal array at 2 mm plate thickness in different da/L .

Figure 3

Convolution filter of different weights.

Figure 4

a. Periodic pattern phantom with plate thickness of 2 mm. b. Placement of ROI on phantom image for ladder method. c. Magnetic resonance image of periodic pattern phantom with plate thickness of 2 mm.

Figure 5

a. Diagram of edge phantom in y–z plane for calculating MTF. b. Detailed diagram of edge phantom shown in three dimensions. c. Custom-developed edge phantom. d. Magnetic resonance image of edge phantom with matrix size of 512. e. Placement of ROI on edge phantom image to calculate MTF.

Figure 6

a. Relations between plate thickness and evaluated value of $m(v)$ in simulation for different da/L values. Evaluated value of $m(v)$ at $da/L = 0.50$ shows the most significant change by changing plate thickness. b. Relationship between spatial frequency and evaluated $m(v)$ in simulation of sharp edge; $m(v)$ did not depend on spatial frequency. c. Relationship between spatial frequency and evaluated $m(v)$ in simulation of blurred edge, and Resolution was changed in four variations; $m(v)$ at high and low spatial frequencies indicated low values.

Figure 7

Relationship between spatial frequency and evaluated $m(v)$ at different resolutions.

Figure 8

Relationship between spatial frequency and MTF obtained using edge method for different resolutions.

Figure 9

a. Relationship between m(v) and MTF cutoff frequency at 0.3. b. Relationship between m(v) and MTF cutoff frequency at 0.5. c. Relationship between m(v) and MTF cutoff frequency at 0.8.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [Table.pdf](#)