

A Personalized FEM Model for Reproducible Measurement of Anti-inflammatory Drugs in Transdermal Administration to Knee

Pasquale Arpaia (✉ pasquale.arpaia@unina.it)

University of Naples Federico II

Federica Crauso

University of Naples Federico II

Mirco Frosolone

University of Naples Federico II

Massimo Mariconda

University of Naples Federico II

Simone Minucci

University of Tuscia

Nicola Moccaldi

University of Naples Federico II

Research Article

Keywords: Anti-Inflammatory Drugs, experimental data, worst-case reconstruction, personalized knee

Posted Date: August 26th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-835685/v1>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published at Scientific Reports on January 13th, 2022.
See the published version at <https://doi.org/10.1038/s41598-021-04718-2>.

A personalized FEM model for reproducible measurement of anti-inflammatory drugs in transdermal administration to knee

Pasquale Arpaia^{*1,2}, Federica Crauso^{3,1}, Mirco Frosolone^{4,1}, Massimo Mariconda⁴, Simone Minucci^{5,1}, and Nicola Moccaldi¹

¹Laboratory of Augmented Reality for Health Monitoring (ARHeMLab), Department of Electrical Engineering and Information Technology, University of Naples Federico II, Naples - Italy, (<http://arhemlab.dieti.unina.it>)

²Interdepartmental Center for Research in Health Management and Innovation in Health (CIRMIS), University of Naples Federico II, Naples - Italy, (www.cirmis.unina.it)

³Department of Advanced Biomedical Sciences, University of Naples Federico II, Naples - Italy (<http://scienze.biomedicheavanzate.dip.unina.it>)

⁴Department of Public Health, University of Naples Federico II, Naples - Italy, (www.sanitapubblica.unina.it.)

⁵Department of Economics, Engineering, Society and Business Organization (DEIM) University of Tuscia, Viterbo - Italy www.deim.unitus.it

*pasquale.arpaia@unina.it

ABSTRACT

A personalized model of the human knee for enhancing the inter-individual reproducibility of a measurement method for monitoring Non-Steroidal Anti-Inflammatory Drugs (NSAIDs) after transdermal delivery is proposed. The model is based on the solution of Maxwell Equations in the electric-quasi-stationary limit via Finite Element Analysis. The dimensions of the custom geometry are estimated on the basis of knee circumference at the patella, body mass index, and sex of each individual. An optimization algorithm allows to find out the electrical parameters of each subject by experimental impedance spectroscopy data. Muscular tissues were characterized anisotropically, by extracting Cole-Cole equation parameters from experimental data acquired with twofold excitation, both transversal and parallel to tissue fibers. A sensitivity and optimization analysis aiming at reducing computational burden in model customization achieved a worst-case reconstruction error lower than 5%. The personalized knee model and the optimization algorithm were validated in vivo by an experimental campaign on **thirty volunteers, 67% healthy and 33% affected by knee osteoarthritis (Kellgren-Lawrence grade [1,4]), with an average error of 3%.**

1 Introduction

Transdermal administration of non-steroidal anti-inflammatory drugs, commonly known as NSAIDs, is widely used for treating knee joint inflammation. Usually, intramuscular infiltrations or less-invasive transdermal **procedures (e.g. ionophoresis)¹ are carried out.**

To date, however, **standardized methods for measuring the drug actually transmitted during the treatment are missing in transdermal delivery and, therefore, personalized medicine approach is prevented².** In the experimental campaign of Spear et al.³, only less than 60% of the therapies turned out to be effective. Conditions of different permeability of the tissues due to both inter-individual (age, sex, ethnicity) and intra-individual factors (different parts of the body, skin diseases, emotional states) prevent from verifying the therapy effectiveness⁴. This limits the clinical use of less-invasive transdermal delivery⁵. In fact, the drug actually administered by transdermal delivery in current clinical practice is assessed empirically, according to the operator experience⁶. **In the case of local/regional drug therapy,** the lack of scientific methods recognized and codified in shared procedures for in-vivo measurement involves significant issues for the definition of the *bioavailability*. **The concept of bioavailability (rate of presence in blood or in urea after a dosage)^{7,8} is fundamental to establish the drug bioequivalence. However, it is not usable in case the drug therapy is not systemic. For this reason, the equivalence between drugs is fundamentally based on the coincidence of the used active ingredient in the case of many current topical medications. However, the fundamental function of carriers and drug chemical formula in ensuring the absorption efficiency (and therefore bioavailability) is neglected in this way.** In this framework, the European Medicine Agency recently underlined the urgency of defining and adopting new methods **for the in-vivo assessment of local/regional drug therapies⁴.**

From the electrical point of view, the introduction of a reference amount of conductive drug into a given volume of biological tissue produces a variation in the equivalent bioimpedance⁹. In previous studies^{10,11}, the Drug Under Skin Meter (DUSM) was

proposed to measure this impedance and so to estimate the amount of substance present in the subcutaneous tissues. In these experimental campaigns, issues related to interindividual reproducibility arose. Indeed, each tissue exhibits a specific electric behaviour because of its dimensions, shape, and electrical properties. It resulted in different responses of different tissues to the same drug amount administration in terms of impedance variation.

In this paper, the reproducibility uncertainty is faced by means of a numerical electrical model of the knee and an optimization algorithm to calculate the pre-infiltrated knee conditions for each individual. Then, these conditions are used as a sound basis for the identification of a personalized relationship between drug and impedance.

This paper is structured as follows: after a background on anatomical structure of the knee and its dielectric characterization in Section 2, Section 3 reports the measurement production diagram and details the underlying numerical model of the knee. Section 4 reports the experimental campaign on human volunteers, the validation of the personalized model, and its customization on some selected volunteers.

2 Background

2.1 Knee Structure

The main bone components of the knee are: patella, tibia, and femur. Patella is a protruding bone, located in the front part of the knee joint. Tibia is a bulky bone placed in the lower part of the leg. Femur rests on it, namely located in the thigh, which is also part of the hip and knee. In¹², tibia and femur dimensions were measured on 118 subjects by magnetic resonance imaging. In particular, average values of Plateau Width (TPW), plateau Medial Width (MW), and plateau Medial Length (ML) were measured for the tibia (Fig. 1A), and the Condyle width (FCW), medial Width (FW), and medial condyle Length (FL) were assessed for the femur (Fig. 1B). In Tab. 1, knee bones dimensions (average and standard deviation), are shown. The average




Figure 1. Knee bones dimensions, (A) anterior and (B) lateral view¹³: Femur Condyle width (FCW), medial Femur Width (FW), tibial plateau Medial Width (MW), Tibial Plateau Width (TPW), medial Femoral condyle Length (FL), tibial plateau Medial Length (ML).

values in male subjects are about 13%-15% higher than in female ones.

Table 1. Typical tibia and femur dimensions for male and female.

	BONES (mm)	GENERAL	MALE	FEMALE
Tibia	TPW	74.9 ± 6.5	80.6 ± 3.9	69.5 ± 3.0
	ML	45.1 ± 4.5	47.9 ± 4.2	42.2 ± 2.9
	MW	30.4 ± 3.0	32.8 ± 2.1	28.1 ± 1.4
Femur	FCW	78.7 ± 6.7	84.2 ± 4.3	73.4 ± 3.6
	FL	52.2 ± 4.9	54.3 ± 4.8	50.2 ± 4.1
	FW	28.9 ± 3.3	30.6 ± 3.4	27.3 ± 2.6

The main muscles of knee are: (i) the quadriceps (upper part), a large fleshy muscle group covering the front and sides of the thigh, (ii) the gastrocnemius, a superficial muscle covering the entire posterior compartment of the leg¹⁴, and (iii) the soleus, located deeply relative to the latter¹⁵. Their thickness strongly depends on the individual morphotype. Their thickness strongly depends on the individual muscular tone.

The skin is the outermost layer of the knee. It is a keratinized multi-layered epithelial tissue consisting of epidermis, dermis, and hypodermis¹⁶. The epidermis is the outermost skin layer, structured in substrates as well: stratum corneum, lucidum, granulosum, spinosum, and basale¹⁷. The skin thickness depends on several factors including the explored body areas^{18,19,19}: epidermis thickness ranges between 20–300 μm ²⁰, whilst dermis ranges between 1–7 mm and differs for male and female^{16,18}.

The hypoderm is a subcutaneous adipose tissue. The Subcutaneous Fat Thickness (SFT) depends on the body district and the physical subject conditions.

2.2 Dielectric modeling of biological tissues

Dielectric properties of biological tissues in the frequency domain are characterized by ohmic and dielectric losses and can be modeled by means of a complex and frequency-dependent dielectric function, the Effective Dielectric Permittivity. Among the many state-of-the-art models of dielectric relaxations, the so called Cole-Cole equation, widely used to model the dielectric properties of biological tissues, was adopted. It expresses the Effective Dielectric Permittivity as (in the following, the subscript "r" standing for "relative" permittivity drops for the sake of the simplicity):

$$\hat{\epsilon}_{eff}(\omega) = \epsilon_{\infty} + \sum_n \frac{\Delta\epsilon_n}{1 + (j\omega\tau_n)^{1-\alpha_n}} + \frac{\sigma_{dc}}{j\omega\epsilon_0} \quad (1)$$

where:

- ϵ_{∞} is the value of permittivity in a frequency range high enough to consider the dielectric as unrelaxed;
- τ_n is the characteristic relaxation time, necessary for the material molecules or dipoles to return to the relaxed state after the application of the electric field;
- α_n is a coefficient affecting the flatness of the frequency spectrum of the n-th relaxation phenomenon;
- $\Delta\epsilon_n$ is the difference between ϵ_s (the static permittivity) and ϵ_{∞} in relation to the $n - th$ relaxation phenomenon;
- σ_{dc} is the static conductivity;
- ϵ_0 the vacuum permittivity;
- and ω the angular frequency.

Several studies were carried out for the electrical characterization of biological tissues^{21,22}. In²³, skin, fat, muscle, and bone were analyzed. In^{20,24}, the forearm skin was characterized on human volunteers by impedance spectroscopy. However, no specific knee study has ever been conducted. The most extensive and in-depth electrical characterization works on biological tissues are certainly those developed by C.Gabriel^{25–27}.

3 Knee model personalization proposal

A numerical model of the knee joint based on "Finite Element Method" is proposed as the core of the measurement production process in NSAIDs transdermal delivery. The diagram of the measurement production is illustrated in Fig.2. *Impedance Spectroscopy* is carried out by generating sinusoidal current spanning over a frequency range and by acquiring the voltage drop across the *Electrodes* and the *Analog Signal Conditioning* stage. A first measurement preceding the drug administration and *Individual Data* (sex, BMI, and knee circumference) are required in input by the *Personalized Knee Model* for the production of individual parameters. During drug administration, the relative change in impedance spectroscopy is measured and employed by the *Drug-in-Knee Model* (calibrated by the individual parameters of the *Personalized Knee Model*) for assessing the drug amount (*Measured Drug*).

The proposed model aims at identifying both healthy and osteoarthritic knees, but does not account for edema effusions. Indeed, literature confirms that a knee inflammation condition has a negligible impact on impedance measurements²⁸. What above does not hold for edemas, due to their high ionic conductivity with respect to other surrounding tissues, especially whenever they are not minimal²⁹.

In this section, (i) the *Assumptions* of the model and (ii) the *Personalization* procedure, are illustrated.




Figure 2. Diagram of measurement production process and particular of the Knee Model

3.1 Default settings

All tissues mentioned in Section 2 but the muscles can be described by means of a unique dielectric function. Muscle tissues of knee junction are skeletal muscles. This type of tissue consists of muscle fibers and fascicles (multiple bundles) of many cells joined together: muscle fibers of a fascicle are mutually parallel, but the fascicles can have variable direction (even different from tendons direction)³⁰. These physiological characteristics immediately point out that skeletal muscle have anisotropic biologic properties, depending on the orientation of the muscle fibers^{31–33}. Therefore, muscles are expected to show also different dielectric properties in two directions, along and perpendicular to their fibers. Indeed, literature surveys confirm such assumption, as reported by C. Gabriel where different frequency spectra for the Effective Dielectric Permittivity along the two directions are shown³⁴. Literature data provide values and expression of the Effective Dielectric Permittivity functions for all the knee joint tissues but the muscles. For them, the only provided dielectric description is with an external electric excitation field perpendicular to muscular fibers.

A more detailed muscles model has been achieved by an identification procedure set up starting from the experimental data of the Effective Dielectric Permittivity function available in³⁴. In this way, the muscles model is able to take into account their anisotropic behaviour. The scalar Effective Dielectric Permittivity function turns out into a diagonal tensor able to describe the muscle behaviour when excited with an external electric field along any direction. This procedure aims at identifying all parameters in Cole-Cole equation (1) of the muscle excited along the fibers direction to let real and imaginary parts of its Effective Dielectric Permittivity frequency spectra fit the experimental data available in³⁴. This plays as a constraint of the following identification problem, stated so that the abovementioned unknown parameters are not “too far” from those of the muscle excited perpendicularly to the fibers. Therefore, the identification problem is stated as:

$$\begin{aligned} & \min \|x_L - x_P\| \\ & \text{subject to : } \hat{\epsilon}_{eff}(\omega_k, x_{LD}) - \hat{\epsilon}_{eff_{exp}}(\omega_k) = 0 \end{aligned} \quad (2)$$

where:

- x_L is the vector of the Cole-Cole parameters of the Effective Dielectric Permittivity when the tissue is excited with an electric field longitudinal to the fibers,
- x_P is the vector of the Cole-Cole parameters of the Effective Dielectric Permittivity when the tissue is excited with an electric field perpendicular to the fibers,
- ω_k is the sampling frequency of the experimental data.

The complex equality constraint is a non linear function of the degrees of freedom and the problem is stated as a convex optimization problem whose solution is afforded via the Interior-Point optimization method. Fig. 3 shows the results obtained in terms of fitting of the frequency spectra of the dielectric permittivity and equivalent conductivity.




Figure 3. Experimental (circles) and fitted (dotted line) dielectric permittivity with experimental (squares) and fitted (solid line) equivalent conductivity of muscles (excitation parallel to fibers).

The electrical parameters listed in Tab. 2 are considered as inputs of the knee model before its optimization and customization. The identified Cole-Cole parameters for the Muscle (parallel) are reported in grey.

Table 2. Parameters of Cole-Cole equation according to equation (2) for Muscles (Parallel) and according to²⁵ for other biological tissues

	ϵ_{∞}	$\Delta\epsilon_1$	τ_1 [ps]	α_1	$\Delta\epsilon_2$	τ_2 [ns]	α_2	$\Delta\epsilon_3$	τ_3 [μ s]	α_3	$\Delta\epsilon_4$	τ_4 [ms]	α_4	σ_{dc} [S/m]
Dry Skin	4.0	32.0	7.23	0.00	1100	32.48	0.00							0.0002
Wet Skin	4.0	39.0	7.96	0.10	280	79.58	0.00	3.0×10^4	1.59	0.16	3.0×10^4	1.592	0.20	0.0004
Fat	2.5	9.0	7.96	0.20	35	15.92	0.10	3.3×10^4	159.15	0.05	1.0×10^7	15.915	0.01	0.0350
Muscle (Transv.)	4.0	50.0	7.23	0.10	7000	353.68	0.10	1.2×10^6	318.31	0.10	2.5×10^7	2.274	0.00	0.2000
Muscle (Paral.)	4.0	50.0	10.57	0.11	7000	19.10	0.01	1.2×10^6	725.47	0.32	2.5×10^7	0.737	0.00	0.2397
Bone	2.5	18.0	13.26	0.22	300	79.58	0.25	2×10^4	159.15	0.20	2.0×10^7	15.915	0.00	0.0700

3.2 Assumptions

3.2.1 Geometry

Three macroscopic knee tissues were considered in the model: skin, muscle, and bone. **The bones geometry faithfully reproduced the actual human shape and was inserted inside a cylindric muscle layer being the innermost part of a layered structure.** The three skin layers, epidermis, dermis, and hypodermis, were partially rearranged. Part of the epidermis was associated with the dermis according to the percentage water content. In fact epidermis water content ranges between [20–70]% from the stratum corneum to basale^{35–37}. Therefore, a structure (Fig. 4) with five layers was realized:

- *Layer 1:* (outermost skin zone) stratum corneum and lucidum, namely *Dry Skin*;
- *Layer 2:* stratum granulosum, spinosum, basale, and dermis, namely *Wet Skin*;
- *Layer 3:* hypoderm, subcutaneous fat tissue, namely *Subcutaneous Fat*;
- *Layer 4:* muscle;
- *Layer 5:* bone.



Figure 4. 3D structure of human knee.

3.2.2 Physics

The proposed model of the human knee is based on the numerical solution of the Direct Current formulation of Maxwell equation in the frequency domain and in the presence of lossy dielectric materials (i.e. tissues listed in section 2) characterized by relaxation phenomena.

Moreover, the time derivative of the flux density field can be neglected, therefore, Maxwell equations can be simplified into the electric quasi-stationary limit, taking into account simultaneously both a curl-free electric field and the current density field in the tissues.

The current density consists of three contributions: the conduction current generated by ohmic conduction, the Maxwell displacement current equal to the rate of change of the electric displacement field, and the possible external current, not considered here owing to the voltage-driven excitation. Taking into account the presence of dielectric materials by means of their effective dielectric permittivity, the following general formulation holds in terms of scalar electric potential, which turns to be the unknown scalar field of the differential problem:

$$\begin{cases} -\epsilon_0 \epsilon \nabla^2 \bar{\varphi} + \nabla \cdot \bar{J}_{ext} = 0 \\ \bar{\varphi}|_{\partial\Omega_T} = \bar{\varphi}_0(\mathbf{r}) \\ -\sigma_{dc} \nabla \bar{\varphi} \cdot \hat{n}|_{\partial\Omega_S} = \bar{J}_n(\mathbf{r}) \end{cases} \quad (3)$$

where:

- $\bar{\varphi}$ is the phasor of the electric scalar potential;
- $\bar{\varphi}_0(\mathbf{r})$ is the value of the Dirichlet boundary condition set on the boundary $\partial\Omega_T$;
- and $\bar{J}_n(\mathbf{r})$ is the value of the Neumann boundary condition set on the boundary $\partial\Omega_S$.

The Dirichlet boundary conditions set the value of the electric scalar potential across voltage source terminals (labelled as Ω_T). The Neumann boundary conditions set the value of the normal component of the total current across some surfaces. In particular, it imposes natural boundary condition (i.e. electric insulation) overall the skin surface Ω_S .

At last, equations in (3) need to be coupled with additional conditions at interface between two different tissues, stating that the normal component of the current density field is continuous there.

3.3 Personalization

3.3.1 Layer thickness customization

An algorithm was conceived to calculate personalized thicknesses of the five knee layers. The input parameters were: sex, Body Mass Index (BMI) and the radius of the mean circumference of knee (*mean-radius*). The radius was obtained as the mean value of three different measurements: at the center of patella, 5 cm above, and 5 cm below. The algorithm consists of the following steps:

- sex-based definition of the thicknesses of *Layer 5* and *Layer 2* according to Tabs. 1 and 3, respectively;
- calculation of the thickness of *Layer 3* (SFT) via linear relation³⁸:

$$BMI = 0,385 \cdot SFT + 16,991; \quad (4)$$

- computation of the thickness of *Layer 4* as the difference between the *mean-radius* and the thicknesses of other layers, having constrained *Layer 1* equal to 60 μm both for males and females^{20,39,40}.

An example of thicknesses [mm] of model layers for males and females, assuming Body Mass Index (BMI) equal to 23 kg/m^2 and radius of mean circumference equal to 6.5 cm, is reported in Tab. 3.

Table 3. Thicknesses [mm] of model layers for males and females (BMI = 23 kg/m^2 , and radius of mean circumference = 6.5 cm.)

	Layer 1	Layer 2	Layer 3	Layer 4	Layer 5
	Dry Skin	Wet Skin	Subcutaneous Fat	Muscle	Bone
Male	0.060	4.501	15.192	15.621	25.640
Female	0.060	3.666	16.971	17.552	22.760

3.3.2 Dielectric parameters customization

The customization of the knee-model electrical behavior is realized by adapting reference values in Tab. 2 for fitting individual experimental data. To this aim, the parameters most affecting the bioimpedance magnitude spectrum were determined by a sensitivity analysis. However, considering their large number (almost eighty), a reduction of the parameters is needed to carry out the customization process at an affordable computational burden.

First, the number of the parameters in Eq. (1) to be optimized for each tissue was reduced by assessing the frequency range and the effects of each relaxation on the spectrum of the actual dielectric permittivity of each tissue. Then, based on data collected in a preliminary experimental campaign, a one-at-a-time-parameter sensitivity analysis was carried out in order to collect the parameters most affecting the bioimpedance magnitude spectrum. The resulting seven Cole-Cole parameters most impacting on the simulated bioimpedance magnitude are reported in Tab. 4.

Table 4. The seven parameters used for personal model customization

	ϵ_∞	$\Delta\epsilon_1$	τ_1	α_1	$\Delta\epsilon_2$	τ_2	α_2	$\Delta\epsilon_3$	τ_3	α_3	$\Delta\epsilon_4$	τ_4	α_4	σ_{dc}
Dry Skin														
Wet Skin														
Fat														
Muscle (Transverse)														
Muscle (Parallel)														
Bone														

At last a Nelder-Mead simplex optimization procedure is used for identifying the actual value of the extracted seven parameters in order to fit the individual bioimpedance magnitude spectrum within a given tolerance. The prediction error is assessed in terms of relative difference between reconstructed and experimental bioimpedance magnitude:

$$\epsilon_{fit} = \frac{\|Z_{exp}(\omega_k) - Z_{sim}(\omega_k)\|}{\|Z_{exp}(\omega_k)\|} \quad (5)$$

where ω_k is the experimental angular frequency.

4 Experimental validation

An experimental campaign was carried out by in-vivo impedance spectroscopy measurements for validating the knee model customization procedure. **Thirty volunteers (27% male and 73% female, average age 36 ± 15 years, mean knee circumference $39 \text{ cm} \pm 4 \text{ cm}$, and BMI $25 \pm 4 \text{ kg}/\text{m}^2$) were included in this study. Ten volunteers were affected by osteoarthritis with Kellgren-Lawrence⁴¹ grade spanning over the range [1,4]. The remaining twenty volunteers exhibited Kellgren-Lawrence grade equal to zero.** All the participants did not present swelling of the knee and were not affected by skin diseases, and did not report previous joint surgeries, nor other kind of pathology. Measurements were performed in accordance with the relevant guidelines and regulations. All the experimental protocols were approved by the ethics committee of the ‘‘Giovanni Paolo II’’ scientific research, hospitalization and healthcare institute (IRCSS) of Bari (Italy). ‘‘Giovanni Paolo II’’ ethics committee was territorial competent for IRCSS Maugeri - Cassano delle Murge: the partner of University Federico II for this research. Informed consent containing all the information on the experiment was provided and signed by all the participants.

4.1 Experimental Setup

Measurements were carried out by means of the Drug Under Skin Meter¹⁰ for exploring thirteen frequency values in the range [1.00 - 49.00] kHz in compliance with safety regulations (IEC-60601 standard). Volunteers were completely isolated from the ground by means of a rubber mat. A battery-powered laptop gathers the data collected by the instrument. Effects of parasitic capacitances are minimized thanks to the equipotential connection between the patient and the instrument chassis realized by using an Electrocardiography clamp electrode, positioned on the right ankle. DUSM was powered by a USB cable from the laptop-PC battery. Therefore, the device was never directly connected to the main supply. Measurements were carried out by pre-gelled cutaneous electrodes (Fiab P500) using a four-wire configuration. The laboratory temperature and humidity were registered. Preliminarily, the patient parameters such as sex, age, measure of knee circumferences and body mass index (BMI) were recorded. The circumference of knee were measured for the correct electrodes positioning as shown in Fig. 5.




Figure 5. Four-wire electrodes applied to the knee opposite sides.

4.2 Impedance spectroscopy measurements and optimization results

DUSM impedance spectroscopy results are reported in Fig. 6 in terms of mean magnitude value at each frequency and errorbars of 1-sigma-reproducibility on all the subjects. High percentage values of standard deviation with respect to the mean values (22% at low frequencies and 16% at high frequencies) revealed a significant inter-individual variability. **In particular, ANOVA revealed a pathology condition not impacting on impedance trend (p -value < 0.002), by confirming the literature results.**

Among the thirty subjects, the optimization procedure achieved a worst-case reconstruction error lower than 5% and an average error of 3%. In Fig. 7, results about four out of the thirty volunteers are shown. Two osteoarthritic volunteers were chosen at the extremes of Kellgren-Lawrence scale, namely of grades 1 and 4. The others were chosen with the highest and the lowest average bioimpedance magnitude among healthy people. The optimization returns the individual impedance spectroscopy curve fitting within a reconstruction error in 1% – 4%. For each individual, the parameters of Tab. 4 were optimized, as reported in Tab. 5.

The results of the optimization proved the feasibility of personalizing the knee model to the specific characteristics of an individual. Seven parameters were good enough to keep the maximum error below 5%. Three parameters belonged to the parallel component of the muscle layer identified through this study.

5 Conclusion

A personalizable FEM model of knee for improving reproducibility in NSAIDs transdermal delivery measurement was presented. The model geometry consist of five layers: bone, muscle, subcutaneous fat, wet skin and dry skin. Thicknesses of layers were personalized according to sex, BMI index, and mean circumference of knee. Electrical characterization of the tissues took into account their non ideal lossy and dispersive behavior by means of Cole-Cole model. Muscle tissue was characterized anisotropically: the parameters of the Cole-Cole equation of the parallel component were calculated from experimental data available in the literature. Personalized models were identified by tuning seven electrical parameters: less than 10% of the initial amount. **Models were validated on thirty volunteers, twenty healthy and ten affected by knee osteoarthritis (Kellgren-Lawrence**




Figure 6. Impedance magnitude spectroscopy: mean (dots) and 1-sigma-reproducibility (errorbars) on all the 30 subjects.

Table 5. Optimized value of Cole-Cole parameters for Arthrosic and Healthy individuals

		Healthy		Arthrosic	
		M_H	F_H	M_A	F_A
Wet Skin	$\Delta\epsilon_3$	5000	5085	15000	30000
	$\Delta\epsilon_4$	75000	5000	15000	30000
Fat	α_4	0.200	0.240	0.436	0.250
	σ_{dc} [S/m]	0.052	0.032	0.167	0.079
Muscle (Parallel)	α_3	0.320	0.313	0.100	0.325
	τ_4 [ms]	5.896	5.421	0.002	0.737
	σ_{dc} [S/m]	0.224	0.015	0.200	0.240

grade [1,4]). The average reconstruction error was 3% and lower than 5% in the worst case. The customization of a simplified knee FEM model with a low computational burden was proven to be feasible both for healthy and osteoarthritic knees. A forthcoming modeling of NSAIDs electrical behaviour in human tissues (*Drug-in-Knee Model*) combined with the personalized models will allow high reproducible assessment of drug amount transdermally delivered.

References

1. Silve, A., Vezinet, R. & Mir, L. M. Nanosecond-duration electric pulse delivery in vitro and in vivo: experimental considerations. *IEEE Transactions on Instrumentation and Measurement* **61**, 1945–1954 (2012).
2. Sanz, R., Calpena, A. C., Mallandrich, M. & Clares, B. Enhancing topical analgesic administration: review and prospect for transdermal and transbuccal drug delivery systems. *Current pharmaceutical design* **21**, 2867–2882 (2015).
3. Spear, B. B., Heath-Chiozzi, M. & Huff, J. Clinical application of pharmacogenetics. *Trends in molecular medicine* **7**, 201–204 (2001).
4. Clarys, P., Alewaeters, K., Lambrecht, R. & Barel, A. Skin color measurements: comparison between three instruments: the chromameter®, the dermaspectrometer® and the mexameter®. *Skin research and technology* **6**, 230–238 (2000).
5. Herkenne, C. *et al.* In vivo methods for the assessment of topical drug bioavailability. *Pharmaceutical research* **25**, 87 (2008).
6. Lademann, J. *et al.* In vivo methods for the analysis of the penetration of topically applied substances in and through the skin barrier. *International journal of cosmetic science* **34**, 551–559 (2012).




Figure 7. Impedance magnitude identification on arthrosic and healthy significant four subjects: average impedance on thirty subjects (black dotted line). Stars: experimental data; dashed line: fitting models. Healthy male subject (M_H): blue. Healthy female subject (F_H): red. Arthrosic male subject (M_A): cyan. Arthrosic female subject (F_A): green.

7. Ashford, M. Assessment of biopharmaceutical properties. *Aulton's Pharmaceutics E-Book: The Design and Manufacture of Medicines* 339 (2017).
8. Oner, Z. G. & Polli, J. E. Bioavailability and bioequivalence. In *ADME Processes in Pharmaceutical Sciences*, 223–240 (Springer, 2018).
9. Grassini, S., Corbellini, S., Angelini, E., Ferraris, F. & Parvis, M. Low-cost impedance spectroscopy system based on a logarithmic amplifier. *IEEE Transactions on Instrumentation and Measurement* **64**, 1110–1117 (2014).
10. Arpaia, P., Cesaro, U. & Moccaldi, N. A bioimpedance meter to measure drug in transdermal delivery. *IEEE Transactions on Instrumentation and Measurement* 1–8 (2018).
11. Arpaia, P., Cesaro, U. & Moccaldi, N. Noninvasive measurement of transdermal drug delivery by impedance spectroscopy. *Nature Scientific Reports* **7**, 44647 (2017).
12. Elsner, J. J., Portnoy, S., Guilak, F., Shterling, A. & Linder-Ganz, E. Mri-based characterization of bone anatomy in the human knee for size matching of a medial meniscal implant. *Journal of biomechanical engineering* **132** (2010).
13. Doctorlib. Atlas of anatomy 25 knee & leg (2015-2019). URL <https://doctorlib.info/medical/anatomy/anatomy.files/image674.jpg>. [Online; accessed 6-April-2020].
14. Maganaris, C. N. Force-length characteristics of the in vivo human gastrocnemius muscle. *Clinical Anatomy: The Official Journal of the American Association of Clinical Anatomists and the British Association of Clinical Anatomists* **16**, 215–223 (2003).
15. Thomason, D. B. & Booth, F. W. Atrophy of the soleus muscle by hindlimb unweighting. *Journal of applied physiology* **68**, 1–12 (1990).
16. Dabrowska, A. *et al.* Materials used to simulate physical properties of human skin. *Skin Research and Technology* **22**, 3–14 (2016).
17. Duck, F. A. *Physical properties of tissues: a comprehensive reference book* (Academic press, 2013).
18. Oltulu, P., Ince, B., Kökbudak, N., Kılıç, F. *et al.* Measurement of epidermis, dermis, and total skin thicknesses from six different body regions with a new ethical histometric technique. *Türk Plastik, Rekonstrüktif ve Estetik Cerrahi Dergisi (Turk J Plast Surg)* **26**, 56–61 (2018).
19. Sandby-Møller, J., Poulsen, T. & Wulf, H. C. Epidermal thickness at different body sites: relationship to age, gender, pigmentation, blood content, skin type and smoking habits. *Acta Dermato Venereologica* **83**, 410–413 (2003).
20. Birgersson, U., Birgersson, E., Nicander, I. & Ollmar, S. A methodology for extracting the electrical properties of human skin. *Physiological measurement* **34**, 723 (2013).

21. Kraszewski, A., Stuchly, M. A. & Stuchly, S. S. A calibration method for measurements of dielectric properties. *IEEE Transactions on Instrumentation and Measurement* **32**, 385–387 (1983).
22. Tofighi, M.-R. & Daryoush, A. S. Biological tissue complex permittivity measured from s_{21} —error analysis and error reduction by reference measurements. *IEEE Transactions on Instrumentation and Measurement* **58**, 2316–2327 (2009).
23. Miklavčič, D., Pavšelj, N. & Hart, F. X. Electric properties of tissues. *Wiley encyclopedia of biomedical engineering* (2006).
24. Birgersson, U., Birgersson, E., Åberg, P., Nicander, I. & Ollmar, S. Non-invasive bioimpedance of intact skin: mathematical modeling and experiments. *Physiological measurement* **32**, 1 (2010).
25. Gabriel, S., Lau, R. & Gabriel, C. The dielectric properties of biological tissues: Ii. measurements in the frequency range 10 hz to 20 ghz. *Physics in medicine & biology* **41**, 2251 (1996).
26. Gabriel, S., Lau, R. & Gabriel, C. The dielectric properties of biological tissues: Iii. parametric models for the dielectric spectrum of tissues. *Physics in Medicine & Biology* **41**, 2271 (1996).
27. Gabriel, C., Gabriel, S. & Corthout, y. E. The dielectric properties of biological tissues: I. literature survey. *Physics in medicine & biology* **41**, 2231 (1996).
28. Hersek, S., Töreyn, H. & Inan, O. T. A robust system for longitudinal knee joint edema and blood flow assessment based on vector bioimpedance measurements. *IEEE Transactions on biomedical circuits and systems* **10**, 545–555 (2015).
29. Hersek, S. *et al.* Wearable vector electrical bioimpedance system to assess knee joint health. *IEEE Transactions on Biomedical Engineering* **64**, 2353–2360 (2016).
30. Hamilton, W. J. *Textbook of human anatomy* (Springer, 1982).
31. Epstein, B. & Foster, K. Anisotropy in the dielectric properties of skeletal muscle. *Medical and Biological Engineering and Computing* **21**, 51 (1983).
32. Aaron, R., Huang, M. & Shiffman, C. Anisotropy of human muscle via non-invasive impedance measurements. *Physics in Medicine & Biology* **42**, 1245 (1997).
33. Čorović, S. *et al.* The influence of skeletal muscle anisotropy on electroporation: in vivo study and numerical modeling. *Medical & biological engineering & computing* **48**, 637–648 (2010).
34. Gabriel, C. Compilation of the dielectric properties of body tissues at rf and microwave frequencies. Tech. Rep., KING'S COLL LONDON (UNITED KINGDOM) DEPT OF PHYSICS (1996).
35. Warner, R. R., Myers, M. C. & Taylor, D. A. Electron probe analysis of human skin: determination of the water concentration profile. *Journal of Investigative Dermatology* **90**, 218–224 (1988).
36. Verdier-Sévrain, S. & Bonté, F. Skin hydration: a review on its molecular mechanisms. *Journal of cosmetic dermatology* **6**, 75–82 (2007).
37. Nakagawa, N., Matsumoto, M. & Sakai, S. In vivo measurement of the water content in the dermis by confocal raman spectroscopy. *Skin Research and Technology* **16**, 137–141 (2010).
38. Nadeem, B., Bacha, R. & Gilani, S. A. Correlation of subcutaneous fat measured on ultrasound with body mass index. *Journal of medical ultrasound* **26**, 205 (2018).
39. Anastasi, G. *Trattato di anatomia umana* (Edi. Ermes, 2007).
40. Douglas, M., Anderson, A. & Michelle, E. Mosby's medical, nursing, & allied health dictionary. *Padova: Piccin* (2004).
41. Kellgren, J. & Lawrence, J. Radiological assessment of osteo-arthrosis. *Annals of the rheumatic diseases* **16**, 494 (1957).

Acknowledgments

The authors thank Prof. C. Petrarca and Dr. A. Quercia for fruitful discussions and gratefully acknowledge the support of D. Cuneo in the inter-individual identification.

Author contributions statement

P.A., S.M., and N.M. conceived and designed the experiments, F.C. performed the experiments, all authors analyzed the results, wrote the manuscript, and reviewed the manuscript.

Competing interests: The authors declare no competing interests.