

Cone Beam Computed Tomography (CBCT) in Periodontal Diseases: A Systematic Review Based on the Efficacy Model

Hassan Assiri (✉ halmuawad@kku.edu.sa)

king khalid university <https://orcid.org/0000-0002-8588-089X>

Ali Dwasaz

King Khalid University

Ahmad AA Alahmari

King Khalid University

Zuhair A Asiri

King Khalid University

Research article

Keywords: Cone-beam computed tomography, Digital volume tomography, Furcation defects, Infra-bony defects, Periodontitis

Posted Date: April 9th, 2020

DOI: <https://doi.org/10.21203/rs.2.18385/v4>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License. [Read Full License](#)

Version of Record: A version of this preprint was published on July 8th, 2020. See the published version at <https://doi.org/10.1186/s12903-020-01106-6>.

Abstract

Objectives The aim of this study was to address the diagnostic efficacy of cone beam computed tomographic (CBCT) imaging in periodontics based on a systematic search and analysis of the literature using the hierarchical efficacy model.

Methods A systematic search of the electronic databases such as PubMed, Scopus, Web of Science, and Cochrane was conducted until February 2019 to identify studies addressing the efficacy of CBCT imaging in Periodontics. The identified studies were subjected to pre-identified inclusion criteria followed by an analysis using a hierarchical model of efficacy (model) designed for an appraisal of the literature on a diagnostic imaging modality. Two examiners performed the eligibility and quality assessment of relevant studies, and consensus was reached where disagreement was found.

Results The search resulted in 64 studies. Of these, 35 publications were allocated to the relevant level of efficacy and quality assessments wherever applicable. The overall diagnostic accuracy studies showed the risk of bias and applicability concerns in the use of CBCT to be low to moderate. In addition, CBCT is accurate in identifying periodontal defects when compared to other modalities. The studies on the level of patient outcomes agreed that CBCT is a reliable tool for the assessment of the outcomes after the treatment of periodontal defects.

Conclusion CBCT was found to be beneficial and accurate in cases of infra-bony defects and furcation involvements.

Background

Periodontal diseases affect the structures surrounding the teeth [1-3]. They range from the mildest form of gingivitis to the most aggressive form of periodontitis. Gingivitis is limited to the inflammation of gingiva without deep involvement of teeth-supporting structures such as the alveolar bone. On the other hand, periodontitis does extend to the alveolar bone [4-7]. It starts with the formation of a periodontal pocket and, consequently, if not treated, leads to bone and tooth loss. Another manifestation of the periodontal diseases in molar-premolar teeth is the formation of furcation defects [8-11]. Since gingivitis affects only the soft tissue, its diagnosis and treatment rely solely on clinical findings including redness, puffiness, and bleeding [12-14]. However, periodontitis could lead to bone resorption depending on its severity; hence, its diagnosis and treatment planning relies on clinical methods supported by radiographic imaging [15-17].

There are several risks to using clinical examination alone, which could prevent the accurate diagnosis of periodontitis, including gingival tissue consistency, inflammation severity, pressure while probing, probe size, probing angulation, and dental restoration existence [18,19]. In dental practice, practitioners routinely utilize conventional radiography such as periapical, bitewing, and panoramic x-ray to evaluate the bone loss and overall condition of the periodontal disease [18]. Nevertheless, the two-dimensional x-ray has some limitations, mainly due to the overlapping of structures [20]. Thus, the detection of bone craters, inter-radicular bone loss, and lingual and buccal marginal bone loss necessitate the consideration of three-dimensional radiography [17,21,22-24].

Cone-beam computed tomography (CBCT) has been used frequently in the last two decades in dentomaxillofacial region [25]. It has many advantages compared to conventional computed tomography (CT) including low price, low radiation dose, and ease of accommodation at dental offices [25-27]. In addition, it has the ability to view the structures in three dimensions [28-30]. CBCT images of periodontal bone lesions offer a highly informative value. The spatial representation of the alveolar bone in all three planes has a significant role in periodontology, as treatment decisions and long-term prognosis rely on it. [11] Accordingly, it can play a potential role as an adjunct to clinical examination in the case of periodontal diseases [28,31-33].

Evidence-based dentistry aims to identify the best available evidence to justify the efficacy and use of any dental imaging or test in actual practice. Accordingly, Fryback and Thornbury came up with a hierarchal model of efficacy in the early nineties to sort out the best available evidence for a diagnostic tool [34].

There are several published studies on the role of CBCT in periodontal diseases in the literature [13-15].

However, the extent to which CBCT is efficient and accurate in the diagnosis, treatment planning, decision-making, and treatment outcomes of periodontal diseases remains ambiguous. On the path to routine use, especially under consideration of higher radiation exposure to patients, the gain in additional information of clinical relevance has to be explored and evaluated. Consequently, we conducted a systematic review to address the efficacy of CBCT in periodontal diseases.

Methods

This review was conducted based on guidelines from Preferred Reporting Items for Systematic Review and Meta-Analysis (PRISMA) [38] and guidance from the center for reviews and dissemination (CRD) for undertaking a systematic review in health care [39]. The eligibility criteria for inclusion and exclusion were set. Then, the included studies were assigned to the suitable level of efficacy. In the meantime, the review question was designed according to the PICO (Population, or Problem, Intervention or Exposure, Comparison, Outcome) element [40]. Finally, each study was evaluated for quality using the predetermined tool for quality assessment (QUADAS 2).

Criteria for inclusion:

- I. Original studies
- II. Systematic reviews
- III. The study must assess the role of CBCT in plaque-induced periodontal disease
- IV. Each study can be on any level of the efficacy model [34]
- V. Studies addressing CBCT accuracy should compare it to clinical or radiographic measurements

Criteria for exclusion:

- I. Case reports
- II. Narrative reviews
- III. Languages other than English
- IV. Studies addressing periapical periodontitis caused by pulpal infection
- V. Studies addressing the bone status for the purpose of dental implant
- VI. Studies highlighting the use of CBCT to address artificially created bone defects

- **Problem specification:**

The research question was defined as “what is the diagnostic efficacy of CBCT in individuals with periodontal diseases?”

- **Literature search:**

Four databases PubMed, Scopus, Cochrane, and Web of Science were searched till February 2019 to identify the relevant studies. The search strategy is shown in Table 1.

- **Study retrieval:**

The resultant studies were subjected to a duplicate check on the RefWorks database. The studies were then reviewed by four authors for relevance based on inclusion and exclusion criteria. After that, the studies meeting the eligibility criteria were assigned for full-text screening. Where uncertainty was present, discussions were conducted between the authors to reach an agreement on whether to include or exclude a study based on the predefined inclusion and exclusion criteria.

- **Data extraction & quality assessment:**

Finally, each of the selected studies was assigned for data extraction and analysis. After that, each study was allocated its suitable level of efficacy. A Revised Tool for the Quality Assessment of Diagnostic Accuracy Studies (QUADAS 2) was used for quality assessment. This tool contains four domains: patient selection, index test, reference standard, and flow and timing. Each domain is assessed in terms of risk of bias and the first three domains are assessed in terms of concerns regarding applicability. Signaling questions are included to help judge the risk of bias [41].

Result

Studies allocation

The search strategy of the four databases yielded 1717 articles: PubMed 539, Scopus 746, Cochrane 71, and Web of Science 555. After a duplicate check using RefWorks, the result came up to 1262. These were subjected to the title and abstract screening by the two authors. A set of 65 studies were linked to the full-text review. A total of 28 articles were excluded because they did not possess at least one of the inclusion criteria. Studies reported by [28,35,42-53] were ex vivo studies and out of our review.

Plaque-induced periodontitis was not addressed, therefore, studies on that issue were excluded. In addition, studies that did not belong to any level of efficacy were disregarded [54-58]. Studies that addressed bone density conducted by Al Zahrani et al.[59] and bone coverage conducted by Ferreira et al. [60] were also excluded. Published studies by Evangelista et al.[61], Sun et al.[62], and Leung et al. [63] discussed only the naturally occurring

dehiscence and fenestration, hence, they were disregarded. Studies reported by Goodarzy et al. [64] and Nagao et al.[65] were excluded because they did not include patients having periodontitis. The case report presented by Naitoh et al.[66] was disregarded as well. Studies published in languages other than English; reported by Deng et al. [67]) was excluded. Figure 1 shows the results for systematic reviews according to the PRISMA flow chart. Table 2 shows the studies that were included and their suitable efficacy level.

- **Quality assessment**

After allocating each study its suitable efficacy level, special tools of quality assessment were used for each one as described in the literature [41].

- **Technical efficacy studies:**

There was no study identified on this level of efficacy.

- **Diagnostic accuracy studies:**

The results revealed eighteen studies [68-85] on diagnostic accuracy. The QUADAS-2 assessment tool was used for quality assessment [38]. Table 3 reveals the results of the quality assessment using QUADAS-2.

There were three studies that included a previously published systematic, manual search of the reference lists of the included articles [67,83,84], among which one study by Deng et al.[67] was found to be published in a Chinese language and hence excluded.

- **Diagnostic thinking efficacy:**

Only one study was found to be on the level of diagnostic thinking efficacy [86]. The author investigated the effect of CBCT on the treatment decision-making after taking into consideration the clinical parameters.

- **Therapeutic efficacy:**

In this level of efficacy, only one study, Pajnigara et al.[87], seemed relevant.

- **Patient outcome efficacy:**

Our research resulted in eight studies in which CBCT was used to address the patients' outcomes in relation to periodontal disease. All of the studies are randomized clinical trials [89-95]. Table 4, the CASP (critical appraisal skills program) checklist, was used to assess outcomes.

- **Societal efficacy:**

Only one study was found to be relevant in this level of efficacy, Walter et al. [82]. The quality assessment was done using the QUADAS 2 tool.

Systematic reviews:

The remaining six studies [6,36,96-99] were found to be systematic reviews for which the AMSTAR-2 assessment tool [100] was used. It is a popular instrument modified from the original AMSTAR, which contains 16 checklist questions. (Refer to Table 5). The two authors meticulously screened each study in order to give a suitable answer for each checklist question.

Discussion

Alveolar bone loss is considered a primary symptom of periodontal diseases. Mostly, the assessment and treatment decisions depend on clinical measurements supported by conventional imaging modalities. However, 2D imaging has its own limitations for detecting bone defects, including overlapping. An estimation of bone loss bucco-lingually has led to the consideration of 3D imaging. However, to what extent the CBCT is effective in

the diagnosis of periodontal diseases is not yet clear. Accordingly, our systematic review was designed to summarize the available evidence according to the hierarchical model of efficacy developed by Fryback et al. [34].

In our systematic review, we decided to exclude studies that are published in any language other than English because of time restriction. In addition, case reports and narrative reviews are considered in the literature as low-evidence studies. Studies addressing periapical conditions and implant-related periodontal problems were also excluded as they are beyond our aspect in this review. In the meantime, it was decided to not include studies conducted ex vivo where the periodontal defects are created artificially since we believe those results will not mimic the CBCT's performance when conducted on humans.

Technical efficacy level:

It seems most of the studies conducted on the use of CBCT in periodontal disease were aimed at performance detection, accuracy estimation, or the treatment outcome assessment. The authors found no study reported in the literature dealt with the technical aspect of CBCT.

Diagnostic accuracy level:

As mentioned earlier in this review, the QUADAS 2 tool was used for the quality assessment of diagnostic accuracy studies. Only studies conducted in vivo were included in this review. Some studies did not use explicit reference standards to compare CBCT with other modalities [69,70,76-78].

Cimbaljevic et al. [66] compared the periodontal probing with CBCT in the terms of furcation involvement in the absence of a reference standard. Likewise, Darby et al. [67] addressed the discrepancies in the clinical measurements obtained from patients' records with their available CBCT images. A study conducted by Suphaanantachat et al. [73] compared CBCT to conventional intraoral radiography. However, they did not use an actual reference standard for comparison. Similarly, Zhu J. et al. [75] has focused on the reproducibility of the different parameters of CBCT for the furcation involvement evaluation, and hence, no reference standard was used.

Diagnostic thinking:

A study published by Walter et al. [86] on decision-making revealed discrepancies between clinically and CBCT-based therapeutic treatment approaches. The discrepancy was found after 59–82% of the teeth were investigated to find out whether less invasive or most invasive treatment should be considered. However, they concluded that CBCT provides informative details in cases of furcation involvement, and hence, it is considered a reliable tool in decision-making regarding treatment of furcation involvement.

Therapeutic efficacy:

According to our interpretation and in correlation with the hierarchical model of efficacy [34], we found that the study conducted by Pajnigara et al. [87] fits on this level. They investigated the pre and post-surgical measurements of clinical and CBCT for furcation defects. Although they reported statistically significant differences between; clinical-presurgery CBCT ($P < 0.0001$, 95% CI) and clinical-post surgery CBCT; the three-dimensional imaging gives dental practitioners the chance to optimize treatment decisions and assess the degree of healing more effectively.

Patient's outcome efficacy:

Our systematic review has revealed eight studies that used CBCT to assess the results of treatment provided for periodontal diseases [88-95]. However, it seems that this study is in disagreement with a previously published review [6]. They did not identify any study on the level of patient outcome. The reason for this could be the difference between our inclusion and exclusion criteria and theirs. All studies agreed that CBCT is a reliable tool in the assessment of the results of treatment using a bone graft.

Societal efficacy:

The study reported by Walter et al. [82] has shown that the use of CBCT decreases the cost and time for periodontitis screening. However, CBCT should only be advised in cases of advanced therapy. Further studies with a sufficient number of patients were suggested.

Systematic reviews:

Our review has resulted in six studies, which are systematic reviews. Each review is supposed to adhere to the criteria provided by AMSTAR and scores YES whenever applicable. The review published by Haas et al. [96] did not elaborate on whether they included the study registries or consulted content experts in the field in terms of comprehensive literature search strategy. Although a meta-analysis was conducted in such a review, the review authors did not assess the potential impact of risk of bias on the results of the meta-analysis or other evidence synthesis. Moreover, the authors did not carry out an adequate investigation of publication bias (small-study bias) or discuss its likely impact on the results of the review. Based on our interpretation, the study has not reported any source of funding or mentioned any conflict of interest.

The study by Walter et al. [94] did not clearly have an explicit statement that the review methods were established prior to the conduct of the review and did not justify any significant deviations from the protocol. In addition, only one database has been searched for relevant studies. According to the AMSTAR2 criteria, the included studies were not described adequately. The study has not reported on the source of funding for the individual studies included in the review. To our knowledge, the risk of bias has not been elaborated upon in the relevant sites in the review. Moreover, the review authors did not account for the risk of bias in individual studies when interpreting or discussing the results of the review. In addition, the authors have not reported any source of conflict including any funding they received for conducting the review.

The review by Anter et al. [33] addressed the accuracy of the CBCT as a tool for the measurement of alveolar bone loss in periodontal defects. However, the authors did not report that they followed PICO, which is a framework for review question formulation [40]. In terms of a comprehensive search strategy, we saw that this review did not fulfill the criteria regarding study registries and expert consultation in the field. Furthermore, the authors did not conduct the search in duplicate for the purpose of study selection. The review authors had also not performed data extraction in duplicates. According to our interpretation, the included studies were not described in appropriate detail. Additionally, the source of funding for each relevant individual study was not reported.

The study reported by Choi et al. [98] did not specify whether if there was a deviation from protocol, meta-analysis plan, or causes of heterogeneity if appropriate. In addition, a list of the excluded study in association with a justification for exclusion of each potential study has not been provided. Regardless of whether it is one of the targets of the review, this review has not discussed any potential risk of bias of the included studies. Moreover, the source of funding of each included study was also not reported. It could be included that this review does fulfill the AMSTAR2 [100] checklist to some extent.

The review by Woelber et al. [99] neither mentions any deviation from protocol whenever applicable nor elaborates on if is a plan for meta-analysis, if appropriate. In addition, a plan for investigating the possible causes, if appropriate, regarding heterogeneity was also not reported. The source of funding for each included study was not reported either. To some extent, the review fulfills the checklist of AMSTAR2.

According to our systematic review and AMSTAR2 tool, we found the review conducted by Nikolic-Jakoba et al. [6] best fulfills the tool criteria. However, the study's authors did not justify the reason for exclusion of each potentially relevant study from the review. As other reviews were included in our study, the source of funding of each included publication was not reported.

Conclusion

We concluded that most of the studies conducted on the rule of CBCT in periodontal diseases were at diagnostic accuracy level followed by the patient outcome level. Accordingly, it was found that CBCT is quite beneficial and accurate in the diagnosis of infra-bony defects and furcation involvement. Similarly, it is reliable in the assessment of the outcome of periodontal surgery and regenerative therapy. Furthermore, more studies with a larger cohort on the level of diagnostic thinking, therapeutic, and societal efficacy are needed to set up a clear guideline and evidence for the usefulness of CBCT.

Abbreviations

CBCT (Cone beam computed tomography)

CT (Computed tomography)

PRISMA (Preferred items for systematic review and meta-analysis)

CRD (Center for reviews and dissemination)

PICO (**P**opulation, or **P**roblem, **I**ntervention or Exposure, **C**omparison (if appropriate), **O**utcome you would like to measure or achieve)

QUADAS2 (quality assessment tool for diagnostic accuracy studies)

CASP (critical appraisal skills program)

AMSTAR (A measurement tool to assess systematic review)

2D (Two-dimensional)

Declarations

Ethics approval and consent to participant: The research committee at the faculty of dentistry in King Khalid University has approved the study proposal. Since it is a review, the consent of participants is not applicable.

Consent to publish: Not applicable.

Availability of data materials: The datasets used and/or analyzed during the current study are available from the corresponding author on reasonable request.

Competing interest: The authors declare they have no competing interests.

Funding: No funding was provided for this systematic review.

Author's contribution: HA: Literature search, analyzing individual studies, manuscript preparation and review; AAD: Literature search, analyzing individual studies, manuscript review; AA: Analyzing individual studies, manuscript review; ZA: Literature search, manuscript review.

Acknowledgments: No other contributors to acknowledge.

References

1. Flemmig TF. Periodontitis. *Ann Periodontol*. 1999 Dec;4(1):32–8.
2. Highfield J. Diagnosis and classification of periodontal disease. *Aust Dent J*. 2009 Sep;54 Suppl 1:S11–26.
3. Pihlstrom BL, Michalowicz BS, Johnson NW. Periodontal diseases. *Lancet (London, England)*. 2005 Nov;366(9499):1809–20.
4. Loe H, Anerud A, Boysen H, Morrison E. Natural history of periodontal disease in man. Rapid, moderate and no loss of attachment in Sri Lankan laborers 14 to 46 years of age. *J Clin Periodontol*. 1986 May;13(5):431–45.
5. Burt B. Position paper: epidemiology of periodontal diseases. *J Periodontol*. 2005 Aug;76(8):1406–19.
6. Nikolic-Jakoba N, Spin-Neto R, Wenzel A. Cone Beam Computed Tomography for Detection of Intrabony and Furcation Defects: A Systematic Review Based on a Hierarchical Model for Diagnostic Efficacy. *J Periodontol*. 2016 Jun;87(6):630–44.
7. Jordan RCK. Diagnosis of periodontal manifestations of systemic diseases. *Periodontol 2000*. 2004;34:217–29.
8. Listgarten MA. Periodontal probing: what does it mean? *J Clin Periodontol*. 1980 Jun;7(3):165–76.
9. Tugnait A, Clerehugh V, Hirschmann PN. The usefulness of radiographs in diagnosis and management of periodontal diseases: a review. *J Dent*. 2000 May;28(4):219–26.
10. Tyndall DA, Rathore S. Cone beam CT diagnostic applications: caries, periodontal bone assessment, and endodontic applications. *Dent Clin North Am*. 2008 Oct;52(4):825–41, vii.
11. Braun X, Ritter L, Jervoe-Storm P-M, Frentzen M. Diagnostic accuracy of CBCT for periodontal lesions. *Clin Oral Investig*. 2014 May;18(4):1229–36.
12. Summers A. Gingivitis: diagnosis and treatment. *Emerg Nurse*. 2009 Apr;17(1):18–20; quiz 35.
13. Bailey DL, Barrow S-Y, Cvetkovic B, Musolino R, Wise SL, Yung C, et al. Periodontal diagnosis in private dental practice: a case-based survey. *Aust Dent J*. 2016 Jun;61(2):244–51.
14. Garcia-Pola M-J, Rodriguez-Lopez S, Fernanz-Vigil A, Bagan L, Garcia-Martin J-M. Oral hygiene instructions and professional control as part of the treatment of desquamative gingivitis. Systematic review. *Med Oral Patol Oral Cir Bucal*. 2019 Mar;24(2):e136–44.
15. Socransky SS, Haffajee AD. The bacterial etiology of destructive periodontal disease: current concepts. *J Periodontol*. 1992 Apr;63(4 Suppl):322–31.
16. Armitage GC. The complete periodontal examination. *Periodontol 2000*. 2004;34:22–33.
17. Taba MJ, Kinney J, Kim AS, Giannobile W V. Diagnostic biomarkers for oral and periodontal diseases. *Dent Clin North Am*. 2005 Jul;49(3):551–71, vi.
18. Wolf DL, Lamster IB. Contemporary concepts in the diagnosis of periodontal disease. *Dent Clin North Am*. 2011 Jan;55(1):47–61.
19. Albandar JM, Brunelle JA, Kingman A. Destructive periodontal disease in adults 30 years of age and older in the United States, 1988-1994. *J Periodontol*. 1999 Jan;70(1):13–29.
20. facial radiology (Evidence-based guidelines) ISSN 1681-6803 MJ-XA-12-001-EN-C Energy Protection Radiation No 172 Cone beam CT for dental and maxillofacial radiology (Evidence-based guidelines) [Internet]. 2012. Available from: http://www.sedentext.eu/files/radiation_protection_172.pdf
21. Offenbacher S. Periodontal diseases: pathogenesis. *Ann Periodontol*. 1996 Nov;1(1):821–78.

22. Ozmeric N, Kostiuoutchenko I, Hagler G, Frentzen M, Jervoe-Storm P-M. Cone beam computed tomography in assessment of periodontal ligament space: in vitro study on artificial tooth model. *Clin Oral Investig*. 2008 Sep;12(3):233–9.
23. Hashimoto K, Kawashima S, Araki M, Iwai K, Sawada K, Akiyama Y. Comparison of image performance between cone beam computed tomography for dental use and four-row multidetector helical CT. *J Oral Sci*. 2006 Mar;48(1):27–34.
24. Loubele M, Maes F, Schutyser F, Marchal G, Jacobs R, Suetens P. Assessment of bone segmentation quality of cone beam CT versus multislice spiral CT: a pilot study. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod*. 2006 Aug;102(2):225–34.
25. Shukla S, Chug A, Afrashtehfar KI. Role of Cone Beam Computed Tomography in Diagnosis and Treatment Planning in Dentistry: An Update. *J Int Soc Prev Community Dent*. 2017 Nov;7(Suppl 3):S125–36.
26. scholar Filler AG 2010.
27. Orentlicher G, Goldsmith D, Abboud M. Computer-guided planning and placement of dental implants. *Atlas Oral Maxillofac Surg Clin North Am*. 2012 Mar;20(1):53–79.
28. Takeshita WM, Vessoni Iwaki LC, Da Silva MC, Tonin RH. Evaluation of diagnostic accuracy of conventional and digital periapical radiography, panoramic radiography, and cone beam computed tomography in the assessment of alveolar bone loss. *Contemp Clin Dent*. 2014 Jul;5(3):318–23.
29. Jeffcoat MK. Radiographic methods for the detection of progressive alveolar bone loss. *J Periodontol*. 1992 Apr;63(4 Suppl):367–72.
30. Swennen GRJ, Mommaerts MY, Abeloos J, De Clercq C, Lamoral P, Neyt N, et al. A cone beam CT based technique to augment the 3D virtual skull model with a detailed dental surface. *Int J Oral Maxillofac Surg*. 2009 Jan;38(1):48–57.
31. Worthington P, Rubenstein J, Hatcher DC. The role of cone beam computed tomography in the planning and placement of implants. *J Am Dent Assoc*. 2010 Oct;141 Suppl:19S–24S.
32. Alqerban A, Hedesiu M, Baciut M, Nackaerts O, Jacobs R, Fieuws S, et al. Pre-surgical treatment planning of maxillary canine impactions using panoramic vs cone beam CT imaging. *Dentomaxillofac Radiol*. 2013;42(9):20130157.
33. scholar S Patil 2015.
34. Fryback DG, Thornbury JR. The efficacy of diagnostic imaging. *Med Decis Making*. 1991;11(2):88–94.
35. Bagis N, Kolsuz ME, Kursun S, Orhan K. Comparison of intraoral radiography and cone beam computed tomography for the detection of periodontal defects: an in vitro study. *BMC Oral Health*. 2015 May;15:64.
36. Anter E, Zayet MK, El-Dessouky SH. Accuracy and precision of cone beam computed tomography in periodontal defects measurement (systematic review). *J Indian Soc Periodontol*. 2016;20(3):235–43.
37. Mohan R, Singh A, Gundappa M. Three-dimensional imaging in periodontal diagnosis - Utilization of cone beam computed tomography. *J Indian Soc Periodontol*. 2011 Jan;15(1):11–7.
38. Moher D, Liberati A, Tetzlaff J, Altman DG. Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. *PLoS Med*. 2009 Jul;6(7):e1000097.
39. Tacconelli E. Systematic reviews: CRD's guidance for undertaking reviews in health care. *Lancet Infect Dis [Internet]*. 2010 Apr 1;10(4):226. Available from: [https://doi.org/10.1016/S1473-3099\(10\)70065-7](https://doi.org/10.1016/S1473-3099(10)70065-7)
40. Speckman RA, Friedly JL. Asking Structured, Answerable Clinical Questions Using the Population, Intervention/Comparator, Outcome (PICO) Framework. *PM R*. 2019 Feb;
41. Whiting PF, Rutjes AWS, Westwood ME, Mallett S, Deeks JJ, Reitsma JB, et al. QUADAS-2: a revised tool for the quality assessment of diagnostic accuracy studies. *Ann Intern Med*. 2011 Oct;155(8):529–36.
42. Almeida VC, Pinheiro LR, Salineiro FCS, Mendes FM, Neto JBC, Cavalcanti MGP, et al. Performance of cone beam computed tomography and conventional intraoral radiographs in detecting interproximal alveolar bone lesions: a study in pig mandibles. *BMC Oral Health*. 2017 Jun;17(1):100.
43. Fleiner J, Hannig C, Schulze D, Stricker A, Jacobs R. Digital method for quantification of circumferential periodontal bone level using cone beam CT. *Clin Oral Investig*. 2013 Mar;17(2):389–96.
44. Kamburoglu K, Kolsuz E, Murat S, Eren H, Yuksel S, Paksoy CS. Assessment of buccal marginal alveolar peri-implant and periodontal defects using a cone beam CT system with and without the application of metal artefact reduction mode. *Dentomaxillofac Radiol*. 2013;42(8):20130176.
45. Misch KA, Yi ES, Sarment DP. Accuracy of cone beam computed tomography for periodontal defect measurements. *J Periodontol*. 2006 Jul;77(7):1261–6.
46. Mol A, Balasundaram A. In vitro cone beam computed tomography imaging of periodontal bone. *Dentomaxillofac Radiol*. 2008 Sep;37(6):319–24.
47. Noujeim M, Prihoda T, Langlais R, Nummikoski P. Evaluation of high-resolution cone beam computed tomography in the detection of simulated interradicular bone lesions. *Dentomaxillofac Radiol*. 2009 Mar;38(3):156–62.
48. Salineiro FCS, Gialain IO, Kobayashi-Velasco S, Pannuti CM, Cavalcanti MGP. Detection of furcation involvement using periapical radiography and 2 cone beam computed tomography imaging protocols with and without a metallic post: An animal study. *Imaging Sci Dent*. 2017 Mar;47(1):17–24.

49. Vandenberghe B, Jacobs R, Yang J. Detection of periodontal bone loss using digital intraoral and cone beam computed tomography images: an in vitro assessment of bony and/or infrabony defects. *Dentomaxillofac Radiol.* 2008 Jul;37(5):252–60.
50. Vandenberghe B, Jacobs R, Yang J. Diagnostic validity (or acuity) of 2D CCD versus 3D CBCT-images for assessing periodontal breakdown. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2007 Sep;104(3):395–401.
51. Kolsuz ME, Bagis N, Orhan K, Avsever H, Demiralp KO. Comparison of the influence of FOV sizes and different voxel resolutions for the assessment of periodontal defects. *Dentomaxillofac Radiol.* 2015;44(7):20150070.
52. Kamburoglu K, Yeta EN, Yilmaz F. An ex vivo comparison of diagnostic accuracy of cone beam computed tomography and periapical radiography in the detection of furcal perforations. *J Endod.* 2015 May;41(5):696–702.
53. Elashiry M, Meghil MM, Kalathingal S, Buchanan A, Elrefai R, Looney S, et al. Application of radiopaque micro-particle fillers for 3-D imaging of periodontal pocket analogues using cone beam CT. *Dent Mater.* 2018 Apr;34(4):619–28.
54. Lim H-C, Jeon S-K, Cha J-K, Lee J-S, Choi S-H, Jung U-W. Prevalence of Cervical Enamel Projection and Its Impact on Furcation Involvement in Mandibular Molars: A Cone Beam Computed Tomography Study in Koreans. *Anat Rec (Hoboken).* 2016 Mar;299(3):379–84.
55. Ozcan G, Sekerci AE. Classification of alveolar bone destruction patterns on maxillary molars by using cone beam computed tomography. *Niger J Clin Pract.* 2017 Aug;20(8):1010–9.
56. Yang Y, Yang H, Pan H, Xu J, Hu T. Evaluation and New Classification of Alveolar Bone Dehiscences Using Cone beam Computed Tomography in vivo. Vol. 33, *International Journal of Morphology.* 2015. 361-368 p.
57. Amorfini L, Migliorati M, Signori A, Silvestrini-Biavati A, Benedicenti S. Block allograft technique versus standard guided bone regeneration: a randomized clinical trial. *Clin Implant Dent Relat Res.* 2014 Oct;16(5):655–67.
58. Alshaer S, Alhaffar I, Khattab R. Evaluating change in radiographic bone density via cone beam computed tomography before and after surgery to patients with chronic periodontitis. Vol. 23, *International Journal of Pharmaceutical Sciences Review and Research.* 2013. 143-147 p.
59. Al-Zahrani MS, Elfirt EY, Al-Ahmari MM, Yamany IA, Alabdulkarim MA, Zawawi KH. Comparison of Cone Beam Computed Tomography-Derived Alveolar Bone Density Between Subjects with and without Aggressive Periodontitis. *J Clin Diagn Res.* 2017 Jan;11(1):ZC118-ZC121.
60. Ferreira PP, Torres M, Campos PSF, Vogel CJ, de Araujo TM, Rebello IMCR. Evaluation of buccal bone coverage in the anterior region by cone beam computed tomography. *Am J Orthod Dentofacial Orthop.* 2013 Nov;144(5):698–704.
61. Evangelista K, Vasconcelos K de F, Bumann A, Hirsch E, Nitka M, Silva MAG. Dehiscence and fenestration in patients with Class I and Class II Division 1 malocclusion assessed with cone beam computed tomography. *Am J Orthod Dentofacial Orthop.* 2010 Aug;138(2):133.e1-7; discussion 133-5.
62. Sun L, Zhang L, Shen G, Wang B, Fang B. Accuracy of cone beam computed tomography in detecting alveolar bone dehiscences and fenestrations. *Am J Orthod Dentofacial Orthop.* 2015 Mar;147(3):313–23.
63. Leung CC, Palomo L, Griffith R, Hans MG. Accuracy and reliability of cone beam computed tomography for measuring alveolar bone height and detecting bony dehiscences and fenestrations. *Am J Orthod Dentofacial Orthop.* 2010 Apr;137(4 Suppl):S109-19.
64. Goodarzi Pour D, Romoozi E, Soleimani Shayesteh Y. Accuracy of Cone Beam Computed Tomography for Detection of Bone Loss. *J Dent (Tehran).* 2015 Jul;12(7):513–23.
65. Nagao J, Mori K, Kitasaka T, Suenaga Y, Yamada S, Naitoh M. Quantification and visualization of alveolar bone resorption from 3D dental CT images. Vol. 2, *Int. J. Computer Assisted Radiology and Surgery.* 2007. 43-53 p.
66. Naitoh M, Yamada S, Noguchi T, Arijii E, Nagao J, Mori K, et al. Three-dimensional display with quantitative analysis in alveolar bone resorption using cone beam computerized tomography for dental use: a preliminary study. *Int J Periodontics Restorative Dent.* 2006 Dec;26(6):607–12.
67. Deng Y, Wang C, Li T, Li A, Gou J. [An application of cone beam CT in the diagnosis of bone defects for chronic periodontitis]. *Zhonghua Kou Qiang Yi Xue Za Zhi.* 2015 Jan;50(1):7–12.
68. Banodkar AB, Gaikwad RP, Gunjekar TU, Lobo TA. Evaluation of accuracy of cone beam computed tomography for measurement of periodontal defects: A clinical study. *J Indian Soc Periodontol.* 2015;19(3):285–9.
69. Cimbajjevic MM, Spin-Neto RR, Miletic VJ, Jankovic SM, Aleksic ZM, Nikolic-Jakoba NS. Clinical and CBCT-based diagnosis of furcation involvement in patients with severe periodontitis. *Quintessence Int.* 2015;46(10):863–70.
70. Darby I, Sanelli M, Shan S, Silver J, Singh A, Soedjono M, et al. Comparison of clinical and cone beam computed tomography measurements to diagnose furcation involvement. *Int J Dent Hyg.* 2015 Nov;13(4):241–5.
71. de Faria Vasconcelos K, Evangelista KM, Rodrigues CD, Estrela C, de Sousa TO, Silva MAG. Detection of periodontal bone loss using cone beam CT and intraoral radiography. *Dentomaxillofac Radiol.* 2012 Jan;41(1):64–9.
72. Feijo CV, Lucena JGF de, Kurita LM, Pereira SL da S. Evaluation of cone beam computed tomography in the detection of horizontal periodontal bone defects: an in vivo study. *Int J Periodontics Restorative Dent.* 2012 Oct;32(5):e162-8.
73. Guo Y-J, Ge Z, Ma R, Hou J, Li G. A six-site method for the evaluation of periodontal bone loss in cone beam CT images. *Dentomaxillofac Radiol.* 2016;45(1):20150265.
74. Li F, Jia PY, Ouyang XY. Comparison of Measurements on Cone Beam Computed Tomography for Periodontal Intrabony Defect with Intra-surgical Measurements. *Chin J Dent Res.* 2015 Sep;18(3):171–6.

75. Qiao J, Wang S, Duan J, Zhang Y, Qiu Y, Sun C, et al. The accuracy of cone beam computed tomography in assessing maxillary molar furcation involvement. *J Clin Periodontol*. 2014 Mar;41(3):269–74.
76. Suphanantachat S, Tantikul K, Tamsailom S, Kosalagood P, Nisapakultorn K, Tavedhikul K. Comparison of clinical values between cone beam computed tomography and conventional intraoral radiography in periodontal and infrabony defect assessment. *Dentomaxillofac Radiol*. 2017 Aug;46(6):20160461.
77. Walter C, Weiger R, Zitzmann NU. Accuracy of three-dimensional imaging in assessing maxillary molar furcation involvement. *J Clin Periodontol*. 2010 May;37(5):436–41.
78. Zhu J, Ouyang XY. Assessing Maxillary Molar Furcation Involvement by Cone Beam Computed Tomography. *Chin J Dent Res*. 2016;19(3):145–51.
79. George Marinescu A, Boariu M, Rusu D, Stratul S-I, Ogodescu A. Reliability of CBCT as an assessment tool for mandibular molars furcation defects. *Prog Biomed Opt Imaging - Proc SPIE*. 2013;8925.
80. Jiro Nagao, Takayuki Kitasaka, Kensaku Mori, Yasuhito Suenaga, Shohzoh Yamada MN. Three-dimensional analysis of alveolar bone resorption by image processing of 3-D dental CT images. *J Med Imaging*. 2006;6144.
81. Padmanabhan S, Dommy A, Guru SR, Joseph A. Comparative Evaluation of Cone beam Computed Tomography versus Direct Surgical Measurements in the Diagnosis of Mandibular Molar Furcation Involvement. *Contemp Clin Dent*. 2017;8(3):439–45.
82. Walter C, Weiger R, Dietrich T, Lang NP, Zitzmann NU. Does three-dimensional imaging offer a financial benefit for treating maxillary molars with furcation involvement? A pilot clinical case series. *Clin Oral Implants Res*. 2012 Mar;23(3):351–8.
83. Moradi Haghgoo J, Shokri A, Khodadoustan A, Khoshhal M, Rabienejad N, Farhadian M. Comparison the Accuracy of the Cone Beam Computed Tomography With Digital Direct Intraoral Radiography, in Assessment of Periodontal Osseous Lesions. *Avicenna J Dent Res*. 2014;6.
84. Raichur PS, Setty SB, Thakur SL, Naikmasur VG. Comparison of radiovisiography and digital volume tomography to direct surgical measurements in the detection of infrabony defects. *J Clin Exp Dent*. 2012 Feb;4(1):e43-7.
85. Zhang W, Foss K, Wang B-Y. A retrospective study on molar furcation assessment via clinical detection, intraoral radiography and cone beam computed tomography. *BMC Oral Health*. 2018 May;18(1):75.
86. Walter C, Kaner D, Berndt DC, Weiger R, Zitzmann NU. Three-dimensional imaging as a pre-operative tool in decision making for furcation surgery. *J Clin Periodontol*. 2009 Mar;36(3):250–7.
87. Pajnigara N, Kolte A, Kolte R, Pajnigara N, Lathiya V. Diagnostic accuracy of cone beam computed tomography in identification and postoperative evaluation of furcation defects. *J Indian Soc Periodontol*. 2016;20(4):386–90.
88. Gupta V, Bains VK, Singh GP, Jhingran R. Clinical and cone beam computed tomography comparison of NovaBone Dental Putty and PerioGlas in the treatment of mandibular Class II furcations. *Indian J Dent Res*. 2014;25(2):166–73.
89. Dutra BC, Oliveira AMSD, Oliveira PAD, Manzi FR, Cortelli SC, Cota LO de M, et al. Effect of 1% sodium alendronate in the non-surgical treatment of periodontal intraosseous defects: a 6-month clinical trial. *J Appl Oral Sci*. 2017;25(3):310–7.
90. Gupta SJ, Jhingran R, Gupta V, Bains VK, Madan R, Rizvi I. Efficacy of platelet-rich fibrin vs. enamel matrix derivative in the treatment of periodontal intrabony defects: a clinical and cone beam computed tomography study. *J Int Acad Periodontol*. 2014 Jul;16(3):86–96.
91. Khosropanah H, Shahidi S, Basri A, Houshyar M. Treatment of Intrabony Defects by DFDBA Alone or in Combination with PRP: A Split-Mouth Randomized Clinical and Three-Dimensional Radiographic Trial. *J Dent (Tehran)*. 2015 Oct;12(10):764–73.
92. Grimard BA, Hoidal MJ, Mills MP, Mellonig JT, Nummikoski P V, Mealey BL. Comparison of clinical, periapical radiograph, and cone beam volume tomography measurement techniques for assessing bone level changes following regenerative periodontal therapy. *J Periodontol*. 2009 Jan;80(1):48–55.
93. Bhavsar N V, Trivedi SR, Dulani K, Brahmbhatt N, Shah S, Chaudhri D. Clinical and radiographic evaluation of effect of risedronate 5 mg as an adjunct to treatment of chronic periodontitis in postmenopausal women (12-month study). *Osteoporos Int*. 2016 Aug;27(8):2611–9.
94. Pajnigara NG, Kolte AP, Kolte RA, Pajnigara NG. Volumetric Assessment of Regenerative Efficacy of Demineralized Freeze-Dried Bone Allograft With or Without Amnion Membrane in Grade II Furcation Defects: A Cone Beam Computed Tomography Study. *Int J Periodontics Restorative Dent*. 2017;37(2):255–62.
95. Nemoto Y, Kubota T, Nohno K, Nezu A, Morozumi T, Yoshie H. Clinical and CBCT Evaluation of Combined Periodontal Regenerative Therapies Using Enamel Matrix Derivative and Deproteinized Bovine Bone Mineral With or Without Collagen Membrane. *Int J Periodontics Restorative Dent*. 2018;38(3):373–81.
96. Haas LF, Zimmermann GS, De Luca Canto G, Flores-Mir C, Correa M. Precision of cone beam CT to assess periodontal bone defects: a systematic review and meta-analysis. *Dentomaxillofac Radiol*. 2018 Feb;47(2):20170084.
97. Walter C, Schmidt JC, Dula K, Sculean A. Cone beam computed tomography (CBCT) for diagnosis and treatment planning in periodontology: A systematic review. *Quintessence Int*. 2016 Jan;47(1):25–37.
98. Choi IGG, Cortes ARG, Arita ES, Georgetti MAP. Comparison of conventional imaging techniques and CBCT for periodontal evaluation: A systematic review. *Imaging Sci Dent*. 2018 Jun;48(2):79–86.
99. Woelber JP, Fleiner J, Rau J, Ratka-Kruger P, Hannig C. Accuracy and Usefulness of CBCT in Periodontology: A Systematic Review of the Literature. *Int J Periodontics Restorative Dent*. 2018;38(2):289–97.

Tables

Table 1. Study search strategy

Database and its supplies	Index terms	Results
Pubmed (US National Library of Medicine (NLM))	Cone beam computed tomography[MeSH]) OR Cone beam computed tomography[Title/Abstract]) OR Cone beam CT[Title/Abstract]) OR CBCT[Title/Abstract]) OR Dental computed tomography[Title/Abstract]) OR Digital volume tomography[Title/Abstract]) OR Volumetric tomography[Title/Abstract]) OR 3D image[Title/Abstract]) OR Three dimensional imaging[Title/Abstract]) OR Imaging, three dimensional[MeSH] AND Periodontitis[MeSH] OR Periodontal disease[Title/Abstract]) OR Periodontal diseases[Title/Abstract]) OR Furcation defects[MeSH]) OR Furcation involvement[Title/Abstract]) OR Alveolar bone loss[MeSH]) OR Intrabony defect[Title/Abstract]	539
Scopus (Elsevier)	TITLE-ABS-KEY(Cone beam computed tomography) or TITLE-ABS-KEY(Cone beam computed tomography) or TITLE-ABS-KEY(Cone beam CT) or TITLE-ABS-KEY(CBCT) or TITLE-ABS-KEY(Dental computed tomography) or TITLE-ABS-KEY(Digital volume tomography) or TITLE-ABS-KEY(Volumetric tomography) or TITLE-ABS-KEY(3D image) or TITLE-ABS-KEY(Three-dimensional imaging) AND TITLE-ABS-KEY(Periodontitis) or TITLE-ABS-KEY(Periodontal disease) or TITLE-ABS-KEY(Periodontal diseases) or TITLE-ABS-KEY(Furcation involvement) or TITLE-ABS-KEY(Furcation defects) or TITLE-ABS-KEY(Intrabony defect)	746
Cochrane (Wiley InterScience)	Cone beam computed tomography[MeSH] or Cone beam computed tomography (word variations ti, ab, kw) or Cone beam CT (word variations ti, ab, kw) or CBCT (word variations ti, ab, kw) or Dental computed tomography (word variations ti, ab, kw) or Digital volume tomography (word variations ti, ab, kw) or Volumetric tomography (word variations ti, ab, kw) or 3D image (word variations ti, ab, kw) or Iamging, three dimensional [MeSH] or AND Periodontitis [MeSH] or Periodontal disease (word variations ti, ab, kw) or Periodontal diseases (word variations ti, ab, kw) or Furcation involvement (word variations ti, ab, kw) or Furcation defects [MeSH] or Alveolar bone loss [MeSH] or Intrabony defect (word variations ti, ab, kw)	71
Web of Science	Cone beam computed tomography (Topic) or Cone beam CT (Topic) or CBCT (Topic) or Dental computed tomography (Topic) or Digital volume tomography (Topic) or Volumetric tomography (Topic) or 3D image (Topic) or Three dimensional imaging (Topic) or AND Periodontitis (Topic) or Periodontal disease (Topic) or Periodontal diseases (Topic) or Furcation defects (Topic) or Furcation involvement (Topic) or Alveolar bone loss (Topic) or Intrabony defect (Topic)	555

Table 2. Studies that were included and their suitable efficacy level

Included studies						Systematic Review	Excluded studies	Exclusion reason
Efficacy levels								
1	2	3	4	5	6			
	Nagao et al. 2006 [80]	Walter et al. 2009 [86]	Pajnigara et al. 2016 [87]	Grimrad et al. 2009 [92]	Walter et al. 2012 [82]	Walter et al. 2016 [97]	Micsh et al. 2006	Ex vivo
	Walter et al. 2010 [77]			Gupta V et al. 2014 [90]		Anter et al. 2016 [33]	Naitoh et al. 2006	Case report
	De Faria et al. 2012 [71]			Gupta SJ et al. 2014 [90]		Nikolic-Jakoba et al. 2016 [6]	Nagao et al. 2007	Did no check for periodontitis
	Fiejo et al. 2012 [72]			Khosropana et al. 2015 [91]		Haas et al. 2018 [96]	Vandenberghe et al. 2007	Ex vivo
	Walter et al. 2012 [82]			Bhavsar et al. 2016 [93]		Choi et al. 2018 [98]	Mol et al. 2008	Ex vivo
	Raichur et al. 2012 [84]			Pajnigara et al. 2017 [94]		Woelber et al. 2018 [99]	Vandenberghe et al. 2008	Ex vivo
	Marinescu et al. 2013 [79]			Dutra et al. 2017 [89]			Noujeim et al. 2009	Ex vivo
	Qiao et al. 2013 [75]			Nemoto et al. 2018 [95]			Leung et al. 2010	Ex vivo
	Haghgoo et al. 2014 [83]						Evangelista et al. 2010	Addressed naturally occurring dehiscence and fenestration in patients with malocclusions
	Banodkar et al. 2015 [68]						Ferreira et al. 2013	Study targets bone coverage
	Cimbaljevic et al. 2015 [69]						DG Pour et al. 2013	Does not involve patients with periodontitis
	Darby et al. 2015 [70]						AlShaer et al. 2013	Does not belong to any level of efficacy
	Li F et al. 2015 [74]						Fliener et al. 2013	Ex vivo
	Guo et al. 2016 [73]						Kamuroglu et al. 2013	Ex vivo
	Zhu et al. 2016 [78]						Amorfini et al. 2014	Does not belong to any level of efficacy
	Suphanantachat et al. 2017 [76]						Sun L et al. 2015	Excluded patients with periodontal diseases
	Padmanabhan et al. 2017 [81]						Zhao et al. 2015	Does not belong to any level of efficacy
	Zhang et al. 2018 [85]						Yang et al. 2015	Does not belong to any level of efficacy
							Bagis et al. 2015	Ex vivo
							Takeshita et al. 2015	Ex vivo
							Kolsuz et al. 2015	Ex vivo
							Kamburoğlu	Animal study and not plaque-

							et al. 2015	induced periodontitis
							Deng et al. 2015	Not in English Language
							Lim et al. 2016	Does not belong to any level of efficacy
							Al-Zahrani et al. 2017	Targets bone density in aggressive periodontitis and normal patients
							Ozcan et al.2017	Does not belong to any level of efficacy
							Saliniero et al.2017	Ex vivo
							Elashiry et al. 2018	Not plaque-induced periodontitis

Table 3. Quality assessment of diagnostic accuracy studies using QUADAS2 tool

Study	Sample	Index test	Reference standard (comparator)	Risk of bias				Applicability concern		
				Patient Selection	Index test	Reference test	Flow and timing	Patient selection	Index test	Reference standard
Nagao et al. 2006 [80]	Developing novel method for alveolar bone resorption assessment caused by periodontitis	Dental CT	Not applicable			N/A				N/A
Walter et al. 2010 [77]	14 patients, five women and nine men, with a mean age of 57.0 years (range 42-81 years) and a diagnosis of generalized chronic periodontitis	CBCT Accuitomo with settings in the range of 74-90 kV, 5-8 mA and voxel sizes in the range of 0.08-0.25 mm	Intrasurgical measurements							
De Faria et al. 2012 [71]	Image records of 39 teeth from 11 adult patients of both genders, aged between 39 and 66 years	CBCT i-CAT with 120 kV and 36.12 mAs. The field of view (FOV) was 6 cm and the voxel size was 0.260.260.2 mm Intraoral radiography (Kodak) with 120 kV and 36.12 mAs. The field of view (FOV) was 6 cm and the voxel size was 0.260.260.2 mm	Not available				N/A			N/A
Fiejo et al. 2012 [72]	6 patients with 12 measurement sites	CBCT i-CAT with voxel size of 0.2 mm and 40 seconds acquisition time	Intrasurgical measurements							
Walter et al. 2012 [82]	12 patients, 3 women and 9 men, with an average age of 57.5 years (range: 41-80 years) and a diagnosis of generalized chronic periodontitis	CBCT 3D Accuotom with volumes of 4 9 4 cm to 6 9 6 cm	Intrasurgical measurements							
Raichur et al. 2012 [84]	7 patient (3 males and 4 females) having moderate to severe periodontitis	Digital volume tomography Kodak 9000 C3D with (exposure parameters were set at 70-74 kV, 10 mA and 10.8 seconds) & Radiovisiography (with a size #2 charged couple device (CCD) intraoral digital sensor\$ and a	Direct measurements with UNC 15 probe							

		standard X-ray unit* operating at 60-63 kV, 8 mA and 0.25-0.32 sec)								
Marinescu et al. 2013 [79]	19 patients with (presenting a total of 25 lower molars with different degrees of furcation defects)	CBCT	Clinical measurements							
Qiao et al. 2013 [75]	15 patients (9 women and 6 men) with an average age of 43.5 years and a diagnosis of generalized chronic periodontitis	CBCT Accuitomo with settings in the range 74–90 kV and 5–8 mA and voxel size of 0.125 × 0.125 × 0.125 mm	Intrasurgical measurements							
Haghgoo et al. 2014 [83]	50 interproximal sites in patients having periodontitis are assessed	CBCT (Newtom 3G, Verona. Italy) and direct digital intraoral radiography (Sopro-La Ciotat-France) was taken.	Intrasurgical findings							
Banodkar et al. 2015 [68]	15 patients with hundred periodontal bone defects	CBCT promax (Planmec) with 90 kv, voxel size-400 μm, current-10 mA, and exposure time-13 s	Intrasurgical measurements							
Cimbaljevic et al 2015 [69]	15 patients (4 men and 11 women aged 35 to 60 years; mean age 44.5 ± 8.4 years) with 174 furcation sites	CBCT unit (SCANORA 3Dx, Soredex) with 80 × 100 mm field of view, 0.25 mm voxel size, 90 kV tube voltage, 10 mA tube current, and 2.4 seconds active scanning time Probing using a Nabers probe (PQ2N, Hu-Friedy)	Not available			N/A				N/A
Darby et al. 2015 [70]	Clinical records (Retrospective) from the Periodontics clinic at the Royal Dental Hospital of Melbourne (RDHM)	CBCT i-CAT with slice thickness 1 mm, voxel size 0.2 mm. 120 kV, 20.7 mAs ⁻¹ , 14.7 s acquisition time)	Not available			N/A				N/A
Li F et al. 2015 [74]	44 patients (22 men and 22 women) with 44 intrabony defects	CBCT New Tom, Verona, Italy (12-17mA and 110 Kv), and digitalperiapical radiography (70kVp and 12-25mA)	Intrasurgical measurements made with a probe (HU-Friedy)							
Guo et al.	6 patients (2	CBCT 3D	Intrasurgical							

2016 [73]	males and 4 females)	Accuitomo with a field of view of 43.4 cm, tube voltage of 75–85 kVp and tube current of 5 mA. The voxel size used was 0.125 × 0.125 × 0.125 mm.	measurements							
Zhu et al. 2016 [78]	11 patients (Thirty-nine sites with degree II FI, classified by probing of 21 maxillary molars, were investigated)	CBCT 3D Accuitomo with volumes of 4 × 4 to 6 × 6 cm, with a setting in the range of 80 kV, 5.0 to 6.3 mA and a voxel size of 0.125 × 0.125 × 0.125 mm	Not available			N/A	N/A			N/A
Suphanantachat et al. 2017 [76]	25 patients	Intraoral radiograph (Kodak) with 70 kV, 7 mA, exposure time 0.2–0.4 s CBCT 3D Accuitomo with volumes of 100 × 100 × 100 mm, 80 kV, 5 mA, exposure time of 17.5 s and a voxel size of 0.25 mm	Not available			N/A				N/A
Padmanabhan et al. 2017 [81]	14 patients (20–60 years) with 25 mandibular molar furcation sites	Intraoral periapical radiography CBCT with 84 kV, 5 mA, 20 s, Voxel size of 180 µm	Intrasurgical measurements							
Zhang et al. 2018 [85]	83 patients with chronic periodontitis	CBCT with a field of view (FOV) of 150 × 90 mm ² . The scans were acquired at 90 kVp, 10 mA, 16 s and a 0.2 mm ³ voxel size with a Kodak 9500 unit (Carestream Health, Inc., Rochester, NY, USA), Intraoral using the unit (Instrumentarium Dental, Charlotte, NC, USA) operating at 70 kVp, 7 mA, and an exposure time corresponding to the exposed area, Clinical measurements	Not available			N/A			N/A	

Good Quality Not clear **N/A** Not applicable

Table 4. CASP checklist for critical appraisal of randomized clinical trials studies

Criteria	Randomized Clinical Trials							
	Gupta V et al. 2014 [90]	Dutra et al. 2017 [89]	Gupta SJ et al. 2014 [90]	Grimrad et al.2009 [92]	Khosropana et al.2015 [91]	Nemoto et al. 2018 [95]	Bhavsar et al. 2016 [93]	Pajnigara et al. 2017 [94]
Did the trial address a clearly focused issue?	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Was the assignment of patients to treatments randomized?	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Were all of the patients who entered the trial properly accounted for at its conclusion?	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Were patients, health workers and study personnel 'blind' to treatment?	Yes	Can't tell*	Can't tell*	Yes	Yes	Yes	Yes	Yes
Were the groups similar at the start of the trial	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Aside from the experimental intervention, were the groups treated equally?	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
How large was the treatment effect?	Satisfactory	Satisfactory	Satisfactory	Satisfactory	Satisfactory	Satisfactory	Satisfactory	Satisfactory
How precise was the estimate of the treatment effect?	Acceptable	Acceptable	Acceptable	Acceptable	Accept-able	Acceptable	Acceptable	Acceptable
Can the results be applied to the local population or in your contex	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Were all clinically important outcomes considered?	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Are the benefits worth the harms and costs?	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes

*Can't tell = cannot tell; criteria in this tool

Table 5. AMSTAR2 checklist for systematic review appraisal

Criteria	Systematic Reviews					
	Haas et al. 2018 [96]	Anter et al. 2016 [33]	Walter et al. 2016 [97]	Nikolic-Jakoba et al. 2016 [6]	Choi et al. 2018 [98]	Woelber et al. 2018 [99]
1. Did the research questions and inclusion criteria for the review include the components of PICO?	Yes	No	Yes	Yes	Yes	Yes
2. Did the report of the review contain an explicit statement that the review methods were established prior to the conduct of the review and did the report justify any significant deviations from the protocol?	Yes	Yes	Partial yes	Yes	Partial yes	Yes
3. Did the review authors explain their reasons for selection of the study designs for inclusion in the review?	Yes	Yes	Yes	Yes	No "Some failed to continue"	Yes
4. Did the review authors use a comprehensive literature search strategy?	Partial yes	Partial yes	No	Yes	Yes	Yes
5. Did the review authors perform study selection in duplicate?	Yes	Yes	No	Yes	Yes	Yes
6. Did the review authors perform data extraction in duplicate?	Yes	No	Yes	Yes	Yes	Yes
7. Did the review authors provide a list of excluded studies and justify the exclusions?	Yes	Yes	Yes	Partial yes	Partial yes	Yes
8. Did the review authors describe the included studies in adequate detail?	Yes	Partial yes	Partial yes	Yes	Yes	Yes
9. Did the review authors use a satisfactory technique for assessing the risk of bias (RoB) in individual studies included in the review?	Yes	Partial yes	No	Yes	No	No
10. Did the review authors report on the sources of funding for the studies included in the review?	No	No	No	No	No	No
11. If meta-analysis was performed, did the review authors use appropriate methods for a statistical combination of results?	Yes	No meta-analysis conducted	No meta-analysis	No meta-analysis conducted	No meta-analysis	No meta-analysis performed
12. If meta-analysis was performed, did the review authors assess the potential impact of RoB on individual studies based on the results of the meta-analysis or other evidence synthesis?	No	No meta-analysis conducted	No meta-analysis	No meta-analysis conducted	No meta-analysis performed	No meta-analysis performed
13. Did the review authors account for RoB in individual studies when interpreting/ discussing the results of the review?	Yes	Yes	No* "No RoB assessed"	Yes	No	No
14. Did the review authors provide a satisfactory explanation for, and discussion of, any heterogeneity observed in the results of the review?	Yes	Yes	Yes	Yes	Yes	Yes
15. If they performed quantitative synthesis, did the review authors carry out an adequate investigation of publication bias (small study bias) and discuss its likely impact on the results of the review?	No	No meta-analysis conducted	No meta-analysis conducted	No meta-analysis conducted	No meta-analysis performed	No meta-analysis performed
16. Did the review authors report any potential sources of conflict of interest, including any funding they received for conducting	No	Yes	No	Yes	No	Yes

*No RoB = no risk of bias assessed

Figures




Figure 1

The results for systematic reviews according to the PRISMA flow chart