

# ISO 10993 Biological Evaluation of Novel Hemostatic Powder – 4SEAL®

**Lukasz Szymanski**

Institute of Genetics and Animal Biotechnology

**Kamila Golaszewska**

Konmex BioLabs

**Anna Wiatrowska**

Konmex Biolabs

**Monika Dropik**

Konmex BioLabs

**Pawel Szymanski**

Konmex BioLabs

**Patrycja Krakowiak**

Konmex BioLabs

**Justyna Wierzchowska**

Konmex BioLabs

**Damian Matak** (✉ [d.matak@konmex.com](mailto:d.matak@konmex.com))

Konmex BioLabs

---

## Research article

**Keywords:** 4SEAL Hemostatic powder, biocompatibility, biological evaluation, ISO 10993, hemostatic, adhesion prevention, polysaccharide

**Posted Date:** October 20th, 2021

**DOI:** <https://doi.org/10.21203/rs.3.rs-957640/v1>

**License:** © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License. [Read Full License](#)

---

**Version of Record:** A version of this preprint was published at Biomaterials Research on April 5th, 2022. See the published version at <https://doi.org/10.1186/s40824-022-00258-6>.

# Abstract

## *Background:*

Hemostasis plays a crucial role during every surgery allowing for a bloodless operating field. Fast and effective surgery leads to a reduced risk of postoperative complications. One of the latest method for achieving homeostasis is the use of natural polysaccharide-based hemostatic powders. The study aimed to evaluate the biocompatibility according to the ISO 10993 standards of 4SEAL® Hemostatic powder.

## *Methods:*

*Chemical characterization (Headspace GC-MS, GC-MS, and ICP-MS), cytotoxicity, genotoxicity (MLA and AMES), endotoxin contamination, sensitization potential, intracutaneous reactivity, acute and subacute systemic toxicity with implantation, and pyrogenicity were evaluated to investigate the biocompatibility of the 4SEAL® Hemostatic powder. Studies were conducted according to ISO 10993 standards.*

## *Results:*

The biocompatibility requirements with accordance to ISO 10993-1 for 4SEAL® Hemostatic powder were met.

## *Conclusions:*

4SEAL® Hemostatic powder is a promising new hemostatic agent with a wide range of potential applications and excellent biocompatibility.

# Introduction

Throughout every surgery, hemostasis plays a crucial role by allowing for a bloodless operating field. Fast and effective surgery leads to a reduced risk of postoperative complications. However, intraoperative bleeding or postoperative hemorrhage is a significant concern, and it presents significant perioperative morbidity [1]. Hemostasis failure leads to prolonged hospitalization, increased infection risk, and potential further surgical interventions. Hemostasis is a process that prevents and stops bleeding. Biologically, it is a complex process requiring precisely coordinated activation of platelets and plasma clotting factors to form a platelet-fibrin clot [2]. It can be divided into two distinct processes, primary and secondary hemostasis. Primary hemostasis results in the formation of soft platelet plugs, which are stabilized and cross-linked during secondary hemostasis. Of central importance in both primary and secondary hemostasis is the activation of the clotting cascade, which can be broken down into two pathways [3]. The intrinsic pathway is activated by collagen, which is exposed when a blood vessel is damaged. The extrinsic pathway is similarly activated by tissue damage and the resultant release of tissue factors. The intrinsic and extrinsic pathways converge into the common pathway, which begins with the conversion of Factor X to Xa, and ultimately results in the conversion of prothrombin to thrombin, which is integral in clot stabilization via fibrin [4,5].

An appropriate method for achieving hemostasis depends on various factors such as location, visibility, the possibility of access, and bleeding intensity. The conventional methods include electrocoagulation, suturing, and pressure application. On the other hand, natural polysaccharide-based hemostatic powders gained popularity due to their ease of application and effectiveness in controlling mild to moderate bleeding. In the present study, the biocompatibility of 4SEAL® Hemostatic powder according to the ISO 10993 has been investigated.

## **Study product**

4SEAL® Hemostatic powder (Grena Biomed LTD) is a hemostatic medical device consisting of natural polysaccharide-based fine powder particles and a delivery applicator. The product is ready to use, thus can be used immediately. 4SEAL® Hemostatic powder is used to stop bleeding during surgical procedures or traumatic injuries. The ultra-hydrophilic particles are derived from purified starch. 4SEAL® Hemostatic powder absorbs water rapidly from the blood, which causes a high concentration of thrombocytes, leukocytes, red blood cells, and coagulation proteins, thus inducing instantly primary hemostasis and accelerating

the blood clotting process. For adhesion prevention, 4SEAL® Hemostat is indicated when the formation of postoperative adhesion is to be prevented after surgical interventions in cavities covered by mesothelium. The product is plant-based and contains no animal or human-derived materials. The absorption process of 4SEAL® Hemostatic powder begins immediately and is dependent on several factors, including the amount applied and the site of use. Generally, its complete absorption takes a few days. Studied product is shown in the Fig. 1.

## Material And Methods

Detailed data for every experiment and additional information regarding materials and methods are presented in Supplement 1. Cell lines were purchased from ATCC, reagents for cell culture were purchased from Thermo Fisher Scientific, Poland, and all chemical compounds were purchased from Sigma, Poland unless otherwise indicated.

### Extraction

Extractions were performed according to ISO 10993-12 [6]. Extraction media for each experiment were chosen based on the appropriate ISO norm for the study. Briefly, the extractions were prepared by incubating the test material with a suitable extraction medium at 50±1°C for 72±2 hours for AMES, ICP MS, GC-MS, and LC-MS studies or at 37±1°C for 72±2 hours for other studies unless otherwise indicated. The extraction volume was derived from Table 1 – Standard surface areas and extract liquid volumes, ISO 10993-12 and determined at 0,2 g/mL [6]. The extracts were not centrifuged, filtered, or otherwise altered prior to dosing. The extract was clear without the presence of any particulates. The extracts were used within 24h of preparation. Since the material absorbs vehicles, the extraction vehicle that each 0,1 g of material absorbs was determined. During the extraction, the additional volume of the extraction vehicle was added to the mixture.

### Chemical characterization

Per ISO 10993-18, a semi-quantitative analysis of VOC (volatile organic compound) in product and SVOC (semi-volatile organic compound) in 4SEAL® Hemostatic powder water extract was performed [7]. QP2010 Ultra gas chromatograph and QP-5000 mass spectrometer (Shimadzu) were used for analysis. In addition, quantitative analysis of elements concentration in 4SEAL® Hemostatic powder water extract was performed using NexION 300D (Perkin Elmer). As per ICH Q3D (R1), the concentration of the following elements was examined: Cd, Pb, As, Hg, Co, V, Ni, Tl, Au, Pd, Ir, Os, Rh, Ru, Se, Ag, Pt, Li, Sb, Ba, Mo, Cu, Sn, Cr [8].

For the analysis of VOCs and SVOCs, the Analytical Evaluation Threshold (AET) was calculated according to the following formula:

$$AET \left( \frac{\mu g}{ml} \right) = \frac{DBT * \frac{A}{BC}}{UF}$$

where:

A – number of devices extracted

B – volume of the extract

C – clinical exposure to medical device per day

DBT – the dose based threshold (TTC)

UF – uncertainty factor

Assuming an uncertainty factor of 2, ten 5 g bottles of 4SEAL® Hemostatic powder per patient usage, and Threshold of Toxicological Concern of 120 µg/day, the AET was calculated to be 0,240 µg/mL.

For the analysis of each element concentration, Total Element Exposure (TEE) was calculated assuming device patient usage of 50 g per day. TEE was calculated as follows:

$$TEE[\mu g] = \frac{\text{Element concentration} * \text{number of devices per patient}}{\text{number of devices used for extraction} * \frac{1000}{\text{extraction volume}}}$$

## Cytotoxicity

Cytotoxicity was evaluated quantitatively using the MTT method based on the ISO 10993-5 and ISO 10993-12 [6,9]. Briefly, 4SEAL® Hemostatic powder was extracted in single strength MEM at 37±1 °C for 24±1 hours. Following the extraction, quadruplicate monolayers of L-929 mouse fibroblast cells were dosed with: 100 %, 50 %, 33 %, and 25 % extracts and incubated at 37±1°C, 5±1% CO<sub>2</sub>, 95% humidity for 24±1 hours. Following the incubation, 50 µL of the MTT solution, prepared just before use, were dispensed in each well and incubated for 120 minutes at 37±1°C, 5±1% CO<sub>2</sub>, 95% humidity. Following the incubation, MTT solution was replaced with 100 µL isopropanol and incubated for 10 min in 37±1°C, 5±1% CO<sub>2</sub>, 95% humidity. Finally, the optical density was measured by absorption at 570 nm (650 nm reference). The percent viability was determined from the blanks

## Genotoxicity

Extraction of 4SEAL® Hemostatic powder for genotoxicity studies.

The amount of extractables was assessed by a pre-experiment "Determination of Extractables" according to ISO 10993-3 [10]. Based on the results, Method C – extraction according to ISO 10993-12 was chosen [6]. The extraction was conducted using an appropriate extraction vehicle.

## Mouse Lymphoma Assay - MLA

Based on the ISO 10993-3, ISO 10993-12, ISO 10993-33, and OECD Test No 490, the 4SEAL® Hemostatic powder genotoxicity was evaluated using Mouse Lymphoma Assay [6,10–12]. In short, mycoplasma-free L5178Y TK+/-3.7.2C cells were cultured (37±1°C, 5±1% CO<sub>2</sub>, 95% humidity) in the F10 medium to the sufficient number and cleansed using THMG for 1 day and then THG for 2 days. The cleansed cells were then used in the experiment. 6 \* 10<sup>5</sup> (for 4h treatment) or 4 \* 10<sup>5</sup> (for 24h treatment) cells were exposed to the 10 ml of the 100% sample extract (worst case scenario), appropriate positive control, or negative control. The cells treatment was performed for 4h with and without the presence of 1% liver Aroclor-induced S9 fraction and for 24h without S9 fraction in 37±1°C, 5±1% CO<sub>2</sub>, 95% humidity. For each condition, duplicate test sample, duplicate negative control, and one positive control were prepared. After the 4h treatment, the cells were centrifuged and washed twice with fresh medium and then resuspended in the 20 ml of the F10 medium. After additional 20 hours, the cells were counted and resuspended in a fresh F10 medium at the concentration of 2 \* 10<sup>5</sup>. The cells were incubated for 24h (37±1°C, 5±1% CO<sub>2</sub>, 95% humidity) and recounted. After the 24h treatment, the cells were counted, washed twice, and resuspended in F10 fresh medium at the concentration of 2 \* 10<sup>5</sup>. The cells were incubated for 24h (37±1°C, 5±1% CO<sub>2</sub>, 95% humidity) and recounted. The number of cells counted was used to calculate total suspension growth (TSG) according to ISO 10993-33. After the expression period, the cell's relative plating efficiency (RPE; percentage plating efficiency of the test group in relation to the negative control) was determined by seeding a statistical number of 1,6 cells/well in two 96-well plates in F20 medium. The cells were incubated for 14 days at 37±1°C in the humidified atmosphere in the presence of 5% CO<sub>2</sub>. Analysis of results was based on the number of cultures without cell growth compared to the total number of cultures seeded. Relative suspension growth and relative total growth (RSG and RTG; RTG = RSG x RPE / 100) of the treated cell cultures were calculated according to the ISO 10993-33. Additionally, cultures were seeded in a selective medium. Cells from each experimental group were seeded in four 96-well plates at a density of 2000 cells/well in 200 µl selective F20 medium with TFT. The plates were scored after an incubation period of 14 days at 37±1°C in the humidified atmosphere in the presence of 5% CO<sub>2</sub>. Small colonies were counted separately. Small colonies are defined as less than a quarter of the diameter of the well, while large colonies are more than a quarter of the diameter of the

well. The mutation frequencies were calculated from the data obtained from cultures used for the plating efficiency (cultures with non-selective medium) and those used for selection (cultures with selective medium) according to the following formula:

$$\text{Mutant frequency per } 10^6 \text{ cells} = \frac{\frac{-\ln(\text{empty wells of mutant selection plates}/384)}{2000}}{\frac{-\ln(\text{empty wells of non selection plates}/192)}{1,6}} * 10^6$$

### Bacterial Reverse Mutation Test - AMES

Genotoxicity of 4SEAL® Hemostatic powder was evaluated using commercially available Bacterial Reverse Mutation Test AMES Penta 2 (Xenometrix) according to ISO 10993-3, ISO 10993-12, ISO 10993-33, and OECD Test No. 471 [6,10,11,13].

Bacteria were exposed to the 25 µl of full strength extracts of the test material as well as positive (Table 1) and negative controls for 135 minutes in a medium containing sufficient histidine (*S. typhimurium*) or tryptophan (*E. coli*) to support approximately two cell divisions. The volume of extract added was based on the ISO 10993-33 and kit manufacturers' documentation. After exposure, the cultures were diluted in a pH indicator medium lacking histidine or tryptophan and aliquoted into 48 wells of a 384-well plate. After two days, cells that have undergone reversion to amino acid prototrophy grow into colonies. Bacterial metabolism reduces the pH of the medium, changing the color of that well. The number of wells containing revertant colonies were counted for each group (test article and positive control) and compared to a solvent (negative) control. Samples were prepared in triplicate to allow for statistical analysis of the data. The mutagenic potential of samples was assessed directly and in the presence of 4,5% of liver Aroclor-induced S9 fraction. Baseline, fold increase over baseline value, and binomial B-value were calculated using an excel spreadsheet provided by the manufacturer. The baseline is calculated as a mean plus standard deviation of the negative control. Fold increase over baseline is calculated by dividing the mean number of positive wells for a sample by the baseline value. The binomial B-value indicates the probability that spontaneous mutation events occur. For example, a binomial B-value  $\geq 0,99$  indicates that chances are  $\leq 1\%$  that this Result is due to spontaneous mutation. If both fold increase  $\geq 2$  and binomial B-value  $\geq 0,99$  occur for a test sample in specific conditions (strain, +/- S9 fraction), it should be considered mutagenic.

Table 1. Positive controls list for AMES test.

Strain	Substance	
	Without S9 fraction	With S9 fraction
<i>Salmonella typhimurium</i> TA98	2-nitrofluorene (2-NF)	2-aminoanthracene (2-AA)
<i>Salmonella typhimurium</i> TA100	4-nitroquinoline N-oxide (4-NQO)	2-AA
<i>Salmonella typhimurium</i> TA1535	N4-aminocytidine (N4-ACT)	2-AA
<i>Salmonella typhimurium</i> TA1537	9-aminoacridine (9-AA)	2-AA
<i>E. coli</i> WP2 uvrA[pKM101]	4-NQO	2-aminofluorene (2-AF)

### Endotoxins

Endotoxins were measured using Pierce Chromogenic Endotoxin Quant K, which is in regard to 85. Bacterial Endotoxin Test, U.S. Pharmacopoeia [14]. The 4SEAL® Hemostatic powder was extracted in water for injection using an extraction ratio of 0,2 g/mL. According to the manufacturer's instruction, the standard curve was prepared ( $R^2=0,9887$ ) and is shown in Figure 2. Internal validation of the experiment was performed by spiking the samples with 0,05 EU/mL of endotoxin. The unspiked sample and the spiked sample were assayed to determine the respective endotoxin concentrations. For the test to be valid, the difference between the two calculated endotoxin values should equal the known (0,5 EU/mL) concentration of the spike  $\pm 25\%$ .

## Sensitization

The sensitization potential of the 4SEAL® Hemostatic powder was analyzed according to the ISO 10993-10 using the Local Lymph Node Assay (LLNA) [15]. Briefly, 4SEAL® Hemostatic powder was extracted using Acetone: Olive Oil 4:1. 15 adult, albino, healthy house mice (*Mus musculus*) of BALB/c strain were randomly assigned to solvent control, positive control, or study group. The test samples, solvents control, and positive control were applied to the dorsal side of both ears of designated mice at a dose of 25 µl / day for three consecutive days. 48h after the last extract application, mice were injected with 0.5 mL of BrdU (10 mg/mL) in phosphate-buffered saline (PBS) solution intra-peritoneally. 24 h after BrdU injection, animals were sacrificed, and auricular lymph nodes were harvested. A single-cell suspension of lymph node cells (LNC) was prepared from each mouse by gentle mechanical disaggregation through a disposable µ70 nylon cell strainer. In each case, the target volume of the LNC suspension was adjusted to 15 mL. The incorporation of BrdU was measured using the Colorimetric BrdU Cell Proliferation ELISA Kit (Abcam) according to the manufacturer's recommendations. 50 µL of each LNC was transferred in triplicate into a 96-well cell culture plate. PBS was used as a blank control. The colored reaction product was quantified using the µQuant spectrophotometer with dual-wavelength of 450/550 nm. BrdU labeling index was calculated according to the formula:

$$\text{BrdU labelling index} = (\text{ABS}_{\text{em}} - \text{ABS blank}_{\text{em}}) - (\text{ABS}_{\text{ref}} - \text{ABS blank}_{\text{ref}})$$

em - emission wavelength; ref - reference wavelength.

For the test sample and positive control Sensitization Index (SI) was calculated according to the formula:

$$SI = \frac{\text{BrdU labelling index of the sample}}{\text{BrdU labelling index of the negative control}}$$

For the test to be valid, the SI of positive control (PC) must be higher than 2.

## Intracutaneous reactivity

The study was conducted according to ISO 10993-10 [15]. The test article was extracted using Sodium Chloride and Cottonseed Oil as described above. Before the treatment, the fur on the animal's back on both sides of the spinal column was closely clipped over a sufficiently large test area, avoiding mechanical irritation and trauma. Then, 0,2 ml of the polar (Sodium Chloride) and non-polar (Cottonseed Oil) extracts were injected intracutaneously at five sites on one side of each rabbit. Similarly, 0.2 ml of the polar and non-polar solvent controls were injected intracutaneously on five sites of the contralateral side of each rabbit. The animals were observed immediately after injection, 24±2, 48±2, and 72±2 hours after the treatment to evaluate the signs of local reaction. Injection sites were examined for evidence of any tissue reaction such as erythema, oedema, and eschar. Tested and control sites were scored according to the Table 2 below.

Table 2. The grading system for intracutaneous (intradermal) reactions.

Reaction	Numerical grading
<b>Erythema and eschar formation</b>	
No erythema	0
Very slight erythema (barely perceptible)	1
Well defined erythema	2
Moderate erythema	3
Severe erythema (beet redness) to eschar formation preventing grading of erythema	4
<b>Oedema formation</b>	
No oedema	0
Very slight oedema (barely perceptible)	1
Well defined oedema (edges of area well defined by define raising)	2
Moderate oedema (edges raised approximately 1 mm)	3
Severe oedema (raised more than 1 mm and extended beyond exposure area)	4

After the 72±2 h grading, all erythema grades plus oedema grades (at 24±2 h, 48±2 h, and 72±2 h) are separately summed for each test sample or blank for each animal. To calculate the score of a test sample or blank on each animal, divide each totals by

15 (3 scoring time points x 5 test or blank sample injection sites). To determine the overall mean score for each test sample and each corresponding blank, add the scores for the three animals and divide them by three. The final test sample score is obtained by subtracting the blank score from the test sample score. The acceptance criteria are met if the final test score is 1.0 or less.

### **Acute Systemic Toxicity**

The study was conducted according to ISO 10993-10 [15]. Four groups of 5 animals were injected with 50 ml/kg of Sodium Chloride extract, Cottonseed Oil extract, the polar and non-polar solvent controls. Polar and non-polar extracts and solvents controls were injected intraperitoneal. Animals underwent a clinical examination and were weighted 24±2 h, 48±2 h, 72±2 h after injection. 72±2 h after injection, animals were euthanized.

### **Subacute toxicity combined with implantation**

Based on ISO 10993-6 and ISO 10993-11, 4SEAL® Hemostatic powder was evaluated for subacute toxicity combined with implantation using Starsil® Hemostat as reference material [16,17]. Cannulas from peripheral venous access catheters were divided into 10 mm pieces. Pieces were divided into two groups – the control and test group. Each piece from the control group was filled with approx. 0,3 g of Starsil® Hemostat while each piece from the test group was filled with 0,3 g of 4SEAL® Hemostatic powder. Before animal treatment, the fur on the animal's back was closely clipped over a sufficiently large test area, avoiding mechanical irritation and trauma. Place of implantation was disinfected by iodine solution. Surgery was performed under general anesthesia using isoflurane when animals received analgesic – subcutaneously injected butorphanol (2 mg/kg). An incision was made on the skin in a paraspinal line. Implants were placed on both flanks of the animal at equal intervals, in separate pockets in subcutaneous tissue. Eight pieces per rat of test or control article were implanted. Control and test materials had contact with surrounding tissue only in the base of cylindrical implants. Wounds were closed using non-resorbable threads. After implantation, each animal was injected subcutaneous meloxicam (1 mg/kg). Animals were housed separately for a week until wounds were healed. For 3 days after implantation, each animal was injected subcutaneous meloxicam (1 mg/kg). Animal's weight was observed 1, 2, 3, 7, 14, 21, 28 days after implantation. On the 27<sup>th</sup> day of the experiment, urine samples were collected. After 28 days, animals were premedicated with Ketamine/Xylazine (100 mg/kg – Ketamine, 10 mg/kg – Xylazine) to collect blood samples and then killed by CO<sub>2</sub>. Routine hematology and clinical chemistry were conducted on all animals at the end of the exposure period. Animals were anesthetized with Ketamine/Xylazine, and blood was drawn into tubes with K2-EDTA for hematology and heparin for clinical chemistry. Hemoglobin, PCV, RBC, reticulocytes, thrombocytes, and total WBC were determined with a hematology analyzer (Scil, Germany). Plasma concentrations of glucose, ALP, ALAT, ASAT, GGT, total protein, albumin, urea, creatinine, total bilirubin, total cholesterol, triglycerides, phospholipids, Ca<sup>2+</sup>, Na<sup>+</sup>, K<sup>+</sup>, Cl<sup>-</sup> and inorganic phosphate were determined using a biochemical analyzer (Fujifilm, Poland).

According to 4SEAL® Hemostatic powder Instruction For Use, maximal patient exposition to the hemostatic powder is 50g. Statistical human weight is 60 kg. Every animal was implanted with 8 implants containing 0,3g of hemostatic powder each. The dose of test articles for a single rat was more than 10x of the maximal human dose.

### **Gross necropsy**

After animal euthanasia, a gross necropsy was performed on all animals. The following organs were weighed (paired organs together) after dissection: adrenals, brain, lungs, heart, kidneys, liver, ovaries, spleen, testes. The organ-to-body weight ratios (relative organ weights) were calculated from the rats' absolute organ weights and the terminal body weight.

Samples of the weighed organs and the colon, lymph nodes, skin, lungs, mammary gland, peripheral nerve (sciatic), esophagus, parathyroid, pituitary, prostate, rectum, small intestines (duodenum, ileum, jejunum), sternum with bone marrow, stomach, thyroid, trachea with bronchi, urinary bladder, uterus, vagina, places of implantation and all gross lesions were preserved in a neutral aqueous phosphate-buffered 4% solution of formaldehyde. Histopathologic analysis from organs: brain, lungs, heart, liver, kidneys, adrenals, ovaries/testis, sternum, muscle, the skin was conducted on 5 µm sections of paraffin-embedded tissues, stained with hematoxylin and eosin, of the preserved organs from two representative animals per sex from control and test group

by light microscopy. Each place of subcutaneous implantation was examined under a microscope and evaluated based on the guidelines provided in Table 3.

Table 3. Guidelines of histological evaluation system of place of implantation – tissue response

Histologic feature	Score				
	0	1	2	3	4
Inflammatory cell type/response — Polymorphonuclear cells – Lymphocytes	0	Rare, 1 to 5/hpf <sup>a</sup>	Rare, 5 to 10/hpf <sup>a</sup>	Moderate infiltrate	Marked infiltrate
Plasma cells					
Macrophages/gitter cells					
Multinucleated giant cells (MGC)	0	Rare, 1 to 2/hpf	Rare, 3 to 5/hpf		
Necrosis	0	Minimal	Mild	Moderate infiltrate	Marked
Neovascularization	0	Minimal capillary proliferation, focal, 1 to 3 buds	Groups of 4 to 7 capillaries with supporting fibroblastic structures	Broad band of capillaries with supporting fibroblastic structures	Extensive band of capillaries with supporting fibroblastic structures
Fibrosis	0	Narrow band	Moderately thick band	Thick band	Extensive band
Astrocytosis/fatty infiltration					

<sup>a</sup> hpf=high-powered (400x) field.

## Pyrogenicity

### Rabbit selection

Rabbits used for the study were submitted to a negative pyrogen test within 14 days preceding the test (with a rest period of a minimum of 3 days following the negative pyrogen test).

### Determination of the Initial Temperature

Previously weighted rabbits were placed in a restrainer, and a thermometric rectal probe was inserted at not less than 7.5 cm but not more than 9 cm. The temperature of each rabbit was recorded every 30 minutes for 90 minutes before injection. The rabbits which showed a temperature variation two successive readings higher than 0.2°C during the initial temperature determination or which showed a temperature higher than 39.6 °C or lower than 38.2 °C were not injected. The initial temperature of each rabbit was determined as the mean of two temperatures recorded at intervals of 30 minutes before the injection. In the group, the difference between the three initial temperatures did not exceed 1°C.

### Rabbit Injection and Follow up

After extraction, the tested solution was warmed to about 38,5°C and injected intravenously via the marginal ear vein at a dose of 10 ml/kg of body weight. The temperature of each rabbit was recorded every 30 minutes for 3 hours after injection. The

maximum rise (compared to the initial temperature) of each rabbit was determined at the end of the test. Acceptance criteria for the test are presented in Table 4.

Table 4. Criteria of acceptance for pyrogenicity test.

Number of rabbits	Product passes if the summary response does not exceed	Product fails if the summary response exceeds
3	1.15° C	2.65° C
6	2.80° C	4.30° C
9	4.45° C	5.95° C
12	6.60° C	6.60° C

## Results

### Chemical characterization

Determination of extraction conditions for exhaustive extraction of 4SEAL® Hemostatic powder has revealed that isopropanol and hexane cause product degradation and vehicle color change. Therefore, as per ISO 10993-18, only water extract was analyzed.

No VOCs above AET were identified.

No SVOCs above AET were identified.

No elements above the limit were identified. The results are presented in Table 5.

Table 5. ICP MS results of 4SEAL® Hemostatic powder.

Analyzed element	Limit of detection (LOD) [µg/L]	Result [µg/L]	Total Element Exposure [µg]	Parenteral PDE [µg/day]
Cd	1	< 1	< 0,25	2
Pb	2,5	< 2,5	< 0,625	5
As	30	< 30	< 7,5	15
Hg	1	< 1	< 0,25	3
Co	2,5	< 2,5	< 0,625	5
V	24	< 24	< 6	10
Ni	10	< 10	< 2,5	20
Tl	5	< 5	< 1,25	8
Au	5	< 5	< 1,25	100
Pd	5	< 5	< 1,25	10
Ir	5	< 5	< 1,25	10
Os	5	< 5	< 1,25	10
Rh	5	< 5	< 1,25	10
Ru	5	< 5	< 1,25	10
Se	10	< 10	< 2,5	80
Ag	5	< 5	< 1,25	10
Pt	5	< 5	< 1,25	10
Li	10	< 10	< 2,5	250
Sb	10	< 10	< 2,5	90
Ba	10	< 10	< 2,5	700
Mo	10	< 10	< 2,5	1500
Cu	130	< 130	< 32,5	300
Sn	10	< 10	< 2,5	600
Cr	60	< 60	< 15	1100

### Cytotoxicity

4SEAL® Hemostatic powder cell culture medium extract and its dilutions showed no cytotoxic potential to L-929 mouse fibroblasts using the quantitative MTT method. The cellular response observed from the positive and negative controls, systemic cell seeding errors, and absolute value of optical density confirmed the suitability of the test system. Results of cytotoxicity testing are presented in Table 6.

Table 6. Cytotoxicity results.

<b>Material</b>	<b>Percent Viability [%]</b>	<b>System Suitability</b>
Positive control	0,60	Met criteria
Negative control	102,81	Met criteria
4SEAL® (1x)	99,75	No Cytotoxic Potential
4SEAL® (2x)	103,2	No Cytotoxic Potential
4SEAL® (3x)	101,7	No Cytotoxic Potential
4SEAL® (4x)	110,1	No Cytotoxic Potential

<b>Quality check of assay</b>	<b>Result</b>	<b>System Suitability</b>
Absolute value of optical density (OD <sub>570</sub> )	0,997	Met criteria
Systematic cell seeding errors	0,04 %	Met criteria

### Genotoxicity (AMES and MLA)

#### AMES

Every strain employed in the test, both with and without S9 fraction, passed internal quality controls. 4SEAL® Hemostatic powder showed an unclear mutagenic effect only when exposed to the TA1535 strain with the presence of the S9 fraction. Results of cytotoxicity testing are presented in Table 7.

Table 7. AMES assay results.

<b>Strain</b>	<b>without S9</b>			<b>with S9</b>		
	Baseline	Fold increase over baseline	Binomial B-value	Baseline	Fold increase over baseline	Binomial B-value
TA98	1,00	1,33	0,82	3,75	0,18	0,12
TA100	9,00	0,81	0,17	14,61	0,64	0,19
TA1535	3,91	0,68	0,32	1,00	17,00	1,00
TA1537	1,82	0,55	0,86	2,00	0,50	0,65
E.coli uvrA[pKM101]	9,39	0,46	0,08	10,0	0,40	0,00

Data has been analyzed and summarized in the Table 8 below. No precipitation or toxicity was observed in this study.

Table 8. AMES overall results.

Strain	Mutagenic data points		Overall Result for sample	Solvent control		Positive control	
	w/o S9	with S9		w/o S9	with S9	w/o S9	with S9
TA98	No	No	Probably not mutagenic	PASS	PASS	PASS	PASS
TA100	No	No	Probably not mutagenic	PASS	PASS	PASS	PASS
TA1535	No	Yes	Unclear, needs further evaluation	PASS	PASS	PASS	PASS
TA1537	No	No	Probably not mutagenic	PASS	PASS	PASS	PASS
E. coli uvrA[pKM101]	No	No	Probably not mutagenic	PASS	PASS	PASS	PASS

### Mouse Lymphoma Assay (MLA)

A sample is considered mutagenic if the increase in MF is above the Global Evaluation Factor that equals 126 ( $\times 10^{-6}$ ) over the negative control. The acceptance criteria for MLA and summarized results of the reliability check of the assay are presented in Supplementary Materials. No precipitation or toxicity was observed in this study. MLA results are presented in Tables 9-14.

Table 9. Toxicity data, 4h exposure, without metabolic activation.

Sample	Number of cells seeded ( $\times 10^5$ )	Number of cells 24h after treatment ( $\times 10^5$ )	Number of cells 48h after treatment ( $\times 10^5$ )	Total suspension growth	Relative Suspension Growth (RSG) [%]	Plating Efficiency [%]	Relative Plating Efficiency (RPE) [%]	Relative Total Growth (RTG) [%]
NC1	3	11,02	9,03	16,59	100,00	93,59	100,00	100,00
NC2	3	10,56	9,11	16,03				
PC	3	8,62	7,67	11,02	67,56	65,80	70,31	47,50
4SEAL® 1	3	10,38	8,97	15,52	95,15	92,08	98,39	93,62
4SEAL® 2	3	10,62	9,02	15,97	97,89	87,96	93,99	92,01

PC – positive control, NC – negative control

Table 10. Toxicity data, 4h exposure, with metabolic activation.

Sample	Number of cells seeded ( $\times 10^5$ )	Number of cells 24h after treatment ( $\times 10^5$ )	Number of cells 48h after treatment ( $\times 10^5$ )	Total suspension growth	Relative Suspension Growth (RSG) [%]	Plating Efficiency [%]	Relative Plating Efficiency (RPE) [%]	Relative Total Growth (RTG) [%]
NC1	3	10,23	9,12	15,55	100,00	103,91	100,00	100,00
NC2	3	9,89	9,21	15,18				
PC	3	8,67	7,27	10,51	68,37	61,30	59,00	40,33
4SEAL® 1	3	9,78	8,93	14,56	94,73	86,64	92,58	87,70
4SEAL® 2	3	9,83	9,01	14,76	96,07	84,09	89,86	86,32

PC – positive control, NC – negative control

Table 11. Toxicity data, 24h exposure, without metabolic activation.

Sample	Number of cells seeded (*10 <sup>5</sup> )	Number of cells 24h after treatment (*10 <sup>5</sup> )	Number of cells 48h after treatment (*10 <sup>5</sup> )	Total suspension growth	Relative Suspension Growth (RSG) [%]	Plating Efficiency [%]	Relative Plating Efficiency (RPE) [%]	Relative Total Growth (RTG) [%]
NC1	2	11,18	9,73	130,95	100,00	90,82	100,00	100,00
NC2	2	10,95	9,62	129,57				
PC	2	8,58	8,72	72,67	55,79	64,87	71,43	39,85
4SEAL® 1	2	10,79	8,96	108,52	83,31	86,64	95,41	79,49
4SEAL® 2	2	10,61	9,11	106,93	82,09	87,96	96,85	79,51

PC – positive control, NC – negative control

Table 12. Mutagenicity data, 4h exposure, without metabolic activation.

Sample	Number of large colonies	Number of small colonies	Mutant frequency (*10 <sup>-6</sup> )	Small colonies [%]	Small colonies mutant frequency (*10 <sup>-6</sup> )	Mutagenicity
NC1	78	2	128,82	2,50	3,22	N/A
NC2	64	2	97,72	3,03	2,96	N/A
PC	130	60	518,83	31,58	163,84	Mutagenic
4SEAL® 1	64	2	102,40	3,03	3,10	Not Mutagenic
4SEAL® 2	58	3	98,34	4,92	4,84	Not Mutagenic

PC – positive control, NC – negative control

Table 13. Mutagenicity data, 4h exposure, with metabolic activation.

Sample	Number of large colonies	Number of small colonies	Mutant frequency (*10 <sup>-6</sup> )	Small colonies [%]	Small colonies mutant frequency (*10 <sup>-6</sup> )	Mutagenicity
NC1	85	2	128,94	2,30	2,96	N/A
NC2	80	2	111,01	2,44	2,71	N/A
PC	54	88	376,58	61,97	233,38	Mutagenic
4SEAL® 1	77	4	136,72	4,94	6,75	Not Mutagenic
4SEAL® 2	62	6	115,89	8,82	10,23	Not Mutagenic

PC – positive control, NC – negative control

Table 14. Mutagenicity data, 24h exposure, without metabolic activation.

Sample	Number of large colonies	Number of small colonies	Mutant frequency (*10 <sup>-6</sup> )	Small colonies [%]	Small colonies mutant frequency (*10 <sup>-6</sup> )	Mutagenicity
NC1	78	2	122,97	2,50	3,07	N/A
NC2	64	2	108,83	3,03	3,30	N/A
PC	130	90	655,71	40,91	268,25	Mutagenic
4SEAL® 1	64	2	108,83	3,03	3,30	Not Mutagenic
4SEAL® 2	58	3	98,34	4,92	4,84	Not Mutagenic

PC – positive control, NC – negative control

### Endotoxins

Endotoxin testing results per mL are presented in Table 15 and per 1 g finished device are presented in Table 16.

Table 15. Endotoxins concentration results.

	4SEAL®	Spike
Concentration [EU/mL]	0,103	0,695

Table 16. Endotoxins concentration per device.

	4SEAL®
Endotoxin content per device (1 g) [EU]	0,514

### Sensitization

Each day animals were observed for signs of toxicity or skin irritation.

None of the test or control animals exhibited overt signs of toxicity, enumerated in Annex C, Table C.1 – Common clinical signs and observations, ISO 10993 – 11, at any observation points. Furthermore, none of the animals treated with the test sample shows a significantly greater biological reactivity during the observation period than animals treated with vehicle control. None of the animals die, none of the animals' lost 10% or more bodyweight. Animals showed minor signs of skin irritation – slight erythema was observed in all animals. Summarized results are presented in Table 17 and Table 18.

Table 17. Change in body weight.

Group	Average body weight change [%]
Negative control – acetone: olive oil (4:1 v/v)	-2,76
Extract of 4SEAL® – acetone: olive oil (4:1 v/v)	-0,40
Positive control - 25% HCA in acetone: olive oil (4:1 v/v)	1,31

Table 18. Absorbance results of the BrdU analysis.

	Mean absorbance		BrdU Index	SI
	A450	A550		
PC	0,177	0,042	0,108	2,440
NC acetone: olive oil	0,106	0,044	0,033	N/A
4SEAL® acetone: olive oil	0,125	0,044	0,048	1,377
Blank	0,054	0,042	N/A	N/A

NC – Negative control; PC – Positive control;

### Intracutaneous reactivity

The Primary Irritation Index for sodium chloride and cottonseed oil extracts was calculated by subtracting the control group's total Primary Irritation Score (PIS) from the total PIS of the study group. For sodium chloride and cottonseed oil extracts of 4SEAL® Hemostatic powder, the Primary Irritation Index was calculated to be 0,00. Results are presented in Table 19.

Table 19. Intracutaneous reactivity results.

Group	24 hours after injection		48 hours after injection		72 hours after injection		Total Primary Irritation Score
	Average Erythema	Average Oedema	Average Erythema	Average Oedema	Average Erythema	Average Oedema	
4SEAL® - Sodium Chloride	0	0	0	0	0	0	0
Solvent control - Sodium Chloride	0	0	0	0	0	0	0
4SEAL® - Cottonseed oil	0	0	0	0	0	0	0
Solvent control - Cottonseed oil	0	0	0	0	0	0	0

### Acute Systemic Toxicity

None of the test or control animals exhibited overt signs of toxicity, enumerated in Annex C, Table C.1 – Common clinical signs and observations, ISO 10993 – 11, at any observation points [17]. Furthermore, none of the animals treated with the test sample shows a significantly greater biological reactivity during the observation period than animals treated with vehicle control. None of the animals die, none of the animals' lost 10% or more bodyweight. Changes of bodyweight are presented in Table 20.

Table 20. Bodyweight changes.

Group	Average body weight change 72h after injection [%]
Solvent control -Sodium Chloride	0,95
4SEAL® - Sodium Chloride	-0,71
4SEAL® - Cottonseed Oil	6,07
Solvent control - Cottonseed Oil	3,51

### Subacute toxicity combined with implantation

None of the test or control animals exhibited overt signs of toxicity, enumerated in Annex C, Table C.1 – Common clinical signs and observations, ISO 10993 – 11, at any observation points [17]. Furthermore, none of the animals treated with the test sample shows a significantly greater biological reactivity during the observation period than animals treated with vehicle control. None

of the animals die, none of the animals' lost 10% or more bodyweight. For statistical analysis, a two-tailed heteroscedastic T-test was used. Bodyweight changes are presented in Table 21.

Table 21. Change in bodyweight.

Group	Average body weight change after 28 days from exposition [%]	Average weight change after 28 days from exposition [g]
Negative control - female	14,50	30,92
4SEAL® - female	17,12	36,14
Negative control - male	24,19	75,06
4SEAL® - male	24,00	74,94

*Gross necropsy findings:*

No abnormalities have been found during gross necropsy. Places of subcutaneous implantation and the surrounding tissues did not show any abnormalities. During gross necropsy, organs (brain, lungs, heart, liver, kidneys, ovaries/testis, spleen) were weighted. Organ weight was divided by the animal's body weight and is given as % of body weight. Test results are presented in Tables 22-33.

Table 22. Organ weight as a [%] of bodyweight.

Group	Average body weight [g]	Average organ as a [%] of body weight							
		Brain	Heart	Lungs	Liver	Kidneys	Adrenal	Ovaries / testis	Spleen
Negative control - female	245,06	0,73	0,38	0,69	4,20	0,86	0,06	0,08	0,25
4SEAL® - female	247,02	0,78	0,33	0,81	4,20	0,82	0,06	0,11	0,27
Negative control - male	385,94	0,51	0,34	0,69	4,66	0,76	0,03	1,07	0,24
4SEAL® - male	388,58	0,53	0,34	0,73	4,82	0,75	0,04	1,00	0,22

Table 23. Statistical comparison of the control group with the test group – P-value results.

Sex	Brain	Heart	Lungs	Liver	Kidneys	Adrenal	Ovaries/testis	Spleen
female	0,02	0,30	0,37	0,98	0,27	1,00	0,08	0,32
male	0,40	0,89	0,59	0,06	0,75	0,59	0,12	0,11

Statistically significant differences were observed in brain weight between the control and test group of female rats. The microscopic and macroscopic observations did not show any abnormalities. The rest of the organs did not show statistically significant differences between the test and the control group.

Table 24. Biochemical finding results.

Group	albumin	ALP	ALT	AST	Ca	Cl	Cholesterol	Creatinine
	[g/dL]	[U/L]	[U/L]	[U/L]	[mg/dL]	[mmol/L]	[mg/dL]	[mg/dL]
Negative control - female	4,12	257,00	38,80	154,80	9,46	102,20	69,00	0,21
4SEAL® - female	3,78	236,00	43,40	97,60	9,92	102,00	73,60	0,17
Negative control – male	3,36	245,80	44,20	81,40	10,28	98,40	73,00	0,20
4SEAL® - male	3,34	247,00	43,80	86,80	10,44	97,00	62,00	0,13

Table 25. Statistical comparison of the control group with the test group – P-value results.

Sex	albumin	ALP	ALT	AST	Ca	Cl	Cholesterol	Creatinine
female	0,17	0,04	0,31	0,16	0,30	0,83	0,55	0,62
male	0,87	0,94	0,88	0,69	0,70	0,43	0,40	0,33

Statistically significant differences were observed in ALP levels between the control and test group of female rats. However, the difference did not impact the clinical picture of the animals. Furthermore, there were no significant differences between the remaining parameters.

Table 26. Biochemical findings results.

Group	GGT*	Glucose	K	P	Na	bilirubin	Total protein	Triglycerides	Blood urea nitrogen
	[U/L]	[mg/dL]	[mmol/L]	[mg/dL]	[mmol/L]	[mg/dL]	[g/dL]	[mg/dL]	[mg/dL]
Negative control - female	*	242,20	4,06	6,08	141,80	0,18	5,64	182,20	22,74
4SEAL® - female	*	247,20	4,44	6,30	141,00	0,24	5,48	101,40	18,88
Negative control – male	*	324,60	5,32	7,70	138,40	0,22	5,32	99,60	18,50
4SEAL® - male	*	333,60	5,14	7,72	136,00	0,18	5,28	146,00	26,50

\*level undetected or very low

Table 27. Statistical comparison of the control group with the test group – P-value results.

Sex	GGT	Glucose	K	P	Na	Bilirubin	Total protein	Triglycerides	Blood urea nitrogen
female	N/A	0,90	0,40	0,87	0,41	0,31	0,62	0,03	0,20
male	N/A	0,82	0,79	0,98	0,12	0,20	0,70	0,21	0,36

Statistically significant differences were observed in triglycerides and alkaline phosphatase levels between the control and test group of female rats. However, the difference did not impact the clinical picture of the animals. There were no significant differences between the remaining parameters.

Table 28. Hematology findings.

Group	PT	APTT	HGB	HCT	platelets	RBC's	WBC	Lymphocytes	Monocytes	Granulocytes
	[sec.]	[sec.]	[g/dL]	[%]	[x 10 <sup>3</sup> /mm <sup>3</sup> ]	[x 10 <sup>6</sup> /mm <sup>3</sup> ]	[x 10 <sup>3</sup> /mm <sup>3</sup> ]	[%]	[%]	[%]
Negative control – female	9,05	17,83	13,42	30,28	783,20	7,39	6,72	59,80	17,54	22,66
4SEAL® - female	7,80	20,08	12,80	29,30	752,40	7,23	5,40	53,50	17,36	29,02
Negative control – male	10,16	17,88	13,68	31,12	778,60	7,36	7,58	54,02	16,58	29,42
4SEAL® - male	11,28	22,08	13,16	30,16	748,80	7,39	6,14	55,24	17,88	26,88

Table 29. Statistical comparison of the control group with the test group – P-value.

Sex	PT	APTT	HGB	HCT	platelets	RBC's	WBC	Lymphocytes	Monocytes	Granulocytes
female	0,01	0,21	0,07	0,51	0,37	0,47	0,21	0,10	0,83	0,08
male	0,21	0,06	0,17	0,40	0,50	0,88	0,24	0,67	0,15	0,48

Statistically significant differences were observed in prothrombin time between the control and test group of female rats. However, the difference did not impact the clinical picture of the animals. In addition, there were no significant differences between the remaining parameters.

Table 30. Urine test results.

Group	BLD	UBG	BIL	PRO	NIT	KET	GLU	pH	SG	LEU
	[Ery/ $\mu$ l]	[ml/dl]	[ $\mu$ mol/l]	[g/l]	[mg/dl]	[mg/dl]	[mg/dl]			[leu/ $\mu$ l]
Negative control - female	40,00	4,00	0,60	44,00	0,00	0,00	30,00	8,00	1,01	125,00
4SEAL® - female	12,00	4,80	1,00	38,00	0,00	0,00	0,00	7,60	1,02	105,00
Negative control – male	14,00	3,60	1,20	86,00	0,00	0,00	0,00	7,50	1,02	401,00
4SEAL® - male	12,00	6,40	1,40	100,00	0,00	15,00	0,00	8,00	1,01	405,00

Table 31. Statistical comparison of the control group with the test group – P-value.

Sex	BLD	UBG	BIL	PRO	NIT	KET	GLU	pH	SG	LEU
female	0,08	0,66	0,52	0,79	N/A	N/A	0,37	0,37	0,51	0,89
male	0,89	0,10	0,81	0,37	N/A	0,07	N/A	0,30	0,46	0,98

No statistically significant differences were observed between the test and the control group.

According to ISO 10993-11, instead of full histopathology, limited analysis was conducted [17]. From each group, control, and study group, two representative animals were chosen. Following organs were examined: lungs, heart, liver, kidneys, ovaries/testis, spleen, bone, bone marrow. No abnormalities were found during histopathology evaluation. The microscopic structure of the organs was normal, with no signs of apoptosis of structural cells of individual organs. No significant differences between control and test groups of animals were detected.

Table 32. Histological evaluation of place implantation of study and control groups.

	Negative control		Study group	
	Female	Male	Female	Male
<b>Cell type response</b>				
Polymorphonuclear cells	0	1	0	1
Lymphocytes	1	3	1	1
Plasma cells	0	1	0	0
Macrophages	3	4	3	3
Giant cells	1	2	0	1
Necrosis	0	1	0	0
<b>Subtotal (x2)</b>	10	24	8	12
<b>Tissue response</b>				
Neovascularization	3	2	3	3
Fibrosis	1	3	1	2
Subcutaneous changes	4	4	4	4
Fatty infiltrate	2	2	0	2
Muscular layer infiltration	3	2	0	3
<b>Subtotal</b>	23	49	16	32
<b>Total</b>	72		48	
<b>Average</b>	36		24	

Table 33. Rating of reaction for implantation.

Grade	Classification
0.0 – 2.9	Minimal or no reaction
3.0 – 8.9	Slight reaction
9.0 – 15.0	Moderate reaction
15.1	Severe reaction

As per ISO 10993-6, doubled cell-type response scores and tissue response scores were summarized and divided by the number of groups (male and female) to calculate the average score for test and control groups [16]. The final reaction rating was calculated by subtracting the average negative control score from the average tested sample score. The rating of reaction for 4SEAL® Hemostatic powder was -12 (0), which is classified as minimal or no reaction.

### Pyrogenicity

At the end of the test, no rabbit showed an individual temperature rise higher or equal to 0.6°C above its initial temperature. Pyrogenicity test results are presented in Table 34.

Table 34. Results of the pyrogenicity test.

Rabbit No.	Rabbit weight [g]	Volume injected [mL]	Initial temperature [°C]	Maximal temperature [°C]	Temperature rise [°C]	Total temperature rise [°C]
1	3250	32,5	38,8	38,9	0,1	0,1
2	3970	39,7	38,75	38,7	-0,05	
3	3380	33,8	38,45	38,3	-0,15	

Summarized results of 4SEAL® Hemostatic powder biocompatibility testing are presented in Table 35 below.

Table 35. 4SEAL® Hemostatic powder biocompatibility testing summary.

Test Performed	Extract(s)	Test and Control(s) Positive control (+) Negative control (-)	4SEAL® results
Chemical characterization ICP-MS ISO 10993-18[7]	Water for injection	Water for injection (-)	Elements <LOD
Chemical characterization Headspace GC-MS ISO 10993-18 [7]	Raw product	Laboratory air (-) EPA VOC Mix 2 (+)	<AET
Chemical characterization GC-MS ISO 10993-18 [7]	Water for injection	Water for injection (-) Octane, decane, tridecane, tetradecane, hexadecane (+)	<AET
Cytotoxicity ISO 10993-5 [9]	MEM	Latex (+) HDPE (-)	No cytotoxicity
Sensitization ISO 10993-10 [15]	Acetone: olive oil	Acetone:olive oil (-) A-Hexylcinnamaldehyde (+)	No sensitization
Intracutaneous reactivity ISO 10993-10 [15]	Saline and CSO	Saline and CSO (-)	No irritation
Material mediated pyrogenicity ISO 10993-11 [17]	Saline	Saline (-)	No pyrogenicity
Endotoxin ISO 10993-11 [17]	Water for injection	Water for injection (-) Endotoxin standard (+)	<20 EU/ 5 g device <2,15 EU/ 3 g device
Acute systemic toxicity ISO 10993-11 [17]	Saline and CSO	Saline and CSO (-)	No signs of toxicity
Subacute toxicity combined with implantation ISO 10993-6 [16] ISO 10993-11 [17]	Direct implantation	Starsil Hemostat (-)	No signs of subacute toxicity No difference in tissue reaction
Genotoxicity (Ames) ISO 10993-3 [10]	Water for injection	Water for injection (-) without S9: 2-NF, 4-NQO, N4-ACT, 9-AA (+) with S9: 2-AA, 2-AF(+)	No mutagenic potential
Genotoxicity (MLA) ISO 10993-3 [10]	F5	F5(-) without S9: methylmethansulfonate (+) with S9: benzo[a]pyrene (+)	No mutagenic potential

## Discussion

The biocompatibility evaluation of medical devices is a complicated, multi-stage approach that aims to predict whether a medical device could present any potential danger in clinical use by evaluating the device's compatibility with different biological systems. The process is regulated by internationally recognized standards such as the International Organization for Standardization (ISO) standard 10993 and a number of additional guidance from the FDA, Japanese Ministry of Health, Labor and Welfare, and other regulatory bodies [18]. The term medical device has a broad meaning and includes devices like a simple wooden spatula and highly complicated and sophisticated spine implant. Therefore, a different set of tests is needed for different categories of medical devices. 4SEAL® Hemostatic powder is degraded by alpha-amylases, glucoamylases, and macrophages in a few days. Therefore, 4Seal is classified as an implant that contacts tissue for a prolonged time (24h to 30d) as per Table A.1 in ISO 10993-1 [19, 20]. The tests performed in this study were chosen based on this classification.

While determining extraction conditions that may be used for exhaustive extraction of 4SEAL® Hemostatic powder, a degradation and vehicle color change was observed for semi-polar and non-polar vehicles. Therefore, only water was used for chemical testing. Analytical Evaluation Thresholds (AET) for 4SEAL® Hemostatic powder was determined based on the maximal dose per patient. No organic compounds above AET were observed therefore, no further toxicological assessment was necessary. ICP-MS analysis revealed that the Total Element Exposure is significantly less than element-specific Parenteral PDE limits present in ICH Q3D(R1) guidelines eliminating the need for further toxicological evaluation [8]. Analysis of extractables and leacheables performed using GC-MS, Headspace GC-MS, and ICP MS methods showed, with high probability, that there is no toxicological risk associated with the composition of 4SEAL® Hemostatic powder. Next, the endotoxin level was evaluated. The endotoxin content of the maximal size of the product (5g) was 2,568 EU/device, which is well below the endotoxin limit for general medical devices. Results also indicate that the 3g 4SEAL® Hemostatic powder may be used in procedures involving the contact with cerebrospinal fluid for which the limit is 2,15 EU/device [21]. To further evaluate the safety of 4SEAL® Hemostatic powder cytotoxicity was evaluated using MTT assay and L-929 cells. The extract and its dilutions showed no cytotoxic potential. *In vitro* assays were also used to assess the potential for genotoxicity of 4Seal. Two independent tests were conducted, AMES test employing procaryotic biological system and MLA using eukaryotic L5178Y TK+/-3.7.2C cell line. In AMES assay, 4SEAL® Hemostatic powder showed unclear mutagenic effect only when exposed to the TA1535 strain with the presence of the S9 fraction. For TA1535 strain without S9 fraction and for other strains, 4SEAL® Hemostatic powder was non-mutagenic. 4SEAL® Hemostatic powder did not show any mutagenic effects in any condition tested in MLA assay and therefore should be considered non-mutagenic.

A battery of *in vivo* tests was also conducted on the 4SEAL® Hemostatic powder to evaluate its biocompatibility. Local lymph node assay showed that the device should be considered non-sensitizing. Intracutaneous reactivity assay also showed that the product should be considered non-irritant. What is more, the testing showed that the 4SEAL® Hemostatic powder does not induce acute systemic toxicity. Furthermore, subacute toxicity studies combined with implantation revealed no subacute toxicity potential of the 4SEAL® Hemostatic powder even though the dose was more than 10x of the human dose. Finally, the lack of pyrogenic potential was confirmed with the *in vivo* Rabbit Pyrogen Test.

## Conclusion

In conclusion, 4SEAL® Hemostatic powder showed excellent biocompatibility and should be considered safe for use. Therefore, 4SEAL® Hemostatic powder is a promising new hemostatic agent with a wide range of potential applications.

## Abbreviations

AET Analytical Evaluation Threshold

AMES Bacterial Reverse Mutation Test

LLNA Local Lymph Node Assay

LNC Lymph Node Cells

MLA Mouse Lymphoma Assay

PC Positive Control

RPE Relative Plating Efficiency

RSG Relative Suspension Growth

RTG Relative Total Growth

SI Sensitization Index

SVOC Semi-Volatile Organic Compound

TEE Total Element Exposure

VOC Volatile Organic Compound

## Declarations

### Ethics approval and consent to participate

The study was conducted according to the guidelines of the Declaration of Helsinki and approved by the I Local Ethics Committee in Warsaw protocol codes 886/2019, 867/2019, and 878/2019.

### Consent for publication

Not applicable.

### Availability of data and materials

The following are available online at [www.xxxxx.com/xxx/s1](http://www.xxxxx.com/xxx/s1), Additional information and data for MLA, intracutaneous toxicity, sensitization studies, acute systemic toxicity, subacute systemic toxicity with implantation and pyrogenicity are presented in Tables S1-S13.

### Competing interests

Dr. Damian Matak and Dr. Lukasz Szymanski have invented 4SEAL® Hemostatic powder. All tests were performed in ISO 17025 accredited and GLP certified laboratory to ensure data integrity.

### Funding

Not applicable.

### Authors' contributions

Conceptualization, LS and DM; Data curation, AW, KG, MD, PK and JW; Formal analysis, AW; Funding acquisition, DM; Investigation, LS, AW, KG, MD, PSz, PK and JW; Methodology, LS, AW, KG, MD, PK, JW and DM; Supervision, LS and DM; Validation, PSz; Writing – original draft, LS; Writing – review & editing, LS and DM.

### Acknowledgements

Not applicable.

## References

1. Karkouti, K.; Dattilo, K.M. Perioperative Hemostasis and Thrombosis. *Can J Anaesth* 2006, *53*, 1260–1262, doi:10.1007/BF03021588.
2. Boucher, B.A.; Traub, O. Achieving Hemostasis in the Surgical Field. *Pharmacotherapy* 2009, *29*, 2S-7S, doi:10.1592/phco.29.pt2.2S.
3. Hess, J.R.; Brohi, K.; Dutton, R.P.; Hauser, C.J.; Holcomb, J.B.; Kluger, Y.; Mackway-Jones, K.; Parr, M.J.; Rizoli, S.B.; Yukioka, T.; et al. The Coagulopathy of Trauma: A Review of Mechanisms. *J Trauma* 2008, *65*, 748–754, doi:10.1097/TA.0b013e3181877a9c.
4. Rubin R, Strayer DS, Rubin E, McDonald (MD), J.M. *Rubin's Pathology: Clinicopathologic Foundations of Medicine*, Lippincott Williams & Wilkins, 2008; ISBN 978-0-7817-9516-6.
5. Franchini, M.; Lippi, G. Factor V Leiden and Hemophilia. *Thromb Res* 2010, *125*, 119–123, doi:10.1016/j.thromres.2009.11.003.
6. ISO 10993-12:2021(En), Biological Evaluation of Medical Devices – Part 12: Sample Preparation and Reference Materials Available online: <https://www.iso.org/obp/ui#iso:std:iso:10993:-12:ed-5:v1:en> (accessed on 16 August 2021).
7. ISO 10993-18:2020(En), Biological Evaluation of Medical Devices – Part 18: Chemical Characterization of Medical Device Materials within a Risk Management Process Available online: <https://www.iso.org/obp/ui#iso:std:iso:10993:-18:ed-2:v1:en> (accessed on 16 August 2021).
8. Q3D(R1) Elemental Impurities; International Council for Harmonisation; Guidance for Industry; Availability Available online: <https://www.federalregister.gov/documents/2020/03/11/2020-04995/q3dr1-elemental-impurities-international-council-for-harmonisation-guidance-for-industry> (accessed on 16 August 2021).
9. ISO 10993-5:2009(En), Biological Evaluation of Medical Devices – Part 5: Tests for in Vitro Cytotoxicity Available online: <https://www.iso.org/obp/ui#iso:std:iso:10993:-5:ed-3:v1:en> (accessed on 16 August 2021).
10. ISO 10993-3:2014(En), Biological Evaluation of Medical Devices – Part 3: Tests for Genotoxicity, Carcinogenicity and Reproductive Toxicity Available online: <https://www.iso.org/obp/ui#iso:std:iso:10993:-3:ed-3:v1:en> (accessed on 16 August 2021).
11. ISO/TR 10993-33:2015(En), Biological Evaluation of Medical Devices – Part 33: Guidance on Tests to Evaluate Genotoxicity – Supplement to ISO 10993-3 Available online: <https://www.iso.org/obp/ui#iso:std:iso:tr:10993:-33:ed-1:v1:en> (accessed on 18 August 2021).
12. Test No. 490: In Vitro Mammalian Cell Gene Mutation Tests Using the Thymidine Kinase Gene Available online: [https://www.oecd-ilibrary.org/environment/test-no-471-bacterial-reverse-mutation-test\\_9789264071247-en](https://www.oecd-ilibrary.org/environment/test-no-471-bacterial-reverse-mutation-test_9789264071247-en) (accessed on 18 August 2021).
13. Test No. 471: Bacterial Reverse Mutation Test Available online: [https://www.oecd-ilibrary.org/environment/test-no-471-bacterial-reverse-mutation-test\\_9789264071247-en](https://www.oecd-ilibrary.org/environment/test-no-471-bacterial-reverse-mutation-test_9789264071247-en) (accessed on 18 August 2021)..
14. Bacterial Endotoxins | USP Available online. <https://www.usp.org/harmonization-standards/pdg/general-methods/bacterial-endotoxins> (accessed on 18 August 2021)..
15. ISO 10993-10:2010(En), Biological Evaluation of Medical Devices – Part 10: Tests for Irritation and Skin Sensitization Available online: <https://www.iso.org/obp/ui#iso:std:iso:10993:-10:ed-3:v1:en> (accessed on 18 August 2021).
16. ISO 10993-6:2016(En), Biological Evaluation of Medical Devices – Part 6: Tests for Local Effects after Implantation Available online: <https://www.iso.org/obp/ui#iso:std:iso:10993:-6:ed-3:v1:en> (accessed on 18 August 2021).
17. ISO 10993-11:2017(En), Biological Evaluation of Medical Devices – Part 11: Tests for Systemic Toxicity Available online: <https://www.iso.org/obp/ui#iso:std:iso:10993:-11:ed-3:v1:en> (accessed on 18 August 2021).
18. Bernard, M.; Jubeli, E.; Pungente, M.D.; Yagoubi, N. Biocompatibility of Polymer-Based Biomaterials and Medical Devices – Regulations, in Vitro Screening and Risk-Management. *Biomater. Sci.* 2018, *6*, 2025–2053, doi:10.1039/C8BM00518D.
19. K132105. - 510(k) Summary PerClot 2014.
20. ISO 10993-1:2018(En), Biological Evaluation of Medical Devices – Part 1: Evaluation and Testing within a Risk Management Process Available online: <https://www.iso.org/obp/ui#iso:std:iso:10993:-1:ed-5:v2:en> (accessed on 16 August 2021).
21. Research C, for DE. and Guidance for Industry: Pyrogen and Endotoxins Testing: Questions and Answers Available online: <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/guidance-industry-pyrogen-and-endotoxins>

## Figures



Figure 1

Photo of 4SEAL® Hemostatic powder.




Figure 2

Endotoxins concentration standard curve.

## Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [Suplementdata.docx](#)