Nanoscopic resolution within a single imaging frame

Esley García

Centro de Investigación en Ciencias, Instituto de Investigación en Ciencias Básicas y Aplicadas, Universidad Autónoma del Estado de Morelos <https://orcid.org/0000-0002-9301-6962>

Raúl Cámara

Centro de Investigación en Ciencias, Instituto de Investigación en Ciencias Básicas y Aplicadas, Universidad Autónoma del Estado de Morelos <https://orcid.org/0000-0001-7528-5758>

Alejandro Linares

Laboratorio Nacional de Microscopía Avanzada, Instituto de Biotecnología, Universidad Nacional Autónoma de México

Damián Martínez

Laboratorio Nacional de Microscopía Avanzada, Instituto de Biotecnología, Universidad Nacional Autónoma de México <https://orcid.org/0000-0002-5906-598X>

Víctor Abonza

Laboratorio Nacional de Microscopía Avanzada, Instituto de Biotecnología, Universidad Nacional Autónoma de México

Eduardo Brito-Alarcon

Laboratorio Nacional de Microscopía Avanzada, Instituto de Biotecnología, Universidad Nacional Autónoma de México <https://orcid.org/0000-0001-5209-5388>

Carlos Calcines-Cruz

Instituto de Investigaciones Biomédicas, Universidad Nacional Autónoma de México

Gustavo Galindo

Departamento de Química de Biomacromoléculas, Instituto de Química. Universidad Nacional Autónoma de México

David Torres

Laboratorio Nacional de Microscopía Avanzada, Instituto de Biotecnología, Universidad Nacional Autónoma de México

Martina Jabłoński

Instituto de Biología y Medicina Experimental

Héctor H. Torres-Martínez

Departamento de Biología Molecular de Plantas, Instituto de Biotecnología, Universidad Nacional Autónoma de México <https://orcid.org/0000-0001-5817-4255>

José Martínez

Departamento de Genética del Desarrollo y Fisiología Molecular, Instituto de Biotecnología, Universidad Nacional Autónoma de México

Haydee Hernández

Instituto de Investigaciones en Matemáticas Aplicadas y en Sistemas, Universidad Nacional Autónoma de México

Joseph Ocelotl

Laboratorio Nacional de Microscopía Avanzada, Instituto de Biotecnología, Universidad Nacional Autónoma de México

Yasel Garcés

Departamento de Genética del Desarrollo y Fisiología Molecular, Instituto de Biotecnología, Universidad Nacional Autónoma de México

Marco Barchi

University of Rome Tor Vergata <https://orcid.org/0000-0003-1104-6234>

Joseph G. Dubrovsky

Departamento de Biología Molecular de Plantas, Instituto de Biotecnología, Universidad Nacional Autónoma de México <https://orcid.org/0000-0002-2072-4650>

Alberto Darszon

Departamento de Genética del Desarrollo y Fisiología Molecular, Instituto de Biotecnología, Universidad Nacional Autónoma de México

Mariano Buffone

Instituto de Biología y Medicina Experimental

Roberto Rodríguez

Instituto de Cibernética, Matemática y Física

Juan Rendón-Mancha

Centro de Investigación en Ciencias, Instituto de Investigación en Ciencias Básicas y Aplicadas, Universidad Autónoma del Estado de Morelos

Christopher Wood

Universidad Nacional Autónoma de México <https://orcid.org/0000-0001-7175-7121>

Armando Hernández

Departamento de Química de Biomacromoléculas, Instituto de Química. Universidad Nacional Autónoma de México

Diego Krapf

Colorado State University <https://orcid.org/0000-0002-2833-5553>

Álvaro Crevenna

European Molecular Biology Laboratory <https://orcid.org/0000-0002-5580-3711>

Adán Guerrero (✉ adan.guerrero@ibt.unam.mx)

Laboratorio Nacional de Microscopía Avanzada, Instituto de Biotecnología, Universidad Nacional Autónoma de México

Article

Keywords: super-resolution microscopy, diffraction limit, single frame, Mean Shift, fluorescence microscopy, live-cell imaging

Posted Date: October 29th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-984659/v1>

License: This work is licensed under a Creative Commons Attribution 4.0 International License.
[Read Full License](#)

1 **Nanoscopic resolution within a single imaging frame**

2 Esley Torres García ^{1,2}, Raúl Pinto Cámara ^{1,2}, Alejandro Linares ², Damián Martínez ², Víctor Abonza ²,
3 Eduardo Brito-Alarcón ², Carlos Calcines-Cruz ³, Gustavo Valdés-Galindo ⁴, David Torres ², Martina
4 Jabłoński ⁵, Héctor H. Torres-Martínez ⁶, José L. Martínez ⁷, Haydee O. Hernández ⁸, José P. Ocelotl-
5 Oviedo ², Yasel Garcés ^{2,7}, Marco Barchi ⁹, Joseph G. Dubrovsky ⁶, Alberto Darszon ⁷, Mariano G.
6 Buffone ⁵, Roberto Rodríguez Morales ¹⁰, Juan Manuel Rendon-Mancha ¹ Christopher D. Wood ²,
7 Armando Hernández-García ⁴, Diego Krapf ¹¹, Álvaro H. Crevenna ¹², and Adán Guerrero ^{2,*}.

8

9 ¹ Centro de Investigación en Ciencias, Instituto de Investigación en Ciencias Básicas y Aplicadas,
10 Universidad Autónoma del Estado de Morelos, Cuernavaca, Morelos, Mexico.

11 ² Laboratorio Nacional de Microscopía Avanzada, Instituto de Biotecnología, Universidad Nacional
12 Autónoma de México, Cuernavaca, Morelos, Mexico.

13 ³ Instituto de Investigaciones Biomédicas, Universidad Nacional Autónoma de México. Ciudad de
14 México, Mexico.

15 ⁴ Departamento de Química de Biomacromoléculas, Instituto de Química. Universidad Nacional
16 Autónoma de México. Ciudad de México, Mexico.

17 ⁵ Instituto de Biología y Medicina Experimental (IBYME-CONICET), Buenos Aires, Argentina.

18 ⁶ Departamento de Biología Molecular de Plantas, Instituto de Biotecnología, Universidad Nacional
19 Autónoma de México, Cuernavaca, Mexico.

20 ⁷ Departamento de Genética del Desarrollo y Fisiología Molecular, Instituto de Biotecnología,
21 Universidad Nacional Autónoma de México, Cuernavaca, Morelos, Mexico.

22 ⁸ Instituto de Investigaciones en Matemáticas Aplicadas y en Sistemas, Universidad Nacional Autónoma
23 de México, Ciudad de México, Mexico.

24 ⁹ Department of Biomedicine and Prevention, Faculty of Medicine, University of Rome Tor Vergata,
25 Rome, Italy.

26 ¹⁰ Instituto de Cibernética, Matemática y Física, Ciudad de la Habana, Cuba.

27 ¹¹ Electrical and Computer Engineering and School of Biomedical Engineering, Colorado State
28 University, Fort Collins, Colorado 80523, U.S.A.

29 ¹² European Molecular Biology Laboratory, Neurobiology and Epigenetics Unit, Monterotondo, Italy.

30

31 * Correspondence contact: adan.guerrero@ibt.unam.mx

Abstract

Mean-Shift Super Resolution (MSSR) is a principle based on the Mean Shift theory that improves the spatial resolution in fluorescence images beyond the diffraction limit. MSSR works on low- and high-density fluorophore images, is not limited by the architecture of the detector (EM-CCD, sCMOS, or photomultiplier-based laser scanning systems) and is applicable to single images as well as temporal series. The theoretical limit of spatial resolution, based on optimized real-world imaging conditions and analysis of temporal image series, has been measured to be 40 nm. Furthermore, MSSR has denoising capabilities that outperform other analytical super resolution image approaches. Altogether, MSSR is a powerful, flexible, and generic tool for multidimensional and live cell imaging applications.

Key Words: super-resolution microscopy, diffraction limit, single frame, Mean Shift, fluorescence microscopy, live-cell imaging.

Introduction

Super-resolution Microscopy (SRM), which encompasses a collection of methods that circumvent Abbe's optical resolution limit, has dramatically increased our capability to visualize the architecture of cells and tissues at the molecular level. There are several approaches to SRM which vary in terms of the final attainable spatial and temporal resolution, photon efficiency, as well as in their capacity to image live or fixed samples at depth [1, 2]. One class of techniques exceed the diffraction limit by engineering the illumination or the point spread function (PSF), such as SIM and STED [3-5]. These techniques can be used for live imaging although they require specialized hardware and dedicated personnel for maintenance and operation. Single-molecule localization methods (e.g., STORM, PAINT, PALM) [6-9] that localize individual emitters with nanometer precision require temporal analysis of several hundred-to-thousands of images and are prone to error due to fast molecular dynamics within live specimens.

Some SRM computational methods have few or no demands on hardware or sample preparation and provide resolution improvements beyond the diffraction limit [10-13]. The quantity and performance of computational methods have both increased over the past decade given the many advantages they present, such as their low barriers to entry and generic applicability to data acquired with any microscopy modality (wide-

field, confocal, or light-sheet). However, these methods also present some limitations, such as the possible introduction of artifacts [14], the requirement for high signal-to-noise ratio (SNR) data and the acquisition of tens to hundreds of frames [10-13], which limit their applicability to reconstruct fast dynamical processes.

Here, we introduce the Mean Shift Super-Resolution principle for digital images ‘MSSR’ (pronounced as *messer*), derived from the Mean Shift (MS) theory [15, 16]. MSSR improves the resolution of any single fluorescence image up to 1.6 times, including its use as a resolution enhancement complement after the application of other super-resolution methods. Additionally, we demonstrate the super-resolving capabilities of MSSR as a standalone method for a variety of fluorescence microscopy applications, through a single-frame and temporal stack analysis, allowing resolution improvements towards a limit of 40 nm.

Open-source implementations of MSSR are provided for ImageJ (as a plugin), R, and MATLAB, some of which take advantage of the parallel computing capabilities of regular desktop computers (Supplementary Note 7). The method operates almost free of parameters; users only need to provide an estimate of the point spread function (PSF, in pixels) of the optical system, choose the MSSR order, and decide whether a temporal analysis will take place (Supplementary material MSSR Manual). The provided open-source implementations of MSSR represent a novel user-friendly alternative for the bioimaging community for unveiling life at its nanoscopic level.

Results

The MSSR principle.

MSSR is tailored around the assumption that fluorescence images are formed by signals collected from point sources (i.e., fluorophores) convolved with the PSF of the microscope (Supplementary Notes 1, 2 and 3). Processing a single image with MSSR starts with the calculation of the MS, which guarantees that large intensity values on the diffraction-limited (DL) image coincide with large positive values in the MSSR image (Supplementary Note 4). Further algebraic transformations then restore the raw intensity distribution and remove possible artifacts caused by the previous step (edge effects and noise dependent artifacts), giving rise to an image that contains objects with a narrower full width at half maximum (FWHM) (Figure 1a). This procedure is denoted by MSSR⁰, as the first stage to shrink emitter distribution.

The MS is locally computed by a kernel window that slides throughout the entire image, subtracts the sample mean (weighted local mean) as well as the central value of the kernel using a spatial-range neighborhood (Supplementary Notes 2 and 3, Figure S4 and S5, Table S1) [15, 16]. The MS is a vector that always points towards the direction of the intensity gradient and its length provides a local measure of the fluorescence density and brightness [17-19]. Since the MS lies within the gradient space, its values depend on the difference between the central pixel of the neighborhood and the surrounding pixels and thus is not necessarily linked to the fluorescence intensity values of the raw image. A mathematical proof, provided in Supplementary Note 4, demonstrates that the minimum MS value, computed from a Gaussian distribution, matches with the point of maximum intensity of the initial distribution (Supplementary Note 4, Figure S6).

The increase in resolution offered by MSSR^0 was evaluated by the Rayleigh and Sparrow limits [20-22], which are two criteria that establish resolution bounds for two near-point sources (Figure 1b). Processing with MSSR^0 of two-point sources located at their resolution limit (2.5σ and 2σ for Rayleigh and Sparrow limit respectively, Figure 1c vertical discontinuous lines) decreases the dip (height at the saddle point) [23] within their intensity distributions (Figure 1b and 1c). Processing a single image with MSSR^0 shifts the resolution limit by 26 % and 20 %, according to the Rayleigh and Sparrow limits respectively (Figure 1c vertical continuous lines). Therefore, processing a single fluorescence image with MSSR^0 will reduce the FWHM of individual emitters. Also, a comparison of the shrinkability of MSSR^0 applied to Gaussian and Bessel distributions are shown in Figure S7. Additionally, the reduction of FWHM of Bessel distribution at different wavelengths of the visible spectrum are shown in Figure S8.

Since the result of MSSR is an image, we used the resulting image to seed an iterative process (Figure 2a). We refer to this as higher-order MSSR (MSSR^n , $n>0$), which delivers a further gain of resolution per n -iteration step (Figures. 2a and S9). As the order of MSSR^n increases, both the FWHM of emitters (Figure S10) and the dip of their intensity distribution decrease (Figure 2b). Numerical approximations indicate that two point-sources separated at 1.6σ are resolvable with MSSR^3 , but not when their separation is 1.5σ (Figure 2b). The separation of 1.6σ sets the theoretical resolution limit of MSSR^n .

To empirically test the ability of MSSRⁿ to achieve super-resolution image within a single frame, a commercial nanoruler sample (GATTA-SIM140B, GATTAquant) was imaged by Structured Illumination Microscopy (SIM) and widefield fluorescence microscopy, which was then processed by MSSRⁿ. The iterative processing of the widefield data with MSSR³ reveals the two fluorescence emitters located at a separation of 140 nm which is consistent with the result obtained by SIM (Figure 2c).

MSSR further increases the resolution of super-resolved images.

Based on the MSSR capabilities to generate a super-resolved micrograph after processing a single fluorescence image, we explored if a pre-existing super-resolved image can be further enhanced by MSSR.

First, we used temporal stack of DL images of tubulin-labeled microtubules collected at high fluorophore density [24] (previously used to test and compare a variety of SRM algorithms) [25], which were subject to ESI, SRRF or MUSICAL analysis [11-13], where each was used to compute a single super-resolved image (Figure 3a). Supplementary note 9 contains an in-depth comparison of MSSR reconstructions combined with other SRM analytical methods (ESI, SRRF and MUSICAL), which achieve super-resolution through a temporal analysis [12-14]. Post-processing of ESI, SRRF or MUSICAL images with MSSR⁰ increases contrast and resolution (Figure 3a).

Second, a sequence of images of randomly blinking emitters placed along a synthetic tubular structure [26] was processed with MSSR⁰ after analysis with MUSICAL. In both reconstructions, three regions (small squares in Figure 3b) were chosen to assess the gain in resolution, visualized in terms of the distance between the normalized intensity distributions peaks. MSSR further resolves the edges of the synthetic structures on the MUSICAL-processed image without changing the position of the distribution peaks (Figure 3b) as predicted by our theory.

Lastly, we applied MSSR on a super-resolved SIM image of sister chromatids of mouse chromosomes (Figure 3c). Similar to the results obtained above, processing with MSSR increases both the contrast and resolution of the final image. Each of the individual SRM methods tested performs optimally under specific experimental

conditions; one can thus choose whichever of them to use based on the available infrastructure, optical setup, and biological or experimental conditions that best fit the specific research goals. Altogether these data show that post-processing with MSSR increases resolution by a factor up to 1.6 times of any tested super-resolution technique.

Temporal analysis of MSSR

In theory, MSSR can be applied to a sequence of images (Supplementary Note 5). Based on the increase in resolution offered by computational methods that rely on temporal analyses (SRRF, ESI, MUSICAL), we investigated whether a further resolution gain could be achieved by applying a temporal analysis to a sequence of single frame MSSR images (t-MSSRⁿ) (Figure 4a). Pixel-wise temporal functions (*PTF*), such as average (*Mean*), variance (*Var*), the temporal product mean (*TPM*), coefficient of variation (*CV*) or auto-cumulant function of orders 2 to 4 (*SOFI*₂, *SOFI*₃, *SOFI*₄) [10], can be used to create an image with enhanced spatial resolution (Supplementary Note 5, Table S2).

To experimentally validate the increase in resolution from single-frame (sf-MSSRⁿ) to t-MSSRⁿ, we used two different nanoruler systems, an in-lab CRISPR/dCas12a nanoruler, used to score nanoscopic distances between individual fluorescent sites down to 100 nm, and a commercial nanoruler with fluorophores positioned at 40 nm of separation (GATTA-PAINT, 40G, and 40RY. Gattaquant).

The CRISPR/dCas12a nanoruler system consists of a dsDNA with four binding sites for dCas12a uniformly distributed every 297 bp (equivalent to ~ 100 nm of separation) (Figure S33a). To validate this system, we imaged the association of the CRISPR/dCas12a complex to the binding sites on the dsDNA by atomic force microscopy (AFM) and measured the distance between each dCas12a complex (Figure S33b).

The CRISPR-dCas12a nanorulers were then imaged in buffer by total internal reflection fluorescence microscopy (TIRFM) for further MSSR analysis. We used a DNA-PAINT approach for fluorescence indirect tagging [27], in which a fluorescent ssDNA probe hybridizes with an extension of the gRNA. The “blinking” of the fluorescence signal is attained by events of association and dissociation between the fluorescent probe and the gRNA on the CRISPR/dCas12a nanoruler at the binding site.

In the DL image, amorphous spot-like fluorescent patterns were observed (Figure 4b). sf-MSSR³ processing of either an isolated frame or an average projection of the corresponding stack of 100 images (DL-AVG) could not resolve individual CRISPR/dCas12a binding sites (Figure 4b), and only after processing by t-MSSR³ did individual binding sites become resolved (Figure 4c). The result of t-MSSR³ varied in relation to the temporal function used (Figure 4c). The best result for this nanoruler was obtained by the pixel-wise temporal variance (*Var*) of the sf-MSSR³ stack (Figure 4c). t-MSSR³-*Var* resolved nearby emitters engineered to recognize binding sites located at 100 nm (Supplementary Movie S1), provided by scoring association-dissociation events between the imaging probe and the gRNA.

To determine the distance between two dCas12a sites along the DNA chain we obtained the distribution of distances between dCas12a binding sites taking in consideration their unidimensional association to a semi-flexible polymer such as the DNA [28]. Estimated distances after t-MSSR³-*Var* in the CRISPR/dCas12a nanoruler are 85 ± 14 nm, 152 ± 21 nm, 232 ± 37 nm (Figure 4d). These results confirm that t-MSSR³ can successfully resolve nanoscopic distances.

To explore the resolution limit attainable by t-MSSRⁿ even further, we looked at a nanoruler system with smaller separation between fluorophore sites (from Gattaquant) (Figure S34a). Analysis with t-MSSR³ of 100 images revealed individual fluorescent spots at 40 nm apart (Figure 4e and Supplementary Figure S34b). The data presented in Figure 4e demonstrate that t-MSSR³ resolves nanoscopic distances in the 30–80 nm range, validating a lower experimental spatial resolution bound of 0.5σ (≈ 40 nm), which depends on the emission wavelength of the fluorophore (Figure 4e, Supplementary Figure S8c). In comparison, SRRF, ESI and MUSICAL were not able to resolve fluorescent emitters located 40 nm apart, consistent with their limit within the range of 50–70 nm (Figure 4f) [11–13].

Single frame nanoscopy, free of noise-dependent artifacts

The theory of image processing by MSSR (Supplementary Note 5), suggests that it should be robust over a wide range of SNR, granted by four factors. First, when processing a single frame, MS works as a local spatial frequency filter (a smoothing filter); regions corresponding to the image background are homogenized by the kernel window, reducing variation in background noise. Second, one of the steps of the MSSR

procedure is to remove the MS negative constraints. This threshold operation exerts influence on structures at σ , at about 65% of the intensity distribution of the emitters; values below this threshold will be considered as noise and set to zero value. Third, when using a PTF, nanoscopic information is enriched due to temporal oversampling of the hidden fluorescent structure. Fourth, the spatial kernel of the MSSR algorithm operates within the subpixel realm; the number of neighboring pixels is digitally increased through bicubic interpolation providing digital oversampling of the emitters' locations (Supplementary Note 6).

We then experimentally assessed the capacity of MSSR to denoise fluorescence images and determine whether it introduces noise-related artifacts. We used a PSFcheck slide [29], which contains an array of regular fluorescent nanoscopic patterns shaped by laser lithography (Figure 5). Analysis with sf-MSSRⁿ or t-MSSRⁿ showed, in comparison to alternative approaches, striking denoising capabilities without introducing noticeable artifacts (Figure 5a) (Supplementary Note 9). These artifacts, resembling amorphous nanoscopic structures around the fluorescent ring or within it, were commonly found at reconstructions generated by other analytical techniques (Figure S22).

Starting at a SNR > 2, sf-MSSR¹ provides reliable SRM reconstructions of comparable quality to other SRM approaches, which demand the temporal analysis of the fluorescence dynamics (Figure 5a and Supplementary Note 9). We quantified the quality of the reconstructions by calculating the Resolution Scaled Pearson (RSP) coefficient and the Resolution Scaled Error (RSE), which provide a global measurement of the quality of the reconstruction by comparing the super-resolution image and the reference image (in this case, the DL image) [14]. Higher RSP and lower RSE values are associated with reliable reconstructions (Supplementary Note 8). When the SNR is above 5, all tested algorithms perform similarly well in quality (Figure 5b), but their global errors differ from each other (Figure 5c). As expected, the RSE increased as a function of the SNR of the input images for any tested algorithm (Figure 5c).

Furthermore, the performance of MSSR in achieving a satisfactory reconstruction was assessed by varying the number of input images using a temporal analysis scheme (Supplementary Note 8). With SNR > 2 input data, RSP reaches near maxima values and RSE near minima values when processing a single frame (Figure S19-20, Supplementary Movie S2). However, when computing MSSR using low SNR

input data ($\text{SNR} \sim 2$) a temporal analysis is required as RSP and RSE values reach a plateau only when a temporal stack of as few as 20 images is used (Figure S20-21, Supplementary Movie S3). These findings illustrate that the minimal number of frames needed by MSSR to provide a reliable reconstruction depends on the information itself, i.e., on the SNR and on the photophysical properties of the specimen (movies S1 - S3); and can be determined by computing RSP and RSE as function of the number of analyzed frames with t-MSSRⁿ (Figure S21).

Nanoscopic resolution with conventional fluorescence imaging

To showcase the versatility of MSSR to super-resolve data acquired from different fluorescence applications, we evaluated its performance over a collection of experimental scenarios (Supplementary Note 10).

Analysis with MSSR provided nanoscopic resolution of rotavirus replication machineries (Figure S25), which were recently described by Garcés *et al* as a layered array of viral protein distributions [30]. Originally, it took the authors several days to weeks to generate a single super-resolution image by means of analyzing several stacks of hundreds of DL images using 3B-ODE SRM. With MSSR, we were able to achieve comparable results, through analyzing single DL frames within seconds with a regular desktop computer with either sf-MSSR¹ or t-MSSR¹ (Supplementary Note 7).

Mouse sperm cells are used to study the acrosomal exocytosis (AE), a unique secretory process which results from fusion events between the plasma membrane and a specialized vesicle called acrosome located in the sperm head [31,32]. Nanoscopic remodeling of both plasma membrane and actin cytoskeleton was imaged during the AE by means of sf-MSSR¹, showing single frame temporal resolution (of milliseconds) (Figs. S26). At the onset of the AE, the FM4-64 fluorescence (a probe that fluoresces when bound to membranes) was confined to the plasma membrane and was visible above of a F-actin cytoskeleton fringe. During the AE, several fenestration events were observed to occur at both the plasma and acrosome membranes, as consequence of that, a notorious increase of FM4-64 was observed close-below the F-actin fringe (Supplementary Movie S5 a-f). The AE is a dynamic remodeling process that takes minutes to occur, sf-MSSR¹ allows the observation of events occurring at the millisecond scales, which are hindered when using other SRM multi-frame analytical approaches, such as SRRF or 3B [11, 33], due to their mandatory need of a temporal

analysis of the fluorescence dynamics to unveil nanoscopic detail (compare Figures S26 and S27).

Background noise is known to be an important issue in single-particle tracking (SPT) applications as it decreases the ability to faithfully localize particles and follow them through time [34, 35]. Moreover, the spatial overlap of PSFs derived from individual particles makes it challenging for SPT algorithms to recognize them as separate entities. The denoising capabilities of sf-MSSR¹ enhanced both the contrast and spatial resolution of freely diffusing in-silico particles previously used as benchmarks to test a variety of SPT algorithms (Figure S28) [36]. Pre-processing of the images with sf-MSSR¹ improved the tracking performance of three commonly employed SPT tracking algorithms: (i) the LAP framework for Brownian motion as in [37, 38], (ii) a linear motion tracker based on Kalman filter [39-41], and (iii) a tracker based on Nearest neighbors [42-44] within a wide range of particle densities and SNR (Figure S29). Additional testing with sf-MSSRⁿ showed an increase in nanoscopic colocalization accuracy in double imaging experiments in single-molecule DNA curtain assays (Figure S30) [45].

Plasmalemma- and nuclear-labeled transgenic *Arabidopsis thaliana* plants are routinely used to study cell fate and proliferation during root development in time-lapse confocal microscopy experiments in two and three dimensions [46, 47]. When applied to lateral root primordium cells, located deep inside the parent root, sf-MSSR¹ demonstrated the capacity to achieve multidimensional nanoscopic resolution as it revealed isolated nanodomains resembling nucleosome clutches, previously reported in mammalian cells [48, 49], within the nuclei of a lateral root primordium cells (Figure S31 and Supplementary Movie S10). Similar observations were performed upon epidermal root tissues visualized via selective plane illumination microscopy (SPIM) after examination of volumetric data with sf-MSSR¹ (Figure S32). In combination, these studies provide evidence for the capabilities of MSSR to resolve biological detail at nanoscopic scales using either simple or advanced fluorescence microscopy technologies.

Discussion

- Novel theoretical contributions

From the historical point of view, since the seminal development of the MS theory [15, 16] and until the present day, few statistical and imaging applications based on the theory of MS compute the MS vector itself [50]. This can be explained, in part, because previous applications of MS are based on finding modes in the features space and did not operate directly in the derivative space. In contrast, MSSR represents an application of MS theory that operates in the second derivative space. By computing the MS vector and estimating densities among pixels, MSSR computes a probability function for the fluorophore estimates whose individual fluorescence distributions are narrowed in comparison with the PSF of the optical system. The exploration of the information stored on the second derivative space of the image can be also achieved by substituting the MS by similar functions that operate in such space, e.g., Laplacian, Hessian, Difference of Gaussians [51] which, in comparison with the MS, offer computational advantages as they can be expressed in the Fourier space and implemented using the FFT algorithm [51]. The information harbored in the second derivative space of the DL image is used by MSSR to compute a super-resolved image with higher spatial frequencies than the corresponding DL image, hence, overcoming both the Rayleigh and Sparrow limits, and setting up an undescribed limit of resolution which deserves further exploration and characterization.

The MS theory is not restricted by the number of dimensions of the information required to compute the kernel windows over which MSSR operates (Supplementary notes 2 and 3). Given that, MSSR parameters are suitable to extend its application to assess data with higher dimensions. For example, in 2D images, the spatial parameter of MSSR, which encompasses the lateral resolution width of the PSF, is defined to be the same in the x and y dimensions of the image. In such case, the shape of the kernel is circle- or square-like, depending on the application used. For three-dimensional (3D) microscopy imaging, the lateral (x-y plane) and axial (x-z and y-z planes) dimensions are affected in different ways by diffraction. The MSSR principle can be further extended for explicit volumetric imaging by means of using an asymmetric kernel which can be defined following the 3D lateral-axial aspect ratio of the PSF. In addition, the definition of the spatial kernel can be refined to also consider possible deformations of axial symmetry of the PSF due to optical aberrations introduced by the imaging

system or by the sample itself. A similar reasoning aimed to extend the portfolio of applications of MSSR can be envisaged considering spatial-range parameter, the latter narrowing down the working intensity space where local calculations of MSSR take place.

Novel contributions to microscopy

We present a new SRM approach capable of achieving multidimensional nanoscopy through single-frame analysis under low SNR conditions and with minimal noise-dependent artifacts. Limited only by the imaging speed of the optical system setup, MSSR increases resolution by analyzing either a single frame, or by applying MSSR to each individual image in a stack followed by the application of a pixel-wise temporal function. MSSR is a powerful stand-alone method for either single or multi-frame SRM approaches, or as post-processing method which can be applied to other analytical multi-frame (restricted to camera-based systems) or hardware dependent SRM methods for further enhancement of resolution and contrast. We demonstrated MSSR compatibility with other SRM methods and showed that its usage improved resolution and overall image quality in all the cases tested.

SRM analytical multi-frame approaches such as SRRF, ESI, MUSICAL and 3B demand a temporal analysis which limits their utility for multi-dimensional imaging of live samples [46]. The need to collect hundreds to thousands of images of the same pseudo-static scene, challenges the applicability of these methods in multidimensional imaging. The temporal multi-frame requirement imposes a tradeoff between the achievable temporal and spatial resolutions. MSSR removes these constraints while maintaining computational efficiency (Supplementary note 7).

We present applications of the MSSR principle that revealed fast molecular dynamics through single-frame analysis of live-cell imaging data, with reduced processing times in comparison with similar SRM approaches (Supplementary notes 7 and 10). Moreover, MSSR greatly improves the tracking efficacy of SPT methods by means of reducing background noise and increasing both the contrast and SNR of noisy SPT movies, enhancing the ability to resolve the position of single emitters. MSSR further pushes the limits of live-cell nanoscopy by its excellent single-frame performance. This flexibility extends its utility to most fluorescence microscopy and alternative SRM methods.

Achieving both high (or sufficient) temporal and spatial resolution within a broad range of fluorescence microscopy applications is a common goal among the bioimaging community. With recent advances in microscopy equipment and imaging protocols, the gap between the highest attainable resolution in the temporal and spatial dimensions within the same experiment, has narrowed. This has been a challenge especially because both parameters often involve mutually exclusive optical instrumentation and experimental strategies. The introduction of MSSR represents one more step in the right direction as it drastically reduces the amount of data needed to reconstruct a single super-resolved micrograph.

No longer having to sacrifice either temporal or spatial resolution over the other, has led some scientists to propose new ways to analyze imaging data. Some approximations have been tailored to study millisecond molecular dynamics and structural feature changes within the same experiment [52], e.g., by taking advantage of the simultaneous use of image correlation spectroscopy (ICS) and SRM methods such as SRRF [11]. In these contexts, MSSR could improve the analysis in three ways: a) it delivers reliable SRM images in low SNR scenarios, which are common in the experimental regimes of ICS due to the relatively fast frame rates of its applications, b) MSSR introduces no noise-dependent artifacts which further refines the quality of the spatial analysis and c) since no temporal binning is necessary for MSSR, there is no restriction in the level of temporal detail retrievable from the ICS analysis.

Sub-millisecond time-lapse microscopy imaging can now be achieved by sCMOS technologies, with applications for particle velocimetry [53], rheometry [54], and optical patch clamp [55]. We envisage further applications for MSSR in these areas through unveiling nanoscopic detail hidden in single DL images. Moreover, MSSR can facilitate correlative nanoscopic imaging through crosstalk with other imaging techniques such as electron microscopy, i.e., CLEM: correlative light electron microscopy [56]; or atomic force microscopy, i.e., CLAFEM: Correlative light atomic force electron microscopy [57]. In addition, MSSR can be applied to nanoscopic volumetric imaging by using it together with expansion microscopy [58], oblique angle microscopy [59], SPIM, and lattice light sheet microscopy [60], extending their capabilities to previous unattainable resolution regimes.

A recent study by Chen R. et al., suggests that deep-learning based artificial intelligence (AI) can reconstruct a super-resolution image from a single frame of a DL image [61]. Such AI-based SRM approaches are promising, however, they are limited to

423 the existence of a maximum likelihood image obtained with another SRM, such as
424 STORM, that is required for neural network training an error minimization. Otherwise,
425 the method it is prompted to bias the final reconstruction toward the topological
426 information used to train the AI - network [61]. Our approach works completely
427 independent of other SRM methods and provides evidence of the existence of a new
428 resolution limit which lies on the second derivative space of the DL image, information
429 inaccessible when using neural networks.

430 MSSR applications might impact far beyond the field of microscopy, as its
431 principles can be applied to any lens-based system such as astronomy [62] and high-
432 resolution satellite imagery [63].

433

References

- 1 Galbraith, C. G. & Galbraith, J. A. Super-resolution microscopy at a glance, *J. cell science*, **124**, 1607–1611, (2011).
- 2 Thorley, J. A., Pike, J. & Rappoport, J. Z. Chapter 14 - Super-resolution Microscopy: A Comparison of Commercially Available Options. In Cornea, A. & Conn, P. M. (eds.) *Fluorescence Microscopy*, 199–212, (Academic Press, Boston, 2014).
- 3 Hell, S. W. & Wichmann, J. Breaking the diffraction resolution limit by stimulated emission: stimulated-emission-depletion fluorescence microscopy. *Opt. Lett.* **19**, 780–782, (1994).
- 4 Gustafsson, M. G. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. *J. microscopy* **198**, 82–87, (2000).
- 5 Wegel, E. et al. Imaging cellular structures in super-resolution with SIM, STED and Localisation Microscopy: A practical comparison. *Sci. Reports* **6**, 27290, (2016).
- 6 Betzig, E. et al. Imaging intracellular fluorescent proteins at nanometer resolution. *Science* **313**, 1642–1645, (2006).
- 7 Bock, H. et al. Two-color far-field fluorescence nanoscopy based on photoswitchable emitters. *Appl. Phys. B* **88**, 161–165, (2007).
- 8 Oi, C., Mochrie, S. G. J., Horrocks, M. H. & Regan, L. Paint using proteins: A new brush for super-resolution artists. *Protein Sci.* **29**, 2142–2149, (2020).
- 9 Oi, C. et al. LIVE-PAINT allows super-resolution microscopy inside living cells using reversible peptide-protein interactions. *Commun. Biol.* **3**, 458, (2020).
- 10 Dertinger, T., Colyer, R., Iyer, G., Weiss, S. & Enderlein, J. Fast, background-free, 3d super-resolution optical fluctuation imaging (sofi). *Proc. Natl. Acad. Sci.* **106**, 22287–22292, (2009).
- 11 Gustafsson, N. et al. Fast live-cell conventional fluorophore nanoscopy with ImageJ through super-resolution radial fluctuations. *Nat. Commun.* **7**, 12471, (2016).
- 12 Agarwal, K. & Machán, R. Multiple signal classification algorithm for super-resolution fluorescence microscopy. *Nat. Commun.* **7**, 13752, (2016).
- 13 Yahiatene, I., Hennig, S., Müller, M. & Huser, T. Entropy-Based Super-Resolution Imaging (ESI): From Disorder to FineDetail. *ACS Photonics* **2**, 1049–1056, (2015).
- 14 Culley, S. et al. Quantitative mapping and minimization of super-resolution optical imaging artifacts. *Nat. Methods* **15**, 263–266, (2018).
- 15 Fukunaga, K. & Hostetler, L. The estimation of the gradient of a density function, with applications in pattern recognition. *IEEE Transactions on Inf. Theory* **21**, 32–40, (1975).
- 16 Yizong Cheng. Mean shift, mode seeking, and clustering. *IEEE Transactions on Pattern Analysis Mach. Intell.* **17**, 790–799, (1995).

- 17 Comaniciu, D. and Meer, P., Mean shift: a robust approach toward feature space analysis, IEEE Transactions on Pattern Analysis Mach. Intell. **24**, 603–619 (2002).
- 18 Emami, E., Fathy, M. and Kozegar, E. Online failure detection and correction for camshifttracking algorithm, 2013 8th Iranian Conference on Machine Vision and Image Processing (MVIP), (2013), pp. 180–183.
- 19 Fazekas, F. J., Shaw, T. R., Kim, S., Bogucki, R. A. & Veatch, S. L. A mean shift algorithm for drift correction in localization microscopy. Preprint at <https://doi.org/10.1101/2021.05.07.443176> (2021).
- 20 Rayleigh, L. On the theory of optical images, with special reference to the microscope. J. Royal Microsc. Soc. **23**, 474–482, (1903).
- 21 Sparrow, C. M. On Spectroscopic Resolving Power. Astrophys. J. **44**, 76, (1916).
- 22 Diaspro, A. & Bianchini, P. Optical nanoscopy. La Rivista del Nuovo Cimento **43**, 385–455, (2020).
- 23 Sharma, K. Optics: Principles and Applications (Academic Press, Elsevier Science, 2006).
- 24 Dataset: Tubulin 2d high density. <http://bigwww.epfl.ch/smlm/challenge2013/>.
- 25 Sage, D. et al. Super-resolution fight club: assessment of 2D and 3D single-molecule localization microscopy software. Nat. Methods **16**, 387–395, (2019).
- 26 Dataset: Bundled tubes high density. <http://bigwww.epfl.ch/smlm/challenge2013/>.
- 27 Schnitzbauer, J., Strauss, M. T., Schlichthaerle, T., Schueder, F. & Jungmann, R. Super-resolution microscopy with DNA-PAINT. Nat. Protoc. **12**, 1198–1228, (2017).
- 28 Wang, H. & Milstein, J. N. Simulation Assisted Analysis of the Intrinsic Stiffness for Short DNA Molecules Imaged with Scanning Atomic Force Microscopy. PLOS ONE **10**, 1–11, (2015).
- 29 Corbett, A. D. et al. Microscope calibration using laser written fluorescence. Opt. Express **26**, 21887–21899, (2018).
- 30 Garcés Suárez, Y. et al. Nanoscale organization of rotavirus replication machineries. elife **8**, e42906, (2019).
- 31 Romarowski, A. et al. Super-resolution imaging of live sperm reveals dynamic changes of the actin cytoskeleton during acrosomal exocytosis. J. Cell Sci. **131**, (2018).
- 32 Balestrini, P. A. et al. Seeing is believing: Current methods to observe sperm acrosomal exocytosis in real time. Mol. Reproduction Dev. **87**, 1188–1198, (2020).
- 33 Cox, S. et al. Bayesian localization microscopy reveals nanoscale podosome dynamics. Nat. Methods **9**, 195–200, (2012).
- 34 Jin, S., Haggie, P. M. & Verkman, A. S. Single particle tracking of membrane protein diffusion in a potential: Simulation, detection, and application to confined diffusion of cfr cl-channels. Biophys. J. **93**, 1079–1088, (2007).

506 35 Yamashita, N. et al. Three-dimensional tracking of plus-tips by lattice light-sheet
507 microscopy permits the quantification of microtubule growth trajectories within the
508 mitotic apparatus. *J. Biomed. Opt.* **20**, 1 – 18, (2015)

509 36 Chenouard, N. et al. Objective comparison of particle tracking methods. *Nat. Methods* **11**,
510 281–289, (2014).

511 37 Jaqaman, K., Loerke, D., Mettlen, M., Kuwata, H., Grinstein, S., Schmid, S. L. and
512 Danuser, G. Robust single-particle tracking in live-cell time-lapse sequences, *Nat.*
513 *Methods* **5**, 695 (2008).

514 38 Fields, A. P. and Cohen, A. E. Optimal tracking of a brownian particle, *Opt. Express* **20**,
515 22585–22601 (2012).

516 39 Kalman, R. E. A New Approach to Linear Filtering and Prediction Problems, *J. Basic*
517 *Eng.* **82**, 35–45 (1960).

518 40 Yüce M. Y., Erdoğan, A., Jonáš, A. & Kiraz, A. Single molecule tracking with kalman
519 filtering. *Frontiers in Optics 2011/Laser Science XXVII, FTuH5*, (Optical Society of
520 America, 2011).

521 41 Tinevez, J.-Y., Perry, N., Schindelin, J., Hoopes, G. M., Reynolds, G. D., Laplantine, E.,
522 Bednarek, S. Y., Shorte, S. L. and Eliceiri, K. W. Trackmate: An open and extensible
523 platform for single-particle tracking, *Methods* **115**, 80–90 (2017).

524 42 Patel, M., Leggett, S. E., Landauer, A. K., Wong, I. Y., and Franck, C. Rapid, topology-
525 based particle tracking for high-resolution measurements of large complex 3D motion
526 fields, *Sci. Reports* **8**, 5581 (2018).

527 43 Gross, J., Köster, M. & Krüger, A. Fast and efficient nearest neighbor search for particle
528 simulations. *Computer Graphics and Visual Computing CGVC*, 55–63, (The
529 Eurographics Association, 2019).

530 44 Cheezum, M. K., Walker, W. F., and Guilford, W. H. Quantitative Comparison of
531 Algorithms for Tracking Single Fluorescent Particles, *Biophys. J.* **81**, 2378–2388 (2001).

532 45 Calcines-Cruz, C., Finkelstein, I. J. & Hernandez-Garcia, A. Crispr-guided programmable
533 self-assembly of artificial virus-like nucleocapsids. *Nano Lett.* **21**, 2752–2757, (2021).

534 46 Federici, F., Dupuy, L., Laplaze, L., Heisler, M. & Haseloff, J. Integrated genetic and
535 computation methods for *in planta* cytometry. *Nat. Methods* **9**, 483–485, (2012).

536 47 Torres-Martínez, H. H., Hernández-Herrera, P., Corkidi, G. & Dubrovsky, J. G. From one
537 cell to many: Morphogenetic field of lateral root founder cells in *arabidopsis thaliana* is
538 built by gradual recruitment. *Proc. Natl. Acad. Sci.* **117**, 20943–20949, (2020).

539 48 Ricci, M. A., Manzo, C., García-Parajo, M. F., Lakadamyali, M., & Cosma, M. P. (2015).
540 Chromatin fibers are formed by heterogeneous groups of nucleosomes in vivo. *Cell*, **160**,
541 1145-1158.

- 49 Rutowicz, K., Lirski, M., Mermaz, B., Teano, G., Schubert, J., Mestiri, I., Kroten, M., A.,
Fabrice, T. N., Fritz, S., Grob, S., Ringli, C., Cherkezyan, L., Barneche, F., Jerzmanowski,
A., and Baroux, C. Linker histones are fine-scale chromatin architects modulating
developmental decisions in arabidopsis, *Genome Biol.* **20**, 157 (2019).
- 50 Wu, G., Zhao, X., Luo, S. & Shi, H. Histological image segmentation using fast mean
shift clustering method. *BioMedical Eng. OnLine* **14**, 24, (2015).
- 51 Szeliski, R. *Computer vision: Algorithms and applications* (Springer Science & Business
Media, 2011).
- 52 Sankaran, J. et al. Simultaneous spatiotemporal super-resolution and multi-parametric
fluorescence microscopy. *Nat. Commun.* **12**, 1748, (2021).
- 53 Lin, Y.-H., Chang, W.-L., and Hsieh, C.-L. Shot-noise limited localization of single 20
nm gold particles with nanometer spatial precision within microseconds, *Opt. Express* **22**,
9159–9170 (2014).
- 54 Glover, Z. J., Ersch, C., Andersen, U., Holmes, M. J., Povey, M. J., Brewer, J. R. and
Simonsen, A. C. Super-resolution microscopy and empirically validated autocorrelation
image analysis discriminates microstructures of dairy derived gels,” *Food Hydrocoll.* **90**,
62–71(2019).
- 55 Betzig, E. Nobel lecture: Single molecules, cells, and super-resolution optics, *Rev. Mod.*
Phys. **87**, 1153–1168 (2015).
- 56 Liss, V., Barlag, B., Nietschke, M. and Hensel, M. Self-labelling enzymes as universal
tags for fluorescence microscopy, super-resolution microscopy and electron microscopy,
Sci. Reports **5**, 17740 (2015).
- 57 de Boer, P., Hoogenboom, J. P. & Giepmans, B. N. G. Correlated light and electron
microscopy: ultrastructure lights up! *Nat. Methods* **12**, 503–513, (2015).
- 58 Cahoon, C. K. et al. Superresolution expansion microscopy reveals the three-dimensional
organization of the Drosophila synaptonemal complex. *Proc. Natl. Acad. Sci.* **114**, 6857–
6866, (2017).
- 59 Kim, J., Wojcik, M., Wang, Y., Moon, S., Zin, E.A., Marnani, N., Newman, Z.L.,
Flannery, J.G., Xu, K. and Zhang, X. Oblique-plane single-molecule localization
microscopy for tissues and small intact animals, *Nat. Methods* **16**, 853–857 (2019).
- 60 Hu, Y. S., Zimmerley, M., Li, Y., Watters, R. and Cang, H. Single-molecule super-
resolution light-sheet microscopy,” *ChemPhysChem* **15**, 577–586 (2014).
- 61 Chen, R., Tang, X., Shen, Z., Shen, Y., Li, T., Wang, J., Cui, B., Guo, Y., Du, S. and Yao,
S. Deep-learning super-resolution microscopy reveals nanometer-scale intracellular
dynamics at the millisecond temporal resolution, Preprint at
<https://doi.org/10.1101/2021.10.08.463746> (2021).

- 578 62 Puschmann, K. G. & Kneer, F. On super-resolution in astronomical imaging. *Astronomy*
579 & *Astrophysics* **436**, 373–378, (2005).
- 580 63 Guo, R., Shi, X., Zhu, Y. & Yu, T. Super-resolution reconstruction of astronomical
581 images using time-scale adaptive normalized convolution. *Chin. J. Aeronaut* **31**, 1752–
582 1763, (2018).

Figure 1. MSSR of zero order increases resolution by reducing the width of the spatial distribution of photons emitted from modelled fluorescent emitters. a) The MS is applied to the initial Gaussian distribution of photons emitted by a point-source (left) resulting in a MS graph (center). Application of further algebraic transformations

590 (see Supplementary Note 5 and Figure S7 (ii-iv)) provides the MSSR⁰ distribution
 591 (right). **b)** Sparrow and Rayleigh limits (blue, diffraction-limited) and the corresponding
 592 MSSR⁰ transformation (brown) for two point-sources. Red dots represent each emitter's
 593 location. The dip is indicated by a vertical black line. The inter-emitter distance is
 594 expressed as σ -times their individual standard deviation before MSSR processing. **c)**
 595 Dip computed for two point-source emitters of Gaussian distribution located away at
 596 distinct σ (blue line) where the corresponding MSSR⁰ result is also depicted (red line).
 597 For Gaussian: Rayleigh limit – gray discontinuous line, Sparrow limit - black
 598 discontinuous line. For MSSR⁰: Rayleigh limit - gray solid line, Sparrow limit - gray
 599 solid line. The solid vertical lines represent the distance between emitters such that
 600 when processed with MSSR⁰, the criteria of Rayleigh and Sparrow are obtained.

601

Figure 2. Single-frame MSSR analysis of higher order attains a resolution limit of 1.6σ for nearby emitters. **a)** The algorithm for computing higher-order MSSR (MSSR^n) is presented. The first iteration of MSSR (MSSR^1) is given by subtracting the MSSR^0 from the original image, resulting in a donut-like region centered at the emitter's location. MSSR^1 is computed after applying further algebraic transformations (see Supplementary Note 5 and Figure S8 (ii-iv) for a full description of the MSSR^n process). The second iteration encompasses the subtraction of MSSR^1 from MSSR^0 and the same algebraic transformations as used for generation of MSSR^1 . The process is repeated by updating consecutive MSSR images which generates higher MSSR orders. **b)** Theoretical limit of resolution achievable by MSSR^n . Dip computed for two Gaussian emitters in accordance with the variation of the inter-emitter distance (expressed as σ -times their standard deviation before MSSR processing). Colored lines represent the dip of MSSR order, from 0 to 3, computed at a given σ distance between emitters. Images on the right are the bidimensional representation of the MSSR^n processing for two single emitters separated at distances of 1.5σ and 1.6σ . Note that, for 1.5σ , emitters are unresolved up to the third order of MSSR. **c)** Experimental demonstration of the resolution increases attainable with higher order MSSR using the GATTA-SIM 140B nanoruler system. The intensity distribution of the emitter shrinks, both in σ and intensity, as the order of the MSSR increases (Figure S8). Nearby emitters (Alexa Fluor® 488) located 140 nm apart are resolved using MSSR^1 , MSSR^2 and MSSR^3 (right side). SIM images collected from the same sample (distinct fields) are shown as a positive control. Scale bar: 100 nm.

Figure 3. MSSR enhances resolution and contrast of single super-resolved images.

a) Comparison of SRM results of ESI, SRRF and MUSICAL alone and after post-processing with MSSR⁰ (ESI + MSSR⁰, SRRF + MSSR⁰, MUSICAL + MSSR⁰), over a temporal stack of 500 DL images of tubulin-labeled microtubules. The average projection of the DL stack is shown on the leftmost side. **b)** Comparison of the increase in spatial resolution of MUSICAL with and without post-processing with MSSR⁰ (MUSICAL + MSSR⁰), over a temporal stack of 361 DL images of modelled fluorophores bounded to a synthetic array of nanotubules (average projection shown on

636 left). The graphs show the intensity profiles along the lines depicted in each of the insets
637 in the images of the upper row; black, blue and red lines correspond to the average DL,
638 MUSICAL and MUSICAL + MSSR⁰ images, respectively. **c)** Sister chromatids of
639 mitotic mouse chromosomes visualized by TIRFM (left), SIM (middle) and SIM +
640 MSSR⁰ (right). Scale bars: **a)** 1 μ m, insets = 200 nm; **b)** 500 nm, insets = 100 nm; **c)**
641 200 nm.

642

Figure 4. The temporal analysis of MSSR provides a further increase in resolution to approximately 40 nm. **a)** Single-frame analysis of MSSR of a given order n is applied to each frame of a sequence, becoming the sf-MSSRⁿ stack. Next, a pixel-wise temporal function (PTF) converts the MSSR stack into a single super-resolved t-MSSRⁿ image. Depending on the temporal entropy of the dataset and on the PTF used, a resolution enhancement can be obtained. **b)** Left: a stack of DL images of a CRISPR/dCas12a nanoruler system. Scale bar: 1 μm . Right: zoomed region of the first frame in the stack, along with the average projection (DL-AVG) of a stack of 100 images, before and after MSSR processing. Scale bar: 400 nm. **c)** PTF applied to a stack of MSSR³ images (t-MSSR³). Fluorescent emitters are separated by 100 nm, as established by the CRISPR/dCas12a nanoruler system. Four types of PTF were computed: *TPM*, *Var*, *Mean* and *SOFI*₄. **d)** Euclidean distances between nearby emitters automatically computed from t-MSSR³-*Var* images, following a worm-like chain model

658 (16 regions of interest used, $1.5 \mu\text{m}^2$ each). **e)** Comparison of the results obtained with
659 each of the PTF analysis available with MSSR (see Table S3), for a commercially
660 available GATTA-PAINT nanoruler system. The *Var* column shows inter-emitter
661 distances resolved in the range 30 –75 nm. Atto 488 (green), Atto 550 (orange) and Atto
662 655 (magenta) fluorescent probes were used. **f)** Same nanorulers shown in **e)** but
663 analyzed with either ESI, SRRF or MUSICAL.

Figure 5. MSSR is robust to image noise and shows high global performance when compared to other SRM analytical procedures. a) sf-MSSR¹ and t-MSSR¹ of 100 images provide consistent reconstructions across a wide range of SNR. The expected feature is a uniform fluorescent ring located at the center of the image with a dark background lacking fluorescence. Each image is displayed to show its full intensity range. The row for DL images (Widefield, WF) exemplifies a stack of 100 frames collected at the corresponding SNR. The central row represents a resolved stack using sf-MSSR¹. The third row shows the super-resolved micrography after t-MSSR¹ analysis of 100 DL images using *TPM* for temporal analysis (see table S2). Scale bar: 1 μ m. **b-c)** Resolution Scaled Pearson (RSP) coefficient (b) and Resolution Scaled Error (RSE) (c), computed for the super-resolution reconstructions provided by SRRF, MUSICAL, ESI, sf-MSSR¹ and t-MSSR¹ (100 frames). **b)** RSP measures a global correlation between reconstruction and reference (input DL image), values closer to 1 indicate a reliable reconstruction. **c)** RSE measures the absolute difference of the reconstructed image and

679 its reference. Lower values of RSE at a particular SNR mean reduced global error in the
680 reconstruction. Scale bar: 1 μm .

681

682 **Data availability**

683 All raw imaging data which support the findings of this study are available from the
684 corresponding author upon request. Source data are provided with this paper.

685 Correspondence and requests for materials should be addressed to A.G.

686

687 **Code availability**

688 Source code for R and MATLAB platforms is available as supplementary materials, the
689 MSSR plugin for FIJI/ImageJ is available at <https://github.com/MSSRSupport/MSSR>.

690

691 **Funding**

692 This research was supported by Dirección General de Asuntos del Personal Académico
693 (DGAPA) – Programa de Apoyo a Proyectos de Investigación e Innovación
694 Tecnológica-UNAM (PAPIIT). and by Mexican Consejo Nacional de Ciencia y
695 Tecnología (CONACyT) grant A1-S-9236 to JGD, grant number IN211821 to AG,
696 IN211216 to CDW, IN204221 to JGD. Microscopy equipment was provided and
697 maintained through CONACyT grants 123007, 232708, 260541, 280487, 293624 and
698 294781. MB thank to IBYME-CONICET: Grants: PICT 2017-3047 (Agencia Nacional
699 de Promoción de la Investigación, el Desarrollo Tecnológico y la Innovación),
700 Fundación Williams; Fundación Rene Barón. DK acknowledges the support of the
701 National Science Foundation grant 2102832.

702 We also acknowledge the Programa de Becas de Posgrado of CONACyT for
703 granting scholarships to ETG, RPC, AL, DM, VA, EBA, CCC, GVG, DT, HTM, JLM,
704 HH, and to JPO. YG acknowledges DGAPA/UNAM for postdoctoral fellowship.

705

706 **Acknowledgements**

707 To the Microscopy Facility at EMBL Rome and IBYME-CONICET for providing
708 samples and materials for experiments. To the ECE and ICIMAF institutes for
709 theoretical and experimental input to this work. The authors thank the LNMA staff for

providing support with microscopy imaging. The OpenSPIM project was technically executed and extended by Oliver Valdez Escalona. ET and AG thank Angelica Flores Navarrete and Yuriney Abonza Amaro for their valuable feedback during the elaboration of the manuscript.

Author contributions

Conceptualization: ET, AG.
Methodology: ET, AG, RP, AHC, RR., JGD.
Investigation: ET, AG, RP, AHC, DM, VA, EBA, CC, GV, AHG, DT, JM, YG, MaB, MJ, HTM, HH, JO.
Visualization: ET, AG, HT.
Supervision: AG, AHC, DK.
Funding Acquisition: AG, AD, MB, JGD, CW.
Project Administration: AG.
Drafting Main Document: ET, AG, AL, AHC, CW.
Drafting Supplementary Material: ET, AG, AL, RR, JR, DK.
All authors contributed to the writing and reviewing of the manuscript.

Competing interests

The authors declare no competing interests.

Additional information

The online version contains supplementary methods, supplementary notes, supplementary movies, the MSSR manual for FIJI/ImageJ, MSSR scripts R and MATLAB, and additional references.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [movies1CRISPRPAINTnanorulersresolvedbyMSSR.mp4](#)
- [movies2LaserwrittenlithographypatternreconstructionatlowSNR5.4.avi](#)
- [movies3LaserwrittenlithographypatternreconstructionathighSNR21.3.avi](#)
- [movies4SuperresolutionMSSRheadofsingleframe.avi](#)
- [movies5SuperresolutionMSSRfenestrationofsingleframes.avi](#)
- [movies6SuperresolutionMSSRfenestrationof100framesatt0.avi](#)
- [movies7SuperresolutionMSSRfenestrationof100framesatt13.avi](#)
- [movies8SuperresolutionMSSRheadof100frames.avi](#)
- [movies9SuperresolutionMSSRfenestrationof100frames.avi](#)
- [movies103DreconstructionofMSSRsuperresolutionofslicesZstackofA.thalianarootnucleosome.avi](#)
- [MSSRSupplementaryInformation.pdf](#)
- [MSSRUserManual.pdf](#)